Direct PCR amplification of 16S rRNA genes offers accelerated bacterial

identification using the MinION[TM] nanopore sequencer

Shinichi Kai[1], Yoshiyuki Matsuo[2], So Nakagawa[3], Kirill Kryukov[3], Shino Matsukawa[1],

Hiromasa Tanaka[2], Teppei Iwai[2], Tadashi Imanishi[3], Kiichi Hirota[2]

1 Department of Anesthesia, Kyoto University Hospital, Kyoto, Japan

2 Department of Human Stress Response Science, Institute of Biomedical Science,

Kansai Medical University, Hirakata, Japan

3 Department of Molecular Life Science, Tokai University School of Medicine, Isehara,

Japan

Shinichi Kai and Yoshiyuki Matsuo equally contributed to this work.

Correspondence:

Yoshiyuki Matsuo and Kiichi Hirota

Department of Human Stress Response Science, Institute of Biomedical Science,

Kansai Medical University 2-5-1 Shin-machi, Hirakata, Osaka 573-1010, Japan

Tel: +81-72-804-2526

E-mail (YM): ysmatsuo-kyt@umin.ac.jp

E-mail (KH): hif1@mac.com

Keywords:

16S rRNA, bacteria, direct PCR, MinION, nanopore

Abbreviations:

CFU, colony forming unit; PCR, polymerase chain reaction; rRNA, ribosomal RNA

Abstract

Rapid identification of bacterial pathogens is crucial for appropriate and adequate antibiotic treatment, which significantly improves patient outcomes. 16S ribosomal RNA (rRNA) gene amplicon sequencing has proven to be a powerful strategy for diagnosing bacterial infections. We have recently established a sequencing method and bioinformatics pipeline for 16S rRNA gene analysis utilizing the Oxford Nanopore Technologies MinION[TM] sequencer. In combination with our taxonomy annotation analysis pipeline, the system enabled the molecular detection of bacterial DNA in a reasonable timeframe for diagnostic purposes. However, purification of bacterial DNA from specimens remains a rate-limiting step in the workflow. To further accelerate the process of sample preparation, we adopted a direct PCR strategy that amplifies 16S rRNA genes from bacterial cell suspensions without DNA purification. Our results indicated that differences in cell wall morphology significantly affected direct PCR efficiency and sequencing data. Notably, mechanical cell disruption preceding direct PCR was indispensable for obtaining an accurate representation of the specimen bacterial composition. Furthermore, 16S rRNA gene analysis of the mock polymicrobial samples revealed that optimization of the primer sequences is required to avoid preferential detection of particular taxa and to cover a broad range of bacterial species.

This study establishes a relatively simple workflow for rapid bacterial identification via MinION^{TM} sequencing that reduces the turnaround time from sample to result and provides a reliable method that does not require trained personnel or specialized techniques.

Introduction

Acute infectious diseases remain one of the major causes of life-threatening conditions

with high mortality, particularly in patients under intensive care. Therefore, rapid and

accurate identification of pathogenic bacteria facilitates the initiation of appropriate and

adequate antibiotic treatment [1, 2]. Although culture-based techniques are still the

forefront of clinical microbial detection, these methods are time-consuming and have

the critical drawback of not being applicable to non-cultivable bacteria [3].

As an alternative approach for overcoming the limitations of traditional culture-based

bacterial identification, metagenomic sequencing analysis has been introduced for the

diagnosis of bacterial infections [4]. Among the sequence-based microbiome studies,

the 16S ribosomal RNA (rRNA) genes have been the most predominantly used

molecular marker for bacterial classification [5]. The bacterial 16S rRNA gene is

approximately 1,500 bp long and contains both conserved and variable regions that

evolve at different rates. The slow evolution rates of the former regions enable the

design of universal PCR primers that amplify genes across different taxa, whereas fast-

evolving regions reflect differences between species and are useful for taxonomic

classification [6].

Targeted amplification of specific regions of the 16S rRNA gene followed by next

generation sequencing (NGS) is a powerful strategy for identifying bacteria in a given

sample. Despite the high-throughput capacity, second generation DNA sequencing

technologies provide relatively short read lengths with limited sequence information,

which often hampers accurate classification of the bacterial species [7]. A portable

sequencing device MinION^{TM} from Oxford Nanopore Technologies offers a number of

advantages over existing NGS platforms [8]. With the ability to generate longer read

lengths, MinION^{TM} analysis targets the whole coding region of the 16S rRNA gene,

showing great potential for rapid pathogen detection with more accuracy and sensitivity

[9-14]. We have previously established a sequencing method and bioinformatics

pipeline for rapid determination of bacterial composition based on 16S rRNA gene

amplicon sequencing via the MinION^{TM} platform [13]. A 5-minute data acquisition

using MinION^{TM} and sequence annotation against our in-house genome database

enabled the molecular detection of bacterial DNA in a reasonable timeframe for

diagnostic purposes.

In the current study, we attempted to further refine and update the protocols for 16S

rRNA gene sequencing analysis. We evaluated the performance of primer sets targeting

the near full-length 16S rRNA gene. To accelerate the process of sample preparation, we

adopted a direct polymerase chain reaction (PCR) strategy to amplify the 16S rRNA

gene from bacterial extracts without DNA purification.

Materials and Methods

Direct PCR amplification of 16S rRNA genes

The number of colony forming units (CFU) of bacteria (*Escherichia coli* and

Staphylococcus aureus) was determined by plating serial dilutions of cultures on agar

plates and counting colonies [15]. For mechanical cell disruption, zirconia beads (EZ-

Beads™; Promega, Madison, WI) were added to the bacterial cell suspensions and the

samples were vortexed for 30 s. The bacterial cell samples with or without mechanical

disruption were added directly to PCRs for amplifying the 16S rRNA genes. Bacterial

DNA was purified using DNeasy Blood & Tissue Kit (Qiagen, Hilden, Germany) and

used as a PCR template. PCR amplification of 16S rRNA genes was conducted using

the 16S Barcoding Kit (SQK-RAB204; Oxford Nanopore Technologies, Oxford, UK)

containing the 27F/1492R primer set [16, 17] and LongAmp™ Taq 2X Master Mix

(New England Biolabs, Ipswich, MA). Amplification was performed using an Applied

Biosystems Veriti™ Thermal Cycler (Thermo Fischer Scientific, Waltham, MA) with

the following PCR conditions, initial denaturation at 95 °C for 3 min, 25 cycles of 95 °C

for 20 s, 55 °C for 30 s, and 65 °C for 2 min, followed by a final extension at 65 °C for

5 min. To determine the effects of human DNA contamination on 16S rRNA gene

amplification, genomic DNA purified from the human monocytic cell line THP-1 was

mixed with *E. coli* DNA and subjected to PCR. To amplify human β-globin gene as an internal control for the human genome, the following primers were used, forward, 5′-GGTTGGCCAATCTACTCCCAGG-3′ and reverse, 5′-TGGTCTCCTTAAACCTGTCTTG-3′.

Genomic DNA from a mock bacterial community

MSA-1000™ 10 Strain Even Mix Genomic Material was obtained from the American Type Culture Collection (ATCC; Manassas, VA). The DNA mixture (1 ng) was used as a template for amplifying 16S rRNA genes. PCR amplification was conducted using the 16S Barcoding Kit and LongAmp™ Taq 2X Master Mix following the thermal cycling protocol as described above. Alternatively, 16S rRNA genes were amplified using KAPA2G™ Robust HotStart ReadyMix PCR Kit (Kapa Biosystems, Wilmington, MA). Amplification conditions for fast PCR using the KAPA2G™ polymerase were as follows, initial denaturation at 95 °C for 3 min, 25 cycles of 95 °C for 15 s, 55 °C for 15 s, 72 °C for 30 s, followed by a final extension at 72 °C for 1 min.

Whole cell mock bacterial community

MSA-3000™ 10 Strain Mix Whole Cell Material was obtained from ATCC.

Lyophilized bacterial cell pellets were suspended in phosphate-buffered saline (PBS) and divided into aliquots. The resulting cell suspensions were then either used for direct PCR to amplify the 16S rRNA genes (2.5×10^4 cells/reaction) or subjected to mechanical cell disruption via bead-beating prior to PCR amplification. Bacterial DNA purified from the cell suspension was also used for 16S rRNA amplicon sequencing.

Sequencing of 16S rRNA gene amplicons

PCR products were purified using AMPure XP (Beckman Coulter, Indianapolis, IN) and quantified by a NanoDrop (Thermo Fischer Scientific). A total of 100 ng DNA was used for library preparation and MinION^{TM} sequencing was performed using R9.4 flow cells (FLO-MIN106; Oxford Nanopore Technologies) according to manufacturer's instructions. MinKNOW software ver. 1.11.5 (Oxford Nanopore Technologies) was used for data acquisition.

Bioinformatics analysis

MinION^{TM} sequence reads (i.e., FAST5 data) were converted into FASTQ files by using Albacore software ver. 2.2.4 (Oxford Nanopore Technologies). Then, the FASTQ files were converted to FASTA files using our own program. In these reads, simple repetitive

sequences were masked using TanTan program ver. 13 with default parameters [18]. To

remove reads derived from humans, we searched each read against the human genome

(GRCh38) using minimap2 with default parameters [19]. Then, unmatched reads were

regarded as reads derived from bacteria. For each read, a minimap2 search with 5,850

representative bacterial genome sequences stored in the GenomeSync database

(http://genomesync.org) was performed. Next, we chose species showing the highest

minimap2 score as the existing species in a sample. Taxa were determined using our in-

house script based on the NCBI taxonomy database [20] and then visualized using

Krona Chart [21]. Sequence data from this article have been deposited with links to

BioProject accession number PRJDB7458 in the DDBJ BioProject database.

Results

Direct PCR approach for amplifying 16S rRNA genes from crude bacterial extracts

To overcome the time-consuming and laborious process of sample preparation for DNA sequencing, we tried to amplify the 16S rRNA gene directly from bacterial suspensions without a DNA purification step (Fig. 1). We used a commercially available kit (16S Barcoding Kit; Oxford Nanopore Technologies) with primers optimized for 16S rRNA amplicon sequencing on the MinION™ platform. The primers were designed to amplify the near full-length 16S rRNA gene for bacterial identification [16, 17]. Each indexing primer has a unique barcode for multiplexing and contains a tag sequence at the 5′-end for attachment of sequencing adapters.

Performance of the barcoded primers for 16S rRNA gene amplification was evaluated by PCR assays using bacterial cell suspensions. *E. coli* was chosen to represent gram-negative pathogens and *S. aureus* was used as a model for hard-to-lyse bacteria with gram-positive cell walls. A defined amount of each bacteria was serially diluted and the resulting cell suspensions were directly added to PCRs with LongAmp™ Taq DNA polymerase. As for *E. coli* suspensions, the lower limit of detection was less than 1×10^2 CFU in agarose gel electrophoresis (Fig. 2A). On the other hand, a higher number

of cells was required for detecting 16S rRNA gene amplicons from *S. aureus*

suspensions, whose detection limit was as low as 1×10^3 CFU (Fig. 2B).

To facilitate the release of bacterial DNA, cells in suspensions were disrupted by

vortexing with zirconia beads before being subjected to PCR amplification. Bead-

beating proved to be effective for the direct amplification of 16S rRNA genes from cell

suspensions of both *E. coli* and *S. aureus*, improving the yield of PCR products (Fig.

2C).

Rapid detection and identification of bacterial strains via direct PCR amplicon

sequencing on MinION™

Having demonstrated the efficacy of the barcoded primers for amplifying 16S rRNA

genes directly from bacterial suspensions, we investigated whether the direct PCR

method can impact MinION™ sequencing results and the accuracy of strain

identification. Bacterial cell suspensions of *E. coli* and *S. aureus* were used for

preparing 16S rRNA gene amplicon libraries and then the samples were sequenced on

MinION™ for 5 minutes (Table 1 and Fig. 3). In addition, sequencing libraries were

prepared using purified bacterial DNA templates as a standard reference for

comparison. Sequencing reads were analyzed using a bioinformatics pipeline based on a

BLAST search against our in-house genome database GenomeSync. MinION™

sequencing data identified the bacteria at the species level with an accuracy of more

than 90% for both *E. coli* and *S. aureus* (Figs 3A and 3B). *Shigella flexneri* was

additionally detected at a low abundance probably due to its high sequence similarity to

E. coli [22, 23]. Moreover, the type of PCR template (purified DNA versus cell

suspension) did not substantially affect the quality of sequence reads nor bacterial

identification results (Table 1). These results demonstrate the utility of the direct PCR

method, which can enable rapid pathogen identification from crude materials without

the need for DNA purification.

Impact of non-bacterial DNA contamination on 16S rRNA gene amplification

Successful identification of infectious pathogens should rely on the specific

amplification of bacterial target sequences in clinical samples, which can often be

contaminated with patient-derived human genetic materials. We tested whether a higher

amount of human DNA would affect the amplification of bacterial 16S rRNA genes. *E.*

coli DNA (0.1 ng) was mixed with increasing amounts of human DNA samples

extracted from the monocytic cell line THP-1, after which the mixture was subjected to

PCR amplification of the 16S rRNA gene (Fig. 4; upper panel). Primers targeting the β-

globin gene were used as internal control for the human genome (Fig. 4; lower panel).

Bacterial 16S rRNA genes were specifically amplified even in a background of high

human DNA concentrations, showing no inhibitory effects on PCR product yield. Thus,

the existence of non-bacterial genetic materials in the sample does not affect the

sensitivity and specificity of 16S rRNA gene detection.

16S rRNA gene sequencing of a mock bacterial community

The performance of the current tools for 16S rRNA gene amplicon sequencing was

further tested with a mixture of DNA prepared from 10 different bacterial species. The

relative abundance of individual bacterial taxa was estimated by genome size and copy

number of the 16S rRNA gene (Table 2). The mock community DNA mixture was used

as a template for PCR and the 16S rRNA gene amplicon libraries were sequenced on

MinION™. Nine out of 10 bacterial strains were successfully identified at the species

level (Fig. 5). Three minutes of run time generating 3,985 reads was sufficient for

identifying the nine species, whereas longer run times (5 minutes, 10,167 reads; 30

minutes, 44,248 reads) did not significantly affect species detection accuracy (Table 3).

There were some biases observed in the taxonomic profile; *Bacillus cereus* was detected

at lower abundances than expected, while *Clostridium beijerinckii* and *E. coli* were

overrepresented. These instances of partially-biased assignment or misidentification of

bacterial species were not resolved by increasing the number of sequencing reads

analyzed (Fig. 5). Our approach failed to identify *Bifidobacterium adolescentis* in the

mock community even when it was represented in the database.

We also evaluated the potential of another DNA polymerase and PCR amplification

protocol for bacterial species identification by MinION™ sequencing. KAPA2G™

Robust DNA Polymerase has a significantly faster extension rate than the standard wild-

type Taq, enabling shorter reaction times (approximately 100 minutes with LongAmp™

Taq versus 45 minutes with KAPA2G™) for amplifying 16S rRNA genes from the

mock bacterial community. The rapid amplification protocol with KAPA2G™ did not

impact the overall taxonomy assignment results of 16S rRNA gene sequence reads

generated by MinION™ (Fig. 5 and Table 3). Three-minute sequencing of KAPA2G™-

amplified 16S rRNA libraries identified nine bacterial species from the mock

community. *B. adolescentis* was not detected as was the case with LongAmp™ Taq

polymerase.

Evaluation of sample preparation methods for accurate bacterial identification via

MinION™ sequencing

Given the successful identification of a broad range of bacterial species from mixed

DNA samples, we further tested the utility of direct 16S rRNA gene amplification and

MinION™ sequencing on a mixture of whole bacterial cells with intact cell walls (Table

4 and Fig. 6). We assessed the effects of DNA extraction procedures on MinION™

sequencing results. Bacterial cell pellets comprising 10 different species were

suspended in PBS and divided into three aliquots. The first aliquot remained untreated

and the second was subjected to bead-beating for mechanical cell disruption. Bacterial

DNA purified from the third aliquot served as a reference for comparison. Regardless of

the extraction procedures, all bacterial species except for *B. adolescentis* were correctly

identified. However, species abundance differed across the three groups. The relative

abundance of *E. coli* was especially high in the 16S rRNA sequencing library amplified

directly from the untreated cell suspension and impaired sensitivity was found for the

detection of several types of bacteria (Fig. 6A). Mechanical disruption of bacteria by

bead-beating can improve results, as it showed patterns of bacterial composition that

were more similar to the reference group (Fig. 6B and 6C).

Discussion

Currently, identification of clinically relevant bacteria largely relies on culture-based

techniques. However, culture-dependent methods are time intensive and potentially lead

to delayed or even incorrect diagnoses [24]. Metagenomic sequencing analysis provides

an alternative approach for identifying bacterial pathogens in clinical specimens [5-7].

As previously reported, we developed a sequencing method and bioinformatics pipeline

for 16S rRNA gene amplicon sequencing and analysis utilizing the nanopore sequencer

MinION[TM] [13]. Although the system offers faster turnaround time than other NGS

platforms, purification of bacterial DNA from samples remains a rate-limiting step in

the workflow. To further facilitate the process of sample preparation, we attempted to

amplify 16S rRNA genes directly from bacterial cell suspensions without DNA

purification [25, 26], and achieved successful identification of bacterial species with a

total analysis time of less than 2 hours. The direct PCR approach revealed that

differences in cell wall morphology (gram status) significantly affected amplification

efficiency and sequencing results. For example, *S. aureus*, a gram-positive bacteria with

thick cell walls, was more resistant to heat lysis for DNA extraction and yielded less

PCR product compared with that of *E. coli*. A mechanical disruption method such as

bead-beating was useful in minimizing sample preparation bias. Indeed, samples

processed by bead-beating prior to PCR amplification exhibited a better representation

of the mock bacterial composition. The differential susceptibility to cell lysis among

bacterial species can affect 16S rRNA gene amplification and may introduce a bias in

the relative abundance of bacterial species in the community. Thus, mechanical cell

disruption preceding direct PCR amplification was indispensable for obtaining an

accurate representation of the sample bacterial composition.

We used new primer sets from Oxford Nanopore Technologies that are optimized for

rapid 16S rRNA gene sequencing on the MinION^{TM} platform. These universal primers

are designed to amplify the near full-length sequence of bacterial 16S rRNA genes. The

specificity and sensitivity of 16S rRNA gene amplification with these primers were not

substantially affected even when human DNA contaminants outweighed bacterial DNA.

Using these primer sets and the updated sequencing protocols, we performed a

metagenomic analysis of the pre-characterized bacterial community consisting of 10

different species. Although the universal primers are expected to bind to regions that are

highly conserved among bacterial species, we did not detect *B. adolescentis* in the mock

community. The 27F forward primer used in this study has three base pair mismatches

against *Bifidobacterium* (27F: AGAGTTTGATCMTGGCTCAG; priming site in *B.*

adolescentis: AG̲GGTTC̲GATT̲CTGGCTCA; mismatched bases are underlined) [27].

We speculate that these sequence mismatches lead to poor amplification of

Bifidobacterium 16S rRNA gene, resulting in the absence of these bacteria in the

sequence data. Consistent with our results, it has been reported that the 27F primer has a

bias toward underrepresentation of *Bifidobacterium* and other bacterial taxa in

microbiome analysis, which is caused by nucleotide variations even in the

phylogenetically highly conserved regions of 16S rRNA genes [27-29]. As shown here

and in previous publications, it should be noted that universal primers (e.g., 27F primer)

commonly used for metagenomic analyses have a limitation related to amplification

bias; thus, modifications of primer sequences are required to avoid preferential

detection of particular taxa and to cover a broad range of bacterial species [30].

In conclusion, direct amplification of 16S rRNA genes from crude bacterial extracts can

further accelerate sample processing for MinION™ sequencing. Direct 16S rRNA gene

amplification combined with MinION™ sequencing provides an attractive option for

accelerating pathogen detection. Further optimization and establishment of the

relatively simple workflow for rapid bacterial identification via MinION™ sequencing

would reduce the turnaround time from sample to result and provide a reliable method

that does not require trained personnel or specialized techniques.

Acknowledgements

This work was supported by grants from the Japanese Society of Anesthesiologists (JSA) Pitch Contest 2017 to SK, JSPS KAKENHI grant number JP16K10975 to YM, MEXT-Supported Program for the Strategic Research Foundation at Private Universities to SN, and Japan Agency for Medical Research and Development (AMED, JP17fm0108023) to TIm.

Author contributions

SK, YM, SN, KK, TIm, and KH designed the study. SK, YM, SM, HT, and TIw conducted the experiments. SK, YM, SN, and KH analyzed and interpreted the data. SK and YM wrote the manuscript.

References

1.   Shorr, AF, Micek, ST, Welch, EC, Doherty, JA, Reichley, RM & Kollef, MH (2011)

Inappropriate antibiotic therapy in Gram-negative sepsis increases hospital length of stay,

Crit Care Med **39**: 46-51

2.   Puskarich, MA, Trzeciak, S, Shapiro, NI, Arnold, RC, Horton, JM, Studnek, JR,

Kline, JA, Jones, AE & Emergency Medicine Shock Research, N (2011) Association

between timing of antibiotic administration and mortality from septic shock in patients

treated with a quantitative resuscitation protocol, *Crit Care Med* **39**: 2066-2071

3.   Fredricks, DN & Relman, DA (1996) Sequence-based identification of microbial

pathogens: a reconsideration of Koch's postulates, *Clin Microbiol Rev* **9**: 18-33

4.   Didelot, X, Bowden, R, Wilson, DJ, Peto, TEA & Crook, DW (2012) Transforming

clinical microbiology with bacterial genome sequencing, *Nat Rev Genet* **13**: 601-612

5.   Srinivasan, R, Karaoz, U, Volegova, M, MacKichan, J, Kato-Maeda, M, Miller, S,

Nadarajan, R, Brodie, EL & Lynch, SV (2015) Use of 16S rRNA gene for identification

of a broad range of clinically relevant bacterial pathogens, *PLoS One* **10**: e0117617

6.   Clarridge, JE, 3rd (2004) Impact of 16S rRNA gene sequence analysis for

identification of bacteria on clinical microbiology and infectious diseases, *Clin Microbiol*

Rev **17**: 840-862

7.   Kuczynski, J, Lauber, CL, Walters, WA, Parfrey, LW, Clemente, JC, Gevers, D & Knight, R (2011) Experimental and analytical tools for studying the human microbiome, *Nat Rev Genet* **13**: 47-58

8.   Leggett, RM & Clark, MD (2017) A world of opportunities with nanopore sequencing, *J Exp Bot* **68**: 5419-5429

9.   Sanderson, ND, Street, TL, Foster, D, Swann, J, Atkins, BL, Brent, AJ, McNally, MA, Oakley, S, Taylor, A, Peto, TEA, Crook, D & Eyre, DW (2017) Real-time analysis of nanopore-based metagenomic sequencing from orthopaedic device infection, *bioRxiv* doi: https://doi.org/10.1101/220616

10.   Leggett, RM, Alcon-Giner, C, Heavens, D, Caim, S, Brook, TC, Kujawska, M, Hoyles, L, Clarke, P, Hall, L & Clark, MD (2017) Rapid MinION metagenomic profiling of the preterm infant gut microbiota to aid in pathogen diagnostics, *bioRxiv* doi: https://doi.org/10.1101/180406

11.   Schmidt, K, Mwaigwisya, S, Crossman, LC, Doumith, M, Munroe, D, Pires, C, Khan, AM, Woodford, N, Saunders, NJ, Wain, J, O'Grady, J & Livermore, DM (2017) Identification of bacterial pathogens and antimicrobial resistance directly from clinical urines by nanopore-based metagenomic sequencing, *J Antimicrob Chemother* **72**: 104-114

12. Benitez-Paez, A, Portune, KJ & Sanz, Y (2016) Species-level resolution of 16S rRNA gene amplicons sequenced through the MinION portable nanopore sequencer, *GigaScience* **5**: 4

13. Mitsuhashi, S, Kryukov, K, Nakagawa, S, Takeuchi, JS, Shiraishi, Y, Asano, K & Imanishi, T (2017) A portable system for rapid bacterial composition analysis using a nanopore-based sequencer and laptop computer, *Sci Rep* **7**: 5657

14. Shin, H, Lee, E, Shin, J, Ko, SR, Oh, HS, Ahn, CY, Oh, HM, Cho, BK & Cho, S (2018) Elucidation of the bacterial communities associated with the harmful microalgae Alexandrium tamarense and Cochlodinium polykrikoides using nanopore sequencing, *Sci Rep* **8**: 5323

15. Miller, JH (1972) Determination of viable cell counts: bacterial growth curves in *Experiments in Molecular Genetics*, pp. 31-36, Cold Spring Harbor Laboratory Press, Cold Spring Harbor, NY.

16. Jiang, H, Dong, H, Zhang, G, Yu, B, Chapman, LR & Fields, MW (2006) Microbial diversity in water and sediment of Lake Chaka, an athalassohaline lake in northwestern China, *Appl Environ Microbiol* **72**: 3832-3845

17. Eden, PA, Schmidt, TM, Blakemore, RP & Pace, NR (1991) Phylogenetic analysis of Aquaspirillum magnetotacticum using polymerase chain reaction-amplified 16S

rRNA-specific DNA, *Int J Syst Bacteriol* **41**: 324-325

18.   Frith, MC (2011) A new repeat-masking method enables specific detection of homologous sequences, *Nucleic Acids Res* **39**: e23

19.   Li, H (2018) Minimap2: pairwise alignment for nucleotide sequences, *Bioinformatics* **34**: 3094-3100

20.   Federhen, S (2012) The NCBI Taxonomy database, *Nucleic Acids Res* **40**: D136-143

21.   Ondov, BD, Bergman, NH & Phillippy, AM (2011) Interactive metagenomic visualization in a Web browser, *BMC Bioinformatics* **12**: 385

22.   Fukushima, M, Kakinuma, K & Kawaguchi, R (2002) Phylogenetic analysis of Salmonella, Shigella, and Escherichia coli strains on the basis of the gyrB gene sequence, *J Clin Microbiol* **40**: 2779-2785

23.   Devanga Ragupathi, NK, Muthuirulandi Sethuvel, DP, Inbanathan, FY & Veeraraghavan, B (2018) Accurate differentiation of Escherichia coli and Shigella serogroups: challenges and strategies, *New Microbes New Infect* **21**: 58-62

24.   Clarridge, JE, 3rd, Raich, TJ, Sjosted, A, Sandstrom, G, Darouiche, RO, Shawar, RM, Georghiou, PR, Osting, C & Vo, L (1996) Characterization of two unusual clinically significant Francisella strains, *J Clin Microbiol* **34**: 1995-2000

25.   Videvall, E, Strandh, M, Engelbrecht, A, Cloete, S & Cornwallis, CK (2017) Direct PCR Offers a Fast and Reliable Alternative to Conventional DNA Isolation Methods for Gut Microbiomes, *mSystems* **2**: e00132-00117

26.   Flores, GE, Henley, JB & Fierer, N (2012) A direct PCR approach to accelerate analyses of human-associated microbial communities, *PLoS One* **7**: e44563

27.   Walker, AW, Martin, JC, Scott, P, Parkhill, J, Flint, HJ & Scott, KP (2015) 16S rRNA gene-based profiling of the human infant gut microbiota is strongly influenced by sample processing and PCR primer choice, *Microbiome* **3**: 26

28.   Frank, JA, Reich, CI, Sharma, S, Weisbaum, JS, Wilson, BA & Olsen, GJ (2008) Critical evaluation of two primers commonly used for amplification of bacterial 16S rRNA genes, *Appl Environ Microbiol* **74**: 2461-2470

29.   Farris, MH & Olson, JB (2007) Detection of Actinobacteria cultivated from environmental samples reveals bias in universal primers, *Lett Appl Microbiol* **45**: 376-381

30.   von Wintzingerode, F, Gobel, UB & Stackebrandt, E (1997) Determination of microbial diversity in environmental samples: pitfalls of PCR-based rRNA analysis, *FEMS Microbiol Rev* **21**: 213-229

Figure legends

Fig. 1. Workflow of 16S rRNA amplicon sequencing on a MinION™ platform and bioinformatics analyses.

Bacterial cells were left untreated or disrupted by bead-beating and then subjected to direct PCR for amplifying the near full-length 16S rRNA genes. Additionally, purified bacteria DNA was used for 16S rRNA gene amplification. The samples were sequenced on a MinION™ platform. The data obtained were analyzed using the computational analysis pipeline Genome Search Toolkit (GSTK) with GenomeSync database.

Fig. 2. Direct PCR amplification of 16S rRNA genes without DNA purification.

(A and B) A known amount of (A) *E. coli* and (B) *S. aureus* were added directly to PCRs (1×10^1–1×10^4 CFU/reaction) and then the amplified products were analyzed by agarose gel electrophoresis. M, molecular weight marker; DNA, 1 ng of bacterial DNA (positive control); -, no template (negative control); CFU, colony forming unit. (C) Bacterial cells were left untreated (-) or mechanically disrupted by bead-beating (+) and then the resulting cell extracts were subjected to 16S rRNA gene amplification.

Fig. 3. Accurate taxonomic assignment of MinION™ sequence reads amplified directly

from bacterial cell suspensions.

(A and B) 16S rRNA gene sequencing libraries were prepared using purified DNA

templates or amplified directly from cell suspensions. The samples were sequenced on a

MinION™ platform and taxonomic assignment was performed with the analytical

pipeline GSTK. The classification accuracy is shown for (A) *E. coli* and (B) *S. aureus*.

Fig. 4. Impact of background human DNA contamination on the specific amplification

of 16S rRNA genes.

E. coli DNA (0.1 ng) was mixed with varying amounts of human DNA samples and

subjected to 16S rRNA gene amplification. β-globin was used as an internal control for

the human genome.

Fig. 5. Taxonomic assignment of the mock community consisting of 10 bacterial

species.

A mixture of DNA from 10 different bacterial species was analyzed by 16S rRNA

amplicon sequencing using MinION™. PCR amplification was performed with

LongAmp™ Taq or KAPA2G™ polymerase. The samples were analyzed by MinION™

sequencing with different runtime conditions and the percentage of reads mapping to

the 10 bacterial species is shown. Expected abundance of individual taxa is based on the

genome size and copy number of 16S rRNA genes.

Fig. 6. Evaluation of sample processing methods for bacterial composition analysis.

(A–C) Whole cell mixtures of 10 bacterial species were (A; direct) left untreated or (B;

processed) disrupted by bead-beating for facilitating DNA release. (C; purified)

Bacterial DNA isolated from the mixed cell suspensions was also used. The samples

were subjected to 16S rRNA amplicon sequencing using MinION™ and the percentage

of reads mapping to the 10 bacterial species is shown.

Table 1 MinION sequencing of 16S rRNA gene amplicons.

Species	Sample	Total reads	Total base pairs	Mean read length	Mapped reads[a] (%)
E. coli	Direct	1692	2612624	1544.1	91.7
	Purified DNA	3470	5299622	1527.3	92.1
S. aureus	Direct	709	1072805	1513.1	98.2
	Purified DNA	3158	4717013	1493.7	96.2

[a] Proportion of reads mapping to E. coli or S. aureus.

Table 2 Mock community of 10 bacterial species.

Species	GenBank ID	Gram status	Genome size (Mb)	16S copies	Expected abundance (%)
Bacillus cereus	NC_003909.8	positive	5.42	12	11.16
Bifidobacterium adolescentis	NC_008618.1	positive	2.09	5	12.06
Clostridium beijerinckii	NC_009617.1	positive	6.49	14	10.88
Deinococcus radiodurans	NC_001263.1	negative	3.29	7	10.73
Enterococcus faecalis	NC_017316.1	positive	3.36	4	6.00
Escherichia coli	NC_000913.3	negative	4.64	7	7.61
Lactobacillus gasseri	NC_008530.1	positive	1.89	6	16.01
Rhodobacter sphaeroides	NZ_AKVW01000001.1	negative	4.60	3	3.29
Staphylococcus epidermidis	NC_004461.1	positive	2.56	5	9.85
Streptococcus mutans	NC_004350.2	positive	2.03	5	12.42

Table 3 MinION™ sequencing of a bacterial DNA mock community.

Polymerase	Run time (min)	Total reads	Total base pairs	Mean read length	Mapped reads[a] (%)
	3	3985	6133979	1539.3	97.2
LongAmp™ Taq	5	10167	15687563	1543.0	97.4
	30	44248	68330985	1544.3	97.2
KAPA2G™	3	4284	6287820	1467.4	96.0

[a] Proportion of reads mapping to the 10 bacterial species in the mock community.

Table 4 MinION™ sequencing of a whole cell mock bacterial community.

Sample	Total reads	Total base pairs	Mean read length	Mapped reads[a] (%)
Direct	3868	5622803	1453.7	93.5
Processed (bead-beating)	3608	5500056	1524.4	97.1
Purified DNA	3965	6035602	1522.2	96.9

[a] Proportion of reads mapping to the 10 bacterial species in the mock community.

Figure 1

Sample preparation   Bacterial cells

Disruption by bead-beating

DNA extraction and purification by spin column-based method

Extracts

DNA

Library preparation   Amplification of 16S rRNA genes
(16S Barcoding Kit SQK-RAB204)

PCR: LongAmp™ Taq / KAPA2G™ Robust

Bead clean-up

Adapter attachment

Sequencing   MinION™ sequencing

Loading: R9.4 Flow Cell [FLO-MIN106]

Data acquisition: MinKNOW

Data analysis   Read mapping and annotation using the GSTK software suite

Base calling

Extraction of non-human reads

Sequence search against the GenomeSync database

Identification of bacterial species

Figure 2

A

E. coli

B

S. aureus

C

Figure 3

Figure 4

Figure 5

Figure 6

A

Direct

B

Processed

C

Purified