A computational solution to improve biomarker reproducibility

during long-term projects

Feng Feng[1][¶], Morgan P. Thompson[2][¶], Beena E. Thomas[2][#], Elizabeth R. Duffy[3],

Jiyoun Kim[3], Shinichiro Kurosawa[3], Joseph Y. Tashjian[2,3], Yibing Wei[2,3],

Chris Andry[2,3], D.J. Stearns-Kurosawa[3][*]

[1] Department of Microbiology, Boston University School of Medicine, Boston,

 Massachusetts, United States of America

[2] Department of Pathology and Laboratory Medicine, Boston Medical Center, Boston,

 Massachusetts, United States of America

[3] Department of Pathology and Laboratory Medicine, Boston University School of

 Medicine, Boston, Massachusetts, United States of America

[¶]These authors should be considered equal first authors

*Corresponding author:

Email: dstearns@bu.edu (DSK)

[#]Current Address; Beth Israel Deaconess Medical Center Boston, MA

Abstract

Biomarkers are fundamental to basic and clinical research outcomes by reporting host responses and providing insight into disease pathophysiology. Measuring biomarkers with research-use ELISA kits is universal, yet lack of kit standardization and unexpected lot-to-lot variability presents analytic challenges for long-term projects. During an ongoing two-year project measuring plasma biomarkers in cancer patients, control concentrations for one biomarker (PF) decreased significantly after changes in ELISA kit lots. A comprehensive operations review pointed to standard curve shifts with the new kits, an analytic variable that jeopardized data already collected on hundreds of patient samples. After excluding other reasonable contributors to data variability, a computational solution was developed to provide a uniform platform for data analysis across multiple ELISA kit lots. The solution (*ELISAtools*) was developed within open-access R software in which variability between kits is treated as a batch effect. A defined best-fit Reference standard curve is modelled, a unique Shift factor "S" is calculated for every standard curve and data adjusted accordingly. The averaged S factors for PF ELISA kit lots #1-5 ranged from -0.086 to 0.735, and reduced control inter-assay variability from 62.4% to <9%, within quality control limits. S factors calculated for four other biomarkers provided a quantitative metric to monitor ELISAs over the 10 month study period for quality control purposes. Reproducible biomarker measurements are essential, particularly for long-term projects with valuable patient samples. Use of research-use ELISA kits is ubiquitous and judicious use of this computational solution maximizes biomarker reproducibility.

Introduction

In virtually every research project with real or potential clinical application, biomarkers provide valuable data to monitor presence or progression of disease, as well as therapeutic susceptibility or efficacy. Biomarker data monitor defined outcomes and there is considerable discussion about whether investigators should disclose incidental research findings to study participants [1-3]. Intrinsic to this discussion is the need for reproducible study data and this presents challenges, particularly with long-term studies. Data rigor and reproducibility is a systemic problem [4] and a priority issue for the NIH. It is an analytic challenge for long-term studies because research laboratories often do not have standardized operations nor validated biomarker assay reagents that adhere to the quality assurance and quality control standards for diagnostic use as required by the Clinical Laboratory Improvement Act (CLIA; www.cdc.gov/clia).

Protein biomarkers are measured frequently in plasma, serum or other matrices by solid phase Enzyme-Linked Immunosorbent Assay (ELISA) methods in which the antigen of interest in the sample is bound by antibodies and the amount of bound antigen is proportional to the signal strength that develops in the assay. There have been only 136 *in vitro* diagnostic ELISA kits or kit components cleared or approved by the FDA since 2000, but there are hundreds of commercially available ELISA kits labeled for "research purposes only" from dozens of vendors. Typically these are vetted by the manufacturer for sensitivity, selectivity, intra/inter-assay variability, stability and storage needs, but they are not required to adhere to federal CLIA guidelines. Lot-to-lot variability between ELISA kits is either not relevant or is manageable for short-term projects, but challenges arise for quality assurance when multiple lots of research ELISA kits are used in long-term

projects. A research laboratory may operate under NIH biomarker recommendations [5]

and Good Clinical Laboratory Practice guidelines with appropriate training, auditing,

assay validation and proficiency testing [6], but it does not have jurisdiction over kit

reagents controlled by the manufacturer. In general practice, expected inter-assay

coefficients of variation (CV) for ELISA standard curves will be within 10~20%, but if a

commercial ELISA standard curve suddenly shifts significantly with a new kit lot and

validated control data shifts outside limits, then the current patient biomarker

concentrations cannot be compared with those quantified months earlier. An experienced

laboratory can rescue data from one ELISA with a failed standard curve [7], but long-term

quality assurance poses other challenges. Either all the samples have to be re-assayed,

which introduces variables such as freeze/thaw issues, or the data discarded, all of which

wastes precious patient samples and resources.

This was encountered by our research group with one commercial ELISA kit during

an on-going two year project to evaluate effects of pre-analytic variables on plasma

thrombosis biomarkers in patients. The project developed thirty-five standard operating

procedures (SOPs) that define and document operations from blood acquisition to

transport, processing, assay and storage. Nine biomarkers are quantified by research-

use ELISA kits. For one biomarker, biomarker "PF", after months of reproducible assays,

we observed a significant shift of our standard curves and internal control results when

assayed with new kits lots. The scope of the problem was revealed by review of data

from 5 kit lots over 10 months and 65 ELISA plates. Rigorous review by the quality

assurance team did not identify laboratory or operational pre-analytic contributions and

similar changes were not observed with the other ELISAs. The project had encountered

an unexpected analytic variable and the data from 420 plasma samples collected over 10

months could not be compared. The manufacturer was responsive but ultimately unable to resolve the problem.

To rescue our patient data, we developed a computational solution with a sufficiently generalized approach such that it may be used by others facing a similar situation. In the solution, the lot-to-lot variability in ELISA kits is treated as a batch effect, and a defined Reference standard curve is modelled with either a four- or five-parameter logistic function. Based on this Reference curve, a Shift factor ("S") can be calculated and applied retrospectively to every standard curve from every ELISA plate over many months, and the biomarker concentrations for that plate are adjusted accordingly. In this way, the data collected from many ELISA plates over many months can be compared on a uniform platform. Once instituted, calculating the Shift factor for each ELISA standard curve or each kit lot provides an expedient way to rapidly monitor standard curves as a quality assurance metric and to facilitate data management.

Materials and methods

ELISA kits

Biomarker ELISAs with at least two kit lots were analyzed for the current study. ELISAs for human P-selectin/CD62P, human myeloperoxidase and human plasminogen activator inhibitor-1/serpin E1 were provided by R&D Systems (Minneapolis, MN, USA). The ELISA kit vendor for the biomarker of focus for the current study (biomarker "PF") is not provided for discretionary reasons. Five lots of the biomarker PF ELISA kits were received over a 10 month time period. All ELISA kits were a standard 96-well format, sandwich antibody-based ELISA designated "for research purposes".

Assays and Equipment

ELISA kits were stored at 4 ± 2^{o}C in a cooler (Helmer Scientific, Noblesville, IN, USA) equipped with alarmed wireless external temperature monitoring (SensoScientific, Inc., Simi Valley, CA). Temperature logs were reviewed and constant temperatures without drift were verified. All kits were used within the manufacturer's expiration date. Plasma samples stored at -80^{o}C were thawed just before assay in a 37^{o}C water bath for <10 minutes, gently mixed, and kept on ice. Plasma samples were assayed according to a detailed standard operating procedure (SOP) for each biomarker ELISA that includes the manufacturer's procedural steps. The SOP also included required documentation for every ELISA plate for operator, date, plate ID, critical reagents (date received, lot number, dilution/concentration, expiration date), incubation times (date, start/stop times, temperature), equipment (manufacturer, model, serial number) and a section to document any deviations from the SOP. Every plate included kit standards prepared according to the manufacturer's instructions. The standards were added to triplicate wells in the first three columns of the provided 96 well plate (columns A-C, rows 1-8 using standard plate designations).

For every biomarker and every plate, an internal spiked plasma-based control sample (BMC Control, see below) was included in triplicate wells. Samples were added to the plates in with calibrated pipettes, washing steps were performed with an automated plate washer (Biotek model Elx50, Winooski, VT) and developed color was quantified by measuring optical density (O.D.) at the appropriate wavelength with a microplate reader (VERSAmax; Molecular Devices, Sunnyvale, CA). The acceptable coefficient of variation (CV) of triplicate wells for each standard, control or unknown was ≤15%. For biomarker

141 PF, incubations with samples and detection antibody were done at 37°C in a dry incubator

142 per the manufacturer's instructions and perimeter wells were not used for patient samples

143 to prevent possible evaporation complications due to the elevated incubation

144 temperature. Other ELISAs were performed at room temperature. Initial data analyses

145 were done with SoftMax Pro version 7 (Molecular Devices).

BMC Control Preparation and Storage

Every ELISA plate included a human pooled plasma control sample that had been

spiked with supplemental biomarker and designated the BMC Control. For each

biomarker, the BMC Control was made in bulk volume according to the respective ELISA

SOP and stored at -80°C in small aliquots for single use. For biomarker PF, lyophilized

human pooled citrated plasma (Sigma Aldrich, catalogue P9523-5ML) was reconstituted

with deionized water at room temperature for at least 15 minutes with gentle mixing,

diluted with appropriate buffer to the same ratio as the patient samples and then spiked

with reconstituted biomarker PF standard prepared from the same manufacturer's kit but

purchased expressly for this purpose. On the assay day, a BMC Control aliquot was

thawed just before assay in a 37°C water bath for <10 minutes, mixed gently and added

to each plate in triplicate wells. Two preparations of BMC Controls were made and

aliquoted for storage: one in August 2017 (C1) and one in October 2017 (C2). Both

preparations were made with PF standard from kit lot #1. The mean O.D. ± S.D. for C1=

0.759 ± 0.095 (CV=13%, n=26 plates) and 0.672 ± 0.062 (CV=9%, n=16 plates) for C2.

Data Analysis and Derivation of Shift Factor "S"

The approach was implemented in the statistical package R, an open software

environment for statistical computing and graphics that accepts ELISA optical density

data and standard concentration data for calculation of a best-fit Reference standard

curve. The lot-to-lot variability is modelled as a fixed batch effect calculated as the

difference between each plate's standard curve and the Reference curve. This difference

is designated the Shift factor "S". An adjusted plate standard curve is derived using the

S factor and used to adjust the biomarker concentrations.

The Reference and standard curves are fitted with four- or five-parameter logistic

functions (4pl, 5pl).   These functions are well established models to relate analyte

concentrations to their response signal intensities in immunoassays [8-10]. The 5pl has

the form of:

175
$$Y = a + \frac{d-a}{\left(1 + e^{\frac{xmid-x}{scal}}\right)^g}$$
(1)

where Y is the signal intensity of measurement (OD in ELISA assays); x is the log-

transformed concentration of analytes; a and d are the lower and upper asymptotes of

signal intensity, respectively; $xmid$ is the x value of the curve's inflexion point; $scal$ is

the scale parameter or the inverse of the slope of the curve at the inflexion point ($x = xmid$

); g is the factor controlling the curve asymmetry. When g takes a value of 1, the 5pl

becomes the 4pl function.

The 5pl could also be written equivalently as a non-logarithm or exponential form,

183
$$Y = A + \frac{D-A}{\left(1 + \left(\frac{x'}{C}\right)^B\right)^g}$$
(2)

where Y and g are identical to the parameters in eq. (1); x' is non-log transformed

concentration and equal to e^x; A and D are identical to its equivalent lower-case letter

parameters in eq.(1); B and C are equivalent to $xmid$ and $scal$, but have an exponential

value of them. In this study, the log-form equation is used for implementation.

To analyze data from multiple plates or lots of ELISA kits, batch effects have to be

modelled and corrected for data comparison. Many biological or technical factors could

impact immunoassay reproducibility and lead to batch effects [11-15]. Lot-to-lot variability

between PF ELISA kits is among such factors. We proposed to model and correct it as a

fixed batch effect. It first assumes the lot-to-lot differences mainly result from the variable

quantitation of standard analyte concentrations, which can expressed as:

194
$$c = c_0 \times N$$
$$\log(c) = \log(c_0) + \log(N)$$
(3)

195 where c_0 and c are the real and provided standard analyte concentrations from the

196 manufacturer, respectively, and N is the fold difference between them. We can use x,

197 x_0 and S to replace $\log(c)$, $\log(c_0)$, and $\log(N)$ and rewrite the above log-transformed

198 equation as:

199
$$x_0 = x - S$$
(4)

200 Therefore the statistical model of the standard curves can be written as:

201
$$Y_{ijk} = a + \frac{d-a}{\left(1 + e^{\frac{xmid - x_{ij} + S_i}{scal}}\right)^g} + \varepsilon_{ijk}$$
$$i = 1, 2....n; j = 1, 2...m; k = 1, 2..l$$
(5)

where a, d, $xmid$, $scal$ and g are the parameters for the 5pl as in eq.(1); Y_{ijk} is the

observed signal intensity (measured OD in ELISA); x_{ij} is the log-transformed

concentration of the j^{th} standard analytes in the i^{th} batch; S_i, the Shifting factor, is the

log fold difference between the known concentration and the true one; i and n are the

i^{th} and total batch number, respectively; j and m are the j^{th} and total number of

standards, respectively; k and l are the k^{th} and total number of measurements; ε_{ijk} is

the random error for each measurements. This equation can be further rewritten into:

$$Y_{ijk} = a + \frac{d - a}{\left(1 + e^{\frac{xmid_i - x_{ij}}{scal}}\right)^g} + \varepsilon_{ijk} \tag{6}$$

where $xmid_i = xmid + S_i$ and all other parameters are the same as in eq.(5). As a result,

the model indicates that the standard curves of the same batch differ from each other as

a result of random errors of measurements, while differences between curves from

different batches is ascribed to inaccurate quantitation of analyte concentrations.

Furthermore, the standard curves from different batches follow the 5pl functions (or 4pl)

with the identical parameters of a, d, $scal$ and g, but different $xmid$. The differences are

defined by the Shift factor, S_i, which is estimated through the non-linear regression

together with other 5pl parameters.

To do the batch normalization/correction, the analyte concentrations in unknown

samples are first estimated based on unadjusted standard curves and then the Shift factor

S of the batch is applied to obtain the final quantities:

$$\left(x_{ij}\right)_{adj} = \hat{x}_{ij} - S_i \tag{7}$$

where x is the log-transformed analyte concentrations in the unknown sample, i and j

are the batch and standard number as in eq.(5), respectively.

It might not be possible to know the precise concentrations of analyte in the

standard samples (c_0 in eq.(3)) without other validation methods, such as proteomics [16].

Therefore, we designated one batch as the Reference batch, in which the analyte

concentrations of the standard samples were treated as accurate, and from that we

estimated the Shift factor S of all other batches relative to it. The data from 28 standard

curves from PF ELISA kit lot #1 (batch #1) was used to calculate a 4pl reference curve

for biomarker PF. For other biomarkers, at least four representative standard curves from

at least two lots and three operators were chosen to model the best-fit 4pl Reference

curve.

Statistical Analysis

236   Student's t-tests were performed for analysis of inter-assay differences. Analysis of

237   variance between groups was performed with Bonferroni post-test. P<0.05 was

238   considered significant.

Software Availability

241   An ELISA data analysis tool (*ELISAtools*) with the ability to correct batch effects

242   has been implemented in the statistical R software (version 3.5.1) and is available freely

243   for academic use at https://github.com/BULQI/ELISAtools. Instructions for calculation of

244   S factors is provided in Supplemental Methods.

Results

Procedurally, the ELISA manufacturer's directions were followed for reagents, buffers, assay temperature, reagent incubations and wash times. Operationally, these instructions were supplemented with documentation for each plate that included operator, kit reagent lot numbers and clock times for reagent additions and incubation periods. Including a BMC control sample on each plate permitted comparison of data over months (Fig 1). Control optical density (O.D.) readings were consistent with time even with different ELISA kit lots (Fig 1A,B). Myeloperoxidase data (lots #1,2) is shown as a comparison with biomarker PF (lots #1-5). Observations were similar for the other seven biomarkers (data not shown). Myeloperoxidase O.D.s were slightly higher for lot #2, but calculated antigen concentrations were stable (Fig 1C). In contrast, calculated PF concentrations in the BMC controls (preparations C1 and C2) decreased by an average 62.4% between ELISA kit lot #1 and lot #5 over the time (Fig 1D), exceeding our quality control limits. This disconnect between O.D. readings and calculated PF biomarker concentrations over time raised problems for the 420 patient samples already analyzed.

Fig 1. Biomarker Controls with Time

ELISAs were completed over ~10 months as described in Methods for biomarkers myeloperoxidase and PF. (A,B) Optical density (O.D. at 450nm) readings and biomarker concentrations calculated from each plate's standard curve (C,D) for the BMC internal control samples are shown *versus* time. Two ELISA kit lots were used for myeloperoxidase, and five kit lots for PF. Two BMC control preparations were used for PF (C1 and C2), and one BMC control preparation for myeloperoxidase. OD readings for PF controls are reasonably constant with time, but unlike myeloperoxidase, PF concentrations decreased by 62.4% over the study period.

Loss of PF antigenicity in the BMC controls during freezer storage could be a

contributor. BMC control preparation C1 was prepared in August 2017 and C2 in October

2017, and stored aliquots were used over the study period. The averaged optical density

values for C1 were higher than C2 (P<0.01; C1 O.D.= 0.759 ± 0.095, CV=9%;n=26 plates;

C2 O.D. = 0.672 ± 0.062, CV=13%; n=16 plates). However, O.D. values over time for

each preparation were reasonably stable (Fig 1B), suggesting PF antigenicity did not

change significantly during storage.   Despite similar O.D. values, a plate with control C1

had a calculated PF concentration of 723.9 pg/mL (mean, triplicate wells) in October 2017

with kit lot #1, but the same control calculated as 238.5 pg/mL in June 2018 with kit lot #5

(Table 1). A similar change was observed for BMC control preparation C2 samples.

Other than the kit itself, a detailed operations review did not identify significant changes

in PF antigenicity, environment, equipment or operator contributions that could explain

the large change in calculated PF concentrations observed in the BMC controls (data not

shown). Notably, no similar changes were observed with the other eight biomarker kits

(1-6 kit lots) over the same period.

Table 1. BMC Controls and ELISA Kit Lot number (#)

Assay Date	BMC Control C1	BMC Control C2	Kit Lot #	O.D. (450nm)[*]	PF Conc. (pg/mL)
10/11/2017	x		1	0.686	723.9
06/19/2018	x		5	0.638	238.5
02/19/2018		x	1	0.717	605.4
06/04/2018		x	5	0.684	225.1

[*] Mean O.D. ± S.D. is 0.759 ± 0.095 for C1 (CV=13%, n=26) and 0.672 ± 0.062 (CV=9%, n=16)

In contrast, comparison of PF standard curves from kit lot #1-5 showed a trend

with time (Fig 2A) that paralleled changes observed with the BMC control concentrations.

Expected variability between standard curves within the lot was observed (Fig 2B), but

lot-to-lot variability showed a left-shift trend. The averaged standard curves for lots #1

and #2 (September, November 2017) were similar, but lots #3-5 (April, May 2018) curves

had increasing O.D. at each standard concentration. The expected consequence of a left-

shift in standard curves will be lower biomarker PF concentrations, which was observed

(Fig 1, Table 1).

Fig 2. Lot-to-Lot Variability in PF ELISA.

(A) The averaged standard curves for each lot of PF ELISA as the mean optical density

O.D. ± S.D. at each standard concentration. N= 28, 19, 8, 4, 9 curves for lots #1-5,

respectively. (B) Variability of standard curve optical density at 450nm within and

between PF ELISA kit lots #1 and #5.

To address this analytic variable and rescue our patient data, a strategy was

developed such that each PF standard curve is compared to a best-fit Reference curve

(O.D. *versus* PF concentration) and a Shift factor ("S") calculated that quantifies the

difference between each plate's standard curve and the Reference curve. An adjusted

plate standard curve equation is derived and the biomarker concentrations are re-

calculated. Thus, every plate has a unique adjustment, based on the calculated S factor

for that plate, and results are generated on a uniform platform. A small S factor value

indicates the original standard curve on that plate is similar to the best-fit Reference curve.

Conversely, a large S factor indicates a larger left- or right-shift relative to the best-fit

Reference curve. The PF Reference curve was fitted with a 4-parameter logistics (4pl)

equation common to many ELISA analyses, but a 5-parameter logistics (5pl) curve fit

option is also available (Supplemental Methods).

Assignment of biomarker PF data to calculate the Reference curve was based on

our available data, the manufacturer's screening data, and judgment. We used data from

28 standard curves (PF kit lot #1) consistent over six months and three operators to

calculate the (4pl) Reference curve. Fig 3 shows the Reference curve with unadjusted

averaged standard curves for lots #2 and #5. Lot #2 curves had an average S factor of

0.0690 (n=19) indicating small differences from the Reference curve. In contrast, lot #5

curves had an average S factor of 0.6994 (n=9), indicating a substantial shift relative to

the Reference curve.

Fig 3. Reference Curve for calculation of S factor.

The best-fit, 4-parameter logistic (4pl) Reference curve (closed circle, R, solid line) from

twenty-eight PF ELISA kit lot #1 standard curves is shown as derived from the indicated

equation and D= -0.01; A= 3.20; C= 1300.00; B= -1.30. The averaged standard curves

for kit lot #2 has an S factor = 0.0690 (open square; n=19 curves) and S= 0.6994 (open

triangle; n=9) for curves from kit lot #5.

S factors for each PF standard curve were calculated (Fig 4). Data from lot #1 was

used to fit the Reference Curve, so the average S factor is close to 0 as expected. A shift

from the Reference curve for lots #3-5 is shown by their higher average S factors. One

lot #2 plate had an S = 0.7032, which is 20.8-fold higher than the average S factor from

the remaining plates (mean S=0.0338 ± 0.158; n=18). Data from this plate is under

review.   Average S factors for lots #3-5 are similar, but those from lot #3 have greater

inter-assay variability. The CV for lot #3 S factors is 44.2% (n=8), compared to 16.7%

(n=4) for lot #4 and 11.5% (n=9) for lot #5.   Whether this is due to manufacturer's

differences, sample size or laboratory-based variables is not known.

Fig 4. Calculated Shift Factors for PF ELISA Standard Curves.

A 4pl Reference curve was derived from PF kit lot #1 standard curves data and a Shift

factor "S" was calculated to quantify the difference between each plate's standard curve

and the best-fit reference curve. The S factors for each curve in lots #1-5 is shown with

the mean (horizontal line) for each lot. S factors for lots #1 and #2 curves are not different,

but those for lots #3-5 differ significantly from lot #1 (***P<0.0001, ANOVA with Bonferroni

post-test).

The BMC control PF concentrations for preparations C1 and C2 were re-calculated

using the S factor for each plate's standard curve. The reference curve was based on kit

lot #1 data and C1 was assayed primarily with kit lots #1 and #2. Thus, the average PF

concentration did not change significantly (P=0.843), but the variance was reduced (P=

0.002), as expected (Fig 5). C2 controls were included primarily on plates from kit lots

#3-5, so the difference after correction with S factors is significant (P<0.0001).

Figure 5. BMC Control PF Concentrations.

A BMC control sample from preparation C1 or C2 was included on each ELISA plate for

the 10 month study period and assayed with kit lots #1-5. C1 was used primarily on plates

from kit lots #1 and #2. C2 was used primarily on plates from kit lots #3-5.   Standard

curves from each plate were adjusted according to their calculated S factor and the control

PF concentrations were re-calculated. The data shows the PF concentrations before and

after correction with the S factors. ***P<0.0001; ns, not significant (Student's t test).

S factors for myeloperoxidase, soluble P-selectin and plasminogen activator

inhibitor-1 were calculated (Table 2). Data for 4pl Reference curves was chosen from

four standard curves performed by 3 different operators over at least 6 months to

represent the composite data and inter-assay variability. One myeloperoxidase curve

had an S = 1.099, and this data is under review. All BMC controls were within quality

control limits. Overall, their calculated S factors agree with consistent BMC Control

values over time and between kit lots.

Table 2. Average S Factors (min, max; N) for Biomarker Standard Curves from Kit Lots

Biomarker	Lot 1	Lot 2	Lot 3
Myeloperoxidase	0.041	0.505	0.289
	(-0.352, 0.347; 10)	(0.085, 1.10; 35)	(0.231, 0.382; 3)
Soluble P-selectin	-0.032	0.105	----
	(-0.294,0.156; 42)	(-0.153, 0.393; 4)	
Plasminogen activator inhibitor-1	-0.226	-0.056	-0.008
	(-0.220,0.107; 22)	(-0.373,0.047; 22)	(-0.043,0.039; 2)

Discussion

Acquiring ELISA data using commercial research kits over a prolonged period presents challenges for quality control needs. There is potential impact of multiple operators, antigen stability, environmental and equipment drift, lot-to-lot reagent variability, and lack of validated controls [15, 17-20]. It is virtually impossible to manufacture a new reagent lot that is identical in all respects to prior lots, and reagent variability occurs even with commercial diagnostic assay reagents [21]. Yet minimizing these influences is necessary so patient or other experimental data can be compared with confidence over the study period. In this study, changing BMC control concentrations for biomarker PF first raised the potential of excessive lot-to-lot variability in the ELISA kits. The optical density data did not change significantly, but calculated PF concentrations greatly decreased, and this disconnect triggered the comprehensive data review. After excluding other reasonable contributors, the S factor strategy was developed to provide a uniform platform for comparison of patient PF data collected over many months.

Calculation of S factors allowed retrospective adjustment of each plate's standard curve. The averaged S factors for PF kit lots #1-5 ranged from -0.086 to 0.735, and reduced the BMC control inter-assay variability to within our quality control limits (<9%). Review of the S factors for each standard curve was useful in that we could rapidly identify possible outliers. One curve each from PF ELISA lot #1 (high) and lot #3 (low), and one for myeloperoxidase (high) were sufficiently different and those plates were tagged for review. Unanticipated variability occurred even with SOPs, required documentation, equipment calibration and heightened operator awareness of pre-analytic variables. Controlling pre-analytic variables for quality assurance and method validation is difficult

[22] and identifying trends on a day-to-day basis is largely subjective. Once a Reference

curve is established, calculating the S factor for each curve provides rapid and

quantitative comparative data.

Few research-use ELISA kits include validated controls, so the early decision to

include spiked plasma controls on every plate was advantageous. One limitation of

laboratory-made controls is that each preparation will be slightly different and those

differences may not be apparent until enough plates have been run to set the quality

control ranges. Access to independently standardized and validated controls in a variety

of matrices (plasma, serum, urine, etc) for established and emerging biomarkers would

be a valuable resource for investigators and facilitate more consistent biomarker results

between laboratories. This in turn supports the goals of improved data rigor and

reproducibility, and the more ethical discussion regarding disclosure of validated data to

study participants [23-25].

The problem of inter-assay variability is not new [26] and various approaches are

proposed to quantify data that are acquired in batches using conversion of a signal from

known samples into a meaningful value for unknown samples [15, 27]. For *in vitro*

diagnostic ELISA kits used in a clinical laboratories, quality assurance is provided by a

validated control sample(s) with defined value limits. Research-only ELISA kits do not

have this foundation, yet are used universally. Calculation of S factors and retrospective

re-analysis with adjusted standard curves is useful when quality control values exceed

limits and environmental, procedural, equipment or operator contributions are ruled out.

This strategy should be used judiciously primarily because data choices for the

Reference curve is partially subjective. We chose data from PF ELISA lot #1 based on

volume of data, relative consistency over many months and similarity to the

manufacturer's data for that lot. We could have chosen standard curve data from lot #5

plates instead, but there were fewer curves over a shorter time period and our collective

experience judged this to be a less favorable choice. We do not anticipate using this

strategy for the other biomarkers. Their inter-assay ELISA variability is acceptable, their

BMC controls remain within limits, and there is no justifiable need to adjust the curves.

That said, we have found the S factors to be a useful monitor of ELISA outcomes to

rapidly identify pre-analytic factors, such as operator differences, that otherwise may be

difficult to identify by visual inspection of the data, particularly for long-term projects.

There are other approaches to adjust for batch effects in immunoassays similar to our

current work. More complicated statistical approaches are employed, such as the mix-

effect model [15] and the iterative maximum likelihood method [27]. They assume a linear

relationship between the measured signals and the analyte concentrations, which is only

an approximation and has its own limitations. Our implementation takes the form of a non-

linear logistic function. Our proposed statistical model is simple and the assumption is

appropriate for the observed lot-to-lot variability (Supplementary Figure 2).

We focused on developing an accessible and generalizable strategy to solve similar

issues as long as the assumptions are met. The software is implemented in R and the

data input uses a format familiar to those who use standard 96 well plates for ELISAs. It

is written for either Mac or PC and a choice of 4pl or 5pl curve fitting is provided. The

software is open source and in the public domain, with instructions for data input in the

Supplemental Methods.

Acknowledgements

The authors gratefully acknowledge Dr. Thomas Kepler (BUSM, Department of Mathematics and Biostatistics, Department of Microbiology) for strategic discussions and Dr. Daniel Remick (BUSM, Department of Pathology and Laboratory Medicine) for critical reading of the manuscript. We gratefully acknowledge the administrative assistance of Lindy Joseph (BUSM) and Dr. Jasmin Bavarva (Leidos Biomedical Research, Inc., Frederick, MD) who contributed to project set-up.

References

1.   Wong CA, Hernandez AF, Califf RM. Return of research results to study participants: Uncharted and untested. JAMA. 2018;320(5):435-6. doi: 10.1001/jama.2018.7898.

2.   Schmidt CO, Hegenscheid K, Erdmann P, Kohlmann T, Langanke M, Volzke H, et al. Psychosocial consequences and severity of disclosed incidental findings from whole-body MRI in a general population study. Eur Radiol. 2013;23(5):1343-51. doi: 10.1007/s00330-012-2723-8. PubMed PMID: 23239059.

3.   Shalowitz DI, Miller FG. Disclosing individual results of clinical research: Implications of respect for participants. JAMA. 2005;294(6):737-40. doi: 10.1001/jama.294.6.737.

4.   Prinz F, Schlange T, Asadullah K. Believe it or not: how much can we rely on published data on potential drug targets? Nat Rev Drug Discov. 2011;10(9):712. doi: 10.1038/nrd3439-c1. PubMed PMID: 21892149.

5.   Dancey JE, Dobbin KK, Groshen S, Jessup JM, Hruszkewycz AH, Koehler M, et

al. Guidelines for the Development and Incorporation of Biomarker Studies in Early

Clinical Trials of Novel Agents. Clinical Cancer Research. 2010;16(6):1745.

6.   Sarzotti-Kelsoe M, Cox J, Cleland N, Denny T, Hural J, Needham L, et al.

Evaluation and Recommendations on Good Clinical Laboratory Practice Guidelines for

Phase   I–III   Clinical   Trials.   PLoS   Medicine.   2009;6(5):e1000067.   doi:

10.1371/journal.pmed.1000067. PubMed PMID: PMC2670502.

7.   Natarajan S, Remick DG. ELISA rescue protocol: recovery of sample

concentrations from an assay with an unsuccessful standard curve. Methods.

2013;61(1):69-72. doi: 10.1016/j.ymeth.2012.08.013. PubMed PMID: 22982520; PubMed

Central PMCID: PMCPMC3604139.

8.   Gottschalk PG, Dunn JR. The five-parameter logistic: a characterization and

comparison with the four-parameter logistic. Anal Biochem. 2005;343(1):54-65. doi:

10.1016/j.ab.2005.04.035. PubMed PMID: 15953581.

9.   Higgins KM, Davidian M, Chew G, Burge H. The effect of serial dilution error on

calibration inference in immunoassay. Biometrics. 1998;54(1):19-32. PubMed PMID:

9544505.

10.   Feng F, Sales AP, Kepler TB. A Bayesian approach for estimating calibration

curves and unknown concentrations in immunoassays. Bioinformatics. 2011;27(5):707-

12. doi: 10.1093/bioinformatics/btq686. PubMed PMID: 21149344; PubMed Central

PMCID: PMCPMC3465100.

11.   Fichorova RN, Richardson-Harman N, Alfano M, Belec L, Carbonneil C, Chen S,

et al. Biological and technical variables affecting immunoassay recovery of cytokines from

human   serum   and   simulated   vaginal   fluid: a   multicenter   study.   Anal   Chem.

2008;80(12):4741-51. doi: 10.1021/ac702628q. PubMed PMID: 18484740; PubMed

Central PMCID: PMCPMC2646866.

12.   Schrijver RS, Kramps JA. Critical factors affecting the diagnostic reliability of

enzyme-linked immunosorbent assay formats. Rev Sci Tech. 1998;17(2):550-61.

PubMed PMID: 9713894.

13.   Tate J, Ward G. Interferences in immunoassay. Clin Biochem Rev.

2004;25(2):105-20.   PubMed   PMID: 18458713;   PubMed   Central   PMCID:

PMCPMC1904417.

14.   Wilson DH, Williams G, Herrmann R, Wiesner D, Brookhart P. Issues in

immunoassay standardization: the ARCHITECT Folate model for intermethod

harmonization. Clin Chem. 2005;51(4):684-7. doi: 10.1373/clinchem.2004.042358.

PubMed PMID: 15788785.

15.   Whitcomb BW, Perkins NJ, Albert PS, Schisterman EF. Treatment of batch in the

detection, calibration, and quantification of immunoassays in large-scale epidemiologic

studies. Epidemiology. 2010;21 Suppl 4:S44-50. doi: 10.1097/EDE.0b013e3181dceac2.

PubMed PMID: 21422966; PubMed Central PMCID: PMCPMC3073366.

16.   Hassis ME, Niles RK, Braten MN, Albertolle ME, Ewa Witkowska H, Hubel CA, et

al. Evaluating the effects of preanalytical variables on the stability of the human plasma

proteome. Anal Biochem. 2015;478:14-22. doi: 10.1016/j.ab.2015.03.003. PubMed

PMID: 25769420; PubMed Central PMCID: PMCPMC4492164.

17.   Algeciras-Schimnich A, Bruns DE, Boyd JC, Bryant SC, La Fortune KA, Grebe SK.

Failure of current laboratory protocols to detect lot-to-lot reagent differences: findings and

possible solutions. Clin Chem. 2013;59(8):1187-94. doi: 10.1373/clinchem.2013.205070.

PubMed PMID: 23592508.

18.   Potter DM, Butterfield LH, Divito SJ, Sander CA, Kirkwood JM. Pitfalls in

retrospective analyses of biomarkers: a case study with metastatic melanoma patients. J

Immunol Methods. 2012;376(1-2):108-12. doi: 10.1016/j.jim.2011.12.003. PubMed

PMID: 22210094; PubMed Central PMCID: PMCPMC3472364.

19.   Kim HS, Kang HJ, Whang DH, Lee SG, Park MJ, Park J-Y, et al. Analysis of

Reagent Lot-to-Lot Comparability Tests in Five Immunoassay Items. Annals of Clinical &

Laboratory Science. 2012;42(2):165-73.

20.   Teunissen CE, Willemse E. Cerebrospinal Fluid Biomarkers for Alzheimer's

Disease: Emergence of the Solution to an Important Unmet Need. EJIFCC.

2013;24(3):97-104. PubMed PMID: 27683444; PubMed Central PMCID:

PMCPMC4975183.

21.   Thaler MA, Iakoubov R, Bietenbeck A, Luppa PB. Clinically relevant lot-to-lot

reagent difference in a commercial immunoturbidimetric assay for glycated hemoglobin

A1c. Clin Biochem. 2015;48(16-17):1167-70. doi: 10.1016/j.clinbiochem.2015.07.018.

PubMed PMID: 26187005.

22.   Andreasson U, Perret-Liaudet A, van Waalwijk van Doorn LJC, Blennow K,

Chiasserini D, Engelborghs S, et al. A Practical Guide to Immunoassay Method

Validation. Frontiers in Neurology. 2015;6:179. doi: 10.3389/fneur.2015.00179. PubMed

PMID: PMC4541289.

23.   Broadstock M, Michie S, Marteau T. Psychological consequences of predictive

genetic testing: a systematic review. Eur J Hum Genet. 2000;8(10):731-8. doi:

10.1038/sj.ejhg.5200532. PubMed PMID: 11039571.

24.   Partridge AH, Wong JS, Knudsen K, Gelman R, Sampson E, Gadd M, et al.

Offering participants results of a clinical trial: sharing results of a negative study. Lancet.

2005;365(9463):963-4.   doi: 10.1016/S0140-6736(05)71085-0.   PubMed   PMID:

15766998.

25.   Gibson LM, Littlejohns TJ, Adamska L, Garratt S, Doherty N, Group UKBIW, et al.

Impact of detecting potentially serious incidental findings during multi-modal imaging.

Wellcome open research. 2018;2:114-41. doi: 10.12688/wellcomeopenres.13181.3.

PubMed PMID: 30009267.

26.   Liao JJ, Lewis JW. Qualifying ELISA data: combining information. J Biopharm Stat.

2000;10(4):545-58. doi: 10.1081/BIP-100101983. PubMed PMID: 11104392.

27.   Andrews SS, Rutherford S. A Method and On-Line Tool for Maximum Likelihood

Calibration of Immunoblots and Other Measurements That Are Quantified in Batches.

PLoS One. 2016;11(2):e0149575. doi: 10.1371/journal.pone.0149575. PubMed PMID:

26908370; PubMed Central PMCID: PMCPMC4764494.

Supporting Information

Supplemental Methods provides instructions on how to load and use ELISAtools for

calculation of S factor(s).

Figure 1

Figure 2

Figure 3

Figure 4

Figure 5