Semi-blind sparse affine spectral unmixing of autofluorescence-contaminated micrographs

Blair J. Rossetti[1,*], Steven A. Wilbert[2], Jessica L. Mark Welch[3], Gary G. Borisy[2], and James G. Nagy[4]

[1]Department of Computer Science, Emory University, Atlanta, GA, USA,
[2]Department of Microbiology, Forsyth Institute, Cambridge, MA, USA
[3]Josephine Bay Paul Center, Marine Biological Laboratory, Woods Hole, MA, USA
[4]Department of Mathematics, Emory University, Atlanta, GA, USA

Abstract—Spectral unmixing methods attempt to determine the concentrations of different fluorophores present at each pixel location in an image by analyzing a set of measured emission spectra. Unmixing algorithms have shown great promise for applications where samples contain many fluorescent labels; however, existing methods perform poorly when confronted with autofluorescence-contaminated images. We propose an unmixing algorithm designed to separate fluorophores with overlapping emission spectra from contamination by autofluorescence and background fluorescence. First, we formally define a generalization of the linear mixing model, called the affine mixture model (AMM), that specifically accounts for background fluorescence. Second, we use the AMM to derive an affine nonnegative matrix factorization method for estimating endmember spectra from reference images. Lastly, we propose a semi-blind sparse affine spectral unmixing (SSASU) algorithm that uses knowledge of the estimated endmembers to learn the autofluorescence and background fluorescence spectra on a per-image basis. When unmixing real-world spectral images contaminated by autofluorescence, SSASU was shown to have a similar reconstruction error but greatly improved proportion indeterminacy as compared to existing methods. The source code used for this paper was written in Julia and is available with the test data at https://github.com/brossetti/ssasu.

Index Terms—spectral, unmixing, microscopy

I. INTRODUCTION

Most conventional fluorescence microscopes use a series of mirrors and optical filters to separate the emission light of different fluorophores. The characteristics of these filters (i.e. what wavelengths of light they let pass) dictate what types of fluorophores and, more importantly, how many fluorophores can be used in an experiment. There exist a number of tools to help optimize the choice of fluorophores based on the configuration of a given microscope, such as the recently released SPEKcheck by Phillips et al. [2018]. These tools attempt to reveal the set of fluorophores that minimize spectral cross talk, a problem where the filter used for one fluorophore does not adequately exclude the fluorescent emission of other fluorophores (see Waters [2009] and references therein). Yet, even with such optimization, most experiments are still limited to three or four fluorescent labels. To make matters worse, many biological samples contain a variety of autofluorescent molecules that emit light in wavelengths that overlap and obscure the desired signal.

Spectral microscopy has become the method of choice when needing to avoid cross talk, mitigate autofluorescence, and simultaneously visualize many biological objects [Harmany et al., 2017, Jonkman et al., 2014, Levenson et al., 2008, Valm et al., 2011, 2016]. Instead of relying on filters, spectral microscopes use specialized optics and computational analysis to measure and separate the light emitted by different fluorophores. A spectral image is a three-dimensional data structure that holds a discrete spectral emission profile for each (x, y) pixel location. The range of detectable wavelengths and the number and width of the spectral bands varies by system and application. However, most commercial spectral microscopes are capable of measuring tens (multispectral) or even hundreds (hyperspectral) of spectral bands at resolutions of 1-15 nm/band [Cole et al., 2013, Gao and Smith, 2015]. For an in-depth treatment of spectral microscopy and its applications, we refer the reader to the "Spectral Imaging" special issue of *Cytometry Part A* [Lerner et al., 2006] and the reviews by Garini and Tauber [2013] and Li et al. [2013].

Since a spectral microscope simply records the spectrum at each pixel, the ability to separate fluorophores with overlapping emission spectra depends on the choice of spectral unmixing algorithm. Spectral unmixing refers to a group of techniques that attempt to determine how much of each fluorophore was present in some observed spectrum. Nearly all of these methods assume that light emitted by different fluorophores mix linearly, and are rightly called linear mixing models (LMM) [Zimmermann et al., 2014]. In the deterministic case, fluorophores are assumed to have one canonical emission spectrum called a *reference spectrum* or *endmember* (a term borrowed from mineralogy that refers to the purest form of an element that exists in a mixture). We can represent the set of N endmembers, each with M spectral bands, as the columns, denoted $\mathbf{s}_n$ for $n = 1, \dots, N$, of the endmember matrix $\mathbf{S} \in \mathbb{R}_+^{M \times N}$. If we denote the endmember concentrations, or weights, as $\mathbf{w} \in \mathbb{R}_+^N$, then we can write the LMM as

$$\mathbf{y} = \sum_{n=1}^{N} \mathbf{s}_n w_n + \boldsymbol{\varepsilon}, \tag{1}$$

where $\mathbf{y} \in \mathbb{R}_+^M$ is the observed spectrum, w_n is the n^{th}

*corresponding author: blair.rossetti@emory.edu

entry of $\mathbf{w}$, and ε is noise (typically assumed to be Gaussian or Poisson distributed). When considering an entire spectral image with P pixels, denoted $\mathbf{Y} \in \mathbb{R}_+^{M \times P}$, we can rewrite the LMM in matrix form as

$$\mathbf{Y} = \mathbf{SW} + \varepsilon, \qquad (2)$$

where $\mathbf{W} \in \mathbb{R}_+^{N \times P}$ is the set of weights for each endmember in each pixel. It is important to highlight that the entries of $\mathbf{W}$ must be nonnegative (i.e. in $\mathbb{R}_+$) since negative combinations of endmembers are not physically meaningful. The endmember weights are visualized by reorganizing the columns of $\mathbf{W}$ to produce a three-dimensional ($\sqrt{P} \times \sqrt{P} \times N$) unmixed image.

When the number and spectra of endmembers in an image is known in advance, spectral unmixing under the LMM is equivalent to the problem of Nonnegative Least Squares (NLS), and it can be written as

$$\mathbf{W}_{\mathrm{NLS}} = \arg\min_{\mathbf{W}} \left\| \mathbf{Y} - \mathbf{SW} \right\|_F^2, \text{ subject to } \mathbf{W} \succeq \mathbf{0}, \quad (3)$$

where $\|\cdot\|_F$ is the Frobenius norm, $\succeq$ is element-wise $\geq$, and $\mathbf{0}$ is the zero matrix. Conveniently, NLS is a convex optimization problem with a unique solution. Many unmixing algorithms that come packaged with commercial microscopes rely on some variant of the NLS active-set algorithm by Lawson and Hanson [1995]. While NLS is sufficient for some applications, it is not always possible to know which endmembers exist in an image. In particular, there are many different types of autofluorescent molecules, and knowing which autofluorescence endmembers are present is often impractical. In addition, endmembers are typically estimated from a reference sample. A poorly prepared reference sample or improper estimation method will lead to undesired unmixing results [Zimmermann et al., 2014]. As such, NLS lacks the flexibility to adequately handle many of the unmixing problems that arise in realistic applications.

Spectral unmixing methods have been used extensively by the remote sensing community for the analysis of hyperspectral geospatial data, and there exists a rich literature on advanced hyperspectral unmixing algorithms [Bioucas-Dias et al., 2012, Drumetz et al., 2016, Heylen et al., 2014, Keshava and Mustard, 2002, Keshava, 2003, Ma et al., 2014]. Although some methods from remote sensing have been directly applied to spectral micrographs [Harris, 2006, Lu and Fei, 2014], there are several key differences between the problem conditions that render unmixing algorithms for geospatial data unsuitable for spectral microscopy. As compared to remote sensing data,

1) spectral micrographs contain fewer spectral bands and, therefore, suffer less from the curse of dimensionality;
2) spectral micrographs have significantly higher spatial resolution relative to the size of the target objects, meaning that individual pixels contain fewer endmembers and neighboring pixels have similar compositions;
3) the number of fluorophores used for labeling is known *a priori*;
4) it is possible to estimate endmembers from reference images;

5) and spectral micrographs are often heavily contaminated by background fluorescence and autofluorescent organic molecules.

As a result of these differences, there is an existing and ongoing need to develop spectral unmixing methods specifically tailored towards spectral microscopy applications [Arena et al., 2017].

Several notable algorithms have been recently introduced with the aim of unmixing images contaminated by autofluorescence. Fereidouni et al. [2012, 2014] introduced a fast spectral phasor analysis algorithm (along with an ImageJ plugin) that utilized phasor analysis methods from lifetime imaging to unmixing endmembers in the presence of autofluorescence. However, phasor analysis is a geometric approach that is computationally impractical when unmixing with more than three endmembers. Megjhani et al. [2017] proposed a powerful morphologically constrained spectral unmixing (MCSU) algorithm using dictionary learning. In addition to learning the endmember spectra from reference images, MCSU builds a dictionary of morphological motifs unique to each fluorophore. Although MCSU shows impressive results for up to eight fluorophores, it requires that the reference images share the same morphologies found in the test images and that the morphologies differ between fluorophores. Perhaps the most popular open source unmixing tool is the PoissonNMF plugin for ImageJ by Neher et al. [2009]. PoissonNMF is based on the nonnegative matrix factorization (NMF) methods popularized by Lee and Seung [2001]. In the context of unmixing, NMF is a blind source separation algorithm that aims to learn the endmembers and their weights by solving

$$\min_{\mathbf{S},\mathbf{W}} \left\| \mathbf{Y} - \mathbf{SW} \right\|, \text{ subject to } \mathbf{S} \succeq \mathbf{0}, \mathbf{W} \succeq \mathbf{0}, \qquad (4)$$

for an appropriate distance metric. As its name suggests, PoissonNMF is intended for use with images affected primarily by Poisson distributed noise (i.e. shot-noise). The PoissonNMF ImageJ plugin can operate in both a blind and semi-blind manner where known endmembers are fixed when solving Eq. 4. However, Neher et al. [2009] report that PoissonNMF yields unsatisfactory results when unmixing more than four endmembers due to the heavily overlapping emission profiles. Similar semi-blind NMF approaches using the Gaussian noise model for read-noise-limited data have been suggested by Huang et al. [2015] and Tong et al. [2016]. These methods have been shown to work well on images with three to five endmembers; however, they both rely on sparsity constraints to help regularize the optimization problem. Although enforcing sparsity for the weight matrix is generally thought to be physically meaningful because endmembers are not expected to exist in all pixels (i.e. in background regions), the sparsity assumption does not hold if sources of background fluorescence are ignored.

Global background fluorescence originates from a variety of sources in the sample and optical path [Waters, 2009]. When neglecting this source of light, the weight matrix becomes dense and sparsity constraints are ineffective. Laurberg and Hansen [2007] introduced a sparse affine NMF method aimed at handling the offset components found in a variety of

data types. Woolfe et al. [2011] used a similar affine model (incorrectly called a linear model) to address autofluorescence in spectral images; however, this model did not impose constraints on sparsity.

In this paper, we formally define an affine mixing model (AMM) that generalizes the LMM by including a term to absorb any background fluorescence. From this model, we derive an affine nonnegative matrix factorization (ANMF) method for estimating endmembers from reference images. We then propose a semi-blind sparse affine spectral unmixing (SSASU) method for images contaminated with both autofluorescence and background fluorescence.

We test the proposed methods on the study of complex bacterial biofilms. As shown by Valm et al. [2011] and Mark Welch et al. [2016], multiplexed labeling and spectral microscopy can be used to explore the spatial relationships within biofilms. In these experiments, fluorophores conjugated to oligonucleotide probes were used to label and differentiate tens of bacterial taxons. Until now, the presence of autofluorescence and background has made separating these overlapping fluorophores difficult for certain biofilm samples. We show that our SSASU approach is able to successfully reduce the impact of autofluorescence and background and unmix seven fluorophores.

II. MATERIALS AND METHODS

A. Sample preparation

To evaluate the proposed method, a set of seven reference samples, ten test samples, and one no-probe control was prepared. The bacteria *Leptotrichia buccalis* was used as the biological target for generating reference samples for each of the seven fluorophores: DY-415, DY-490, ATTO 520, ATTO 550, Texas Red-X, ATTO 620, and ATTO 655 (Dyomics GmbH; ATTO-TEC GmbH; Thermo Fisher Scientific Inc.). *L. buccalis* cells were cultured, fixed, and then separately hybridized using custom fluorophore-conjugated oligonucleotide probes (biomers.net GmbH) as described by Mark Welch et al. [2016].

Test samples consisted of biofilms that were collected from the dorsum of the tongue, chemically fixed, and hybridized using a set of probes specific to different taxa of bacteria (see Table I). Two samples were taken from each of five human subjects (A-E). An additional biofilm sample was collected from subject D, chemically fixed, and hybridized without a fluorophore to generate a no-probe control sample (used to measure autofluorescence). After hybridization, the reference samples, test samples, and no-probe control were mounted on slides as described by Mark Welch et al. [2016].

B. Imaging and preprocessing

All spectral micrographs were acquired on a Zeiss LSM 880 using a 63×/1.4NA Plan-Apochromat objective. Point scanning was performed simultaneously for the 405 nm, 488 nm, 561 nm, and 633 nm lasers using the 405 and 488/561/633 dichroic mirrors. Spectral data was collected from a range of 410 nm to 696 nm using 8.6 nm steps. Each reference image was acquired using dimensions of 512×512 pixels at

TABLE I
OLIGONUCLEOTIDE PROBES AND THEIR TAXONOMIC TARGETS

Probe ID	Taxon	Target	Fluorophore
Smit651-DY415-2	Species	*Streptococcus mitis*	DY-415
Ssal372-DY490-2	Species	*Streptococcus salivarius*	DY-490
Prv392-AT520-2	Genus	*Prevotella*	ATTO 520
Vei488-AT550-1	Genus	*Veillonella*	ATTO 550
Act118-TRX-1	Genus	*Actinomyces*	Texas Red-X
Nei1030-AT620-2	Family	*Neisseriaceae*	ATTO 620
Rot491-AT655-2	Genus	*Rothia*	ATTO 655

0.263 μm/pixel resolution. Test images were were acquired at 2048×2048 pixels and down-sampled to 1024×1024 pixels at a resolution of 0.220 μm/pixel.

The use of dichroic mirrors blocked the detection of emitted light near the excitation wavelengths. Since these dark bands contained little information, they were removed from the reference and test images prior to any analysis. In total, 6 of the 32 spectral bands were removed.

C. Affine mixing model

The vast majority of spectral unmixing methods assume that the emitted light from different fluorophores combines according to the linear mixing model (LMM) described in Eq. 2. The LMM is so widely used that the phrase "linear unmixing" is regularly used interchangeable with "spectral unmixing." As microscopy data is predominately read-noise-limited [Lambert and Waters, 2014], the noise component is expected to follow a Gaussian distribution with zero mean. In reality, nearly all microscopy images are contaminated by background fluorescence that offsets the noise profile [Waters, 2009, Waters and Wittmann, 2014] and breaks the assumption of zero-centered noise. To explicitly accommodate for the presence of background fluorescence, we define an affine mixing model (AMM) as

$$\mathbf{Y} = \mathbf{SW} + \mathbf{b1}^T + \varepsilon, \tag{5}$$

where $\mathbf{b} \in \mathbb{R}^M_+$ is the nonnegative background spectrum and $\mathbf{1}$ is a vector of ones. Conveniently, the AMM becomes equivalent to the LMM when no background fluorescence exists (i.e. $\mathbf{b} = \mathbf{0}$).

D. Estimating endmembers

Spectral unmixing by NLS carries the assumption that the endmember spectra of both fluorophores and autofluorescence are known in advance. Our proposed unmixing method relaxes this assumption by only requiring that the endmember spectra of fluorophores be known. This claim is valid because one can control which fluorophores are used to label the sample, and endmember spectra can be estimated from reference images. But how to estimate endmembers from their corresponding reference images is not widely discussed in the literature. This open problem is important to emphasize because even slight inaccuracies in endmember estimates are known to dramatically effect the determination of the endmember weights—a problem called proportion indeterminacy [Zare and Ho, 2014].

We address this open question by proposing an affine non-negative matrix factorization (ANMF) method for estimating endmembers.

For the following discussion of endmember estimation, we will denote the set of reference images as $\mathbf{R}_1, \ldots, \mathbf{R}_N$ with $\mathbf{R}_n \in \mathbb{R}_+^{M \times Q}$, where Q is the number of pixels in the reference image. The corresponding endmembers, denoted $\mathbf{s}_n$ for $n = 1, \ldots, N$, make up the columns of the endmember matrix $\mathbf{S}$.

1) Mean estimated endmember: Spectral microscopy practitioners estimate endmembers from reference images using a variety of methods based on the arithmetic mean. In general, endmembers are determined by the average spectral signature over all foreground pixels or some user-defined region of interest within the reference image. Since the illumination conditions can vary between reference and test images, it is common to normalize the mean endmembers by their ℓ^∞- or ℓ^1-norm. We can write this in matrix form as

$$\bar{\mathbf{s}}_n = \frac{1}{\|\mathbf{h}_n\|_1} \mathbf{R}_n \mathbf{h}_n$$
$$\mathbf{s}_n = \frac{1}{\|\bar{\mathbf{s}}_n\|_\infty} \bar{\mathbf{s}}_n \text{ for } n = 1, \ldots, N, \tag{6}$$

where $\mathbf{h}_n \in \mathbb{R}^Q$ is a binary vector with 1 indicating the foreground and 0 indicating the background of the n^{th} reference image. Foreground/background thresholding can be performed using any number of different thresholding algorithms; we use the Triangle algorithm for its robustness to different illumination conditions [Zack et al., 1977].

2) ANMF estimated endmember: Using the arithmetic mean of foreground pixels assumes that the noise is Gaussian distributed with zero mean. Yet, even reference images contain some level of background fluorescence. We, therefore, define an affine model similar to Eq. 5 for reference images as

$$\mathbf{R}_n = \mathbf{s}_n \bar{\mathbf{w}}_n^T + \bar{\mathbf{b}}_n \mathbf{1}^T + \boldsymbol{\varepsilon}, \tag{7}$$

where $\bar{\mathbf{w}}_n \in \mathbb{R}_+^Q$ and $\bar{\mathbf{b}}_n \in \mathbb{R}_+^M$ are the weights and background spectrum for the n^{th} reference image. From Eq. 7, we can formulate an endmember estimation method based on ANMF. For each endmember spectrum $\mathbf{s}_n \in \mathbb{R}_+^M$ from $n = 1, \ldots, N$ we minimize the objective function

$$\min_{\mathbf{s}_n, \bar{\mathbf{w}}_n, \bar{\mathbf{b}}_n} \left\| \mathbf{R}_n - \mathbf{s}_n \bar{\mathbf{w}}_n^T + \bar{\mathbf{b}}_n \mathbf{1}^T \right\|_F^2,$$
$$\text{subject to } \mathbf{s}_n, \bar{\mathbf{w}}_n, \bar{\mathbf{b}}_n \succeq \mathbf{0}. \tag{8}$$

Since $\mathbf{s}_n$, $\bar{\mathbf{w}}_n$, and $\bar{\mathbf{b}}_n$ are all unknown, Eq. 8 is a nonconvex optimization problem. We also note that the solutions for $\mathbf{s}_n$ and $\bar{\mathbf{w}}_n$ are non-unique as they may be scaled by a nonzero constant (i.e. $\mathbf{s}_n \bar{\mathbf{w}}_n^T = (c\mathbf{s}_n)(c^{-1}\bar{\mathbf{w}}_n^T)$). To control the uniqueness of the solution, we normalize $\mathbf{s}_n$ to the ℓ^∞-norm at each iteration of the optimization scheme. This normalization is also required later to properly scale the unmixed image to the bit-depth of the observed image.

Since ANMF operates on the entire image, there is no need to threshold foreground from background or define a region of interest as with the Mean method. This is important to emphasize because the Mean estimated spectrum can change dramatically from one thresholding method to another.

E. Semi-blind sparse affine spectral unmixing

Although it is possible to accurately characterize endmembers for each fluorophore, estimating autofluorescence endmembers is difficult for several reasons. First, it may not always be possible to create a reference sample for each type of autofluorescence that occurs in the test samples. Second, the structure of the autofluorescent tissue may influence its spectrum (i.e. thicker samples may induce more light scattering). Therefore, we believe it is better to learn the K separate autofluorescence endmembers, denoted $\mathbf{S}_A \in \mathbb{R}_+^{M \times K}$, directly from the test images. This approach provides the flexibility necessary to adjust to varying numbers and types of autofluorescence spectra on an image-by-image basis.

Since we are confident in the estimation of our fluorophore endmembers, it would not make sense to impose the same flexibility on them. In fact, this flexibility can cause problems during unmixing. Although a sample can be labeled with a set of fluorophores, there is no guarantee that each fluorophore will exist in every field-of-view. In this rank-deficient case (i.e. $rank(\mathbf{Y}) < rank(\mathbf{S})$), the columns of $\mathbf{S}$ that are associated with the missing fluorophores will learn some other spectrum from the image that may not be physically meaningful. Therefore, we propose a semi-blind unmixing that separately considers the fluorophore endmembers, $\mathbf{S}$, and the autofluorescence endmembers, $\mathbf{S}_A$. To further reduce overfitting, we include a term to enforce sparsity of the weight matrix. We define our semi-blind sparse affine spectral unmixing (SSASU) method as

$$\min_{\mathbf{S}_A, \tilde{\mathbf{W}}, \mathbf{b}} \left\| \mathbf{Y} - [\mathbf{S}|\mathbf{S}_A]\tilde{\mathbf{W}} + \mathbf{b}\mathbf{1}^T \right\|_F^2 + \gamma \|\tilde{\mathbf{W}}\|_1,$$
$$\text{subject to } \mathbf{S}_A, \tilde{\mathbf{W}}, \mathbf{b} \succeq \mathbf{0} \tag{9}$$

where $\tilde{\mathbf{W}} \in \mathbb{R}_+^{(N+K) \times P}$ is the combined weight matrix for the fluorophore and autofluorescence endmembers, $\|\cdot\|_1$ is the sum over all matrix entries, and γ is the parameter controlling sparsity.

As with our ANMF method for estimating endmembers, SSASU is a nonconvex optimization problem. However, we note that SSASU is convex in $\mathbf{S}_A$ when $\tilde{\mathbf{W}}$ is fixed, and convex in $\tilde{\mathbf{W}}$ when $\mathbf{S}_A$ is fixed. Therefore, SSASU can be solved using a block coordinate descent algorithm similar to that suggested by Cichocki et al. [2009]. Algorithm 1 describes the multiplicative update method used to minimize Eq. 9. Note that element-wise multiplication and division are denoted by $\odot$ and $\oslash$, respectively. As with all multiplicative update algorithms, SSASU is sensitive to values becoming zero during the optimization. Therefore, we set all non-positive entries to some small value before each iteration. These small values were removed by thresholding after reaching a stopping condition.

III. RESULTS

We assessed the robustness of endmember estimation by ANMF and arithmetic mean across seven different reference images: DY-415, DY-490, ATTO 520, ATTO 550, Texas Red-X, ATTO 620, and ATTO 655. Each endmember estimate was compared against a ground truth endmember by measuring the spectral angle.

Algorithm 1: SSASU

Input : spectral data $\mathbf{Y} \in \mathbb{R}_+^{M \times P}$, initial endmembers $[\mathbf{S}|\mathbf{S}_A] = \tilde{\mathbf{S}} \in \mathbb{R}_+^{M \times (N+K)}$, indices of the autofluorescence endmembers ι, and sparsity parameter γ

Output: optimal solutions $\tilde{\mathbf{S}}^* \succeq \mathbf{0}$, $\tilde{\mathbf{W}}^* \succeq \mathbf{0}$, and $\mathbf{b}^* \succeq \mathbf{0}$

$\tilde{\mathbf{W}} \leftarrow$ randomly initialize;

repeat

   $\hat{\mathbf{Y}} \leftarrow \tilde{\mathbf{S}}\tilde{\mathbf{W}} + \mathbf{b}\mathbf{1}^T$;

   $\tilde{\mathbf{W}} \leftarrow \tilde{\mathbf{W}} \odot (\tilde{\mathbf{S}}^T\mathbf{Y}) \oslash (\tilde{\mathbf{S}}^T\hat{\mathbf{Y}} + \gamma)$;

   $\tilde{\mathbf{S}}(:,\iota) \leftarrow \tilde{\mathbf{S}}(:,\iota) \odot (\mathbf{Y}\tilde{\mathbf{W}}(\iota,:)) \oslash (\hat{\mathbf{Y}}\tilde{\mathbf{W}}(\iota,:))$;

   foreach $\mathbf{s}_i \in \tilde{\mathbf{S}}(:,\iota)$ **do**

      $\mathbf{s}_i \leftarrow \mathbf{s}_i/\|\mathbf{s}_i\|_\infty$;

   end

   $\mathbf{b} \leftarrow \mathbf{b} \odot (\mathbf{Y}\mathbf{1}) \oslash (\hat{\mathbf{Y}}\mathbf{1})$;

until *a stopping condition is met*;

Fig. 1. Comparison of seven Mean and ANMF estimated endmember spectra with fluorometer measurements. The shaded regions represent the fluorometer data, the dotted lines represent the Mean estimates, and the dashed lines represent the ANMF estimates. The gray vertical lines show the wavelength where dichroic mirrors blocked the measurement of emitted light (i.e. locations of missing spectral data).

TABLE II
SPECTRAL ANGLE BETWEEN ESTIMATED ENDMEMBERS AND THEIR CORRESPONDING FLUOROMETER SPECTRA

Method	DY-415	DY-490	ATTO 520	ATTO 550	TRX	ATTO 620	ATTO 655
Mean	0.484	0.324	0.455	0.088	0.065	0.203	0.053
ANMF	0.421	0.293	0.455	0.087	0.052	0.203	0.024

We evaluated our SSASU method by unmixing a set of ten real-world spectral images each of which were labeled with seven fluorophores (see Table I) and contaminated by background and autofluorescence. For comparison, we performed the same evaluation with the two most commonly used unmixing methods—NLS and PoissonNMF. The success of each method was measured by the Relative Reconstruction Error (RRE) and Proportion Indeterminacy (PI).

A. Endmember estimation performance

Since poorly estimated endmembers can degrade the overall performance of unmixing, it remains important to evaluate the accuracy of the estimates. Yet, defining a ground truth set of endmembers is difficult because all images will contain some level of background fluorescence. Instead, we compare the endmember estimates to fluorometer data reported in the literature [McNamara et al., 2006]. While the fluorometer data is measured under different optical and environmental conditions, these data still provide a useful baseline to check estimates made from reference images.

While neither the Mean nor the ANMF estimation method perfectly matched the fluorometer data due to differing environmental conditions and the spectral accuracy of the microscope [Cole et al., 2013], Figure 1 shows that ANMF endmembers are less contaminated by background fluorescence than Mean endmembers. As expected, fluorophores with higher-energy emission spectra are effected more by background generated from mounting media and other sources. We evaluated this quantitatively by calculating the spectral angle between each Mean and ANMF endmember and its corresponding fluorometer-measured spectrum. The spectral angle, which is related to cosine similarity, is calculated as $\theta(\mathbf{s},\hat{\mathbf{s}}) = \arccos\left(\frac{\mathbf{s}\cdot\hat{\mathbf{s}}}{\|\mathbf{s}\|_2\|\hat{\mathbf{s}}\|_2}\right)$, where $\mathbf{s}$ is the fluorometer spectrum and $\hat{\mathbf{s}}$ is the estimated spectrum. As reported in Table II, ANMF endmembers are as good or better at approximating the true endmember spectra than estimation by arithmetic mean.

B. Parameters

SSASU requires two parameters to be determined prior to unmixing: the sparsity parameter, γ, and the rank of the autofluorescence endmember matrix, $K = rank(\mathbf{S}_A)$. While these parameters can be tuned on a per-image basis, we kept the parameters constant across all test images to show the flexibility of the algorithm. The parameters were empirically chosen to be $\gamma = 0.009$ and $K = 1$. PoissonNMF also requires the user to set a sparsity parameter γ_P and the autofluorescence endmember rank K_P. These parameters were set empirically to $\gamma_P = 6$ and $K_P = 1$ for all test images. The fluorophore endmembers, $\mathbf{S}$, used for SSASU, NLS, and PoissonNMF were set to the values estimated by the ANMF method. The autofluorescence endmember for NLS was estimated by ANMF from a no-probe control sample (i.e. an unlabeled sample taken from the dorsum of the tongue).

C. Unmixing performance

The unmixing performance of SSASU, NLS, and Poisson-NMF was evaluated on the basis of two relative error metrics that ranged from zero (better) to one (worse). The ability to reconstruct the observed signal was determined by the relative reconstruction error (RRE), and was defined by

$$RRE(\mathbf{Y},\hat{\mathbf{Y}}) = \frac{\|\mathbf{Y} - \hat{\mathbf{Y}}\|_F}{\|\mathbf{Y}\|_F}, \qquad (10)$$

where $\mathbf{Y}$ was the observed image and $\hat{\mathbf{Y}}$ was the reconstructed image. It is worth noting that the RRE provides a measure of underfitting/overfitting and does not indicate the quality of a

Fig. 2. Comparison of unmixing performance for SSASU, NLS, and PoissonNMF across ten test images taken from five samples. The relative reconstruction error (top) evaluates each method's ability to reconstruct the observed spectra image. The proportion indeterminacy (bottom) measures the non-orthogonality of the weight matrices and illustrates how well each method separates the fluorophore endmembers in the presence of autofluorescence.

solution. For example, an algorithm can produce a physically meaningless solution to the unmixing problem with a near-zero RRE by fitting the noise in addition to the signal. Therefore, we use the RRE in conjunction with a metric that evaluates the ability of the unmixing algorithm to separate autofluorescence from the fluorophore endmembers.

Since the fluorophores used to label the test images were known to bind to distinct bacteria, each pixel in the image contains at most one fluorophore (with the rare exception of areas where different bacteria overlap). In this case, a poor solution for an observed spectrum will produce positive weights for more than one endmember. This is a type of overfitting known as proportion indeterminacy (PI). We measured PI by checking the non-orthogonality of the weight matrix. Since each pixel contains only one endmember, the property $\tilde{\mathbf{W}}\tilde{\mathbf{W}}^T = \mathbf{D}$ should hold, where $\mathbf{D} = diag(\|\tilde{\mathbf{w}}_1\|_2^2, \ldots, \|\tilde{\mathbf{w}}_N\|_2^2)$ and $\tilde{\mathbf{w}}_n$ is the n^{th} row of $\tilde{\mathbf{W}}$. From this property, we define a measure of PI as

$$PI(\tilde{\mathbf{W}}) = \frac{\|\mathbf{D} - \tilde{\mathbf{W}}\tilde{\mathbf{W}}^T\|_F}{\|\mathbf{D}\|_F}. \tag{11}$$

Together, RRE and PI indicate the fit and quality of each unmixing solution.

SSASU, NLS, and PoissonNMF were each able to effectively reconstruct the test images with RREs below 0.14 (see top of Figure 2). Across all test images, NLS had the

highest RRE. This was expected because NLS has the fewest free variables and is therefore more limited in its ability to fit the observed data. Although PoissonNMF had the lowest RRE in most cases, both PoissonNMF and SSASU performed similarly and neither exhibited signs of overfitting.

Despite having similar RRE across all test images, the quality of the solutions varied dramatically between the different methods. SSASU outperformed both NLS and PoissonNMF in all but one test case at reducing proportion indeterminacy (see bottom of Figure 2). For eight of the ten test images, Poisson-NMF was least able to cleanly separate the fluorophores. When comparing the worst PI for each method, SSASU showed a clear improvement over both NLS and PoissonNMF at 0.55, 0.89, and 0.90, respectively.

The performance illustrated by these metrics can be observed qualitatively in the unmixing results for test image E2 (see Figure 3). In the results for NLS (Figure 3 A-H), autofluorescence has contaminated nearly all of the unmixed channels. This same autofluorescence light was efficiently captured in the autofluorescence channel of the SSASU unmixed image (Figure 3 I). We also note that *Prevotella* (ATTO 520) was not present in image E2, yet the ATTO 520 channel of the NLS unmixed image contained a significant amount of light (Figure 3 D). The composite view of all unmixed fluorophore channels (i.e. all channels excluding autofluorescence) clearly illustrates the ability of SSASU to efficiently separate both the autofluorescence and the fluorophores (Figure 3 Q & R).

D. Characterizing autofluorescence

Samples can contain many different types of autofluorescent molecules, and exactly which autofluorescence endmembers exist in an image is difficult to ascertain in advance. For applications using NLS, it is common to estimate the autofluorescence endmember from a no-probe control. Figure 4 shows how the autofluorescence endmember estimated from the no-probe control image compare to those learned by SSASU. It is clear that the no-probe control endmember poorly characterized the autofluorescence encountered in the test images. The ability to learn the autofluorescence spectrum directly from an image allows SSASU to adapt to highly varied samples. This flexibility lets SSASU fit the observed data with a sparse set of endmember weight, thereby reducing proportion indeterminacy as compared to NLS.

IV. CONCLUSION

Spectral microscopy and unmixing make it possible to visualize biological samples labeled with a large set of fluorophores. However, the choice of unmixing algorithm is important for achieving the desired results. In this paper, we proposed and evaluated a semi-blind sparse affine spectral unmixing (SSASU) algorithm aimed at separating fluorescence endmembers in the presence of autofluorescence and background fluorescence. In all but one test case, SSASU was able to outperform both NLS and PoissonNMF in mitigating autofluorescence. While our method is more flexible than NLS, we note that, like other NMF methods, SSASU is more computationally expensive and does not guarantee a unique

Fig. 3. Montage of unmixed images for NLS (top) and SSASU (bottom). Panels A-P show the unmixed channels for autofluorescence (A, I); *S. mitis*/DY-415 (B, J); *S. salivarius*/DY-490 (C, K); *Prevotella*/ATTO 520 (D, L); *Veillonella*/ATTO 550 (E, M); *Actinomyces*/Texas Red-X (F, N); *Neisseriaceae*/ATTO 620 (G, O); and *Rothia*/ATTO 655 (H, P). A larger composite view of the non-autofluorescence unmixed channels is shown for NLS in panel Q and for SSASU in panel R. The scale bar in panel R indicates 10 μm.

Fig. 4. Comparison of the autofluorescence endmember estimated from the no-probe control reference image (gray region) to the autofluorescence endmembers learned by SSASU.

reference images. We showed that ANMF estimation is as good or better than the Mean method across all test cases. In addition, ANMF does not depend on a thresholding of foreground and background. This makes ANMF more robust to images with uneven illumination profiles.

There are several obvious extensions of this work. First, formulating a version of SSASU for tensors would allow for the unmixing of spectral images that use sequential excitation. Second, allowing minor adjustments to fluorophore endmember spectra on an image-by-image basis would allow the algorithm to accommodate endmember variability (i.e. changes in the endmember spectra as a result of the microenvironment).

ACKNOWLEDGEMENTS

The authors would like to thank Andrew Kempchinsky for his help with sample preparation; Dr. Lars Ruthotto, Dr. Yuanzhe Xi, Yunyi Hu, and Kelvin Kan for their thoughtful advice and guidance; and Louis Kerr for his microscopy support.

FUNDING

This material is based upon work supported by the National Science Foundation (NSF) Graduate Research Fellowship Program under Grant No. DGE-1444932, NSF Grant No. DMS-1522760, and the National Institutes of Health (NIH) National

solution. Therefore, we recommend SSASU for situations where spectral micrographs are contaminated with one or many sources of autofluorescence.

In addition, we described an affine nonnegative matrix factorization (ANMF) method for estimating endmembers from

Institute of Dental and Craniofacial Research (NIDCR) under Grant No. DE022586. Any opinions, findings, and conclusions or recommendations expressed in this material are those of the author and do not necessarily reflect the views of the NSF or the NIH.

REFERENCES

Ellen T Arena, Curtis T Rueden, Mark C Hiner, Shulei Wang, Ming Yuan, and Kevin W Eliceiri. Quantitating the cell: turning images into numbers with imagej. *Wiley Interdisciplinary Reviews: Developmental Biology*, 6(2):e260, 2017.

José M Bioucas-Dias, Antonio Plaza, Nicolas Dobigeon, Mario Parente, Qian Du, Paul Gader, and Jocelyn Chanussot. Hyperspectral unmixing overview: Geometrical, statistical, and sparse regression-based approaches. *IEEE journal of selected topics in applied earth observations and remote sensing*, 5(2):354–379, 2012.

Andrzej Cichocki, Rafal Zdunek, Anh Huy Phan, and Shunichi Amari. *Nonnegative matrix and tensor factorizations: applications to exploratory multi-way data analysis and blind source separation*. John Wiley & Sons, 2009.

Richard W Cole, Marc Thibault, Carol J Bayles, Brady Eason, Anne-Marie Girard, Tushare Jinadasa, Cynthia Opansky, Katherine Schulz, and Claire M Brown. International test results for objective lens quality, resolution, spectral accuracy and spectral separation for confocal laser scanning microscopes. *Microscopy and Microanalysis*, 19(6):1653–1668, 2013.

Lucas Drumetz, Jocelyn Chanussot, and Christian Jutten. Variability of the endmembers in spectral unmixing: recent advances. In *Hyperspectral Image and Signal Processing: Evolution in Remote Sensing (WHISPERS), 2016 8th Workshop on*, pages 1–5. IEEE, 2016.

Farzad Fereidouni, Arjen N Bader, and Hans C Gerritsen. Spectral phasor analysis allows rapid and reliable unmixing of fluorescence microscopy spectral images. *Optics express*, 20(12):12729–12741, 2012.

Farzad Fereidouni, Arjen N Bader, Anne Colonna, and Hans C Gerritsen. Phasor analysis of multiphoton spectral images distinguishes autofluorescence components of in vivo human skin. *Journal of biophotonics*, 7(8):589–596, 2014.

Liang Gao and R Theodore Smith. Optical hyperspectral imaging in microscopy and spectroscopy–a review of data acquisition. *Journal of biophotonics*, 8(6):441–456, 2015.

Yuval Garini and Elad Tauber. Spectral imaging: methods, design, and applications. In *Biomedical Optical Imaging Technologies*, pages 111–161. Springer, 2013.

Zachary T Harmany, Farzad Fereidouni, and Richard M Levenson. Spectral unmixing methods and tools for the detection and quantitation of collagen and other macromolecules in tissue specimens. In *Fibrosis*, pages 491–509. Springer, 2017.

A Thomas Harris. Spectral mapping tools from the earth sciences applied to spectral microscopy data. *Cytometry Part A*, 69(8):872–879, 2006.

Rob Heylen, Mario Parente, and Paul Gader. A review of nonlinear hyperspectral unmixing methods. *IEEE Journal of Selected Topics in Applied Earth Observations and Remote Sensing*, 7(6):1844–1868, 2014.

Shaosen Huang, Yong Zhao, and Binjie Qin. Two-hierarchical nonnegative matrix factorization distinguishing the fluorescent targets from autofluorescence for fluorescence imaging. *Biomedical engineering online*, 14(1):116, 2015.

James Jonkman, Claire M Brown, and Richard W Cole. Quantitative confocal microscopy: beyond a pretty picture. In *Methods in cell biology*, volume 123, pages 113–134. Elsevier, 2014.

Nirmal Keshava. A survey of spectral unmixing algorithms. *Lincoln laboratory journal*, 14(1):55–78, 2003.

Nirmal Keshava and John F Mustard. Spectral unmixing. *IEEE signal processing magazine*, 19(1):44–57, 2002.

Talley J Lambert and Jennifer C Waters. Assessing camera performance for quantitative microscopy. In *Methods in cell biology*, volume 123, pages 35–53. Elsevier, 2014.

Hans Laurberg and Lars Kai Hansen. On affine non-negative matrix factorization. In *Acoustics, Speech and Signal Processing, 2007. ICASSP 2007. IEEE International Conference on*, volume 2, pages II–653. IEEE, 2007.

Charles L Lawson and Richard J Hanson. *Solving least squares problems*, volume 15. Siam, 1995.

Daniel D Lee and H Sebastian Seung. Algorithms for non-negative matrix factorization. In *Advances in neural information processing systems*, pages 556–562, 2001.

Jeremy Lerner, Steve Lockett, and Robert Zucker. Spectral imaging. *Cytometry Part A*, 69A(8):711–946, 2006. special issue.

Richard M Levenson, Alessandro Fornari, and Massimo Loda. Multispectral imaging and pathology: seeing and doing more. *Expert Opinion on Medical Diagnostics*, 2(9):1067–1081, 2008.

Qingli Li, Xiaofu He, Yiting Wang, Hongying Liu, Dongrong Xu, and Fangmin Guo. Review of spectral imaging technology in biomedical engineering: achievements and challenges. *Journal of biomedical optics*, 18(10):100901, 2013.

Guolan Lu and Baowei Fei. Medical hyperspectral imaging: a review. *Journal of biomedical optics*, 19(1):010901, 2014.

Wing-Kin Ma, José M Bioucas-Dias, Tsung-Han Chan, Nicolas Gillis, Paul Gader, Antonio J Plaza, ArulMurugan Ambikapathi, and Chong-Yung Chi. A signal processing perspective on hyperspectral unmixing: Insights from remote sensing. *IEEE Signal Processing Magazine*, 31(1):67–81, 2014.

Jessica L Mark Welch, Blair J Rossetti, Christopher W Rieken, Floyd E Dewhirst, and Gary G Borisy. Biogeography of a human oral microbiome at the micron scale. *Proceedings of the National Academy of Sciences*, 113(6):E791–E800, 2016.

George McNamara, Amit Gupta, James Reynaert, Thomas D Coates, and Carl Boswell. Spectral imaging microscopy web sites and data. *Cytometry Part A: The Journal of the International Society for Analytical Cytology*, 69(8):863–871, 2006.

Murad Megjhani, Pedro Correa de Sampaio, Julienne Leigh Carstens, Raghu Kalluri, and Badrinath Roysam.

Morphologically constrained spectral unmixing by dictionary learning for multiplex fluorescence microscopy. *Bioinformatics*, 33(14):2182–2190, 2017.

Richard A Neher, Mišo Mitkovski, Frank Kirchhoff, Erwin Neher, Fabian J Theis, and André Zeug. Blind source separation techniques for the decomposition of multiply labeled fluorescence images. *Biophysical journal*, 96(9): 3791–3800, 2009.

Mick A Phillips, David M Susano Pinto, and Ian M Dobbie. Spekcheck-fluorescence microscopy spectral visualisation and optimisation: a web application, javascript library, and data resource. *Wellcome open research*, 3, 2018.

Lei Tong, Jun Zhou, Yuntao Qian, Xiao Bai, and Yongsheng Gao. Nonnegative-matrix-factorization-based hyperspectral unmixing with partially known endmembers. *IEEE Transactions on Geoscience and Remote Sensing*, 54(11):6531–6544, 2016.

Alex M Valm, Jessica L Mark Welch, Christopher W Rieken, Yuko Hasegawa, Mitchell L Sogin, Rudolf Oldenbourg, Floyd E Dewhirst, and Gary G Borisy. Systems-level analysis of microbial community organization through combinatorial labeling and spectral imaging. *Proceedings of the National Academy of Sciences*, 108(10):4152–4157, 2011.

Alex M Valm, Rudolf Oldenbourg, and Gary G Borisy. Multiplexed spectral imaging of 120 different fluorescent labels. *PloS one*, 11(7):e0158495, 2016.

Jennifer C Waters. Accuracy and precision in quantitative fluorescence microscopy. *Journal of Cell Biology*, 185(7): 1135–1148, 2009.

Jennifer C Waters and Torsten Wittmann. Concepts in quantitative fluorescence microscopy. In *Methods in cell biology*, volume 123, pages 1–18. Elsevier, 2014.

Franco Woolfe, Michael Gerdes, Musodiq Bello, Xiaodong Tao, and Ali Can. Autofluorescence removal by non-negative matrix factorization. *IEEE Transactions on image processing*, 20(4):1085–1093, 2011.

GW Zack, WE Rogers, and SA Latt. Automatic measurement of sister chromatid exchange frequency. *Journal of Histochemistry & Cytochemistry*, 25(7):741–753, 1977.

Alina Zare and KC Ho. Endmember variability in hyperspectral analysis: Addressing spectral variability during spectral unmixing. *IEEE Signal Processing Magazine*, 31(1):95–104, 2014.

Timo Zimmermann, Joanne Marrison, Karen Hogg, and Peter O'Toole. Clearing up the signal: spectral imaging and linear unmixing in fluorescence microscopy. In *Confocal microscopy*, pages 129–148. Springer, 2014.