Profiling embryonic stem cell differentiation by MALDI TOF mass spectrometry: development of a reproducible and robust sample preparation workflow

Rachel E. Heap[1], Anna Segarra Fas[2], Alasdair P. Blain[3], Greg M. Findlay[2] and Matthias Trost[1*]

[1]Institute for Cell and Molecular Biosciences, Newcastle University, Newcastle-upon-Tyne, UK; [2]MRC Protein Phosphorylation & Ubiquitylation Unit, University of Dundee, Dundee, Scotland, UK; [3]Wellcome Trust Centre for Mitochondrial Research, Newcastle University, Newcastle-upon-Tyne, UK;

*To whom correspondence should be addressed:

Matthias Trost (matthias.trost@ncl.ac.uk)

Mailing address:

M.T.: Newcastle University, Institute for Cell and Molecular Biosciences, Framlington Place, Newcastle-upon-Tyne, NE2 4HH, Phone: +44 191 2087009

Keywords:

MALDI TOF, mass spectrometry, stem cells, whole-cell profiling, phenotypic screen, drug discovery

Abstract

MALDI TOF mass spectrometry (MS) is widely used to characterize and biotype bacterial samples,

but a complimentary method for profiling of mammalian cells is still underdeveloped. Current

approaches vary dramatically in their sample preparation methods and are not suitable for high-

throughput studies. In this work, we present a universal workflow for mammalian cell MALDI TOF

MS analysis and apply it to distinguish ground-state naïve and differentiating mouse embryonic stem

cells (mESCs), which can be used as a model for drug discovery. We employed a systematic approach

testing many parameters to evaluate how efficiently and reproducibly each method extracted

unique mass features from four different human cell lines. This data enabled us to develop a unique

mammalian cell MALDI TOF workflow involving a freeze-thaw cycle, methanol fixing and CHCA

matrix to generate spectra that robustly phenotype different cell lines and are highly reproducible in

peak identification across replicate spectra. We applied our optimised workflow to distinguish naïve

and differentiating populations using multivariate analysis and reproducibly identifying unique

features. We were also able to demonstrate the compatibility of our optimised method for current

automated liquid handling technologies. Consequently, our MALDI TOF MS profiling method enables

identification of unique features and robust phenotyping of mESC differentiation in under 1 hour

from culture to analysis, which is significantly faster and cheaper when compared with conventional

methods such as qPCR. This method has the potential to be automated and can in the future be

applied to profile other cell types and expanded towards cellular MALDI TOF MS screening assays.

1   Introduction

Matrix-assisted laser desorption/ionization time of flight mass spectrometry (MALDI TOF MS) is an

versatile technique with many different applications ranging from protein identification by peptide

mass fingerprinting and small molecule analysis to imaging of tissues.[1–3] Although conventionally

considered a low-throughput technology, recent advances in MS and liquid handling technologies

and liquid handling tools has enabled MALDI TOF MS to emerge as a powerful tool for label-free

high-throughput screening (HTS) within both the pharmaceutical industry and academic sphere.[4–6]

This platform is already well established for *in vitro* assays to monitor post-translational

modifications such as ubiquitylation,[7,8] phosphorylation[9,10] and methylation,[11,12] as the read-out is

relatively simple with often just a single substrate and product. Similar to MALDI, laser desorption

ionization can also be combined with self-assembled monolayers (SAMs), also known as SAMDI.[13,14]

Here, substrates are first immobilised on a surface before treatment with an enzyme, thus

determining activity and kinetic parameters. Interestingly, SAMDI has been shown to be not only

compatible with peptide substrates for protein specificity,[15,16] but also carbohydrates and

glycosyltransferase activity.[17]

Whole cell analysis or cellular assays for evaluating compound efficacy affecting a cellular phenotype

presents an interesting challenge for MALDI TOF MS analysis as the system becomes inherently

more complex. A well-established application for whole cell MALDI TOF MS is the profiling of micro-

organisms, also known as biotyping.[18,19] Profiling of protein biomarkers specific to a bacterial

taxonomy by MALDI TOF MS was first performed by Claydon *et al* and enabled reproducible and

robust identification of gram-positive and gram-negative species.[20] Since then, bacterial genera have

been identified through various approaches from spectral mass fingerprinting, to more complex

approaches that involve comparing peaks identified in MALDI spectra to predictive masses from

proteomic and genomic data sets.[21,22] This in turn enabled the generation of reference protein

databases that list biomarkers specific to different bacterial species.[23] Combined with automated

spectral acquisition and novel algorithms to tackle data analysis, biotyping has become a powerful,

high-throughput tool for rapidly profiling bacterial genera in both academic and clinical settings.[24]

However, inter-lab studies revealed surprising discrepancies in *E. coli* fingerprints as experimental variables such as sample preparation and instrument parameters can affect spectral quality and reproducibility.[25,26] Several studies have therefore scrutinised sample preparation methods for bacterial biotyping, looking at solvent extraction or direct analysis, sample handling and also matrix choice affects spectra quality with the aim of developing a standardised method to enable universal identification of micro-organisms.[27–29]

While bacterial biotyping has been very successful and has become a standard tool in the clinic, profiling of mammalian cells by MALDI TOF MS has not yet reached this level. High-resolution MALDI-FT-ICR mass spectrometry has been used by Sweedler *et al* to characterise lipids within 30,000 individual rodent cerebellar cells.[30] This study enabled the identification of 520 lipid features and classification of neuron-like and astrocyte-like cells, thus allowing lipid profiles to be assigned to particular cellular functions. Characterising the protein signatures of mammalian cells by MALDI-MS is less common when compared with lipid analysis, however it has been used for phenotypic screening of human cancer cell lines,[31,32] identification of cells within a co-culture[33,34] or tissues[35] and detection of transient changes within a specific cell type, such as immune cells.[36–39] However, many of these studies list dramatically different experimental procedures with several being adapted from existing biotyping protocols. The huge range of experimental parameters could therefore be problematic for translation of published assays to the pharmaceutical industry.

To address the variation in experimental workflows we have systematically tested different methods at key steps in preparing mammalian cells for whole cell MALDI TOF MS analysis. We have generated a robust and sensitive sample preparation workflow by studying four commonly used human cell lines, followed by application of our final method to a pharmacologically controlled biological system, where we applied our optimised method to profile differences between naïve ground-state mouse embryonic stem cells (mESCs) and those undergoing differentiation. Thus, we have

established a sample preparation method that is highly reproducible, robust and sensitive with

respect to both biological and experimental variances and would be suitable for expansion to a HTS

platform.

Materials and Methods

Human cell line culture. HEK293 and U2OS cell lines were cultured in DMEM media supplemented

with 10% FBS, 1% pen/strep and 1% L-glutamine. MCF7 and THP-1 cells were cultured in RPMI-1640

media supplemented with the same and 50 µM β-mercaptoethanol was added to THP-1 cells.

Adherent cell lines were lifted from 10 cm culture plates by addition of trypsin-EDTA solution. All cell

lines were incubated in a controlled atmosphere at 5% CO_2 and 37°C. Cells were harvested and

centrifuged at 300 xg for 3 minutes before resuspension in PBS and counted using a

haemocytometer. Cells were then aliquoted at a concentration 1×10^6 into 1.5 mL microtubes and

centrifuged at 300 xg, 4 °C for 10 minutes.

Mouse Embryonic Stem Cell (mESC) Culture. CGR8 mESCs were cultured in 0.1% gelatin [w/v]

coated plates in N2B27 medium (DMEM/F12-Neurobasal (1:1), 0.5% N2, 1% B27 (ThermoFisher

Scientific), 1% L-glutamine, 100 µM β-mercaptoethanol) containing "2i",[40] 3 µM CHIR99021 (Axon

Medchem) and 1 µM PD0325901, in a controlled atmosphere at 5% CO_2 and 37°C. To induce multi-

lineage differentiation,[41] cells were plated at 4×10^4 cells/cm^2 in N2B27 medium without CHIR99021

and PD0325901 and incubated for 48h at 5% CO_2 and 37°C.

RNA extraction and qPCR. Total RNA extraction was performed by a column-based system

(Omega) and then subjected to reverse transcription using iScript reverse transcriptase (Bio-Rad)

according to the manufacturer's guidelines. qPCR reactions were carried out using SYBR® Premix Ex

Taq™ II Supermix (Takara) in a CFX384 real-time PCR system (Bio-Rad). Samples were analysed for

gene expression in 2i release conditions relative to 2i medium culture using the ΔΔCt method, and

GAPDH expression was analysed as a loading control. Data from three independent biological

replicates, with two technical replicates for each, were analysed in Excel software (Microsoft) and

plotted in GraphPad Prism v.6.00 software (GraphPad). Primers used are listed in Table S-1.

Statistical significance was determined using an unpaired Student's t test, and significant differences

were considered when $p < 0.05$.

Cell microscopy and diameter analysis. The four cell lines were measured for number and cell

diameter by light microscopy using an Evos XL Core Cell Imaging System (Invitrogen). Optimal cell

numbers were calculated by a cell titration, whose values are reported in Table S-2, and these

concentrations were used for subsequent experiments. To assess permeability, cell pellets were

resuspended in PBS before mixing 1:1 with trypan blue. Trypan blue positive cells were then

automatically counted using the same microscope to calculate cell viability. For mESC phenotype

visualization, brightfield light microscopy was used in a Leica DM IL LED microscope at 10X

magnification.

BCA Protein quantitation. Cells were titrated from 300,000 to 9,000 in a 96 well plate format. BCA

reagent (Pierce) was then prepared to the manufacturer's instruction by missing Reagent A and

Reagent B at a ratio of 50:1 respectively. 20 µL of the mixed reagent was added to 180 µL of sample

and incubated at 37 °C for 30 minutes. The plate was then read on plate reader measuring

absorbance at 562 nm and protein concentration calculated from these values.

Sample preparation for MALDI TOF MS analysis. Cell pellets were processed in one of three ways:

(a) Direct analysis where cell pellets were washed twice with PBS and centrifuged at 300 xg, 4 °C for

10 minutes. Cell pellets were then resuspended in 0.1% TFA before subsequent spotting.

(b) Cell pellets were snap frozen on dry ice and stored at -80°C until required. Thawed cell pellets

were centrifuged at 300 xg, 4 °C for 10 minutes before either being washed 1x with PBS or fixed in

4% paraformaldehyde solution or methanol on ice. Cell suspensions were then centrifuged at 300 xg,

4°C for 10 minutes before being resuspended in 0.1% TFA.

(c) Cell pellets were washed twice with PBS and centrifuged at 300 xg, 4 °C for 10 minutes. Cell

pellets were then either washed 1x with PBS or fixed in 4% paraformaldehyde solution or methanol

on ice. Cell suspensions were then centrifuged at 300 xg, 4°C for 10 minutes before being

resuspended in 0.1% snap frozen on dry ice and stored at -80°C until required.

Matrix preparation and spotting. Sinapinic Acid (SA), α-cyano-4-hydroxycinnamic acid (CHCA) and

dihydroxybenzoic acid (DHB) were used as matrices for all MALDI TOF cellular analysis. All matrix

solutions were prepared in 50% ACN, 0.1% TFA at varying concentrations and ratios of matrix solute:

2.5, 10, 20 mg/mL or saturated. For manual deposition cell suspensions were mixed at a 1:1 ratio

with matrix solution and 1 μL spotted onto a ground steel MALDI target before ambient drying.

Automated target spotting was performed using a Mosquito liquid handling robot (TTP Labtech) by

first mixing 1200 nL of matrix solution with 1200 nL of whole cell samples before subsequent

deposition of 200 nL of the matrix: sample mixture an AnchorChip MALDI. The target allowed to

ambient dry before analysis.

MALDI TOF MS analysis. A RapifleX PharmaPulse MALDI TOF/TOF mass spectrometer (Bruker

Daltonics) equipped with a Smartbeam 3D laser was used in positive ion mode with Compass 2.0

control for all data acquisition. Samples were acquired in automatic mode (AutoXecute; Bruker

Daltonics), totalling 10,000 shots at a 10-kHz frequency per spot. Full MALDI-TOF MS parameters can

be found in supplementary methods. MALDI TOF data processed by the FlexAnalysis 4.0 software

where a peak picking threshold of 3 S/N was set before being exported as a .csv file using

FlexAnalysis Batch Process (Compass 2.0) and further processed in Microsoft Excel and/or Perseus.[42]

Spectra based PCA plots were generated using ClinPro Tools (Bruker Daltonics). Data was then

formatted using both GraphPad Prism 7.0 and Adobe Illustrator.

Bootstrap statistical analysis. Hierarchical cluster analysis was performed on all biological and

technical replicates of the two conditions using the R package pvclust, with multiscale bootstrap

resampling of 10,000 iterations used to assess statistical significance by approximately unbiased

p-values.[43,44] Clustering was implemented using the average agglomeration method with a

correlation distance metric. The results are presented alongside a z-score heatmap of detected

mass features.[45]

Results and Discussion

Optimizing the workflow for mammalian cell MALDI TOF MS.

In order to optimize the sample preparation for whole cell MALDI TOF MS, we focused initially on

four different human cell lines (U2OS, MCF7, THP1 and HEK293) (Figure 1A). Cells were washed once

with PBS, which was sufficient to remove the culture medium contaminants as high levels of foetal

bovine serum (FBS) and salts from the culture medium affect MALDI TOF MS ionization.[36] To

determine optimal cell concentration, we spotted 25 to 20,000 cells on target. Surprisingly, there

was a narrow window where good spectra could be acquired, with large numbers of cells on-target

proving to be detrimental to ionization.

Further to this, we observed that the best spectral intensity varied for each cell line (Figure 1B) and

hypothesised that the number of ionisable biomolecules from cells was dependent on the cell size.

This relationship was also consistent between protein concentration and cell number (Figure S1).

Therefore, we measured the diameters of all four cells lines in solution (Table S-2) and plotted these

values against the optimal cell number derived from the titration to identify an optimal cells number

on-target for MALDI TOF MS analysis (Figure 1C). This generated a linear relationship with a very

good correlation of R^2 = 0.99 indicating that to obtain optimal and reproducible spectra from

mammalian cells by MALDI TOF MS, cell numbers on target need to be optimised and this number is

dependent on cell size. This generated a linear relationship with a very good correlation of R^2 = 0.99

indicating that to obtain optimal and reproducible spectra from mammalian cells by MALDI TOF MS,

cell numbers on target need to be optimised and this number is dependent on cell size.

We tested next if the biomolecules detected from mammalian cells derive from intact cells or if mild

breakage of cells enhanced the occurrence of unique mass features. In our experience, harsh lysing

conditions resulted in spectra that were less distinguishable, which has also been observed by lysing

with increasing acidity.[37] It was indicated before that freeze-thawing of cell pellets prior to MALDI

TOF MS analysis may have beneficial effects with respect to number of features identified and

overall spectral intensity.[33,46] This is likely due to the freeze/thaw cycle permeating the cell

membrane, thus allowing the cytoplasmic contents of the cells to become exposed and more easily

ionised. We therefore decided to test whether a freeze/thaw cycle improved MALDI TOF MS analysis

of mammalian cell lines and whether freezing before or after a wash with PBS affected sensitivity

and spectral quality compared with direct analysis (Figure 1D).

Both methods of freeze/thawing permeated the cell membrane of about 50-80% of the cells (Figure

1E). This led to a significant increase in the number of peaks identified compared to "intact" cell

samples (Figure 1F). As well as this, software analysis did not result in a significant difference

between cells frozen before or after further treatment and manual inspection of spectra resulted in

the same conclusion (Figure S2). We conclude that a freeze/thaw cycle is critical to improve MALDI

TOF MS quality of mammalian cells as it increases the number of features identified. However, the

order in which this step is performed, does not affect the final readout.

Suitability of mammalian cell fixing techniques for whole cell MALDI TOF-MS

Next, we took inspiration from cell and tissue fixing techniques and examined how different fixing

techniques influence the preparation of mammalian cells for MALDI TOF MS analysis. We

hypothesised that the use of these techniques would enable preservation of a specific cellular

phenotype and could be incorporated into a whole-cell MALDI TOF-MS compatible workflow. We

chose to initially test formaldehyde and alcohol fixing methods, methanol specifically, as they have

been used previously in MALDI imaging workflows.[47] Initial experiments revealed that samples

treated with 4% PFA generated spectra that were 5 – 10x less intense than other methods, (Figure

S3- 6) therefore we chose to continue by studying methanol fixation versus the previously reported

PBS washes. We systematically evaluated how both these methods performed with respect to the

number of identified peaks, quality of the acquired spectra, as well as technical reproducibility when

analysed by MALDI TOF MS.

Both methanol and PBS washing steps distinguished each of the four cell lines by both manual

spectra interrogation and principle component analysis (PCA) (Figure 2A-F) and, each extraction

method was able to generate a unique set of peaks for each of the four cell lines, thus allowing

classification of the different populations. As phenotype distinguishing peaks are often not the most

intense features and peaks identified with a lower S/N and intensity are less likely to be quantified

accurately, we therefore looked at how the relative intensity of peaks was distributed for the top 10

most intense peaks for each cell line (Figure 2G & H). This is important for high-throughput analysis,

and both PBS and methanol treatment showed a generally even intensity distribution. Finally, and

arguably most importantly, we tested at how reproducible peaks were identified over six technical

replicates (Figures 2 I & J, S2-5). Methanol was the most consistent, with the majority of all peaks

being identified in all six spectra, whereas PBS and water (pH 7) were slightly more variable. Taken

together, our data suggests that methanol fixation is comparable if not superior with PBS washing

for whole cell analysis and may be ideal for classification of subtle phenotypes.

Choosing a suitable matrix for mammalian cell MALDI TOF MS.

Next, we tested which type of matrix allows for the best MALDI TOF MS analysis of mammalian cells.

The three matrices mostly used in MALDI TOF MS of proteins and peptides are sinapinic acid (SA), α-

cyano-4-hydroxycinnamic acid (CHCA) and dihydroxybenzoic acid (DHB), which are often categorised

for the analysis of proteins, peptides, and glycans, lipids and peptides, respectively. However, when

analysing whole mammalian cells by MALDI TOF MS the origin of the biomolecules being ionised is

often unknown, and we therefore hypothesised that the choice of matrix will have a significant

influence on the resulting mass spectrum.

As expected, when each cell line sample was prepared with either saturated SA, CHCA or DHB,

significantly different mass profiles of the same cell line were observed (Figure S7). Using DHB matrix

resulted in more variable spectra over technical replicates, with the PCA analysis using ClinPro Tools

software showing wider distances and grouping only three of the six technical replicate spectra for

MCF7 cells (Figure S8A). Moreover, we could classify unique peaks to each of the matrices (Figure

S8B), which indicates that different biomolecules are being ionised and mammalian cell profiles are

matrix-dependent.

We then chose to look at how each matrix performs with respect to concentration and increasing

laser power. In all four cell lines, CHCA performed significantly better with respect to both

parameters. Although each matrix performed optimally at a saturated concentration, CHCA was able

to yield much more informative spectra and with more peaks identified at a third of the

concentration of DHB and SA. Following this, we observed that CHCA matrix was also able to ionise

cellular biomolecules at much a lower laser power than DHB and SA. We were able to fit non-linear

curves to these data, thereby identifying optimal minimal laser power and approximate saturation

points of each matrix (Figure 3, Table 1). As CHCA is known for ionising peptides and small proteins,

these data indicate that the biomolecules are either peptidic or small molecular weight proteins.

We chose to take these data further to understand the dynamics of matrix behaviour in a mock

screen and evaluate their performance across a 1536 target. We chose to use methanol treated THP-

1 cells that were then mixed with each matrix using a Mosquito HTS and spotted in 200 nL aliquots.

Each target was then analysed at the approximate saturation points described in Table 1: 60%, 80%

and 90% for CHCA, SA and DHB, respectively. When compared a laser power of 60%, this

corresponded to a laser energy fold change of 1.74x and 2.30x for 80% (DHB) and 90% (SA),

respectively. From manual inspection as well as a positive MALDI response we determined a spotting

accuracy of >96% for each target (Figure 4A). This infers that methanol-fixed whole-cell samples are

compatible with current liquid handling technologies, as well as MALDI TOF-MS.

We were also able to show that for CHCA and SA, the top five most intense features were identified

in >98% of spots, showing robustness for high-throughput screening, whereas for samples spotted

with DHB peak identification is much more variable (Figure 4B). This is somewhat expected as DHB

crystallises into long, needle-like structures that produce a heterogeneous surface. This combined

with an inherently heterogeneous sample such as fixed cells may account for this variability. Overall

spectral intensity based on the top five peaks as well as signal to noise also varied significantly

between the matrix conditions. Samples spotted with CHCA exhibited much greater spectral

intensity compared to SA, and an almost 10-fold increase when compared with DHB (Figure 4C), as

well as significantly better S/N for these top 5 features (Figure 4D). Interestingly, we observed

different lens contamination patterns for each of the three matrices after accumulation of 3072

individual spectra with each matrix (Figure S9). Both DHB and SA yielded significant deposition of

matrix onto the lens compared with CHCA. From our data set this does not appear to negatively

impact whole cell ionisation, however we do suspect that prolonged ionisation and exposure to

samples co-crystallised with either SA or DHB will impact studies that have a greater number of

samples such as those in high throughput screens. From this data, we conclude that CHCA would be

the matrix choice for whole cell analysis at a small and large scale due to its superiority across the

parameters discussed above. However, we do report that for masses greater than >10,000 Da, peak

resolution is significantly improved by using SA (Figure S10) and therefore may have a role to play in

studies that identify significant features in this mass range. Together, these results show how

phenotyping cells by MALDI-TOF MS can be scaled up to a high-throughput platform and still enable

robust identification of cell specific features.

MALDI TOF MS profiling of pharmacologically controlled stem cell differentiation.

Finally, we wanted to apply our optimised workflow to phenotypically profile cells in an

physiologically relevant system that is employed in drug screening and toxicity testing and that has

been used as a drug discovery model.[48] We used mouse embryonic stem cells (mESC) maintained in

a naïve ground state pluripotency using the 2i kinase inhibitor system (PD0325901 and CHIR99021),

which inhibit the kinases ERK1/2 and GSK3, respectively (Figure S11).[40,49]

Efficient exit from naïve ground state pluripotency towards differentiation upon 2i release was

confirmed by suppressed mRNA expression of the naïve pluripotency factors *Nanog* and *Klf4*, and

induction of the lineage priming/differentiation marker *Fgf5* (Figure 5A). As expected, the

pluripotency factor Pou5f1/Oct4, which is expressed in both naïve and lineage primed mESC states,

is not significantly altered upon acute 2i release (Figure 5A). Using MALDI TOF MS, we could robustly

identify unique features to each population, as well as quantify changes in common peaks. For all

spectra, the base peak was identified at *m/z* 4875, which made subsequent analysis simpler, as the

raw spectral intensity can vary significantly from spot-to-spot (Figure 5D & Figure S12). Utilizing *m/z*

4875, as a normalizing control, we identified a number of peaks that were unique to 2i and 2i

release, such as *m/z* 5566, 2i and 2i release conditions could be well differentiated as two

populations by PCA (Figure 5B) and we observed good grouping of biological replicates. A similar

distribution was identified when using a jackknife method (Figure S14, Table S3).[50] To further

understand how relative intensity of specific peaks changed across the three biological and five

technical replicates, we generated a Z-score averaged based heatmap of detected mass features

relative intensity (Figure 5C). For robustness, data were first filtered to include features that were

identified in 30% of all technical replicates. A complete heatmap of all features is presented in Figure

S15. This hierarchical clustering approach allowed us to look at the unique and common features

combined across all independent biological and technical replicates. The three biological replicates

clustered well together and 2i and 2i release conditions were separated efficiently and two discrete

row clusters emerged – peaks that were up-regulated and those that were down-regulated upon

release from PD0325901 and CHIR99021 inhibitors. This clustering was significant in the hierarchical

dendogram (Figure S16) showing classification of the two different cell populations. Through this, we

identified several peaks that changed significantly between the conditions and these can now be

used as features of phenotypic screening of mESC differentiation (Table 2). Our MS approach has

significant advantages over the conventional qPCR approach with respect to time, cost and

automation possibilities. Using MALDI-TOF MS and liquid handling robots, samples could be

processed in a 384 well plate format and analysed within one hour, approximately three times faster

than using qPCR. Furthermore, only 1000 cells are required to phenotype their differentiation state,

comparable with qPCR, but the consumable cost per sample is significantly reduced. Consumables for our MALDI-TOF MS assay are about £0.05-0.10 per sample, requiring only basic plastic ware, low solvent volume, CHCA matrix and stainless steel MALDI targets. These are significantly cheaper per sample compared to qPCR plate kits that typically cost >£2-5 per sample. Finally, our MALDI-TOF MS approach has the capability to be automated to a high-throughput scale using already established technologies such as the Mosquito HTS.

Conclusion

Due to its speed and its relative simplicity, MALDI TOF MS has become increasingly popular for the application of bacterial biotyping. However, a complimentary methodical approach to phenotypic screening of mammalian cells has not been well characterised. Here we presented a systematic study that explores initial sample handling, matrix choice and suitability of fixing techniques with whole cell MALDI TOF-MS analysis. We found that at all three steps had a profound impact on the resulting mass spectra and subsequent data analysis. We also applied a unique way of analysing the efficacy of each method by looking at not only spectral quality and observable changes but evaluating performance over technical replicate spots. This enabled us to gain a deeper understanding of how each step of the sample preparation impacts subsequent analysis and consequently an insight as to how each method would perform with higher throughput analyses. Our optimised method was validated by our observation of distinct MALDI TOF MS profiles for naïve ground state mESCs compared to differentiating mESCs in a pharmacologically controlled system. Using hierarchical clustering, we could visualize and identify a subset of peaks that are unique to each condition. We therefore present here a novel sample preparation method that enables robust, reproducible and rapid profiling of mammalian cells and is suitable for expansion to high-throughput platform.

ASSOCIATED CONTENT

Supporting Information

Supplementary figures, tables and texts (PDF)

Complete Table of 2i and 2i release relative intensities (.xlsx)

AUTHOR INFORMATION

Corresponding Author

*Email: matthias.trost@newcastle.ac.uk; phone: +44 191 2087009

Author Contributions

The manuscript was written through contributions of all authors. / All authors have given approval to
the final version of the manuscript.

Notes
The authors declared no potential conflicts of interest with respect to the research, authorship,
and/or publication of this article.

ACKNOWLEDGMENT

We thank the Bruker Daltonics, particularly Meike Hamester, Anja Resemann and Astrid Erdmann,
for their support. We thank Anetta Härtlova, Adam Moore, Shin Lim and Julien Peltier for help. The
authors disclosed receipt of the following financial support for the research, authorship, and/or
publication of this article: This work was funded by generous start-up funds from Newcastle
University to MT, an iCASE studentship to REH by the BBSRC and Bruker Daltonics. GMF is funded by
a Wellcome Trust/Royal Society Sir Henry Dale Fellowship, ASF is supported by a studentship from
the MRC.

References
1. Pappin. D. J, Hojrup. P & Bleasby, A. J. *Curr. Biol. CB* **3**, 327–332 (1993).
2. Cohen, L. H. and Gusev, A. I. *Anal. Bioanal. Chem.* **373**, 571–586 (2002).
3. Aichler, M. and Walch, A. *Lab. Invest.* **95**, 422–431 (2015).
4. Haslam. C, Hellicar. J, Dunn. A, Fuetterer. A, Hardy. N, Marshall. P, Paape. R, Pemberton. M,
Resemannand. A, Leveridge. M. *J. Biomol. Screen.* **21**, 176–186 (2016).
5. Chandler. J, Haslam. C, Hardy. N, Leveridge. M and Marshall, P. *SLAS Discov. Adv. Life Sci. R D* **22**,
1262–1269 (2017).
6. Winter. M, Ries. R, Kleiner. C, Bischoff. D, Luippold. A. H, Bretschneider. T, Büttner. F. H. *SLAS*
Technol. Transl. Life Sci. Innov. (2018).
7. Ritorto. M. S, Ewan. R, Perez-Oliva, A.B, Knebel. A, Buhrlage. S. J, Wightman. M, Kelly. S. M,
Wood. N. T, Virdee. S, Gray. N. S, Morrice. N. A, Alessi. D. R, Trost. M. *Nat. Commun.* **5**, 4763
(2014).
8. Cesare. V. D, Johnson. C, Barlow. V, Hastie. J, Knebel. A, Trost. M. *Cell Chem. Biol.* **0**, (2018).
9. Heap. R. E, Hope. A. G, Pearson. L, Reyskens. K. M. S. E, McElroy. R. P, Hastie. J, Porter. D. W,
Arthur. S, Gray. D. W, Trost. M. *SLAS Discov. Adv. Life Sci. R D* **22**, 1193–1202 (2017).
10. Winter. M, Bretschneider. T, Kleiner. C, Ries. R, Hehn. J. P, Redemann. N, Luippold. A. H, Bischoff.
D, Büttner. F. *SLAS Discov. Adv. Life Sci. R D* **23**, 561–573 (2018).
11. Guitot. K, Drujon. T, Burlina. F, Sagan. S, Beaupierre. S, Pamlard. O, Dodd. R. H, Guillou. C,
Bolbach. G, Sachon. E, Guianvarc'h. D. *Anal. Bioanal. Chem.* **409**, 3767–3777 (2017).
12. Guitot. K, Scarabello. S, Drujon. T, Bolbach. G, Burlina. F, Jeltsch. A, Sagan. S, Guianvarc'h. D.
Anal. Biochem. **456**, 25–31 (2014).

13. Gurard-Levin. Z. A, Scholle. M. D, Eisenberg. A. H. and Mrksich. M. *ACS Comb. Sci.* **13**, 347–350 (2011).

14. Patel. K, Sherrill. J, Mrksich. M and Scholle. M. D. *J. Biomol. Screen.* **20**, 842–848 (2015).

15. O'Kane. P. T. and Mrksich. M. *J. Am. Chem. Soc.* **139**, 10320–10327 (2017).

16. Wood. S. E, Sinsinbar. G, Gudlur. S, Nallani. M, Huang. C, Liedberg. B, Mrksich. M. *Angew. Chem. Int. Ed Engl.* **56**, 16531–16535 (2017).

17. Ban. L, Pettit. N, Li. L, Stuparu. A. D, Cai. L, Chen. W, Guan. W, Han. W, Wang. P. G, Mrskisch. M. *Nat. Chem. Biol.* **8**, 769–773 (2012).

18. Fenselau, C. and Demirev, P. A. *Mass Spectrom. Rev.* **20**, 157–171 (2001).

19. Croxatto. A, Prod'hom. G. and Greub. G. *FEMS Microbiol. Rev.* **36**, 380–407 (2012).

20. Claydon. M. A, Davey. S. N, Edwards-Jones. V. and Gordon. D. B. *Nat. Biotechnol.* **14**, 1584–1586 (1996).

21. Holland. R. D, Wilkes. J. G, Rafii. F, Sutherland. J. Bm Persons. C. C, Voorhees. K. J, Lay. J. O. *Rapid Commun. Mass Spectrom. RCM* **10**, 1227–1232 (1996).

22. Wang. Z, Dunlop. K, Long. S. R. and Li. L. *Anal. Chem.* **74**, 3174–3182 (2002).

23. Rahi. P, Prakash. O. and Shouche. Y. S. *Front. Microbiol.* **7**, (2016).

24. Jarman. K. H, Cebula. S. T, Saenz. A. J, Peterson. C. E, Valentine. N. B, Kingsley. M. T, Wahl. K. L. *Anal. Chem.* **72**, 1217–1223 (2000).

25. Williams. T. L, Andrzejewski. D, Lay. J. O. and Musser. S. M. *J. Am. Soc. Mass Spectrom.* **14**, 342–351 (2003).

26. Wunschel. S. C, Jarman. K. H, Petersen. C. E, Valentine. N. B, Wahl. K. L, Schauki. D, Jackman. J, Nelson. C. P, White. E. *J. Am. Soc. Mass Spectrom.* **16**, 456–462 (2005).

27. Alatoom. A. A, Cunningham. S. A, Ihde. S. M, Mandrekar. J. and Patel. R. *J. Clin. Microbiol.* **49**, 2868–2873 (2011).

28. Toh-Boyo. G. M, Wulff. S. S. and Basile. F. *Anal. Chem.* **84**, 9971–9980 (2012).

29. Bizzini. A, Durussel. C, Bille. J, Greub. G. and Prod'hom. G. *J. Clin. Microbiol.* **48**, 1549–1554 (2010).

30. Neumann.E, Ellis. J. J. F, Triplett. A. E, Rubakhin. S. S, Sweedler. J. V. *Anal. Chem.*, **91**, 12, 7871-7878 (2019)

31. Serafim. V, Shah. A, Puiu. M, Andreescu. N, Coricovac. D, Nosyrev. A, Spandidos. D. A, Tsatsakis. A. M, Dehelean. C, Peinzaru. I. *Int. J. Mol. Med.* **40**, 1096–1104 (2017).

32. Stubiger. G, Nairn. M. D, Abban. K. T, Openshaw. M. E, Mancera. L, Herzig. B, Wuczkowski. M, Senfter. D, Mader. R. M. *Anal Chem*, **90**, 13178-13182 (2018)

33. Chen. X, Wo. F, Chen. J, Tan. J, Wang. T, Liang. X, Wu. J. *Sci. Rep.* **7**, (2017).

34. Petukhova. V. Z, Young. A. N, Wang. J, Wang. M, Ladanyi. A, Kothari. R, Burdette. J. E, Sanchez. L. M. *J. Am. Soc. Mass. Spectrom.*, **30** (2019)

35. Holzlechner. M, Strasser. K, Zareva. E, Steinhäuser. L, Birnleitner. H, Beer. A, Bergmann. M, Oehler. R, Marchetti-Deschmann. M. *J. Proteome Res.* **16**, 65–76 (2017).

36. Ouedraogo. R, Flaudrops. C, Ben Amara. A, Capo. C, Raoult. D, Mege. J. L, Rénia. L. *PLoS ONE* **5**, (2010).

37. Munteanu. B, von Reitzenstein. C, Hänsch. G. M, Meyer. B and Hopf, C. *Anal. Bioanal. Chem.* **404**, 2277–2286 (2012).

38. Portevin, D. Pflüger. V, Otieno. P, Brunisholz. R, Vogel. G, Daubenberger. C. *BMC Biotechnol.* **15**, (2015).

39. Ouedraogo. R, Daumas. A, Ghigo, E, Capo. C, Mege, J. L, Textoris. J. *J. Proteomics* **75**, 5523–5532 (2012).

40. Ying, Q.-L. *et al.* The ground state of embryonic stem cell self-renewal. *Nature* **453**, 519–523 (2008).

41. Kalkan. T, Olova. N, Roode. M, Mulas. C, Lee. H. J, Nett. I, Marks. H, Walker. R, Stunnenberg. H. G, Lilley. K. S, Nichols. J, Reik. W, Bertone. P, Smith. *Dev. Camb. Engl.* **144**, 1221–1234 (2017).

42. Tyanova. S, Tikira. T and Cox. J. *Nat. Protcols*. **11**, 2301-2319 (2016).
43. Suzuki. R, Shimodaira H. *Bioinforma Oxf. Engl.* **22**, 1540-1542 (2006)

44. R Core Team. R: A language and environment for statistical computing. R Foundation for
Statistical Computing, Vienna, Austria. (2018)

45. Gu. Z, Eils. R, Schelsner. M. *Bioinforma. Oxf. Engl.,* **32**, 2847-2849 (2016)

46. Schwamb. S, Munteanu. B, Meyer. B, Hopf. C, Hafner. M, Wiedemann. P. *J. Biotechnol.* **168**, 452–
461 (2013).
47. Suthipintawong. C, Leong. A. S and Vinyuvat. S. *Diagn. Cytopathol.* **15**, 167–174 (1996).
48. Pouton. C. W. and Haynes. J. M. *Nat. Rev. Drug Discov.* **6**, 605–616 (2007).
49. Silva. J, Barrandon. O, Nichols. J, Kawaguchi. K, Theunissen. T. W, Smith. A. *PLOS Biol.* **6**, e253
(2008).
50. Nyamundanda. G, Brennan. L, Gormley. I. C. *BMC Bioinformatics*, **11**, 571, (2010)

Table 1. Laser power parameters determined from non-linear regression curves shown in Figure 4B.

Laser power parameters	U2OS			MCF7			THP1			HEK293		
Matrix	CHCA	DHB	SA	CHCA	DHB	SA	CHCA	DHB	SA	CHCA	DHB	SA
Optimal laser power (%)	38	83	70	32	79	73	35	71	76	31	73	78
Saturation point (%)	60	~90	90	60	~90	80	50	~90	80	60	~90	90

17 **Table 2.** Significant features identified in both 2i and 2i release spectra that exhibit a greater than 2
18 fold change.

m/z	Relative fold increase	
	2i	2i release
2437	x	2.03
5438	x	2.70
7148	x	2.08
7412	x	8.82
3097	4.57	x
4628	2.34	x
6197	3.18	x
11174	2.86	x

Figure 1: Optimizing cell numbers for whole cell MALDI TOF MS. (A) Light microscopy images of four

human cell lines U2OS, MCF7, THP1 and HEK293. (B) 2D plot of normalized mass spectrum intensity

at different cell numbers on target for each of the four cell lines. Plots have maxima indicating

optimal cell numbers. (C) Plotting optimal cell numbers derived from (B) against their respective

measured cell diameter shows a high correlation. Plots in (B) derived from 6 technical replicates.

Error bars represent standard deviation. (D) Schematic showing workflow. (E) Trypan blue staining of

the four cell lines shows that the freeze/thaw cycle significantly increases percentage of trypan blue

positive cells and therefore cell membrane permeability. (F) Number of peaks identified across six

technical replicates for each of the experimental workflows shown in (D). Data show that the

number of features identified are significantly increased upon inclusion of a freeze/thaw cycle. Error

bars represent standard deviation of six replicates. *** and **** represent p<0.001 and p<0.0001,

respectively, student's t-test.

Figure 2: Methanol fixation and PBS washing are suitable for identifying reproducible features in

whole cell MALDI TOF MS. (A & B) Number of identified unique features for any of the four cells

after methanol fixation (A) and PBS washes (B). (C & D) PCA plots showing how multivariate analysis

can distinguish each of the four cell lines after methanol fixation (C) and PBS washes (D). (E & F)

Number of peaks identified for each cell line over six technical replicates. (G & H) Relative intensity

distribution of the top 10 peaks identified over six technical replicates for each cell line. (I & J) Pie

charts displaying the % of peaks identified within each of the six replicates per cell line.

Figure 3: Evaluating matrix performance when analysing whole cell samples by MALDI TOF MS. (A)

Dependence of matrix concentration from 2.5 – 30 mg/mL (saturated) on number of peaks identified

for each of the four human cell lines and three matrices. (B) Laser power dependence of each of the

three matrices for each of the four cell lines. Non-linear regressions were fitted with a Sigmoidal, 4PL

curve with good fits of R^2 >0.85 – 0.99. All error bars are given as standard deviations over six

technical replicates.

Figure 4: CHCA is the most suitable matrix for studies with a large sample number. (A) Mosquito

spotting accuracy determined by positive signal spectra and manual inspection for each of the

matrices. Percentage of 1536 spots. (B) Percentage peak identification of the top 5 most intense

peaks over a 1536 target for each of the three matrices. (C) Frequency distribution of the top 5 most

intense peak intensities of a 1536 target for each of the three matrices. (D) Violin plots showing

distribution of the signal-to-noise of the top 5 most intense peaks over a 1536 target for each of the

three matrices. Solid lines indicate median values and dotted lines the upper and lower quartiles.

2   **Figure 5: Whole cell MALDI TOF MS distinguishes between naïve and differentiating mESCs** (A)

3   qPCR data showing changes in key mESC pluripotency and differentiation genes. Error bars represent

standard deviation of two technical replicates from three biological replicates. *** and ****

represent p<0.001 and p<0.0001, respectively, student's t-test. (B) PCA plot showing how the two

cell populations can be distinguished by multivariate analysis over three biological and five technical

replicates. (C) Z-score heat map showing how each normalized peak intensity changes with respect

to 2i and 2i release condition over three biological and five technical replicates. (D) Selected

examples of mass spectral regions that showed changes between mESCs cultured in 2i and 2i release

conditions where the base peak (4875 *m/z*) is of identical intensity, thus allowing relative intensity

comparison.

10