µECoG Recordings Through a Thinned Skull

Sarah K. Brodnick[1], Jared P. Ness[1], Thomas J. Richner[1], SanittaThongpang[3], Joseph Novello[1],
Mohammed Hayat[1], Kevin P Cheng[1], Lisa Krugner-Higby[2], Aaron J. Suminski[1,] Kip A.
Ludwig[1], Justin C. Williams[1]*

[1]Department of Biomedical Engineering, University of Wisconsin-Madison, 1550 Engineering Drive,
Madison, WI USA 53706

[2]Laboratory Department of Surgical Sciences, School of Veterinary Medicine, University of
Wisconsin Madison, 2015 Linden Drive, Madison, WI 53706

[3]Department of Biomedical Engineering, Mahidol University, 25/25 Puttamonton 4 Road, Salaya,
Puttamonton, Nakorn Pathom 73170, Thailand

* Correspondence:
Justin Williams
jwilliams@engr.wisc.edu

Keywords: Thinned skull$_1$, µECoG$_2$, Local Field Potenitals$_3$, Optogenetics$_4$, Somatosensory
Evoked Potentials$_5$. (Min.5-Max. 8)

Abstract

The studies described in this paper for the first time characterize the acute and chronic performance
of optically transparent thin-film µECoG grids implanted on a thinned skull as both an
electrophysiological complement to existing thinned skull preparation for optical
recordings/manipulations, and a less invasive alternative to epidural or subdurally placed µECoG
arrays. In a longitudinal chronic study, µECoG grids placed on top of a thinned skull maintain
impedances comparable to epidurally placed µECoG grids that are stable for periods of at least one
month. Optogenetic activation of cortex is also reliably demonstrated through the optically
transparent ECoG grids acutely placed on the thinned skull. Finally, spatially distinct
electrophysiological recordings were evident on µECoG electrodes placed on a thinned skull
separated by 500-750µm, as assessed by stimulation evoked responses using optogenetic activation
of cortex as well as invasive and epidermal stimulation of the sciatic and median nerve at chronic
time points. Neural signals were collected through a thinned skull in multiple species, demonstrating
potential utility in neuroscience research applications such as *in vivo* imaging, optogenetics, calcium
imaging, and neurovascular coupling.

31   1   Introduction

Electrophysiological recordings of brain activity using high density electrode arrays are a staple of
neuroscience research and have become increasingly prevalent for the clinical diagnosis of epileptic
seizure foci as well as the clinical deployment of Brain-Machine Interfaces (BMI) [1; 2; 3].
Traditional electrophysiological recording methods involve the implantation of invasive electrode

arrays either indwelling within cortex [4; 5], beneath the dura (subdural) [6; 7; 8], on top of the dura
(epidural) [9; 10; 11], or non-invasively on the skin directly on above the exterior of the skull [12]. It
is generally accepted that electrode placement closer to the neural signal sources of interest within the
brain yields a more information rich and spatially distinct signal [13], whereas activity measured at a
distance non-invasively is attenuated in part by the high impedance skull, yielding less spatially
distinct information in the recorded signal from electrode to electrode [14; 15].

More recently, there has been a growing appreciation that surgical methods to open the skull, and/or
the placement of an indwelling electrode grid on or within cortex, may cause adverse effects that
impact the neural circuitry of interest [13; 16; 17]. For example, increased glial scarring [15; 18; 19],
large increases in temperature of cortex [20], changes in intracranial pressure [21], intracranial
hemorrhage and/or physical depression of cortex [18; 21; 22], and bacterial infection [21] have all
been linked to the surgical procedure and implantation of electrocorticography (ECoG) or indwelling
cortical arrays. These adverse events cause subtle changes to the neural circuitry of interest that have
been shown to cause long-lasting deficit in performance of fine motor tasks amongst other
consequences[16].

Concurrently there has been growing interest in neuroscience experiments that thin the skull instead
of opening the skull to perform experiments using optical methods to record and manipulate both
neuronal and non-neuronal cells within the brain [23; 24; 25; 26]. Removal of the skull in rodents
has been shown to create glial scarring in the area under the craniotomy (Yang 2010 Nature
Protocols, Yan 2009 J. Neurosci). Unlike the outer compact layer of the skull which has low
conductivity, the spongy bone of the skull closer to the brain is low impedance [27] and if thinned
appropriately is optically transparent [28]. However, the performance of µECoG grids placed
chronically on a thinned skull preparation has yet to be evaluated.
59
To address this gap, a series of acute and chronic studies was performed where the skull was thinned
to a translucent layer and implanted with a µECoG array. µECoG arrays were used in the
experiments described because of their flexibility, transparency, and well characterized epidural
signal profile [9; 10].   In rats chronically implanted for one month, impedance values and evoked
somatosensory evoked potentials (SSEPs) were recorded at regular intervals to assess stability of
electrical function and spatial resolution of recordings through the thinned skull. Cortical signals
from optogenetic stimulation in a ChR2 mouse were recorded in an acute terminal session through a
thinned skull and were compared to recordings through the dura after removal of the thinned skull.
These studies tested multiple common stimulation paradigms for neuroscience research in multiple
species to characterize the reliability and spatial resolution of electrophysiological recordings through
a thinned skull.

2   Materials and Methods

2.1   Ethics Statement

All animal procedures were approved by the Institutional Animal Care and Use Committee (IACUC)
at the University of Wisconsin - Madison. All efforts were made to minimize animal discomfort.

2.2   Device Fabrication

µECoG devices were fabricated following protocols previously described for polyimide [10] and
Parylene C [29] arrays (Figure 1). Rat sized polyimide (750 um spacing, 250 um site diameter)
(Figure 1b) or Parylene C (750 um spacing, 200 um site diameter) (Figure 1c) based µECoG

electrode arrays were custom fabricated with 16 platinum sites (one or two 4mm x 4mm grids) and
implanted unilaterally or bilaterally (Figure 1c) between bregma and lambda in Sprague-Dawley rats.
Similarly, for experiments with mice, a smaller, 2mm x 2mm 16 platinum site Parylene-C µECoG
array (500 um spacing, 150 um site diameter), was fabricated and used for optogenetic experiments
(Figure 2c). Parylene C was chosen for optogenetic and imaging studies due to its flexibility and
translucent properties.

2.3   Surgical Preparation

Chronic Experiments

Male Sprague-Dawley rats (n=7, Envigo, Indianapolis IN) 2-4 months old were chronically
implanted with custom built µECoG arrays (Figure 1a). Three rats were implanted with bilateral
arrays over thinned skull, three rats were implanted with bilateral arrays over the dural surface, and
one rat was implanted with a bilateral array, one over thinned skull and one over the dura. The
electrode array on the dural hemisphere for this animal was damaged during insertion and not viable
therefore electrophysiology and impedance data were not included in this paper, however,
histological staining was performed and included. Surgical procedures were based on previously
published methods [30]. Before surgery, buprenorphine hydrochloride (0.05 mg kg−1, Reckitt
Benckiser Healthcare) was administered for analgesia and dexamethasone (2 mg kg−1, AgriLabs) to
prevent cerebral edema. Rats were induced with 5% isoflurane gas in O2 and maintained on 1.0-
2.5% throughout the duration of the surgery. Following induction, rats were placed into a stereotaxic
frame with the scalp shaved and prepped with alternating povidone iodine and alcohol. The skin was
incised, and the exposed skull was cleaned and dried. Three stainless steel screws (stainless steel,
00–80 × 1/8 in.), two for attachment of a ground wire, and another for reference and mechanical
support, were attached to the rostral and caudal areas of the skull. Next, UV curable dental acrylic
(Fusio, Pentron Clinical) was placed on the periphery of the exposed skull to provide an anchor for
the attachment of future acrylic, and two craniotomies (~5mm x 5mm) or thin skull areas were drilled
over somatosensory cortex. A thinned skull area was made by drilling through the top layer of
compact bone, through the spongy layer, and slightly into the lower compact bone where it became
transparent. The µECoG arrays were placed epidurally or over a thinned skull area and covered with
a thin layer of GelFoam (Pharmacia and Upjohn Co, New York, NY) and saline before being covered
by dental acrylic. GelFoam was used to prevent acrylic from covering the electrode array and was
only placed on top of the arrays and not beneath. The ZIF connector was then secured to the skull
and a purse string suture (3-0 vicryl) closed the skin wound. Triple antibiotic ointment was applied to
the wound during closing to prevent infection. Rats were monitored post-surgically until they were
ambulatory and showed no signs of pain or distress.   Another dose of buprenorphine was
administered 8-12 hours after the initial dose to relieve any pain the animal may have been
experiencing following the surgery. Ampicillin (50 mg kg−1 SC, Sage Pharmaceutical) was
administered twice daily for seven days postoperatively to prevent infection.

2.4   Periodic Chronic Electrophysiology Testing

Sensorimotor evoked potentials were recorded periodically for up to one month under sedation in rats
with chronic µECoG implants to assess signal stability and uniqueness/spatial resolution of
information recorded on nearby sites. Dexmedetomidine (50 µgrams/kg subcutaneous injection
(SC)) was used to achieve sedation. Atipamezole (0.5 mg/kg SC) was administered at the end of the
procedure as a reversal agent. Dexmedetomidine sedation was supplemented with small amounts of
isoflurane (0-0.5%) throughout the procedure to deepen sedation. The sciatic or median nerve,
hindlimb or forelimb respectively, were stimulated weekly to evoke somatosensory evoked potentials

(SSEPs). Needle or surface stimulation electrodes were used. Needle electrodes were placed on
either side of the sciatic nerve, 3mm apart. Surface electrodes were placed on shaved skin above the
sciatic or median nerve, with a reference electrode placed below the leg. Stimulation pulses were
applied using needle electrodes (monophasic 0-0.8mA for 2ms), or surface electrodes (monophasic
0.5-3.5mA for 1ms) both at approximately 0.5 Hz. The cortical response was recorded
simultaneously at 24KHz using a multichannel recording system (TDT, Alachua, Florida).

2.5   Acute Terminal Terminal Experiments

Three Thy1::ChR2/H134R-YFP (ChR2) mice (Jackson Laboratory; stock number 012350) ~6-16
weeks old were implanted during acute terminal recording sessions with µECoG arrays implanted
over the dura or a thinned skull area to compare neural signals recorded from light stimulation
(Figure 2a, 2b). Evoked potential data from optogenetic stimulation were collected from three mice
and strength duration curve data were collected from one mouse. Arrays were placed onto a thinned
skull first, and then placed epidurally after removing the thinned skull. Mouse surgical procedures
were similar to previously published methods [31]. Briefly, mice were administered buprenorphine
hydrochloride (0.05 mg kg−1) and dexamathasone (1 mg kg−1 SC) before induction, induced, and
maintained with 1-2.5% isoflurane. The animal was placed in a stereotaxic-like frame and a
craniotomy or thinned skull was performed. A µECoG array was placed on the dura or thinned skull
and ground and reference wires were coiled and placed on a small area of thinned skull on the
contralateral hemisphere. GelFoam was not used in optogenetic studies. Instead the cortical surface
was continually kept wet with a saline drip.
144
Heart rate and blood oxygen concentration in both species were monitored throughout the surgery
using a pulse oximeter. Body temperature was monitored with a digital thermometer and regulated
with a water-circulated heating blanket.

2.6   Acute Terminal Optogenetic Electrophysiological Testing

Optogenetically evoked potentials were recorded during a terminal procedure by shining light though
a fiber coupled LASER system or LED through an optically transparent parylene µECoG onto the
dura or thinned skull of ChR2 mice using previously reported methods (15) (figure 2).
Photostimulation was accomplished by using an optical fiber (200µm in diameter, 0.22 NA, flat
cleaved and polished, Thorlabs, Newton, NJ) connected to a 100mW 473 nm LASER (Laserglow,
Toronto, ON) and controlled by a multichannel system (TDT, Alachua, FL). 2.5ms pulses, varying
power settings, and random interstimulus intervals were used. Power at the tip of the optical fiber
was approximately 80 mW/mm^2 and was placed approximately 1mm from the cortical or thinned
skull surface. Recordings were taken with a Tucker-Davis Technologies RZ2 amplifier (TDT,
Alachua, FL), and sampled with a high impedance headstage. A photostimulus delivered by an LED
(465 nm, RGB MC-E, Cree, Durham, NC) approximately 2 cm away from the cortical or thinned
skull surface was used to create photostimulus duration versus amplitude peak to peak potential
contour plots (figure 8). Voltage pulses were changed to current pulses (0–1000 mA, 0.5–12 ms)
with an LED driver (BuckBlock, LEDdynamics, Randolph, VT). Irradiance was calculated by
measuring optical power (PM100D, S130C, Thor Labs, Newton, NJ) 2 cm from the LED, and the
result was divided by the commercially available photo sensor's area (S370 Optometer, United
Detector Technology, Hawthorne, CA).

2.7   Electrophysiology Analysis

Chronic electrical evoked responses were recorded through a thinned skull preparation.
Optogenetically evoked responses from terminal mice experiments were recorded both epidurally and
through a thin layer of skull. Local field potentials (LFPs) were bandpass filtered using a
combination of a 2nd order, Butterworth lowpass filter (cutoff frequency = 1000 Hz), a Butterworth
high pass filter (cutoff frequency = 3Hz), and a 3rd order notch filter (cutoff frequencies =55Hz and
65Hz) to remove line noise. Evoked potentials were computed from the average of evoked responses
from the same stimulus amplitude and channel. To increase signal-to-noise ratios, two known post
processing referencing techniques, common average referencing (CAR) and small Laplacian
referencing, were employed and compared (Figure 3).   Each were incorporated as described in the
literature [32; 33]. After small Laplacian referencing, heatmaps of the electrical and optogenetic
evoked responses (Figures 4 and 7), were created using the maximum positive and negative peaks to
visualize the spatial organization of the cortical responses. Each peak was defined as the average of
7 data points centered at the maximum and minimum points of the response in a window defined as
roughly 10ms to 35 ms post stimulus onset and subtracted [34; 35]. Channels above 600kOhms were
considered to be non-functional and removed from analysis.

2.8   Impedance Recordings and Analysis

Electrical impedance spectrum data was collected from arrays before implantation, and periodically
after implantation to assess electrical characteristics using a potentiostat (Autolab PGSTAT 128N,
Metrohm, Riverview, FL) and following previously published methods (Figure 5) [36]. Arrays that
were determined viable for implantation had values of approximately 50-100 kΩ at 1 kHz. Animals
were trained to sit still with treats and were not anesthetized or sedated for chronic impedance
measurements. Analysis consisted of data from six rats, three with thinned skull implants and three
with epidural implants for comparison. Each rat had a bilateral implant consisting of 32 electrode
sites. Impedance measurements were gathered from each electrode for 30 days post implantation.
Resistive values at 1kHz were plotted for each of the 32 channels corresponding to length of time of
the implant. Single channels with a resistance above 600 kΩ were considered outliers and eliminated
from calculations for that day. Outliers were considered to be broken or due to an inadequate
connection. Average resistance was plotted for each day and fitted to a curve across days using cubic
spline interpolation to account for measurements potentially not lining up exactly on individual days
across animals. The thinned skull implants interpolation curves were averaged together and plotted
against the epidural implants averaged interpolation curves. Impedance values were not recorded
after implantation in acute mice experiments, although pre-implantation impedance values were
comparable to those of the devices implanted for chronic recordings in rats.

3.0   Results and Discussion

3.1   Chronic periodic sensory evoked potential recordings in rats

Chronic SSEP recordings were obtained weekly during electrical stimulation of the sciatic or median
nerve in three rats to compare the spatial resolution of thinned skull µECoG arrays versus traditional
epidural arrays. Thinned skull µECoG arrays were implanted bilaterally in sensorimotor cortex in
each rat and SSEPs were recorded in each contralateral hemisphere from a cutaneous electrical
stimulus of hindlimb or forelimb. A representative plot of thinned skull SSEPs on the left
hemisphere from right hindlimb stimulation is shown in Figure 3.   Using the stainless-steel bone
screw as a reference, which was implanted cranial and contralateral to the µECoG array (Figure 3a),
recorded signals contained common noise and differences in SSEPs from nearby electrode locations
were not readily apparent. Two post process referencing techniques were used to reduce both
common noise and common signal to highlight spatially distinct differences in neural signals.

Employing a Common Average Reference (CAR) (Figure 3b) successfully recovered spatially
distinct hindlimb SSEPs on adjacent electrode sites. Similarly, employing a small Laplacian (Figure
3c) reference post-hoc further highlighted spatially distinct SSEP responses on adjacent sites.
Consequently, we chose to use the small Laplacian post-hoc referencing strategy for the remainder of
the recording data, because it visually increased unique highlighted spatial information present in the
SSEP on adjacent sites [32].
218
Distinct somatotopic signals were recorded 38 days post-implantation from µECoG arrays placed on
a thinned skull area of the rat's right sensorimotor cortex from both left hindlimb and forelimb
stimulation (Figure 4). Small Laplacian referencing methods were also applied. Highest peak to peak
SSEP values from forelimb stimulation, according to the heatmap, are positioned at the anterior
portion of the electrode with peaks spanning both medially and laterally (Figure 4a). When switching
the area of stimulation to the hindlimb, SSEPs shifted medially similar to previously mapped rat
sensorimotor cortex (Figure 4b) [35]. The recorded sensory responses are consistent with the
response latencies for myelinated sensory fiber conduction, around 13ms for forelimb and 17ms for
hindlimb according to previously published data [34; 35].
228
Thinned skull µECoG electrode arrays not surprisingly have lower signal amplitudes recorded during
evoked responses in comparison to historical studies using the same arrays placed epidurally which
are closer to the source of the neural signal [9]. As a result, it becomes more important to CAR and
small Laplacian referencing strategies to eliminate common signal/noise to uncover spatially distinct
spatial information for neuroscience applications. Given a similar SSEP was recorded across all
electrode sites with appropriate conduction latency prior to post-hoc referencing, this may suggest the
common signal was recorded at the stainless-steel bone screws in contact with the surface of the
brain used for the reference and ground respectively. Although the ECoG signal recorded from the
bone screws has been insignificant compared to signals recorded epidurally from µECoG in previous
studies and therefore post-hoc referencing was not required to reveal spatially distinct information
from site to site, the attenuation of signal through the thin skull made the small common signal
putatively recorded from the stainless-steel screws more problematic. Consequently, post-hoc
referencing was necessary before spatially distinct SSEPs were observed on adjacent electrode sites.
242
Thinned skull µECoG electrode arrays also have been shown in this study to record information on a
temporal scale similar to epidurally placed arrays. For example, thinned skull SSEPs were recorded
at a temporal resolution of roughly 15ms. Currently, GCaMP6f is a popular genetically coded
calcium indication (GECI) that is commonly used to observe neural activity at an onset of
approximately 45ms [37]. This suggests that the incorporation of an optically transparent µECoG
array with common thinned skull experiments for optical imaging would provide unique,
complementary temporal information.
250
251   **3.2   Chronic periodic impedance spectra recordings in rats**
252
To compare the electrical performance of epidural vs thinned skull placed electrodes in rats over
time, we measured the impedance spectra of electrodes on each array at 1kHz periodically over the
chronic implantation period (Figure 5). Impedance plots from µECoG arrays implanted on a thinned
skull preparation showed significantly different patterns of change over time than those implanted
epidurally (Figure 5). Initial electrode impedances were similar when measured in 0.9 % w/v
phosphate buffered NaCl saline (~25-125mOhms at 1KHz). After approximately 14 days of
implantation as shown in Figure 5, the impedances of the electrodes on the epidural surface were
higher on average than that of the electrodes on the thinned skull surface. The epidural impedance

interpolation curve shows rise in impedance around one week after implantation and lasting for
approximately 14 days, similar to other microelectrodes implanted in or on cortex in other studies
[36; 38]. This may be attributed to a central nervous system immune response and new tissue
formation and follows previous intracortical and epidural implantation impedance results [9; 36].
Impedances of epidural implants reached a steady state between 2-4 weeks post-implant reflecting
decelerated wound healing. In contrast, the thinned skull impedance interpolation curve remained
relatively stable for approximately 21 days post-surgery, only rising slightly towards the end of the
30-day period. This suggests the chronic electrode/tissue interface is different in composition than the
epidural grids, but also demonstrates that bone regrowth/scarring under the thinned skull electrodes
does not significantly increase the impedance by comparison. Supplementary figure 1 shows line
plots of individually recorded impedance values from each rat over a time period of one month.
272
Decreased impedances during the first few weeks of thinned skull electrode implantation may
suggest edema, and that fluid remained at the electrode/tissue interface without clearing. Extra fluid
could have hypothetically caused shunting of current and increased distance between the electrode
array and the thinned skull. Regardless, we were still able to record spatially and temporally accurate
SSEPs and optogenetically induced filed potentials from thinned skull electrodes in rats and mice
with relatively low impedance values (impedance values not acquired post-implantation in acute
mice studies).
280
281 **3.3 Comparison of thinned skull vs epidural recordings from optogenetic light stimulation in**
282 **acute terminal mice**
283
To further investigate the spatial resolution of information on nearby electrodes given a thinned skull
recording approach, a light stimulus was applied through a clear Parylene C μECoG array and
thinned skull to optogenetically activate neurons expressing light sensitive proteins. Optogenetically
evoked potentials were recorded in ChR2 mice through a thinned skull (figure 6a) and epidurally
using a smaller 2mm by 2mm clear μECoG array to generate a consistent focal activation of cortex
for comparison. Evoked responses through the thinned skull showed the highest peak responses near
the foci of optogenetic stimulation after small Laplacian referencing, further demonstrating spatiality
the spatial recording ability of the preparation (Figure 6a). Increasing 473nm laser power also
increased evoked potential peaks amplitudes. A similar response paradigm occurs with epidural
stimulation (figure 6b), however, we obtain a much larger signal possibly due to lack of spatial
filtration of signal through the skull. Figures 6c and 6d use a 2D interpolated heat map to show
differences in peak to peak amplitudes at a stimulation laser power of 545.5mW/mm2. Both thinned
skull and epidural heat maps display spatial distinct recordings on nearby electrode sites. Due to the
presumed filtration/attenuation of signal through the skull and other tissues, the thinned skull
recording (Figure 6c) is approximately 10 times less in peak-to-peak amplitude than the epidural
recording (Figure6d). The thinned skull signal also seems to be slightly more diffuse given
appropriate referencing strategies.
301
One ChR2 mouse underwent an acute procedure where the skull was thinned, and a 465nm LED was
positioned 2 cm away from the thinned skull with the light power and duration values varied to
generate photostimulus strength vs duration curves (Figure 7). The resulting illumination covered
most of the cortical area under the μECoG array. Stimuli strength and duration were applied
randomly, and peak amplitudes of signals recorded. The thinned skull was then removed exposing
the dural surface and stimulation procedure repeated. Figure 7a depicts a contour plots for signals
recorded from the dura, whereas figure 7b depicts the same plot from the thinned skull.
309

The optogenetically evoked µECoG signal on both the epidural and thinned skull grids demonstrated
spatially distinct information, with waveform reversals often apparent on two adjacent sites. These
reversals, in conjunction with the waveshape of the evoked response, demonstrated that the
electrophysiological recordings were not photoelectric artifacts. Although the magnitude in µVolts
of the evoked signal was approximately 10x less with the thinned skull preparation than with the
epidurally placed grids, the spatial information as assessed by differences in recordings at adjacent
electrode sites was highly similar after post-hoc Small Laplacian referencing.

3.4   Imaging of immune response in neural tissue to thinned skull and epidural µECoG
implantations

Given the impedance responses over time in all animals, histology was performed on rat M32 to
compare histology to the impedance measurement of approximately 50kOhms at timepoint 32 days
post-implantation. Histologic sectioning and immunochemical staining for glial fibrillary acidic
protein (GFAP) for astrocytes and Iba-1 for microglia and/or infiltrating macrophages was performed
with perfused neural tissue in the single rat with bilateral thinned skull/epidural µECoG arrays.
(Supplementary figure 2). The cortical region directly beneath the epidural preparation showed
putative increase in GFAP immunoreactivity and projection of astrocytic processes towards the
cortical surface (Supplementary figure 2c). Iba-1 staining did not reveal any obvious increases in
microglial immunoreactivity within the brain tissue on either the epidural or thinned skull
hemispheres (Supplementary figure 2d). However, an apparent thickening of the dura on the epidural
side was observed (Supplementary figure 2f) which contained a higher density of Iba-1 positive cells,
either microglia or infiltrating macrophages, that was not present on the thinned skull side of the
animal (Supplementary figure 2e).   Previous studies have also reported thickening of the dura under
the µECoG array consisting of collagen [22; 39].

Gross visual inspection of the brain surface post-mortem directly under the thinned skull where a
µECoG array was chronically placed was also directly compared to the surface of the brain directly
under a chronically implanted epidural µECoG array (see Supplementary Figure 3). Consistent with
data reported by Onal 2003 [21] and Degenhart 2016 [22], the epidurally placed chronic µECoG
array leaves a visible dent/impression on the surface of the brain upon removal. Similarly, denting is
also visible beneath the two stainless steel mechanical and reference screws placed posterior to the
olfactory bulbs. In contrast, no denting of the brain is evident under the thinned skull where the
µECoG array was chronically placed.

The main benefit of the thinned skull preparation is that the skull remains partially intact. When the
skull is completely removed many side effects can occur which may impact the interpretation of
behavioral results. Previous studies in the field of *in vivo* imaging have showed increased glial
reaction (microglia and astrocytes) under an open craniotomy window preparation compared to a
thinned skull window preparation in mice [23]. Pneumocephalus can occur after craniotomy in a
clinical setting which involves air being trapped in the cranial cavity [40]. Also, dendritic spine
plasticity has been shown to differ in thinned versus open-skull window preparations, emphasizing
that the neural environment under a craniotomy may be changed by the craniotomy itself [23; 41].
Another benefit the thinned skull recording technique might offer is improved implant mechanical
stability, and the lack of direct contact with the surface of the brain or dura. The latter may reduce
the risk of injury to neural tissue or device failure due to the lack of device movement on the surface
of the brain, although this will need to be investigated in further studies.

4.0   Conclusion

In summary, the studies described in this paper for the first time demonstrate that µECoG grids
placed on a thin skulled can provide stable, spatially distinct electrophysiological information out to
periods in excess of a month. This method may be particularly useful as a complement to other
studies which thin the skull for optical recordings/modulation of neural activity. µECoG recording
grids may also provide a useful balance between invasiveness, information content, and day to day
stability that may be important for neuroprosthetics applications. In addition, this novel method may
be critically enabling for neuroscience studies in which minimizing the trauma to the underlying
neural or non-neuronal cells of interest is necessary to avoid potential confounds given the
fundamental hypothesis to be tested.

5.0 Conflict of Interest

JCW is a scientific board member and has stock interests in NeuroOne Medical Inc., a company
developing next generation epilepsy monitoring devices. He also has an equity interest in
Neuronexus, a company that supplies electrophysiology equipment and multichannel probes to the
neuroscience research community. KAL has stock interests in NeuroOne Medical Inc., a company
developing next generation epilepsy monitoring devices.

6.0 Author Contributions

SB, JN and TR: experimental design, data analysis, and manuscript preparation. AS, KL and JW:
experimental design and manuscript preparation. KC, MH, and JN: data collection and data analysis.
ST for device fabrication, and data collection. LKH contributed to manuscript preparation. All
authors contributed to draft the manuscript and have read and approved the final manuscript.,

7.0 Funding

This work was sponsored by the Defense Advanced Research Projects Agency (DARPA) MTO
under the auspices of Dr. Jack Judy and Dr. Doug Weber through the Space and Naval Warfare
Systems Center, Pacific Grant/Contract No. N66001-11-1-4013 and No. N66001 -12-C-4025.

9.0 References

[1] E.C. Leuthardt, G. Schalk, J.R. Wolpaw, J.G. Ojemann, and D.W. Moran, A brain-computer
interface using electrocorticographic signals in humans. J Neural Eng 1 (2004) 63-71.

[2] E.C. Leuthardt, G. Schalk, D. Moran, and J.G. Ojemann, The emerging world of motor
neuroprosthetics: a neurosurgical perspective. Neurosurgery 59 (2006) 1-14; discussion 1-14.

[3] I. Osorio, M.G. Frei, J. Giftakis, T. Peters, J. Ingram, M. Turnbull, M. Herzog, M.T. Rise, S.
Schaffner, R.A. Wennberg, T.S. Walczak, M.W. Risinger, and C. Ajmone-Marsan,
Performance reassessment of a real-time seizure-detection algorithm on long ECoG series.
Epilepsia 43 (2002) 1522-35.

[4] D.R. Kipke, R.J. Vetter, J.C. Williams, and J.F. Hetke, Silicon-substrate intracortical
microelectrode arrays for long-term recording of neuronal spike activity in cerebral cortex.
IEEE Trans Neural Syst Rehabil Eng 11 (2003) 151-5.

[5] R.A. Normann, and E. Fernandez, Clinical applications of penetrating neural interfaces and Utah
Electrode Array technologies. J Neural Eng 13 (2016) 061003.

[6] A.R. Wyler, G.A. Ojemann, E. Lettich, and A.A. Ward, Jr., Subdural strip electrodes for
localizing epileptogenic foci. J Neurosurg 60 (1984) 1195-200.

[7] D. Khodagholy, J.N. Gelinas, T. Thesen, W. Doyle, O. Devinsky, G.G. Malliaras, and G.
Buzsáki, NeuroGrid: recording action potentials from the surface of the brain. Nature
Neuroscience 18 (2014) 310.

[8] C. Henle, M. Raab, J.G. Cordeiro, S. Doostkam, A. Schulze-Bonhage, T. Stieglitz, and J. Rickert,
First long term in vivo study on subdurally implanted micro-ECoG electrodes, manufactured
with a novel laser technology. Biomed Microdevices 13 (2011) 59-68.

[9] D.-W. Park, A.A. Schendel, S. Mikael, S.K. Brodnick, T.J. Richner, J.P. Ness, M.R. Hayat, F.
Atry, S.T. Frye, R. Pashaie, S. Thongpang, Z. Ma, and J.C. Williams, Graphene-based
carbon-layered electrode array technology for neural imaging and optogenetic applications.
Nature Communications 5 (2014) 5258.

[10] S. Thongpang, T.J. Richner, S.K. Brodnick, A. Schendel, J. Kim, J.A. Wilson, J. Hippensteel, L.
Krugner-Higby, D. Moran, A.S. Ahmed, D. Neimann, K. Sillay, and J.C. Williams, A Micro-
Electrocorticography Platform and Deployment Strategies for Chronic BCI Applications. Clin
EEG Neurosci 42 (2011) 259-65.

[11] M. Spuler, A. Walter, A. Ramos-Murguialday, G. Naros, N. Birbaumer, A. Gharabaghi, W.
Rosenstiel, and M. Bogdan, Decoding of motor intentions from epidural ECoG recordings in
severely paralyzed chronic stroke patients. J Neural Eng 11 (2014) 066008.

[12] A. Myrden, and T. Chau, Effects of user mental state on EEG-BCI performance. Front Hum
Neurosci 9 (2015).

[13] E. Fernández, M.H.U.o.E. Bioengineering Institute, Elche, Spain, Z. CIBER-BBN, Spain,
e.fernandez@umh.es, B. Greger, A.S.U. School of Biological and Health Systems
Engineering, Tempe, AZ, USA, P.A. House, U.o.U. Department of Neurosurgery, Salt Lake
City, UT, USA, I. Aranda, H.G.U. Department of Pathology, Alicante, Spain, C. Botella,
H.L.F. Department of Neurosurgery, Valencia, Spain, J. Albisua, F.J.D.a.H.R.J.C.
Department of Neurosurgery, Madrid, Spain, C. Soto-Sánchez, M.H.U.o.E. Bioengineering
Institute, Elche, Spain, Z. CIBER-BBN, Spain, A. Alfaro, M.H.U.o.E. Bioengineering
Institute, Elche, Spain, Z. CIBER-BBN, Spain, R.A. Normann, and U.o.U. Department of
Bioengineering, Salt Lake City, UT, USA, Acute human brain responses to intracortical
microelectrode arrays: challenges and future prospects. Frontiers in Neuroengineering 7
(2014).

[14] J.A. Uriguen, and B. Garcia-Zapirain, EEG artifact removal-state-of-the-art and guidelines. J
Neural Eng 12 (2015) 031001.

[15] W.M. Grill, S.E. Norman, and R.V. Bellamkonda, Implanted neural interfaces: biochallenges
and engineered solutions. Annu Rev Biomed Eng 11 (2009) 1-24.

[16] M. Goss-Varley, K.R. Dona, J.A. McMahon, A.J. Shoffstall, E.S. Ereifej, S.C. Lindner, and J.R.
Capadona, Microelectrode implantation in motor cortex causes fine motor deficit:
Implications on potential considerations to Brain Computer Interfacing and Human
Augmentation. Scientific Reports 7 (2017) 15254.

[17] J.D. Falcone, S.L. Carroll, T. Saxena, D. Mandavia, A. Clark, V. Yarabarla, and R.V.
Bellamkonda, Correlation of mRNA Expression and Signal Variability in Chronic
Intracortical Electrodes. Front Bioeng Biotechnol 6 (2018).

[18] C. Marin, I.o.B. University Miguel Hernandez, Elche, Spain, cmarin@umh.es, E. Fernandez,
B.a.N. Center for Biomedical Research Network in Bioengineering, Elache, Spain, I.o.B.
University Miguel Hernandez, Elche, Spain, and e.fernandez@umh.es, Biocompatibility of

intracortical microelectrodes: current status and future prospects. Frontiers in
Neuroengineering 3 (2010).

[19] J.W. Salatino, K.A. Ludwig, T.D.Y. Kozai, and E.K. Purcell, Glial responses to implanted
electrodes in the brain. Nature Biomedical Engineering 1 (2017) 862.

[20] A.J. Shoffstall, J.E. Paiz, D.M. Miller, G.M. Rial, M.T. Willis, D.M. Menendez, S.R. Hostler,
and J.R. Capadona, Potential for thermal damage to the blood-brain barrier during
craniotomy: implications for intracortical recording microelectrodes. J Neural Eng 15 (2018)
034001.

[21] C. Onal, H. Otsubo, T. Araki, S. Chitoku, A. Ochi, S. Weiss, I. Elliott, O.C. Snead, 3rd, J.T.
Rutka, and W. Logan, Complications of invasive subdural grid monitoring in children with
epilepsy. J Neurosurg 98 (2003) 1017-26.

[22] A.D. Degenhart, J. Eles, R. Dum, J.L. Mischel, I. Smalianchuk, B. Endler, R.C. Ashmore, E.C.
Tyler-Kabara, N.G. Hatsopoulos, W. Wang, A.P. Batista, and X.T. Cui, Histological
Evaluation of a Chronically-implanted Electrocorticographic Electrode Grid in a Non-human
Primate. J Neural Eng 13 (2016) 046019.

[23] G. Yang, F. Pan, C.N. Parkhurst, J. Grutzendler, and W.B. Gan, Thinned-skull cranial window
technique for long-term imaging of the cortex in live mice. Nat Protoc 5 (2010) 201-8.

[24] A.Y. Shih, C. Mateo, P.J. Drew, P.S. Tsai, and D. Kleinfeld, A Polished and Reinforced
Thinned-skull Window for Long-term Imaging of the Mouse Brain, J Vis Exp, 2012.

[25] D.E. Bonder, and K.D. McCarthy, Astrocytic Gq-GPCR-Linked IP3R-Dependent Ca2+
Signaling Does Not Mediate Neurovascular Coupling in Mouse Visual Cortex In Vivo, J
Neurosci, 2014, pp. 13139-50.

[26] G. Yang, F. Pan, C.N. Parkhurst, J. Grutzendler, and W.-B. Gan, Thinned-skull cranial window
technique for long-term imaging of the cortex in live mice. Nature Protocols 5 (2010) 201.

[27] M. Akhtari, H.C. Bryant, A.N. Mamelak, E.R. Flynn, L. Heller, J.J. Shih, M. Mandelkern, A.
Matlachov, D.M. Ranken, E.D. Best, M.A. DiMauro, R.R. Lee, and W.W. Sutherling,
Conductivities of three-layer live human skull. Brain Topogr 14 (2002) 151-67.

[28] P.J. Drew, A.Y. Shih, J.D. Driscoll, P.M. Knutsen, P. Blinder, D. Davalos, K. Akassoglou, P.S.
Tsai, and D. Kleinfeld, Chronic optical access through a polished and reinforced thinned
skull. Nature Methods 7 (2010) 981.

[29] A.A. Schendel, S. Thongpang, S.K. Brodnick, T.J. Richner, B.D. Lindevig, L. Krugner-Higby,
and J.C. Williams, A cranial window imaging method for monitoring vascular growth around
chronically implanted micro-ECoG devices. J Neurosci Methods 218 (2013).

[30] D.-W. Park, S.K. Brodnick, J.P. Ness, F. Atry, L. Krugner-Higby, A. Sandberg, S. Mikael, T.J.
Richner, J. Novello, H. Kim, D.-H. Baek, J. Bong, S.T. Frye, S. Thongpang, K.I. Swanson,
W. Lake, R. Pashaie, J.C. Williams, and Z. Ma, Fabrication and utility of a transparent
graphene neural electrode array for electrophysiology,. Nature Protocols 11 (2016) 2201.

[31] T.J. Richner, S. Thongpang, S.K. Brodnick, A.A. Schendel, R.W. Falk, L.A. Krugner-Higby, R.
Pashaie, and J.C. Williams, Optogenetic micro-electrocorticography for modulating and
localizing cerebral cortex activity. J Neural Eng 11 (2014) 016010.

[32] D.J. McFarland, L.M. McCane, S.V. David, and J.R. Wolpaw, Spatial filter selection for EEG-
based communication. Electroencephalogr Clin Neurophysiol 103 (1997) 386-94.

[33] K.A. Ludwig, R.M. Miriani, N.B. Langhals, M.D. Joseph, D.J. Anderson, and D.R. Kipke,
Using a Common Average Reference to Improve Cortical Neuron Recordings From
Microelectrode Arrays, J Neurophysiol, 2009, pp. 1679-89.

[34] F.A. Bazley, C. Hu, A. Maybhate, A. Pourmorteza, N. Pashai, N.V. Thakor, C.L. Kerr, and A.H.
All, Electrophysiological evaluation of sensory and motor pathways after incomplete
unilateral spinal cord contusion. J Neurosurg Spine 16 (2012) 414-23.

[35] K. Sakatani, H. Iizuka, and W. Young, Somatosensory evoked potentials in rat cerebral cortex
before and after middle cerebral artery occlusion. Stroke 21 (1990) 124-32.

[36] J.C. Williams, J.A. Hippensteel, J. Dilgen, W. Shain, and D.R. Kipke, Complex impedance
spectroscopy for monitoring tissue responses to inserted neural implants. J Neural Eng 4
(2007) 410-23.

[37] W. Wang, C.K. Kim, and A.Y. Ting, Molecular tools for imaging and recording neuronal
activity. Nature Chemical Biology 15 (2019) 101.

[38] K.A. Ludwig, J.D. Uram, J. Yang, D.C. Martin, and D.R. Kipke, Chronic neural recordings
using silicon microelectrode arrays electrochemically deposited with a poly(3,4-
ethylenedioxythiophene) (PEDOT) film. J Neural Eng 3 (2006) 59-70.

[39] A.A. Schendel, M.W. Nonte, C. Vokoun, T.J. Richner, S.K. Brodnick, F. Atry, S. Frye, P.
Bostrom, R. Pashaie, S. Thongpang, K.W. Eliceiri, and J.C. Williams, The effect of micro-
ECoG substrate footprint on the meningeal tissue response. J Neural Eng 11 (2014) 046011.

[40] D.K. Reasoner, M.M. Todd, F.L. Scamman, and D.S. Warner, The incidence of
pneumocephalus after supratentorial craniotomy. Observations on the disappearance of
intracranial air. Anesthesiology 80 (1994) 1008-12.

[41] H.T. Xu, F. Pan, G. Yang, and W.B. Gan, Choice of cranial window type for in vivo imaging
affects dendritic spine turnover in the cortex. Nat Neurosci 10 (2007) 549-51.

9.0   Figure Legends

Figure 1: (a) Diagram illustrating implantation placement of bilateral 32 channel µECoG and
connector scheme on thinned bone surface over sensorimotor cortex. **(b)** Polyimide based platinum
µECoG array with 16 channels (750 µm spacing, 250 µm site diameter). **(c)** Surgical photograph of
bilateral Parylene C based platinum µECoG array being placed over a thinned skull (top of
photograph) and on the dural surface (bottom of photograph).

Figure 2: (a) Illustration of the placement of a µECoG electrode array over a thinned skull with
optical fiber positioning. (b) Optogenetic stimulation of cortex with optical fiber placed over a
µECoG array and either thinned skull or dural surface. (c) Parylene C based platinum µECoG array
with 16 channels (500 µm spacing, 150 µm site diameter) and ZIF connector.

Figure 3: Somatosensory evoked potentials (SSEPs) recorded on day 20 post-implantation from a
16-channel µECoG array placed over thinned skull and left hemisphere sensorimotor cortex in a
chronically implanted rat. Biphasic current pulses (1 ms, varied amplitude) were used to stimulate
the right hindlimb with surface electrodes over the sciatic nerve. (a) Normal skull screw, (b)
Common Average and (c) Small Laplacian referencing strategies are shown to increase the signal-to-
noise ratio, and to reveal spatial signaling from the predicted hindlimb anatomical region.

527
Figure 4: Somatosensory evoked potentials (SSEPs) on day 38 post-implantation with small
Laplacian referencing from forelimb and hindlimb electrical surface stimulation using a 16-channel
µECoG array placed over a thinned skull portion of rat sensorimotor cortex. Plots represent spatial
recordings from the same electrode array, demonstrating LFPs from (a) biphasic forelimb stimulation
and (b) monophasic hindlimb stimulation. Stimuli were applied for 1ms at 1.25mA. Activity is
represented by 2D interpolated heat maps. The portions closer to the red spectrum show evoked
activity higher than baseline when averaged over at least 25 trials, and closer to blue shows negative
activity. The *x* scale bar, 20 ms; *y* scale bar, 20 µV.

Figure 5: (a) Electrical impedance spectral data at 1kHz from thinned skull (blue) and epidurally
(red) implanted µECoG electrodes. Each interpolation curve represents three animals in per group,
and 32 electrode sites per animal. Individual data points represent individual electrode site
impedance spectra measurements.
541
542 **Figure 6:** Optically evoked local field potentials from thinned skull (a) and epidurally (b) implanted
543 µECoG arrays in a Thy1-ChR2 mouse. Amplitude heat maps show the 545.5 mW/mm2 optically
544 evoked potentials referenced using the small Laplacian from both the (c) thin skull and (d) epidural
545 preparations can each be processed to show spatial resolution, although the difference in scale is
546 smaller in the thinned skull preparation by approximately a magnitude of 10.
547
548 **Figure 7:** Photostimulus duration versus amplitude peak potential 2D interpolated contour plot.
549 Stimulus strength is plotted against stimulus duration and peak depolarization amplitude curves in µV
550 are shown for (a) epidural and (b) thinned skull micro-ECoG recordings in a Thy1-ChR2 mouse.
551 Longer stimulus durations and stimulus strength (power) are needed to evoke similar sized neural
552 signal amplitudes in the thinned skull versus epidural preparations.
553

a Normal   b Common Average   c Small Laplacian

50 μV
20 ms

10 μV
20 ms

5 μV
20 ms

2.5mA —— 2.0mA —— 1.5mA —— Stimulus ——

a Left Forelimb

b Left Hindlimb

a

connector

optical fiber

electrode array

bone

brain

b

c

1 mm