The effect of diffusion on asymmetric spin echo based quantitative BOLD:

An investigation of the origin of deoxygenated blood volume overestimation

Alan J Stone[1], Naomi C Holland[2], Avery J L Berman[3], Nicholas P Blockley[1*]

[1]Wellcome Centre for Integrative Neuroimaging, FMRIB, Nuffield Department of Clinical

Neurosciences, University of Oxford, Oxford, UK, [2]Department of Physics, University of

Oxford, Oxford, UK, [3]Athinoula A. Martinos Center for Biomedical Imaging, Department of

Radiology, Harvard Medical School, Massachusetts General Hospital, Charlestown, MA

Running Title: The effect of diffusion on ASE qBOLD

Address for correspondence: [*]Nicholas Blockley, School of Life Sciences, The University

of Nottingham Medical School, Queen's Medical Centre, Nottingham, NG7 2UH, UK.

Telephone No.: +44 115 8230146

E-mail: nicholas.blockley@nottingham.ac.uk

Twitter: @nicblockley

Word Count: 5949

Figures: 8 figures and 1 table (+ 7 supplementary figures)

References: 33

Funding Acknowledgements: This work was supported by the Engineering and Physical

Sciences Research Council [grant number EP/K025716/1].

Abstract

Quantitative BOLD (qBOLD) is a technique for mapping oxygen extraction fraction (OEF)

and deoxygenated blood volume (DBV) in the human brain. Recent measurements using an

asymmetric spin echo (ASE) based qBOLD approach produced estimates of DBV which

were systematically higher than measurements from other techniques. In this study, we

investigate two hypotheses for the origin of this DBV overestimation and consider the

implications for *in vivo* measurements. Investigations were performed by combining Monte

Carlo simulations of extravascular signal with an analytical model of the intravascular signal.

Hypothesis 1: DBV overestimation is due to the presence of intravascular signal

which is not accounted for in the analysis model. Intravascular signal was found to have a

weak effect on qBOLD parameter estimates.

Hypothesis 2: DBV overestimation is due to the effects of diffusion which aren't

accounted for in the analysis model. The effect of diffusion on the extravascular signal was

found to result in a vessel radius dependent variation in qBOLD parameter estimates. In

particular, DBV overestimation peaks for vessels with radii from 20 to 30 μm and is OEF

dependent. This results in the systematic underestimation of OEF.

In vivo implications: The impact on *in vivo* qBOLD measurements was investigated

by simulating a distribution of vessel sizes with a small number of discrete radii.

Overestimation of DBV consistent with previous experiments was observed, which was also

found to be OEF dependent. This results in the progressive underestimation of the measured

OEF. Furthermore, the relationship between the measured OEF and the true OEF was found

to be dependent on echo time and spin echo displacement time.

The results of this study demonstrate the limitations of current ASE based qBOLD

measurements and provides a foundation for the optimisation of future acquisition

approaches.

1 Introduction

The quantitative BOLD (qBOLD) technique is a relaxometry based approach for mapping

oxygen extraction fraction (OEF) and deoxygenated blood volume (DBV) in the human brain

(He and Yablonskiy, 2007). An elevated OEF is indicative of tissue at risk of infarction, such

as the penumbral tissue surrounding the core infarct of an ischaemic stroke (Astrup et al.,

1981). When combined with a measurement of cerebral blood flow (CBF), the cerebral

metabolic rate of oxygen consumption ($CMRO_2$) can also be estimated (Kety and Schmidt,

1948). Since qBOLD can provide this valuable information in a non-invasive and rapidly

acquired manner, it has a great deal of potential for providing these quantitative physiological

measurements in clinical research applications.

The analytical model used to analyse qBOLD data assumes that the signal decay

behaves as though it were in the static dephasing regime (SDR) i.e. the diffusion of water in

tissue does not influence the signal decay due to magnetic field inhomogeneity (Yablonskiy

and Haacke, 1994). However, simulations of the Gradient Echo Sampling of Spin Echo

(GESSE) pulse sequence, which is often used to acquire qBOLD data, have shown that this is

not the case and that diffusion introduces a vessel size dependent effect on the signal decay

(Dickson et al., 2010; Pannetier et al., 2014). However, qBOLD data can also be acquired

using the Asymmetric Spin Echo (ASE) pulse sequence, and it is unclear whether a similar

effect is observed in these experiments (An and Lin, 2003; Stone and Blockley, 2017).

Interestingly estimates of DBV made using this ASE based acquisition are systematically

higher than those reported for GESSE based measurements (He and Yablonskiy, 2007),

suggesting that different effects may be at play.

The overestimation of DBV by ASE based qBOLD is at least partially responsible for

the underestimation of the OEF (Stone and Blockley, 2017). This overestimation has

previously been suggested to be due to the presence of intravascular blood signal, which is

not accounted for in the analytical qBOLD model, with flow crushing gradients proposed as a

solution (An and Lin, 2003). However, since it has been shown that diffusion results in

additional signal attenuation (Dickson et al., 2010), which is similarly unaccounted for in the

analytical qBOLD model, this may also provide a mechanism for DBV overestimation.

In this study, we investigate both mechanisms to discover whether either can account

for the overestimation of DBV in ASE based qBOLD. The effect of diffusion on the

extravascular tissue signal was examined using Monte Carlo simulations (Boxerman et al.,

1995a) and the intravascular blood signal was simulated using a recently published analytical

model (Berman and Pike, 2018). Whilst these effects are initially considered using

simulations with vessels of a single radius, these results are also integrated using an *in vivo*

vessel size distribution to investigate sources of systematic error in real world measurements.

Theory

Transverse signal decay results from dephasing of the net magnetisation due to the presence

of magnetic field inhomogeneity at multiple scales. The effect of these scales on the qBOLD

signal can be considered with reference to a spin echo pulse sequence. At the microscopic

scale spins experience local magnetic field inhomogeneities caused by neighbouring spins

that are rapidly varying. Due to this rapid magnetic field variation, the phase evolution cannot

be rewound by the application of a refocussing pulse. The resulting signal decay is described

by the irreversible transverse relaxation rate R_2. The macroscopic scale describes magnetic

field inhomogeneity on the scale of the head i.e. due to the nasal sinuses or ear canals. This

effect can be reversed by a refocussing pulse due to its static nature, enabling the phase

evolution in the period before the refocussing pulse to be rewound by the time the spin echo

is formed. At this scale the signal decay is described by the reversible relaxation rate $R_2{}'$ and

referred to as the SDR. At intermediate scales, often referred to as the mesoscopic scale,

diffusion becomes increasingly important as the so called Diffusion Narrowing Regime is

approached. The transition between these regimes is dependent on the scale of the magnetic

field inhomogeneity and the distance the spin travels due to diffusion. More precisely, the

characteristic diffusion time ($\tau_d \propto r^2/D$), which is dependent on the radius (r) of the

deoxygenated blood vessel and the diffusion coefficient (D), is on the order of the time taken

by a water molecule to diffuse a distance equivalent to the radius of the vessel (Yablonskiy

and Haacke, 1994). This results in an averaging of the magnetic field distribution surrounding

the vessels and a loss of phase history, meaning that signal cannot be efficiently recovered by

a refocussing pulse.

<u>Modelling the qBOLD signal</u>

The qBOLD model relies on the known relationship between R_2' and the baseline OEF, E_0,

and deoxygenated blood volume, V_0, for a network of randomly oriented blood vessels

approximated as infinite cylinders (Yablonskiy and Haacke, 1994),

$$R_2' = \frac{4}{3}\pi\,\gamma\,B_0\,\Delta\chi\,V_0\,Hct\,E_0 \tag{1}$$

where γ is the proton gyromagnetic ratio, B_0 is the main magnetic field, $\Delta\chi$ is the difference

in volume magnetic susceptibility between fully oxygenated and fully deoxygenated blood

and Hct is the haematocrit. However, further modelling has shown that the R_2'-weighted

signal is not purely monoexponential, displaying a quadratic exponential behaviour around

the spin echo. Under the assumption that the signal decays in the SDR, the following solution

has been found,

$$S(\tau) = S_0\,e^{-t_E\,R_2}e^{-Vf_c(\tau\,R_2'/V)} \tag{2}$$

$$f_c(\tau\,R_2'/V) = \frac{1}{3}\int_0^1 du\,(2+u)\,\sqrt{1-u}\,\frac{1 - J_0\,(1.5\,\tau\,(R_2'/V)\,u)}{u^2} \tag{3}$$

1 where t_E is the echo time, τ is the spin echo displacement time and J_0 is the zeroth-order

2 Bessel function (He and Yablonskiy, 2007). For simplicity this continuous function can be

3 broken down into two asymptotic solutions and applied piece wise (An and Lin, 2000).

$$S^S(\tau) = S_0 e^{-t_E\,R_2} e^{-0.3\,\left(\tau\,R_2'\right)^2/V}, \qquad \tau < \frac{1.5\,V}{R_2'} \tag{4}$$

$$S^L(\tau) = S_0 e^{-t_E\,R_2} e^{-\tau\,R_2'} e^{V}, \qquad \tau > \frac{1.5\,V}{R_2'} \tag{5}$$

4 In the long τ regime (Eq. (5)) the signal decay takes a monoexponential form, whilst in the

5 short τ regime (Eq. (2)) the signal follows a quadratic exponential form. A log-linear fit to

6 long τ data enables R_2' to be estimated. Furthermore, comparison of the measured signal at

7 τ=0 ($S_{meas}^S(0)$) with the intercept extrapolated from long τ data ($S_{extrap}^L(0)$) enables V_0 to be

8 calculated.

$$V_0 = \ln S_{extrap}^L(0) - \ln S_{meas}^S(0) \tag{6}$$

Henceforth we will refer to this as the SDR qBOLD model.

<u>Simulating the effect of diffusion</u>

Monte Carlo simulations of the qBOLD signal were performed by repeating the following

three steps for each simulated proton.

Step 1: Generate a system of vessels. The vessel system was defined as a sphere with

radius R_s. Vessel origin points (O) were randomly selected, with half placed on the surface of

the sphere and half within the sphere (Dickson et al., 2010). A uniform distribution of points

over the surface of the sphere was ensured by generating a unit vector (X_i) from a normally

distributed random number generator (mean 0, standard deviation 1) and scaling by R_s

(Muller, 1959). Within the sphere, uniform density was maintained by taking account of the

increased volume occupied by points far from the centre of the system. This scaling factor, U,

is selected from a uniform distribution of random numbers (range 0 to 1).

$$(O_1, O_2, O_3) = \begin{cases} R_s \, \dfrac{(X_1, X_2, X_3)}{\sqrt{X_1^2 + X_2^2 + X_3^2}}, & \text{on sphere surface} \\[4mm] R_s \, U \, \dfrac{(X_1, X_2, X_3)}{\sqrt{X_1^2 + X_2^2 + X_3^2}}, & \text{within sphere} \end{cases} \tag{7}$$

Vessels are modelled as randomly oriented infinitely long cylinders with a single radius, R_c,

placed at the vessel origin points described by Eq. (7) and added until the target volume

fraction (V_f) is reached. Random orientation was ensured by generating a unit vector from a

normally distributed random number generator (mean 0, standard deviation 1).

Step 2: Proton random walk. Protons are initially placed at the centre of the vessel

system. Each step taken by the proton is independently selected along each dimension from a

normal distribution of random numbers with mean 0 and standard deviation σ with diffusion

coefficient, D, and time interval between steps, Δt.

$$\sigma = \sqrt{2 \, D \, \Delta t} \tag{8}$$

9   Step 3: Estimate the phase accrued at each step. The phase, $\Delta\phi$, accumulated by the

10 proton during each time interval is calculated by summing over the field contributions from

11 all N vessels (Boxerman et al., 1995a),

$$\Delta\phi = 2\pi \, \gamma \, B_0 \, \Delta t \, (1 - Y) \, Hct \, \Delta\chi \sum_{i=1}^{N} \left(\frac{R_c}{r_i}\right)^2 \cos 2\varphi_i \sin^2 \theta_i \, , \qquad r \geq R \tag{9}$$

where θ is the angle of the vessel with respect to B_0, φ is the angle with respect to the

projection of B_0 onto a plane orthogonal to the vessel, r_i is the perpendicular distance to the

vessel and Y is the blood oxygen saturation. Only the equation for the magnetic field outside

of the vessel is presented, since only extravascular signal was simulated.

By appropriate combination of the phase accrued in each interval it is possible to

simulate the phase evolution of the ASE and GESSE pulse sequences as a function of τ,

$\phi(\tau)$.

$$\Phi(\tau) = \sum_{j=1}^{m} \Delta\phi_j - \sum_{j=m+1}^{n} \Delta\phi_j \qquad (10)$$

where m defines the transition from signal decay to signal recovery due to the refocussing

pulse, n is the point at which the signal is acquired and where $0 \le m \le n$. For ASE $m =$

$(t_E - \tau) / 2 \, \Delta t$ and $n = t_E / \Delta t$, whilst for GESSE $m = t_{SE} / 2 \, \Delta t$ and $n = (t_{SE} + \tau) / \Delta t$.

Here, t_E is defined as the timing of the centre of the readout and t_{SE} is the time at which the

spin echo forms (see Fig. 1). These definitions reflect an important distinction between the

ASE and GESSE pulse sequences, whereby t_E is fixed for ASE and variable for GESSE

whilst t_{SE} is variable for ASE and fixed for GESSE.

The phase evolution of P protons is then summed to simulate the decay of the

extravascular ASE or GESSE signal (Boxerman et al., 1995a),

$$S_{EV}(t_E, \tau) = \left| \frac{1}{P} \sum_{k=1}^{P} e^{i\,\Phi(\tau)} \right| e^{-\frac{t_E}{T_{2,t}}} \qquad (11)$$

where $T_{2,t}$ is the underlying tissue T_2.

Intravascular signal has traditionally been difficult to simulate, with empirical

measurements of blood R_2 and R_2^* commonly used (Griffeth and Buxton, 2011). However,

simulating the R_2'-weighted signal using the difference between R_2 and R_2^* is likely to be

inaccurate in the short τ regime. Recently an analytical model of the blood signal during a

Carr-Purcell Meiboom-Gill (CPMG) pulse sequence was extended to capture the signal

evolution between an arbitrary number of spin echoes (Berman and Pike, 2018), i.e. the

conditions that exist for ASE and GESSE pulse sequences. Using this model, the

intravascular signal, S_{IV}, is described by,

$$S_{IV}(t_E, \tau) = \exp\left\{ -\frac{\gamma^2}{2} G_0 \tau_D^2 \left[\frac{t_E}{\tau_D} + \left(\frac{1}{4} + \frac{t_E}{\tau_D} \right)^{\frac{1}{2}} + \frac{3}{2} - 2\left(\frac{1}{4} + \frac{t_E - t_{SE}/2}{\tau_D} \right)^{\frac{1}{2}} - \right. \qquad (12)$$

$$2\left(\frac{1}{4}+\frac{t_{SE}/2}{\tau_D}\right)^{\frac{1}{2}}\Bigg]\Bigg\}\exp\left(-\frac{t_E}{T_{2,b|0}}\right).$$

Here $\tau_D=R_{rbc}^2/D_b$, where R_{rbc} is the characteristic size of red blood cells and D_b is the

diffusion coefficient of blood, $T_{2,b|0}$ is the intrinsic T_2 of blood and G_0 is the mean square field

inhomogeneity in blood,

$$G_0 = \frac{4}{45}Hct\,(1-Hct)(4\,\pi\,\Delta\chi\,(0.95-Y)\,B_0)^2. \tag{13}$$

The value of t_{SE} is fixed for GESSE but is variable for ASE with $t_{SE}=t_E-\tau$. By definition

t_E is fixed for ASE and varying for GESSE.

Finally, the total signal, S_{TOT}, is calculated by taking a volume weighted sum of the

intra- and extravascular signals.

$$S_{TOT} = (1-V_f)\,S_{EV} + V_f\,S_{IV} \tag{14}$$

Methods

Simulations

Simulations of the tissue signal were performed following the theory outlined above. Firstly,

extravascular signal decay was simulated using Monte Carlo simulations (B_0=3 T,

γ=267.5×10^6 rad s^{-1} T^{-1}). The radius of the spherical system of vessels, U, was chosen to

maintain a similar number of vessels, N, regardless of the vessel radius (N~1,300). For each

proton, a complete random walk was generated with a step size, Δt, of 20 μs, which was

downsampled to 200 μs, and D=1 μm^{2}ms^{-1}. The perpendicular distance, r_i, to each vessel in

the system was then calculated. For protons that passed close to vessels, defined as $R_c^2/r_i^2 >$

0.04, the perpendicular distance was recalculated using the original 20 μs time step to better

sample the rapid magnetic field variation expected close to vessels (Dickson et al., 2010).

Walks that moved the proton inside a vessel were flagged to be discarded in order to simulate

non-permeable blood vessels. The phase of each proton was allowed to evolve for 120 ms

after the excitation with $\Delta\chi$=0.27 ppm (Spees et al., 2001). Phase accrual was stored for each

proton in 2 ms intervals, Δt. A new system of vessels was generated for each proton and a

total of 10,000 protons were simulated for each vessel radius investigated. However, the

number of protons that passed within a vessel increased rapidly for smaller vessel radii.

Therefore, only the first P=5,000 protons that did not pass within a vessel were used to

calculate S_{EV} using Eq. (11) with $T_{2,t}$=80 ms. Secondly, intravascular signal decay was

simulated using Eqs. (12) and (13), which are independent of vessel radius. Based on

previous work the following parameters were used (Berman et al., 2017): $T_{2,b|0}$=189 ms,

R_{rbc}=2.6 μm and D_b=2 $\mu m^2 ms^{-1}$. The total signal was then calculated using Eq. (14).

Whilst the intravascular simulations are rapid to perform, Monte Carlo simulations of

the extravascular signal are time consuming. Therefore, the following approaches were taken

to accelerate these simulations, with examples presented as supplementary figures. We have

previously shown that different oxygenation levels can be simulated by scaling the accrued

phase of a nominal oxygenation value by the target value (Blockley et al., 2008). This is

made possible by saving the phase of each proton and the fact that phase is a linear function

of blood oxygenation for a network of vessels with the same oxygenation (Fig. S1). Different

volume fractions can be simulated from the signal magnitude generated by Eq. (11). It has

been shown that the extravascular signal, S_{EV}, can be described as a radius dependent shape

function, $f(R_c, \tau)$, scaled by the volume fraction (Dickson et al., 2011; Kiselev and Posse,

1999) (Fig. S2).

$$S_{EV}(R_c, \tau) = \exp\left[-V_f(R_c)f(R_c, \tau)\right] \tag{15}$$

22 Finally, it is possible to simulate the effect of a system with multiple vessel radii by

23 combining multiple single vessel radius simulations of the extravascular signal (Dickson et

24 al., 2011; Kiselev and Posse, 1999). The resulting combined signal, S_{EV}^{MULTI}, can be calculated

1   as the product of the signals of M single vessel simulations which have already been scaled

2   for blood oxygenation and volume fraction as described above (Fig. S3).

$$S_{EV}^{MULTI} = \prod_{k=1}^{M} S_{EV}(k) \tag{16}$$

Parameter quantification

The following framework was used to quantify the parameters of the qBOLD model from the

simulated decay curves. The parameters of the SDR qBOLD model (R_2' and DBV) were

organised as a vector of unknowns (x) in a linear system (A·x=B) (Stone et al., 2019). The

first row of the matrix A represents Eq. (4) when τ=0 with subsequent rows representing Eq.

(5) with values of τ beyond the transition between the quadratic and linear exponential

regime. In this case only values of τ greater than 15 ms were used. Vector B contains the

ASE signals, $S(\tau)$.

$$\begin{bmatrix} 0 & 0 & 1 \\ 1 & -\tau_1 & 1 \\ 1 & -\tau_2 & 1 \\ \vdots & \vdots & \vdots \\ 1 & -\tau_n & 1 \end{bmatrix} \begin{bmatrix} V_0 \\ R_2' \\ \log(S_0) - t_E \cdot R_2 \end{bmatrix} = \begin{bmatrix} \log(S(0)) \\ \log(S(\tau_1)) \\ \log(S(\tau_2)) \\ \vdots \\ \log(S(\tau_n)) \end{bmatrix} \tag{17}$$

12 Parameters were estimated via Eq. (17) using the least square solution, with the error in each

13 parameter determined from the covariance matrix. Finally, OEF can be estimated by

14 rearranging Eq. (1).

$$E_0 = \frac{3 \cdot R_2'}{4\pi \cdot \gamma B_0 \cdot \Delta\chi_0 \cdot Hct \cdot V_0} \tag{18}$$

16 Effect of diffusion on ASE measurements

17 Initial simulations were performed for a selection of vessel radii (R_c=5, 10, 50, 1000 µm), a

18 venous Y of 60%, a Hct of 40% and a DBV of 3%. Simulations of the ASE pulse sequence

19 were performed with t_E=60 ms and -60 ms ≤ τ ≤ 60 ms for both extra- and intravascular

signal, where τ=60 ms corresponds to pure gradient echo decay. For validation purposes,

similar simulations were performed for the GESSE pulse sequence using t_{SE}=60 ms and -30

ms $\leq \tau \leq$ 60 ms.

Effect of diffusion on qBOLD parameters

A further set of synthetic ASE signal decay curves were generated for vessel radii

logarithmically spaced between 1 and 1,000 μm. All other parameters were set consistent

with previous qBOLD measurements (Stone and Blockley, 2017). In the context of these

simulations this required t_E=80 ms with τ=0 and τ=16 to 64 ms in 4 ms steps. The *apparent*

value of R$\square'$, DBV and OEF were then estimated using Eq. (17) and (18). The effect of

diffusion on the estimation of qBOLD parameters was investigated by first fixing OEF and

varying DBV and then by fixing DBV and varying OEF. In the former case a fixed OEF of

40% was coupled with DBV values of 1, 3 and 5%, whilst in the latter case DBV was fixed at

3% and OEF took values of 20, 40 and 60%. These values are considered to be the *true*

parameters in both cases. The results of varying DBV were also used to consider the

percentage error in DBV as a function of vessel radius. In these single vessel simulations

arterial blood is assumed to have an oxygen saturation, *Y*, of 100% hence the venous

saturation, $Y_v = 1 - E_0$.

The effect of intravascular signal on qBOLD parameter estimates was investigated by

repeating these simulations, but excluding the intravascular compartment. In this way it was

possible to quantify the percentage of the parameter estimate (PE) which results from the

presence of intravascular signal i.e. $(PE_{EV} - PE_{EV+IV})/PE_{EV+IV}$.

Further investigation of the effect of diffusion on DBV estimates was pursued based

on a consideration of Eq. (6), which suggests that errors must be due to either the signal

measured at τ=0 ($S_{meas}^S(0)$) or the extrapolated estimate of the signal at τ=0 from the R$_2'$ fit

$(S_{extrap}^{L}(0))$, or both. However, given the analysis approached represented by Eq. (17)

$S_{extrap}^{L}(0)$ is not estimated and $S_{meas}^{S}(0)$ is confounded by T_2 decay. The latter was corrected

by calculating the signal decay relative to the value at R=1,000 µm, where previous

simulations would suggest the SDR applies and hence signal attenuation should be zero. The

former was estimated by subtracting this relative measure of $S_{meas}^{S}(0)$ from the estimated

value of apparent DBV.

Effect of an *in vivo* vessel radius distribution

The *in vivo* effect of a distribution of vessel radii was investigated by integrating the results

from single radius simulations. A compartmental model of the vasculature derived from the

morphology of the sheep brain was selected (Sharan et al., 1989). This model has five orders

of arterial and venous vessels, with a range of radii, and a capillary compartment with a

single vessel radius (Table 1). Additional Monte Carlo simulations for this range of vessel

radii were performed and combined using the acceleration techniques described above.

Arterial vessels were assigned an arterial oxygen saturation, Y_a, of 98%, which was used to

calculate the venous saturation, Y_v, for a given OEF.

$$Y_v = Y_a(1 - E_0) \tag{19}$$

17   The capillary compartment was an intermediate oxygen saturation, Y_c, calculated as an

18   average of the arterial and venous saturations weighted by a factor, κ, equal to 0.4

19   representing a weighting towards the venous saturation (Griffeth and Buxton, 2011; Tsai et

20   al., 2003).

$$Y_c = \kappa Y_a + (1 - \kappa)Y_v \tag{20}$$

Relative blood volume fractions for each vessel type were calculated by estimating the

volume of each vessel radius population as cylinders with the properties described in Table 1.

These relative blood volume fractions were then scaled by the total cerebral blood volume

(CBV). Pairs of OEF and CBV values were drawn from a uniform random number generator

within the following ranges: OEF 0-100%, CBV 0-10%. The qBOLD parameters were

quantified for 1,000 random OEF-CBV pairs to examine the effect of diffusion across the

physiological range. In the absence of a strict definition of DBV, the ground truth was

assumed to be equal to the combined blood volume occupied by capillary and venous vessels.

This is therefore only a working assumption, since it is likely the *true* DBV is weighted by

blood oxygenation and vessel radius. Deoxyhaemoglobin content , dHb, was calculated based

on the same assumption for DBV and a value for the density of brain tissue $\rho=1.04$ g/ml

(Rempp et al., 1994) using the following equation.

$$dHb = 100 \, \frac{V_0}{\rho} \, \frac{Hct}{0.03} \, E_0 \qquad\qquad (21)$$

For comparison these simulations were also repeated for the original ASE based qBOLD

implementation with t_E=64 ms with τ=0 and τ=10 to 18 ms in 4 ms steps (An and Lin, 2003).

Details on how to access the simulation code, simulation results and analysis code

that underlie this study can be found in Appendix A.

Results

Figure 2 presents simulations of the signal generated by the ASE pulse sequence in the

absence of T_2 decay. The extravascular signal (Fig. 2a) was found to be symmetric with

respect to the spin echo (τ=0) regardless of vessel radius. Similarly, the intravascular signal

(Fig. 2b) was symmetric, but displayed a relatively weak signal decay as a function of τ. In

contrast, simulations of the GESSE pulse sequence demonstrated increasing asymmetry with

reducing vessel radius for the extravascular signal and strong asymmetry for the intravascular

signal (Fig. S4).

Figure 3 displays the effect of vessel size on the parameter estimates from the SDR

qBOLD model. The apparent values of R_2' plateau above a critical vessel radius of

approximately 40 μm (Fig. 3a,d) and are then consistent with predictions from the SDR

qBOLD model (dashed lines calculated using Eq. (1)). The apparent DBV is found to be

strongly dependent on vessel radius, peaking between 20 and 30 μm (Fig. 3b,e). Estimates of

the apparent OEF increase monotonically with vessel radius reaching the value predicted by

the SDR qBOLD model as the vessel radius approaches 1,000 μm (Fig. 3c,f). When the true

OEF was fixed whilst DBV was varied (Fig. 3c) estimates of apparent OEF were consistent

across DBV levels, suggesting that the error in DBV is a linear scale factor. Likewise, it can

be seen that the profile of apparent DBV when the true DBV was fixed and OEF was varied

(Fig. 3e) peak at different vessel radius values, suggesting that the error in DBV is OEF

dependent. Furthermore, this effect can be seen to result in a reduced dynamic range for the

estimates of apparent OEF as vessel size is reduced (Fig. 3f). Figure 4 confirms that the

percentage error in DBV is constant for a given combination of OEF and vessel radius (Fig.

4a), but differs for different OEF values (Fig. 4b).

Figure 5 considers the contribution of intravascular signal to the parameter estimates

in Fig. 3 as a function of vessel radius. This contribution is generally small for R_2' and DBV

at around ±1% for vessel radii greater than 10 μm. However, the intravascular signal appears

to reflect a larger contribution when OEF is low, conditions where qBOLD contrast is low.

Despite this the effect of the intravascular signal appears to be largely cancelled in the

estimation of OEF (Fig. 5c,f). A reproduction of Fig. 3 without intravascular signal is

included in the supplementary material for comparison and shows little discernible difference

by eye (Fig. S5).

Figure 6 investigates the origin of the DBV estimation error attributed either to an

error in the measured signal at τ=0 (orange markers) or an error in the intercept extrapolated

from long τ data (green markers). In the case of the former, $-\ln S_{meas}^{S}(0)$ is plotted such that

the sum of the two curves representing the apparent DBV (represented by grey shading).

When interpreting these curves, it is useful to consider the orange markers as a reflection of

the deviation of the spin echo from perfect refocusing (with positive values representing

increased signal attenuation) and the green markers as a reflection of the deviation of the

measured R_2' from the SDR qBOLD estimate of R_2'. The former is found to be subject to

increasing signal attenuation as vessel size is reduced, which is strongly affected by blood

oxygenation via OEF. The latter is found to plateau and is relatively consistent with the SDR

qBOLD model for vessel radii greater than approximately 20 μm.

Figure 7 explores the combined effect of a distribution of vessel radii on parameter

estimates from the SDR qBOLD model. The apparent R_2' is plotted against values of R_2'

predicted by the SDR model via Eq. (1), with DBV estimated according to the working

assumption described above (Fig. 7a). Data points are colour coded to reflect the true voxel

deoxyhaemoglobin content in $ml^{dHb}/100\ g^{tissue}$. A linear dependence is maintained, albeit with

a shallower gradient than predicted by the SDR qBOLD model. A large amount of

uncertainty is observed in estimates of apparent DBV over the large physiological range

tested (Fig. 7b), with data points colour coded by true OEF value. However, this level of

uncertainty does not propagate into estimates of apparent OEF (Fig 7c) where data points are

colour coded by true DBV. Apparent OEF increases monotonically between 0 and 50%, but

reaches a plateau for higher values, and is inappropriately scaled compared with the true

OEF. In a similar manner to Fig. 4, the percentage error in the apparent DBV can be plotted

as a function of true OEF (Fig. 8). As noted for the single vessel radius simulations, this error

is strongly OEF dependent.

These simulations were repeated for different ASE pulse sequence parameters,

namely variations in t_E and τ, and included in supplemental material. The results in Fig. S6

largely mirror those in Fig. 7 with the following variations. The slope of the relationship

between apparent R_2' and SDR qBOLD predicted R_2' is slightly reduced for the alternative

parameters (Fig. S6a). More noticeable is the reduction in the range of apparent DBV values

(Fig. S6b), with the error in the apparent DBV reduced by more than a half (Fig. S7). Whilst

the apparent OEF is also inappropriately scaled, the relationship with true OEF is more

monotonic in nature.

Discussion

In this study numerical simulations were used to investigate the effect of diffusion on ASE

based qBOLD measurements and the origin of DBV overestimation in such measurements. In

contrast to the previously observed shift of the GESSE signal maximum due to the effect of

diffusion, the ASE signal was observed to maintain its symmetry as vessel radius is reduced

and the effect of diffusion is increased. Two hypotheses for the origin of the observed DBV

overestimation were tested: (i) the effect of intravascular blood signal and (ii) the effect of

diffusion on the extravascular tissue signal. The presence of intravascular blood signal was

found to have a minor effect on qBOLD parameter estimates. It is therefore unlikely to be

responsible for the majority of the overestimation observed in DBV measurements. In

contrast, the extravascular signal was shown to have a very strong dependence on vessel

radius providing the potential for a large error in DBV and is considered to be the dominant

cause of DBV overestimation. Furthermore, the error in DBV is predicted to be blood oxygen

saturation level dependent. Integration of these single vessel radius simulations via an *in vivo*

vessel distribution revealed three main findings. Firstly, that the relationship between the

apparent R_2' and deoxyhaemoglobin content is retained. Secondly, there is an inherent

uncertainty in estimates of DBV. Finally, this uncertainty is not propagated to apparent OEF

estimates, but results in inappropriate scaling of these estimates. Furthermore, the monotonic

behaviour of the relationship between apparent and true OEF was found to be dependent on

the pulse sequence parameters t_E and τ. These results provide new directions for improving

the modelling of ASE qBOLD signal and the reduction of systematic error in parameter

estimates of OEF and DBV.

<u>Effect of diffusion on ASE measurements</u>

Whilst several studies have investigated the qBOLD signal as acquired by the GESSE pulse

sequence (Christen et al., 2014; Dickson et al., 2011, 2010; Pannetier et al., 2014), this study

considered whether the signal decay under an ASE acquisition behaves in the same way. One

particular characteristic of the GESSE pulse sequence concerns the maximum of the qBOLD

signal decay curve. This would ordinarily be expected to coincide with the spin echo ($\tau = 0$),

but has been shown to be shifted towards negative τ values (Fig. S4a) for the GESSE

sequence in the presence of diffusion (Dickson et al., 2010; Pannetier et al., 2014). However,

this effect is not observed in simulations of the extravascular signal acquired using an ASE

pulse sequence (Fig. 2a), where the signal maximum was found to be close to the spin echo

($\tau = 0$). However, the GESSE and ASE sequences differ in an important way. The t_E of each

successive τ value increases in the GESSE experiment and hence the time for protons to

diffuse around blood vessels increases. Whilst the t_E is constant for all τ values in the ASE

method and hence the time for diffusion is also constant. This would suggest that there is a t_E

dependent component of the R_2'-weighted signal decay. Such a component has previously

been included as a correction to estimates of R_2' (Berman et al., 2017).

20   This study also considered the R_2'-weighted contribution of the blood to the qBOLD

signal using a recently proposed model (Berman and Pike, 2018). In common with the

extravascular results, the ASE blood signal is symmetric with respect to the spin echo, but

decays far less as a function of τ (Fig. 2b). However, the signal is heavily attenuated at all τ

values compared with the extravascular simulations. This is in contrast to simulations of the

GESSE blood signal, which are highly shifted to negative τ values and present largely as an

exponential decay (Fig. S4b).

<u>Origin of DBV overestimation</u>

Simulations of the combined intravascular and extravascular signal revealed a vessel radius

dependent overestimation of DBV (Fig. 3b,e). The error in the apparent DBV was found to

be OEF dependent (Fig. 4). However, at larger radii (approaching 1 mm) estimates of DBV

were consistent with ground truth values. The contribution of intravascular signal to these

parameter estimates was performed by comparing simulations with (Fig. 3) and without (Fig.

S5) an intravascular compartment. A small and largely vessel radius independent effect (for

$R_c > 10$ μm) was observed (Fig. 5b,e). The effect of intravascular signal was more pronounced

for smaller vessel radii and low OEF, where the relative contribution of intravascular signal is

increased by weak extravascular contrast. Despite this, the overestimation of DBV is

dominated by the effect of diffusion on the extravascular signal.

15   Finally, Eq. (6) provides the opportunity to consider whether the systematic error in

DBV originates in the measurement of the spin echo ($S_{meas}^{S}(0)$), the intercept extrapolated

from the long τ regime ($S_{extrap}^{L}(0)$) or a combination of both, as explored in Fig. 6. For

vessel radii greater than 20 μm additional signal attenuation of $S_{meas}^{S}(0)$ is the main driver of

overestimation of DBV. However, for vessels with radii below 20 μm, errors in $S_{extrap}^{L}(0)$

provide an additional confound to DBV estimation. These results are consistent with the

characteristics of gradient echo versus spin echo BOLD vessel size sensitivity, which

correspond to $\ln S_{extrap}^{L}(0)$ and $\ln S_{meas}^{S}(0)$, respectively (Boxerman et al., 1995b). For the

smallest vessel radii the apparent R_2' is reduced relative to the value expected by the SDR

qBOLD model (Fig. 3a,d) due to diffusional narrowing, such that $\ln S_{extrap}^{L}(0)$ is also

reduced (Fig. 6). Similarly, additional unrecoverable signal decay due to diffusion narrowing

results in a decrease in the value of $\ln S_{meas}^S(0)$, which is analogous to an increase in

apparent R⬜ and is strongest for capillary sized vessels (Note that Fig. 6 plots

$-\ln S_{meas}^S(0)$). With increasing vessel radius, R⬜$'$ approaches the SDR qBOLD model

prediction and the value of $\ln S_{extrap}^L(0)$ approaches a constant value. Similarly the

attenuation of the spin echo is reduced as the SDR is approached and $\ln S_{meas}^S(0)$ reaches its

minimum. Therefore, when the differing profiles of these phenomena are combined the form

of the apparent DBV as a function of vessel radius can be described.

<u>Effect of an *in vivo* vessel radius distribution</u>

Having established the vessel radius dependence of the qBOLD signal, the implications for

the *in vivo* condition were considered. In order to integrate the single vessel radius results, a

vessel distribution with a small number of discrete vessel radii was selected. This enabled

different oxygenation levels to be associated with different vessel types. A wide

physiological range was investigated by randomly selecting pairs of OEF and CBV values.

The apparent R_2' was found to be tightly correlated with the R_2' predicted by SDR qBOLD

model (Fig. 7a). This is important as it demonstrates that the relationship between R_2' and the

voxel deoxyhaemoglobin content (proportional to the product of deoxyhaemoglobin

concentration and DBV) is maintained despite the effects of diffusion. It should therefore be

possible to quantify maps of R_2' in terms of deoxyhaemoglobin content with appropriate

scaling. Likewise with improved quantification of DBV, either through improvements to the

qBOLD technique or via an additional experimental technique (Blockley et al., 2013; Lee et

al., 2018), accurate measurements of OEF are possible. A large amount of uncertainty was

observed in the apparent DBV (Fig. 7b). This was demonstrated to be blood oxygenation

dependent i.e. a function of OEF (Fig. 8). This is consistent with the results of the single

vessel simulations (Fig. 4) and demonstrates the important contribution of smaller vessel

radii. This also explains why this uncertainty does not propagate into the apparent OEF, since

the percentage error in apparent DBV is constant at each OEF level (Fig. 7c). However, the

increasing percentage error in apparent DBV with OEF (Fig. 8) results in a progressive

underestimation of apparent OEF. A plateau in the apparent OEF limits the maximum

measured OEF to approximately 50%. Despite this the remaining range covers the majority

of the expected healthy physiological range (Marchal et al., 1992). These simulation were

repeated for an alternative set of ASE pulse sequence parameters, replicating the effects

observed for R_2' and DBV (Fig S6a,b and Fig. S7). However, a monotonic relationship

between apparent and true DBV was revealed suggesting that a shorter t_E and/or maximum τ

value could enable the technique to be sensitive to a broader range of OEF values.

Finally, the results of these multi-radius simulations appear to be consistent with

previous measurements of OEF=21±2% and DBV=3.6±0.4% (Stone and Blockley, 2017).

Under the assumption that a true OEF of 40% is healthy, Fig. 7 would predict an apparent

OEF of 23%. Likewise Fig. 8 would predict the percentage error in the apparent DBV is

100%, which would reduce the measured value above to 1.8%. This would bring these

measurements in line with other MR based measurements of DBV at 1.75% (He and

Yablonskiy, 2007) and venous CBV at 2.2% (Blockley et al., 2013).

<u>Future work and Limitations</u>

Whilst the simulation methodology used in this study has identified some limitations of the

current implementation of ASE based qBOLD, it also offers an opportunity to optimise future

implementations. Further simulations could be used to identify optimal values of t_E and τ

which maximise the linearity of the relationship between apparent OEF and the ground truth.

They could also be used to estimate a more appropriate scale factor for OEF estimation by

treating the $\frac{4}{3}\pi$ geometry factor in Eq. (1) as an arbitrary scale factor. Such an approach has

previously been used in calibrated BOLD to great effect (Griffeth and Buxton, 2011).

The results of this study rely on a detailed model of the qBOLD signal. However, in

this implementation it only accounts for the intra- and extravascular signal contributions of a

single distribution of blood vessels in grey matter. Whilst this two compartment model was

sufficient to investigate the origin of DBV overestimation in ASE based qBOLD, a more

realistic model might include the signal contributions of cerebral spinal fluid (Dickson et al.,

2009), the myelin in white matter (Bouvier et al., 2013), desaturated arterial blood vessels

(Boas et al., 2008), the effect of iron deposition (Wismer et al., 1988) or different vessel

radius distributions (Germuska et al., 2013; Lauwers et al., 2008). As such these

contributions to the qBOLD signal may also provide fertile ground for future exploration.

Conclusion

The ASE qBOLD signal decay was found to be symmetric with respect to the spin echo, in

contrast to previous simulation of the GESSE pulse sequence. Overestimation of DBV by

ASE based qBOLD was found to be dominated by the effect of diffusion on extravascular

signal decay, with the presence of intravascular blood signal having only a small

contribution. Integrating the results of single vessel simulations using an *in vivo* distribution

of vessel radii revealed several limitations of current measurements and provides a

foundation for future optimisation of ASE based qBOLD acquisitions.

Acknowledgements

This work was supported by the Engineering and Physical Sciences Research Council [grant

number EP/K025716/1].

Appenix A

The results of the simulations performed in this study can be accessed via the Oxford

University Research Archive, doi: https://doi.org/10.5287/bodleian:mvPY99a9D.

Furthermore, the code used to generate these simulation results and to analyse experimental

data can be downloaded from the Zenodo repository, doi:

https://doi.org/10.5281/zenodo.2586624. Supplementary data related to this article can be

found at <DOI>.

References

An, H., Lin, W., 2003. Impact of intravascular signal on quantitative measures of cerebral oxygen extraction and blood volume under normo- and hypercapnic conditions using an asymmetric spin echo approach. Magn. Reson. Med. 50, 708–716.

An, H., Lin, W., 2000. Quantitative measurements of cerebral blood oxygen saturation using magnetic resonance imaging. J. Cereb. Blood Flow Metab. 20, 1225–1236.

Astrup, J., Siesjö, B.K., Symon, L., 1981. Thresholds in cerebral ischemia - the ischemic penumbra. Stroke 12, 723–725.

Berman, A.J.L., Mazerolle, E.L., MacDonald, M.E., Blockley, N.P., Luh, W.M., Pike, G.B., 2017. Gas-free calibrated fMRI with a correction for vessel-size sensitivity. Neuroimage 169, 176–188.

Berman, A.J.L., Pike, G.B., 2018. Transverse signal decay under the weak field approximation: Theory and validation. Magn. Reson. Med. 32, 710–749.

Blockley, N.P., Griffeth, V.E.M., Germuska, M.A., Bulte, D.P., Buxton, R.B., 2013. An analysis of the use of hyperoxia for measuring venous cerebral blood volume: Comparison of the existing method with a new analysis approach. Neuroimage 72, 33–40.

Blockley, N.P., Jiang, L., Gardener, A.G., Ludman, C.N., Francis, S.T., Gowland, P.A., 2008. Field strength dependence of R1 and R2* relaxivities of human whole blood to prohance, vasovist, and deoxyhemoglobin. Magn. Reson. Med. 60, 1313–1320.

Boas, D.A., Jones, S.R., Devor, A., Huppert, T.J., Dale, A.M., 2008. A vascular anatomical network model of the spatio-temporal response to brain activation. Neuroimage 40, 1116–1129.

Bouvier, J., Castellani, S., Debacker, C., Pannetier, N., Troprès, I., Krainik, A., Barbier, E.L., 2013. Evalutation of multiparametric qBOLD in white matter: a simulation study, in: Proc. Intl. Soc. Mag. Reson. Med. . p. 1.

Boxerman, J.L., Bandettini, P.A., Kwong, K.K., Baker, J.R., Davis, T.L., Rosen, B.R., Weisskoff, R.M., 1995a. The intravascular contribution to fMRI signal change: Monte Carlo modeling and diffusion-weighted studies in vivo. Magn. Reson. Med. 34, 4–10.

Boxerman, J.L., Hamberg, L.M., Rosen, B.R., Weisskoff, R.M., 1995b. MR contrast due to intravascular magnetic susceptibility perturbations. Magn. Reson. Med. 34, 555–566.

Christen, T., Pannetier, N.A., Ni, W.W., Qiu, D., Moseley, M.E., Schuff, N., Zaharchuk, G., 2014. MR vascular fingerprinting: A new approach to compute cerebral blood volume,

mean vessel radius, and oxygenation maps in the human brain. Neuroimage 89, 262–270.

Dickson, J.D., Ash, T.W.J., Williams, G.B., Harding, S.G., Carpenter, T.A., Menon, D.K., Ansorge, R.E., 2010. Quantitative BOLD: the effect of diffusion. J. Magn. Reson. Imaging 32, 953–961.

Dickson, J.D., Ash, T.W.J., Williams, G.B., Sukstanskii, A.L., Ansorge, R.E., Yablonskiy, D.A., 2011. Quantitative phenomenological model of the BOLD contrast mechanism. J. Magn. Reson. 212, 17–25.

Dickson, J.D., Williams, G.B., Harding, S.G., Carpenter, T.A., Ansorge, R.E., 2009. Nulling the CSF Signal in Quantitative fMRI, in: Proc. Intl. Soc. Mag. Reson. Med. 17. p. 1640.

Germuska, M.A., Meakin, J.A., Bulte, D.P., 2013. The influence of noise on BOLD-mediated vessel size imaging analysis methods. J. Cereb. Blood Flow Metab. 33, 1857–1863.

Griffeth, V.E.M., Buxton, R.B., 2011. A theoretical framework for estimating cerebral oxygen metabolism changes using the calibrated-BOLD method: Modeling the effects of blood volume distribution, hematocrit, oxygen extraction fraction, and tissue signal properties on the BOLD signal. Neuroimage 58, 198–212.

He, X., Yablonskiy, D.A., 2007. Quantitative BOLD: Mapping of human cerebral deoxygenated blood volume and oxygen extraction fraction: Default state. Magn. Reson. Med. 57, 115–126.

Kety, S.S., Schmidt, C.F., 1948. The nitrous oxide method for the quantitative determination of cerebral blood flow in man: Theory, procedure and normal values. J. Clin. Invest. 27, 476–483.

Kiselev, V.G., Posse, S., 1999. Analytical model of susceptibility-induced MR signal dephasing: effect of diffusion in a microvascular network. Magn. Reson. Med. 41, 499–509.

Lauwers, F., Cassot, F., Lauwers-Cances, V., Puwanarajah, P., Duvernoy, H., 2008. Morphometry of the human cerebral cortex microcirculation: General characteristics and space-related profiles. Neuroimage 39, 936–948.

Lee, H., Englund, E.K., Wehrli, F.W., 2018. Interleaved quantitative BOLD□: Combining extravascular R 2 □ - and intravascular R 2 -measurements for estimation of deoxygenated blood volume and hemoglobin oxygen saturation. Neuroimage 174, 420–431. https://doi.org/10.1016/j.neuroimage.2018.03.043

Marchal, G., Rioux, P., Petit-Taboué, M.C., Sette, G., Travère, J.M., Le Poec, C., Courtheoux, P., Derlon, J.M., Baron, J.C., 1992. Regional cerebral oxygen consumption, blood flow, and blood volume in healthy human aging. Arch. Neurol. 49, 1013–1020.

Muller, M.E., 1959. A note on a method for generating points uniformly on n -dimensional spheres. Commun. ACM 2, 19–20.

Pannetier, N.A., Sohlin, M., Christen, T., Schad, L., Schuff, N., 2014. Numerical modeling of susceptibility-related MR signal dephasing with vessel size measurement: phantom validation at 3T. Magn. Reson. Med. 72, 646–658.

Rempp, K.A., Brix, G., Wenz, F., Becker, C.R., Gückel, F., Lorenz, W.J., 1994. Quantification of regional cerebral blood flow and volume with dynamic susceptibility contrast-enhanced MR imaging. Radiology 193, 637–641.

Sharan, M., Jones, M.D., Koehler, R.C., Traystman, R.J., Popel, A.S., 1989. A compartmental model for oxygen transport in brain microcirculation. Ann. Biomed. Eng. 17, 13–38.

Spees, W.M., Yablonskiy, D.A., Oswood, M.C., Ackerman, J.J., 2001. Water proton MR properties of human blood at 1.5 Tesla: magnetic susceptibility, T1, T2, T2*, and non-Lorentzian signal behavior. Magn. Reson. Med. 45, 533–542.

Stone, A.J., Blockley, N.P., 2017. A streamlined acquisition for mapping baseline brain

oxygenation using quantitative BOLD. Neuroimage 147, 79–88.

Stone, A.J., Harston, G.W.J., Carone, D., Okell, T.W., Kennedy, J., Blockley, N.P., 2019. Prospects for investigating brain oxygenation in acute stroke: Experience with a non-contrast quantitative BOLD based approach. Hum. Brain Mapp.

Tsai, A.G., Johnson, P.C., Intaglietta, M., 2003. Oxygen gradients in the microcirculation. Physiol. Rev. 83, 933–963.

Wismer, G.L., Buxton, R.B., Rosen, B.R., Fisel, C.R., Oot, R.F., Brady, T.J., Davis, K.R., 1988. Susceptibility induced MR line broadening: applications to brain iron mapping. J. Comput. Assist. Tomogr. 12, 259–265.

Yablonskiy, D.A., Haacke, E.M., 1994. Theory of NMR signal behavior in magnetically inhomogeneous tissues: the static dephasing regime. Magn. Reson. Med. 32, 749–763.

Tables

Table 1. Vascular compartment model described by (Sharan et al., 1989). Radius, length and number of vessels were used to calculate the relative volume fractions for each compartment with and without arteriolar vessels.

	Arterioles					Capillary		Venules			
	a1	a2	a3	a4	a5	c	v5	v4	v3	v2	v1
Radius (µm)	60	30	15	10	5	2.8	7.5	15	22.5	45	90
Length (µm)	5390	2690	1350	900	450	600	450	900	1350	2690	5390
Number of vessels	1880	1.5×10^4	1.15×10^5	3.92×10^5	3.01×10^6	5.92×10^7	3.01×10^6	3.92×10^5	1.15×10^5	1.5×10^4	1880
Relative vol. frac. (%) (all vessel types)	4.3	4.3	4.1	4.1	4.0	32.6	8.9	9.3	9.2	9.6	9.6
Relative vol. frac. (%) (excluding arterioles)						41.2	11.3	11.7	11.6	12.1	12.1

Figures

Fig. 1. By combining the phases generated by the Monte Carlo simulations different pulse sequences can be simulated. The cumulative sum of the phase accumulated after the 180° refocussing pulse is subtracted from the cumulative sum of the phases accumulated prior to the refocussing pulse. In a standard spin echo pulse sequence (a) the refocussing pulse is placed midway between the 90° excitation pulse and the echo time (t_E), which is equal to the spin echo time (t_{SE}). The GESSE pulse sequence (b) introduces R_2'-weighting through the parameter τ by altering t_E, whilst keeping t_{SE} constant. Note: each value of τ is acquired at a different t_E. The ASE sequence (c) introduces R_2'-weighting by shifting the refocussing pulse by a time $\tau/2$ leading to a change in t_{SE}, although t_E is kept constant. By convention positive values of τ occur when the $t_E > t_{SE}$ and negative values occur when $t_E < t_{SE}$.

Fig. 2. Examples of the signal decay from the ASE pulse sequence as a function of vessel radius. (a) The extravascular signal (S_{EV}) decay is observed to be symmetric with respect to τ=0 regardless of vessel radius. (b) The intravascular signal (S_{IV}) decay shows considerable signal attenuation which is symmetric and varies weakly with τ.

Fig. 3. Investigation of the effect of vessel radius on the parameter estimates derived from ASE based qBOLD. Simulations were first performed with a fixed OEF (E_0=40%) and three DBV values (top) then with a fixed DBV (V_0=3%) and three values of OEF (bottom). The apparent R_2' (left) is estimated for each OEF-DBV pair and presented alongside the R_2' values predicted by the SDR qBOLD model (dashed lines). Likewise the apparent DBV (centre) and apparent OEF (right) are presented alongside the true DBV and OEF, respectively, (dashed lines).

Fig. 4. Estimation of the percentage error in DBV at each of the three simulated values with (a) fixed E_0=40%, varying DBV and (b) fixed V_0=3%, varying E_0.

Fig. 5. Investigation of the contribution of intravascular signal to qBOLD parameter estimates (PE) presented in Fig. 3. This contribution was quantified as the percentage difference between PEs simulated with and without intravascular signal i.e. $(PE_{EV} - PE_{EV+IV})/PE_{EV+IV}$. Extravascular only PE results can be found in supplementary materials (Fig. S5).

Fig. 6. Investigation of the origin of the overestimation of the measured DBV for three different OEF values; (a) E_0=60%, (b) E_0=40%, (c) E_0=20% (true V_0=3%). The orange markers represents natural log of the measured signal at τ=0, plotted here as $-\ln S_{meas}^{S}(0)$, whilst the green markers represent the log of the intercept extrapolated from long τ data points ($\ln S_{extrap}^{L}(0)$). The sum of these curves is the apparent DBV as in Fig. 3 and represented here by the grey shaded area. Dashed lines display the prediction made by the SDR qBOLD model.

Fig. 7. The effect of multiple vessel radii simulations on the qBOLD parameter estimates was considered by generating many pairs of OEF and CBV values. ASE pulse sequence parameters were t_E=80 ms with τ=0 and τ=16 to 64 ms in 4 ms steps following the work of (Stone and Blockley, 2017). (a) The apparent R_2' is linearly dependent on the R_2' predicted by the SDR model, but with a different gradient. (b) A large amount of uncertainty in the apparent DBV is observed. (c) The apparent OEF appears to plateau beyond 50%, but monotonically increases with true OEF for lower values. Markers are coloured to reflect true dHb content, true OEF and true DBV for parts (a), (b) and (c), respectively.

Fig. 8. The uncertainty in DBV in Fig. 7 was investigated by plotting apparent DBV as a function of true OEF. ASE pulse sequence parameters follow the work of (Stone and Blockley, 2017). The results suggest that the error in the apparent DBV is OEF dependent. Markers are coloured to reflect their true DBV.