A rapid and automated sample-to-result *Candida auris* real-time PCR assay for high throughput testing of surveillance samples with BD MAX™ open system

L. Leach[1], A. Russell[1], S. Chaturvedi[1,2*], V. Chaturvedi[1,2*]

Mycology Laboratory[1], Wadsworth Center New York State Department of Health, and Department of Biomedical Sciences[2], University at Albany School of Public Health, Albany, NY, USA

*Corresponding authors: Mycology Laboratory, Wadsworth Center, New York State Department of Health, 120 New Scotland Ave., Albany, NY 12208, USA

Tel: +1-518-474-4177; FAX: +1-518-486-7971;

E-mail: Sudha.Chaturvedi@health.ny.gov

E-mail: Vishnu.Chaturvedi@health.ny.gov ⓘ orcid.org/0000-0002-3922-9676

ABSTRACT

The multidrug-resistant yeast pathogen *Candida auris* continues to cause

outbreaks and clusters of clinical cases worldwide. Previously, we developed a real-

time PCR test for the detection of *C. auris* from surveillance samples (Leach et al. JCM.

2018: 56, e01223-17); the test played a crucial role in the ongoing investigations of *C.*

auris outbreak in New York City. To ease the implementation of the test in other

laboratories, we developed an automated sample-to-result real-time *C. auris* PCR using

BD MAX™ open system, a platform proven to be adaptable for direct detection of other

fungal pathogens. We optimized sample extraction at three different temperatures and

four incubation periods, sensitivity was determined using eight pools of patient

samples, and specificity was calculated using four clades of *C. auris*, and closely and

distantly related yeasts. Three independent extractions and testing of two patient

sample pools in the quadruplicate yielded assay precision. BD MAX™ optimum assay

conditions were: DNA extraction at 75°C, 20 min, and the use of PerfeCTa Multiplex

qPCR Toughmix. The limit of detection (LOD) of the assay was one *C. auris* CFU/PCR

reaction; we detected all four clades of *C. auris* without cross-reactivity to other yeasts.

Of the 110 patient surveillance samples tested, 49 were positive for *C. auris* using the BD

MAX™ System with 94% clinical sensitivity and 94% accuracy compared to the manual

assay. BD MAX™ assay allows high throughput screening of 180 *C. auris* surveillance

samples in a 12-hour workday.

Candida auris, an emerging multidrug-resistant yeast pathogen, continues to

cause outbreaks and clusters of clinical cases worldwide[1, 2]. There are ongoing efforts to

devise better diagnostic approaches for the rapid detection of *C. auris* in clinical and

surveillance samples[3-5]. Previously, we developed and validated a manual real-time

PCR test for the detection of *C. auris* from surveillance samples at the New York State

Department of Health (NYSDOH) Mycology Laboratory (Leach et al. JCM. 2018: 56,

e01223-17). The laboratory-developed test (LDT) enabled NYSDOH laboratory scientists

and epidemiologists to carry out unprecedented surveillance and testing in hospitals

and healthcare facilities in New York City. To date, over 13,000 clinical samples from

151 facilities, and over 1,000 *C. auris* isolates were processed from likely the largest,

sustained outbreak documented globally [2]. *Candida auris* LDT, standard operating

procedures, and validation results were shared extensively with the hospital,

commercial, and public health laboratories (S. Chaturvedi, unpublished data).

However, the adoption of *C. auris* LDT assay is progressing slowly while the affected

facilities and sample numbers continue to grow. *Candida auris* real-time PCR assay was

not amenable to automation and high throughput screening. Therefore, we developed

and validated a real-time PCR for *C. auris* using BD MAX™ System, a fully integrated

and automated platform amenable to direct detection of other fungal pathogens[6, 7].

As we aimed to migrate the manual real-time PCR test to BD MAX™, the

primers, and probe, and PerfeCTa Multiplex qPCR ToughMix (Quanta Biosciences),

were evaluated as per details reported earlier along with BD MAX™ DNA MMK SPC

master mix[3]. We optimized BD MAX™ ExK™ DNA 1 (Plasma/Serum/Urine) DNA

Extraction Kit on the BD MAX™ using three different temperatures and four incubation

times. We carried out the entire sample-to-result procedure in the BD MAX™ PCR

Cartridges. BD MAX™ assay sensitivity was determined using eight pools of patient

samples in parallel testing with the manual real-time PCR test. We determined assay

specificity by the analysis of four clades of *C. auris*, and closely and distantly related

yeasts at high and low concentrations. The primers and probe set were previously

screened against a more extensive panel comprising of closely and distantly related

fungi, bacteria, viruses, and parasites[3]. The assay precision was determined by testing

samples in three independent extractions and testing two patient pools in the

quadruplicate. Following validation, we used the assay for the screening of 110

surveillance samples. GraphPad Prism 8 software (GraphPad Software, Inc., La Jolla,

CA) was used for statistical analysis. We used student t-test for analysis of the means

and considered a *P* value of <0.05 statistically significant. The manual *C. auris* real-time

PCR assay was selected as the "gold standard" to assess the diagnostic value of BD

MAX™ assay in *C. auris* detection from surveillance samples. BD MAX™ assay hands-

on time estimation involved the sum of the time required to complete sample

preparation, device loading, and cleaning, and result review and reporting.

We determined the optimum sample extraction condition to be 75°C for 20 min

on the BD MAX™ (Supplementary figures 1). The PerfeCTa Multiplex qPCR Toughmix

allowed for more efficient amplification of *C. auris* DNA than the BD MAX™ DNA

MMK (SPC) master mix (Supplementary figures 2). The BD MAX™ assay was highly

sensitive with the limit of detection (LOD) of one *C. auris* CFU/PCR reaction within Ct

range of 18-37 (Figure 1A). The linear regression analysis showed y = -3.66,446x + 37.49

and R^2= 0.9995. The assay detected all four known clades of *C. auris* with similar Ct

values for the same concentration of cells (Supplementary Table 1). The assay was

highly specific as no cross-reactivity was seen against two closely related yeasts, *C.*

haemulonii and *C. duobushaemulonii*, and 2 distantly related yeasts, *C. albicans*, and *C.*

glabrata, tested at high (1 x10^4 cells/PCR reaction) and low (1 x 10^1 cells/PCR reaction)

concentrations (Supplementary Table 1). The assay was highly reproducible as it

produced consistent Ct values for a sample pool on different days and within the same

day of testing (Supplementary Table 2). Of the 110 surveillance samples tested, 49 were

positive for *C. auris* using the BD MAX™ System, resulting in 94% clinical sensitivity

and 94% accuracy when compared to the manual real-time PCR assay (Table 1; Figure

1B). We estimated the automated sample-to-result *C. auris* BD MAX™ assay to allow

screening of 180 samples in a 12-hour workday.

We developed and validated the first automated sample-to-result real-time PCR

assay for high throughput testing of *C. auris*. The highlights of this study are: easy

migration of the manual real-time PCR assay to an automated platform, the consistent

performance of primers, probes, and master mix from our manual assay, comparable

sensitivity and specificity, and easy integration in the laboratory workflow. These

findings also confirm the suitability of BD MAX™ platform for direct detection of

fungal pathogens from clinical specimens as reported earlier for *Pneumocystic jirovecii*

and *Coccidioides immitis* [6, 7]. Ahmad et al. (2019) partially modified our manual real-time

PCR assay by carrying out DNA extraction from suspected *C. auris* surveillance samples

on MagNA Pure 96 automated extraction system (Roche Diagnostic System,

Indianapolis, IN, USA); these authors reported a good correlation of the modified test

findings with the results obtained by cell culture and MALDI-TOF-MS identification [4].

The authors further claimed that partial automation of the real-time PCR assay allowed

testing of approximately 200 samples per day. However, the trend in infectious disease

diagnosis is towards full automation, multiplexing, and miniaturization of assay

systems to facilitate point-of-care (POC) diagnosis rapidly [8, 9]. It is encouraging to note

that another group has successfully migrated manual real-time PCR assays to multiple

sample-to-results platforms such as BD MAX, ELITe InGenius (ELITechGroup), and

ARIES (Luminex) simultaneously[10]. In summary, the development and validation of the

rapid and automated sample-to-result *C. auris* BD MAX™ open system assay would

allow for the wider adaptability and availability of surveillance testing at the front-line

laboratories.

ACKNOWLEDGMENTS

This work was supported by the Centers for Disease Control and Prevention-Antibiotic

Resistance Lab Network grant (NU50CK000423) and Clinical Laboratory Reference

System, Wadsworth Center, New York State Department of Health.

L. Leach assisted in the design of the assay, performed the experiments and tabulated

and analyzed the data. A. Russel assisted in the design of the assay, performed the

experiments and participated in the analysis of data. S. Chaturvedi conceived and

designed the study, interpreted data, and edited the manuscript. V. Chaturvedi,

conceived and designed the study, interpreted data and wrote the manuscript. All

authors commented upon the final manuscript.

Figure legends

Figure 1. BD MAX™ *Candida auris* assay sensitivity. The pool of surveillance samples

ranging in *C. auris* CFU from 5×10^{-1}-10^5/PCR reactions were run in duplicate on three

different days. The assay was linear over six orders of magnitude, and the limit of

detection limit was one CFU/PCR reaction on BD MAX and manual assay.

Figure 2. Correlation between *Candida auris* BD MAX™ real-time PCR and manual real-

time PCR. Ct distribution of 110 patient surveillance samples showed that BD MAX

real-time PCR assay had an excellent correlation for the detection of *C. auris* DNA with

manual real-time PCR assay for most of the samples (r = 0.9955).

Supplementary figures

Supplementary Figure 1. Comparison of efficacy of two different master mixes with

different extraction conditions on the BD MAX™ platform.

Supplementary Figure 2. Performance of BD MAX™ at three different temperatures and

four incubation periods determined with two patient pools.

Table 1. Comparison of Test Performance

BD MAX Real-Time PCR Result	Manual Real-Time PCR Result		Accuracy (%)	Sensitivity (%)	Specificity (%)	PPV (%)	NPV (%)
	Positive	Negative					
Positive	45	4	94%	94%	94%	92%	95%
Negative	3	58					

154   Figure 1

157   Figure 2

158
159

160   Supplementary Figure 1.

162   Supplementary Figure 2.

163 Supplementary Table 1. Specificity testing BD MAX™ real-time PCR assay

Fungus (accession no.)*	CFU/PCR Reaction	Manual Ct	BD Max Ct
C. auris South Asia Clade (M5658)	1×10^4	21.51	23.3
	1×10^1	33.35	34.4
C. auris South Africa Clade (M5948)	1×10^4	22.45	23.5
	1×10^1	37.02	34.8
C. auris South America Clade (M5952)	1×10^4	21.17	21.3
	1×10^1	32.26	33.1
C. auris East Asia Clade (M5956)	1×10^4	17.84	19.5
	1×10^1	31.71	34.2
C. haemulonii (M8541)	1×10^4	Undet	-1
	1×10^1	Undet	-1
C. duobushaemulonii (M3051)	1×10^4	Undet	-1
	1×10^1	Undet	-1
C. albicans (M3223)	1×10^4	Undet	-1
	1×10^1	Undet	-1
C. glabrata (M8984)	1×10^4	Undet	-1
	1×10^1	Undet	-1

166 * NYSDOH Mycology Laboratory Culture Collection Repository

167   Supplementary Table 2. Intraassay reproducibility with two pools of samples test in

168   four lanes of BD Max.

Sample type	BD Max Ct	BD Max Ct SD	%CV 171
Pooled Axilla/Groin Swabs	27.1	0.48	1.79
	27.2		
	27.4		
	26.3		
Pooled Nares Swabs	22.8	0.39	1.70
	22.2		
	22.9		
	23.1		

References

1.   Lockhart SR, Etienne KA, Vallabhaneni S et al. Simultaneous Emergence of Multidrug-Resistant

Candida auris on 3 Continents Confirmed by Whole-Genome Sequencing and Epidemiological Analyses.

Clinical Infectious Diseases 2017; **64**: 134-40.

2.   Adams E, Quinn M, Tsay S et al. *Candida auris* in Healthcare Facilities, New York, USA, 2013-

2017. *Emerg Infect Dis* 2018; **24**: 1816-24.

3.   Leach L, Zhu Y, Chaturvedi S. Development and Validation of a Real-Time PCR Assay for Rapid

Detection of *Candida auris* from Surveillance Samples. *J Clin Microbiol* 2018; **56**.

4.   Ahmad A, Spencer JE, Lockhart SR et al. A High-Throughput and Rapid Method for Accurate

Identification of Emerging Multidrug-Resistant *Candida auris*. *Mycoses* 2019.

5.   Arastehfar A, Fang W, Daneshnia F et al. Novel multiplex real-time quantitative PCR detecting

system approach for direct detection of *Candida auris* and its relatives in spiked serum samples. *Future*

Microbiol 2019; **14**: 33-45.

6.   Dalpke AH, Hofko M, Zimmermann S. Development and Evaluation of a Real-Time PCR Assay for

Detection of *Pneumocystis jirovecii* on the Fully Automated BD MAX Platform. *Journal of Clinical*

Microbiology 2013; **51**: 2337-43.

7.   Mitchell M, Dizon D, Libke R et al. Development of a Real-Time PCR Assay for Identification of

Coccidioides immitis by Use of the BD Max System. *Journal of Clinical Microbiology* 2015; **53**: 926-9.

8.   Dincer C, Bruch R, Kling A et al. Multiplexed Point-of-Care Testing – xPOCT. *Trends in*

Biotechnology 2017; **35**: 728-42.

9.   Zarei M. Infectious pathogens meet point-of-care diagnostics. *Biosensors and Bioelectronics*

2018; **106**: 193-203.

10.   Frans G, Beuselinck K, Peeters B et al. Migrating a lab-developed MERS-CoV real-time PCR to 3

"Sample to Result" systems: experiences on optimization and validation. *Diagnostic Microbiology and*

Infectious Disease 2019.