Smartphone screening for neonatal jaundice via ambient-subtracted sclera chromaticity: *neoSCB* app pilot study

Felix Outlaw[1*], Miranda Nixon[1], Oluwatobiloba Odeyemi[1,#a], Lindsay W. MacDonald[2], Judith Meek[3], Terence S. Leung[1]

[1]Department of Medical Physics and Biomedical Engineering, University College London, London, UK

[2]Department of Civil Environmental and Geomatic Engineering, University College London, London, UK

[3]The Neonatal Care Unit, Elizabeth Garrett Anderson Wing, University College London Hospitals Trust, London, UK

[#a]Current Address: UCL Medical School, Faculty of Medical Sciences, University College London, London, UK

*Corresponding author

Email: felix.outlaw.15@ucl.ac.uk (FO)

Abstract

Jaundice is a major cause of mortality and morbidity in the newborn. Globally, early identification and home monitoring are significant challenges in reducing the incidence of jaundice-related neurological damage. Smartphone cameras are promising as colour-based screening tools as they

19are low-cost, objective and ubiquitous. We propose a novel smartphone method and app called

20neoSCB to screen for neonatal jaundice by imaging the sclera. It does not rely on colour calibration

21cards or accessories, which may facilitate its adoption at scale and in less economically developed

22regions. Our approach is to explicitly address three confounding factors in relating colour to

23jaundice: (1) skin pigmentation, (2) ambient light, and (3) camera spectral response. (1) The variation

24in skin pigmentation is avoided by imaging the sclera. (2) With the smartphone screen acting as an

25illuminating flash, a flash/ no-flash image pair is captured using the front-facing camera. The

26contribution of ambient light is subtracted. (3) This permits a device- and ambient-independent

27measure of sclera chromaticity following a one-time calibration. We introduce the concept of

28Scleral-Conjunctival Bilirubin (SCB), in analogy with Transcutaneous Bilirubin (TcB). The scleral

29chromaticity is mapped to an SCB value. A pilot study was conducted in the UCL Hospital Neonatal

30Care Unit. 51 neonates were imaged using the neoSCB app and had a blood test for total serum

31bilirubin (TSB). The better of two models for SCB based on ambient-subtracted sclera chromaticity

32achieved r = 0.75 (p<0.01) correlation with TSB. Ambient subtraction improved chromaticity

33estimates in laboratory tests and screening performance within our study sample. Using an SCB

34decision threshold of 190µmol/L, the sensitivity was 100% (specificity 61%) in identifying newborns

35with TSB>250µmol/L (area under receiver operating characteristic curve, AUROC, 0.86), and 92%

36(specificity 67%) in identifying newborns with TSB>205µmol/L (AUROC 0.85). These results are

37comparable to modern transcutaneous bilirubinometers.

38Introduction

39Smartphones present unique opportunities for healthcare thanks to their processing and sensing

40capabilities coupled with their ubiquity and connectivity. Researchers are investigating smartphone-

41based tools to track and manage ongoing conditions such as diabetes [1] and bipolar disorder [2], as

42well as point-of-care, one-time screening or diagnostic applications such as assessing pupil reflex

43after traumatic brain injury [3]. In some cases, smartphones are being used to replicate the

44functionality of specialised medical devices such as the heart rate and blood oxygenation

45measurements performed by pulse oximeters [4,5]. Many concepts involve using an external sensor

46or wearable to collect data which is then processed and stored on a connected smartphone. Others

47endeavour to use only the in-built sensors of the phone. For example, the in-built camera can be

48used to assess wounds [6], the in-built microphone to monitor lung function [7], and the in-built

49accelerometer to detect falls [8].

50In this work we explore the possibility of screening for jaundice in newborns using a smartphone

51camera. Jaundice refers to the yellow discolouration of skin and sclerae caused by a build-up of

52bilirubin, a naturally occurring breakdown product of the haemoglobin in red blood cells. Our

53technique relies on quantifying the colour of the sclera, as the degree of scleral yellowing is

54indicative of the systemic concentration of bilirubin. In fact, the overlaying conjunctiva more readily

55accumulates bilirubin than the sclera [9]. Here we use 'sclera' as a shorthand for 'sclera and

56conjunctiva'. The project goal is to achieve a screening performance comparable to that of

57commercially available transcutaneous bilirubinometers using only the smartphone itself (without

58any additional hardware or add-ons). To this end we introduce our screening app, which we call

59'neoSCB' for 'neonatal Scleral-Conjunctival Bilirubinometer'.

60Jaundice can be caused by an increase in the rate of haemolysis or a decrease in the rate of bilirubin

61excretion. A majority of newborns are affected by jaundice due to their immature livers and a high

62load of fetal haemoglobin breakdown. Other factors such as preterm birth and certain genetic

63disorders can significantly increase the risk of severe jaundice. When a neonate is severely jaundiced

64for an extended time, a neurotoxic form of bilirubin crosses the blood-brain barrier and can lead to

65permanent neurological dysfunction (kernicterus) or death [10].

66Early identification is crucial: once diagnosed, jaundice can be effectively treated using phototherapy

67or, in extreme cases, exchange transfusion. In hospitals, newborns suspected of being at risk can be

68given a blood test. This is the gold standard diagnostic technique for determining the concentration

69of bilirubin in the blood. However, jaundice may only become problematic some weeks after birth,

70by which time the mother and newborn have typically been discharged. Screening for jaundice in

71the home environment presents several challenges, and current screening methods have inherent

72limitations. The most reliable screening tool is the transcutaneous bilirubinometer (TcB), a point-of-

73care device that takes a contact-based optical measurement of the skin to estimate bilirubin

74concentration. However, this device is expensive and not widely available, especially in less

75economically developed countries. The next most reliable screening approach is a visual examination

76by a trained healthcare professional, who can also look for behaviours associated with jaundice.

77Unfortunately, visual identification of jaundice is unreliable, with significant inter-observer variation

78[11,12]. The UK's National Institute for Health and Care Excellence (NICE) explicitly recommends

79against using visual inspection as the sole means of screening [13].

80In low-resource and remote areas, it may be that neither TcB screening nor expert visual assessment

81are options. In such cases, new parents may suffer anxiety and unnecessary, costly trips to the

82hospital. In the worst cases, severe jaundice goes unnoticed and leads to permanent disability or

83death.

84The global burden of jaundice was estimated in 2010 to amount to 114,100 deaths and 75,400 cases

85of kernicterus each year [14]. The regions worst affected by jaundice-related death and disability are

86Sub-Saharan Africa and South Asia. These two regions account for three quarters of global mortality.

87Long-term impairments from kernicterus such as hearing loss and cerebral palsy are an order of

88magnitude more common in Sub-Saharan Africa than in developed countries [14]. Reasons for this

89include a lack of contact with healthcare professionals, a greater prevalence of certain genetic risk

90factors, such as G6PD deficiency, and the difficulty of spotting yellow discolouration in newborns

91with darker skin. The need for an accessible and reliable means of screening is particularly acute in

92these regions.

93A smartphone camera-based screening method promises to be both objective and contactless.

94Importantly, smartphones are now ubiquitous even in resource-poor parts of the world. The impact

95of turning these devices into globally accessible home screening devices would be significant.

96Research interest in smartphone-based jaundice screening is growing, and recent work is promising

97[15–22]. The most extensively validated system is the BiliCam app that estimates total serum

98bilirubin (TSB) based on the skin colour of the sternum [15]. Taylor et al. showed BiliCam performed

99at least as well as the latest models of TcB over a multi-ethnic sample of 530 newborns [18]. BiliCam

100uses a custom calibration card to standardise measurements, and machine learning on six images to

101estimate the TSB.

102neoSCB differs in two main respects. Firstly, we image the sclera rather than the skin. Leung et al.

103compared sclera and skin regions using digital photography and found a stronger correlation

104between TSB values and sclera colour [23]. The sclera accumulates bilirubin while lacking the

105melanin and haemoglobin chromophores found in skin. It has been suggested that this permits a

106more sensitive measure of jaundice [17,23]. Secondly, we introduce a novel ambient subtraction

107method using the screen of the phone to illuminate the subject. This technique removes the need

108for a calibration card in the scene by explicitly subtracting the influence of ambient light on the

109measurement. This is important as the need for a calibration card may represent a significant barrier

110to adoption in a global context.

111This paper introduces the neoSCB app and presents results from a pilot study testing the app in a

112neonatal population at University College London Hospital. One contribution of this paper is the

113ambient subtraction method using screen illumination, which serves as an equipment-free means of

114removing the variation due to ambient light. Another contribution is the demonstration that it is

115feasible to use this method in a clinical environment to image the newborn sclera. A third is the

116validation of the predictive utility of the Jaundice Eye Colour Index (JECI) proposed in a previous

117publication [24], and the comparison of this metric against another sclera-chromaticity-based TSB

118prediction model. We also introduce the concept of a Scleral-Conjunctival Bilirubin (SCB)

119measurement, analogous to the commonly used Transcutaneous Bilirubin (TcB) measurement.

120This paper is structured as follows:

121In the 'Theoretical Background' section, the process of image formation, the motivation for the

122ambient subtraction method, and the concept of a device-independent colour space are outlined.

123In 'Methods', the development and functioning of the neoSCB app is described, and an experiment

124in a controlled environment shows that ambient subtraction can improve the chromaticity estimate

125of a range of colours typical of a jaundiced eye. The clinical data collection procedure is explained.

126The processing of the image data to arrive at TSB estimates via sclera chromaticity estimates is

127outlined.

128In 'Results', we compare a simple linear regression model based on the JECI metric to a multiple

129linear regression model based on chromaticity. We quantify the screening performance in our study

130population.

131In 'Discussion', we summarise the challenges in attaining a reliable colour descriptor, and how our

132method seeks to address them. The performance of the neoSCB app is compared to standard

133screening tools. We discuss the advantages and drawbacks of the approach, and some limitations of

134the pilot study.

135Theoretical background

136Image formation

137When an object is imaged by an RGB digital camera, the signal recorded at each pixel sensor

138depends on the product of three independent factors: the spectral power distribution of the light

139incident on the object, the object's reflectance, and the spectral sensitivity of the camera. For a

140given point on the object, Equation 1Error: Reference source not found shows the recorded value, I_k,

141for the corresponding pixel sensor in terms of the incident light spectrum, *E(λ)*, the object's spectral

142reflectance profile, *S(λ)*, and the relative camera sensitivity, *Q_k(λ)*.

$$143 \quad I_k = a \int_{\lambda_{min}}^{\lambda_{max}} E(\lambda) S(\lambda) Q_k(\lambda) d\lambda$$

$$\left(SEQ\ Equation\ \text{¿}\ ARABIC\ 1\right)$$

144where *k = {R,G,B}* denotes the channel and *a* is a scale factor which accounts for geometry-

145dependent effects such as object shading, as well as scaling due to camera optics and sensor

146efficiency. The integral limits are determined by the sensitivity range of the camera.

147To extract useful information about the object properties, one must isolate spectral reflectance *S(λ)*

148from the other factors that influence the recorded RGB values. This is problematic for several

149reasons. The light arriving at the camera sensor, sometimes called the colour signal *C(λ)*, is a product

150of two spectra, *C(λ) = E(λ)S(λ)*. It is impossible to isolate their individual contributions without

151contextual information; the same colour signal may result from two different objects given the right

152illuminations. Furthermore, the camera sensitivity functions *Q_k(λ)* reduce a colour signal *C(λ)* to a

153single RGB triplet. This is a many-to-one operation: different colour signals may produce the same

154camera response. Finally, not all camera sensitivity profiles *Q_k(λ)* are the same. There are variations

155between different makes and models, and even between different examples of the same device.

156This means that RGB triplets from different cameras cannot be directly compared, requiring

157conversion to a common colour space.

158Ambient subtraction

159In most situations the ambient illumination spectrum is unknown. There is a large literature on

160methods to discount the effect of the ambient illumination on digital images [25–27]. Human vision

161exhibits colour constancy and so does this automatically. Image processing pipelines must first

162estimate the effect of the illumination and then scale the output RGBs values to remove it, such that

163a neutral object has equal RGB values. This is known as white balance. The aim of ambient

164subtraction is to circumvent the problem by providing an additional, known illumination, which we

165call the flash here. Two images are taken, one with flash and one without. By subtracting the output

166linear RGB values of the second image from the first image, it is possible to estimate the image as it

167would have been recorded under only the flash illumination. White balance can then be performed

168using *a priori* knowledge of the camera and flash without the need for any estimation from image

169information.

170The ambient subtraction technique is summarised by Equation 2Error: Reference source not found

171for a pixel at image coordinate *x*. $I_k^{A+F}(x)$ refers to the pixel value under the ambient and flash

172illuminations (from the first image captured), $I_k^A(x)$ refers to the pixel value under only the ambient

173illumination (from the second image captured), and $\hat{I}_k^F(x)$ refers to the estimate for the pixel value

174under only the flash illumination.

$$175 \quad I_k^{A+F}(x) - I_k^A(x) = \hat{I}_k^F(x)$$
$$(2)$$

176Device-independent colour

177A digital camera records RGB values according to its spectral sensitivity functions, which are device

178specific. These values are said to be in the internal or raw space of the camera. To be displayed or

179compared to other colour values, these raw triplets must be converted to a device independent

180colour space. A device independent colour space allows a description of a colour that is independent

181of the method of capture or display.

182CIE XYZ colour space

183Based on the experiments of Wright and Guild in the 1920's, the 1931 International Commission on

184Illumination (CIE) defined a standard observer that represents the average human response to

185colour stimuli within a 2-degree arc inside the fovea [28]. The XYZ space derived from this standard

186observer is a 'reference' colour space, meaning that it encompasses all the colours visible to the

187human eye. Most colour spaces are defined in relation to the CIE XYZ space, either directly or

188indirectly. It is often used as an intermediary space when transforming from a device-specific raw

189colour space to device-independent 'output' spaces, such as the commonly used sRGB space.

190Developing a mapping between raw RGB values and XYZ values involves measuring or simulating a

191set of raw RGB values with known XYZ equivalents and then deriving a transformation matrix that

192minimises some colour error function. Typically, this is done using a specialised colour card with

193known XYZ values. Because illumination affects the colour of an object, the resulting transformation

194is only optimal for scenes under the same illumination used to develop the mapping. In general, no

195error free transformation from raw RGB to XYZ can exist because camera spectral sensitivity

196functions cannot be described as linear combinations of human cone sensitivities. In practice,

197however, 3x3 linear matrices suffice for most applications.

198Methods

199Ambient subtraction using screen illumination

200The use of a screen as a source of illumination when using the front-facing camera was first

201implemented in smartphones by Apple Inc., with their so-called Retina Flash [29]. The neoSCB app

202uses the screen instead of the back-facing flash because the screen is not uncomfortably bright for

203the newborn. It provides an even and diffuse illumination when held near to the face.

204neoSCB was developed for Android devices in the Android Studio IDE. It takes two raw format

205captures in succession using the front-facing camera. For the first, the screen is fully white and at

206maximum brightness. For the second, the screen is off. These will be referred to as the flash and no-

207flash images, respectively. The screen is white by default, turning off only briefly to take the second,

208no-flash image. This means the eye can adapt to the brightness, avoiding exposure to an over-

209stimulating flash. The image capture is triggered using the smartphone volume keys. Fig 1 shows the

210neoSCB app being used on a doll.

Fig 1. Demonstration of the use of the neoSCB app to capture flash/ no-flash sclera image pair. (A) Screen illuminates face for flash image capture. (B) Screen off for no-flash image capture.

There are several image processing techniques that rely on taking a flash and no-flash image pair of the same scene. These include estimating the ambient illumination, red eye removal, de-noising, and identifying foreground objects [30–32]. For our purposes, we are concerned with removing the influence of ambient light on the sclera colour recorded. By subtracting the colour values of the sclera under only the ambient illumination from the values obtained from the flash image, we can estimate the colour as it would appear under only the flash illumination. This ensures results are independent of the surrounding lighting environment.

Data collection

51 neonates (35 male, 16 female) from the UCL Hospital Neonatal Care Unit outpatient clinic and Postnatal Ward were imaged using the neoSCB app over a 14-month period from January 2017. Images were captured at a convenient time during assessment and treatment. Typically, this was before the blood test while the baby was supine, but some images were taken while the baby was upright in the arms of a parent. All images were captured within 20 minutes of a blood draw for TSB determination. No additional blood tests that were not part of the routine care of the newborns were performed. None had received phototherapy within the preceding 24 hours. Written and verbal consent was obtained from all parents. Ethics approval was granted by the London – City Road and Hampstead NHS Research Ethics Committee. The study was conducted in accordance with the World Medical Association's Declaration of Helsinki. The screen illumination was measured and confirmed to be safe for newborn eye (using appropriate aphakic hazard function) as per the International Commission on Non-Ionizing Radiation Protection Guidelines on Limits of Exposure to Incoherent Visible and Infrared Radiation [33].

234Gestational age ranged from 35 weeks and six days to 41 weeks and one day. Postnatal age was

235distributed as follows: 17 subjects less than one week old; 24 subjects between one and three weeks

236old; 10 subjects greater than three weeks old.

237The images were captured with a LG Nexus 5X smartphone held 10-20cm from the eye. The capture

238sequence took approximately one second. Multiple image pairs were captured when possible. Each

239image is saved in two formats: JPEG and DNG (a lossless raw image format by Adobe).

241Data analysis

242Data analysis was performed in MATLAB R2018a (The Mathworks, Inc., USA). Fig 2 shows the

243processing pipeline applied to the images captured by the neoSCB app. It can be divided into two

244stages: first, the subtraction method is used to determine an xy chromaticity estimate for the sclera;

245next, a chromaticity-based prediction model is used to predict TSB. We call this TSB prediction from

246the sclera the Scleral-Conjunctival Bilirubin, or SCB, creating a concept analogous with the commonly

247used Transcutaneous Bilirubin, or TcB.

248**Fig 2. Processing pipeline to estimate scleral-conjunctival bilirubin from raw format flash/ no-flash**

249**image pairs.** In Step 1, a color descriptor is estimated (ambient-subtracted sclera chromaticity). In

250Step 2, this is mapped to a scleral-conjunctival bilirubin (SCB) value.

251Chromaticity estimation via ambient subtraction

252First, sclera regions of interest (ROIs) in the flash and no-flash images are manually selected. This is

253done on each image separately to allow for small movements between the two captures. Blood

254vessels and specular highlights are avoided, and same region of the eye is selected in each image.

255dcraw (v. 9.27) [34], an open source raw conversion software, is used to convert the DNG files into

25616-bit, linear TIFF images. The images are minimally pre-processed: only black level subtraction and

257scaling to hardware saturation levels takes place. Using the ROIs and knowledge of the sensor Bayer

258pattern, the RGB values are extracted directly from the colour filter array output data. The median

259RGB values within the ROI are then calculated for flash and no-flash images. The median is used

260because it is robust when parts of the sclera ROI are not representative of the bulk colour. For

261example, part of a blood vessel or eyelash may be included by mistake.

262The raw RGB values from the two images are then subtracted to provide an estimate of the raw RGB

263values as they would appear without ambient illumination. This subtraction is valid because the

264values are still in the linear colour space of the camera. If there are several image pairs available for

265one subject, each is used to generate a new ambient-subtracted RGB estimate. If there are several

266usable ROIs in a given image (for example, sclera regions either side of the iris), these are treated as

267independent measurements too. Sometimes, the signal from the flash is too low. This happens when

268the screen is not held close enough to illuminate the eye sufficiently, or when the ambient light

269signal is greater in the no-flash image. The latter can happen if there is movement between the flash

270and no-flash captures or the ambient light changes during image acquisition. Values that are less

271than one percent of the total bit depth after subtraction are discarded, as the signal from the flash is

272too low or the subtraction is invalid due to environmental changes. The final RGB estimate for a

273given subject is calculated as the median of the remaining post-subtraction RGB triplets generated. A

274linear 3x3 matrix transformation optimised for the screen illumination spectrum is then used to map

275these RGB values to device-independent XYZ values.

276The phone-eye distance and absolute radiance of the phone screen are unknown, and therefore the

277subtracted XYZ values are only known up to a scale factor. Chromaticities (Equation 2Error:

278Reference source not found), denoted with lowercase letters xyz, are the XYZ values normalised by

279the total (X+Y+Z). Chromaticities are invariant to anything that affects all channels equally, such as

280exposure time or phone-eye distance. By definition, the three chromaticities sum to one, and so only

281two chromaticity values are needed to fix the third: the chromaticity space is two-dimensional.

$$282 \quad x = \frac{X}{X+Y+Z} \; ; \; y = \frac{Y}{X+Y+Z} \; ; \; z = \frac{Z}{X+Y+Z}$$

$$(2)$$

284 Chromaticity-based models for estimating Scleral-Conjunctival Bilirubin (SCB)

We investigate two SCB models based on the estimated chromaticity values. The first, SCB_{JECI}, is

based on the Jaundice Eye Colour Index (JECI) proposed by Leung et al. [24]. JECI is defined in

Equation 2 Error: Reference source not found and was conceived as a proxy for the yellowness of the

sclera. This SCB model is derived from a simple linear regression between JECI and TSB values. It

assumes that the predictive utility of the chromaticity measurement is a function of how far along

the yellow-blue axis the measurement is found. The JECI SCB model has one independent variable,

so, necessarily, some chromaticity information is lost (chromaticity is two-dimensional). The second

SCB model, SCB_{xy}, uses both x and y chromaticity as independent variables. It is derived via a multiple

linear regression of x and y chromaticity against TSB. The form of the SCB_{JECI} and SCB_{xy} models are

shown in Equation 3 Error: Reference source not found and Equation 4 Error: Reference source not

found.

$$296 \quad JECI = z_{D65} - z$$

$$(SEQ\ Equation ¿ ARABIC\ 2)$$

$$297 \quad SCB_{JECI} = m \cdot JECI + n$$

$$(SEQ\ Equation ¿ ARABIC\ 3)$$

$$298 \quad SCB_{xy} = p \cdot x + q \cdot y + r$$

$$(SEQ\ Equation ¿ ARABIC\ 4)$$

299 Validation of ambient subtraction using screen as illumination

300 What follows is a demonstration that ambient subtraction can be used to improve chromaticity

301 estimates under various ambient lightings. The test target was a printed grid of 11 yellow patches

302 ranging from 0 to 0.1 on the JECI scale, where 0 JECI is white and positive values are yellow. A JECI

303value corresponds to a chromaticity value rather than a colour value, so the luminance Y was
304selected to be the largest that would fit within the printer gamut. The printed target was measured
305using a X-Rite Inc. ColorMunki Photo to determine a ground truth value for the test target
306chromaticities.

307A TaoTronics TT-DL09 LED desk lamp was used to provide a controlled ambient light of two different
308correlated colour temperatures, 2700K (warm) and 6500K (cool). The smartphone was mounted at
309an angle of 45° and a separation 15cm from the test target centre for capture using the neoSCB app.

310The pipeline shown in Step 1 of Fig 2 was used to determine chromaticity values from flash/ no-flash
311image pairs.

312Fig 3 shows the results of the subtraction in the xy chromaticity space. For clarity, only data for every
313second patch is displayed. The human visual system chromaticity gamut is included for visual
314context. The ground truth chromaticities are plotted as filled circles in their respective colours (as
315measured using the ColorMunki). The empty circles represent the chromaticities measured under
316the no-flash condition, when the screen was off and only the lamp was illuminating the test target.
317The filled circles represent the chromaticities after the ambient subtraction. For both warm (Fig 3A)
318and cool (Fig 3B) ambient illuminations, the agreement is significantly improved by the subtraction
319method.

320**Fig 3. Visualising ambient subtraction in CIE 1931 chromaticity diagram for Jaundice Eye Colour**
321**Index (JECI) colours.** JECI values are 0.01, 0.03, 0.05, 0.07, 0.09. After ambient subtraction, values
322come into alignment with ground truth values. (A) Under warm ambient illumination, correlated
323colour temperature is 2700K. (B) Under cool ambient illumination, correlated colour temperature is
3246500K.

325Results

326For two of the 51 subjects, no TSB was recorded. For a further 12, a flash/ no-flash image pair with

327the sclera visible was not successfully captured, or the post-subtraction values were less than the

328minimum threshold. SCB values were calculated for the remaining 37 subjects.

329The correlation between TSB and SCB for the two SCB models are summarised in Table 1, alongside

330their root-mean-square residuals. A comparison of the models with and without the use of ambient

331subtraction shows that subtraction decreases RMS residuals and greatly increases correlation.

332**Table 1. Comparison of SCB models in terms of correlation with TSB and RMS residuals with/without ambient**

333**subtraction.**

SCB Model	Without Ambient Subtraction		With Ambient Subtraction	
	Pearson's r	RMS Residual (µmol/L)	Pearson's r	RMS Residual (µmol/L)
SCB_{JECI}	0.38 (p>0.01)	70	0.70 (p<0.01)	54
SCB_{xy}	0.56 (p<0.01)	63	0.75 (p<0.01)	51

334The best correlation is obtained using SCB_{xy}, the multiple linear regression model for SCB, with

335ambient subtraction. Fig 4 shows the correlation between SCB_{xy} and TSB with and without ambient

336subtraction. Results under the white screen illumination show a correlation of 0.56 (p < 0.01). When

337the subtraction method is applied to suppress the effect of ambient light, the correlation increases

338to 0.75 (p < 0.01). The root-mean-square residual is 51 µmol/L.

339**Fig 4. Scatter plot of scleral-conjunctival bilirubin (SCB) against TSB with and without ambient**

340**subtraction, 37 subjects.** The use of ambient subtraction greatly improves the correlation. (A)

341Correlation without ambient subtraction, r=0.56. (B) Correlation with ambient subtraction, r=0.75.

342Fig 5 displays the Bland-Altman plot for the ambient-subtracted SCB_{xy} method. The Bland-Altman

343plot shows that SCB_{xy} is systematically lower than TSB at high values, and vice versa. However, the

344important quantities are the sensitivity and specificity achieved as this technique is intended as a

345screening tool. Fig 6 shows Receiver Operating Characteristic curves for the ambient-subtracted

346SCB$_{xy}$ method at two clinically relevant screening thresholds: 250µmol/L (Fig 6A) and 205µmol/L (Fig

3476B). 250µmol/L is the threshold TcB measurement for babies greater than 35 weeks gestational age

348and older than 24 hours above which a blood test is recommended, and 205µmol/L is the treatment

349threshold for term babies at 24 hours, according to NICE guidelines [13].

350**Fig 5. Bland-Altman plot for TSB and SCB$_{xy}$ derived from ambient-subtracted scleral chromaticity.**

351**Fig 6. Receiver Operating Characteristic plots screening TSB at two thresholds using SCB$_{xy}$ as the**

352**decision metric.** (A) Screening threshold of 250µmol/L. (B) Screening threshold of 205µmol/L.

353For a screening threshold of 250µmol/L, the area under curve (AUC) – which quantifies how effective

354the tool is at separating cases over and under the screening threshold – is 0.86 (a perfect tool would

355have AUC of 1, a random decision would have AUC of 0.5). The maximal Youden's Index (true

356positive rate – false positive rate) is achieved for a cut off threshold of 190µmol/L. In this case, the

357screening method (with 95% confidence interval in brackets) has sensitivity 100% (88% - 100%),

358specificity 61% (43% - 76%), positive predictive value 45% (29% - 62%), and negative predictive value

359100% (88% - 100%).

360For a screening threshold of 205µmol/L, the AUC is 0.85. The maximal Youden's Index is achieved for

361a cut off threshold of 190µmol/L. In this case, the screening method (with 95% confidence interval in

362brackets) has sensitivity 92% (78% - 99%), specificity 67% (49% - 81%), positive predictive value 60%

363(43% - 75%), and negative predictive value 94% (80% - 100%). To successfully classify every baby

364with TSB greater than 205µmol/L in our study sample, the SCB$_{xy}$ cut off threshold is lowered to

365170µmol/L. In this case, the screening method has sensitivity 100% (88% - 100%) and specificity 54%

366(37% - 70%).

368Discussion

369neoSCB can be divided into two steps: first, attain a reliable colour descriptor for the tissue, and,

370second, map this to a SCB value. A screening decision can then be made based on the SCB value. In

371this work we have used the ambient-subtracted scleral chromaticity as our tissue colour descriptor,

372and a linear model to calculate SCB from it. This approach can be generalised: chromaticity could be

373mapped to SCB using more sophisticated models (the relatively small dataset prohibited this in our

374case), or a colour descriptor other than chromaticity could be used. Indeed, the linear model used to

375calculate SCB in this work systematically underestimates high TSB values, which could imply that a

376higher order model is required. Regardless of the SCB model used in step two, the performance will

377depend on the accuracy of the colour descriptor output by step one. In what follows we discuss

378some of the challenges in attaining an accurate colour descriptor and how they were addressed in

379this work.

380Attaining a reliable colour descriptor

381There are several challenges in attaining a reliable and representative tissue colour descriptor by

382digital imaging: (1) Ambient illumination affects recorded colour (2) Devices may respond in different

383ways to the same light signal, both at the hardware and software level (3) The selected ROI should

384be 'representative': a specular region or a blemish or bleed is not a good area from which to

385measure the degree of yellowing due to hyperbilirubinemia.

386Discounting ambient illumination

387To avoid the effect of ambient light, we have proposed a subtraction method. Our results indicate

388that the correlation between our chromaticity-based SCB models and TSB can be improved by

389subtracting the ambient component of the signal. There are several necessary requirements if

390ambient subtraction is to be effective. Crucially, the camera must have a linear response to light.

391Furthermore, no region of either image can be saturated, and both images must be captured with

392the same exposure time (or have values scaled to correct for this). The ambient light falling on the

393ROI cannot change in intensity or chromaticity between the flash and no-flash captures, otherwise

394the post-subtraction values will be spurious. Finally, ambient subtraction requires that the same ROI

395is visible in both images.

396Device independence

397To ensure the technique is device independent, we capture in raw format to avoid the software-

398specific post-processing algorithms found in all smartphones. However, smartphone cameras can

399vary at the hardware level too because sensors have different spectral sensitivity profiles [35]. To

400mitigate this, we use a transform to a device-independent colour space, the CIE 1931 XYZ space.

401Avoiding specular reflection and 'unrepresentative' regions

402By imaging the sclera, neoSCB avoids the variability introduced by different skin tones. However,

403despite the lack of melanin, the sclera colour signal may be influenced by factors other than bilirubin

404concentration. Non-representative and specular regions can sometimes be avoided by careful image

405segmentation. However, manual and automatic ROI selection system will both sometimes fail. To

406mitigate this source of error, multiple image pairs and regions can be analysed (with the assumption

407that only a minority of regions will suffer from these problems). Pixel values can be summarised with

408summary statistics robust to outliers such as the median.

409Comparison to transcutaneous bilirubinometers

410With a sample size of 630, Romangoli et al. evaluated the BiliChek (Philips Healthcare) and Jaundice

411Meter JM-103 (Draeger Medical Inc.) transcutaneous bilirubinometers for a screening threshold of

412205µmol/L [36]. The sensitivity and specificity achieved were 0.99 and 0.30, respectively, for

413BiliChek, and 1.00 and 0.42, respectively, for JM-103. neoSCB results are comparable for this

414screening threshold, with a sensitivity of 1.00 and a specificity of 0.54. However, neoSCB AUC was

415smaller at 0.85 compared to 0.89 for BiliCheck and 0.94 for JM-103.

In areas with poorer healthcare resources TcBs are not always available, even in hospital settings.

Smartphone-based methods offer an accessible and cheap alternative that can provide an objective

indicator of jaundice severity. Using smartphones could reduce the amount of equipment that

visiting healthcare professionals need to transport, allow for home monitoring by parents, and

increase the frequency of measurements in underequipped hospitals. In some hospitals blood tests

are only available infrequently (e.g. once per day) and TcBs are not used. In these situations, neoSCB

could be used to monitor response to treatment. It is also worth noting that TcBs are not used for

babies undergoing phototherapy treatment, as the skin exposed is no longer thought to be

representative of systemic bilirubin levels [37]. As the eyes are covered during phototherapy, this

hypothesised bleaching effect may not affect the sclera colour. Another reason to prefer sclera

colour measurement is the fact that it cannot be influenced by skin pigment. There is a body of

evidence that TcBs overestimate jaundice levels in dark-skinned infants [38–41]. This may lead to

unnecessary treatment which can block access to phototherapy beds and cause dehydration.

429 Limitations and future work

One of the aims of the ambient subtraction approach is to minimise the need for equipment such as

colour cards for calibration to lower the barrier to adoption. However, variation between the screen

spectral power distributions and camera spectral sensitivities of different smartphone models means

that some characterisation of the smartphone is needed to get a device-independent colour

measurement. This would involve a one-time colour card characterisation performed by the user

before use of the smartphone.

neoSCB relies on raw capture. Even if there is a trend towards allowing access to the raw values

output by the front-facing camera, only a few phones currently meet this requirement.

The post-subtraction signal is determined by the proportion of the signal which comes from the

flash. In this work, we discarded post-subtraction values less than one percent of the total bit depth

during offline analysis. In a future version, an automatic system will warn users if the screen

441illumination is insufficient, requesting a less bright environment or the screen held closer to the face.

442In some cases, movement between flash and no-flash captures rendered the pair of images

443unusable. The capture sequence lasts less than one second, so this is not typically an issue.

444Currently, segmentation of the sclera and thresholding of the post-subtraction pixel value results are

445both done after capture during an after-the-fact analysis. To provide a TSB prediction for the user at

446the time of capture would require the automation of these two steps. Sclera segmentation

447algorithms have been developed by researchers interested in biometric identification using sclera

448vasculature [42,43]. Mariakakis et al. developed their own segmentation using accessories as

449fiducials [17].

450Although here we have processed images after the time of capture on a desktop computer, the

451pipeline is computationally simple enough to be performed in real time on the smartphone. This is

452advantageous because it removes the need for internet connection for cloud-based processing,

453which is often unavailable or too expensive for users in less economically developed regions.

454Imaging the eye of the infant can be challenging due to unpredictable movements and the amount

455of time spent asleep. This study has shown that it is possible to collect this data using a smartphone

456in a time-constrained clinical environment. It was observed that infants typically open their eyes

457when feeding, and this can provide a suitable opportunity for measurement. In a home setting there

458is less time pressure than in a clinical assessment and so the measurement could be done at any

459convenient point.

460Although sclera colour has less variation than skin colour, there can be confounding influences.

461Neonates can have a blue tint in the sclera because it is thinner than the adult sclera. Blood vessels

462and bleeds must be identified so they do not lead to spurious TSB estimates.

463 Advantages of ambient subtraction for point-of-care colorimetry

The principle advantage of using the ambient subtraction approach is to remove the need of a

calibration colour card present in the frame of every image. Especially in realistic settings in the

home, in the clinic, and in the field, it can be difficult to position a colour card in the shot and almost

impossible to keep it in pristine condition for an extended period of use. By applying an ambient

subtraction, we can know *a priori* the effective illumination of the post-subtraction scene. This

means the transformation to XYZ can be determined by a one-time calibration in controlled

conditions and optimised to provide more accurate chromaticity information.

471 Conclusion

In this work, we show that by explicitly addressing the confounding factors in colour measurement

of jaundice (namely ambient light, camera characteristics, skin tone), a linear model based on sclera

chromaticity can predict total serum bilirubin accurately enough to be useful as a screening tool. By

employing a novel ambient subtraction technique using the screen as illumination, this can be done

without the need for add-ons or colour calibration cards in the shot. Our app, neoSCB, is tested in a

clinical environment, demonstrating that it is feasible to capture the necessary flash/ no-flash image

pair of the newborn sclera using the smartphone's front-facing camera. We have shown that the

correlation between a simple colour metric (JECI) and the measured TSB can be much improved by

ambient subtraction with this flash/ no-flash image pair.

We have proposed the concept of Scleral-Conjunctival Bilirubin (SCB), analogous to Transcutaneous

Bilirubin (TcB). It is not yet clear whether the scleral-conjunctival or transcutaneous bilirubin

concentration is a better proxy for serum bilirubin concentrations, or indeed whether serum or

extravascular tissue concentrations are better predictors of neurological damage from bilirubin. We

hope that by introducing the concept of SCB measurement further research into measuring jaundice

via the sclera can be stimulated.

487We have investigated the screening utility of two xy chromaticity-based SCB models. One based on

488the Jaundice Eye Colour Index (JECI) proposed by Leung et al. achieved a correlation of 0.70 (p<0.01)

489with TSB. Another based on a linear model of x and y chromaticity values achieved a correlation of

4900.75 (p<0.01). Using the latter approach, we were able to achieve a sensitivity of 100% for a

491specificity of 61% when screening infants with TSB above 250μmol/L and a sensitivity of 100% for a

492specificity of 54% when screening infants with TSB above 205μmol/L, a performance comparable to

493that of transcutaneous bilirubinometers.

494Acknowledgements

495The authors would like to thank all the study participants and their parents.

496References

4971.   Kirwan M, Vandelanotte C, Fenning A, Duncan MJ. Diabetes self-management smartphone

498   application for adults with type 1 diabetes: Randomized controlled trial. J Med Internet Res.

499   2013; doi:10.2196/jmir.2588

5002.   Grünerbl A, Muaremi A, Osmani V, Bahle G, Öhler S, Tröster G, et al. Smartphone-based

501   recognition of states and state changes in bipolar disorder patients. IEEE J Biomed Heal

502   Informatics. 2015; doi:10.1109/JBHI.2014.2343154

5033.   Mariakakis A, Baudin J, Whitmire E, Mehta V, Banks MA, Law A, et al. PupilScreen: Using

504   Smartphones to Assess Traumatic Brain Injury. Proc IMWUT '17. 2017. doi:10.1145/3131896

5054.   Scully CG, Lee J, Meyer J, Gorbach AM, Granquist-Fraser D, Mendelson Y, et al. Physiological

506   parameter monitoring from optical recordings with a mobile phone. IEEE Trans Biomed Eng.

507   2012; doi:10.1109/TBME.2011.2163157

5085.   Jonathan E, Leahy M. Investigating a smartphone imaging unit for photoplethysmography.

Physiol Meas. 2010; doi:10.1088/0967-3334/31/11/N01

6.	Wang L, Pedersen PC, Strong DM, Tulu B, Agu E, Ignotz R. Smartphone-based wound assessment system for patients with diabetes. IEEE Trans Biomed Eng. 2015; doi:10.1109/TBME.2014.2358632

7.	Larson EC, Goel M, Boriello G, Heltshe S, Rosenfeld M, Patel SN. SpiroSmart. Proceedings of the 2012 ACM Conference on Ubiquitous Computing - UbiComp '12. New York, New York, USA: ACM Press; 2012. p. 280. doi:10.1145/2370216.2370261

8.	He Y, Li Y, Bao S Di. Fall detection by built-in tri-accelerometer of smartphone. Proceedings - IEEE-EMBS International Conference on Biomedical and Health Informatics: Global Grand Challenge of Health Informatics, BHI 2012. 2012. doi:10.1109/BHI.2012.6211540

9.	Tripathi RC, Sidrys LA. 'Conjunctival Icterus,' Not 'Scleral Icterus.' JAMA: The Journal of the American Medical Association. 1979. doi:10.1001/jama.1979.03300230014016

10.	Rennie JM. Rennie & Robertson's Textbook of Neonatology. Elsevier Health Sciences; 2012.

11.	Riskin A, Tamir A, Kugelman A, Hemo M, Bader D. Is Visual Assessment of Jaundice Reliable as a Screening Tool to Detect Significant Neonatal Hyperbilirubinemia? J Pediatr. 2008; doi:10.1016/j.jpeds.2007.11.003

12.	Moyer VA, Sneed S. Accuracy of clinical judgment in neonatal jaundice. Arch Pediatr Adolesc Med. 2000; doi:10.1001/archpedi.154.4.391

13.	National Institute for Health and Care. Jaundice in newborn babies under 28 days. NICE Qual Stand. 2010;

14.	Bhutani VK, Zipursky A, Blencowe H, Khanna R, Sgro M, Ebbesen F, et al. Neonatal hyperbilirubinemia and Rhesus disease of the newborn: incidence and impairment estimates for 2010 at regional and global levels. Pediatr Res. 2013;74: 86–100. doi:10.1038/pr.2013.208

15.	de Greef L, Goel M, Seo MJ, Larson EC, Stout JW, Taylor JA, et al. Bilicam: using mobile

phones to monitor newborn jaundice. Proceedings of the 2014 ACM International Joint

Conference on Pervasive and Ubiquitous Computing. 2014. doi:10.1145/2632048.2632076

53516.   Aydın M, Hardalaç F, Ural B, Karap S. Neonatal Jaundice Detection System. J Med Syst.

536   2016;40. doi:10.1007/s10916-016-0523-4

53717.   Mariakakis A, Banks MA, Phillipi L, Yu L, Taylor J, Patel SN. BiliScreen: Smartphone-Based

538   Scleral Jaundice Monitoring for Liver and Pancreatic Disorders. Proc ACM Interactive, Mobile,

539   Wearable Ubiquitous Technol. 2017; doi:10.1145/3090085

54018.   Taylor JA, Stout JW, de Greef L, Goel M, Patel S, Chung EK, et al. Use of a Smartphone App To

541   Assess Neonatal Jaundice. Pediatrics. 2017; doi:10.1542/peds.2017-0312

54219.   Munkholm SB, Krøgholt T, Ebbesen F, Szecsi PB, Kristensen SR. The smartphone camera as a

543   potential method for transcutaneous bilirubin measurement. PLoS One. 2018;

544   doi:10.1371/journal.pone.0197938

54520.   Rizvi MR, Alaskar FM, Albaradie RS, Rizvi NF, Al-Abdulwahab K. A Novel Non-invasive

546   Technique of Measuring Bilirubin Levels Using BiliCapture. Oman Med J. Oman Medical

547   Specialty Board; 2019;34: 26–33. doi:10.5001/omj.2019.05

54821.   Padidar P, Shaker M, Amoozgar H, Khorraminejad-Shirazi M, Hemmati F, Najib KS, et al.

549   Detection of Neonatal Jaundice by Using an Android OS-Based Smartphone Application. Iran J

550   Pediatr. Kowsar; 2019; doi:10.5812/ijp.84397

55122.   Outlaw F, Meek J, MacDonald LW, Leung TS. Screening for Neonatal Jaundice with a

552   Smartphone. Proceedings of the 2017 International Conference on Digital Health - DH '17.

553   New York, New York, USA: ACM Press; 2017. pp. 241–242. doi:10.1145/3079452.3079488

55423.   Leung TS, Kapur K, Guilliam A, Okell J, Lim B, MacDonald LW, et al. Screening neonatal

jaundice based on the sclera color of the eye using digital photography. Biomed Opt Express.

2015; doi:10.1364/BOE.6.004529

24. Leung TS, Outlaw F, MacDonald LW, Meek J. Jaundice Eye Color Index (JECI): quantifying the yellowness of the sclera in jaundiced neonates with digital photography. Biomed Opt Express. Optical Society of America; 2019;10: 1250. doi:10.1364/BOE.10.001250

25. Agarwal V, Abidi B, Koschan A, Abidi MA. An Overview of Color Constancy Algorithms. J Pattern Recognit Res. 2013; doi:10.13176/11.9

26. Hordley SD. Scene illuminant estimation: Past, present, and future. Color Res Appl. 2006; doi:10.1002/col.20226

27. Gijsenij A, Gevers T, Van De Weijer J. Computational color constancy: Survey and experiments. IEEE Trans Image Process. 2011; doi:10.1109/TIP.2011.2118224

28. Fairman HS, Brill MH, Hemmendinger H. How the CIE 1931 color-matching functions were derived from Wright-Guild data. Color Res Appl. 1997; doi:10.1002/(SICI)1520-6378(199702)22:1<11::AID-COL4>3.0.CO;2-7

29. Aybes A, Westen P, Christie GN, Forstall S, others. Image capture using display device as light source. Google Patents; 2010.

30. Petschnigg G, Szeliski R, Agrawala M, Cohen M, Hoppe H, Toyama K. Digital photography with flash and no-flash image pairs. ACM Trans Graph. 2004; doi:10.1145/1015706.1015777

31. Lu C, Drew MS. Practical Scene Illuminant Estimation via Flash / No-Flash Pairs. Fourteenth Color Imaging Conf. 2006;

32. Sun J, Kang SB, Xu Z Ben, Tang X, Shum HY. Flash cut: Foreground extraction with flash and no-flash image pairs. Proceedings of the IEEE Computer Society Conference on Computer Vision and Pattern Recognition. 2007. doi:10.1109/CVPR.2007.383080

33. Ziegelberger G. ICNIRP guidelines on limits of exposure to incoherent visible and infrared radiation. Health Physics. 2013. doi:10.1097/HP.0b013e318289a611

34.   Coffin D. DCRAW: Decoding raw digital photos in Linux. 2008.

35.   Jiang J, Liu D, Gu J, Susstrunk S. What is the space of spectral sensitivity functions for digital

582   color cameras? Proceedings of IEEE Workshop on Applications of Computer Vision. 2013.

583   doi:10.1109/WACV.2013.6475015

584   36.   Romagnoli C, Zecca E, Catenazzi P, Barone G, Zuppa AA. Transcutaneous bilirubin

585   measurement: Comparison of Respironics BiliCheck and JM-103 in a normal newborn

586   population. Clin Biochem. 2012; doi:10.1016/j.clinbiochem.2012.03.014

587   37.   Tan KL, Dong F. Transcutaneous bilirubinometry during and after phototherapy. Acta Paediatr

588   Int J Paediatr. 2003; doi:10.1080/08035250310009257

589   38.   Slusher TM, Angyo IA, Bode-Thomas F, Akor F, Pam SD, Adetunji AA, et al. Transcutaneous

590   Bilirubin Measurements and Serum Total Bilirubin Levels in Indigenous African Infants.

591   Pediatrics. 2004;113: 1636–1641. doi:10.1542/peds.113.6.1636

592   39.   Wainer S, Rabi Y, Parmar SM, Allegro D, Lyon M. Impact of skin tone on the performance of a

593   transcutaneous jaundice meter. Acta Paediatr. 2009;98: 1909–1915. doi:10.1111/j.1651-

594   2227.2009.01497.x

595   40.   Samiee-Zafarghandy S, Feberova J, Williams K, Yasseen AS, Perkins SL, Lemyre B. Influence of

596   skin colour on diagnostic accuracy of the jaundice meter JM 103 in newborns. Arch Dis Child -

597   Fetal Neonatal Ed. 2014;99: F480–F484. doi:10.1136/archdischild-2013-305699

598   41.   Afanetti M, Eleni dit Trolli S, Yousef N, Jrad I, Mokhtari M. Transcutaneous bilirubinometry is

599   not influenced by term or skin color in neonates. Early Hum Dev. 2014;90: 417–420.

600   doi:10.1016/j.earlhumdev.2014.05.009

601   42.   Alkassar S, Woo WL, Dlay SS, Chambers JA. Robust Sclera Recognition System with Novel

602   Sclera Segmentation and Validation Techniques. IEEE Trans Syst Man, Cybern Syst. 2017;

603   doi:10.1109/TSMC.2015.2505649

60443. Das A, Pal U, Blumenstein M, Ballester MAF. Sclera Recognition - A Survey. Proceedings - 2nd

605 IAPR Asian Conference on Pattern Recognition, ACPR 2013. 2013.

606 doi:10.1109/ACPR.2013.168

608Supporting information

609S1 Data. Clinical data for all 51 subjects.

Figure One

Figure Two

Figure Three

Figure Four

Figure Five

Figure Six