1
2
3 Lights, Camera, Path Splitter: A New Approach for Truly Simultaneous Dual Optical Mapping of the
4 Heart with a Single Camera
5
6

7 Rafael Jaimes III[1,2] (rjaimes@childrensnational.org), Damon McCullough[1]
8 (dmcculloug@childrensnational.org) , Bryan Siegel[2] (bsiegel@childrensnational.org), Luther Swift[1,2]
9 (lswift2@childrensnational.org), James Hiebert[1] (jhiebert1691@gmail.com), Daniel McInerney[1]
10 (dmcinerney21@gmail.com), Nikki Gillum Posnack (nposnack@childrensnational.org)[1,2,3]
11

[1]Sheikh Zayed Institute for Pediatric and Surgical Innovation: Children's National Health System, 111
Michigan Avenue NW, Washington DC 20010
[2]Children's National Heart Institute: Children's National Health System, 111 Michigan Avenue NW,
Washington DC 20010
[3]Department of Pediatrics, Department of Pharmacology & Physiology, School of Medicine and Health
Sciences: George Washington University, 2300 I Street NW, Washington DC 20037
18
19
Corresponding Author:
Nikki Gillum Posnack, Ph.D.
Sheikh Zayed Institute, 6th floor, M7708
111 Michigan Avenue, NW
Washington, DC, USA 20010
Tel: (202) 476-2475
Email: nposnack@childrensnational.org
27

28 **Running title:** Dual optical mapping
29
30

ABSTRACT

Background: Optical mapping of transmembrane voltage and intracellular calcium is a powerful tool for investigating cardiac physiology and pathophysiology. However, simultaneous dual mapping of two fluorescent probes remains technically challenging. We introduce a novel, easy-to-use approach that requires a path splitter, single camera and excitation light to simultaneously acquire voltage and calcium signals from whole heart preparations, which can be applied to other physiological models – including neurons and isolated cardiomyocytes.

Results: Complementary probes were selected that could be excited with a single wavelength light source. Langendorff-perfused hearts (rat, swine) were stained and imaged using a sCMOS camera outfitted with an optical path splitter to simultaneously acquire two emission fields at high spatial and temporal resolution. Voltage (RH237) and calcium (Rhod2) signals were acquired concurrently on a single sensor, resulting in two 384x256 images at 814 frames per second. At this frame rate, the signal-to-noise ratio was 47 (RH237) and 85 (Rhod2). Imaging experiments were performed on small rodent hearts, as well as larger pig hearts with sufficient optical signals. In separate experiments, each dye was used independently to assess crosstalk and demonstrate signal specificity. Additionally, the effect of ryanodine on myocardial calcium transients was validated – with no measurable effect on the amplitude of optical action potentials. To demonstrate spatial resolution, ventricular tachycardia was induced –resulting in the novel finding that spatially discordant calcium alternans can be present in different regions of the heart, even when electrical alternans remain concordant. The described system excels in providing a wide field of view and high spatiotemporal resolution for a variety of cardiac preparations.

Conclusions: We report the first multiparametric mapping system that simultaneously acquires calcium and voltage signals from cardiac preparations, using a path splitter, single camera and excitation light. This approach eliminates the need for multiple cameras, excitation light patterning or frame interleaving. These features can aid in the adoption of dual mapping technology by the broader cardiovascular research community, and decrease the barrier of entry into panoramic heart imaging, as it reduces the number of required cameras.

Keywords:

Optical mapping, calcium cycling, transmembrane voltage, electrophysiology

BACKGROUND

Cardiovascular research has been propelled by the advent of parameter-sensitive probes, which can be

used to monitor transmembrane voltage and intracellular calcium within live cardiac preparations (1–5).

Optical mapping is an imaging technique that measures fluorescence signals across a cardiac

preparation with high spatiotemporal resolution. Optical mapping of voltage-sensitive probes (6–8)

allows for the measurement of action potential morphology and the spread of electrical activity, as well

as the identification of tissue heterogeneities that can promote arrhythmias. Whereas intracellular

calcium probes (9) are used to investigate modifications in excitation-contraction coupling, which can

alter action potential duration, elicit after-depolarizations, and promote electrical/mechanical alternans.

Accordingly, simultaneous imaging of both transmembrane voltage and intracellular calcium is a

powerful integrative tool for cardiac research (for extensive reviews see(1–4)). Yet, assembling a dual

optical mapping system remains technically challenging (4,10,11), which has limited the use of this

technique to a relatively small number of research laboratories.

<u>Dual camera configurations</u>

Simultaneous dual mapping systems have traditionally used a dual-sensor design, wherein the

emission of each complementary probe (voltage, calcium) is separated by wavelength and diverted to

two separate detectors (1,5,12–16). Such a design was described by Fast and Ideker, in which

cardiomyocytes monolayers were stained with complementary probes (RH237 – voltage, Fluo-3AM –

calcium) and the emitted fluorescent signals were focused onto two 16x16 photodiode arrays(15). A

similar approach was employed by Choi and Salama to simultaneously record transmembrane voltage

(RH237) and intracellular calcium (Rhod2) signals from isolated, whole guinea pig hearts (13).

Subsequently, dual-sensor optical configurations have been expanded to include EMCCD and sCMOS

sensors with improved spatial resolution (see **Table 1** for example configurations (4,5,17)).

RH237/Rhod2-AM probes are still commonly used for dual imaging (1,18–22), although dual-dye

combinations that separate fluorescence signals by emission have also been developed. They include:

Di-4-ANEPPS with Indo-1(16), Di-2-ANEPEQ and Calcium green (23), and RH237 with Fluo-3/4/5N

(12,15,24,25). Importantly, these dye combinations can have spectral overlap, which necessitates non-

ideal emission bandpass to negate spectral overlap and/or the inclusion of a calcium probe with an

inferior dissociation constant(26). A dual-sensor optical mapping system offers many advantages,

including the full spatial and temporal resolution of each individual camera. However, a dual-camera

optical setup can be both technically challenging and cost-prohibitive for basic-science and teaching

laboratories (see **Table 1** (3,4,10,11)). Dual-sensor systems also require proper alignment to ensure

that fluorescence signals are being analyzed from the same tissue region on each individual detector,

which could lead to erroneous results. Finally, the physical footprint of two cameras in a 90 degree

orientation can be limiting and reduce its versatility.

Single camera configurations

An alternative optical mapping approach includes the use of a single sensor to *sequentially* image each
complementary probe (voltage, calcium) in time using excitation light patterning (10,11,27,28) (see
Table 2). A single-sensor, sequential imaging approach was described by Lee et al., which achieved
rapid excitation light switching by utilizing recently developed high-power light emitting diodes (10).
Accordingly, single-sensor designs use dual-dye combinations that require two (or more) excitation light
sources, but share a single emission band, including: di-4-ANBDQPQ and Fura-2 (10) or Rhod2-AM
(29), Di-4-ANBDQBS and Fluo-4 (28), or Di-4-ANEPPS and X-Rhod-1(11) (for review see (26)). Single-
sensor optical mapping systems offer a cost advantage, since the camera sensor is often one of the
most expensive components of an optical mapping setup. However, a single-sensor platform design is
technically challenging since the use of two different excitation light wavelengths necessitates light
source triggering, camera synchronization and frame interleaving (10,27,28). A microcontroller is
typically used to properly time light triggering and to coordinate image acquisition, since each
(alternating) probe signal is acquired in sequential frames. This alternating configuration decreases the
frame rate for each fluorescent probe of interest, since voltage/calcium signals typically reside on
odd/even images. Excitation light ramp up/down times and shutter open/close times also have to be
taken into consideration to avoid overlap. Accordingly, single-sensor setups (with interleaved frames
per dye of interest) necessitate shortened exposure times compared with dual-sensor setups. The latter
can diminish signal-to-noise quality, without offering the same temporal fidelity.

In the described study, we describe the first multiparametric mapping system that simultaneously
acquires calcium and voltage signals from cardiac preparations, using a commercially available optical
path splitter, single camera and single excitation light (Figure 1,2). Specifically, we have taken
advantage of recently developed large field of view sCMOS sensors that are faster and more sensitive
(2048 x 2048 pixels, 18.8 mm diagonal, Zyla 4.2 plus, Andor Technology PLC, Belfast, UK). Our
configuration separates the two emission bands for Rhod2 and RH237, however, we negate the need
for a bulkier footprint and costly two camera setup by simultaneously directing each emission band to
different sides of the single, large camera sensor. To date, optical path splitters have largely been
limited to microscopy applications that allow for a slow speed of acquisition (longer exposure time) and
utilize bright immunofluorescent samples. However, such an approach has not yet been described for
multiparametric imaging of whole heart preparations that require 1) fast acquisition speeds, 2) high
spatial resolution, and 3) utilize fluorescent probes with low signal to noise ratios.

Notably, the described approach enables truly simultaneous dual imaging of cardiac preparations, while
eliminating the need for two cameras and/or multiple excitation light sources, light patterning and frame
interleaving. The described imaging platform is composed entirely of off-the-shelf components, which

can aid in the adoption and successful assembly of this setup by other laboratories. The described

protocol also employs a commonly used dual-dye combination (RH237, Rhod2-AM) that is widely

available, thereby negating the need for genetically-encoded indicators (30–32) or fluorescent probes

that are not yet commercially available (33,34). We validate the utility of our approach by performing

high-speed simultaneous dual imaging with sufficient signal-to-noise ratio for calcium and voltage

signals and specificity of emission signals with negligible cross-talk. Ventricular tachycardia was

induced to demonstrate high spatiotemporal resolution, which resulted in a novel finding that spatially

discordant calcium alternans can be present in different regions of the heart, even when electrical

alternans remain concordant. Furthermore, we highlight the versatility of our imaging platform by

seamlessly maneuvering our optical setup between a Langendorff-perfused rat heart setup (2-3 cm in

length, laying down) and Langendorff-perfused piglet heart setup (5-8 cm in length, suspended) with

different orientations. Due to technical challenges, dual optical mapping of a larger pig heart has only

been previously described by one other group (35); albeit this study employed a traditional dual camera

configuration with limited spatial resolution (see **Table 1**). Taken together, the described dual imaging

platform may be of interest to a wide variety of basic science and clinical researchers who utilize

diverse models.

RESULTS

<u>Demonstration of distinct optical emission paths</u>

Multiparametric imaging depends upon negligible cross-talk between probes, since interference

between two fluorescent dyes can lead to erroneous calculations from the acquired signals. Lack of

optical crosstalk along the light path was evaluated using a blanking plate, which was used to block

either the short/585nm light path or long/710nm light path (Figure 2D). To test for potential dye

crosstalk, experiments were performed in which Langendorff-perfused rat hearts were loaded with a

single fluorescent probe (Rhod2-AM <u>or</u> RH237), and the degree of cross-talk between the two optical

paths was assessed independently (Figure 2E, F). Representative optical signals recorded after

staining with Rhod2 are shown in Figure 2E. Note that calcium transients are distinctly visualized with

no detectable signal in the 710 nm long-pass channel (RH237 channel). Conversely, staining with

RH237 only resulted in robust voltage signals with negligible signal in the 585/40 nm channel (Rhod2

channel) with only a small max amplitude of <3 counts (Figure 2F). There was no detected

fluorescence from Rhod2 on the 710 nm long pass channel (0 ± 0 counts) compared to the maximum of

17 ± 1 counts on the 585 nm centered channel (p < 0.0001). With RH237 staining only, there was an

average maximum of 37 ± 5 counts compared to the 2 ± 1 counts on the 710 nm longpass and 585 nm

centered band, respectively, p < 0.0001). These tests demonstrate that RH237 and Rhod2 signals

selectively correspond to transmembrane voltage and intracellular calcium, respectively, with minimal

cross-talk between the two probes. The latter is comparable to previously reported dual-sensor

configurations (5,13).

<u>Use of ryanodine receptor antagonist to demonstrate signal specificity</u>

Ryanodine, a ryanodine receptor antagonist, is known to significantly impact intracellular calcium

transients (36) triggered by action potentials, with minimal effect on action potential morphology (13).

To further demonstrate signal specificity, a subset of experiments were performed on juvenile rats to

illustrate the effect of ryanodine on calcium transient versus action potential characteristics (Figure 3).

As anticipated, ryanodine-supplementation resulted in a marked reduction in the calcium transient peak

amplitude by 60% (Figure 3A,C) with no effect on action potential amplitude (Figure 3B,D). Also

expected (37), ryanodine-treatment lengthened both the calcium transient duration by 42% (CaD80,

Figure 3G) and the action potential duration by 40% (APD80, Figure 3H). This study further

demonstrates the specificity of our optical configuration, and that the acquired RH237 and Rhod2

signals are accurately separated.

<u>Simultaneous optical mapping of transmembrane voltage and intracellular calcium</u>

Experimental studies were performed to demonstrate the spatiotemporal performance of our optical

mapping system, with suitable signal-to-noise ratio (SNR) for simultaneous dual mapping. Langendorff-

perfused rat (2-3 cm length, laying down) or piglet hearts (5-8 cm length, suspended) were loaded with

both fluorescent probes (RH237, Rhod2) and optical signals were acquired concurrently (Figure 4). Dye

loading can be quickly verified by measuring regions on either image without de-interlacing, which is

necessary with other single sensor systems. Following image acquisition (1msec exposure time, 1000

fps, 768x208), optical signals were spatially averaged using a pixel radius of 30 for SNR

measurements. These acquisition settings resulted in a SNR of 74 for Rhod2 and 39 for RH237. A

slight increase in the exposure time (1.2 msec, 814 fps, 768x256) improved the SNRs to 85 for Rhod2

and 47 for RH237 (see Supplemental Table 2 for SNR calculations). In comparison, when the exposure

time was doubled (2 msec, 500 fps, 768x384), the SNR improved to 121 for Rhod2 and 58 for RH237.

No difference in fluorescence signal quality was observed within the short image acquisition time (2

sec). Repetitive imaging over the course of 1hr reduced the voltage peak fractional fluorescence

($\Delta F(p)/F_0$) from 3% to 1.9%, and the calcium peak fractional fluorescence from 4% to 3% (data not

shown, n=3).

Examples of voltage and calcium signals acquired simultaneously from an isolated rodent heart are

shown in Figure 4A, and in Supplemental video 1 & 2. As expected, action potential activation preceded

calcium transients for all pacing rates (150, 200, 220 msec cycle length). To demonstrate the versatility

of our optical mapping system, supplementary studies were performed on isolated piglet hearts that

were imaged in an upright and suspended orientation (180, 200, 220 msec cycle length, Figure 4B,
Supplemental video 3 & 4). These experimental studies highlight the utility of our setup for
simultaneous transmembrane voltage and intracellular calcium recordings of cardiac preparations, with
excellent SNR and temporal resolution.

Requisite spatiotemporal resolution for whole heart optical mapping

To demonstrate spatiotemporal resolution of our optical mapping system, we assessed excitation-
contraction coupling and measured epicardial conduction velocity. Calcium and voltage images were
briefly processed by convolution with a uniform kernel for box blurring of increasing size (Figure 5A-H),
as previously described(38). The blurring process decreased salt and pepper noise and improved the
overall quality of isochronal maps and videos with minimal effect on spatial resolution (Figure 5I). Box
blurred images were used for subsequent image analysis, such as conduction velocity measurements
(Figure 6A-D). Calcium and voltage signals were acquired with high temporal resolution upstrokes
(Figure 6C), and as expected, isochrones show an elliptical pattern of action potential activation and
calcium release that originated at the pacing site (Figure 6A, B). A shorter time delay was observed
between voltage and calcium signals at the apex of the heart compared with the base (Figure 6C) - an
anatomical difference in excitation-contraction coupling that has previously been described (13).
Notably, apex-to-base differences can result in slower conduction velocity measurements when calcium
measurements are used as a surrogate for voltage wavefront velocity in isolated whole heart
preparations (Figure 6D), as compared to cell monolayer preparations (11,39).

As an example, an isolated rat heart preparation is shown wherein longitudinal voltage conduction
velocity was 67 ± 4.6 cm/sec compared to the 45 ± 2.0 cm/sec for calcium (Figure 6). Transverse
voltage and calcium velocities were 41 ± 2.3 and 34 ± 1.6 cm/sec. In this example, spatial
heterogeneity was readily observed as the anisotropic ratio (longitudinal/transverse) was 1.62 for
voltage conduction velocity, compared to 1.32 for calcium velocity, which illustrates discontinuity in
wavefront propagation (Figure 6E,F). Indeed, viewing the time course of the voltage and calcium
activation patterns of this heart pinpointed the activity in left, basal region. At t=40 msec, an area
showed a discontinuity activation/repolarization pattern (Figure 6E), with a wavebreak clearly visible on
the calcium channel (Figure 6F). This anatomical region was consistent with the increased lag of the
calcium release following depolarization, and may suggest underlying myocardial damage. Indeed, the
described imaging platform provides sufficient spatial resolution to visualize the spread of electrical
activity and calcium cycling in cardiac preparations.

Induction of cardiac alternans to demonstrate spatial resolution

Arrhythmias are relatively uncommon in the rodent heart under basal conditions, due to species

differences in cardiac size and ion channel expression (40). Therefore, a burst pacing protocol was

employed to induce ventricular tachyarrhythmias for subsequent pathophysiological imaging. Following

burst pacing, concordant alternans were readily observed, wherein an alternating rhythm was observed

in the action potential duration and/or intracellular calcium concentration (41–44) (data not shown).

Alternans can be detected as T-wave alternans on an electrocardiogram (Figure 7A), or identified

optically, from voltage and/or calcium signals acquired from a single region of interest (Figure 7B-G).

Conversely, discordant alternans arise when alternating rhythms in different spatial regions are out of

phase. Such heterogeneities can be identified with optical mapping of fluorescence signals from

multiple sites across a cardiac preparation with high spatiotemporal resolution (45). One such example

is shown in Figure 7, in which ventricular tachycardia was encountered concomitantly with spatially

discordant calcium transient alternans following dynamic pacing (S1-S1, 70 msec cycle length, isolated

rat heart). With higher spatiotemporal resolution than previously reported (Table 1), different tissue

regions can be identified and investigated using the described single-sensor setup.

Studies suggest that calcium alternans are mechanistically responsible for action potential duration

alternans (46), however the association has not been fully elucidated. Depending on the underlying

condition and cellular function, two different variants of alternans can manifest. "Positive coupling"

describes the scenario in which a small calcium transient amplitude occurs after a shortened action

potential (42). Conversely, "negative coupling" describes a tall calcium transient amplitude that is

associated with a shortened action potential. In the current study, we show examples of both positive

and negative electromechanical coupling (in- and out-of-phase, respectively, Figure 7E,F). We report a

novel finding that calcium transient alternans can be spatially discordant, while electrical alternans

remain concordant. The described system excels in providing both a wide field of view and high spatial

resolution, which facilitated the detection of discordant alternans. Importantly, discordant alternans are

considered a precursor for reentry arrhythmias (41), as repolarization gradients become amplified when

neighboring cells are out of phase (41,43).

DISCUSSION

In the current study, we present a novel and innovative approach to dual mapping of cardiac tissue that

employs recent advancements in imaging technology. Key advantages include the simplicity of

construction (path splitter, single sensor, single light), the elimination of artifacts common for dual light

sources and dual camera designs, use of optically compatible dyes, and reduced cost barrier. This dual

mapping platform does not require the coordination of exposure frame rate and LED light triggering,

which can be technically challenging and also divides the acquisition frame rate by the number of

probes (e.g. dual mapping calcium and voltage would otherwise decrease the effective frame rate by

50%). With this configuration, individual images of calcium or voltage are each acquired,

simultaneously, at 384x256 resolution at 814 fps. If necessary for a specific application, the acquisition

rate can be increased with vertical cropping or the field of view can be expanded to decrease the frame

rate, all while maintaining SNR for the detection of calcium and voltage signals.

We validated the spatial and temporal performance of our optical mapping system using isolated whole

rat and piglet hearts. Proof-of-principle experiments were performed to measure optical action

potentials and calcium transients, spatial distribution of excitation-contraction coupling, conduction

velocity measurements in transverse/longitudinal directions, and spatiotemporal resolution of electrical

and mechanical alternans. Importantly, accurate separation of transmembrane voltage and calcium

signals was shown with single-dye loading experiments and also by application of ryanodine. In the

latter, ryanodine administration decreased calcium transient amplitude – with no discernible effect on

action potential amplitude (13). We also demonstrated that a larger heart size (piglet, 5-8 cm length)

could be accommodated without sacrificing acquisition speed by using wider strips (1024x256) and/or

utilizing a wide-angle lens. Notably, the described system has a small footprint that expands its

versatility of use between cardiac preparations of different size and orientation. The system

configuration can also be outfitted with other camera models to achieve specific temporal and spatial

resolution needs. As an example, the N256 camera (SciMedia, Costa Mesa, CA) has a 256x256 sensor

at a frame rate of 1818 fps, which could result in dual 256x128 images using the described

configuration. Finally, the described approach reduces the investment for a panoramic imaging setup

(47–49) by decreasing the total number of required cameras.

One limitation of our approach is the sensitivity of the front-illuminated sCMOS sensor, which peaks at

82% quantum efficiency, but has a roll-off at 700 nm, in the emission spectrum of RH237. We see this

limitation as an acceptable trade-off compared to the increased speed that the camera provides

compared to back-illuminated options. We did not characterize the loss in quantum yield from the

optical path splitter, though we anticipate losses that are comparable to other dichroic based emission

splitting systems. The optical path was optimized by using high transmission band (>95%) filters and a

high-speed lens (f/0.95 or f/1.4). Another potential limitation of the camera is the small pixel size of 6.5

μm, which is not necessarily required for applications at the tissue level, such as whole heart

preparations. We have mitigated the small pixel size by employing box blur algorithms during post-

acquisition processing when necessary to improve signal fidelity (binning could also be used). A single

excitation light source, single camera platform can complicate the control of dual emission intensities;

although we did not encounter this problem due to the similar SNR of the dyes and wide (16-bit)

dynamic range of the camera. Finally, a slight difference in the focal plane between 585nm versus

710nm can occur due to chromatic focal shift – although this did not impact fluorescence signals at the

tissue level. Color-corrected lenses can be employed to minimize chromatic aberration.

CONCLUSION

We developed a novel platform to simultaneously acquire voltage and calcium signals from intact,

whole heart preparations using an optical path splitter, single camera and single excitation light.

Notably, the described platform is composed entirely of off-the-shelf components, which can aid in the

adoption and successful implementation of this setup by other laboratories. The described protocol also

employs a readily available dye combination (RH237, Rhod-2AM), thereby negating the need for

genetically-encoded indicators or fluorescent probes that are not commercially available. We confirm

the specificity of our emission signals with negligible cross-talk and demonstrate high spatiotemporal

resolution for excitation-contraction coupling, conduction velocity, and concordant/discordant alternans

research. The described platform is easy to align and has a small footprint, which provides flexibility in

its use for tissue preparations of varying size and/or orientation. In summary, the described platform

requires minimal technical complexity, yet provides the spatiotemporal resolution and signal-to-noise

ratio necessary for the investigation of voltage/calcium kinetics, conduction velocity and arrhythmia

incidence.

METHODS

Animal procedures were approved by the Institutional Animal Care and use Committee of the Children's

Research Institute, and followed the National Institutes of Health's *Guide for the Care and Use of*

Laboratory Animals. All animals were euthanized by exsanguination under anesthesia during heart

excision, as detailed below.

Isolated rodent heart preparation

Unless otherwise noted, experiments were conducted using adult male Sprague-Dawley rats (2-3

months of age, 250-350g, n=11, Taconic Biosciences). Animals were anesthetized with 3-5%

isoflurane, the heart was excised and then transferred to a temperature-controlled (37°C) constant-

pressure (70 mmHg) Langendorff perfusion system. Excised hearts were perfused with Krebs-Henseleit

buffer throughout the duration of the experiment, as previously described (<1 hour)(12,50,51). Three

electrodes were positioned to acquire a far-field electrocardiogram in the lead II configuration. In a

subset of experiments, male juvenile rats (postnatal day 5, 10 grams, n=6) were used to demonstrate

the effects of ryanodine-supplementation on electromechanical coupling.

Isolated piglet heart preparation

Yorkshire piglets (2-4 weeks of age, n=4) were used in a supplementary study to demonstrate the
versatility of the optical setup. Briefly, an intravenous bolus injection of fentanyl (50µg/kg) and
rocuronium (1mg/kg) was administered, and anesthesia was maintained with isoflurane (1-3%),
fentanyl (10-25µg/kg) and pancuronium (1mg/kg). The heart was excised and submerged in
cardioplegia (4°C) and then flushed with Krebs-Henseleit solution by aortic cannulation. The heart was
transferred (~10 minutes at room temperature) to a temperature-controlled (37°C) constant-pressure
(70 mmHg) Langendorff-perfusion system. To avoid ischemic-injury in these larger tissue preparations,
the heart was suspended in contrast to the "laying" rat heart. Accordingly, the dual mapping platform
(camera, path splitter) was relocated in proximity to the larger capacity perfusion system.

Once a baseline heart rate was established (10 min), the perfusate was supplemented with 10 µM (-/-)
blebbistatin (Sigma-Aldrich) to reduce motion artifact for subsequent imaging experiments (52,53).
Fluorescent dyes were added sequentially, as a bolus, through a bubble trap located proximal to the
aortic cannula. Based on predetermined myocardial staining time for each dye, 50 µg Rhod-2AM (1,50)
was added first and allowed to stabilize for 10 min, followed by 62.1 µg RH-237 staining for 1 min (12).
The myocardial tissue was re-stained by RH237 if needed throughout the duration of the experiment
(54). Homogenous dye loading is shown in Supplemental Figure 1. For electrical stimulation, a coaxial
stimulation electrode was positioned on the ventricular epicardium, which was driven by a Bloom
Classic electrophysiology stimulator (Fischer Medical). Pacing current was set to 1.5x the threshold
(resulting in typically 1.8 mA), with 1 msec monophasic pulses.

Ryanodine administration
To demonstrate negligible cross-talk between voltage and calcium signals, a subset of experiments
was performed in the presence of ryanodine (n=3). Ryanodine has previously been shown to
selectively affect calcium transient upstroke and amplitude(13). 10 µM ryanodine was added as a
bolus, through a bubble trap located proximal to the aortic cannula. Hearts were subjected to
continuous pacing during imaging (180 msec), and images were captured before and immediately after
application of ryanodine (30 seconds, 1 and 3 minutes).

Instrumentation
The overall system configuration is shown in Figure 1. The epicardium was illuminated using broad light
emitting diode (LED) spotlights centered at 530 nm (Mightex), equipped with an optical filter to constrict
the excitation band (ET530/40x nm, Chroma Technologies). LED radiant power varied between 50 –
200 mW, as needed. At the onset of imaging, the excitation LEDs were automatically enabled via a
direct TTL connection from the camera output. Due to the use of a single light source for excitation,

complex light triggering was not needed to coordinate alternating LED light sources with individual
frames.

An image splitting device (OptoSplit II, Cairn Research, Ltd) was positioned in front of a sCMOS
camera (Andor Technology, PLC, Zyla 4.2 PLUS); the corresponding light path is shown in Figure 1B.
The beam splitter was configured with a dichroic mirror (660+nm, Chroma Technologies, see
Supplemental Table 1) that passed RH237 emission and reflected Rhod2 fluorescence. High
transmission emission filters were used for Rhod2 (ET585/40m, Chroma Technologies) and RH237
emitted light (long pass ET710, Chroma Technologies). A fixed focal length 17mm/F0.95 lens was
attached to the front of the imaging splitting device (Schneider, #21-010456) for rat hearts, and a wide-
angle 6mm/F1.2 lens (Fujinon, #DF6HA-1B) was used for pig hearts. Details of the optical configuration
are shown in Figure 1B,C and the experimental workflow is displayed in Figure 1D.

MetaMorph v7.10.2.240 (Molecular Devices, LLC) was used for camera configuration, optosplit image
alignment and LED on/off triggering. To guide manual optosplit alignment, *MetaMorph* overlays the live
images as contrasting colors or as subtractive greyscales to highlight misalignment. With this live
feedback, images are quickly aligned (<1 minute) using the optosplit's "long" and "short" control knobs
(Figure 2). As an alternative to *MetaMorph*, images could also be aligned with the free software
µManager (55). After alignment, any standard image acquisition software can be used such as
MetaMorph, µManager, or *Solis* (Andor Technologies, software supplied with camera). The acquired
image will include two fields, which can be separated using *MetaMorph*, *µManager*, or alternative
imaging software that includes automated tools. LEDs were attached to a controller (SLC-SA04-U/S,
Mightex) that was triggered "on" 1 second prior to image acquisition. The computer consisted of a Xeon
CPU E3-1245 v5 3.50 GHz (Intel corporation), 32 GB of RAM, and a non-volatile memory express solid
state disk (NVMe SSD, Samsung 960 Pro). A frame grabber was used for imaging control and
acquisition (Karbon #KBN-PCE-CL4-F, BitFlow). Because of the high data rate of acquisition (due to
high spatial and temporal resolution and bit depth), the NVMe SSD disk was essential for reducing data
rate bottlenecks. The frame grabber with 10-tap CameraLink^{TM} connection was necessary to achieve
the fastest frame rates possible; a USB connection would result in much slower frame rates.

To maximize spatiotemporal resolution, the sCMOS camera sensor was cropped and set to an
acquisition rate of 814 frames per second (fps). The exposure time to achieve this frame rate was 1.206
msec, with a 98.4% duty cycle. This configuration resulted in two images from the splitting device, each
384x256 pixels, with a 16 pixel boundary between each image to negate optical crosstalk (Figure 2D).
The field of view was approximately 2.1 x 1.4 cm, which was sufficient to fully image pediatric rat hearts
(0.3 g average weight, 1.7 cm length from aortic root to apex). For the larger adult rat hearts (1 g

average weight, 2.7 cm in length), the field of view was extended by increasing the working distance.

The actual pixel size on the sensor is 6.5 μm with a projected pixel size typically 45-80 μm – depending

on working distance and field of view from a given lens. The front of the image splitter is a standard C-

mount, which allows a user to select any standard lens to use in combination with accessories (back

extension tubes, diopters) to adjust field of view and working distance. The projected pixel size was

measured for each study and individually calibrated for conduction velocity measurements. If needed,

the frame rate could be increased further by vertical cropping; the feasibility of which was tested with

208 vertical pixels at 1000 fps, which resulted in adequate SNR on the epicardial surface (see results

section for details). For supplemental piglet heart studies (80-120 g average weight, 5-8 cm length), a

wide-angle lens was used in conjunction with a larger area of the sensor to attain a field of view

approximately 5.9 x 4.7 cm.

<u>Signal and image processing</u>

Following image acquisition, signal or image processing was performed as outlined in Figure 1D. Signal

processing and data analysis were performed using a custom MATLAB script(26,51). A circular region

of interest was taken in the center of the raw image with a 30-pixel radius (5-20 mm^2 area on the heart,

depending on lens), averaged, and plotted against time. Drift removal and temporal smoothing were

applied when necessary. Drift removal was performed by subtraction of a polynomial fit (0th, 1st, or 2nd

order). To remove high frequency noise, a 5th order Butterworth low-pass filter was applied to the

resulting signals with a cut-off frequency adjusted between 50-150 Hz. After pre-processing, a peak

detector was used to measure the total number of action potentials or calcium transients in the file

across time. Characteristics from each event are measured and averaged, including action potential

duration, calcium transient duration, time to 90% peak, and amplitude as described

previously(26,50,56,57).

Image processing was performed and isochrone maps were constructed in the *rhythm* software(38).

The background was removed, convolved with a uniform kernel for box blurring(38) and then time

signals were low-pass filtered below 100 Hz. A similar approach was previously described by Laughner,

et al. 2012 in which a 3x3 kernel size was used to process images acquired from a MiCAM ULTIMA

(100x100 pixels, 100 μm pixel size). We chose a larger uniform kernel size (15x15) since our camera

sensor (Zyla 4.2) has sharper resolution (1024x1024 pixels, 6.5 μm pixel size). The activation time of

every pixel on the heart was defined as the maximum derivative of the action potential or transient

upstroke, which was plotted for both voltage and calcium, respectively. As proof of concept, we also

tested the feasibility of measuring conduction velocity across the epicardial surface by exporting the

activation maps and analyzing via ORCA(58). Subsequent electrical wave propagation images were

constructed using custom Python scripts and plotted with matplotlib(59). Signal-to-noise ratio (SNR) of

the time-series was calculated as the action potential amplitude (ΔF) over the standard deviation of the

baseline during the diastolic interval(60). The number of pixels used to calculate the SNR was noted

when necessary; importantly, the SNR can be improved considerably after box blurring and/or spatial

averaging.

Statistical analysis

Statistical tests were performed using the R software package. Datasets from the ryanodine

experiments were compared pairwise using Welch's T-test and significance reported as $p < 0.05$.

ABBREVIATIONS

APD_{30} action potential duration at 30% repolarization

Ca   intracellular calcium signal

CaD_{30} calcium transient duration at 30% relaxation

MEHP   mono-2-ethylhexyl phthalate

Rr   ryanodine

SNR   signal to noise ratio

Vm   transmembrane voltage

DECLARATIONS

Ethics approval and consent to participate: Animal procedures were approved by the Institutional

Animal Care and use Committee of the Children's Research Institute, and followed the National

Institutes of Health's *Guide for the Care and Use of Laboratory Animals*.

Consent for publication: Not applicable

Availability of data and material: The datasets used and/or analyzed during the current study available

from the corresponding author on reasonable request.

Competing Interests: The authors declare that they have no competing interests.

Funding: This work was supported by the National Institutes of Health (R00ES023477 to N.G.P,

R01HL139472 to N.G.P), Children's Research Institute and Children's National Heart Institute. We

thank the generosity of the NVIDIA corporation for the graphics processing unit to perform CUDA-

enabled image processing. Funding sources had no role in the design of the study, collection, analysis,

interpretation of the data or writing the manuscript.

<u>Author contributions:</u> RJ, DM, BS, LS, JH, DM, NGP performed experiments; RJ, DM, BS, NGP
analyzed data; RJ, DM, NGP prepared figures; RJ and NGP drafted manuscript, RJ and NGP
conceived and designed experiments; RJ, DM, BS; LS, JH, DM and NGP approved manuscript.
491
492 <u>Acknowledgements:</u> The authors gratefully acknowledge Narine Sarvazyan, Charles Berul, and Yu-Ling
493 Shao for helpful discussions, Manelle Ramadan and Morgan Burke for technical assistance, and
494 Nobuyuki Ishibashi and Takuya Maeda for biological materials.
495

REFERENCES

1. Lang D, Sulkin MS, Lou Q, Efimov IR. Optical Mapping of Action Potentials and Calcium
Transients in the Mouse Heart. J Vis Exp. 2011;(55).
2. Efimov IR, Nikolski VP, Salama G. Optical imaging of the heart. Circ Res. 2004 Jul;95(1):21–33.
3. Herron TJ, Lee P, Jalife J. Optical imaging of voltage and calcium in cardiac cells & tissues. Circ
Res. 2012 Feb;110(4):609–23.
4. Entcheva E, Bien H. Macroscopic optical mapping of excitation in cardiac cell networks with
ultra-high spatiotemporal resolution. Prog Biophys Mol Biol. 2006 Oct;92(2):232–57.
5. Salama G, Hwang SM. Simultaneous optical mapping of intracellular free calcium and action
potentials from langendorff perfused hearts. 2009/07/04. Vol. Chapter 12, Current Protocols in
Cytometry. 2009. p. Unit 12 17-Unit 12 17.
6. Salama G, Morad M. Merocyanine 540 as an optical probe of transmembrane electrical activity in
the heart. Science. 1976 Feb;191(4226):485–7.
7. Fluhler E, Burnham VG, Loew LM. Spectra, membrane binding, and potentiometric responses of
new charge shift probes. Biochemistry. 1985 Oct;24(21):5749–55.
8. Efimov IR, Huang DT, Rendt JM, Salama G. Optical mapping of repolarization and refractoriness
from intact hearts. Circulation. 1994 Sep;90(3):1469–80.
9. Fabiato A, Fabiato F. Use of chlorotetracycline fluorescence to demonstrate Ca2+-induced
release of Ca2+ from the sarcoplasmic reticulum of skinned cardiac cells. Nature. 1979 Sep
13;281(5727):146–8.
10. Lee P, Bollensdorff C, Quinn TA, Wuskell JP, Loew LM, Kohl P. Single-sensor system for
spatially resolved, continuous, and multiparametric optical mapping of cardiac tissue. Heart
Rhythm. 2011 Sep;8(9):1482–91.
11. Scull JA, McSpadden LC, Himel HD, Badie N, Bursac N. Single-detector simultaneous optical
mapping of V(m) and [Ca(2+)](i) in cardiac monolayers. Ann Biomed Eng. 2012 May;40(5):1006–
17.
12. Swift L, Martell B, Khatri V, Arutunyan A, Sarvazyan N, Kay M. Controlled regional hypoperfusion
in Langendorff heart preparations. Physiol Meas. 2008 Feb;29(2):269–79.

13.   Choi BR, Salama G. Simultaneous maps of optical action potentials and calcium transients in
guinea-pig hearts: mechanisms underlying concordant alternans. J Physiol. 2000 Nov;529 Pt
1(2000):171–88.
14.   Atherton PJ, Babraj J, Smith K, Singh J, Rennie MJ, Wackerhage H. Selective activation of
AMPK-PGC-1alpha or PKB-TSC2-mTOR signaling can explain specific adaptive responses to
endurance or resistance training-like electrical muscle stimulation. FASEB J. 2005
May;19(7):786–8.
15.   Fast VG, Ideker RE. Simultaneous optical mapping of transmembrane potential and intracellular
calcium in myocyte cultures. J Cardiovasc Electrophysiol. 2000;11(5):547-56.
16.   Laurita KR, Singal A. Mapping action potentials and calcium transients simultaneously from the
intact heart. Am J Physiol Hear Circ Physiol. 2001;280(5):H2053--60.
17.   Omichi C, Lamp ST, Lin S-F, Yang J, Baher A, Zhou S, et al. Intracellular Ca dynamics in
ventricular fibrillation. Am J Physiol Heart Circ Physiol. 2004 May;286(5):H1836-44.
18.   Lang D, Holzem K, Kang C, Xiao M, Hwang HJ, Ewald GA, et al. Arrhythmogenic Remodeling of
Beta2 Versus Beta1 Adrenergic Signaling in the Human Failing Heart. Circ Arrhythmia
Electrophysiol. 2015 Feb;8(2):409–19.
19.   Myles RC, Wang L, Kang C, Bers DM, Ripplinger CM. Local β-Adrenergic Stimulation
Overcomes Source-Sink Mismatch to Generate Focal Arrhythmia. Circ Res. 2012/04/28. 2012
Apr;110(11):1454–64.
20.   Lou Q, Li W, Efimov IR. Multiparametric Optical Mapping of the Langendorff-perfused Rabbit
Heart. Journal of Visualized Experiments. 2011.
21.   Holcomb MR, Woods MC, Uzelac I, Wikswo JP, Gilligan JM, Sidorov VY. The potential of dual
camera systems for multimodal imaging of cardiac electrophysiology and metabolism. Exp Biol
Med (Maywood). 2009 Nov;234(11):1355–73.
22.   Hwang G-S, Hayashi H, Tang L, Ogawa M, Hernandez H, Tan AY, et al. Intracellular Calcium
and Vulnerability to Fibrillation and Defibrillation in Langendorff-Perfused Rabbit Ventricles.
Circulation. 2006 Dec 12;114(24):2595–603.
23.   Bullen A, Saggau P. Indicators and optical configuration for simultaneous high-resolution
recording of membrane potential and intracellular calcium using laser scanning microscopy.
Pfl☐gers Arch Eur J Physiol. 1998 Aug 24;436(5):788–96.
24.   Wang L, Myles RC, De Jesus NM, Ohlendorf AKP, Bers DM, Ripplinger CM. Optical mapping of
sarcoplasmic reticulum Ca2+ in the intact heart: ryanodine receptor refractoriness during
alternans and fibrillation. Circ Res. 2014 Apr;114(9):1410–21.
25.   Kong W, Walcott GP, Smith WM, Johnson PL, Knisley SB. Emission ratiometry for simultaneous
calcium and action potential measurements with coloaded dyes in rabbit hearts: reduction of
motion and drift. J Cardiovasc Electrophysiol. 2003 Jan;14(1):76–82.

26.   Jaimes R, Walton RD, Pasdois PLC, Bernus O, Efimov IR, Kay MW. A Technical Review of

561   Optical Mapping of Intracellular Calcium within Myocardial Tissue. Am J Physiol Heart Circ

562   Physiol. 2016 Mar 25;ajpheart.00665.2015.

563   27.   Yamanaka T, Arafune T, Shibata N, Honjo H, Kamiya K, Kodama I, et al. Single camera system

564   for multi-wavelength fruorescent imaging in the heart. In: Proceedings of the Annual International

565   Conference of the IEEE Engineering in Medicine and Biology Society, EMBS. IEEE; 2012. p.

566   3716–9.

567   28.   Lee P, Yan P, Ewart P, Kohl P, Loew LM, Bollensdorff C. Simultaneous measurement and

568   modulation of multiple physiological parameters in the isolated heart using optical techniques.

569   Pflugers Arch Eur J Physiol. 2012 Aug;

570   29.   Lee P, Taghavi F, Yan P, Ewart P, Ashley EA, Loew LM, et al. In situ optical mapping of voltage

571   and calcium in the heart. PLoS One. 2012;7.

572   30.   Vogt CC, Bruegmann T, Malan D, Ottersbach A, Roell W, Fleischmann BK, et al. Systemic gene

573   transfer enables optogenetic pacing of mouse hearts. Cardiovasc Res. 2015 May 1;106(2):338–

574   43.

575   31.   Bruegmann T, Malan D, Hesse M, Beiert T, Fuegemann CJ, Fleischmann BK, et al. Optogenetic

576   control of heart muscle in vitro and in vivo. Nat Methods. 2010 Nov;7(11):897–900.

577   32.   Ambrosi CM, Sadananda G, Klimas A, Entcheva E. Adeno-associated virus mediated gene

578   delivery: Implications for scalable in vitro and in vivo cardiac optogenetic models. bioRxiv. 2017

579   Aug 31;183319.

580   33.   Matiukas A, Mitrea BG, Pertsov AM, Wuskell JP, Wei M, Watras J, et al. New near-infrared

581   optical probes of cardiac electrical activity. Am J Physiol Heart Circ Physiol. 2006

582   Jun;290(6):H2633-43.

583   34.   Matiukas A, Mitrea BG, Qin M, Pertsov AM, Shvedko AG, Warren MD, et al. Near-infrared

584   voltage-sensitive fluorescent dyes optimized for optical mapping in blood-perfused myocardium.

585   Heart Rhythm. 2007 Nov;4(11):1441–51.

586   35.   Lee P, Calvo CJ, Alfonso-Almazán JM, Quintanilla JG, Chorro FJ, Yan P, et al. Low-Cost Optical

587   Mapping Systems for Panoramic Imaging of Complex Arrhythmias and Drug-Action in

588   Translational Heart Models. Sci Rep. 2017 Feb 27;7:43217.

589   36.   Escobar AL, Ribeiro-Costa R, Villalba-Galea C, Zoghbi ME, Pérez CG, Mejía-Alvarez R.

590   Developmental changes of intracellular Ca2+ transients in beating rat hearts. Am J Physiol Heart

591   Circ Physiol. 2004 Mar 1;286(3):H971-8.

592   37.   Fauconnier J, Bedut S, Le Guennec J-Y, Babuty D, Richard S. Ca2+ current-mediated regulation

593   of action potential by pacing rate in rat ventricular myocytes. Cardiovasc Res. 2003 Mar

594   1;57(3):670–80.

38.   Laughner JI, Ng FS, Sulkin MS, Arthur RM, Efimov IR. Processing and analysis of cardiac optical

mapping data obtained with potentiometric dyes. Am J Physiol Circ Physiol. 2012

Oct;303(7):H753-65.

39. Gillum N, Karabekian Z, Swift LMM, Brown RPP, Kay MWW, Sarvazyan N. Clinically relevant

concentrations of di (2-ethylhexyl) phthalate (DEHP) uncouple cardiac syncytium. Toxicol Appl

Pharmacol. 2009 Apr;236(1):25–38.

40. Kaese S, Verheule S. Cardiac electrophysiology in mice: a matter of size. Front Physiol. 2012

Jan;3:345.

41. Wilson LD, Rosenbaum DS. Mechanisms of arrythmogenic cardiac alternans. Europace. 2007

Nov;9(Supplement 6):vi77–82.

42. Sato D, Shiferaw Y, Garfinkel A, Weiss JN, Qu Z, Karma A. Spatially discordant alternans in

cardiac tissue: role of calcium cycling. Circ Res. 2006 Sep;99(5):520–7.

43. Qu Z, Xie Y, Garfinkel A, Weiss JN. T-wave alternans and arrhythmogenesis in cardiac diseases.

Front Physiol. 2010 Jan;1:154.

44. Garfinkel A. Eight (or more) kinds of alternans. J Electrocardiol. 2007;40(6 Suppl):S70-4.

45. Uzelac I, Ji YC, Hornung D, Schröder-Scheteling J, Luther S, Gray RA, et al. Simultaneous

Quantification of Spatially Discordant Alternans in Voltage and Intracellular Calcium in

Langendorff-Perfused Rabbit Hearts and Inconsistencies with Models of Cardiac Action

Potentials and Ca Transients. Front Physiol. 2017;8:819.

46. Chudin E, Goldhaber J, Garfinkel A, Weiss J, Kogan B. Intracellular Ca(2+) dynamics and the

stability of ventricular tachycardia. Biophys J. 1999 Dec;77(6):2930–41.

47. Gloschat C, Aras K, Gupta S, Faye NR, Zhang H, Syunyaev RA, et al. RHYTHM: An Open

Source Imaging Toolkit for Cardiac Panoramic Optical Mapping. Sci Rep. 2018 Dec;8(1):2921.

48. Kay MW, Amison PM, Rogers JM. Three-dimensional surface reconstruction and panoramic

optical mapping of large hearts. IEEE Trans Biomed Eng. 2004 Jul;51(7):1219–29.

49. Lou Q, Ripplinger CM, Bayly P V, Efimov IR. Quantitative panoramic imaging of epicardial

electrical activity. Ann Biomed Eng. 2008 Oct;36(10):1649–58.

50. Jaimes R, Kuzmiak-Glancy S, Brooks D, Swift LM, Posnack NG, Kay MW. Functional response

of the isolated, perfused normoxic heart to pyruvate dehydrogenase activation by dichloroacetate

and pyruvate. Pflügers Arch - Eur J Physiol. 2016 Jul;468(1):131–42.

51. Posnack NGNG, Jaimes R, Asfour H, Swift LMLM, Wengrowski AMAM, Sarvazyan N, et al.

Bisphenol A Exposure and Cardiac Electrical Conduction in Excised Rat Hearts. Environ Health

Perspect. 2014 Jan;122(4):384–90.

52. Swift LMLM, Asfour H, Posnack NGNG, Arutunyan A, Kay MWMW, Sarvazyan N. Properties of

blebbistatin for cardiac optical mapping and other imaging applications. Pflugers Arch Eur J

Physiol. 2012 Sep;464(5):503–12.

53. Fedorov V V, Lozinsky IT, Sosunov EA, Anyukhovsky EP, Rosen MR, Balke CW, et al.

Application of blebbistatin as an excitation-contraction uncoupler for electrophysiologic study of

rat and rabbit hearts. Hear Rhythm. 2007 May;4:619–26.

54.   Del Nido PJ, Glynn P, Buenaventura P, Salama G, Koretsky AP. Fluorescence measurement of

calcium transients in perfused rabbit heart using rhod 2. Am J Physiol. 1998 Feb;274(2 Pt

2):H728--41.

55.   Edelstein AD, Tsuchida MA, Amodaj N, Pinkard H, Vale RD, Stuurman N. Advanced methods of

microscope control using μManager software. J Biol Methods. 2014 Nov 7;1(2):10.

56.   Posnack NGNG, Idrees R, Ding H, Jaimes Iii R, Stybayeva G, Karabekian Z, et al. Exposure to

phthalates affects calcium handling and intercellular connectivity of human stem cell-derived

cardiomyocytes. PLoS One. 2015 Jan;10(3):e0121927–e0121927.

57.   Ramadan M, Sherman M, Jaimes R, Chaluvadi A, Swift L, Posnack NG. Disruption of neonatal

cardiomyocyte physiology following exposure to bisphenol-a. Sci Rep. 2018;8(1).

58.   Doshi AN, Walton RD, Krul SP, de Groot JR, Bernus O, Efimov IR, et al. Feasibility of a semi-

automated method for cardiac conduction velocity analysis of high-resolution activation maps.

Comput Biol Med. 2015;65:177–83.

59.   Hunter JD. Matplotlib: A 2D Graphics Environment. Comput Sci Eng. 2007;9(3):90–5.

60.   Lee P, Calvo CJ, Alfonso-Almazán JM, Quintanilla JG, Chorro FJ, Yan P, et al. Low-Cost Optical

Mapping Systems for Panoramic Imaging of Complex Arrhythmias and Drug-Action in

Translational Heart Models. Sci Rep. 2017 Feb;7:43217.

FIGURE LEGENDS
654

Figure 1: **Overview of the optical design and workflow.** (**A**) Optical system configuration with image
splitting device positioned in front of a sCMOS camera (Zyla 4.2 PLUS, Andor Technologies), (**B**)
Emission of each complementary probe (Vm, Ca) is separated by wavelength using an image splitting
device (**C**) Dichroic cube setup with the two emission filters and a dichroic mirror. (**D**) Experimental
workflow includes image acquisition, followed by signal processing using a custom MATLAB script or
image processing in Rhythm(38) and conduction velocity analysis in ORCA(58). Vm = transmembrane
voltage, Ca = intracellular calcium signal.

Figure 2: **Optical configuration results in negligible crosstalk between fluorescent probes.** (**A**)
Passbands of all the filters used in the optical configuration. (**B**) Two sides of a single, large camera
sensor display a duplicated ruler image from each emission band with negligible focal shift. (**C**)
Illustration of misaligned (right) and proper alignment (left) of images. (**D**) **i**: 800x256 pixel image of a
back illuminated ruler is shown before splitting into two 384x256 pixel images; left: long wavelength
path (710+nm), middle: black boundary between images (16 pixels), right: short wavelength path
(585$\pm$20nm). **(ii)** 800x256 image when the a plate is inserted to block the light path on the left
(710+nm). **(iii)** 800x256 image when the path on the right is blocked (585 $\pm$ 20nm). **(iiii)** Profile traces
corresponding to the full image (black), left path blocked (blue, 710+nm) and right path blocked (red,
585$\pm$20nm) demonstrate a lack of crosstalk along the light paths. Arrow denotes location of the
boundary between images without crosstalk. (**E**) Left: Detected fluorescence on the epicardial surface
after single probe administration (Rhod2). Center: High quality calcium transients through the 585 nm
emission filter with no detectable signal through the 710 nm longpass filter. Right: Maximum counts
from each channel after Rhod2 (only) administration. (**F**) Left: Detected fluorescence on the epicardial
surface after single probe administration (RH237). Center: High quality optical action potentials through
the 710 nm emission filter, with minimal signal through the 585 nm emission. Right: Maximum counts
from each channel after RH237 (only) administration. Note: Bar plots (E,F) on the right collected from
different rat hearts, each loaded with a single fluorescent probe (n = 3 each plot). Scale bar = 1cm.

Figure 3: **Use of ryanodine receptor antagonist to demonstrate signal specificity.** Representative
optical action potential **(A)** and calcium transient **(B)** recorded under control perfusion (Ctrl) or
immediately after administration of ryanodine (Ry, 30 sec − 3 min). Ryanodine (Ry) administration
significantly reduced peak calcium transient amplitudes (**A, C**), with no effect on optical action potential
amplitudes (**B, D**). Normalized fluorescent signals were used for time measurements. Calcium transient
(**C, E, G**) and action potential (**D, F, H**) parameters are shown before and 1 minute after ryanodine
administration at a pacing frequency of 180 msec cycle length. Mean $\pm$ SEM, n=3 juvenile rat hearts.

Figure 4: Multiparametric optical signals acquired from Langendorff-perfused rat and piglet hearts. **(A)**

Normalized transmembrane voltage (RH237, red) and intracellular calcium (Rhod2, orange)

fluorescence signals acquired simultaneously from an excised rat heart during electrical pacing (150,

200, 220 msec pacing cycle length). In this example, each individual image represents 384x256 pixels

acquired at 814fps. **(B)** Normalized fluorescence signals acquired simultaneously from excised piglet

hearts during electrical pacing (180, 200, 220 msec pacing cycle length). In this specific example, each

individual image represents 640x512 pixels acquired at 406fps. Circle denotes region of interest (rat: 45

pixel region, piglet: 80 pixel region in this example). Note: Schneider 17mm f/0.95 lens used for rat

hearts, Fujinon 6mm f/1.2 lens used for piglet hearts. Vm = transmembrane voltage, Ca = intracellular

calcium signal. ROI = region of interest. Scale bar = 1 cm.

Figure 5: Spatial filtering of optical signals using box blurring. The fluorescence intensity of each

pixel is averaged with neighboring pixels. **(A-F)** 384x256 pixel images were convolved with an odd

uniform kernel for box blurring to improve SNR for isochrone analysis. Salt and pepper noise is

decreased with increasing kernel size for both Vm (top) and Ca (bottom) signals. (**G, H**) Traces show

increased SNR with increasing kernel size. Note: The displayed signal is measured from a single pixel

(ROI = region of interest), whereas typically for signal analysis an averaged region of interest is used

consisting of several pixels from images that have not been box blurred. **(I)** Spatial resolution is

demonstrated in the raw image (purple), 5x5 box blurred image (red) and 15x15 box blurred image

(blue). Corresponding edge profile trace illustrates spatial resolution of approximately 0.8mm (raw, 5x5

box blurring), 0.86mm (9x9 box blurring) to 1.2mm (15x15 box blurring). Scale bar = 1cm

Figure 6: Images from simultaneous acquisition of voltage and calcium signals. (**A**) Voltage

activation isochronal map. (**B**) Calcium release isochronal map. (**C**) Voltage and calcium upstrokes

display anatomical differences in excitation-contraction coupling (apex versus base). (**D**) Transverse (T)

and longitudinal (L) conduction velocity computed from voltage and calcium wavefront propagation.

Longitudinal propagation velocity of the calcium wavefront was significantly slower than the voltage

conduction velocity (t-test, p<0.001). Slower calcium conduction velocities coincide with increased lag

time between the action potential and calcium transient near the base of the heart. Conduction velocity

was calculated from 6 angles for transverse and 2 angles for longitudinal measurements. (**E**)

Sequential images of impulse propagation show the voltage wavefront originating from the pacing

electrode (arrow) and propagating toward the base. Spatial hetereogeneity was observed in the

activation and repolarization pattern at t = 40 msec (**B**) Sequential images of the calcium wavefront

follows the voltage wavefront, with an example of folding observed at t = 40 msec. Arrows denote

pacing electrode; Vm = transmembrane voltage, Ca = intracellular calcium signal. Scale bar = 1cm

Figure 7: **Simultaneous calcium and voltage imaging during ventricular tachycardia.** (**A**) T-wave alternans (arrows) on alternating waveforms recording from an isolated, intact heart. **(B, C**) APD30 map shows repolarization heterogeneity during an odd "O" and even "E" beats. (**D,E**) CaD30 signals also show reuptake heterogeneity across the epicardial surface during odd and even beats. Optical action potentials and calcium transients display alternans that can be in-phase **(F)** or out-of-phase **(G)**. In-phase alternans are represented by a small calcium transient amplitude following a shortened action potential duration. Out-of-phase alternans describe a tall calcium transient amplitude that follows a shortened action potential duration. APD30 = Action potential duration at 30% repolarization; CaD30 = Calcium transient duration at 30% relaxation; R1 & R2 = regions of interest; Vm = transmembrane voltage; Ca = intracellular calcium signal. Scale bar = 1cm

738 **SUPPLEMENTAL FILE LIST**
739
Supplemental Video 1

Transmembrane voltage mapping of rat heart epicardium. An adult rat heart was imaged during
ventricular pacing at cycle length 250 msec. The voltage wavefront can be seen originating from the
center of the posterior ventricle and propagating across the surface, followed by a slight delay and
retrograde atrial conduction. Calcium imaging was performed simultaneously (see Supplement Video
2). The images were box blurred using a 15 x 15 uniform kernel and the length of the heart is
approximately 2.5 cm from base to apex.
747

748 **Supplemental Video 2**

749 **Simultaneous calcium mapping of rat heart epicardium.** The calcium activity of the adult rat heart
750 was mapped concurrently with transmembrane voltage. The calcium wavefront can be seen following
751 the voltage propagation from Supplement Video 1. The images were box blurred using a 15 x 15
752 uniform kernel and the length of the heart is approximately 2.5 cm from base to apex.
753

Supplemental Video 3

Transmembrane voltage mapping of pig heart epicardium. A juvenile pig heart was imaged during
an episode of ventricular tachycardia (cycle length = 150 msec). Circus movement with wave collision
can be observed. Calcium imaging was performed simultaneously (see Supplement Video 4). The
images were box blurred using a 15 x 15 uniform kernel and the length of the heart is approximately 5.0
cm from base to apex.
760

Supplemental Video 4

Simultaneous calcium mapping of pig heart epicardium. A juvenile pig heart was imaged during an
episode of ventricular tachycardia (cycle length = 150 msec). Circus movement with wave collision can
be observed. Transmembrane voltage imaging was performed simultaneously (see Supplement Video
3). The images were box blurred using a 15 x 15 uniform kernel and the length of the heart is
approximately 5.0 cm from base to apex.
767

768
769
770

771 **TABLES**

773 **Table 1:** Example of optical mapping configurations using dual sensors

SENSOR	*Andor Zyla 4.2 PLUS	Photometrics EMCCD	SciMedia N256	IDS UI-3220CP-M-GL	IDS UI-3360CP-NIR-GL
PEAK QE	82%	>92%	55%	50%	60%
SENSOR SIZE (MM)	1.7 x 2.5	3.1 x 3.1	17.6 x 17.6	2.9 x 4.5	6 x 11.3
PIXEL SIZE (UM)	6.5	24	69	6	5.5
NO. OF PIXELS	256 x 384 (rat) 640 x 512 (pig)	128 x 128	256 x 256	120 x 160	160 x 220
DIGITIZATION	16-bit	16-bit	14-bit	10-bit	12-bit
EFFECTIVE FRAME RATE	814 (rat) 406 (pig)	530	1,000	400 (pig)	1,000
TOTAL COST, # CAMERAS	$18k, 1 camera	$70k, 2 cameras	$107k, 2 cameras	$1,600, 2 cameras	$2,400, 2 cameras

*note: novel configuration described in this manuscript

776 **Table 2**: Optical mapping experiments employing a single sensor design.

REFERENCE	Jaimes et al. (2019)*	Lee et al. (2011)	Bachtel et al. (2011)	Yamanaka et al. (2012)	Scull et al. (2012)
DYES USED	Rhod2 and RH237	Fura-2 and Di-4-ANBDQPQ	Di-4-ANEPPS	Rhod2 and RH237	X-Rhod-1 and Di-4-ANEPPS
BIOLOGICAL SAMPLE	Isolated heart	Isolated heart	Isolated heart	Isolated heart	Cell monolayer
NO. OF PIXELS	256 x 384 (rat) 640 x 512 (pig)	64 x 64	128 x 64	512 x 512	504 bundled optical fibers and PDA
EFFECTIVE FRAME RATE PER CHANNEL	814 (rat) 406 (pig)	465 (variable)	375	125	488.3
SIGNAL-TO-NOISE	~12	50-200	5.8	NA	21
NECESSARY HARDWARE	Single, constant wavelength light-source. Path splitter required	Four wavelength LEDs with custom-built microcontroller	Blue and cyan LEDs mounted on PCBs, custom hardware and software for rapid switching	LED ring light consisting of blue and yellow LEDs	Dual wavelength LED matrix, custom passband filter, A/D triggering board, custom LabVIEW program
KEY DIFFERENCES	Simultaneous imaging of Vm and Ca, off-the-shelf components	Interleaved, Ratiometric Vm and ratiometric Ca	Sequential imaging, Ratiometric Vm measurements only	Sequential imaging, Emission filter wheel, allows for more possible dye combinations	Sequential imaging, Crosstalk reduction algorithm

*note: novel configuration described in this manuscript; calculated signal-to-noise ratio measurements are listed in Supplemental Table 2

A

sCMOS

Image splitter

530 nm LED

Lens

530 nm ± 20 nm

Heart

B

Vm   Ca

530 nm LED

530 nm ± 20 nm

Spectrally mixed emitted light

C

710+ nm

660+ nm

585 nm ± 20

D

CUSTOM MATLAB SCRIPT

ROI spatial averaging → Signal Processing
• low-pass filter
• drift removal
• normalization
→ Vm & Ca Trace Analysis

Acquire Image Stack

RHYTHM

Image Processing
• background subtraction
• box blur
• normalization
→ Generate isochrone → ORCA
Conduction velocity analysis

A

Rat, Vm

Rat, Ca

PCL: 150 ms

PCL: 200 ms

PCL: 220 ms

Normalized Vm & Ca
Fluorescence @ ROI

Time (ms)

B

Pig, Vm

Pig, Ca

PCL: 180 ms

Vm
Ca

PCL: 200 ms

PCL: 220 ms

Normalized Vm & Ca
Fluorescence @ ROI

Time (ms)

A **Vm, 5x5 Box Blur**

C **Vm, 9x9 Box Blur**

E **Vm, 15x15 Box Blur**

G

B **Ca, 5x5 Box Blur**

D **Ca, 9x9 Box Blur**

F **Ca, 15x15 Box Blur**

H

Activation Time (ms)
5 10 15 20 25 30 35

I   Raw   5x5   15x15

A Vm

Base

Apex

B Ca

Activation Time (ms)

1 cm

C

- – – Vm Apex
- – – Ca Apex
- — Vm Base
- — Ca Base

Time (msec)

D

Conduction Velocity (cm/sec)

Vm T   Vm L   Ca T   Ca L

E Vm

F Ca

t=0 ms   t=10 ms   t=20 ms   t=30 ms   t=40 ms

A

B APD30 Odd Beat

R1

Apex   Base

R2

C APD30 Even Beat

D CaD30 Odd Beat

E CaD30 Even Beat

Duration (ms)

20 25 30 35 40 45 50 55

F R1 (In-Phase)

Vm   O   E

50 ms

Ca   O   E

APD70 (ms)

0   5   10   15
Beat #

CaT Amp. (F1/F0)

13.0
12.5
12.0
11.5
11.0
10.5
10.0
9.5
9.0

0   5   10   15
Beat #

G R2 (Out-of-Phase)

Vm   O   E

50 ms

Ca   O   E

APD70 (ms)

0   5   10   15
Beat #

CaT Amp. (F1/F0)

2.5
2.4
2.3
2.2
2.1
2.0
1.9
1.8
1.7
1.6

0   5   10   15
Beat #

25 msec