1 **Generating high-quality libraries for DIA-MS with empirically-corrected peptide**
2 **predictions.**
3

4 Brian C. Searle,[1,2,*] Kristian E. Swearingen,[1] Christopher A. Barnes,[3] Tobias Schmidt,[4] Siegfried
5 Gessulat,[4,5] Bernhard Kuster,[4,6] and Mathias Wilhelm[4]
6
[1] Institute for Systems Biology, Seattle, WA, USA
[2] Proteome Software, Inc. Portland, OR, USA
[3] Novo Nordisk Research Center Seattle, Inc. Seattle, WA, USA
[4] Chair of Proteomics and Bioanalytics, Technical University of Munich, Freising, Germany
[5] SAP SE, Potsdam, Germany
[6] Bavarian Center for Biomolecular Mass Spectrometry, Freising, Germany
13
14 *Corresponding author, email: bsearle@systemsbiology.org
15

ABSTRACT

Data-independent acquisition approaches typically rely on sample-specific spectrum

libraries requiring offline fractionation and tens to hundreds of injections. We demonstrate a new

library generation workflow that leverages fragmentation and retention time prediction to build

libraries containing every peptide in a proteome, and then refines those libraries with empirical

data. Our method specifically enables rapid library generation for non-model organisms, which

we demonstrate using the malaria parasite *Plasmodium falciparum*, and non-canonical

databases, which we show by detecting missense variants in HeLa.

INTRODUCTION

Data-independent acquisition (DIA) mass spectrometry (MS) is a powerful label-free

technique for deep, proteome-wide profiling.[1,2] With DIA, mass spectrometers are tuned to

systematically acquire tandem mass spectra at regular retention time and m/z intervals, making

the method free of intensity-triggering biases introduced by data-dependent acquisition (DDA).

To accomplish this, precursor isolation windows are widened such that multiple peptides are

usually co-fragmented in the same MS/MS scan. DIA methods generally identify peptides with

library search engines[3–5] using sample-specific spectrum libraries[6] from DDA experiments. In

peptide-centric searching,[7] library entries are scored according to retention time, such that the

best-scoring time point for each peptide is reported. Only peptides present in the libraries can

be detected, and the peptide detection reports must be corrected to limit the number of potential

false discoveries.[8] Most importantly, these libraries are built at the expense of time, sample, and

considerable effort with offline fractionation, especially considering that they are typically not

reusable across laboratories or instrument platforms.[9]

When sample-specific spectrum library generation is either impossible or impractical, as

is frequently the case with non-model organisms, sequence variants, splice isoforms, or scarce

sample quantities, software tools such as Pecan[10] and DIA-Umpire[11] can detect peptides from

DIA experiments without a spectrum library by directly searching every peptide in FASTA

databases. Gas-phase fractionation[12] (GPF) improves detection rates with these tools[10] by

injecting the same sample multiple times with tiled precursor isolation windows, allowing each

injection to have narrower windows (and thus fewer co-fragmented peptides) with the same

instrument duty cycle. While this method is often prohibitively expensive because it requires

enough instrument time and protein content for multiple injections for each sample, the use of

multiple GPF injections can be applied just to pooled samples to generate DIA-only

chromatogram libraries that make it easier to detect peptides in single-injection DIA

experiments.[13] However, even when using GPF these tools still generally detect fewer peptides

than library search engines, which can leverage previously acquired instrument-specific

fragmentation and measured retention times. Recently it has become possible to accurately

predict spectra from peptide sequences,[14,15] but direct searching of single-injection DIA data has

remained problematic, in part due to the false discovery rate (FDR) correction required when

considering all possible tryptic peptides in a FASTA database. Proteins show 3-4 orders of

magnitude difference in intensity between the best- and worst-responding tryptic peptides,[16] and

only considering the best-responding peptides in libraries can improve detection rates by

lessening the required FDR correction.

RESULTS

Generating empirically-corrected libraries from peptide predictions. Here we report on a

DIA-only workflow that produces higher-quality libraries than those generated by DDA while

simultaneously supplanting the need for any offline fractionation. Our workflow (Fig. 1a, Online

Methods) uses a recently developed deep neural network, Prosit,[14] to predict a library of

fragmentation patterns and retention times for every peptide in FASTA databases.

Fragmentation prediction in Prosit adjusts based on normalized collision energy (NCE), and we

tune the NCE parameter for each peptide charge state to account for DIA-specific

fragmentation. Building on the chromatogram library method,[13] we collect 6 GPF-DIA runs of a

sample pool using 90-minute gradients (12 total hours of MS acquisition). We then search these

GPF-DIA runs against the predicted library using EncyclopeDIA to detect peptides in the pool.

Assuming that virtually every consistently quantifiable peptide in an experiment is detectable

from the pool using GPF-DIA, we filter the predicted library to remove peptides and charge

states that cannot be found in the pool. We also replace the retention times and fragmentation

patterns in the predicted library with sample- and instrument-specific values found in the 6 runs.

The resulting empirically-corrected library typically contains only 10-15% of the entries in the

predicted library, requiring less stringent FDR correction when later searching single-injection

DIA runs.

Validating the empirically-corrected library methodology. Although searching single-injection DIA runs directly with predicted libraries is highly dependent on prediction accuracy, we found that our workflow produced high-quality libraries even if the predicted libraries were not precisely tuned for the instrument, suggesting broad cross-platform applicability. We demonstrated this by modulating the NCE setting in Prosit (Fig. 1b) and found that with a wide range of NCE settings, searching GPF-DIA spectra of yeast acquired on a Fusion Lumos MS produced between 33% and 60% larger empirically-corrected libraries than a sample-specific 10-fraction DDA library containing 40,025 unique peptides (Fig. 1c). Additionally, searches were less sensitive to prediction accuracy after empirical correction (Fig. 1d), in part due to consistently improved fragmentation patterns (Supp. Fig. 1).

While offline fractionation requires additional liquid chromatography (LC) using orthogonal separation modes to online LC-MS, GPF occurs completely within the mass spectrometer, making it both more reproducible and easier to perform. Also unlike offline fractionation, GPF also maintains the same LC matrix complexity as each experimental sample, ensuring more precise retention time consistency. This process produced libraries with better retention-time accuracy (80% of peptides within 35 seconds) than both the predicted (80% within 5.4 minutes) and DDA libraries (80% within 4.6 minutes), even when the DDA libraries were acquired on the same instrument (Supp. Fig. 2). Coupled with smaller library size and improved fragmentation patterns, these 3 factors had roughly equal and orthogonal improvements over directly searching single-injection DIA with predicted libraries (Supp. Fig. 3). Overall, searching single-injection DIA runs with empirically corrected libraries detected 66% more yeast peptides than searching the 10-fraction DDA library, and 37% more peptides than a 39-injection DDA library in reanalyzed HeLa datasets[13] from a Q-Exactive HF (QE-HF) MS (Fig. 2a).

Finding missense sequence variants with non-canonical libraries. Most publicly available spectrum libraries[17–19] are built by searching millions of spectra against a canonical genome. Our approach facilitates the analysis of sequence variants and splice junctions by simplifying exome-specific library construction, and we demonstrate this by analyzing HeLa-specific missense variants. RNA-seq expression across different HeLa strains[20] suggests that 12.4k genes containing 127 missense variants determined by the COSMIC[21] cell line project are typically expressed (>=10 counts, >=0.1 RPKM). Recently a custom database including COSMIC variants was used to reanalyze 40 public HeLa datasets containing 1,206 total DDA runs,[22] globally detecting 21 missense variants. In contrast, from 6 GPF-DIA runs we built an empirically-corrected HeLa-specific library with 7,484 protein groups including 37 missense variants (Supp. Data 1), approximately 30% of the expected expressed variants reported by COSMIC. With the library we detected an average of 24 missense variants from single-injection DIA runs (Fig. 2b). While in general the additionally detected variant-containing peptides will not greatly affect the overall number of quantitative measurements, quantifying key peptides in specific genes, such as the peptide antigen binding region of HLA[23] (Fig. 2c-e), can have a profound effect on biological interpretation. While improved fragmentation and retention time accuracy in empirically-corrected libraries can buffer search engines from over-reporting missense variants, care must be taken when attempting to distinguish between homologous sequences with DIA. Heterozygous peptides with similar retention times, such as those associated with G623S from the kinase EEF2K (Supp. Fig 4a), can share fragment ions (Supp. Fig 4b) and must be localized as if they contained post-translational modifications[24,25] should they fall in the same precursor isolation window.

Detecting and quantifying peptides from non-model organisms. We further applied our workflow to analyze cultures of *Plasmodium falciparum*, the parasite responsible for 50% of all malaria cases. We performed DDA and DIA on late-stage (stage IV/V) gametocytes, the form of

the parasite that is transmitted from an infected human to the mosquito vector that spreads the

disease. Using a QE-HF MS, we produced a 3,415-protein empirically corrected library,

increasing the previously known proteome[26] size by 58% (Fig. 3a) while only missing 2.5% of

proteins detected with other methods. Using this library we measured 2,740 proteins in single-

injection DIA experiments of *P. falciparum* peptides. Because *Plasmodium* is an obligate

parasite, even samples produced by *in vitro* culture are frequently contaminated with the host

proteome. To study the robustness of our method we diluted *P. falciparum* peptides into

peptides derived from uninfected red blood cells, and found we could detect >80% of the

parasite peptides in up to 1:99 dilution (Fig. 3b, Supp. Fig. 5a). We found that with our method,

DIA maintained higher quantification precision and depth than comparable DDA experiments at

each dilution step (Fig. 3c, Supp. Fig. 5b), which were analyzed with MaxQuant[27] with "match

between runs" enabled. Leveraging the highly accurate retention times in our empirically-

corrected library, we found that we could in part recover contaminated *P. falciparum* samples

(Supp. Fig. 6), even when MaxQuant/DDA reported the majority of peptides as missing values.

We analyzed 2,444 *P. falciparum* protein groups with at least 2 peptides (Supp. Data 2)

that were detected in every DIA run of 3 culture replicates with 3 technical replicates each (9

runs total). After cross-referencing against the PlasmoDB[28] compendium of 20 different studies

at various asexual, sexual, and mosquito stages, we observed 20 proteins previously

undetected by mass spectometry in any life stage of *P. falciparum*. We found an additional 396

proteins that had never been observed in late-stage gametocytes, including 55 previously

detected only in immature (stage I/II) gametocytes. Experiments such as these have the

potential to redefine the protein expression signature for each stage of the *Plasmodium* life

cycle, information that may be vital to identifying targets to disrupt parasite maturation. For

example, among those proteins previously undetected in gametocytes is the calcium-dependent

protein kinase CDPK2. *P. falciparum* parasites lacking CDPK2 develop normally through

asexual stages, but male gametocytes are incapable of undergoing exflagellation to become

gametes, thereby preventing transmission to the mosquito vector.[29] Our work validates that

CDPK2 is indeed present at measurable levels in gametocytes, paving the way to monitor

dynamic expression of this kinase over the course of parasite maturation.

DISCUSSION

In conclusion, empirical correction of predicted spectrum libraries enables rapid library

generation for non-canonical proteomes or non-model organisms without offline fractionation.

To encourage the reuse of this method, we have released a growing repository of pre-generated

predicted libraries compatible with EncyclopeDIA, Skyline, and Scaffold DIA, which are available

for download at ProteomicsDB (https://www.proteomicsdb.org/prosit/libraries). We also

developed a graphical user interface in EncyclopeDIA (Supplementary Note) to facilitate making

empirically-corrected libraries for new proteomes from any FASTA database.

Data availability statement. Prosit (https://www.proteomicsdb.org/prosit) and EncyclopeDIA

1.0 (https://bitbucket.org/searleb/encyclopedia) are both available under the Apache 2 open-

source license. The raw data from the yeast and *P. falciparum* studies are available at MassIVE

(MSV000084000). The raw data from the HeLa reanalysis are available as originally published

at MassIVE (MSV000082805).

Acknowledgements. We thank S. Kappe for insightful discussions, N. Carmago for providing

the gametocyte samples, L. Pino for providing yeast lysates, B. Kim for technical assistance.We

also thank M. MacCoss, K. Grove, and M. Guldbrandt for providing instrument time. B.C.S. is

supported by the Translational Research Fellows Program (TRFP) from the Institute for

Systems Biology. K.E.S. is supported by K25 AI119229. This work was in part funded by the

German Federal Ministry of Education and Research (BMBF, grant no. 031L0008A and no.

031L0168) and the EU Horizon 2020 grant EPIC-XS (grant no. 823839).

Author contributions. B.C.S. conceived the study. B.C.S., K.E.S., and C.A.B. performed the

experiments. B.C.S., T.S., and S.G. developed the software. B.C.S., B.K., and M.W. supervised

the work. All authors wrote and approved the manuscript.

Competing financial interests. B.C.S. is a founder and shareholder in Proteome Software,

which operates in the field of proteomics. M.W. and B.K. are founders and shareholders of

OmicScouts, but have no operational role in the company.

Figure 1: Evaluating empirically-corrected libraries made with peptide predictions. (**a**)

Workflow for generating empirically-corrected libraries from Prosit predictions of peptides in

FASTA databases. GPF-DIA runs of a sample pool are searched with a spectrum library of

fragmentation and retention time predictions. Peptides detected in these runs are compiled with

fragmentation patterns and retention times extracted from the GPF-DIA data into an empirically-

corrected library. (**b**) Violin plots showing spectral correlation between predicted library for yeast

peptides and single-injection DIA at various NCE settings (circles indicate medians). (**c**) The

total numbers of empirically-corrected library entries detected from GPF-DIA at various NCE

settings. (**d**) The fraction of peptides detected in single-injection DIA relative to the optimal NCE

for empirically-corrected libraries (blue line) are more consistent than predicted libraries (red

dashed line) across a wide range of NCE settings.

Figure 2: Empirically-corrected libraries improve peptide and missense variant detection

rates. (a) The average number of yeast (N=4) and HeLa (N=3) peptides in single-injection DIA

detected using library-free searching with Pecan and searching either a fractionated DDA

spectrum library, a predicted spectrum library, or an empirically-corrected library. (b) The

number of HeLa missense variants detected in the total DIA library, single-injection DIA

(triplicate experiments), or a meta-analysis of 40 published DDA experiments,[22] each filtered to

a 1% peptide/protein FDR. (c) Retention time predictions differ somewhat from empirical data

from GPF-DIA for homologous peptides: YFYTSMSRPGR from HLA-A (F33Y,V36M),

YFYTAMSRPGR from HLA-B, and YFYTAVSRPGR from HLA-C. (d) Relative +1H y-type and

b-type fragmentation patterns above 200 m/z for the same peptides are shown as butterfly plots

with empirical intensities (blue) above predicted intensities (red). (e) All 3 peptides are in the

1   HLA peptide binding/presentation region, as indicated by YFYTSMSRPGR (red) in HLA-A (blue)

2   relative to an antigenic peptide (green).

Figure 3: Rapid library generation and quantitation of *Plasmodium falciparum* proteins.

(a) A proportionally-sized Venn diagram showing the overlap between proteins detected in the

empirically-corrected library (blue), the full single-injection dilution curve with DDA (red), and the

Stage IV/V mixed male/female sexual *P. falciparum* proteome published by Lasonder *et al.*[26]

from 20x GeLCMS-fractionated DDA runs (yellow). Quantitative ratios of (**b**) 2,740 DIA- and (**c**)

1,491 DDA-detected malaria parasite *P. falciparum* proteins (gray lines) after dilution with

uninfected RBC lysate relative to the undiluted sample (N=1). Dots indicate the lowest dilution

level of measurement for each protein, and the blue dashed line indicates the ideal

measurement. Empirically-corrected libraries enabled DIA quantification of 2,152 *P. falciparum*

proteins in the last dilution sample (1:99), even when no proteins could be detected or quantified

using DDA.

METHODS

***Saccharomyces cerevisiae* culture and sample preparation.** As described in Searle *et al.*,[13]

S. cerevisiae strain BY4741 (Dharmacon) was cultured at 30°C in YEPD and harvested at mid-

log phase. Cells were pelleted and lysed in a buffer of 8 M urea, 50 mM Tris (pH 8), 75 mM

NaCl, 1 mM EDTA (pH 8) followed by 7 cycles of 4 minutes bead beating with glass beads.

After a 1 minute rest on ice, lysate was collected by piercing the tube and centrifuging for one

minute at 3,000 x g and 4°C into an empty eppendorf. After further centrifugation at 21,000 x g

and 4°C for 15 minutes, the protein content of the supernatant was removed and estimated

using BCA. Proteins were then reduced with 5 mM dithiothreitol for 30 minutes at 55°C,

alkylated with 10 mM iodoacetamide in the dark for 30 minutes at room temperature, and diluted

to 1.8 M urea, before digestion with sequencing grade trypsin (Pierce) at a 1:50 enzyme to

substrate ratio for 16 hours at 37°C. 5N HCl was added to approximately pH 2 to quench the

digestion, and the resulting peptides were desalted with 30 mg MCX cartridges (Waters).

Peptides were dried with vacuum centrifugation and brought to 1 µg / 3 µl in 0.1% formic acid

(buffer A) prior to mass spectrometry acquisition.

***Plasmodium falciparum* culture and red blood cell sample preparation.** Three flasks of

stage IV/V *P. falciparum* NF54 gametocytes were prepared. Asexual cultures were

synchronized with sorbitol and set up at 5% hematocrit and 1% parasitemia.

Gametocytogenesis was induced by withholding fresh blood and allowing parasitemia to

increase. N-acetyl glucosamine was added to media for 4 days beginning 7 days after set-up in

order to remove asexual parasites. Gametocyte-infected erythrocytes (giRBC) were enriched

from uninfected erythrocytes (uiRBC) by magnetic-activated cell sorting at stage III. Stage IV/V

gametocytes were collected on day 15 post-set-up. Additional uiRBC were also prepared by

washing multiple times in RPMI and stored at 50% hematocrit. giRBC and uiRBC cells were

lysed in a buffer of 10% sodium dodecyl sulfate (SDS), 100 mM ammonium bicarbonate (ABC),

cOmplete EDTA-free Protease Inhibitor Cocktail (Sigma), and Halt Phosphatase Inhibitor

Cocktail (Thermo Scientific). Proteins were then reduced with 20-40 mM tris(2-

carboxyethyl)phosphine (TCEP) for 10 minutes at 95°C and alkylated with 40-80 mM

iodoacetamide in the dark for 20 minutes at room temperature. After centrifugation at 16,000 x g

to pellet insoluble material, proteins were purified with methanol:chloroform extraction[30] and

dried and resuspended in 8M urea buffer before the content was estimated using BCA. After

dilution to 1.8 M urea, proteins were digested with sequencing-grade trypsin (Promega) at a

1:40 enzyme-to-substrate ratio for 15 hours at 37°C. The resulting peptides were desalted with

Sep-Pak cartridges (Waters), dried with vacuum centrifugation, and brought to 1 μg / 3 μl in

0.1% formic acid (buffer A) prior to mass spectrometry acquisition. In addition, several digested

peptide mixtures were made by diluting peptides from one flask of giRBC cells with peptides

from uiRBC cells at ratios of 1:0, 2:1, 7:8, 4:15, 1:9, 2:41, 2:91, and 1:99 giRBC:uiRBC.

Liquid chromatography mass spectrometry (*S. cerevisiae*). Tryptic *S. cerevisiae* peptides

were separated with a Thermo Easy nLC 1200 on self-packed 30 cm columns packed with 1.8

μm ReproSil-Pur C18 silica beads (Dr. Maisch) inside of a 75 μm inner diameter fused silica

capillary (#PF360 Self-Pack PicoFrit, New Objective). The 30 cm column was coiled inside of a

Sonation PRSO-V1 column oven set to 35°C prior to ionization into the MS. The HPLC was

performed using 200 nl/min flow with solvent A as 0.1% formic acid in water and solvent B as

0.1% formic acid in 80% acetonitrile. For each injection, 3 μl (approximately 1 μg) was loaded

and eluted with a linear gradient from 7% to 38% buffer B over 90 minutes. Following the linear

separation, the system was ramped up to 75% buffer B over 5 minutes and finally set to 100%

buffer B for 15 minutes, which was followed by re-equilibration to 2% buffer B prior to the

subsequent injection. Data were acquired using data-independent acquisition (DIA).

The Thermo Fusion Lumos was set to acquire 6 GPF-DIA acquisitions of a biological sample

pool using 120,000 precursor resolution and 30,000 fragment resolution. The automatic gain

control (AGC) target was set to 4e5, the maximum ion inject time (IIT) was set to 60 ms, the

normalized collision energy (NCE) was set to 33, and +2H was assumed as the default charge

state. The GPF-DIA acquisitions used 4 m/z precursor isolation windows in a staggered window

pattern with optimized window placements (i.e., 398.4 to 502.5 m/z, 498.5 to 602.5 m/z, 598.5 to

702.6 m/z, 698.6 to 802.6 m/z, 798.6 to 902.7 m/z, and 898.7 to 1002.7 m/z). Individual samples

for proteome profiling acquisitions used single-injection DIA acquisitions (120,000 precursor

resolution, 15,000 fragment resolution, AGC target of 4e5, max IIT of 20 ms) using 8 m/z

precursor isolation windows in a staggered window pattern with optimized window placements

from 396.4 to 1004.7 m/z.

For generation of an *S. cerevisiae* spectral library, 80 μg of the same tryptic digests described

above were separated into 10 total fractions using the Pierce high-pH reversed-phase peptide

fractionation Kit (Thermo, #84868). Briefly, peptides were loaded onto hydrophobic resin spin

column and eluted using the following 8 acetonitrile steps: 5%, 7.5%, 10%, 12.5%, 15%, 17.5%,

20.0%, and 50%, keeping both the wash and flow-through. The resulting peptide fractions were

injected into the same Thermo Fusion Lumos using the same chromatography setup and

column described above, but configured for data-dependent acquisition (DDA). After adjusting

each fraction to an estimated 0.5 to 1.0 μg on column, the fractions were measured in a top-20

configuration with 30 second dynamic exclusion. Precursor spectra were collected from 300-

1650 m/z at 120,000 resolution (AGC target of 4e5, max IIT of 50 ms). MS/MS were collected

on +2H to +5H precursors achieving a minimum AGC of 2e3. MS/MS scans were collected at

30,000 resolution (AGC target of 1e5, max IIT of 50 ms) with an isolation width of 1.4 m/z with a

NCE of 33.

Liquid chromatography mass spectrometry (*P. falciparum*). Tryptic *P. falciparum* and RBC

peptides were separated with a Thermo Easy nLC 1000 and emitted into a Thermo Q-Exactive

HF. In-house laser-pulled tip columns were created from 75 μm inner diameter fused silica

capillary and packed with 3 µm ReproSil-Pur C18 beads (Dr. Maisch) to 30 cm. Trap columns

were created from Kasil fritted 150 µm inner diameter fused silica capillary and packed with the

same C18 beads to 2 cm. The HPLC was performed using 250 nl/min flow with solvent A as

0.1% formic acid in water and solvent B as 0.1% formic acid in 80% acetonitrile. For each

injection, 3 µl (approximately 1 µg) was loaded and eluted using a 84-minute gradient from 6 to

40% buffer B, followed by steep 5-minute gradient from 40 to 75% buffer B and finally set to

100% buffer B for 15 minutes, which was followed by re-equilibration to 0% buffer B prior to the

subsequent injection. Data were acquired using either DDA or DIA.

The Thermo Q-Exactive HF was set to acquire DDA in a top-20 configuration with "auto"

dynamic exclusion. Precursor spectra were collected from 400-1600 m/z at 60,000 resolution

(AGC target of 3e6, max IIT of 50 ms). MS/MS were collected on +2H to +5H precursors

achieving a minimum AGC of 1e4. MS/MS scans were collected at 15,000 resolution (AGC

target of 1e5, max IIT of 25 ms) with an isolation width of 1.4 m/z with a NCE of 27. Additionally,

6 GPF-DIA acquisitions were acquired of a biological sample pool (60,000 precursor resolution,

30,000 fragment resolution, AGC target of 1e6, max IIT of 60 ms, NCE of 27, +3H assumed

charge state) using 4 m/z precursor isolation windows in a staggered window pattern with

optimized window placements (i.e., 398.4 to 502.5 m/z, 498.5 to 602.5 m/z, 598.5 to 702.6 m/z,

698.6 to 802.6 m/z, 798.6 to 902.7 m/z, and 898.7 to 1002.7 m/z). Individual samples used

single-injection DIA acquisitions (60,000 precursor resolution, 30,000 fragment resolution, AGC

target of 1e6, max IIT of 60 ms) using 16 m/z precursor isolation windows in a staggered

window pattern with optimized window placements from 392.4 to 1008.7 m/z.

FASTA databases and predicted spectrum libraries. Species-specific reviewed FASTA

databases for *Homo sapiens* (April 25 2019, 20415 entries) and *Saccharomyces cerevisiae*

(January 25 2019, 6729 entries) were downloaded from Uniprot. The *Plasmodium falciparum*

FASTA database[31] was downloaded from PlasmoDB[28] version 43 (April 24 2019, 5548 entries).

The Ensembl-based HeLa-specific FASTA database[22] was downloaded from the ACS

Publications website and modified to be compatible with EncyclopeDIA (47305 entries, including

both canonical and variant protein sequences). Each database was digested *in silico* to create

all possible +2H and +3H peptides with precursor m/z within 396.43 and 1002.70, assuming up

to one missed cleavage. Peptides were further limited to be between 7 and 30 amino acids to

match the restrictions of the Prosit tool.[14] In general, NCE were assumed to be 33 (yeast was

processed using NCE from 15 to 42 in 3 NCE increments) but modified to account for charge

state. Since DIA assumes all peptides are of a fixed charge, we adjusted the NCE setting as if

peptides were fragmented at the wrong charge state using the formula:

$Adjusted\ NCE = NCE * \frac{factor(default\ charge)}{factor(peptide\ charge)}$

where the factors were 1.0 for +1H, 0.9 for +2H, 0.85 for +3H, 0.8 for +4H, and 0.75 for +5H and

above.[32] After submitting to Prosit, predicted MS/MS and retention times were converted to the

EncyclopeDIA DLIB format for further processing. Scripts to produce Prosit input from FASTAs

and build EncyclopeDIA-compatible spectrum libraries from Prosit output are available as

functions in EncyclopeDIA 1.0.

DDA data processing. All Thermo RAW files were converted to .mzML format using the

ProteoWizard package[33] (version 3.0.18299) using vendor peak picking. DDA data was

searched with Comet[34] (version 2017.01 rev. 1), allowing for fixed cysteine

carbamidomethylation, variable peptide n-terminal pyro-glu, and variable protein n-terminal

acetylation. Fully tryptic searches were performed with a 50 ppm precursor tolerance and a 0.02

Da fragment tolerance permitting up to 2 missed cleavages. High-pH reverse-phase fractions

were combined and search results were filtered to a 1% peptide-level FDR using

PeptideProphet[35] from the Trans-Proteomic Pipeline[36] (TPP version 5.1.0). A yeast-specific

Bibliospec[37] DDA spectrum library was created from Thermo Q-Exactive DDA data using Skyline[38] (Daily version 19.0.9.149).

P. falciparum and RBC DDA data was additionally processed with MaxQuant[27] (version 1.6.5.0) to perform label-free quantitation with precursor ion integration. MaxQuant was configured to use default parameters, briefly fixed cysteine carbamidomethylation, variable methionine oxidation, and variable protein n-terminal acetylation. Fully tryptic searches were performed with a 20 ppm fragment tolerance using both the human and *P. falciparum* FASTA databases, as well as common contaminants and filtered to a 1% peptide-level FDR. Quantification was performed using unique and razor peptides with the match between runs setting turned on.

DIA data processing. DIA data was overlap demultiplexed[39] with 10 ppm accuracy after peak picking in ProteoWizard (version 3.0.18299). Searches were performed using EncyclopeDIA (version 0.8.3), which was configured to use default settings: 10 ppm precursor, fragment, and library tolerances. EncyclopeDIA was allowed to consider both B and Y ions and trypsin digestion was assumed. EncyclopeDIA search results were quantified and filtered to a 1% peptide-level using Percolator[40] (version 3.1) and 1% protein-level FDR assuming parsimony.

Empirically-corrected library generation. Predicted libraries were corrected with EncyclopeDIA using the chromatogram library method described previously.[13] Briefly, GPF-DIA samples for a given study were loaded into EncyclopeDIA using the above parameters, where the search library was set to the appropriate predicted spectrum library. Peptides detected by EncyclopeDIA were exported as a chromatogram library.

A peptide entry in a chromatogram library is similar to a peptide entry in a spectrum library in that it contains a precursor mass, retention time, and a fragmentation spectrum. In addition, a chromatogram library entry also contains the peptide peak shape, and a correlation score for every fragment ion indicating the agreement between the overall peptide peak shape and the

fragment peak shape. This score provides an indication of the likelihood the fragment ion was

interfered with in the GPF-DIA, with the expectation that it will also likely be interfered with in

single-injection DIA as well. Missense variants in the HeLa empirically-corrected library were

manually validated.

This process created a new empirically-corrected library containing only peptides found in the

GPF-DIA samples, and also retained empirical fragment ion intensities and retention times

observed from the DIA data. These libraries were made to be compatible with both

EncyclopeDIA and Skyline and were used for downstream analysis of single-injection DIA.

REFERENCES

1. Gillet, L. C. *et al.* Targeted data extraction of the MS/MS spectra generated by data-
independent acquisition: a new concept for consistent and accurate proteome analysis.
Mol. Cell. Proteomics **11**, O111.016717 (2012).

2. Doerr, A. DIA mass spectrometry. *Nat. Methods* **12**, 35 (2014).

3. Weisbrod, C. R., Eng, J. K., Hoopmann, M. R., Baker, T. & Bruce, J. E. Accurate peptide
fragment mass analysis: multiplexed peptide identification and quantification. *J. Proteome*
Res. **11**, 1621–1632 (2012).

4. Röst, H. L. *et al.* OpenSWATH enables automated, targeted analysis of data-independent
acquisition MS data. *Nat. Biotechnol.* **32**, 219–223 (2014).

5. Wang, J. *et al.* MSPLIT-DIA: sensitive peptide identification for data-independent
acquisition. *Nat. Methods* **12**, 1106–1108 (2015).

6. Schubert, O. T. *et al.* Building high-quality assay libraries for targeted analysis of SWATH
MS data. *Nat. Protoc.* **10**, 426–441 (2015).

7. Ting, Y. S. *et al.* Peptide-Centric Proteome Analysis: An Alternative Strategy for the
Analysis of Tandem Mass Spectrometry Data. *Mol. Cell. Proteomics* **14**, 2301–2307 (2015).

8. Rosenberger, G. *et al.* Statistical control of peptide and protein error rates in large-scale
targeted data-independent acquisition analyses. *Nat. Methods* **14**, 921–927 (2017).

9. Bruderer, R. *et al.* Optimization of Experimental Parameters in Data-Independent Mass
Spectrometry Significantly Increases Depth and Reproducibility of Results. *Mol. Cell.*
Proteomics **16**, 2296–2309 (2017).

10. Ting, Y. S. *et al.* PECAN: library-free peptide detection for data-independent acquisition
tandem mass spectrometry data. *Nat. Methods* **14**, 903–908 (2017).

11. Tsou, C.-C. *et al.* DIA-Umpire: comprehensive computational framework for data-
independent acquisition proteomics. *Nat. Methods* **12**, 258 (2015).

12. Panchaud, A. *et al.* Precursor acquisition independent from ion count: how to dive deeper
into the proteomics ocean. *Anal. Chem.* **81**, 6481–6488 (2009).

13. Searle, B. C. *et al.* Chromatogram libraries improve peptide detection and quantification by
data independent acquisition mass spectrometry. *Nat. Commun.* **9**, 5128 (2018).

14. Gessulat, S. *et al.* Prosit: proteome-wide prediction of peptide tandem mass spectra by
deep learning. *Nat. Methods* **16**, 509–518 (2019).

15. Tiwary, S. *et al.* High-quality MS/MS spectrum prediction for data-dependent and data-
independent acquisition data analysis. *Nat. Methods* **16**, 519–525 (2019).

16. Searle, B. C., Egertson, J. D., Bollinger, J. G., Stergachis, A. B. & MacCoss, M. J. Using
Data Independent Acquisition (DIA) to Model High-responding Peptides for Targeted
Proteomics Experiments. *Mol. Cell. Proteomics* **14**, 2331–2340 (2015).

17. Rosenberger, G. *et al.* A repository of assays to quantify 10,000 human proteins by

SWATH-MS. *Sci Data* **1**, 140031 (2014).

18. Zolg, D. P. *et al.* Building ProteomeTools based on a complete synthetic human proteome. *Nat. Methods* **14**, 259–262 (2017).

19. Wang, M. *et al.* Assembling the Community-Scale Discoverable Human Proteome. *Cell Syst* **7**, 412–421.e5 (2018).

20. Liu, Y. *et al.* Multi-omic measurements of heterogeneity in HeLa cells across laboratories. *Nat. Biotechnol.* **37**, 314–322 (2019).

21. Tate, J. G. *et al.* COSMIC: the Catalogue Of Somatic Mutations In Cancer. *Nucleic Acids Res.* **47**, D941–D947 (2019).

22. Robin, T., Bairoch, A., Müller, M., Lisacek, F. & Lane, L. Large-Scale Reanalysis of Publicly Available HeLa Cell Proteomics Data in the Context of the Human Proteome Project. *J. Proteome Res.* **17**, 4160–4170 (2018).

23. Garboczi, D. N. *et al.* Structure of the complex between human T-cell receptor, viral peptide and HLA-A2. *Nature* **384**, 134–141 (1996).

24. Rosenberger, G. *et al.* Inference and quantification of peptidoforms in large sample cohorts by SWATH-MS. *Nat. Biotechnol.* **35**, 781–788 (2017).

25. Searle, B. C., Lawrence, R. T., MacCoss, M. J. & Villén, J. Thesaurus: quantifying phosphopeptide positional isomers. *Nat. Methods* **16**, 703–706 (2019).

26. Lasonder, E. *et al.* Integrated transcriptomic and proteomic analyses of P. falciparum gametocytes: molecular insight into sex-specific processes and translational repression. *Nucleic Acids Res.* **44**, 6087–6101 (2016).

27. Tyanova, S., Temu, T. & Cox, J. The MaxQuant computational platform for mass spectrometry-based shotgun proteomics. *Nat. Protoc.* **11**, 2301–2319 (2016).

28. Aurrecoechea, C. *et al.* PlasmoDB: a functional genomic database for malaria parasites. *Nucleic Acids Res.* **37**, D539–43 (2009).

29. Bansal, A., Molina-Cruz, A., Brzostowski, J., Mu, J. & Miller, L. H. Plasmodium falciparum Calcium-Dependent Protein Kinase 2 Is Critical for Male Gametocyte Exflagellation but Not Essential for Asexual Proliferation. *MBio* **8**, (2017).

30. Wessel, D. & Flügge, U. I. A method for the quantitative recovery of protein in dilute solution in the presence of detergents and lipids. *Anal. Biochem.* **138**, 141–143 (1984).

31. Gardner, M. J. *et al.* Genome sequence of the human malaria parasite Plasmodium falciparum. *Nature* **419**, 498–511 (2002).

32. Orsburn, B. Normalized collision energy calculation for Q Exactive. (2014). Available at: http://proteomicsnews.blogspot.com/2014/06/normalized-collision-energy-calculation.html. (Accessed: 13th June 2019)

33. Chambers, M. C. *et al.* A cross-platform toolkit for mass spectrometry and proteomics. *Nat. Biotechnol.* **30**, 918–920 (2012).

34. Eng, J. K., Jahan, T. A. & Hoopmann, M. R. Comet: an open-source MS/MS sequence database search tool. *Proteomics* **13**, 22–24 (2013).

35. Keller, A., Nesvizhskii, A. I., Kolker, E. & Aebersold, R. Empirical statistical model to estimate the accuracy of peptide identifications made by MS/MS and database search. *Anal. Chem.* **74**, 5383–5392 (2002).

36. Deutsch, E. W. *et al.* Trans-Proteomic Pipeline, a standardized data processing pipeline for large-scale reproducible proteomics informatics. *Proteomics Clin. Appl.* **9**, 745–754 (2015).

37. Frewen, B. E., Merrihew, G. E., Wu, C. C., Noble, W. S. & MacCoss, M. J. Analysis of peptide MS/MS spectra from large-scale proteomics experiments using spectrum libraries. *Anal. Chem.* **78**, 5678–5684 (2006).

38. MacLean, B. *et al.* Skyline: an open source document editor for creating and analyzing targeted proteomics experiments. *Bioinformatics* **26**, 966–968 (2010).

39. Amodei, D. *et al.* Improving Precursor Selectivity in Data-Independent Acquisition Using Overlapping Windows. *J. Am. Soc. Mass Spectrom.* (2019). doi:10.1007/s13361-018-2122-8

40. Käll, L., Canterbury, J. D., Weston, J., Noble, W. S. & MacCoss, M. J. Semi-supervised learning for peptide identification from shotgun proteomics datasets. *Nat. Methods* **4**, 923–925 (2007).