Matrix-Assisted Laser Desorption-Ionization Time-of-Flight mass spectrometry using a custom-made database, biomarker assignment or mathematical classifiers does not differentiate *Shigella spp.* and *Escherichia coli*

Authors: Maaike J.C. van den Beld[1,2](orcid:0000-0001-8720-8434), John W.A. Rossen[2](orcid: 0000-0002-7167-8623), Arie Evers[1], A.M.D. (Mirjam) Kooistra-Smid[2,3](orcid:0000-0002-8856-581X), Frans A.G. Reubsaet[1]

Affiliations:

[1]Infectious Disease Research, Diagnostics and laboratory Surveillance, Centre for Infectious Disease Control National Institute for Public Health and the Environment (RIVM), , Bilthoven, the Netherlands

[2]Department of Medical Microbiology and Infection Prevention, University of Groningen, University Medical Center Groningen, Groningen, the Netherlands

[3]Department of Medical Microbiology, Certe, Groningen, the Netherlands

Corresponding author: Maaike van den Beld

Email: maaike.van.den.beld@rivm.nl

Abstract

Purpose: *Shigella spp.* and *E. coli* are closely related, and cannot be distinguished using Matrix-Assisted Laser Desorption-Ionization Time-of-Flight mass spectrometry (MALDI-TOF MS) with commercially available databases. Here, three alternative approaches using MALDI-TOF for identification and distinction of *Shigella spp.*, *E. coli* and its pathotype EIEC were explored.

Methods: A custom-made database was developed, biomarkers were assigned and classification models using machine learning were designed and evaluated using spectra of 456 *Shigella spp.*, 42 *E. coli* and 61 EIEC isolates, obtained by the direct smear method and the ethanol-formic acid extraction method.

Results: Identification with a custom-made database resulted in >94% *Shigella* identified at the genus level, and >91% *S. sonnei* and *S. flexneri* at the species level, but distinction of *S. dysenteriae*, *S. boydii* and *E. coli* was poor. Moreover, 10-15% of duplicates rendered discrepant results. With biomarker assignment, 98% *S. sonnei* isolates were correctly identified, although the *S. sonnei* biomarkers were not specific as other species were also identified as *S. sonnei*. Discriminating markers for *S. dysenteriae*, *S. boydii*, and *E. coli* were not assigned at all. Classifiers identified *Shigella* in 96% of isolates correctly, but most *E. coli* isolates were also assigned to *Shigella*.

Conclusion: None of the proposed alternative approaches is suitable for use in clinical diagnostics for the identification of *Shigella spp.*, *E. coli* and EIEC because of their poor distinctive properties. We suggest the use of MALDI-TOF MS for identification of the *Shigella spp.*/*E. coli* complex, but other tests should be used for distinction.

Keywords: MALDI-TOF, *Shigella spp.*, *Escherichia coli*, custom-made database, biomarker assignment, machine-learning classifiers

Introduction

The *E. coli* pathotype entero-invasive *E. coli* (EIEC) is thought to cause the same disease as *Shigella spp.* [1]. This pathotype consists of isolates that possess some of the *E. coli* phenotypic characteristics, but has also the invasive nature of *Shigella spp.* [2,3]. EIEC harbors the same virulence markers as *Shigella spp.* that are used in molecular diagnostics for detection of both *Shigella spp.* and EIEC, but are not suitable to distinguish them [4]. Differentiation of *E. coli* from *Shigella spp.* is historically performed using phenotypical tests, serotyping and the determination of virulence markers using PCR [5,6].

Currently, most clinical laboratories use Matrix-Assisted Laser-Desorption Ionization Time-of-Flight Mass Spectrometry (MALDI-TOF MS) for the identification of bacteria in a routine diagnostic setting. Commercially available databases, as MALDI biotyper® in combination with the MALDI Security-Relevant (SR) library® (Bruker Daltonik GmbH, Bremen, Germany) and VITEK® MS (BioMérieux, Marcy-l'Etoile, France) are able to distinguish *Shigella spp.* and *E. coli* from other *Enterobacteriaceae*. However, they are not able to distinguish between the different *Shigella* species and *E. coli*, including the EIEC pathotype [7].

The development of custom-made databases for identification of bacteria using MALDI-TOF MS as alternative to commercially available databases, proved successful for multiple species before [8-10]. Most notably, an earlier study developed a custom-made database for identification and distinction of *Shigella spp.* and *E. coli* specifically, however EIEC isolates were not included in their database [11]. Using a database approach, comparisons of unknown isolates to isolates in a database comprise the whole spectra for species identification. However, for closer related groups a more subtle approach can be essential in which variations within the spectra are examined in the form of presence or absence of specific peaks as biomarkers [12,13]. The biomarker approach was used to type *E. coli* isolates before,

with varying success rates [13]. These approaches mainly targeted the pathotype entero- hemorrhagic

E. coli (EHEC) or the highly virulent ST131 clone [13], although two studies used biomarker typing

specifically for *Shigella spp.* and *E. coli*, without EIEC isolates [14,15]. One of those studies identified

biomarkers outside the mass range of 2,000-20,000 Da, which is used in routine applications [15], and

the other did not specify in which species the biomarkers were present or absent [14]. Besides the

determination of the presence or absence of single biomarkers, patterns of these biomarkers can be

investigated and recognized with machine-learning algorithms [16]. These machine-learning based

methods can be used to establish classifiers for identification of groups within species of bacteria

[17,18]. Moreover, these classifiers were developed for the identification of *Shigella spp*. and *E. coli*

before, although EIEC isolates were not included [14].

In this study, the ability of MALDI-TOF MS was assessed for the distinction of the four *Shigella* species,

EIEC and non-invasive *E. coli* using alternatives for the commercially available databases. First, a

custom-made database, including all species of *Shigella*, *E. coli*, including EIEC isolates, was

developed and evaluated. Second, biomarkers were assigned and evaluated, and third, classifier

models based on machine learning were defined, applied and evaluated.

Material and Methods

Bacterial isolates

A total of 559 isolates consisting of 36 *S. dysenteriae*, 156 *S. flexneri*, 32 *S. boydii*, 232 *S. sonnei*, 61

EIEC, and 42 other *E. coli* of human and animal origin comprising phylogroups A, B1, B2 and D [19]

was used (Table 1). All isolates, except the references, were identified using a previously described

culture-based identification algorithm [20]. They were divided into a set of training isolates (n=288) and

a set of test isolates (n=271), both having similar species and serotype distributions. The training set

was used to construct the custom-made database, assign biomarkers, and for defining and training of

machine-learning classifier models. The test set was used to test all of these algorithms in duplicate,

with both direct smear and ethanol-formic acid extraction application methods.

MALDI-TOF preparation of isolates

All isolates were grown overnight on Columbia Sheep Agar (CSA, Biotrading, Mijdrecht, the

Netherlands) at 37°C and were subsequently subjected to the direct smear method as well as to the

ethanol- formic acid extraction with silica beads as previously described [21]. Colonies or 1 µl extract

were applied onto a polished steel plate, air-dried and overlaid with 1 µl α-Cyano-4-hydroxycinnamic

acid in 50% acetonitrile-2.5% trifluoroacetic acid (HCCA matrix). The samples were analyzed using a

Bruker Microflex LT (Bruker Daltonik GmbH, Bremen, Germany) in a linear positive mode, with 30-40%

laser power and within a mass-range of 2,000-20,000 Da.

The isolates from the training set were used to produce Main Spectrum Profiles (MSPs) according to

manufacturer's instructions, using FlexControl V3.4 (Bruker Daltonik). The isolates from the training set

were analyzed using Maldi Biotyper 3.0 Real Time Classification (RTC, Bruker Daltonik).

Database development

The MSPs produced from 288 isolates in the training set were used to build a custom-made database

with Maldi Biotyper OC V3.1.66 (Bruker Daltonik). In addition, a dendrogram to assess the relatedness

of these MSPs was inferred using default settings. The isolates in the test set were identified using this

custom-made database. Additionally, the test isolates were also identified using the commercially

available Bruker MALDI Biotyper database (V8.0.0.0) and the Bruker Security-Relevant Library

(V1.0.0.0), and using a combination of the commercial and custom-made databases. Quality of the

results was indicated by a log-score, calculated by Maldi Biotyper 3.0 RTC: a log-score of 2.000-2.300

corresponds to "secure genus identification, probable species identification" and a log-score of >2.300

corresponds to a "highly probable species identification". Both duplicate spots were analyzed, the

Table 1 Isolates used in this study, divided in training set and test set

Species and serotype/O-type	Training set		Test set	
	n	**Origin**	**n**	**Origin**
S. dysenteriae serotype 1	2	CIP 57.28[T]; A1	1	1 ci[a]
S. dysenteriae serotype 2	5	A2, 4 ci[a]	4	4 ci[a]
S. dysenteriae serotype 3	5	AMC-43-G-93; 4 ci[a]	3	3 ci[a]
S. dysenteriae serotype 4	2	AMC 43-G-86; 1 ci[a]	0	
S. dysenteriae serotype 5	1	AMC 43-G-84	0	
S. dysenteriae serotype 6	1	AMC 43-G-81	1	1 ci[a]
S. dysenteriae serotype 7	1	AMC 43-G-76	1	1 ci[a]
S. dysenteriae serotype 9	2	A58: 1646; 1 ci[a]	1	1 ci[a]
S. dysenteriae serotype 10	1	A2050-52	0	
S. dysenteriae serotype 12	2	2 ci[a]	1	1 ci[a]
S. dysenteriae serotype 14	1	NCTC 11867	0	
S. dysenteriae serotype 15	1	NCTC 11868	0	
Total number of *S. dysenteriae*	**24**		**12**	
S. flexneri serotype 1a	3	B1A; 2 ci[a]	0	
S. flexneri serotype 1b	5	B1B; 4 ci[a]	5	5 ci[a]
S. flexneri serotype 1c	4	4 ci[a]	3	3 ci[a]
S. flexneri serotype 2a	32	CIP 82.48[T]; B2A; 30 ci[a]	32	32 ci[a]
S. flexneri serotype 2b	1	B2B	3	3 ci[a]
S. flexneri serotype 3a	2	B3A; 1 ci[a]	14	14 ci[a]
S. flexneri serotype 3b	2	B3B; B3C	3	3 ci[a]
S. flexneri serotype 4a	1	B4A	4	4 ci[a]
S. flexneri serotype 4av	4	5 ci[a]	0	
S. flexneri serotype 4b	1	B4B	0	
S. flexneri serotype 4c	3	3 ci[a]	0	
S. flexneri serotype 5b	1	B5	1	1 ci[a]
S. flexneri serotype 6	10	B6; 9 ci[a]	10	10 ci[a]
S. flexneri serotype X	1			
S. flexneri serotype Y	2	2 ci[a]	2	2 ci[a]
S. flexneri serotype Yv	2	2 ci[a]	0	
S. flexneri provisional	5	5 ci[a]	0	
Total number of *S. flexneri*	**79**		**77**	
S. boydii serotype 1	2	AMC-43-G-58; 1 ci[a]	3	3 ci[a]
S. boydii serotype 2	3	CIP 82.50[T]; P288; 1 ci[a]	4	4 ci[a]
S. boydii serotype 3	1	D1	0	
S. boydii serotype 4	2	AMC-43-G-63; 1 ci[a]	2	2 ci[a]
S. boydii serotype 5	2	P143; 1 ci[a]	0	
S. boydii serotype 6	1	CDC 9771 (D19)	0	
S. boydii serotype 7	1	AMC 4006 (Lavington)	0	
S. boydii serotype 8	0		1	1 ci[a]
S. boydii serotype 9	1	1296/7	0	
S. boydii serotype 10	1	430	1	1 ci[a]
S. boydii serotype 11	1	34	0	
S. boydii serotype 12	0		1	1 ci[a]
S. boydii serotype 13	0		1	1 ci[a]
S. boydii serotype 14	0		1	1 ci[a]
S. boydii serotype 15	1	CDC C-703	0	
S. boydii serotype 18	1	1 ci[a]	1	1 ci[a]
Total number of *S. boydii*	**17**		**15**	
S. sonnei	**117**	CIP 82.49[T]; 116 ci[a]	**115**	115 ci[a]
EIEC	**30**	DSM 9027; DSM 9028; CCUG 11335; CCUG 38080; CCUG 38092; CCUG 38093; EW227; 1624-56; 1184-68; 145/46; L119B-10; 19 ci[a]	**31**	31 ci[a]
Other *E. coli* pathotypes (human)	**11**	7 STEC ci[a], 4 EPEC ci[a]	**11**	8 STEC ci[a]; 3 EPEC ci[a]
Other *E. coli* pathotypes (animal)[b]	**10**	5 mussel, 3 pigeon, 2 turkey	**10**	4 mussel, 3 pigeon, 2 turkey, 1 oyster

123   [a]*ci = clinical isolate.* [b]*isolated from animals, all other numbers are reference isolates.*

124   highest log-score of at least 2.000 was considered as the definitive MALDI-TOF identification, as is

125   done in a routine workflow. If an isolate had a log-score < 2.000, it was disregarded from further

126   analysis. Isolates were then assigned to different discrimination levels "genus", "pathotype", "group" and

127   "species" as displayed in Figure 1.

Figure 1   The classes in the different discrimination levels to which isolates were assigned

For accurate identification, only matches with database MSPs from the same species within a log-score

range of 2.000-2.300 or >2.300 should be expected in one spot. To assess this, the ten MSPs from the

database that produced the highest scores within a log-score of 2.000-2.300 or >2.300 per spot were

determined. For each species identified with the culture dependent identification algorithm, the median

number of species and quartile ranges per spot that had a log-score of 2.000-2.300 or >2.300 were

calculated and visualized using SPSS 24.0.0.1.

Biomarker assignment and principal component analysis

Spectra files from MSPs of 288 isolates in the training set were exported as mzXML files using Compassxport CXP3.0.5. (Bruker Daltonik) or exported via a batch process in Flexanalysis (Bruker Daltonik). A new database was created in Bionumerics v7.6.3 (Applied Maths NV, http://www.applied-maths.com/) according to the manufacturers' instructions. All raw spectra were imported into the Bionumerics database with x-axis trimming to a minimum of 2000 m/z. Baseline subtraction, noise computing, smoothing, baseline detection and peak detection were performed with default settings. Spectra summarizing, peak matching and peak assignment was performed according to instructions from Bionumerics [22]. In short, all raw spectra were summarized into isolate spectra. Peak matching was performed on isolate spectra using a constant tolerance of 1.9, a linear tolerance of 550 and a peak detection rate of 10%. Binary peak matching tables were exported to summarize the presence of peak classes on all discrimination levels as depicted in Figure 1. For the levels genus, pathotype and groups, decision diagrams were produced (Online Resource 1a, 1b and 1c). The spectra files of isolates from the test set were imported and preprocessed in Bionumerics, using the same methods and settings as with spectra from the isolates in the training set. Peak matching with a constant tolerance of 1.9 and a linear tolerance of 550 on spectra of test isolates was performed using the option "existing peak classes only" to compare the presence of peaks in the test isolates with the presence of peaks in the isolates from the training set. Decision diagrams (Online Resource 1) and the presence or absence of peak masses as depicted in Table 2 were applied to assign unknown isolates from the test set according to the different levels as shown in Figure 1.

By assigning biomarkers, only presence and absence of peaks was investigated. To assess also quantitative peak data such as peak intensity and peak area, a principal component analysis (PCA) was performed on all isolates in the training set to visualize the position of isolates in three dimensions.

Table 2 Discrimination scheme of biomarkers, percentage of isolates in the training set with presence of specific biomarkers

Biomarkers (m/z)		2691	2877	3129	3636	3647	3930	3939	4163	4189	4368	4501	4769	4775
S. dysenteriae	(n=24)	92	4	100	0	0	0	100	0	100	100	100	0	100
S. flexneri	(n= 46)	100	0	100	0	63	53	18	1	94	97	99	22	97
S. boydii	(n=17)	88	0	100	18	0	0	94	0	100	88	100	18	100
S. sonnei	(n=117)	56	49	59	22	0	1	56	17	89	68	56	23	98
EIEC	(n=31)	100	0	100	3	6	0	97	0	97	100	94	26	97
Other E. coli	(n=21)	52	24	52	71	0	5	62	38	90	67	57	67	100

Biomarkers (m/z)	4784	5156	5239	5386	5415	6262	6322	6412	6488	7275	7295	7715	7868	787
S. dysenteriae	100	100	8	92	52	100	100	0	0	0	0	0	0	100
S. flexneri	76	97	55	99	45	99	99	4	42	1	83	0	41	23
S. boydii	76	94	18	88	59	100	88	18	6	18	0	6	12	88
S. sonnei	69	86	27	73	0	74	62	13	1	17	0	28	18	75
EIEC	71	100	39	100	42	94	97	19	0	23	6	3	16	84
Other E. coli	19	86	0	67	0	57	67	48	10	52	0	24	43	33

Biomarkers (m/z)	8326	8370	8379	9002	9227	9535	9546	9563	9739	10300	10310	10488	10934
S. dysenteriae	0	0	100	100	0	0	100	100	0	0	100	0	0
S. flexneri	5	4	90	94	4	17	82	92	8	9	86	0	0
S. boydii	0	18	82	88	12	18	82	88	6	18	82	0	12
S. sonnei	15	15	82	38	12	16	85	54	15	13	70	34	36
EIEC	6	16	77	87	13	23	77	81	6	19	77	0	0
Other E. coli	43	38	38	24	48	48	62	38	43	48	48	0	5

Presence of biomarkers identified in previous studies

All isolates in the training set as well as in the test set were examined for the unique masses (± 500 ppm) that were found in biomarker assignment to *Shigella spp.* and *E. coli* in previous studies [15,14]. Additionally, because peaks in our study were assigned at m/z values instead of masses only and masses could be potentially charged with two electrons, this is corrected for by also examining the previously published masses divided by two (±500 ppm) [14,15].

Classifier models based on machine learning

Peak data of the summarized isolate spectra of the 288 isolates in the training set was used to define and train machine learning based classifiers using Bionumerics v7.6.3 according to the manufacturers' instructions. In short, peak matching with a constant tolerance of 1.9 and a linear tolerance of 550 was performed on isolate spectra on the different levels: genus, pathotype, group and species. Classifiers were created at all levels using character values. Support Vector Machine (linear) learning was used as

193 a scoring method in which p-values were used as a rank score. The classifiers were trained and cross-

194 validated to check their performance for identification. Subsequently, the classifier models were used to

195 classify the unknown isolates in the test set at the different discrimination levels to evaluate their

196 performance.

Results

199 **Database development**

200 All MSPs of 288 training isolates were added to a custom-made database, the relatedness of these

201 MSPs is shown in a dendrogram (Figure 2). The Maldi Biotyper OC software recognized three large

202 clusters of MSPs that are not species-specific within this custom database. This did not change if

203 clusters were assigned manually with a lower distance level at 50-100 relative units, indicating that

204 similarity in spectrum profiles is distributed over the species level (Figure 2).

206 **Figure 2**  Dendrogram of MSPs of training isolates
207 *Blue = cluster 1; green = cluster 2; red = cluster 3. Yellow/blue vertical band = manual cluster distinction at distance level 50-*
208 *100 relative units with species designation using the culture-based identification algorithm.*

Additionally, the duplicate spots of test isolates using either the direct smear or extraction method

resulted in a different species designation in 10-15% of the samples. Furthermore, with an accurate

distinction of species one would not expect assignment to multiple species above the threshold of log-

score 2.000. However, with both application methods, most isolates were assigned to several species

with a log-score of 2.000-2.300 or >2.300 per spot (Figure 3).

One isolate from the test set (*S. boydii* serotype 13) had a low-quality spectrum (log score 1.574-1.930)

and one isolate *(S. dysenteriae* serotype 1) was originally incorrectly stored, as this isolate was

identified as *Corynebacterium diphtheriae* using the Bruker databases. Both these isolates were

ignored in further analyses. All other isolates had log-scores higher than 2.000, and percentages of

Figure 3   Number of different species in the first 10 matches per spot with the direct smear method
Identity (x-axis) was assigned using the culture-based identification algorithm. Black horizontal bars represent the median
number of species, the 25%-75% interquartile ranges are indicated by the blue vertical bars, and 5%-95% intervals by the
black vertical lines. Outliers are indicated with blue dots.

MALDI-TOF identification that were concordant with the original identification on all discrimination levels

were as displayed in Table 3. With the Bruker databases only, percentages of correctly identified

Shigella spp. on all discrimination levels is low, ranging from 6% to 45% correct designations, both for

the direct smear method as for the extraction method (Table 3). In contrast, 90%-100% of *E. coli*

isolates were correctly identified. When identification was based on the custom-made database with or

without the Bruker databases, the percentage of correctly identified *E. coli* isolates decreased from

29%-71%, while *Shigella spp.* were correctly identified in 94%-99% of cases on the genus, pathotype

and group levels. In addition, 91%-97% of *S. flexneri* and *S. sonnei* were correctly identified at the

species level, in contrast to *S. dysenteriae* and *S. boydii*, for which the percentages of correct

identification were poor (Table 3).

Biomarker assignment and principal component analysis

The decision diagrams based on biomarkers assigned to the isolates in the training set were used for

identification of unknown isolates in the test set. Distinctive peaks on the species levels were

summarized in Table 2. High percentages for correct identification of *S. sonnei* isolates were achieved

at the species level using both the direct smear as the extraction method. However, the biomarkers are

not specific for *S. sonnei* as other species also contain them. For other species the identified

biomarkers correctly identified isolates below 38%. Specific biomarkers were not detected for all the

classes at the different discrimination levels as depicted in Figure 1. Consequently, it was not possible

to identify *S. dysenteriae, S. boydii,* and *E. coli* isolates at all, because of the absence of discriminating

peaks for these species (Table 3).

In the PCA of the detected peaks in the isolates of the training set, one large cluster was formed, with

at both ends a few outliers (Figure 4). If the isolates were colored according to their identity based on

the culture-based identification method, in none of the discrimination levels separate groups of isolates

were seen (Figure 4a-d).

Table 3 Correct identification results of isolates from test set

	Correct identification with MALDI-TOF, direct smear:					Correct identification with MALDI-TOF, ethanol-formic acid extraction:				
	Bruker databases[a], n (%)	Custom-made database, n (%)	Bruker databases[a] + custom-made, n (%)	Biomarker assignment, n (%)	Classifier models, n (%)	Bruker databases[a], n (%)	Custom-made database, n (%)	Bruker databases[a] + custom-made, n (%)	Biomarker assignment, n (%)	Classifier models, n (%)
Genus										
Shigella (n =217)	19 (9)	205 (94)	205 (94)	10 (5)	209 (96)	12 (6)	207 (95)	205 (94)	15 (7)	217 (100)
E. coli (n = 52)	49 (94)	26 (50)	29 (56)	NA	11 (21)	47 (90)	35 (67)	37 (71)	NA	4 (8)
Unassigned[b]	2 (1)	1 (0.4)	3 (1)	257 (96)	0 (0)	1 (0.4)	0 (0)	3 (1)	250 (93)	0 (0)
Pathotype										
Shigella/EIEC (n = 248)	NA	233 (94)	241 (97)	217 (88)	145 (58)	NA	245 (99)	242 (98)	225 (92)	147 (59)
Other *E. coli* (n =21)	21 (100)	6 (29)	10 (48)	NA	14 (67)	21 (100)	11 (52)	13 (62)	NA	6 (29)
Unassigned[b]	2 (1)	1 (0.4)	3 (1)	46 (17)	0 (0)	0 (0)	0 (0)	3 (1)	27 (10)	0 (0)
Group										
Shigella (n = 217)	19 (9)	205 (94)	205 (94)	193 (89)	131 (60)	12 (6)	207 (95)	205 (94)	195 (90)	134 (62)
EIEC (n = 31)	NA	9 (29)	8 (26)	NA	2 (6)	NA	19 (61)	19 (61)	NA	0 (0)
Other *E. coli* (n =21)	21 (100)	6 (29)	10 (48)	NA	13 (62)	21 (100)	11 (52)	13 (62)	NA	7 (33)
Unassigned[b]	2 (1)	1 (0.4)	3 (1)	49 (23)	0 (0)	1 (0.4)	0 (0)	3 (1)	36 (13)	0 (0)
Species										
S. dysenteriae (n =11)	5 (45)	5 (45)	5 (45)	0 (0)	0 (0)	4 (36)	7 (64)	6 (55)	0 (0)	0 (0)
S. flexneri (n =77)	NA	70 (91)	70 (91)	24 (31)	6 (8)	NA	73 (95)	73 (95)	30 (39)	3 (4)
S. boydii (n =14)	NA	1 (7)	0 (0)	0 (0)	0 (0)	NA	0 (0)	0 (0)	0 (0)	0 (0)
S. sonnei (n = 115)	NA	110 (96)	110 (96)	113 (98)	92 (80)	NA	112 (97)	112 (97)	108 (94)	101 (88)
EIEC (n = 31)	NA	9 (29)	8 (26)	0 (0)	1 (3)	NA	19 (61)	19 (61)	1 (3)	3 (10)
Other *E. coli (n =21)*	21 (100)	6 (29)	10 (48)	0 (0)	12 (57)	21 (100)	11 (52)	13 (62)	0 (0)	4 (19)
Unassigned[b]	2 (1)	1 (0.4)	3 (1)	85 (32)	0 (0)	1 (0.4)	0 (0)	3 (1)	97 (36)	0 (0)

252 *NA = not applicable, as no discriminating peaks were assigned to these classes. [a]Bruker MALDI Biotyper database (V8.0.0.0) and the Bruker Security-*
253 *Relevant Library (V1.0.0.0). [b]Number of isolates that could not be assigned to a class.*

Figure 4   PCA of isolates in the training set

A. Colored at genus level: beige = Shigella; teal = Escherichia. B. Colored at pathotype level: black = Shigella/EIEC; green = E. coli (other than EIEC). C. Colored at group level: orange = Shigella spp.; yellow = EIEC; purple = E. coli (other than EIEC). D. Colored at species level: light blue = S. dysenteriae; red = S. flexneri; green = S. boydii; pink = S. sonnei; blue = EIEC; light grey = Other E. coli.

Presence of biomarkers identified in previous studies

The specific biomarkers for *S. flexneri*, *S. sonnei* and *E. coli* assigned by Everley *et al.* [15] were not

present in any of the 559 isolates in this study when using an error limit of ± 500 ppm. Even if they were

corrected for a charge with 2 electrons, they were not present. A few biomarkers for *Shigella spp.* and

E. coli described by Khot and Fisher [14] were present within a range of 500 ppm in isolates used in

this study, i.e., 4163 Da, 7157 Da, 8326 Da, and 9227 Da, and corrected for a charge of 2 electrons,

5096 Da and 5752 Da.

Classifier models based on machine learning

Using the internal cross-validation of the classifiers at all discrimination levels, all but one class had an

accuracy of more than 87.5%. The only class with a lower accuracy (77%) was the class "*Escherichia*"

at the genus discrimination level.

When using machine learning based classifiers for identification, 96% of *Shigella spp.* isolates and 21%

of the *E. coli* isolates from the test set were correctly identified at the genus level, using the direct

smear application method and respectively 100% and 8% using the ethanol-formic acid extraction

method (Table 3). Correct identification percentages for the pathotype, group and the species level

were displayed in Table 3. Although more than 80% of *S. sonnei* isolates were correctly identified with

the species classifier, specificity is low, as more than 70% of *S. flexneri* isolates were also identified as

S. sonnei.

Discussion

Current commercially available MALDI-TOF databases cannot distinguish between *Shigella spp.* and

E. coli. Therefore, three different alternatives were explored in this study. A custom-made database

was developed, biomarkers were identified and classification models using machine learning were

designed. All methods were tested using spectra of *Shigella spp.*, *E. coli* and EIEC isolates obtained by

the direct smear method and the ethanol-formic acid extraction method. The latter method resulted in

slightly more correctly identified isolates at all discrimination levels. However, in routine clinical

diagnostics, identification relies on the direct smear method, as the extraction method is more time-

consuming and laborious, especially for large sample quantities.

With the custom-made database with or without combination with the Bruker databases, >91% of *S.*

flexneri and *S. sonnei* were correctly identified at the species level. For *S. boydii* and *S. dysenteriae*,

our custom-made database performed poor as has been shown before [11]. However, compared to a

previous study our custom-made database assigned less *E. coli* isolates correctly [11]. This indicates

that the inclusion of EIEC isolates in the custom-made database as well as in the test set complicates

the identification. Half of the EIEC isolates were assigned to one of the *Shigella* species, thereby

decreasing the percentage of correctly identified *E. coli.* Additionally, the poor performance of

identifying *E. coli* with our custom-made database can be the result of an overrepresentation of *S.*

flexneri and *S. sonnei*. To investigate this hypothesis, a second custom-made database was

developed, based on 17 isolates of each species, representing the diversity in serotypes. This database

did not perform better or worse than the custom-made database that contained 288 MSPs (see Online

Resource 2), indicating that a more evenly distribution of species in the database does not improve

identification of *E. coli*. Although percentages of correct species assignments to *S. flexneri* and *S.*

sonnei were high, other species were falsely assigned to them, both in our study as in a previous study

[11]. In the latter study, correct species identification was based on the majority rule that three out of

four spots should indicate the same species. Besides the fact that the interpretation of four spots per

isolate is not feasible in clinical diagnostics, this indicates that distinction of spectra is poor and

assignment of species is based on probabilities rather than actual variations in spectra. Our study

confirms this phenomenon, because multiple species identifications within the same log-score range

were made per spot. Moreover, 10-15% of duplicate spots resulted in different species assignments

using the commercially available databases as well as the custom-made database. Another designation

for poor distinction was shown in the dendrogram that was inferred from the MSPs in the custom-made

database. Because the MSPs of the same species were not clustering together, one could expect on

forehand that the resulting database would not be capable to identify the isolates from the test set

correctly.

Another alternative approach for the use of commercially available databases is the detection of

discriminating biomarkers. However, in our study many isolates had an inconclusive identification, as

specific biomarkers were not detected for most classes. Although more than 90% of *S. sonnei* isolates

were identified at the species level, other species are also frequently falsely identified as *S. sonnei*,

indicating poor distinctive properties of the assigned biomarkers.

Moreover, when analyzing also peak intensity and peak area rather than just peak presence, the PCA

showed that *Shigella spp.* and *E. coli* did not represent separated groups based on their biomarkers.

One large cluster was formed, containing *Shigella spp.*, *E. coli* and EIEC, with a few outliers consisting

only of *S. sonnei* isolates. Furthermore, only six out of 40 (15%) of the unique masses described for

Shigella spp./E. coli in a former study were also present in our biomarker list, indicating that these

biomarkers are stably present in the *E. coli* and *Shigella* population [14]. The absence of 85% of the

masses that were assigned as biomarkers in this former study in our isolates, indicates that the

detected biomarkers vary amongst isolate sets tested, and that a stable variation per species is not

observed. Consequently, we anticipate that assignment of biomarkers based on yet another additional

set of isolates will lead to even more diversity in biomarkers, indicating their unsuitability for distinct

identification of *Shigella spp.*, *E. coli* and EIEC.

The use of classifier models based on machine learning resulted in comparable percentages of

correctly identified *Shigella* on the genus level, i.e. ≥ 94%, as reported in other studies [14]. Our

classifier models also achieved a comparable percentage of correctly identified *E. coli* as earlier

published semi-automated classifiers without EIEC isolates. In our classifier model designed on the

pathotype level, EIEC isolates were not incorporated in the class *E. coli* and correct identification was

67%, compared to 56% in a previous study [14]. Nonetheless, the other remaining *E. coli* isolates were

falsely classified as *Shigella* both with our classifiers and with previously published ones [14].

Therefore, this decreases the specificity for the distinction of *Shigella,* making the classifiers unsuitable

to use in clinical diagnostics. At the group and species level, classifiers performed even less, and most

species could not be identified at all. It is possible that the poor performance of the classifier models

was caused by an overrepresentation of *S. flexneri* and *S. sonnei,* as discussed for the custom-made

database in our study. Therefore, the selection of 17 isolates of each species was used again and

alternative classifiers were designed. These classifiers did not perform better or worse than the

classifiers designed using all 288 isolates in the training set, indicating that an absence of an evenly

distribution of species was not the cause for poor identification with classifiers (Online Resource 2).

Compared to previous studies, we used a substantial larger set of isolates and included the *E. coli*

pathotype EIEC. Another strength was that multiple alternative approaches for the identification of

Shigella spp. and *E. coli* using MALDI-TOF MS were explored. Although *S. sonnei* and *S. flexneri*

isolates were overrepresented in both the training set and the test set, this distribution is representative

for high-resource settings.

In conclusion, none of our explored alternative approaches for identification of *Shigella spp.*, *E. coli* and

EIEC with MALDI-TOF MS was suitable to use in clinical diagnostics as all rendered a poor distinction

based on spectra or biomarkers. We anticipate that with the use of an even larger and more diverse set

of isolates, distinction of *Shigella spp.*, *E. coli* and EIEC based on spectra using MALDI-TOF MS will

remain challenging. Therefore, we propose an identification algorithm in which MALDI-TOF is used for

the identification and differentiation of *Shigella/E. coli* as a group from other *Enterobacteriaceae*,

followed by tests other than MALDI-TOF MS to distinguish between the different *Shigella species*, *E.*

coli, and specific *E. coli* pathotypes including EIEC.

Author contributions:

MB, JR, MK, and FR conceived and designed the project. MB and AE performed the analyses. MB, JR, MK, and FR interpreted results. MB wrote the manuscript. All authors read, reviewed and approved the final manuscript.

Potential conflict of interest

The authors declare that they have no conflict of interest.

Data availability statement

The datasets generated during and/or analyzed during the current study are available from the corresponding author on reasonable request.

Electronic Supplementary Material

ESM_1.pdf   Decision diagrams of assigned biomarkers

ESM_2.pdf   Comparison of performance of a custom-made database and classifiers based on all 288 isolates or based on an evenly distribution of 17 isolates for each species

References

1. DuPont HL, Formal SB, Hornick RB, Snyder MJ, Libonati JP, Sheahan DG, LaBrec EH, Kalas JP (1971) Pathogenesis of *Escherichia coli* diarrhea. N Engl J Med 285 (1):1-9. doi:10.1056/NEJM197107012850101

2. Lan R, Alles MC, Donohoe K, Martinez MB, Reeves PR (2004) Molecular evolutionary relationships of enteroinvasive *Escherichia coli* and *Shigella spp*. Infect Immun 72 (9):5080-5088. doi:10.1128/IAI.72.9.5080-5088.2004 72/9/5080 [pii]

3. Pettengill EA, Pettengill JB, Binet R (2015) Phylogenetic analyses of *Shigella* and enteroinvasive

Escherichia coli for the identification of molecular epidemiological markers: whole-genome comparative

analysis does not support distinct genera designation. Front Microbiol 6:1573.

doi:10.3389/fmicb.2015.01573

4. Kaper JB, Nataro JP, Mobley HL (2004) Pathogenic *Escherichia coli*. Nat Rev Microbiol 2 (2):123-

140. doi:10.1038/nrmicro818

5. Bopp CAB, F.W.; Fields, P.I.; Wells, J.G.; Strockbine, N.A. (2003) *Escherichia*, *Shigella* and

Salmonella. In: Murray PR (ed) Manual of Clinical Microbiology, vol 1. 8th edn. ASM Press, Washington

D.C., pp 654-671

6. van den Beld MJ, Reubsaet FA (2012) Differentiation between *Shigella*, enteroinvasive *Escherichia*

coli (EIEC) and noninvasive *Escherichia coli*. Eur J Clin Microbiol Infect Dis 31 (6):899-904.

doi:10.1007/s10096-011-1395-7

7. Martiny D, Busson L, Wybo I, El Haj RA, Dediste A, Vandenberg O (2012) Comparison of the

Microflex LT and Vitek MS systems for routine identification of bacteria by matrix-assisted laser

desorption ionization-time of flight mass spectrometry. J Clin Microbiol 50 (4):1313-1325.

doi:10.1128/JCM.05971-11

8. Cameron M, Perry J, Middleton JR, Chaffer M, Lewis J, Keefe GP (2018) Short communication:

Evaluation of MALDI-TOF mass spectrometry and a custom reference spectra expanded database for

the identification of bovine-associated coagulase-negative staphylococci. J Dairy Sci 101 (1):590-595.

doi:10.3168/jds.2017-13226

9. Seuylemezian A, Aronson HS, Tan J, Lin M, Schubert W, Vaishampayan P (2018) Development of a

custom MALDI-TOF MS database for species-level identification of bacterial isolates collected from

spacecraft and associated surfaces. Front Microbiol 9:780. doi:10.3389/fmicb.2018.00780

10. Forero Morales MP, Hopprich R, Wise R, Shephard L, Weldhagen GF (2018) Cost-effective

implementation of a custom MALDI-TOF library for the identification of South Australian *Nocardia*

isolates. Pathology 50 (7):753-757. doi:10.1016/j.pathol.2018.08.006

11. Paauw A, Jonker D, Roeselers G, Heng JM, Mars-Groenendijk RH, Trip H, Molhoek EM, Jansen

HJ, van der Plas J, de Jong AL, Majchrzykiewicz-Koehorst JA, Speksnijder AG (2015) Rapid and

reliable discrimination between *Shigella species* and *Escherichia coli* using MALDI-TOF mass

spectrometry. Int J Med Microbiol 305 (4-5):446-452. doi:10.1016/j.ijmm.2015.04.001

12. Josten M, Reif M, Szekat C, Al-Sabti N, Roemer T, Sparbier K, Kostrzewa M, Rohde H, Sahl HG,

Bierbaum G (2013) Analysis of the matrix-assisted laser desorption ionization-time of flight mass

spectrum of *Staphylococcus aureus* identifies mutations that allow differentiation of the main clonal

lineages. J Clin Microbiol 51 (6):1809-1817. doi:10.1128/JCM.00518-13

13. Sauget M, Valot B, Bertrand X, Hocquet D (2017) Can MALDI-TOF Mass Spectrometry Reasonably

Type Bacteria? Trends Microbiol 25 (6):447-455. doi:10.1016/j.tim.2016.12.006

14. Khot PD, Fisher MA (2013) Novel approach for differentiating *Shigella species* and *Escherichia coli*

by matrix-assisted laser desorption ionization-time of flight mass spectrometry. J Clin Microbiol 51

(11):3711-3716. doi:10.1128/JCM.01526-13

15. Everley RA, Mott TM, Wyatt SA, Toney DM, Croley TR (2008) Liquid chromatography/mass

spectrometry characterization of *Escherichia coli* and *Shigella species*. J Am Soc Mass Spectrom 19

(11):1621-1628. doi:S1044-0305(08)00578-3 [pii]10.1016/j.jasms.2008.07.003

16. De Bruyne K, Slabbinck B, Waegeman W, Vauterin P, De Baets B, Vandamme P (2011) Bacterial

species identification from MALDI-TOF mass spectra through data analysis and machine learning. Syst

Appl Microbiol 34 (1):20-29. doi:https://doi.org/10.1016/j.syapm.2010.11.003

17. Mather CA, Werth BJ, Sivagnanam S, SenGupta DJ, Butler-Wu SM (2016) Rapid Detection of

Vancomycin-Intermediate *Staphylococcus aureus* by Matrix-Assisted Laser Desorption Ionization-Time

of Flight Mass Spectrometry. J Clin Microbiol 54 (4):883-890. doi:10.1128/JCM.02428-15

18. Ho PL, Yau CY, Ho LY, Chen JHK, Lai ELY, Lo SWU, Tse CWS, Chow KH (2017) Rapid detection

of cfiA metallo-beta-lactamase-producing *Bacteroides fragilis* by the combination of MALDI-TOF MS

and CarbaNP. J Clin Pathol 70 (10):868-873. doi:10.1136/jclinpath-2017-204335

19. Clermont O, Bonacorsi S, Bingen E (2000) Rapid and simple determination of the *Escherichia coli*

phylogenetic group. Appl Environ Microbiol 66 (10):4555-4558. doi:10.1128/aem.66.10.4555-4558.2000

20. van den Beld MJC, de Boer RF, Reubsaet FAG, Rossen JWA, Zhou K, Kuiling S, Friedrich AW,

Kooistra-Smid M (2018) Evaluation of a culture dependent algorithm and a molecular algorithm for

identification of *Shigella spp.*, *Escherichia coli,* and enteroinvasive *E. coli* (EIEC). J Clin Microbiol

56:e00510-00518. doi:10.1128/JCM.00510-18

21. Saleeb PG, Drake SK, Murray PR, Zelazny AM (2011) Identification of mycobacteria in solid-culture

media by matrix-assisted laser desorption ionization-time of flight mass spectrometry. J Clin Microbiol

49 (5):1790-1794. doi:10.1128/JCM.02135-10

22. Maths A Bionumerics: seven tutorials, spectra. http://www.applied-maths.com/tutorials#Spectra.

Accessed 16-07-2019