Single-cell radioluminescence microscopy with two-fold higher sensitivity using dual scintillator

configuration

Tae Jin Kim[1], Qian Wang[2], Mark Shelor[3], and Guillem Pratx[1]

[1] Department of Radiation Oncology, Stanford School of Medicine, Stanford, California,

USA

[2] Department of Bioengineering, University of California, Davis, California, USA

[3] Department of Biomedical Engineering, University of California, Merced, California, USA

Abstract

Radioluminescence microscopy (RLM) is an imaging technique that allows quantitative analysis

of clinical radiolabeled drugs and probes in single cells. However, the modality suffers from slow

data acquisition (10 – 15 minutes), thus critically affecting experiments with short-lived

radioactive drugs. To overcome this issue, we suggest an approach that significantly accelerates

data collection. Instead of using a single scintillator to image the decay of radioactive molecules,

we sandwiched the radiolabeled cells between two scintillators. As proof of concept, we imaged

cells labeled with [^{18}F]FDG, a radioactive glucose popularly used in oncology to image tumors.

Results show that the double scintillator configuration increases the microscope sensitivity by two-

fold, thus reducing the image acquisition time by half to achieve the same result as the single

scintillator approach. The experimental results were also compared with Geant4 Monte Carlo

simulation to confirm the two-fold increase in sensitivity with only minor degradation in spatial

resolution. Overall, these findings suggest that the double scintillator configuration can be used to perform time-sensitive studies such as cell pharmacokinetics or cell uptake of short-lived radiotracers.

Introduction

Positron emission tomography (PET) is a molecular imaging technique that enables interrogation of biophysical processes in living subjects in a non-invasive manner. It is popularly used in the clinic to diagnose and characterize various diseases such as cancer, cardiovascular disorders, and neurological disorders using a wide range of radiotracers [1–4], including radioactive glucose 2-deoxy-2-[^{18}F]fluorodeoxyglucose or [^{18}F]FDG. [^{18}F]FDG is widely used in the clinic for detecting and staging cancer [5–8].

Due to their clinical significance, novel radiopharmaceuticals are actively investigated for diagnostic and therapeutic purposes. However, the biological activity of radiopharmaceuticals is difficult to confirm—the spatial resolution of current clinical and pre-clinical imaging systems is limited to the tissue level. This makes it difficult to acquire detailed information of how these radioactive molecules interact with target cells that present heterogeneous molecular characteristics. Techniques that enable single-cell radionuclide detection (e.g., micro-autoradiography [9,10]) are challenging to implement and not applicable to live cells, and are thus rarely used.

Radioluminescence microscopy (RLM) was introduced to enable *in vitro* radionuclide imaging of live single cells. In RLM, a scintillator crystal is placed directly under or above the radiolabeled cells to observe optical flashes resulting from the decay of single radioactive molecules. By

capturing a series of image frames and individually counting the scintillation flashes within each

frame, RLM can quantify how many radioactive molecules are present within individual cell with

high sensitivity – down to fewer than 1,000 molecules per cell. The unique capabilities of RLM

have revealed previously unattainable information on single-cell response to clinical radiotracers

[11–13]. However, one limitation of this modality is the relatively long image acquisition time,

which ranges from 15 to 30 minutes per sample.

In this study, we introduce a simple method to significantly reduce image acquisition time.

Compared to the original approach, which uses a single scintillator to capture radioluminescence

signals, we sandwich the radiolabeled cells between two scintillators. This allows complete

geometric coverage for increased detection efficiency. As proof-of-concept, human breast cancer

cells (MDA-MB-231) grown on a cadmium tungstate ($CdWO_4$) scintillator are incubated with

[^{18}F]FDG and imaged with both single and dual scintillator configurations. The experimental

results are further compared with Monte Carlo simulations to demonstrate that the detection

sensitivity of RLM doubles when using two scintillators.

Materials and methods

Sample preparation

Cell culture

A cadmium tungstate scintillator ($CdWO_4$, MTI Co.) with a dimension of $10 \times 10 \times 0.5$ mm (width

$\times$ length $\times$ height) was coated with 100 µl of fibronectin (5 µg/ml) to promote cell adhesion to the

crystal surface. The scintillators were incubated for ~2 hours and washed three times with sterile

distilled water. The treated scintillator was placed on a 35 mm diameter cover-slip bottom dish (µ-

Dish, ibidi GmbH), and MDA-MB-231 human breast cancer cells suspended in DMEM were

dispensed in the dish with cell density of 10^4 cells/ml. The cells were then incubated in a CO_2

incubator for 24 h prior to the experiment to allow them to attach to the scintillator surface.

Radiolabeling cells

The MDA-MB-231 cells were labeled with radioactive glucose analogue [18F]FDG, a radiotracer

that is commonly used in the clinic to detect and stage cancer in patients. The cells were first

incubated in glucose-free DMEM for 30 – 40 minutes. The culture media was then replaced with

a solution of glucose-free DMEM containing 22 – 24 KBq/ml of [18F]FDG, and the cells were

incubated for an additional 45 minutes. After the radiolabeling process, the cells were washed with

clear glucose-free DMEM.

Radionuclide imaging

Radioluminescence imaging was performed with a low-light microscope developed in-house [14].

The microscope consists of a 20× / 0.75 NA microscope objective lens (Nikon, CFI Plan Apo

Lambda) coupled to a 36 mm tube lens. This yields an effective magnification of 3.6× while

maintaining the native numerical aperture of the objective lens [15]. The microscope is mounted

on top of a highly sensitive EMCCD camera (C9100-13, Hamamatsu Co.). To protect the EMCCD

camera from stray light and to minimize the background signal, the entire microscope was enclosed

in a customized light-tight box.

The culture dish containing the scintillator with MDA-MB-231 cells was then placed on the

microscope stage. A brightfield image was first captured with the camera settings set to standard

mode (non-EMCCD mode). The first RLM image was then acquired (Figure 1a), with the camera

parameters set to 1,060 EM gain, 4×4 binning, 30 ms exposure time, and 10,000 acquisition

frames. After the first set of images, another scintillator was gently placed on top of the cells (Fig

1b). A second set of RLM images was captured with identical camera parameters. After both

images were acquired, the culture dish was removed from the microscope stage and a separate dark

reference sequence was captured (1,000 frames) for background subtraction.

Fig 1. Schematic diagrams. Radioluminescence microscopy with (a) single scintillator and (b)

double scintillator configurations.

The radioactive decay signals were analyzed using the ORBIT toolbox (optical reconstruction of

the beta-ionization track) [16]. In ORBIT, the reconstruction process starts by subtracting the

background noise from each RLM image. Individual scintillation flashes are then isolated within

each image frame and converted into (x, y) event coordinates. Once this process is repeated for the

entire set of 10,000 images, the positions of all detected events are aggregated into a single image,

where each pixel represents the number of radioactive decay events detected at that location.

Detailed reconstruction procedures and radioluminescence imaging can also be found from our

previous papers [14,16,17].

Monte Carlo Simulation

Experimental results were also compared by simulating the radioluminescence process using a Monte Carlo software package (Geant4), as described in our previous work [18,19]. A single ^{18}F point source with 1 Bq of radioactivity was generated in virtual space to represent a single radiolabeled cell. For the single scintillator experiment, the source was positioned 5 μm above a 100 μm-thick $CdWO_4$ slab. For the double scintillator configuration, a second $CdWO_4$ scintillator was placed 5 μm above the source, i.e. with the ^{18}F point source equidistant from the two scintillators. A total of 20,000 radioactive decay events were simulated, and the corresponding radioluminescence images were generated. The raw images were then reconstructed using the same methods from the previous section with the ORBIT toolbox.

Results

Radioluminescence images of single vs. double scintillator

As expected, experimental results demonstrate that the double scintillator configuration yields significantly higher radioactive decay count. RLM images acquired with the two configurations are shown in Figure 2a and 2b, respectively. The figures are shown in color using the same intensity scale, where red represents higher radioactive decay counts measured at each pixel. Also, the double scintillator configuration can detect radioactivity from cells that are otherwise undetectable using only one scintillator (highlighted circles in Figs 2c and 2d). It should be noted that the radioactive decay counts of the double scintillator case were adjusted to correct for the

half-life of ^{18}F ($\tau_{1/2}$ ~110 minutes). Since the double scintillator image was acquired 21 minutes

after the single scintillator image, the displayed count values were increased by 14%.

Fig 2. RLM images of [^{18}F]FDG uptake by MDA-MB-231 cells. Raw RLM data acquired with

(a) single scintillator and (b) double scintillator configuration. Composite image of

radioluminescence and brightfield image of (c) single scintillator and (d) double scintillator. Red

circles represent cells that are detectable with double scintillator only. Scale bar, 200 μm.

Sensitivity comparison

The number of radioactive decay counts detected for each cell was estimated by drawing a circular

region of interest (ROI; diameter, 90 μm) around individual cells ($N = 66$) and summing the

number of decay counts within each ROI. Each ROI measurement was then corrected by

subtracting a background signal, which was obtained by averaging the radioactive decay counts

detected in empty areas devoid of cells ($N = 66$ ROIs).

The number of decay events measured in a single cell, defined as D, is related to the number of

[^{18}F]FDG molecules N_0 within that cell at a reference time point. The relation can be expressed

by,

$$D = SYN_0\left(1 - \exp\left(-\frac{\ln 2}{\tau_{1/2}}t\right)\right) \tag{1}$$

144   where S is the detection sensitivity of the RLM system, Y is the radioactive yield for particulate

145   radiation (0.97 for [18]F), $\tau_{1/2}$ is the half-life (~110 minutes for [[18]F]FDG), and t is the elapsed time.

146   The sensitivity increase can thus be expressed in terms of the ratio,

$$\frac{S_{\text{double}}}{S_{\text{single}}} = \frac{D_{\text{double}} \cdot \exp\left(-\frac{\ln 2}{\tau_{1/2}}\Delta t\right)}{D_{\text{single}}} \tag{2}$$

where $\Delta t = t_{\text{double}} - t_{\text{single}}$ is the time delay between the two measurements. The exponent term

represents the decay correction due to time delay between the two experiments (and is equal to

1.14 for $\Delta t = 21$ min). This allows us to assume that the two measurements are simultaneously

performed on the same ROI. The increase in sensitivity was computed both for background ROIs

and ROIs containing single cells. For background signals (ROIs devoid of cells), we individually

quantified the radioluminescence signals (Fig 3a). While the data scatter is relatively large ($r^2 >$

0.69), results demonstrate that the double scintillator configuration is indeed more sensitive than

the single scintillator by a factor of 2. The average uptake values also show a two-fold increase in

sensitivity, 73 ± 34 counts/ROI (mean $\pm$ one standard deviation) and 37 ± 14 counts/ROI for double

and single scintillator, respectively.

Cell ROI are shown in Fig 3b for single and double scintillator configurations. As expected,

[[18]F]FDG uptake demonstrated a heterogeneous behavior, with some cells taking up the radiotracer

$3 - 4$ times more than other cells. Furthermore, the measured average cell uptake was increased by

two-fold when the second scintillator was added. The average uptake values were 407 ± 156

counts/cell for the double scintillator configuration and 200 ± 90 counts/cell for the single

scintillator case. Linear regression (dotted red line) of the two datasets found a linear relationship

between the two datasets, $y = 1.59\,x + 87.86$ ($r^2 > 0.83$).

Fig 3. Graphs representing single vs. double scintillator results. Scatter plot of decay counts detected in 90 μm ROI ($N = 66$) for (a) background devoid of cells and (b) areas containing single cells. The solid blue line and dotted red lines represent the linear regression fit.

Spatial Resolution

The spatial resolution of the single and dual scintillator cases was quantitatively assessed by drawing a line profile across two radioactive cells (Fig 4a). The profile was normalized according to the maximum count values for both configurations (Fig 4b). Results clearly show two peaks separated by a valley, which correspond to two cells with [^{18}F]FDG uptake along the line profile. Furthermore, no qualitative discrepancies in peak, valley, and background values can be observed between the two configurations. This simple comparison suggests that the sensitivity of RLM can be increased without significant degradation in spatial resolution.

Fig 4. Radioactive decay profile of two closely positioned cells. (a) A profile is drawn between two radioactive cells in the radioluminescence image. The image is represented in grayscale for presentation. Scale bar 200 μm. (b) Graph of decay count profiles comparing radioluminescence image captured with single scintillator and double scintillator. Each plot is normalized by their respective maximum count values.

To further assess the effects on spatial resolution, a Monte Carlo simulation of the system was

implemented using the GEANT4 package. A single ^{18}F point source (1 Bq) was simulated for both

configurations. The point-spread function measured by using the single and double scintillator

configuration is shown in Figs 5a and 5b, respectively. The simulation results agree well with the

experimental results – the double scintillator is more sensitive than the single scintillator system

by a factor of two. It can be seen from the simulation, however, that the single scintillator system

provides slightly higher spatial resolution than the double scintillator system. Using a circularly

symmetric 2-D Gaussian curve fit, we estimate the FWHM of both systems and find that the single-

scintillator system achieves a spatial resolution of 44 μm, compared with 53 μm for the double-

scintillator system.

Fig 5. Monte Carlo simulation results of a single radioactive point source. Simulated RLM

image of point source and corresponding 2D Gaussian fits with the (a, c) single and (b, d) double

scintillator, respectively. Scale bar 200 μm.

Discussion

200   The main finding of this study is that the double scintillator configuration increases the detection

201   sensitivity by a factor of two compared to the single scintillator case. While there are clear

202   advantages of using the double scintillator configuration over a single scintillator, a few

203   considerations must be made. The first is the depth of field of the microscope, which must be large

204   enough to accommodate both scintillators. The depth of field d_{LLM} of a microscope is,

$$d_{LLM} = \frac{n \cdot \lambda}{NA^2} + \frac{n \cdot b \cdot e}{M \cdot NA},$$ (3)

where $\lambda = 475$ nm is the emission peak of the scintillation light, n is the refractive index of the medium, NA is the numerical aperture of the objective lens, e is the pixel size, b is the binning number, and M is the effective magnification. As explained in a previous publication [15], the low-light microscope equipped with a 20× Nikon CFI Plan Apochromat λ has the following characteristics: $M_{effective} = 3.6$, $n = 1$, and $NA = 0.75$. Considering the parameters of the EMCCD camera (pixel size of 16 µm × 16 µm and binning of 4 × 4), d_{LLM} is estimated to be ~24.5 µm. Assuming that the distance between the bottom and top scintillator is ~ 10 µm [19], the depth of focus is sufficiently large to simultaneously focus on top and bottom edges of the scintillators. However, this is not true for all objective magnifications. For instance, a 40×/1.3 NA oil lens may not be compatible with the double-scintillator configuration. While $M_{effective}$ will increase to 7.2, d_{LLM} will be reduced by half or ~12.3 µm. This depth of field may not be enough to simultaneously capture the radioluminescence signals from both scintillators.

It should be pointed out that there is [18F]FDG efflux from the live cells, a well-known phenomenon of the glucose analog radiotracer [20,21]. Since the double scintillator image was captured 21 minutes after the single scintillator configuration, [18F]FDG efflux can be visualized by directly comparing the two RLM images. The pixel values of the double scintillator image were divided by two to match the single scintillator sensitivity, and the image was then subtracted by the single scintillator image. The resulting difference image is shown as Fig 6. In the figure, a significant number of radiolabeled cells with decreased activity are observed (shown in blue). On the other hand, there is a relative increase in radioactivity at the immediate vicinity of the cells. This is a clear indication that biological [18F]FDG efflux occurred between the two measurements,

which also explains why the linear regression coefficient was less than ×2 in Figure 3b ($y = 1.59$

$x + 87.86$).

Fig 6. Differential RLM image comparing double and single scintillator configurations. The

decay-corrected double scintillator image was divided in half, followed by subtracting the single

scintillator result. The black circles are ROIs around single cells.

Another confounding factor is that the double scintillator configuration may have increased signal

contamination due to stray gamma rays, which may cause scintillation far away from the cell of

origin. This may degrade the quality of radioluminescence images, particularly in areas of high

cell density. Further studies regarding the effect of radiation contamination from closely positioned

cells is required.

Conclusion

We demonstrated a method to double the sensitivity of radioluminescence microscopy or RLM.

While RLM provides quantitative information on the biological response of tissues and cells to

radiopharmaceuticals, the technique is limited by slow data acquisition time. To resolve this issue,

we sandwiched the radiolabeled cells between two parallel scintillators. Results show that the

sensitivity increased by a factor of two compared to the original single scintillator method, thus

effectively reducing the acquisition time by half. While simulation results show that the double

scintillator geometry has slightly lower spatial resolution, this effect was not noticeable during

experimental cell imaging. Moreover, we were able to visualize radiotracer efflux around the cells,

suggesting a new application that our technique can be used to simultaneously quantify radiotracer

uptake and efflux. In conclusion, the significant reduction in image acquisition time will increase

the experimental throughput and enabling pharmacokinetic analysis of short-lived clinical

radiotracers. In addition, the dual scintillator configuration may be a suitable candidate for

miniaturized radiobioassay devices, which is an emerging technology that enables evaluation of

biological responses to novel radiotracers with minimal volume requirements [22].

References

1.   Din AS, Kalapparambath TP, Tomas MB, Palestro CJ, Love C. Radionuclide Bone
      Imaging: An Illustrative Review. RadioGraphics. 2007; doi:10.1148/rg.232025103

2.   Bruehlmeier M, Roelcke U, Schubiger PA, Ametamey SM. Assessment of hypoxia and
      perfusion in human brain tumors using PET with 18F-fluoromisonidazole and 15O-H2O. J
      Nucl Med. 2004;

3.   Dobrucki LW, Sinusas AJ. PET and SPECT in cardiovascular molecular imaging. Nature
      Reviews Cardiology. 2010. doi:10.1038/nrcardio.2009.201

4.   Zhu L, Ploessl K, Kung HF. PET/SPECT imaging agents for neurodegenerative diseases.
      Chemical Society Reviews. 2014. doi:10.1039/c3cs60430f

5.   Gambhir SS. Molecular imaging of cancer with positron emission tomography. Nat Rev
      Cancer. 2002;2: 683–693.

267 6. Bar-Shalom R, Yefremov N, Frenkel A, Keidar Z, Israel O, Guralnik L, et al. Clinical

268   performance of PET/CT in evaluation of cancer: Additional value for diagnostic imaging

269   and patient management. J Nucl Med. 2003;

270 7. Kwee TC, Kwee RM. Combined FDG-PET/CT for the detection of unknown primary

271   tumors: Systematic review and meta-analysis. European Radiology. 2009.

272   doi:10.1007/s00330-008-1194-4

273 8. Young H, Baum R, Cremerius U, Herholz K, Hoekstra O, Lammertsma AA, et al.

274   Measurement of clinical and subclinical tumour response using [18F]- fluorodeoxyglucose

275   and positron emission tomography: Review and 1999 EORTC recommendations. Eur J

276   Cancer. 1999; doi:10.1016/S0959-8049(99)00229-4

277 9. Puncher MRB, Blower PJ. Radionuclide targeting and dosimetry at the microscopic level:

278   the role of microautoradiography. Eur J Nucl Med. 1994;21: 1347–1365.

279 10. Nielsen JL, Nielsen PH. Advances in microscopy: Microautoradiography of single cells.

280   Methods in Enzymology. 2005. doi:10.1016/S0076-6879(05)97014-6

281 11. Kiru L, Kim TJ, Shen B, Chin FT, Pratx G. Single-Cell Imaging Using

282   Radioluminescence Microscopy Reveals Unexpected Binding Target for [18F]HFB. Mol

283   Imaging Biol. 2017; doi:10.1007/s11307-017-1144-0

284 12. Sengupta D, Pratx G. Single-cell characterization of FLT uptake with radioluminescence

285   microscopy. J Nucl Med. 2016; jnumed. 115.167734.

286 13. Turkcan S, Kiru L, Naczynski DJ, Sasportas LS, Pratx G. Lactic acid accumulation in the

287   tumor microenvironment suppresses 18 F-FDG uptake. Cancer Res. 2019;

288   doi:10.1158/0008-5472.CAN-17-0492

289   14.   Kim TJ, Türkcan S, Pratx G. Modular low-light microscope for imaging cellular

290   bioluminescence and radioluminescence. Nat Protoc. 2017;12: 1055–1076.

291   15.   Kim TJ, Tuerkcan S, Ceballos A, Pratx G. Modular platform for low-light microscopy.

292   Biomed Opt Express. 2015;6: 4585–4598.

293   16.   Pratx G, Chen K, Sun C, Axente M, Sasportas L, Carpenter C, et al. High-Resolution

294   Radioluminescence Microscopy of 18F-FDG Uptake by Reconstructing the β-Ionization

295   Track. J Nucl Med. 2013;54: 1841–1846.

296   17.   Klein JS, Sun C, Pratx G. Radioluminescence in biomedicine: Physics, applications, and

297   models. Physics in Medicine and Biology. 2019. doi:10.1088/1361-6560/aaf4de

298   18.   Wang Q, Sengupta D, Kim TJ, Pratx G. In silico optimization of radioluminescence

299   microscopy. J Biophotonics. 2018; doi:10.1002/jbio.201700138

300   19.   Wang Q, Sengupta D, Kim TJ, Pratx G. Performance evaluation of 18 F

301   radioluminescence microscopy using computational simulation. Med Phys. 2017;

302   doi:10.1002/mp.12198

303   20.   Adonai N, Nguyen KN, Walsh J, Iyer M, Toyokuni T, Phelps ME, et al. Ex vivo cell

304   labeling with 64Cu-pyruvaldehyde-bis(N4-methylthiosemicarbazone) for imaging cell

305   trafficking in mice with positron-emission tomography. Proc Natl Acad Sci. 2002;

306   doi:10.1073/pnas.052709599

307   21.   Botti C, Negri DRM, Seregni E, Ramakrishna V, Arienti F, Maffioli L, et al. Comparison

308   of three different methods for radiolabelling human activated T lymphocytes. Eur J Nucl

309   Med. 1997; doi:10.1007/s002590050081

310 22.   Liu Z, Lan X. Microfluidic radiobioassays: a radiometric detection tool for understanding

311   cellular physiology and pharmacokinetics. Lab Chip. 2019; doi:10.1039/c9lc00159j

(a)

Beta particles

(b) Second scintillator

Scintillation flashes

Figure 1

Figure 2

Figure 4

(a) Single scintillator

(b) Double scint

(c) FWHM = 44 μm
$R^2 = 95\%$

(d) FWHM = 53 μ
$R^2 = 94\%$

Figure 5

Figure 6

Figure 3