A Low-Cost, Open Source, Self-Contained Bacterial EVolutionary biorEactor (EVE)

Vishhvaan Gopalakrishnan [1,2,*] Nikhil P. Krishnan [2] Erin McClure [3,+] Julia Pelesko [4]
Dena Crozier [1,2] Drew F.K. Williamson [3,†] Nathan Webster [3] Daniel Ecker [5] Daniel
Nichol [6] Soumyajit Mandal [7] Robert A. Bonomo [8] Jacob G, Scott [3,4,9,10,*]

1 Lerner College of Medicine, Cleveland Clinic, Cleveland, Ohio, United States of
America
2 Case Western Reserve University School of Medicine, Cleveland, Ohio, United States
of America
3 Department of Translational Hematology and Oncology Research, Cleveland Clinic,
Cleveland, Ohio, United States of America
4 Department of Physics, Case Western Reserve University, Cleveland, Ohio, United
States of America
5 Hawken School, Gates Mills, Ohio, United States of America
6 Centre for Evolution and Cancer, The Institute of Cancer Research, London, United
Kingdom
7 Integrated Circuits and Sensor Physics Lab, Case Western Reserve University School
of Engineering, Cleveland, Ohio, United States of America
8 Department of Medicine, Louis Stokes Cleveland Department of Veteran Affairs
Medical Center, Cleveland, Ohio, United States of America
9 Department of Radiation Oncology, Cleveland Clinic, Cleveland, Ohio, United States
of America
10 Center for Proteomics and Bioinformatics, Case Western Reserve University School
of Medicine, Cleveland, Ohio, United States of America

* vxg135@case.edu, scottj10@ccf.org
+ Current Address: University of South Florida Morsani School of Medicine, Tampa,
Florida, United States of America
†Current Address: Department of Pathology, Brigham and Women's Hospital, Boston,
Massachusetts, United States of America

Abstract

Recently, a concerted effort has been made to study the evolution of drug resistance in
organisms at increasingly smaller time scales and in a high-throughput manner. One
effective approach is through the use of customized bioreactors – devices that can
continuously culture bacteria and monitor this growth in real time. These devices can
be technically challenging and expensive to implement for scientists, let alone students
or teachers who seek an innovative and intuitive way of studying evolution. We seek to
provide a flexible and open source automated continuous culture device framework for
the academic setting to study biological concepts such as population dynamics and
evolution; a framework that is capable of replicating the functionality of many
prominent and expensive bioreactors in the market today. Within the educational
environment, our goal is to foster interaction and interest between the engineering and
biological fields by allowing teachers and students to build their own systems and design
experiments on the proposed open framework. We present a continuous culture device

designed for bacterial culture that is easily and inexpensively constructed, lends itself to evolution experiments, and can be used both in the academic and educational environments.

Author summary

The continuous monitoring of population growth in the presence of cytotoxic selective pressures can reveal new insights into resistance development and corresponding susceptibilities. Bioreactors have been proposed to accomplish this task yet are costly and without formal build instructions or software. In this article, we present a framework for a bioreactor, called the EVolutionary biorEactor (EVE), that will enable users to economically implement hardware and create circuits through diagrams. Hardware communicates with open-source software, written in Python, to create a flexible yet fully featured bioreactor that incorporates many modes of operation. A single EVE controls many Culture Units simultaneously, each with the capability of running its own experiment or a replicate of the same experiment to measure stochastic differences during evolution. While currently built for bacterial culture, this framework can be adapted in many ways from continuous mammalian cell culture to the measurement of multi-population dynamics in various environments. In the educational setting, this easy-to-implement framework will enable educators to use a hands-on approach to evolution in their lessons and extracurricular activities. For scientists, we propose this open framework as a tool that can be used and modified to investigate new areas in population dynamics and evolution.

Introduction

Continuous microbial culture is a powerful tool for observing and directing evolution in both research and industrial settings. Recently a shift towards open-source science has brought about a new wave of Do-It-Yourself, customized continuous culture devices to accommodate a wide variety of experimental designs across several research areas [1–6]. Within these devices, there are 3 main classes of continuous culture devices proposed in literature - each with a different objective during culture growth. The first is a chemostat – a bioreactor that continuously passages media into an active culture while simultaneously removing volume from the culture at an equal rate to maintain volume homeostasis [7]. This rate of inflow and outflow is called the dilution rate; a parameter that has to be calibrated to prevent culture overgrowth or extinction [8]. The second device is called a turbidostat, which is, a special case of an auxostat. An auxostat is a chemostat that measures a parameter of the culture and is capable of modifying the dilution rate through a feedback mechanism. Monitored parameters in an auxostat can include pH and concentration of a chemical among others. In turbidostats, the monitored parameter is the turbidity of the culture. Commercially available systems are generally unavailable, and professional systems can cost approximately $10,000 for academic units to $25,000 for the general public [9]. By increasing the dilution rate when the culture is growing rapidly, this continuous culture device can facilitate steady state growth at a constant cell density [10]. Morbidostats were recently introduced to study the evolution of drug resistance. Similar to a chemostat, the morbidostat maintains a fixed dilution rate. However, in contrast to a chemostat which keeps culture cell density relatively high as predicated by nutrient availability, the morbidostat keeps the cell density low by exposing the population to selective pressures such as cytotoxic drugs. Through the control algorithm that dictates morbidostat functionality, the drug concentration inside the culture is able to be modified over time to prevent extinction of

the culture [11].

Commercially available continuous culture machines are often expensive, specialized to a field of research, and resistant to modifications. A few such designs with detailed protocols have already been published in the literature; however, barriers in electrical, software, and biological expertise, or cost still leaves the assembly and operation of these machines out of reach of students or scientists in resource-poor settings [11–13]. In addition, increasing functionality of these machines while maintaining cost-effectiveness is an enduring challenge [14–17].

Here we present a protocol for the assembly and operation of a continuous culture device called the the EVolutionary biorEactor (EVE). It is a framework for a system which can support multiple individual bioreactors which we call culture units (CUs). Each of these CUs are capable of simultaneously running their own experiment with different optical density sampling rates, pump timings, optical density thresholds, antibiotics, and many other factors. While the EVE is able to replicate the morbidostat functionality defined in Toprak *et al.*, it is also able to replicate chemostat and turbidostat functionality [11]. To continuously monitor the cultures, the system proposed here utilizes low-cost consumer electronics, such as the Raspberry Pi microcomputer, as well as cheap integrated circuits that communicate with the Pi. We have designed a printed circuit board (PCB) that can be assembled for faster setup but also support a breadboard implementation of the circuit for educational applications and to save cost. This system makes use of an open source programming language (Python) to power our software on the Pi. We have programmed a setup utility to allow users to quickly setup the software in the device. After the setup process, users can utilize a web interface to easily customize, start, and monitor experiments. Since the source code is readily available on GitHub (https://github.com/vishhvaan/eve-pi), users can implement quick modifications to the flexible framework. Modifiable plans for 3D printable parts for vial housings, housing bases, and circuit board files are freely available on GitHub. The 3D parts can all be printed in 24 hours without the need for special materials and at low cost.

Our device also features the capability of conducting multiple experiments (up to 16) simultaneously with one Raspberry Pi (referred to as multiplexing). The web interface is built to allow users to configure multiple experiments with a single interface. The live monitoring system automatically detects the number of systems running and optimizes the memory usage of the device to allow for monitoring all the experiments simultaneously.

Ease of use and extensibility were highly valued and prioritized when designing the system. We hope that the cost-effectiveness of the system and its ability to be setup by those with novice engineering and biological knowledge can lend to the system's utility both in the regular classroom environment and outreach evolution teaching efforts such as EvolvingSTEM [18]. Keeping in mind that users can design their own experiments, we demonstrate a few simple experiments that showcase the capability of the system that are ideal for users becoming familiar with the system, as well as students learning about evolutionary biology in an interactive way.

Description of the EVE

The physical functionality of the EVE is centered around a culture vial and is inspired by prominent bioreactors featured in literature [11–13, 19, 20]. For optimal growth conditions of bacterial culture, the culture vial is stored in an incubator which is kept at 37°C for the duration of the experiment. Within the incubator, the vial is on top of a holder which incorporates a fan with magnets glued to its top surface. When the fan is powered, magnets rotate to spin a magnetic stir-bar at the bottom of the culture vial.

The stir bar has two functions – to facilitate diffusion of oxygen into the culture media and to keep the culture's turbidity uniform across the culture vial. The vial holder also houses two diodes (one LED and one photodiode). Throughout the experiment, the LED shines a light through the culture. Depending on the growth and corresponding turbidity of the culture, light is variably scattered through the culture, changing the amount of light that reaches the photodiode (PD). The LED and PD are connected to a circuit board (either a breadboard or PCB) that powers the devices and converts the analog information from the PD into digital information. The circuit board is connected to the Raspberry Pi with a ribbon cable, enabling the two boards to interact digitally. The culture vial is also connected to three liquid pumps with three plastic tubes. Each pump in turn connects to either a fresh media reservoir, a media reservoir containing cytotoxic drugs (antibiotics in this paper), or a waste reservoir. When powered by the circuit board, a specific pump activates which pumps either media or media mixed with drug into the vial. The waste pump also activates whenever the media pumps activate to prevent culture overflow. The waste pump can also provide the opportunity for easy temporal sampling. The software on the Raspberry Pi is responsible for triggering these pumps at specific points during the culture to achieve the (user defined) goal of the experiment. A schematic of the setup is shown in Fig 1.

Before the experiment, the culture vials and reservoirs are made sterile with autoclave or other sterilization techniques described in the discussion. During the experiment, sterile air filters prevent the culture media in the vial or fresh media in the reservoirs from getting contaminated. The silicone tubes that is used to transport the media to from the reservoir to the culture also needs to be sterilized prior to the experiment. To sterilize these tubes, solutions of 10% bleach, 70% ethanol, and sterile water in reservoirs are attached to the pumps and are used to wash the tubes. These sterilization steps are easily done with the EVE Web Interface.

Of note, absorbance or the optical density of the culture is not directly reported to the Pi. Rather, the voltage of a resistor connected to the PD is measured. The voltage of this resistor increases as more light strikes the PD and therefore increases as more light is scattered - a surrogate of the absorbance of the culture. Since the voltage of the resistor is related to the optical density of the solution, we refer to the voltage value as the optical density of the culture in this article. The exact relationship of the voltage of the resistor and the optical density of the culture (A_{600}) can be found with the use of a calibration curve. Steps to generate calibration curves will be included in the discussion. This voltage value should not be interpreted as absorbance since they are numerically different. In many scenarios, the voltage value provides a clear picture of the state of the culture. However, if a true absorbance is needed, a simple calibration curve correlating the voltage value with an optical density can be performed. A sample calibration curve is included in the supplemental figures. Steps to generate the calibration curve with the software is included in the software section.

Fig 1. Hardware schematic of the EVE system. The culture vial is located inside the incubator. As the culture grows, an LED shines a light through the culture. Light that is reflected by the culture is received by the photodiode. A circuit board interprets and conveys photodiode readings to the Pi which runs control software. Based on user-configured directions, and optical density readings, the Pi uses the circuit board to power pumps connected to the vial. This strategy enables precise and dynamic control of the culture.

Assembly

Assembly of the EVE can be separated into hardware and software tasks. We recommend beginning with hardware assembly as it requires more time to complete. Once the hardware is assembled, software can be installed on the Raspberry Pi and then seamlessly incorporated into the hardware as the last step of the assembly.

Hardware

A list of required hardware components can be found in the supplemental information. The vial housing complex, fan with attached magnets, PD, and LED for each CU are housed in the incubator. The positive and negative (ground) wires that supply power to the LED, PD, and fan connect to the circuit board outside the incubator and extend into the incubator.

The 3D printable designs (.stl and .sldpart files) of the vial holder, base, and diode caps can be found on the device's GitHub repository. We used an Ultimaker S5 to print the components used in the experiments detailed below - though any 3D printer could be used. The vial holder is built to contain the experimental vial as well as the LED and PD. Figure 2h shows one of the vial holders holding a culture vial. The diodes fit in openings on the side of the vial holder, which was adapted from Wong *et al.* [13]. The base of the vial holder (a separate component) houses the fan affixed with magnets and the vial holder. Figure 2k shows 3 fully assembled culture units inside the incubator with LED and PD attached. Figure 2l is a close up of one of these culture units. The vial holder base also raises the fan away from the metallic base of the incubator so there is no interaction between the base of the incubator and the solenoids of the fan.

Fig 2. Hardware components of the EVE. (a) Square fan with magnets attached to a spacer. (b) Three pumps for three functions (pumping in media, pumping in the solution of media and drug, and pumping out waste). All three are connected to the circuit board. (c) Silicone tubes as they are connected to the pumps along with terminal leur connections. (d) Pump in the 3D printed pump holder. (e) Three culture vials with 12mL of media and stirbars. The longest PEEK tube touches the culture. The other PEEK tubes meant to introduce media solutions to the vial do not touch the surface of the media. (f) Media reservoir with three small silicone tubes to transport media to the vials. (g) Terminal connections of the media reservoir. These connect to the terminal connections of the pump shown in (c). (h) 3D printed vial holder with vial inside. Diodes are held in place with 3D printed caps. (i) Top of the reservoir displaying the small silicone tubes attached with the silicone sealant (j) 70% Ethanol reservoir used for sterilization of the small silicone tubes prior to the experiment. Terminal connections and the attached air filter of the reservoir is displayed. (k) An incubator with three Culture Units. (l) The vial holder and fan with magnets fit into the base. Diodes are attached to the vial holder. Together, these components comprise a Culture Unit.

If an incubator other than the one listed in the parts list is used for our experimental setup is to be used, the base file can be modified to ensure a fit with the incubator using any CAD software of choice Supplemental Information. Finally, the caps fit on the diode openings to push the diodes into the openings and ensure a secure fit. These caps minimize diode positioning differences between each CU, therefore improve replicability of experiments. Both figure 2h and figure 2l show the holders with caps for the LED and PD attached.

Electrical connections between components can be made with solder or with heat shrink tubing. The components needed to solder (soldering iron, solder, etc.) are not included in the parts list can be easily found online or at a hardware storeSupplemental

Information. Heat shrink tubing requires a heat gun to shrink the tubing. While the experiments below utilized soldered components, heat shrink tubing and a heat gun were used to test connections made with this method. Hardware assembly for a single CU system are as follows. These steps are influenced in many aspects by the work done by Toprak *et al.* [11]. For an experiment done in triplicate, repeat the steps detailed three times.

Pump Assembly

1. Label a set of three pumps with "Media", "Drugs", and "Waste" (one label per pump)

2. Cut six pieces of 5 cm large silicone tubing.

3. Cut small silicone tubing into six 45 cm long pieces. Ensure that the length of the small tubing is able to reach the culture vial or the media reservoirs from the pumps.

4. Attach two pieces of large silicone tubing to the two tubes protruding from each pump.

5. Fit a male leur lock on the free end of the large silicon tubes.

6. Screw a female leur onto each of the male leur locks

7. Fit a small silicone tubing onto the female leur locks.

8. With the pumps facing up with the large silicone tubes at the top, label the large silicone tube on the left as the input.

9. Fit female leur locks at the end of the small silicone tubes attached to the "input" large silicone tubes.

10. Obtain three 60 cm red hook-up wire and three 60 cm black hook-up wire.

11. Curl one end of a red hook up wire around the positive terminal of a pump. Repeat for all three pumps.

12. Curl one end of a black hook up wire around the negative terminal of a pump. Repeat for all three pumps.

13. Use solder or heat shrink tubing to securely attach the wires to the pump.

14. Label the end of the wires with the pumps function (e.g. media, drug, or waste).

Figure 2b shows the wires solders to the pumps. Figure 2c shows the large and small silicone tubes attached to each other with leur locks.

Diode Assembly

1. Obtain a LED and photodiode.

2. Obtain two 60 cm red hook-up wires and two 60 cm black hook-up wires.

3. To each diode, curl the red wire around the positive terminal and curl the black wire around the negative terminal.

4. Solder or use heat-shrink tubing to securely fasten the four wires to the terminal it is curled around.

5. Label the end of the wires with the diode function (e.g. LED, PD).

<u>Fan Assembly</u>   188

1. Place one spots glue at the center of the axle of the fan.   189

2. Place the 3D printed fan spacer on the superglue and apply pressure so that the   190
   spacer is flat.   191

3. Place a spots of superglue on each of the spacers.   192

4. Place two magnets on the superglue of each spacer and apply pressure so that the   193
   magnets are flat.   194

5. Use solder or heat shrink tubing to securely attach the long black and red wires to   195
   the short black and red wires protruding from the fan, respectively.   196

6. Label the end of the wires with "fan".   197

7. Allow the fan to dry for three hours before use.   198

The final product of the assembly is shown in figure 2a.   199

<u>Combine the prepared fan with the 3D printed components.</u>   200

1. 3D print the vial holder, the base, and the caps (Fig 2h).   201

   - 3d printer settings: 202
     (a) Material: Black PLA   203
     (b) Layer height: 0.1   204
     (c) Print speed: 50 mm/s   205
     (d) Build plate temperature: 60°C   206
     (e) Printing Temperature: 195°C   207
     (f) 30% infill   208
     (g) No support with components oriented upwards   209
     (h) No adhesion   210

2. Place the base inside the incubator.   211

3. Place the fan inside the incubator and route the cables through the hole at the   212
   top of the incubator.   213

4. Fit the fan on the base by sliding the fan through the four columns on the edges   214
   of the base.   215

5. Fit the vial holder on the base so that it rests above the fan on the columns.   216

6. Place the LED and PD in the incubator and route the wires through the hole at   217
   the top of the incubator.   218

7. Fit the LED into the hole in the vial holder marked LED.   219

8. Fit the photodiode into the hole in the vial holder marked PD.   220

9. Ensure that the diode terminals are bent and protrude from the notches on the   221
   diode holes.   222

10. Cover the diode holes with the 3D printed caps. Ensure a secure fit by pressing   223
    down the cap firmly (Fig 2h).   224

The vials that are used during experiments need to be modified so that they can accept media and the drug solution and eject waste during the experiment. Steps to modify the vials for experiments are as follows:

Vial Assembly

1. Cut a 7.5 cm piece of the small silicone tubing.

2. Cut four pieces of PEEK tubing, each 7.5 cm long.

3. Remove the cap of the vial and penetrate the cap with the PEEK tubing so that the insertion points are equidistant from each other.

4. Ensure that the PEEK tubes are at the appropriate height:

    (a) Two of the PEEK tubes should extend 1.25 cm below the cap and extend 3.8 cm above the cap.

    (b) One PEEK tube should extend 2.5 cm below the cap and 6.5 cm above the cap.

    (c) One PEEK tube should extend a 1 cm below the cap and 2.5 cm above the cap.

5. Apply the silicone sealant to both sides of the cap and around the protruding PEEK tubes to make the cap water and air tight.

6. Cut a 10 cm long piece of small silicone tubing.

7. Fit one end of the tubing to PEEK tube that extends 2.5 cm above the cap.

8. Fit the other end of the tubing to a female leur.

9. Push the female leur over the tapered end of a syringe filter.

There are four tubes connected to each vial. Each tube is meant for a specific task and are distinguished from each other by their heights as seen in figure 2e which shows 3 modified vials. The height differences of the PEEK tubes are important. As seen in the figure, the longest PEEK tubes extend into the surface of media so that it is able to remove waste from the culture. The shorter PEEK tubes are near the cap and are far from the surface of the culture to prevent contamination. The PEEK tubes connected to the air filter is the shortest so that it does not interact with the solutions pumped into the vial. In the figure, this tube is connected to the small silicone tubes.

Reservoirs of media and other liquids (for sanitation of the silicone tubes) also need to be made. These need to interface with the tubes from the pumps. Perform the following steps to create one reservoir:

Reservoir Assembly

1. Obtain a 500 mL beaker with cap

2. Remove the cap from the beaker and drill 4 holes with a diameter of 3 mm in the cap.

3. Cut three pieces of 45 cm small silicone tube.

4. Cut one piece of 15 cm small silicone tube.

5. Insert one 45 cm silicone tube into a hole in the cap so that tube can touch the bottom of the beaker when the cap is placed on top of the beaker.

6. Apply silicone sealant to both sides of the cap to secure the silicone tube (Figure 2i).

7. Repeat the previous two steps for each of the 45 cm silicone tubes.

8. Insert the 15 cm silicone tube into the last hole so that only 2.5 cm is below the cap.

9. Apply silicone sealant to both sides of the cap to secure the 15 cm tube.

10. Insert male leur locks at the end of the three 45 cm silicone tubes above the cap.

11. Insert a female leur lock at the end of the 15 cm silicone tube above the cap.

12. Push the female leur over the tapered end of a syringe filter .

Figure 2f shows the completed media reservoir with the small silicone tubes submerged in the media. Figure 2j shows a similar reservoir filled with 70% ethanol. This reservoir is used to sterilize the silicone tubes that interface with the culture vial. The silicone used to secure the tubes is also seen on the top especially in Figure 2i. Figure 2g shows the male leur locks on the end of the silicone tubes. The sterile air filter is also shown at the end of the 15 cm silicone tube.

Circuit Board

Schematic and PCB files can be found on the project's GitHub (https://github.com/vishhvaan/eve-pi). Surface-mounted components including integrated circuits, resistors, diodes, and others can be found in the electronics section of the supplementary parts list Supplemental Information. PCBs can be printed from the PCB file and the necessary circuit components soldered to board to create a functional circuit board. Four simultaneous evolution experiments can be conducted from one circuit board connected to one Raspberry Pi.

Steps to setup the circuit board are as follows:

1. Connect the circuit board to the 12-volt power adapter. Ensure that the LED on the board is on after it has been plugged in.

2. Connect the circuit board to the Raspberry Pi with the ribbon cable.

3. Ensure that the top of the Raspberry Pi's GPIO ('General purpose input/output pins' located in an array-like configuration on one side of the Pi) is connected to the top of the GPIO port on the circuit board.

4. Collect the labeled wires of the diodes, pumps, and fan.

5. Push the wire into its appropriate position on the labeled Phoenix block terminal connector (red wires to the positive terminal and black wires in the negative terminal)

If assembling the circuit board is not desired, a breadboard can also be used. The PCB schematics can be utilized to build a circuit on the breadboard. When purchasing resistors, diodes, and integrated circuits to build a breadboard circuit, through-hole components should be bought instead of their surface-mounted counterparts. A Raspberry Pi GPIO breakout board included in the parts list as an optional component recommended when building the circuit with a breadboard Supplemental Information. In tests of building the machine, given the necessary components, an undergraduate engineering student (NW) was able to assemble the circuit in the course of five hours.

Software

The software in this device is built on the Python programming language and can be run from a single Raspberry Pi. Users interact with a simple web interface that is capable of controlling the hardware, running experiments, and editing configuration files, as shown in Fig 3. The web interface used in this project is forked from another open source script server (https://github.com/bugy/script-server) and modified to integrate with EVE control applications. This web interface can be utilized by multiple users connected to the local network, allowing for remote monitoring and control of experiments. This network capability of the device also allows it to save data to network locations, preventing lack of local space on the Pi from limiting experiment duration. The software allows data to be saved to a USB attached to the Pi or to network locations mounted to the Pi's file system.

The Pi controls chips (the general-purpose input/output extenders, and analog-to-digital converters) digitally through the use of its I2C interface, a method of serial communication. These chips have associated hardware addresses that can be assigned by connecting address pins on the chips to a voltage source or to the ground. When creating the circuit on a breadboard, these hardware addresses can be set by following the circuit schematic or by assigning addresses of choice. Proper control of the hardware relies not only on knowledge of the hardware address of the chips but also the pins of the general-purpose input/output extenders (GPIO) to which the pumps, LEDs, fans were attached to and the pins of the analog-to-digital converters (ADC) to which the photodiodes were attached to.

Fig 3. Software schematic of the EVE system. The device's software resides on the Raspberry Pi and center around a Python-based web-server which serves as the GUI for experimental design and control. This implementation offers built-in remote monitoring and multi-user control of the device. More EVE-centric Python-based applications are also installed Pi, accessible via the web interface. These applications are able to share data, and perform experiments based user defined configurations. The experiment controller application directly interfaces with the circuit board. All of these applications are open source and are built with extensibility in mind. Users have access to all the underlying programs and code. We encourage them to fork our GitHub repository or download programs to their devices and make modifications as they see fit. See Fig 1 for the hardware schematic of the EVE system.

Interaction and direct modification of the underlying Raspbian operating system or the core programs is unnecessary for most users who are unfamiliar with Linux or Python. However, since all the open-source software is contained in a clone of the GitHub repository, Python-savvy users can still be able to access the location of the files (/eve) and modify the programs to their specific needs.

When users run an experiment or control the hardware through the web interface, the executed programs rely on a configuration file that provides information about the hardware configuration. These parameters include the hardware address of the chips, the pins of the GPIOs pumps, LED, photodiode, and fan, duration of the experiment, time between pumps, thresholds for introduction of drug and media, and many others. All of these parameters can be found in the supplemental document titled "Configuration File Parameter Definitions". This document can be found in the GitHub within the "Start Building" folder and can be directly downloaded with the link in the supplemental documents. This configuration file can be edited through a built-in configuration file editor that can be accessed through the web interface.

Since the circuit board has a pre-wired set of hardware addresses for the four culture units, a pre-populated configuration file for the circuit board is included in the software.

For those who wish to build a breadboard version of the circuit, a sample configuration   346
file is also included. If the breadboard circuit was built exactly as described within the   347
circuit schematic, the hardware addresses included in the configuration file will be   348
sufficient to communicate with the chips, pumps, fans, and LEDs. If the circuit was   349
built per user specifications, they can replace keys with their own parameters depending   350
on the setup of their breadboard.   351

The software offers experiment monitoring in two ways: through Slack and through   352
a web server that is deployed automatically when the experiment begins. The web   353
server, based on Dash by Plotly, allows any user on the local network access to real-time   354
progress of the experiment through a web browser [21]. With both Slack and this web   355
server, users can monitor: 356

1. Voltage of a resistor connected to the photodiode - a surrogate of the optical   357
   density of the solution inside culture vial(s). This voltage value can optionally be   358
   correlated to the OD600 of the solution through a calibration curve.   359

2. Time and duration of specific pump activation.   360

3. Calculated drug concentration inside the culture vial(s). These calculations do not   361
   account for drug consumed inside the vial. It is based on the concentration of the   362
   drug solution reservoir and the duration of pump activity.   363

4. Thresholds of optical density before media or drug pumps are activated.   364

5. Temperature of the incubator.   365

6. Concurrent threads of activity in the program during the experiment.   366

For academic labs that use Slack as a means of communication, the EVE's Slack   367
integration may be a reliable way to monitor experiments. The program relies on the   368
Slack API to creates a bots for each CU. These bots communicate with specific Slack   369
channels and regularly update the channel with live updates from the experiment.   370
Controls for the Slack integration can be modified within the configuration file. Steps to   371
access the web server are detailed below.   372

Installation   373

The software was tested on the Raspbian Buster operating system installed on a   374
Raspberry Pi model 3B. The software instructions are written for this combination of   375
software. However, these instructions should also work on any Debian-based operating   376
system installed on versions of the Raspberry Pi 3 or 4. To maintain ease of use, the   377
software installer can be run on an empty installation of Raspbian which will download   378
and setup all software requirements. It will also ensure that the web interface launches   379
every time the Raspberry Pi is restarted. Alternatively, a Docker image compiled on the   380
ARM architecture is available on the GitHub repo. This image and associated   381
Dockerfile can streamline the installation process and also allow for deployment across   382
many ARM devices. The instructions below detail the setup process with the installer   383
from the GitHub. Instructions for the installation of the Docker image can be found on   384
the image's Docker Hub.   385

Steps to install the software on the Raspberry Pi are as follows: 386

1. Download and flash the Raspbian Buster OS on an SD card.   387

2. Insert the SD Card into the Pi and connect the Pi to a display, keyboard, and the   388
   local area network.   389

3. Note the IP address of the device with the command: `ifconfig`. ₃₉₀

4. Enter super user mode with the command: `sudo su`. ₃₉₁

5. Run the setup script by entering the command: `bash <(curl -s https:` ₃₉₂
   `//raw.githubusercontent.com/vishhvaan/eve-pi/master/st_eve.sh)` ₃₉₃

6. Follow the on-screen instructions to install the software. ₃₉₄

7. If the Pi has a browser, navigate to the web interface by accessing: ₃₉₅
   `http://localhost`. ₃₉₆

8. On a browser on the network, navigate to the web interface by accessing the IP ₃₉₇
   address or hostname of the device (e.g. `http://eve.local_domain.net`). ₃₉₈

Usage ₃₉₉

The EVE web interface has several tabs that perform different functions. The functions ₄₀₀
of the tabs are as follows: ₄₀₁

1. Pump control – This tab allows users to turn on specific pumps of specific CUs ₄₀₂
   connected to the circuit board. The 'CU #' field allows users can select multiple ₄₀₃
   devices simultaneously by entering a comma separated list (i.e. 1,2,3) or using the ₄₀₄
   text "all" to select all the enabled CUs. Users are able to input time and the ₄₀₅
   pump type (e.g.: media and drugs alone). The program relies on the configuration ₄₀₆
   file to find the hardware addresses of the CUs. If the CU is disabled in the ₄₀₇
   configuration file, the program will not activate the pumps for the system. This is ₄₀₈
   a safety mechanism that prevents activation of mis-configured pumps. ₄₀₉

2. One of the main functions of this program, in the context of the EVE, is ₄₁₀
   sterilization of the silicone tubes prior to the experiment. When reservoirs of 10% ₄₁₁
   bleach and 70% ethanol are connected to the pumps, this program can be run to ₄₁₂
   pump fluid through the tubes, cleaning it prior to the exposure to media during ₄₁₃
   the experiment. Instructions for these steps are detailed below. ₄₁₄

3. Fan Control – This tab allows users to toggle power to specific or multiple fans. ₄₁₅

4. Run Experiment – This tab runs the experiment on the EVE based on the ₄₁₆
   settings in the configuration file. ₄₁₇

5. Configuration Editor – This tab runs the editor for the configuration file. Users ₄₁₈
   can view the configuration file and edit specific keys. This program offers an ₄₁₉
   alternative for users who are not familiar with the Linux terminal and don't want ₄₂₀
   to use SSH to edit the configuration file. ₄₂₁

6. PD Test – This tab tests the functionality of the LED and photodiodes of specific ₄₂₂
   or multiple CUs. Users have the ability to modify the time between queries and ₄₂₃
   number of times each PD is queried. This program can be used to test the circuit ₄₂₄
   and debug the setup. ₄₂₅

7. Pi Power Control – This tab allows users to restart or power off the Pi. ₄₂₆

8. Pumps and LED Off – This tab turns off all the pumps and LED in all the CUs ₄₂₇
   connected to the EVE. This program provides a method to stop pumps or LED ₄₂₈
   activity before the original program finishes running. In the event of an error, this ₄₂₉
   program can also be used to an efficient way to halt the pumps the LED. ₄₃₀

9. Resource Viewer – This tab displays system statistics for the Pi, including real-time CPU and memory usage.

After the experiment is begun, the live plotting interface is automatically deployed after the first set of data is written to the file (approximately 1 minute after the experiment starts). To visit the live plotting interface, users edit the web address based on the address of the web interface by adding the port 8050. For instance, if the users accessed http://localhost for the web interface, the live plotting interface can be found at `http://localhost:8050`. If the users accessed the device by the hostname or the IP address, the same rule applies (e.g.: `http://eve.local_domain.net:8050` if the web interface is found at `http://eve.local_domain.net`)

The live plotting interface, shown in Fig 4, can be navigated with the tabs at the top of the website. Every time the web page is refreshed, the Dash server processes the data from the output file and displays the most recent data. This *ad hoc* method of processing data for graphing proves to be less of a computational burden on the Pi compared to a strategy where data is processed on a regular basis. The first tab provides an overview of all the data while the other tabs (Pumps, Drug Concentration, Threads, and others) provide focused comparisons of specific data types. Users can select specific Culture Units to compare in the second row of tabs.

Fig 4. Web interface of the EVE running on the Pi. A variety of tabs are available to users to configure and run experiments, selectively turn on the pumps (useful for sterilization prior to the experiment), and control the Pi directly. This server is run from the Pi itself and is automatically installed after the user runs the setup script. This interface is accessible from any device connected to the same network the EVE resides on. It was built to allow ease of use and particularly, to provide non-programmers with a simple way of accessing and using the device. In this screen shot, specific pumps in multiple CUs have been turned on for twenty seconds.

Calibration

Calibration can be performed to map voltage of the photodiode (a surrogate of the optical density of the culture) to the true absorbance of the culture at 600nm. This is an optional step as this mapping is linear and therefore does not change the fundamental shape of the growth curve. Calibration data can be easily recorded through the calibration program of the the web interface. The steps to obtain calibration data are as follows:

1. Transfer a vial of culture saturated with bacteria to a Culture Unit.

2. Use the calibration program in the web interface to record 30 measurements of the voltage reported by the ADC. Data is recorded with in tandem with a user specifed unique identifier and Culture Unit number.

3. Repeat steps 1 2 until the voltage reported from all three Culture Units have been measured.

4. Transfer 100μL of the concentrated solution to a 96 well plate in triplicate.

5. Label the plate so that the three samples are tagged with the same unique identifier used in the data.

6. Remove 1mL of the concentrated culture solution and replace that volume with fresh media.

7. Repeat steps 1 - 6 until the voltage reported by the cultures units do not change between dilutions.

8. Measure the absorbance of the samples on the 96 well plate with a spectrophotometer.

9. Collect data from the EVE and the spectrophotometer.

10. Match data with the same unique identifiers.

11. Create columns for each Culture Unit with with averages of the voltage measurements at each unique identifier.

12. Create a column for the true optical density of each unique identifier.

13. Perform linear regression on each voltage column with the true optical density column to obtain the calibration line for each Culture Unit.

The spectrophotometer that we used to measure the true optical density of the samples was the Biotek Synergy H1 Hybrid Reader. Fig 5 demonstrates an example of a calibration curve that was generated based on EVE. Based on the regression values of the models generated, the relationship between the voltage recorded by EVE and the true optical density of the culture at 600nm is linear. The calibration lines are similar across CUs indicating similar voltage response of each Culture Unit. This implies that any voltage differences measured between CUs is a result of differences in the behaviour of the population.

Fig 5. Sample Calibration Curve of the EVE. (a) Linear regression is performed in each Culture Unit (CU) between measurements of the raw voltage to the true absorbance as determined by a spectrophotometer. The resultant relationship between the true optical density and raw voltage can be mapped to the voltage readings of growth and selection experiments to attain a true optical density measurement. (b) The relationship between voltage and absorbance at 600nm is similar across CUs indicating that any voltage differences measured between CUs is a result of differences in the behaviour of the population.

Experimental Use Cases

Prominent bioreactors featured in literature are very versatile and can be used to perform a variety of experiments. The EVE aims to follow suit by being capable of performing a variety of experiments. In this section, two basic biological experiments that can be carried out by the EVE are described.

Bacterial Growth

In this experiment, the real time bacterial growth of common K12 *Escerichia coli* can be followed by tracking optical density. This experiment can be especially enticing for students or educators since it requires less hardware to be setup and can provide a cross-disciplinary approach to biology. For this experiment the following list of parts are needed for single vial:

1. An incubator that can reach 37$^{\text{o}}$C.

2. 3D printed vial holder, base, and caps.

3. 12V computer fan and magnets.

4. LED and photodiode as setup above.

5. A breadboard circuit or a circuit board.

6. Soldering or heat-shrink technology.

7. K12 E. Coli.

8. Broth to culture bacteria

- If an autoclave is accessible:
  (a) Bacterial growth media powder. Options include:
      i. Lysogeny Broth (LB) powder. Mueller Hinton (MH) Broth powder.
  (b) Distilled Water.
- If an autoclave is not accessible:
  (a) Sterile pre-prepared broth

9. 70% ethanol

10. Bunsen burner

11. Culture Vials modified to only have 1 PEEK tube for air flow.

12. Magnetic stir-bar.

For a triplicate experiment, three times the 3D printed parts, diodes, fans, and magnets are needed. Note that no pumps or tubes are needed for this experiment. Once the system has been setup, as described above, users setup the vial to support bacterial growth. To do this, Sigma Aldrich MH Broth is first made by dissolving 21 g of powder in 1 liter of deionized or distilled water. For other brands of MH broth, brand-specific preparation instructions should be followed. It is always recommended to autoclave media after it is created to prevent contamination of the media itself. However, if an autoclave is not available, there are several alternatives proposed by the WHO in their "Methods of Sterilization" [22]. Sterilize the bench area and gloves with 70% ethanol. Work near a lit Bunsen burner to maintain a sterile field. Then, 12 mL of the broth is transferred to the culture vial with stir-bar inside. The vial is then inoculated with 0.5 g of the K12 *E. coli* and is placed in the holder. The experiment can then be started from the web interface.

To build a system that can multiplex multiple vials simultaneously, multiple 3D printed parts, computer fans, diodes, and vials are needed. The rest of the parts do not have to be multiplied.

As the experiment progresses, users can monitor the bacteria's growth via the live web interface or through the included Slack integration. At the end of the experiment, users can parse the recorded data and visualize the data using their own methods. Graphs of a trial we performed are included in Fig 6.

Evolution of Antibiotic Resistance

The second experiment demonstrates the EVE's full functionality by applying a selective pressure to bacteria as it grows. An antibiotic, in this case ampicillin, is introduced to the culture vial as the bacteria grows by passaging the drug mixed with media. Additionally, the culture is kept in the log-phase of growth during the duration of the experiment by passing fresh media into the vial while removing waste

Fig 6. Sample graphs for a triplicate bacterial growth experiment. Triplicate bacterial growth experiments sharing the same axis. (a) Technical replicates (individual CUs) show different voltage readings over time despite being connected to the same circuit board and made identically. This variation is due to subtle differences in the positioning of the diodes in the holder. (b) When the readings are scaled by the maximum voltage reading, all curves overlap, exemplifying the triplicate nature of the experiment. Diode differences are constant over time, therefore raw voltage values can be correlated with OD600 values as seen in Fig 5. For simple experiments, this step is unnecessary. In general, this growth experiment can be readily performed by scientists to test their system or by students and teachers in scientific classes.

simultaneously. This experiment builds upon the setup of the simpler bacterial growth experiments. Pumps and silicone tubing would need to be setup for this phase of the experiment. In addition to the parts described in the bacterial growth experiment the following parts would be needed for a single vial antibiotic evolution experiment:

1. 3 pumps setup as described in the pump setup section.

2. Vials modified as described in the Assembly section.

3. Reservoir of MH broth (prepared as explained above).

4. Reservoir of MH broth with added antibiotics.

5. Reservoir of 10% bleach.

6. Reservoir of 70% ethanol.

7. Reservoir of distilled water.

8. 2L Erlenmeyer Flask

9. Bunsen Burner

To create a reservoir of MH broth with antibiotics, antibiotics are simply added to a reservoir of MH broth. The concentration of antibiotics in the reservoir is important. Toprak *et al.* found that 10 times the MIC inhibited growth sufficiently [11]. Since the duration of the activation of the drug pump is modifiable in the software, this is also a factor when calculating the dose of antibiotics in the reservoir. The shorter the pump time, higher concentrations of antibiotics are needed to produce the same inhibitory effect on the culture. Generally, with the default pump time of 0.75 seconds, we found that 10 times the MIC of ampicillin (much like the findings of the Toprak paper) inhibited growth in our cultures. When approximately 1 mL of the total 12 mL culture volume is substituted with drug solution at 10 times the MIC, assuming that the culture hasn't been exposed to drug prior, the total concentration of the drug inside the vial becomes 83.3% of the MIC. As more drug solution is substituted with the culture volume, the concentration of drug (and percentage of MIC) will increase. Conversely, if media is substituted with the culture volume, the percentage of MIC will decrease. Users can optimize these parameters as needed depending on the antibiotics they use.

Again, for a triplicate experiment, three times the pumps and vials are needed and the steps below also need to be repeated thrice. The same reservoirs can be used for multiple replicates. When running an experiment on the EVE, especially one that will last several days, it is important to maintain a sterile environment – especially the silicone tubes, which are exposed to the media. During long-term (on the order of weeks) experiments, some groups have been known to clean the tubes between pump activations or pause the software. To sterilize the tubing:

1. Spray all luer connectors with 70% ethanol. ₅₇₆

2. Attach the three outputs of the 10% bleach reservoir to the inputs of the media, bleach, and waste pumps. ₅₇₇ ₅₇₈

3. Run all three pumps for 60 seconds. ₅₇₉

4. Disconnect the 10% bleach reservoir from all three pumps. ₅₈₀

5. Attach the three outputs of the 70% ethanol reservoir to the inputs of the media, bleach, and waste pumps. ₅₈₁ ₅₈₂

6. Run all three pumps for 60 seconds. ₅₈₃

7. Disconnect the 70% ethanol reservoir from all three pumps. ₅₈₄

8. Attach the three outputs of the distilled water reservoir to the inputs of the media, bleach, and waste pumps. ₅₈₅ ₅₈₆

9. Run all three pumps for 60 seconds. ₅₈₇

10. Disconnect the distilled reservoir from all three pumps. ₅₈₈

11. Attach one output of the media reservoir to the input of the media pump. ₅₈₉

12. Run the media pump for 20 seconds to preload media in the tube. ₅₉₀

13. Attach one output of the media reservoir to the input of the drug pump. ₅₉₁

14. Run the drug pump for 20 seconds to preload drug solution in the tube. ₅₉₂

After this step, the vial is ready to be placed in the holder and attached to the tubes for the media, drug solution, and waste: ₅₉₃ ₅₉₄

1. Obtain a 2-liter Erlenmeyer flask and label it as waste. ₅₉₅

2. In a sterile field, transfer 12 mL of MH broth to the culture vial with a stir-bar inside. ₅₉₆ ₅₉₇

3. Inoculate the vial with 0.5 g of the K12 E. coli. ₅₉₈

4. Close the vial with the cap. ₅₉₉

5. Place the vial on the holder. ₆₀₀

6. Attach one output of the media reservoir to the input of the media pump. ₆₀₁

7. Attach one output of the media reservoir to the input of the drug pump. ₆₀₂

8. Insert the input tube of the waste pump into the incubator through the opening at the top. ₆₀₃ ₆₀₄

9. Connect the tube to the longest PEEK tube of the vial. ₆₀₅

10. Insert the output of the media pump and the output of the drug pump into the incubator through the opening at the top. ₆₀₆ ₆₀₇

11. Connect these two tubes to the remaining two short PEEK tubes of the vial. ₆₀₈

After these steps are completed, the experiment is ready to begin and can be started ₆₀₉ from the web interface. To keep track of the concentration inside the vial, the ₆₁₀ concentration of the drug inside the reservoir is modified in the configuration file before ₆₁₁ the experiment starts. Based on the amount of time each pump is run (another variable ₆₁₂ that can be modified in the configuration), the total amount of drug in the vial can be ₆₁₃ tracked. Sample graphs of a singleton experiment are shown in Fig 6 and a triplicate ₆₁₄ shown in Fig 7. This experiment demonstrates the selection of an *E. coli* culture with ₆₁₅ ampicillin in a CU. In figure 6, there are small dilutional effects as a result of the media ₆₁₆ being added to the culture vial. The effect of the initial addition of the drug solution is ₆₁₇ significant and leads to a 33.5% reduction of the optical density. According to the ₆₁₈ selection algorithm, the drug concentration is then lowered to allow for the recovery of ₆₁₉ the population. This is done by repetitively adding fresh media to the culture while ₆₂₀ removing waste. In the experiment, the bacterial population is able to recover in 12 ₆₂₁ hours. The drug solution is then re-introduced to the culture. After this second ₆₂₂ introduction, the optical density decreases 15%, indicating development of ₆₂₃ tolerance/resistance to the drug. The population is again allowed to recover for a ₆₂₄ shorter period – around 6 hours. After this recovery period, the addition of more ₆₂₅ ampicillin does not kill the population. At the end of the experiment, the cells are ₆₂₆ effectively resistant to ampicillin. Resistance of the resultant population can be ₆₂₇ confirmed with an MIC assay [23]. The experiment can be performed with a variety of ₆₂₈ different drugs, however drug stability in cell culture should be considered. For instance, ₆₂₉ ampicillin is stable in culture for 3 days [24]. If the experiment lasts longer than the ₆₃₀ stability of the drug in culture, the drug solution can be refreshed by replacing the drug ₆₃₁ reservoir with one that is freshly made. ₆₃₂

Fig 7. Sample singleton growth experiment with an ampicillin selective pressure applied by the Toprak selection algorithm [11] All three graphs are generated in real-time during the experiment. Users can monitor the activity of a culture unit through the graphs. (a) Voltage over time, as measured from CU photodiode. These values can be optionally co-related to calibration curves in the supplemental documents to achieve OD600 readings. (b) Pump activation shown over time. The red lines (D) represent the activity of the pump that passages a solution of media and drugs into the vial. The green lines (M) represent the activity of the pump that passages media into the vial. Whenever the drug and media pump activate, a waste pump also activates to maintain volume equilibrium. The activity of this pump is not shown in the figure as it matches the sum of the activity of the drug and media pumps. All pump activity is recorded in the software. (c) Calculated drug concentration plotted over time. Values are calculated from the concentration of the drug solution reservoir and the duration of pump activity. These values do not take into account consumption of the drug in the vial and only account for dilution effects.

Discussion ₆₃₃

While extensibility was of prime importance, device affordability was also an important ₆₃₄ consideration during the design of the device. Based on the parts used in our ₆₃₅ experiment, an EVE build for simple growth experiments with triplicate CUs can be ₆₃₆ built for approximately $72, while an EVE build for directed growth experiments with ₆₃₇ selective agents will cost approximately $154.Supplemental Information This cost is an ₆₃₈ estimate for all the parts excluding the incubator, 3D printer costs, and any glassware ₆₃₉ needed to prepare the solutions for the experiment. A system meant for growth ₆₄₀ experiments is markedly cheaper since it does not require pumps or tubing – parts ₆₄₁

Fig 8. Sample triplicate growth experiment with an ampicillin selective pressure. (a) Three separate Culture Units inoculated from the same population showing different drug responses over time in the presence of Ampicillin. The blue lines show the raw voltage (related to the optical density of the system) and the orange lines show the concentration of the drug over time. Flat lines demonstrate no pump activity as the system prevents further dilution of the population by intermittently passaging media into the vial. (b) A representation of pump activity over time. Gold lines (M) represent the activity of the media pump while the red lines (D) represent activity of the drug pump. (c) A graph combining the response curves of the separate populations within each culture vial. Both the manner of the response to initial exposure and reponse to subsequent exposure to ampicilin differ within the Culture Units. (d) Temperature of the incubator during the experiment.

which can add significantly to the total cost of the system. Vendor costs are also listed on the parts list, so costs can be driven down significantly for users who buy their parts from commercial vendors such as eBay or Amazon. This is particularly relevant for those who wish to implement the system in the educational setting. This low cost enables labs to explore concepts of evolution without investing much capital into tools. The cost also makes this project feasible for the educational environment, where cost is not a burden for educators to investigate the feasibility of this project in their classroom or extra-curricular setting. The two experiments described above serve as tests of the hardware and software and demonstrate the most essential components of the system. Successful completion of the first experiment confirms proper operation of the LED sensors and the connections from the microcontroller. In addition, it provides ample opportunity to identify sources of contamination.

The dilution rate of the EVE can be calculated by expanding the equation $D = F/V$ where D is the dilution rate of the bioreactor, F is the flow rate of fresh media, and V is the total culture volume of the reactor. Based on parameters used in this system, the dilution rate can be re-written as $D = (P \times t_p)/(V \times t_i)$ where P is the pump flow rate of the media pump, t_p is the time the media pump is active, V is the total volume of the culture vial (usually 12 mL), and t_i is the interval of time between pump activations (usually 12 minutes). This formula is takes into account the periodic activation of the media pump in the EVE as opposed to the continuous infusion of fresh media that is commonplace in conventional bioreactors. The parameters in the formula can be modified in the configuration file, and in some cases (such as turbidostat mode) can be modified automatically by the EVE depending on the state of the culture. As a result, the software measures and graphs the dilution rate throughout the experiment for easy monitoring. In steady state conditions (where the dilution of the culture is equal to the amount of growth in the culture), users can approximate the doubling time of the bacteria in the culture unit by taking the inverse of the dilution rate and multiplying this number with the natural logarithm of two.

Bacteria is commonly used in the educational environment to promote awareness of scientific concepts. Generation of simple bacterial growth curves require fewer resources and is generally very suited for education. Experiments in the EVE to evolve resistance are more suited to the academic environment due to the relative availability of antibiotics and proper pathways to dispose of resistant organisms. However, students and teachers can perform the evolution experiment in the EVE with the use of natural antibiotics such as garlic or ginger extract. [25] In these scenarios, students and teachers mix these natural products into their drug reservoir when creating it. These natural products provide a safe way to explore evolution in bacteria and are also easily accessible by students or teachers.

The second experiment demonstrates the pumps and the 'morbidostat' algorithm in

action. The algorithm incorporated into the system is a modification from Toprak *et* ₆₈₁
al. [11] In its origtnal conception, maximum and minimum OD thresholds are ₆₈₂
determined from pre-experiment bench tests and are dependent on a number of ₆₈₃
variables including the species of bacteria used. Different trials of the same bacteria ₆₈₄
might also yield different maximal optical density readings and the bacteria might enter ₆₈₅
the log phase at different optical densities owing to the stochasticity of evolution [26] as ₆₈₆
well as variation in ambient light annother experimental conditions. To account for this ₆₈₇
variation, the control program is capable of shifting the threshold automatically. It does ₆₈₈
this by estimating the first derivative and second derivative of the real-time optical ₆₈₉
density (real-time growth rate and real-time growth acceleration respectively). Based on ₆₉₀
these readings, it shifts the threshold to the optical density when the growth rate ₆₉₁
crosses a certain pre-specified threshold and the growth acceleration is approximately ₆₉₂
zero. In the results and conclusion presented here though 'optical density' is discussed, ₆₉₃
the true absorbance was not calculated or derived, with the data reported being raw ₆₉₄
voltage data from the photodiode. A simple calibration of this data using a standard ₆₉₅
spectrophotmeter is recommended to report data from the machine in terms of true ₆₉₆
optical density. ₆₉₇

When antibiotics are added to the solution, there is a corresponding reduction in ₆₉₈
optical density. Ostensibly, this reduction is due to the cytotoxic effect of the ₆₉₉
antibiotics, leading to a reduction to the number of alive cells in the vial. However, ₇₀₀
other factors need to be considered in this scenario. The effect that dead cells within ₇₀₁
the culture have on optical density is yet unknown. Indeed, death may not be ₇₀₂
happening at all and the antibiotics simply limit the cell division of the bacteria ₇₀₃
(bacteriostatic behavior) while the pumps dilute the culture by periodically pumping in ₇₀₄
fresh media. Nevertheless, reductions in the optical density post antibiotic infusion are ₇₀₅
certianly due to the actions of the antibitiocs regardless of the mechanism of action. ₇₀₆
Subsequent reduction in the reaction of the population to addition of antibiotics can be ₇₀₇
classified as development of resistance. ₇₀₈

A further, novel application of this system is the ability to manage multiple ₇₀₉
protocols at once. The capability of the EVE system to manage multiple protocols in ₇₁₀
this manner makes it the only system of its kind that is fully open source and of ₇₁₁
low-cost. Such versatility allows this system to be used as a comprehensive automated ₇₁₂
evolution experiment management tool for numerous diverse experiments in a ₇₁₃
self-contained integrated manner. In the experiment above, the default algorithm used ₇₁₄
is identical to that used in the original paper by Toprak and colleagues (Figure 1 in that ₇₁₅
paper). In this algorithm, the culture is allowed to recover by sequentially passing fresh ₇₁₆
media after exposure to the drug. However, there is an alternative control algorithm ₇₁₇
that can be used during the experiment. Users can choose to maintain a constant ₇₁₈
concentration of drug once the bacterial population is sufficiently healthy inside the ₇₁₉
tube. Biologically, since the selective pressures of the different populations were applied ₇₂₀
differently, the way resistance is reached between the two populations is different. ₇₂₁

In contrast to other pressured bioreactors, EVE does not keep the culture vial ₇₂₂
pressurized or rely on air pumps to transfer liquid. Liquid pumps are used to both ₇₂₃
pump waste out of the vial as well as pump fresh media and drug solution into the vial. ₇₂₄
To maintain atmospheric pressure, the vials and reservoirs are fitted with sterile air ₇₂₅
filters. These filters maintain atmospheric pressure across the system while ensuring ₇₂₆
sterile conditions within the stock of fresh media and preventing contamination of the ₇₂₇
culture units. ₇₂₈

The maximum theoretical drug concentration inside the vial is equal to the ₇₂₉
concentration of the drug inside the drug solution reservoir. Maximizing this theoretical ₇₃₀
drug concentration, however, is at odds with slowly raising the concentration of the ₇₃₁
drug inside the vial. If the reservoir is too concentrated and if the drug solution pump ₇₃₂

replaces 1/12th of the solution, it is likely that the drug concentration will be extremely ₇₃₃
high inside the vial and kill all the bacteria. A solution to this problem is to allow the ₇₃₄
drug solution pump to run for different lengths of time depending on the stage of the ₇₃₅
experiment. If instead of replacing 1/12th of the solution at the beginning, if the pumps ₇₃₆
run for a shorter time, a lower concentration of drug can be achieved inside the vial. ₇₃₇
Later in the experiment, the drug pumps can be run for longer, ensuring that the ₇₃₈
population is exposed to the maximal possible drug concentration. A maximum ΔV ₇₃₉
(such as the 1mL proposed by Toprak and colleagues) can be specified so that the ₇₄₀
culture population does not get washed away by the drug solution. This way, ₇₄₁
development of resistance is assured and the population faces strong selection pressures ₇₄₂
as the experiment progresses. ₇₄₃

The EVE also easily lends itself to fluorescence experiments where LED and PDs ₇₄₄
corresponding to excitation and emission frequencies can be incorporated into the ₇₄₅
design. By using fluorescence instead of optical densities, growth of several populations ₇₄₆
can be distinguished. In the context of varying selective pressures (provided by the ₇₄₇
antibiotics), interactions of different populations of bacteria can be studied to further ₇₄₈
elucidate the effect of these interactions on the generation of drug resistance. Game ₇₄₉
theory and other methods of mathematical modeling can be used to draw conclusions ₇₅₀
about the interactions of the different species/types of bacteria [27, 28]. ₇₅₁

Insights for the Future ₇₅₂

There are several limitations of the device. Naturally, since all the software is executed ₇₅₃
from the Raspberry Pi, memory and processing power is limited during the experiment. ₇₅₄
Many of the programs, especially the master program that interfaces with the circuit ₇₅₅
board during the experiment, have been streamlined to utilize as little of the memory ₇₅₆
and processing power as possible. A queuing system was implemented to prevent many ₇₅₇
threads that utilize high processing power from executing at the same time. ₇₅₈
Furthermore, as more culture units and corresponding sensors are configured, an ₇₅₉
increasingly lower resolution list of the data is generated in addition to the ₇₆₀
high-resolution optical density data. This lower-resolution data is used for the live ₇₆₁
graphing to prevent the greater processing power that's required to graph high ₇₆₂
resolution data. ₇₆₃

To circumvent processing power and memory constraints from preventing users to ₇₆₄
implement more CUs, future work on a server-based version will be implemented to take ₇₆₅
advantage of remote processing power while allowing the Raspberry Pi simply interface ₇₆₆
with the circuit board and relay information to the server. In this scenario, the web ₇₆₇
interface and all the data are managed by the server. Docker integration in this step ₇₆₈
will be crucial to streamline the process and create a simplistic but powerful and robust ₇₆₉
way to deploy the EVE system across the local network. ₇₇₀

Another limitation of the current design is that the circuit board only supports 4 ₇₇₁
simultaneous culture units. In this scenario, the hardware addresses are "hard-coded" ₇₇₂
into the circuit board design. An alternative, but more complicated design, would allow ₇₇₃
for jumpers on the circuit board, allowing users to easily select four different hardware ₇₇₄
addresses. The user is then able to attach circuit board together with one circuit board ₇₇₅
interfacing with the Raspberry Pi. Ultimately, for users who would like to increase the ₇₇₆
level of multiplexing, with 4 circuit boards connected to each other, a maximum of 16 ₇₇₇
CUs will be able to be controlled simultaneously. An advantage of this design is its ₇₇₈
modular capabilities, allowing users to only print as many circuit boards as they need ₇₇₉
and limit the number of components needed to make a functional circuit board and ₇₈₀
reduce overall system cost. A future version of the circuit board will implement this ₇₈₁
design improvement. ₇₈₂

Another future direction is to allow the machine to sequentially inject multiple drugs ₇₈₃

autonomously. The current version of the proposed system only passages one drug   784
solution and therefore can only explore the pathways to specific drug resistance.   785
However, there are many biologically and clinically relevant questions that can be asked   786
when different drugs are given sequentially or together. Concepts like collateral   787
sensitivity and cross-resistance can be effectively studied – especially when combined   788
with frequent sampling of the bacterial population [26, 29]. Tracking the concentrations   789
of multiple drugs in the vial and creating algorithms for drug scheduling can be added   790
in future updates or merged to the main repository from community-driven efforts. In   791
this scenario, a microfluidic solution such as that implemented by Wong *et al.* would be   792
useful [13].   793

Conclusion   794

We present a low-cost, open-source, extensible, and functionally-flexible bioreactor   795
design. The utility of this machine is not just limited to the lab environment where the   796
evolution of drug resistance can be studied with high temporal resolution, but extends   797
to the educational environment where students and educators can engage in an   798
interdisciplinary project to study biological concepts through technology.   799

The proposed framework can have many applications, including many not discussed   800
here. Indeed, use of this device may be used to study a wide variety of biological   801
concepts – even outside the field of evolutionary biology. The open source nature of the   802
project welcomes modification to the hardware and the software for more specialized   803
applications of the technology. We hope the versatile and accessible nature of this   804
device makes it useful for scientists and educators from all areas.   805

Supplemental Information   806

- **Parts list.** List of all parts that comprise the EVE system with estimated price: 807
  https://github.com/vishhvaan/eve-   808
  pi/raw/master/Start%20Building/Parts%20List.xlsx.   809
    810

- **Configuration File Parameter Definitions.** List of all the modifiable   811
  variables and experiment parameters for each Culture Unit: 812
  https://github.com/vishhvaan/eve-   813
  pi/raw/master/Start%20Building/Configuration%20File%20Parameter%20Definitions.xlsx.   814
    815

- **3D Hardware Files.** All the STLs and SLDPRTs for EVE hardware: 816
  https://github.com/vishhvaan/eve-   817
  pi/tree/master/Start%20Building/3D%20Files.   818
    819

- **Circuit Schematics.** Circuit schematics and files required to print or modify a   820
  Printed Circuit Board (PCB): https://github.com/vishhvaan/eve-   821
  pi/tree/master/Start%20Building/Circuit%20Board%20Files.   822

Acknowledgments   823

Rick, our 3D printer. We love you mate. In addition, we'd like to thank VeloSano for   824
their generous support of this work.   825

Contributions

Indicates contribution

References

1. Nash DB. Open-Source Science. Biotechnology healthcare. 2005;2(2):8.

2. Matteau D, Baby V, Pelletier S, Rodrigue S. A small-volume, low-cost, and versatile continuous culture device. PLoS One. 2015;10(7):e0133384.

3. Gill JM. The Internet of Things in the Life Sciences Laboratory; 2018.

4. Cressey D. The DIY electronics transforming research. Nature News. 2017;544(7648):125.

5. Perkel JM. The Internet of Things comes to the lab. Nature News. 2017;542(7639):125.

6. Pilizota T, Yang YT. "Do It Yourself" Microbial Cultivation Techniques for Synthetic and Systems Biology: Cheap, Fun, and Flexible. Frontiers in microbiology. 2018;9.

7. Novick A, Szilard L. Description of the chemostat. Science. 1950;112(2920):715–716.

8. Voulgarelis D, Velayudhan A, Smith F. Stability of two competing populations in chemostat where one of the population changes its average mass of division in response to changes of its population. PloS one. 2019;14(3):e0213518.

9. eVOLVER Full Store;. Available from: https://www.fynchbio.com/evolver-full.

10. McGeachy AM, Meacham ZA, Ingolia NT. An Accessible Continuous-Culture Turbidostat for Pooled Analysis of Complex Libraries. ACS synthetic biology. 2019;8(4):844–856.

11. Toprak E, Veres A, Yildiz S, Pedraza JM, Chait R, Paulsson J, et al. Building a morbidostat: an automated continuous-culture device for studying bacterial drug resistance under dynamically sustained drug inhibition. Nature protocols. 2013;8(3):555.

12. Yoshida M, Reyes SG, Tsuda S, Horinouchi T, Furusawa C, Cronin L. Time-programmable drug dosing allows the manipulation, suppression and reversal of antibiotic drug resistance in vitro. Nature communications. 2017;8:15589.

13. Wong BG, Mancuso CP, Kiriakov S, Bashor CJ, Khalil AS. Precise, automated control of conditions for high-throughput growth of yeast and bacteria with eVOLVER. Nature biotechnology. 2018;36(7):614.

14. Takahashi CN, Miller AW, Ekness F, Dunham MJ, Klavins E. A low cost, customizable turbidostat for use in synthetic circuit characterization. ACS synthetic biology. 2014;4(1):32–38.

15. Liu PC, Lee YT, Wang CY, Yang YT. Design and use of a low cost, automated morbidostat for adaptive evolution of bacteria under antibiotic drug selection. JoVE (Journal of Visualized Experiments). 2016;(115):e54426.

16. Miller AW, Befort C, Kerr EO, Dunham MJ. Design and use of multiplexed chemostat arrays. JoVE (Journal of Visualized Experiments). 2013;(72):e50262.

17. Hoffmann SA, Wohltat C, Müller KM, Arndt KM. A user-friendly, low-cost turbidostat with versatile growth rate estimation based on an extended Kalman filter. PloS one. 2017;12(7):e0181923.

18. Cooper VS, Warren TM, Matela AM, Handwork M, Scarponi S. EvolvingSTEM: a microbial evolution-in-action curriculum that enhances learning of evolutionary biology and biotechnology. Evolution: Education and Outreach. 2019;12(1):12.

19. Milias-Argeitis A, Rullan M, Aoki SK, Buchmann P, Khammash M. Automated optogenetic feedback control for precise and robust regulation of gene expression and cell growth. Nature communications. 2016;7:12546.

20. Dößelmann B, Willmann M, Steglich M, Bunk B, Nübel U, Peter S, et al. Rapid and consistent evolution of colistin resistance in extensively drug-resistant Pseudomonas aeruginosa during morbidostat culture. Antimicrobial agents and chemotherapy. 2017;61(9):e00043–17.

21. Inc PT. Collaborative data science; 2015. Available from: `https://plot.ly`.

22. WHO - Methods of Sterilization;. Available from: `https://apps.who.int/phint/pdf/b/7.5.9.5.8-Methods-of-500sterilization.pdf`).

23. Andrews JM. Determination of minimum inhibitory concentrations. Journal of antimicrobial Chemotherapy. 2001;48(suppl_1):5–16.

24. Perlman D. [7] Use of antibiotics in cell culture media. In: Methods in enzymology. vol. 58. Elsevier; 1979. p. 110–116.

25. Miller GJ, Cunningham AM, Iwase Y, Lautensack NL, Sattley WM. A Laboratory Activity Demonstrating the Antibacterial Effects of Extracts from Two Plant Species, Moringa oleifera and Allium sativum (Garlic). Journal of microbiology & biology education. 2017;18(3).

26. Nichol D, Rutter J, Bryant C, Hujer AM, Lek S, Adams MD, et al. Antibiotic collateral sensitivity is contingent on the repeatability of evolution. Nature communications. 2019;10(1):334.

27. Kaznatcheev A, Peacock J, Basanta D, Marusyk A, Scott JG. Fibroblasts and alectinib switch the evolutionary games played by non-small cell lung cancer. Nature ecology & evolution. 2019;3(3):450.

28. Friedman J, Higgins LM, Gore J. Community structure follows simple assembly rules in microbial microcosms. Nature ecology & evolution. 2017;1(5):0109.

29. Maltas J, Wood KB. Pervasive and diverse collateral sensitivity profiles inform optimal strategies to limit antibiotic resistance. bioRxiv. 2018; p. 241075.

Figure 1: **Hardware schematic of the EVE system.** The culture vial is located inside the incubator. As the culture grows, an LED shines a light through the culture. Light that is reflected by the culture is received by the photodiode. A circuit board interprets and conveys photodiode readings to the Pi which runs control software. Based on user-configured directions, and optical density readings, the Pi uses the circuit board to power pumps connected to the vial. This strategy enables precise and dynamic control of the culture.

Figure 2: **Hardware components of the EVE.** (a) Square fan with magnets attached to a spacer. (b) Three pumps for three functions (pumping in media, pumping in the solution of media and drug, and pumping out waste). All three are connected to the circuit board. (c) Silicone tubes as they are connected to the pumps along with terminal leur connections. (d) Pump in the 3D printed pump holder. (e) Three culture vials with 12mL of media and stirbars. The longest PEEK tube touches the culture. The other PEEK tubes meant to introduce media solutions to the vial do not touch the surface of the media. (f) Media reservoir with three small silicone tubes to transport media to the vials. (g) Terminal connections of the media reservoir. These connect to the terminal connections of the pump shown in (c). (h) 3D printed vial holder with vial inside. Diodes are held in place with 3D printed caps. (i) Top of the reservoir displaying the small silicone tubes attached with the silicone sealant (j) 70% Ethanol reservoir used for sterilization of the small silicone tubes prior to the experiment. Terminal connections and the attached air filter of the reservoir is displayed. (k) An incubator with three Culture Units. (l) The vial holder and fan with magnets fit into the base. Diodes are attached to the vial holder. Together, these components comprise a Culture Unit.

Figure 3: **Software schematic of the EVE system.** The devices software resides on the Raspberry Pi and center around a Python-based web-server which serves as the GUI for experimental design and control. This implementation offers built-in remote monitoring and multi-user control of the device. More EVE-centric Python-based applications are also installed Pi, accessible via the web interface. These applications are able to share data, and perform experiments based user defined configurations. The experiment controller application directly interfaces with the circuit board. All of these applications are open source and are built with extensibility in mind. Users have access to all the underlying programs and code. We encourage them to fork our GitHub repository or download programs to their devices and make modifications as they see fit. See Fig 1 for the hardware schematic of the EVE system.

Figure 4: **Web interface of the EVE running on the Pi.** A variety of tabs are available to users to configure and run experiments, selectively turn on the pumps (useful for sterilization prior to the experiment), and control the Pi directly. This server is run from the Pi itself and is automatically installed after the user runs the setup script. This interface is accessible from any device connected to the same network the EVE resides on. It was built to allow ease of use and particularly, to provide non-programmers with a simple way of accessing and using the device. In this screen shot, specific pumps in multiple CUs have been turned on for twenty seconds.

Figure 5: **Sample Calibration Curve of the EVE.** (a) Linear regression is performed in each Culture Unit (CU) between measurements of the raw voltage to the true absorbance as determined by a spectrophotometer. The resultant relationship between the true optical density and raw voltage can be mapped to the voltage readings of growth and selection experiments to attain a true optical density measurement. (b) The relationship between voltage and absorbance at 600nm is similar across CUs indicating that any voltage differences measured between CUs is a result of differences in the behaviour of the population.

Figure 6: **Sample graphs for a triplicate bacterial growth experiment.** Triplicate bacterial growth experiments sharing the same axis. (a) Technical replicates (individual CUs) show different voltage readings over time despite being connected to the same circuit board and made identically. This variation is due to subtle differences in the positioning of the diodes in the holder. (b) When the readings are scaled by the maximum voltage reading, all curves overlap, exemplifying the triplicate nature of the experiment. Diode differences are constant over time, therefore raw voltage values can be correlated with OD600 values as seen in Fig 5. For simple experiments, this step is unnecessary. In general, this growth experiment can be readily performed by scientists to test their system or by students and teachers in scientific classes.

Figure 7: **Sample singleton growth experiment with an ampicillin selective pressure applied by the Toprak selection algorithm [?]** All three graphs are generated in real-time during the experiment. Users can monitor the activity of a culture unit through the graphs. (a) Voltage over time, as measured from CU photodiode. These values can be optionally co-related to the calibration curves in the supplemental documents to achieve OD600 readings. (b) Pump activation shown over time. The red lines (D) represent the activity of the pump that passages a solution of media and drugs into the vial. The green lines (M) represent the activity of the pump that passages media into the vial. Whenever the drug and media pump activate, a waste pump also activates to maintain volume equilibrium. The activity of this pump is not shown in the figure as it matches the sum of the activity of the drug and media pumps. All pump activity is recorded in the software. (c) Calculated drug concentration plotted over time. Values are calculated from the concentration of the drug solution reservoir and the duration of pump activity. These values do not take into account consumption of the drug in the vial and only account for dilution effects.

Figure 8: **Sample triplicate growth experiment with an ampicillin selective pressure.** (a) Three separate Culture Units inoculated from the same population showing different drug responses over time in the presence of Ampicillin. The blue lines show the raw voltage (related to the optical density of the system) and the orange lines show the concentration of the drug over time. Flat lines demonstrate no pump activity as the system prevents further dilution of the population by intermittently passaging media into the vial. (b) A representation of pump activity over time. Gold lines (M) represent the activity of the media pump while the red lines (D) represent activity of the drug pump. (c) A graph combining the response curves of the separate populations within each culture vial. Both the manner of the response to initial exposure and reponse to subsequent exposure to ampicilin differ within the Culture Units. (d) Temperature of the incubator during the experiment.