Detection and characterization of copy number variants based on whole-genome sequencing by MGI platforms

Junhua Rao[1, *], Lihua Peng[2, *], Fang Chen[1], Hui Jiang[1], Chunyu Geng[1], Xia Zhao[1], Xin Liu[3], Xinming Liang[1, †], Feng Mu[1, 4, †]

[1] MGI, BGI-Shenzhen, Shenzhen 518083, China
[2] BGI-Shenzhen, Shenzhen 518083, China
[3] BGI-Qingdao, BGI-Shenzhen, Qingdao 266555 Shandong, China
[4] MGI-Wuhan , BGI-Shenzhen, Wuhan 430074, China

* These authors contributed equally to this work.
† Correspondence authors: liangxinming@genomics.cn (Xinming Liang) and mufeng@genomics.cn (Feng Mu)

Abstract

Background
Next-generation sequence (NGS) has rapidly developed in past years which makes whole-genome sequencing (WGS) becoming a more cost- and time-efficient choice in wide range of biological researches. We usually focus on some variant detection via WGS data, such as detection of single nucleotide polymorphism (SNP), insertion and deletion (Indel) and copy number variant (CNV), which playing an important role in many human diseases. However, the feasibility of CNV detection based on WGS by MGI platforms was unclear. We systematically analysed the genome-wide CNV detection power of MGI platforms and Illumina platforms on NA12878 with five commonly used tools, respectively.

Results
MGI platforms showed stable ability to detect slighter more CNVs on genome-wide (average 1.24-fold than Illumina platforms). Then, CNVs based on MGI platforms and Illumina platforms were evaluated with two public benchmarks of NA12878, respectively. MGI and Illumina platforms showed similar sensitivities and precisions on both two benchmarks. Further, the difference between tools for CNV detection was analyzed, and indicated the selection of tool for CNV detection could affected the CNV performance, such as count, distribution, sensitivity and precision.

Conclusion
The major contribution of this paper is providing a comprehensive guide for CNV detection based on WGS by MGI platforms for the first time.

Background

Copy number variation (CNV) is becoming more and more important in genomic variation since first reported in 2004[1, 2]. Currently, the large structure variation, especially CNV, have been heavily studied to clearly demonstrated them playing an important role in human diseases, such as autism[3-6], schizophrenia[7],

[Insert Running title of <72 characters]

Parkinson[8], Hirschsprung[9] and cancer[10]. In the same time, many sequencing
technologies and methods about CNV come into being. Fluorescence *in situ*
hybridization (FISH), array-based comparative genomic hybridization (array
CGH)[11, 12] and SNP array were used for CNV detection at the beginning. Later,
whole-exome sequencing (WES) and whole-genome sequencing (WGS), two
methods based on next-generation sequencing (NGS), were introduced for CNV
detection and were widely used in a large number of biological researches. Without
the limitations of specified target regions associated with hybridization or array, WES
and WGS become more advisable in CNV detection.
Illumina developed many sequencing platforms using reversible terminator
chemistry and Illumina`s sequencing by synthesis (SBS) technology, such as
HiSeq2500, HiSeq X Ten and NovaSeq6000[13]. WES and WGS sequenced on
Illumina platforms were widely used in biological research, including CNV detection.
Meanwhile, many tools were developed for CNV detection with WES[14] or
WGS[15, 16]. There are five main strategies used in tools for CNV detection: (1) read
pair (RP), (2) read depth (RD), (3) split read (SR), (4) de novo assembly (AS), and (5)
combination of approaches (CA). And each strategy has its own advantage and
limitation. RD-based tools can get accurate copy number but have no ability to detect
the breakpoints of CNV. SR-based tools can get breakpoints at base-pair resolution
but have low ability in low-complexity regions. AS-based tools could detect CNVs
without a known reference but have high-requirement of computational resources
[15]. Moreover, different tools were designed for different sample conditions. For
example, BreakDancer[17] was developed for RP strategy and was built for single
sample, HYDRA[18] based on CA strategy was suit for multiple samples and
CNVnorm[19] based on RD strategy was designed for case-control condition.
MGI Tech Co., Ltd. (MGI), a subsidiary of BGI Group and the parent company of
Complete Genomics^{TM}(CG), which is the instruments manufacturing business of BGI,
developed several desktop sequencing platforms using DNA nanoball (DNB) and
combinational probe anchor synthesis (cPAS) since 2016, such as BGISEQ-500,
DNBSEQ-G400 (MGISEQ-2000) and DNBSEQ-T7. Compared with Illumina`s
sequencing platform, BGISEQ-500 also showed high performance on detection of
SNP and Indel on WGS[20, 21] or WES[22]. Also, MGI platform is a good choose
for transcriptome analysis in plants[23].
Accompanied by the improvement of NGS, primarily for decrease of time and cost,
increase of data size per run and sequence quality, more CNV researches are carried
out based on WGS by NGS platforms. Many CNV detection analysis based on WGS
by Illumina platforms were evaluated, but there is no comprehensive guide for CNV
detection based on WGS by MGI platform. Here, we presented a performance
analysis of CNV detection by analysing eight WGS demo data sequenced on two
MGI platforms with five common tools based on four strategies. For comparison, two
Illumina platforms were used as control platforms and we indicated that both the
sensitivity and the precision of CNVs based on MGI platforms were consistent with
those of CNVs based on Illumina platforms.

87   Results

88   Selection of tools and benchmarks
89   We selected five tools for CNV detection from published CNV tools in three
90   factors: single sample pattern, widely used and keep date (Supplementary Table S1).
91   These five tools were designed for genome-wide CNV detection based on different

[Insert Running title of <72 characters]

strategies: BreakDancer[17] was built on RP, CNVnator[24] was built on RD,
Pindel[25] was built on SR, DELLY[26] and LUMPY[27] were built on CA.
Moreover, two available CNV benchmarks of NA12878 were introduced for
further evaluation of CNV: 2,819 CNVs from Layer, et al. in 2014
(LUMPY_2014)[27] and 2,171 CNVs from the 1000 Genome Project WGS data in
2015 (1KG_2015)[28]. A CNV in benchmark would be considered as specific if the
proportion of its region size which overlapped with other CNVs in benchmark
reciprocally was less than a threshold. When the threshold was 90.00%, 56.05% of
CNVs in LUMPY_2014 and 32.43% of CNVs in 1KG_2015 were specific,
respectively (Supplementary Figure S1a). While the threshold was 50%, the ratio of
specific CNVs was 47.82% in LUMPY_2014 and 30.49% in 1KG_2015, respectively
(Supplementary Figure S1b). In a word, there were large proportion specific CNVs in
both LUMPY_2014 and 1KG_2015, so these two benchmarks both were used in this
study.
106
CNV detection and filtration
To detect CNV, we used ten WGS datasets of NA12878 sequenced on two MGI
sequencing platforms (three datasets sequenced on BGISEQ-500, five datasets
sequenced on DNBSEQ-G400 (MGISEQ-2000) and two Illumina sequencing
platforms (one dataset sequenced on HiSeq2500 and one dataset sequenced on
NovaSeq6000) (Supplementary Table S2, See Methods). Each dataset was down-
sampled and processed for BAM following our previous approach[20]. We got eight
BAMs based on MGI platforms with average 31.43x and 99.02% whole genome
coverage, and two BAMs based on Illumina platforms with average 30.60x and
99.19% whole genome coverage (Table 1). Then, these BAMs of ten datasets were
used to detect CNVs across genome by five tools: BreakDancer, CNVnator, Pindel,
DELLY and LUMPY, respectively. In total, we obtained 50 files of CNV results.
To reduce the false positives, firstly, low-quality CNVs were filtered out from raw
CNVs (See Materials and Methods). Still, high ratio ambiguous CNVs were remained
and these ambiguous CNVs might decreased the performance of CNV detection. Of
which, a CNV in each CNV result was defined as ambiguous if it had at least 1bp
overlap with other CNVs in the same CNV result, and the rest of CNVs were defined
as unambiguous CNVs. We found average 9.82% (rang from 3.75% to 20.36%)
CNVs detected by BreakDancer for ten datasets were ambiguous, while the ratio of
ambiguous CNVs was 39.43% (range from 30.97% to 49.96%) for Pindel, 19.87%
(range from 16.55% to 26.81%) for DELLY, 30.23% (range from 23.73% to 43.67%)
for LUMPY and 0% for CNVnator, respectively (Supplementary Figure S2a).
Further, we calculated the precision of ambiguous CNVs and unambiguous CNVs
with two benchmarks, respectively (See Materials and Methods). For LUMPY_2014
benchmark, the precision of unambiguous CNVs was significantly higher than that of
ambiguous CNVs (average 51.17% of unambiguous CNVs vs. average 12.83% of
ambiguous CNVs, p = 5.64e-13 with t-test, Supplementary Figure S2b). Similarly, for
1KG_2015 benchmark, we also found the precision of unambiguous CNVs was
significantly higher than ambiguous CNVs (average 39.23% vs. average 9.08%, p =
1.11e-16 with t-test, Supplementary Figure S2c). Those results indicated the negative
effect of ambiguous CNVs and suggested the filtering of ambiguous CNVs to
improve performance of CNV detection.
In a word, after removing low-quality and ambiguous CNVs from raw CNVs, we
got final average 2,586 CNVs (range from 897 to 5,778) on genome-wide using five
tools for ten datasets (Figure 1). The data size, major alignment statistics and total

[Insert Running title of <72 characters]

142   CNV calls using five tools were summarized in Table 1 (more statistics of CNVs were
143   list on Supplementary Table S3).
144
High quality CNVs produced from MGI platforms
To evaluate the performance of CNVs based on MGI platforms, two Illumina
platforms were used as control platforms. Firstly, we compared the number of CNVs
detected by each tool between MGI and Illumina sequencing datasets. We got average
1,827 (range from 1,519 to 2,329) CNVs using BreakDancer for eight datasets from
two MGI platforms, while the CNV count was 3,523 (range from 2,209 to 5,778)
using CNVnator, 4,506 (range from 4,172 to 4,924) using Pindel, 1,838 (range from
1,624 to 2,209) using DELLY and 1,750 (range from 1,422 to 2,378) using LUMPY
(Table 1). While in two datasets from two Illumina platforms, we got 933 (897 and
968) CNVs using BreakDancer, 1,945 (1,229 and 2,660) using CNVnator, 4,888
(4,045 and 5,730) using Pindel, 1,741 (1,709 and 1,773) using DELLY and 1,365
(1,350 and 1,380) using LUMPY (Table 1). We found the average CNVs from MGI
platforms was 1.96-fold of average CNVs from Illumina platforms using
BreakDancer, 1.81-fold using CNVnator, 0.92-fold using Pindel, 1.06-fold using
DELLY and 1.28-fold using LUMPY (Table 1). Those results showed that WGS data
sequenced on MGI platforms could produce more CNVs on genome wide.
Secondly, we annotated CNVs by locating CNVs to seven genomic regions:
Upstream, Exonic, Intronic, Downstream, Intergenic, CpG island (CGI) and CGI-
shore. We found similar distribution of all 50 CNV result files (average 37.28%
CNVs based on MGI platforms and 36.25% CNVs based on Illumina platforms were
in Intergenic, 28.32% and 31.47% in Exonic, 27.49% and 25.19% in Intronic, 3.53%
and 3.56% in Downstream, 3.38% and 3.53% in Upstream, 5.01% and 4.53% in CGI-
shore,2.85% and 2.05% in CGI, Supplementary Figure S6, Supplementary Table S5),
which suggested CNVs based on MGI platforms were highly consistent with Illumina
platforms.
Further, we evaluated the sensitivity of CNVs based on MGI platforms and
Illumina platforms. When compared with LUMPY_2014 benchmark, the average
sensitivity was 54.44% of CNVs based on MGI platforms vs. 26.61% of CNVs based
on Illumina platforms using BreakDancer, while 13.36% vs. 12.73% using CNVnator,
67.06% vs. 47.62% using Pindel, 38.64% vs. 30.08% using DELLY and 43.52% vs.
29.57% using LUMPY when compared with LUMPY_2014 (Figure 2,
Supplementary Figure S5). And when the benchmark 1KG_2015 was used, the
average sensitivity was 51.57% vs. 30.26% using BreakDancer, 25.98% vs. 24.53%
using CNVnator, 70.47% vs. 52.85% using Pindel, 39.16% vs. 33.21% using DELLY
and 42.96% vs. 32.66% using LUMPY (Figure 2, Supplementary Figure S5). Those
results suggested that there were higher sensitivities of CNVs based on MGI
platforms than CNVs based on Illumina platforms using different tool and different
benchmark.
Meanwhile, we evaluated the precision of CNVs between MGI platforms and
Illumina platforms and found the average precision was 82.67% vs. 78.53% using
BreakDancer, 11.29% vs. 20.55% using CNVnator, 41.77% vs. 28.25% using Pindel,
58.97% vs. 48.12% using DELLY and 69.27% vs. 60.58% using LUMPY when
compared with LUMPY_2014. While 58.26% vs. 64.22% using BreakDancer, 15.24%
vs. 26.90% using CNVnator, 33.23% vs. 23.75% using Pindel, 43.23% vs. 38.23%
using DELLY and 50.13% vs. 47.88% using LUMPY when compared with
1KG_2015. Those results showed that with two benchmarks, there were higher

[Insert Running title of <72 characters]

191   precisions of MGI platforms using Pindel, DELLY, LUMPY, but lower precisions of
192   MGI platforms using CNVnator.
193
Tool affects CNV detection
To investigate the characteristic of CNV detection by different tools, we firstly
compared the number or the length of CNVs detected by different tools from the same
datasets. We found that there were average 4,582 (range from 4,045 to 5,730) CNVs
from ten datasets using Pindel, which 87.01% (39,869 / 45,819) CNVs were smaller
than 1,000bp (Table 1, Supplementary Figure S3), verified that Pindel is an effective
tool for small CNVs[15]. Meanwhile, we got average 3,207 (range from 1,229
to5,778) CNVs using CNVnator, which 95.78% (30,717 / 32,072) were larger than
1,000bp (Table 1, Supplementary Figure S3), showed that CNVnator with RD-
strategy is more suitable to detect large CNVs. About BreakDancer, DELLY and
LUMPY, we got CNVs with similar average CNV count (average 1,648, 1,818 and
1,673 CNVs for BreakDancer, DELLY and LUMPY, respectively) and similar size
distributions which were major range from 100 to 5,000bp (Table 1, Supplementary
Figure S3).
Secondly, we compared the sensitivity and precision between different tools. As the
sensitivity and precision of CNVs detection described above, we found BreakDancer
showed the highest precision but unstable sensitivity, while CNVnator had both
lowest precision and sensitivity (Figure 2). Meanwhile, Pindel showed higher
sensitivity but lower precision, while DELLY and LUMPY had both balanced
sensitivity and precision (Figure 2).
Further, we studied the region distribution of CNVs across genomic from different
datasets by different tools. We found CNVnator could detected more CNVs in
Exonic, CpG Island (CGI) and CGI-shore than other tools (average 37.04% with
1,150 CNVs in Exonic, 5.43% with 198 in CGI and 8.84% with 291 in CGI-shore,
Supplementary Figure S6, Supplementary Table S5), while BreakDancer and Pindel
could detected more CNVs in Intronic and Intergenic (average 29.36% and 38.58% of
CNVs in Intronic and Intergenic using BreakDancer, 32.03% and 39.00% using
Pindel). No visible bias appeared in CNVs detected by DELLY or LUMPY.
In a word, we found similar results CNV count and length when analyzing different
datasets with the same tool but variety when analyzing the same dataset with different
tools. And distributions of length and genomics regions and performance would be
helpful for more critical selection of CNV detection tool.

Discussion

Detecting CNV using WGS sequenced on Illumina platforms were widely used and
many tools were designed based on those requirements, but the feasibility of CNV
detection on MGI platforms is unclear. This study represents a first systematic
investigation and characterization of CNV detection based on MGI sequencing
platforms and we found that both the sensitivity and the precision of CNVs based on
MGI platforms were consistent with those of CNVs based on Illumina platforms, if
not better. Furthermore, by comparing different CNV tools, we found that there were
different features of CNV detected by different tools, such as CNV count, the length
region distribution across genome, CpG Island (CGI), CGI-shore, sensitivity and
precisionof CNVs, indicating that this study may help researchers to choose tools
according to different conditions.
There are some critical parameters affect the performance of CNV detection. First
parameter is the choice of CNV detection tools. Dozens of CNV detection tools were

[Insert Running title of <72 characters]

available and designed for different sample conditions. We chose five common tools
for CNV detection on single WGS and still got obvious differences. BreakDancer
built on RP could only detect mediocre deletions influenced by the insert size of
library. CNVnator built on RD gets large CNVs with low sensitivity and precision.
Pindel build on SR would get high proportion overlapped CNVs with worse precision.
Those results indicated deficiency of current CNV detection tools and complicated
tool selection for CNV detection.
The second parameter is the lack of a high-quality benchmark. We used two public
benchmarks of NA12878 here. However, high proportion specific CNVs between
benchmarks and opposite comparisons with two benchmarks between sensitivity and
precision showed the influence of benchmark for CNV detection, and the necessity of
a high-quality benchmark.
Finally, the uncertain saturation is another parameter in evaluation performance of
CNV detection. A combination of current tools, benchmarks and data size could
confuse the saturation and performance of CNV detection.
In summary, the various performance of CNV detection on MGI platforms indicate
the situation of CNV detection and the need for developing more efficient and precise
CNV detection and evaluation methods. CNV detection tools with high performance
and efficiency can enhance CNV researches using WGS sequenced on NGS.

Conclusions

WGS sequenced on NGS is an efficient approach for genome researches. The
applies of WGS sequenced on Illumina platforms were widely-used for CNV
detection with numerous achievements. However, the feasibility of CNV detection on
MGI platforms were not clear. Our results indicate CNV detection using WGS
sequenced on MGI platforms is accurate and efficient. Our study provides a
comprehensive guide for CNV researchers using MGI platforms with tool selection,
benchmark, performance.

Materials and Methods

Public data used.

All fastq data of NA12878 were downloaded from website: Gigadb, NCBI and
CNSA (Supplementary Table S2). All data were down-sampled for approximately
30x and processed following out previous WGS approach.
Two public benchmarks of CNVs of NA12878: 2,819 CNVs from Layer, et al. in
2014 (LUMPY_2014)[27] and 2,171 CNVs from the 1000 Genome Project WGS data
in 2015[28] were used. Both benchmarks were extracted from EXCEL and processed
with BEDtools (v2.16.2).

CNV detection using BreakDancer

278   BreakDancer (ver.1.4.5) was used for CNV calling with default parameters with the
279 read-pair strategy. Raw CNVs from BreakDancer were filtered out according to: (1)
280 SV type is not 'DEL', (2) confidence score < 90, (3) supporting read pairs < 3, (4) not
281 autosomal or chrX, (5) overlap a gap in the reference genome,

CNV detection using CNVnator

284   CNVnator (ver.0.3.3) was used for CNV detection using a read-depth strategy. The
285 optimal bin size for each data was chosen according to the authors` recommendations,
286 such that the ratio of the average read-depth signal to its standard deviation was

[Insert Running title of <72 characters]

between 4 and 5 (Supplementary Table S4). Raw CNVs from CNVnator would be

filtered by those factors: (1) $q0 \geq 0.5$ or $q0 < 0$, (2) e-val1 ≥ 0.05, (3) not autosomal or

chrX, (4) overlap a gap in the reference genome,

CNV detection using Pindel

Pindel (ver.0.2.5b9) was used with default parameters with the split read strategy.

Raw CNVs from Pindel were filtered with SVTYPE is not 'DEL' or

'DUP:TANDEM' and supporting read pairs < 3. Also, CNVs with: (1) not autosomal

or chrX, (2) overlap a gap in the reference genome were filtered out.

CNV detection using DELLY and LUMPY

Both DELLY (ver.0.7.8) and LUMPY (ver.0.2.13) were used to detect CNVs with

default parameters with a combination of approaches. CNVs from DELLY were

filtered out if: (1) SVTYPE is not 'DEL' or 'DUP', (2) FILTER is 'LowQual', (3)

supporting read pairs < 3. CNVs from LUMPY were filtered out if: (1) SVTYPE is

not 'DEL' or 'DUP', (2) PE < 3. Also, CNVs with: (1) not autosomal or chrX, (2)

overlap a gap in the reference genome were filtered out.

Filtration of raw CNVs

Firstly, a raw CNV detected by tool was filtered for removing low-quality CNV.

Then, a CNV was defined as ambiguous or unambiguous if it had at least 1bp overlap

with other CNVs in the same CNV result or not. All ambiguous CNVs were filtered

out for decreasing false positive.

Evaluation of CNVs

The evaluation of CNV compared with benchmark was processed by in-house

scripts. A CNV was considered valid if either it overlapped a single CNV in

benchmark by $\geq 50\%$ reciprocally in size, or there existed a set of CNVs in

benchmark such that each CNV in benchmark had $\geq 50\%$ size overlapped with the

CNV and $\geq 50\%$ of the CNV overlapped with this set of CNVs in benchmark.

The sensitivity or precision were calculated using the following equations:

$$\text{Precision} = \text{TP\_tool} / \text{SUM\_tool}$$

$$\text{Sensitivity} = \text{TP\_benchmark} / \text{SUM\_benchmark}$$

where TP_ tool was the number of validated CNVs, SUM_tool was the total

number of CNVs, TP_ benchmark was the number of a set of CNVs in benchmark

that overlapped with validated CNVs, and SUM_ benchmark was the number of

CNVs in benchmark.

Annotation of CNVs

Genomic regions and CGI were downloaded from UCSC

(http://hgdownload.soe.ucsc.edu/goldenPath/hg19/database/). The region of 2,000bp

frank of CGI in each direction were defined as CGI-shore. CNVs were located to any

of seven regions (Upstream, Exonic, Intronic, Intergenic, Downstream, CGI and CGI-

shore) if the midpoint of CNV region was in any regions.

Authors' contributions

333 J.R., L.P., X.L. and F.M. designed the project. J.R. and L.P. carried out the data

[Insert Running title of <72 characters]

334 analysis and wrote the manuscript, with critical comments and suggestions from the
335 rest of the authors. All authors read and approved the final manuscript.

Acknowledgements

337 This research was supported by the National Key R&D Program of China
338 (2017YFC0906501), the Key R&D Program of Guangdong Province
339 (2019B020226001).
340

References

1.   Iafrate AJ, Feuk L, Rivera MN, Listewnik ML, Donahoe PK, Qi Y, Scherer
SW, Lee C: **Detection of large-scale variation in the human genome**. *Nat*
Genet 2004, **36**(9):949-951.
2.   Sebat J, Lakshmi B, Troge J, Alexander J, Young J, Lundin P, Maner S,
Massa H, Walker M, Chi M *et al*: **Large-scale copy number polymorphism**
in the human genome. *Science* 2004, **305**(5683):525-528.
3.   Levy D, Ronemus M, Yamrom B, Lee YH, Leotta A, Kendall J, Marks S,
Lakshmi B, Pai D, Ye K *et al*: **Rare de novo and transmitted copy-number**
variation in autistic spectrum disorders. *Neuron* 2011, **70**(5):886-897.
4.   Sebat J, Lakshmi B, Malhotra D, Troge J, Lese-Martin C, Walsh T, Yamrom
B, Yoon S, Krasnitz A, Kendall J *et al*: **Strong association of de novo copy**
number mutations with autism. *Science (New York, NY)* 2007,
316(5823):445-449.
5.   Pinto D, Pagnamenta AT, Klei L, Anney R, Merico D, Regan R, Conroy J,
Magalhaes TR, Correia C, Abrahams BS *et al*: **Functional impact of global**
rare copy number variation in autism spectrum disorders. *Nature* 2010,
466(7304):368-372.
6.   Chung BH, Tao VQ, Tso WW: **Copy number variation and autism: new**
insights and clinical implications. *J Formos Med Assoc* 2014, **113**(7):400-
408.
7.   Tam GW, Redon R, Carter NP, Grant SG: **The role of DNA copy number**
variation in schizophrenia. *Biol Psychiatry* 2009, **66**(11):1005-1012.
8.   Pankratz N, Dumitriu A, Hetrick KN, Sun M, Latourelle JC, Wilk JB, Halter
C, Doheny KF, Gusella JF, Nichols WC *et al*: **Copy number variation in**
familial Parkinson disease. *PLoS One* 2011, **6**(8):e20988.
9.   Jiang Q, Ho YY, Hao L, Nichols Berrios C, Chakravarti A: **Copy number**
variants in candidate genes are genetic modifiers of Hirschsprung disease.
PLoS One 2011, **6**(6):e21219.
10.   Shlien A, Malkin D: **Copy number variations and cancer**. *Genome Med*
2009, **1**(6):62.
11.   Pinkel D, Segraves R, Sudar D, Clark S, Poole I, Kowbel D, Collins C, Kuo
W-L, Chen C, Zhai Y *et al*: **High resolution analysis of DNA copy number**
variation using comparative genomic hybridization to microarrays.
Nature Genetics 1998, **20**:207.
12.   Pinkel D, Albertson DG: **Array comparative genomic hybridization and its**
applications in cancer. *Nat Genet* 2005, **37 Suppl**:S11-17.
13.   Bentley DR, Balasubramanian S, Swerdlow HP, Smith GP, Milton J, Brown
CG, Hall KP, Evers DJ, Barnes CL, Bignell HR *et al*: **Accurate whole human**

[Insert Running title of <72 characters]

		genome sequencing using reversible terminator chemistry. *Nature* 2008, **456**(7218):53-59.
381		**genome sequencing using reversible terminator chemistry**. *Nature* 2008,
382		**456**(7218):53-59.
383	14.	Zare F, Dow M, Monteleone N, Hosny A, Nabavi S: **An evaluation of copy**
384		**number variation detection tools for cancer using whole exome**
385		**sequencing data**. *BMC Bioinformatics* 2017, **18**(1):286.
386	15.	Zhao M, Wang Q, Wang Q, Jia P, Zhao Z: **Computational tools for copy**
387		**number variation (CNV) detection using next-generation sequencing**
388		**data: features and perspectives**. *BMC bioinformatics* 2013, **14 Suppl**
389		**11**(Suppl 11):S1-S1.
390	16.	Pirooznia M, Goes FS, Zandi PP: **Whole-genome CNV analysis: advances in**
391		**computational approaches**. *Front Genet* 2015, **6**:138.
392	17.	Chen K, Wallis JW, McLellan MD, Larson DE, Kalicki JM, Pohl CS,
393		McGrath SD, Wendl MC, Zhang Q, Locke DP *et al*: **BreakDancer: an**
394		**algorithm for high-resolution mapping of genomic structural variation**.
395		*Nat Methods* 2009, **6**(9):677-681.
396	18.	Quinlan AR, Clark RA, Sokolova S, Leibowitz ML, Zhang Y, Hurles ME,
397		Mell JC, Hall IM: **Genome-wide mapping and assembly of structural**
398		**variant breakpoints in the mouse genome**. *Genome Res* 2010, **20**(5):623-
399		635.
400	19.	Gusnanto A, Wood HM, Pawitan Y, Rabbitts P, Berri S: **Correcting for**
401		**cancer genome size and tumour cell content enables better estimation of**
402		**copy number alterations from next-generation sequence data**.
403		*Bioinformatics* 2012, **28**(1):40-47.
404	20.	Huang J, Liang X, Xuan Y, Geng C, Li Y, Lu H, Qu S, Mei X, Chen H, Yu T
405		*et al*: **A reference human genome dataset of the BGISEQ-500 sequencer**.
406		*Gigascience* 2017, **6**(5):1-9.
407	21.	Galli A, Patch A-M, Nones K, Kazakoff SH, Newell F, Wood S, Leonard C,
408		Holmes O, Xu Q, Addala V *et al*: **Germline and somatic variant**
409		**identification using BGISEQ-500 and HiSeq X Ten whole genome**
410		**sequencing**. *Plos One* 2018, **13**(1).
411	22.	Xu Y, Lin Z, Tang C, Tang Y, Cai Y, Zhong H, Wang X, Zhang W, Xu C,
412		Wang J *et al*: **A new massively parallel nanoball sequencing platform for**
413		**whole exome research**. *BMC Bioinformatics* 2019, **20**(1):153.
414	23.	Zhu FY, Chen MX, Ye NH, Qiao WM, Gao B, Law WK, Tian Y, Zhang D,
415		Zhang D, Liu TY *et al*: **Comparative performance of the BGISEQ-500 and**
416		**Illumina HiSeq4000 sequencing platforms for transcriptome analysis in**
417		**plants**. *Plant Methods* 2018, **14**:69.
418	24.	Abyzov A, Urban AE, Snyder M, Gerstein M: **CNVnator: an approach to**
419		**discover, genotype, and characterize typical and atypical CNVs from**
420		**family and population genome sequencing**. *Genome Res* 2011, **21**(6):974-
421		984.
422	25.	Ye K, Schulz MH, Long Q, Apweiler R, Ning Z: **Pindel: a pattern growth**
423		**approach to detect break points of large deletions and medium sized**
424		**insertions from paired-end short reads**. *Bioinformatics* 2009, **25**(21):2865-
425		2871.
426	26.	Rausch T, Zichner T, Schlattl A, Stutz AM, Benes V, Korbel JO: **DELLY:**
427		**structural variant discovery by integrated paired-end and split-read**
428		**analysis**. *Bioinformatics* 2012, **28**(18):i333-i339.

[Insert Running title of <72 characters]

27. Layer RM, Chiang C, Quinlan AR, Hall IM: **LUMPY: a probabilistic**
framework for structural variant discovery. *Genome biology* 2014,
15(6):R84-R84.
28. Sudmant PH, Rausch T, Gardner EJ, Handsaker RE, Abyzov A, Huddleston J,
Zhang Y, Ye K, Jun G, Fritz MH *et al*: **An integrated map of structural**
variation in 2,504 human genomes. *Nature* 2015, **526**(7571):75-81.
435

Figures

Figure 1. CIRCOS of CNVs. Circos plot depicting the landscape of CNVs on whole
genome. The outermost circle shows the ideogram of chromosomes. The (i) circle
displays the location of two benchmarks: LUMPY_2014 and 1KG_2015. The (ii),
(iii), (iv), (v) and (vi) circle denote the distribution of CNV of ten datasets using
BreakDancer, CNVnator, Pindel, DELLY and LUMPY, respectively.

Figure 2. Evaluation of CNVs. Dot plot shows the sensitivity and precision of ten
datasets using five tools by compared with LUMPY_2014 (left) and 1KG_2015
(right). Ten datasets are marked with different colors, and five tools are marked with
different point types.

Tables

451 Table 1. Statistics of data and CNVs.

Additional files

Supplementary Figure S1. Venn of benchmarks. Venn showing the number and ratio
of specific CNVs in each benchmark with 90.00% threshold (a) or 50.00% threshold
(b).

Supplementary Figure S2. Summary of the distribution and precision of ambiguous
CNVs. (a) shows the proportion distribution of ambiguous in 50 CNV results of ten
datasets using five tools after low-quality CNV filtration. The comparison of precision
between ambiguous and unambiguous CNVs with LUMPY_2014 (b) and 1KG_2015
(c) is display bellow.

Supplementary Figure S3. Length distribution. The distribution of CNV length for ten
datasets using five tools.

Supplementary Figure S4. Evaluation between benchmarks. Comparison of sensitivity
(a) and precision (b) between two benchmarks.

Supplementary Figure S5. Evaluation between platforms. Comparison of sensitivity
and precision between Illumina platforms and MGI platforms with two benchmarks.

[Insert Running title of <72 characters]

474   Supplementary Figure S6. Annotation of CNVs. Histogram shows the location of
475   CVNs in genomic regions, CGI and CGI-shore. CpG Island: CGI. CpG island-shore:
476   CGI-shore.
477
478   Supplementary Table S1. Tools for CNV detection.
479
480   Supplementary Table S2. Demo data.
481
482   Supplementary Table S3. Statistics of CNVs.
483
484   Supplementary Table S4. Bin size in CNVnator.
485
486   Supplementary Table S5. Annotation of CNVs.
487
488

[Insert Running title of <72 characters]

(i)   Benchmark
(ii) BreakDancer
(iii) CNVnator
(iv) Pindel
(v)   DELLY
(vi) LUMPY

a   LUMPY_2014
b   1KG_2015

I   BGISEQ-500_PE100
II   BGISEQ-500_PE150-1
III   BGISEQ-500_PE150-2
IV   DNBSEQ-G400_PE100
V   DNBSEQ-G400_PE150_PCR-1
VI   DNBSEQ-G400_PE150_PCR-2
VII   DNBSEQ-G400_PE150_PCRfree-1
VIII   DNBSEQ-G400_PE150_PCRfree-2
IX   HiSeq2500_PE150
X   NovaSeq6000_PE150

deletion
duplication

	Sequencing strategy	Clean bases (G)	Average sequencing	Mapping rate (%)	Duplicate rate (%)	Coverage (%)	BreakDancer	CNVnator	#CNV Pindel	DELLY	LUMPY
BGISEQ-500_PE100	PE100	113.6	37.44	99.22	2.45	99.12	2,329	5,778	4,924	1,904	2,378
BGISEQ-500_PE150-1	PE150	94.15	29.63	99.59	6.90	98.99	1,646	3,617	4,463	1,624	1,422
BGISEQ-500_PE150-2	PE150	94.20	29.51	99.57	7.49	98.97	2,265	4,241	4,310	2,209	2,202
DNBSEQ-G400_PE100	PE100	94.11	31.09	99.56	2.22	99.09	2,160	4,354	4,688	1,745	1,889
DNBSEQ-G400_PE150_PCR-1	PE150	94.11	30.83	99.54	2.19	98.99	1,576	2,234	4,577	1,902	1,609
DNBSEQ-G400_PE150_PCR-2	PE150	94.11	30.77	99.55	2.24	98.99	1,600	2,209	4,689	1,900	1,533
DNBSEQ-G400_PE150_PCRfree-1	PE150	94.11	31.13	99.60	1.24	98.98	1,519	2,890	4,221	1,705	1,463
DNBSEQ-G400_PE150_PCRfree-2	PE150	94.11	31.07	99.60	1.25	98.99	1,523	2,860	4,172	1,713	1,505
HiSeq2500_PE150	PE150	96.28	31.91	99.31	1.46	99.00	968	2,660	4,045	1,709	1,380
NovaSeq6000_PE150	PE150	93.69	29.30	99.06	5.91	99.06	897	1,229	5,730	1,773	1,350

Note: Pair-end (PE).