# Self-blinking dye restores efficient use of nanobodies in single-molecule localization microscopy

Samrat Basak[1,2], Kaushik Inamdar[3,4], Yoav G. Pollack[5], László Albert[6,7], Daniel C. Jans[3,4], Stefan Jakobs[3,4,8,9], Jörg Enderlein[1,9], Roman Tsukanov[1,10,*] and Felipe Opazo[6,7,11,*]

1. III. Institute of Physics - Biophysics, Georg-August University, 37077 Göttingen, Germany.

2. Department of Chemistry and Center for NanoScience, Ludwig-Maximilians-Universität München, Butenandtstr. 5-13, 81377 München, Germany.

3. Department of NanoBiophotonics, Max Planck Institute for Multidisciplinary Sciences, 37077 Göttingen, Germany.

4. Department of Neurology, University Medical Center Göttingen, 37073 Göttingen, Germany.

5. Institute for the Dynamics of Complex Systems, Georg-August University, 37077 Göttingen, Germany.

6. Center for Biostructural Imaging of Neurodegeneration (BIN), University Medical Center Göttingen, 37075 Göttingen, Germany.

7. Institute of Neuro-and Sensory Physiology, University Medical Center Göttingen, 37073 Göttingen, Germany.

8. Translational Neuroinflammation and Automated Microscopy, Fraunhofer Institute for Translational Medicine and Pharmacology ITMP, 37073 Göttingen, Germany.

9. Cluster of Excellence "Multiscale Bioimaging: from Molecular Machines to Networks of Excitable Cells" (MBExC), Georg-August University, Göttingen, Germany.

10. Multiscale Biology, Faculty of Biology and Psychology, Georg-August-University Göttingen, 37077 Göttingen, Germany.

11. NanoTag Biotechnologies GmbH, 37079 Göttingen, Germany.

* Corresponding Authors´ emails fopazo@gwdg.de; rtsukan@gwdg.de

**Abstract**

Direct stochastic optical reconstruction microscopy (dSTORM) relies on controlled fluorophore blinking to achieve nanometer-scale resolution, yet the field's benchmark dye, Alexa Fluor 647, underperforms when conjugated to nanobodies, limiting the practical use of minimal-linkage labeling strategies. Here, we show that the self-blinking dye JF635b overcomes this limitation by maintaining robust, photostable blinking upon conjugation to nanobodies under buffer-independent conditions. This enables reliable single-molecule localization microscopy without the need for complex switching buffers. Using JF635b-labeled nanobodies, we demonstrate consistent performance across multiple imaging modalities, including wide-field dSTORM, fluorescence lifetime SMLM, and MINFLUX nanoscopy, achieving localization precisions from ~15 nm down to the sub-nanometer regime. In addition, JF635b supports long-term sample preservation and efficient blinking even in pure water, enabling minimally perturbative imaging conditions. Together, these results establish self-blinking dSTORM as a robust and accessible platform for quantitative nanoscopy across experimental contexts.

**Introduction**

Single-molecule localization microscopy (SMLM)[1] enables nanometer-scale localization precision. Among its core implementations, direct stochastic optical reconstruction microscopy (dSTORM)[2] routinely achieves localization precision below 20 nm, but performance critically depends on the fluorophore blinking characteristics. Alexa Fluor 647 (AF647) has long been the benchmark dye, exhibiting robust blinking in buffers such as GLOX-ME[2]. The reductive properties of the blinking buffers alter disulfide bonds in antibodies and nanobodies, weakening them or rendering them dysfunctional. Therefore, careful optimization of buffer composition for each sample is required, especially in multitarget imaging, which often requires compromises, leading to the development of sequential-labeling strategies such as NanoPlex[3] or microfluidics-enhanced Exchange-PAINT[4].

Despite its widespread adoption, dSTORM remains highly sensitive to imaging conditions, with fluorophore blinking typically requiring carefully optimized chemical environments. Small variations in buffer composition, oxygen-scavenging efficiency, or

reducing agents can substantially affect blinking behavior, localization density, and
photostability, posing a major challenge to reproducibility across laboratories and imaging
facilities. These constraints are particularly problematic for multitarget imaging, long
acquisition times, and workflows involving repeated imaging or sample storage.

Linkage error in indirect immunofluorescence (IF) labeling occurs when the primary
antibody (1.Ab) targeting the protein of interest (POI) is labeled with a fluorescent secondary
antibody (2.Ab). Indirect IF displaces fluorophores by up to 25 nm from the POI. Nanobodies
(Nbs) reduce this linkage error to approximately 3 nm, making them ideal probes[5] for SMLM
nanoscopy. However, a practical limitation has emerged: when directly conjugated to
nanobodies, commonly used dyes such as Alexa Fluor 647 often exhibit compromised
blinking behavior and reduced photostability. This undermines the very advantages
nanobodies offer for high-precision SMLM and restricts their use in workflows that rely on
consistent and sustained photoswitching.

Self-blinking dyes provide an elegant solution by spontaneously cycling between
bright and dark states without the need for blinking-reducing buffers. Several such
fluorophores have been explored for super-resolution microscopy, including specialized dyes
for live-cell super-resolution imaging[6], Thioflavin T for fibrils[7], rhodamine-coumarin
hybrids[8], silicon rhodamine analogs for STORM-MINFLUX[9], and rhodamine-based dyes for
SOFI[10].

JF635b is a newly developed, spontaneously blinking rhodamine from the
hydroxymethyl-SiR Janelia Fluor series, described in Holland et al.[11]. It exhibits an excellent
duty cycle, is ideally suited to SMLM applications, and operates without reducing-blinking
buffers, significantly simplifying the imaging workflow and making it compatible with a
wide range of labeling strategies. Although its photophysical properties have been
characterized when conjugated to HaloTags[11], its versatility and potential for SMLM for
optimal blinking and minimal linkage error due to conjugation to different types of Nbs have
not been explored. Here, we benchmark JF635b against AF647 across Ab and Nb conjugates
using wide-field self-blinking dSTORM (sb-dSTORM), confocal FL-SMLM[12,13], and
MINFLUX[14].

**Results**

To define how fluorophore choice impacts the performance of nanobody-based single-molecule localization microscopy (SMLM), we benchmarked the blinking behavior of Alexa Fluor 647 (AF647) and the self-blinking dye JF635b in a direct dSTORM implementation. COS-7 cells expressing TOMM70–EGFP–ALFA were immunostained using two widely applied nanobodies, NbGFP (binding EGFP) and NbALFA (binding the ALFA-tag), each conjugated to either AF647 or JF635b. Samples labeled with AF647 were imaged under standard GLOX–MEA switching conditions, whereas JF635b-labeled samples were imaged in phosphate-buffered saline without the use of chemical blinking buffers (Fig. 1).

Nanobodies conjugated to AF647 exhibited strongly impaired blinking behavior, characterized by reduced photon yields, shortened on-times, and rapid photobleaching, with most emitters bleaching within approximately 5 minutes of acquisition in wide-field SMLM under optimal imaging conditions (Supp. Fig. S1 and Supp. Table S1). Compared to a control condition using conventional indirect immunofluorescence to detect the ALFA-tag, NbALFA-AF647 conjugates produced ~70% fewer localizations, an effect consistently observed for both NbGFP and NbALFA. Although both Nbs display comparable binding affinities, NbGFP-AF647 yielded marginally better reconstructions than NbALFA-AF647, while NbALFA-AF647 showed particularly poor blinking and rapid signal loss, resulting in clearly suboptimal mitochondrial reconstructions (Fig. 1a-b). Importantly, these differences were independent of the target interaction and were consistently observed across nanobody types, indicating that the reduced performance originates from the dye–nanobody conjugate rather than from binding affinity or labeling efficiency.

In contrast, Nbs conjugated to JF635b exhibited stable, sustained self-blinking behavior when imaged in PBS. Individual emission bursts yielded approximately 500–1500 photons with on-times of 40–60 ms and remained detectable over tens of thousands of frames without appreciable photobleaching (Supp. Movies 1–5). Although individual blinking events were shorter-lived than those observed for AF647 under reducing conditions, the markedly higher number of blinking cycles resulted in higher localization densities, enabling faster and more complete reconstructions. As a result, JF635b-labeled Nbs produced continuous, densely sampled super-resolution images with localization precision comparable to or exceeding that obtained with conventional antibody-based labeling, confirming the high photostability and robustness of JF635b under buffer-independent sb-dSTORM conditions (Fig. 1a-d; Supp. Table S1).

**Figure 1. dSTORM using nanobodies conjugated with AF647 or JF635b.** Primary nanobody (1.Nb) NbALFA and NbGFP, each conjugated with 2 AF647 (**a**) or JF635b (**b**), were used to directly label transfected COS-7 cells displaying EGFP and ALFA-tag on their mitochondria. (**c**) As a control, conventional 1.Ab anti-ALFA-tag, revealed by secondary antibodies (2.Ab-AF647). Samples with AF647 were imaged under GLOX-MEA buffer, while samples with JF635b were imaged in PBS. Average localization precisions (NeNa) are indicated on the images. (**d**) Localization per frame was plotted, with the blinking differences between conditions shown as tau ($\tau$) for each curve (details provided in Methods).

To test whether buffer-independent self-blinking nanobody probes are compatible with widely used antibody-based labeling workflows, we next examined secondary nanobody (2.Nb) detection strategies, which offer flexibility, multiplexing capacity, and broad applicability across established primary antibodies[15]. Microtubules in COS-7 cells were labeled using a monoclonal anti-α-tubulin primary antibody, revealed with secondary nanobodies conjugated to two JF635b molecules per nanobody, and imaged by wide-field sb-dSTORM in phosphate-buffered saline (Fig. 2; Supp. Fig. S2 and Supp. Movie 6).

JF635b-labeled 2.Nbs yielded continuous and densely sampled reconstructions of microtubule filaments with a mean localization precision of $12.8 \pm 4.5$ nm (Fig. 2a). The high localization density enabled reliable segmentation of the filament network, revealing a characteristic mesh organization with a mean area of $0.77 \pm 8.89$ µm² (Fig. 2b; segmentation workflow in Supp. Fig. S3). Applying the same labeling strategy to vimentin intermediate filaments using a polyclonal rabbit primary antibody produced similarly robust reconstructions with a localization precision of $15.2 \pm 4.7$ nm and, as expected, a denser network architecture characterized by a smaller mean mesh area of $0.15 \pm 0.60$ µm² (Fig. 2c,d; Supp. Fig. S4 and Supp. Movie 7). Comparable high labeling efficiency and image quality were observed for additional targets, including peroxisomes (Supp. Fig. S5), underscoring the versatility and general applicability of this approach.

124
**Figure 2. Applications of sb-dSTORM across labeling strategies and SMLM modalities.** (**a-d**) Wide-field
sb-dSTORM in COS-7 cells labeled with 1.Ab and 2.Nb-JF635b. (**a, c**) Full reconstruction with zoom-ins of
microtubule and vimentin. (**b, d**) Segmentation analysis for (a, c) and area histograms for all datasets combined
(microtubule, vimentin, respectively).                                   ).

## Self-blinking nanobody probes enable robust imaging in a buffer-independent manner

To assess long-term stability, COS-7 samples labeled with a dye were stored in PBS
at 4°C and re-imaged after extended periods. Over storage times of up to one year, JF635b-
labeled nanobodies maintained stable blinking behavior and localization precision, with only
a modest reduction in labeling density observed after 12 months (Fig. 3a). In contrast, AF647
typically exhibits a significant decay in blinking performance within weeks. The pronounced
stability of Nb–JF635b conjugates enables long-term sample preservation, repeated re-
imaging, and high-throughput data acquisition without the need for buffer exchange or re-
labeling.

We next asked whether the intrinsic blinking behavior of JF635b could support
SMLM under conditions that are typically incompatible with conventional dSTORM
fluorophores. Remarkably, 2.Nb-JF635b conjugates exhibited stable, reversible blinking even
when imaged in pure water, in the complete absence of salts or additives (Fig. 3b; Supp. and
Supp. Movie 8). Despite these minimally perturbed conditions, localization densities and
reconstruction quality remained sufficient for reliable super-resolution imaging, highlighting
the exceptional robustness of the self-blinking mechanism and its compatibility with
chemically sensitive workflows such as expansion microscopy (ExM)[16].

**Figure 3. Long-term storage of the sample and the water performance of the Nb-JF635.** (**a**) dSTORM images of the same sample imaged immediately after labeling (Fresh) and after 3, 6, and 12 months of storage at +4°C in PBS. Blinking performance and localization precision (NeNa) as well as the photon output of 2.Nb-JF635b remained stable throughout the 12 months. Above the distributions of photons/100 ms, the average (x̄) is displayed. (**b**) dSTORM image of microtubules labeled with JF635b and imaged in pure water. The average localization precision (NeNa) and photon output were comparable to or higher than those in PBS.

## Self-blinking nanobody probes are compatible with diverse SMLM modalities

Because different single-molecule localization microscopy (SMLM) modalities impose distinct constraints on fluorophore photophysics, photon budgets, and temporal dynamics, we next evaluated whether self-blinking nanobody probes could serve as a general labeling strategy across complementary localization techniques.

Using the same antibody–secondary nanobody labeling strategy as for wide-field sb-dSTORM, we performed MINFLUX[17] imaging of microtubules and vimentin filaments labeled with JF635b-conjugated nanobodies. Under identical sample preparation conditions, JF635b-supported MINFLUX localization yielded mean combined localization precisions of 0.90 nm for microtubules and 0.97 nm for vimentin (Fig. 4a, b), demonstrating that the spontaneous blinking kinetics of JF635b are fully compatible with high-precision MINFLUX localization.

We further tested the applicability of self-blinking nanobody probes in fluorescence-lifetime SMLM (FL-SMLM)[12] using confocal laser scanning microscopy. JF635b-labeled microtubules produced high-quality lifetime-resolved super-resolution reconstructions with a mean localization precision of 11.9 nm and a fluorescence lifetime of $3.16 \pm 0.67$ ns (Fig. 4c–e). These results demonstrate that JF635b is fully compatible with confocal SMLM while simultaneously providing robust lifetime information, enabling applications such as lifetime-based multiplexing.

Together, these data demonstrate that self-blinking nanobody probes are compatible

with a broad range of SMLM implementations, spanning wide-field, confocal, and

MINFLUX-based localization microscopy, and can serve as a unified labeling strategy across

modalities.

**Figure 4. Evaluating JF635b in MINFLUX and fluorescence lifetime SMLM.** MINFLUX imaging of the
microtubule (**a**) and vimentin (**b**) staining using 2.Nb-JF635b. Combined localization precision plots yield mean
precisions of 0.90 nm (microtubules) and 0.97 nm (vimentin). (**c**) Confocal fluorescence lifetime SMLM
imaging of microtubules labeled with 1.Ab and 2.Nb-JF635b. (**d**) localization precision histogram, and (**e**)
fluorescence lifetime distribution (mean ± s.d. indicated).

## Discussion

dSTORM relies critically on fluorophore blinking, yet most implementations remain

dependent on specialized chemical environments that limit robustness, reproducibility, and

long-term sample stability. Although nanobodies offer clear advantages in labeling accuracy

and epitope accessibility, reducing linkage error from ~20-25 nm for conventional antibodies

to ~3 nm, their practical use in SMLM has been constrained by the poor blinking

performance of commonly used dyes when directly conjugated. Our results identify

fluorophore blinking as a central bottleneck in nanobody-based dSTORM and demonstrate

that this limitation can be effectively overcome using self-blinking fluorophores.

By conjugating nanobodies to the spontaneously blinking dye JF635b, we enable

robust and sustained localization microscopy under buffer-independent conditions. Self-

blinking nanobody probes restore high localization densities and nanometer-scale precision for both direct and secondary labeling strategies, eliminating the need for reducing or oxygen-scavenging buffers. This simplification not only enhances experimental robustness but also lowers technical barriers, particularly in multi-user imaging facilities and high-throughput settings where reproducibility across experiments and users is critical.

A defining feature of JF635b-labeled nanobodies is their exceptional photophysical stability. Stable blinking in pure water and sustained performance after long-term storage for up to one year highlight the intrinsic robustness of the self-blinking mechanism. In contrast to conventional dSTORM approaches, where reactive imaging buffers degrade over time and introduce variability, samples labeled with JF635b can be stored in simple PBS and re-imaged over extended periods with minimal loss of performance. This enables repeated imaging of the same specimen under comparable conditions, supporting longitudinal studies, large-scale comparisons, and the development of standardized reference samples. By removing buffer-related variability, self-blinking dSTORM directly addresses a major limitation in SMLM: reproducibility across samples and laboratories.

Beyond long-term stability, JF635b can sustain efficient blinking even in pure water, expanding SMLM's operational range. Unlike conventional fluorophores that require carefully tuned environments, JF635b blinks reversibly in the absence of salts or additives, enabling imaging with minimal perturbation and simplifying workflows. This also opens opportunities for integration with chemically sensitive techniques such as expansion microscopy, where buffer composition affects structural preservation. Self-blinking nanobody probes work across various SMLM methods, including wide-field self-blinking dSTORM, fluorescence-lifetime SMLM, and MINFLUX, achieving localization precisions below a nanometer. Their cross-modality compatibility makes them a unifying labeling strategy for super-resolution microscopy, providing consistent probe behavior across platforms.

In summary, self-blinking nanobody probes combine minimal linkage error, photophysical robustness, and operational simplicity. By restoring the practical advantages of nanobody labeling and removing buffer dependence, they provide a broadly enabling solution that enhances reproducibility, flexibility, and accessibility in nanoscale super-resolution microscopy.

**Methods**

*Cell culture*

COS-7 cells were cultured in Petri dishes at 37 °C with 5% CO2 inside a humidified incubator. They were grown in complete Dulbecco's Modified Eagle Medium (DMEM, ThermoFisher Scientific), supplemented with 10% fetal bovine serum (FBS, ThermoFisher), 4 mM L-glutamine, and 1% penicillin/streptomycin (ThermoFisher). For sample preparation, cells were seeded onto 18 mm poly-L-lysine-coated coverslips in 12-well plates and incubated at 37 °C with 5% CO2 in a humidified environment. For transfection, a plasmid encoding TOM70-EGFP-ALFA-tag driven by a CMV promoter was mixed with Lipofectamine 2000 (Thermo Fisher) and added to the seeded cells according to the manufacturer's instructions. Transfection and expression were typically left for 16-18h, before cell fixation.

*Immunostaining of cells*

COS-7 cells were first treated for 30 s with pre-warmed Extraction Buffer (EB, 10 mM MES, 138 mM KCl, 3 mM $MgCl_2$, 2 mM EGTA, 320 mM sucrose, pH 6.8) supplemented with 0.1% Saponin. This was directly followed by fixation in a pre-warmed solution of 4% paraformaldehyde and 0.1% Glutaraldehyde in EB at RT for 15 min. After removing the fixatives and rinsing with PBS, unreacted aldehyde groups were quenched using 100 mM glycine in PBS for 15 min. Cells were blocked and permeabilized with 3% serum albumin (BSA) and 0.1% Triton X-100 (Sigma-Aldrich) in PBS for 30 min at RT. For Fig. 1, NbGFP and NbALFA labeled with AF647 or JF635b were purchased as FluoTag®-X2 from NanoTag Biotechnologies (clone 1H1, Cat#: N0304; Cat#: N1502) and used at a 1:1000 dilution for 45 minutes. Recombinant anti-ALFA antibody carrying a mouse IgG1 Fc domain (NanoTag Biotechnologies, Cat# N1582) was used at a 1:500 dilution and detected by Goat anti-Mouse IgG (H+L) Cross-Adsorbed Secondary Antibody, Alexa Fluor™ 647, also used in a 1:500 dilution (ThermoFisher, Cat# A-21235). Microtubules and vimentin structures were labeled using a one-step immunofluorescence[15] with a mouse monoclonal (SySy, Cat# 302211) and a rabbit monoclonal (Abcam, Cat# ab92547), respectively. Secondary detection was performed using FluoTag®-X2 anti-Mouse IgG1 (Cat# N2002) or FluoTag®-X2 anti-Rabbit IgG (Cat# N2402). See Supp. Table S3 for details on the used Abs and Nbs. All samples, after staining and washing, were post-fixed in 4% PFA for 15-20 minutes at room temperature, the fixative was removed, and the samples were quenched again with 100 mM glycine in PBS.

*Wide-field sb-dSTORM imaging and data analysis*

Wide-field sb-dSTORM measurements were performed on a custom-built setup as described

previously[5] and in the Supplementary Information (Supp Figure S1). Regions of interest

(ROIs) of up to $50 \times 50 \ \mu m^2$ were recorded with a pixel size in the object plane of 103.5 nm,

an exposure time of 30 ms, and an EMCCD gain of 500. Each dataset consisted of 50,000

frames. The laboratory temperature was maintained at 23.0°C to ensure thermal stability of

the setup. Linear sample drifts of typically 1-2 pixels were observed during acquisition and

were corrected by a cross-correlation algorithm. Excitation was performed using a highly

inclined and laminated optical sheet (HILO) configuration with a laser power of ~20 mW,

yielding an optimal signal-to-noise ratio. Data analysis has been performed using free

software FiJi[18], plugin ThunderSTORM[19]. Same analysis parameters were applied as in

previous works[4,5]. Photostability was quantified by tracking the number of localizations per

frame over time (Fig. 1d). Localization tables were exported from the ThunderSTORM

plugin as .csv files and processed with a custom Python script to generate photostability plots

(cumulative localizations in time, see Fig. 1d). The photostability curves were fitted with an

exponential decay, from which characteristic photobleaching times were extracted.

Segmentation of reconstructed super-resolution images was carried out using a standard

Python image pipeline. Images were preprocessed using intensity filtering and morphological

operations to ensure filament continuity and remove non-filament objects, and then inverted

to enable analysis of voids rather than filaments. Void regions were then identified (and

scaled to 10.35 nm/pixel), border-adjacent partial objects were removed, and regions were

hole-filled to avoid biasing the measurement of their spacing area. Quantitative parameters,

including void area distributions, were extracted (see Supplementary Information (Fig. S3)

for details).

*Confocal FL-SMLM sb-dSTORM imaging and data analysis*

CLSM measurements were performed using a custom-built confocal microscope as described

elsewhere[12,20] and in the Supplementary Information (Fig. S6). Regions of interest (ROIs) of

$7 \times 7 \ \mu m^2$ were scanned, each subdivided into $70 \times 70$ virtual pixels (100nm/pixel). Similar

image acquisition parameters were used as described previously[12]. During analysis, temporal

binning of four consecutive frames was applied to improve signal-to-noise ratio. Data

analysis has been performed using custom Matlab-based software TrackNTrace

LifetimeEdition[12], freely available at https://github.com/scstein/TrackNTrace/releases/tag/v2.0.

*MINFLUX sb-dSTORM imaging*

MINFLUX sb-dSTORM measurements were performed on a commercial MINFLUX nanoscope (Abberior Instruments GmbH, Göttingen) controlled by *Imspector* software. Data were acquired using the standard 2D acquisition protocol described by Sahl *et al.*[21], see Supplementary Table S2 for details. Gold colloid particles (150 nm; BBI Solutions, catalogue no. EM.GC150/4, Crumlin, UK) were added to the samples as fiducial markers for active beam-path drift correction.

Regions of interest (ROIs) of $5 \times 5$ µm² were acquired over a total measurement time of ~6 h. Excitation was performed with the 642 nm laser at 5-10% output (~25-50 µW) as part of the MINFLUX localization routine. Image data were saved as MSR files and processed in MATLAB (MathWorks, USA).

Localization precisions were calculated as described previously[22]. In brief, localizations from binding events with more than four detections were combined. For enhanced visualization, MINFLUX images were rendered in Blender 3.4.1: raw localizations were displayed as overlapping blue spheres (20 nm) to illustrate the spread of individual events, while combined localizations were rendered as gold spheres (10 nm) to indicate the most likely positions of fluorophores.

**Data Availability**

All data supporting the conclusions of this study are available from the corresponding authors upon reasonable request.

**Code availability**

FiJi: https://fiji.sc/

Fiji ThunderSTORM plugin: https://github.com/zitmen/thunderstorm

TrackNTrace https://github.com/NRadmacher/TrackNTrace/tree/ISM;

Abberior Imspector commercially available from Abberior Instruments GmbH, Göttingen, Germany

**Funding**

J.E. and S.B. acknowledge financial support from the DFG through Germany's Excellence
Strategy EXC 2067/1390729940. J.E. and S.B. are grateful to the European Research Council
(ERC) through the project "smMIET" (Grant agreement No. 884488) under the European
Union's Horizon 2020 research and innovation program. F.O. acknowledges the support by
Deutsche Forschungsgemeinschaft (DFG) through the SFB1286 (project Z04). Y.G.P.
acknowledges support from the DFG, Project-ID 449750155, RTG 2756. S.J. acknowledges
support from the DFG (FOR2848, P04). The MINFLUX system was funded by the DFG
(grant no. INST 186/1303-1 to S.J.).

## Acknowledgements

The authors are grateful to Jannik Hentze for their help with sample preparation and labeling.
The authors thank Dr. Peter Ilgen (Fraunhofer ITMP, Göttingen) for MINFLUX data
visualization using Blender. The authors are grateful to Dr. Antonio Politi (MPI-NAT,
Göttingen) for advice in developing the segmentation analysis pipeline. The authors thank Dr.
Steffen Frey (NanoTag Biotechnologies GmbH) for advice. We thank Prof. Philip
Tinnefeld's lab (LMU, Munich) for support with JF635b blinking experiments in water. The
authors are grateful to Dr. Ingo Gregor for technical assistance.

## Competing interests

343 F.O. is a shareholder of NanoTag Biotechnologies GmbH. All other authors declare no
344 competing interests.

## Additional information

347 Supplementary information is available for this paper.

## References

1. Lelek, M. *et al.* Single-molecule localization microscopy. *Nat Rev Methods Primers* **1**, 1–27 (2021).
2. Dempsey, G. T., Vaughan, J. C., Chen, K. H., Bates, M. & Zhuang, X. Evaluation of fluorophores for
optimal performance in localization-based super-resolution imaging. *Nat Methods* **8**, 1027–1036
(2011).

3. Mougios, N. *et al.* NanoPlex: a universal strategy for fluorescence microscopy multiplexing using
nanobodies with erasable signals. *Nat Commun* **15**, 8771 (2024).
4. Basak, S. *et al.* Versatile Microfluidics Platform for Enhanced Multitarget Super-Resolution
Microscopy. *ACS Nano* https://doi.org/10.1021/acsnano.5c18697 (2026)
doi:10.1021/acsnano.5c18697.
5. Sograte-Idrissi, S. *et al.* Nanobody Detection of Standard Fluorescent Proteins Enables Multi-
Target DNA-PAINT with High Resolution and Minimal Displacement Errors. *Cells* **8**, 48 (2019).
6. Uno, S. *et al.* A spontaneously blinking fluorophore based on intramolecular spirocyclization for
live-cell super-resolution imaging. *Nature Chem* **6**, 681–689 (2014).
7. Yang, Q. *et al.* Self-Blinking Thioflavin T for Super-resolution Imaging. *J. Phys. Chem. Lett.* **15**,
7591–7596 (2024).
8. Jradi, F. M. *et al.* Coumarin as a general switching auxiliary to prepare photochromic and
spontaneously blinking fluorophores. *bioRxiv* 2024.05.12.593749 (2024)
doi:10.1101/2024.05.12.593749.
9. Remmel, M., Scheiderer, L., Butkevich, A. N., Bossi, M. L. & Hell, S. W. Accelerated MINFLUX
Nanoscopy, through Spontaneously Fast-Blinking Fluorophores. *Small* **19**, 2206026 (2023).
10. Grußmayer, K., Lukes, T., Lasser, T. & Radenovic, A. Self-Blinking Dyes Unlock High-Order and
Multiplane Super-Resolution Optical Fluctuation Imaging. *ACS Nano* **14**, 9156–9165 (2020).
11. Holland, K. L. *et al.* A series of spontaneously blinking dyes for super-resolution microscopy. *Nat*
*Methods* 1–5 (2026) doi:10.1038/s41592-026-03062-5.
12. Thiele, J. C. *et al.* Confocal Fluorescence-Lifetime Single-Molecule Localization Microscopy. *ACS*
*Nano* **14**, 14190–14200 (2020).
13. Oleksiievets, N. *et al.* Fluorescence lifetime DNA-PAINT for multiplexed super-resolution imaging
of cells. *Commun Biol* **5**, 1–8 (2022).
14. Balzarotti, F. *et al.* Nanometer resolution imaging and tracking of fluorescent molecules with
minimal photon fluxes. *Science* **355**, 606–612 (2017).
15. Sograte-Idrissi, S. *et al.* Circumvention of common labelling artefacts using secondary
nanobodies. *Nanoscale* **12**, 10226–10239 (2020).
16. Chen, F., Tillberg, P. W. & Boyden, E. S. Expansion microscopy. *Science* **347**, 543–548 (2015).
17. Masullo, L. A. *et al.* Pulsed Interleaved MINFLUX. *Nano Lett.* **21**, 840–846 (2021).
18. Schindelin, J. *et al.* Fiji: an open-source platform for biological-image analysis. *Nat Methods* **9**,
676–682 (2012).
19. Ovesný, M., Křížek, P., Borkovec, J., Švindrych, Z. & Hagen, G. M. ThunderSTORM: a
comprehensive ImageJ plug-in for PALM and STORM data analysis and super-resolution imaging.
*Bioinformatics* **30**, 2389–2390 (2014).
20. Oleksiievets, N. *et al.* Single-Molecule Fluorescence Lifetime Imaging Using Wide-Field and
Confocal-Laser Scanning Microscopy: A Comparative Analysis. *Nano Lett.* **22**, 6454–6461 (2022).
21. Sahl, S. J. *et al.* Direct optical measurement of intramolecular distances with angstrom precision.
*Science* **386**, 180–187 (2024).
22. Ostersehlt, L. M. *et al.* DNA-PAINT MINFLUX nanoscopy. *Nat Methods* 1–4 (2022)
doi:10.1038/s41592-022-01577-1.
395