Quantifying *Drosophila melanogaster* Feeding Behavior Using flyPAD and optoPAD

Nicholas J. Collins[1], Madison Endres[1], Irina Sinakevitch[1], & Lisha Shao[1]
[1] Department of Biological Sciences, University of Delaware, Newark, DE

Address correspondence to: shaol@udel.edu, or nicollin@udel.edu

Keywords: *Drosophila melanogaster*, Feeding behavior, Optogenetics, flyPAD, optoPAD

Summary

Quantifying feeding behavior with high temporal and spatial precision is critical for understanding how internal state, sensory cues, and neural activity shape food intake and dietary choice. Here, we describe a detailed protocol for performing consumption and dietary choice assays in *Drosophila melanogaster* using the flyPAD/optoPAD system. This method enables simultaneous measurement of feeding events across multiple arenas while allowing precise control of sensory and optogenetic stimulation. We provide step-by-step instructions for assay food preparation, flyPAD arena setup, data acquisition, and downstream data organization with suggested analyses. This approach is suitable for studying nutrient preference, learning, and state-dependent modulation of feeding behavior, and can be readily adapted for optogenetic manipulations and comparative choice assays.

Highlights

- High throughput, quantitative measurement of *Drosophila melanogaster* feeding behavior.
- Easily integrates optogenetic stimulation for precise control of neural circuitry.
- Compatible in open-loop and closed-loop (system feedback from the animals behavior) formats.
- Can measure easily interpretable nutrient preference, learning, and state-dependent feeding behavior.
- Manuscript includes step-by-step instructions from arena acquisition and calibration, through statistical modeling.

76
77 **Before you Begin**
78
79 **Background**
80
Feeding behavior in *Drosophila melanogaster* is a widely used and biologically
informative variable for understanding how genetic factors, cellular mechanisms, neural
circuits, and internal physiological states shape behavior. Alterations in feeding have
been used to characterize the effects of social stress (Li et al., 2021), investigate how
nutrient valuation contributes to metabolic disease and obesity (Kelly et al., 2022;
Strilbystka et al., 2022), map associative learning circuits that guide food choice (Coban
et al., 2024), and reveal links between dietary intake, neurodegeneration, and oxidative
stress (Catalani et al., 2021). Because feeding integrates sensory evaluation, motivation,
reward, and metabolic demand, disruptions in this behavior often signal underlying
changes in neuronal function.
91
Accordingly, identifying the neural circuits and cellular processes that regulate feeding
has become a major research focus. Manipulating these pathways can rescue aberrant
feeding phenotypes (e.g., Brüning & Fenselau, 2023; He et al., 2023; Pool et al., 2014;
Rulifson, 2002) and offers a tractable window into how physiological state modulates
decision-making in the fly brain (e.g., Münch et al., 2022; reviewed in Pool & Scott,
2014). However, many existing feeding assays lack the temporal precision or neural
specificity needed to dissect how circuit activity dynamically shapes feeding decisions.
The flyPAD/optoPAD assay addresses this gap.
100
101   As outlined below, flyPAD/optoPAD provides high-resolution detection of feeding on
102   solid food while enabling closed-loop, real-time optogenetic manipulation, making it
103   uniquely suited to reveal how specific neurons influence moment-to-moment feeding
104   behavior.
105
106 **Important Definitions**
107
108   **flyPAD:** For high-resolution, capacitance based detection of feeding on solid food.
109   Measures total consumption and choice between two substrates, all without the use of any
110   optogenetic stimulation or manipulation (Itskov et al., 2014).
111
112   **optoPAD:** Provides the same utility as the flyPAD, with the addition of an optogenetic
113   module for neural circuit manipulation. Also allows for closed-loop assays, whereby the
114   flies own behavior facilitates feedback from the optogenetic module (Moreira et al.,
115   2019).

Materials and Equipment

TIMING: Acquire reagents, consumables, and equipment 1-2 months prior to planned experiments.

Before beginning, the below reagents and consumables (Table 1), as well as equipment (Table 2) should be purchased in advance and set up according to our recommended protocol outlined below. For a summary of equipment and supplies needed to conduct these experiments daily, see Figure 1.

Table 1. Reagents and Consumables

Reagent or Consumable	Source/Vendor	Notes/Quantity/Concentration
Nutrifly® Cornmeal Fly Food	Genesee Scientific	200 g per 1 L water; prepare as per fly food husbandry protocol
Agar	Thermo Fisher	0.7% for food prep, 1% for starvation vials
Sucrose	Fisher Scientific	7% (weight/volume) in assay food
Yeast Flakes	Any standard lab supplier	7% (weight/volume) in assay food
Propionic Acid	Fisher Scientific	4.8 ml per 1 L of food
Tegosept	Genesee Scientific	10 mL per 1 L food
Retinal	Sigma-Aldrich	Optional, 0.4 mL per 100 mL for optogenetic assays
Fly Vials	Genesee Scientific	Filled as per protocol
Eppendorf Tubes (5 mL)	Any standard lab supplier	For assay food; 2.5 mL per tube
Lab Tissue	Any standard lab supplier	For cleaning arenas and vials
Cheese Cloth	Any standard lab supplier	For food cooling overnight
Plastic Film	Any standard lab supplier	For food storage

Table 2. Equipment

Equipment	Source/Vendor	Notes/Purpose
FlyPAD 4 Fly Arena Module	www.flypad.rocks	Each module has 8 sensitive electrodes, two per fly chamber. These allow precise detection of feeding behavior.
LED Optogenetic Addon Module	www.flypad.rocks	LED module for open and closed-loop optogenetic stimulation.
LED Controller Optogenetic Addon Module	www.flypad.rocks	LED controller that gates power to the LEDs contingent on the real time feeding data (closed-loop) or preprogrammed stimulation pattern(open-loop stimulation).
FlyPAD Feeding Multiplexor	www.flypad.rocks	PC interface and an acquisition board for up to 12 of 4-fly arena modules.
Mug Warmer	Amazon	Keep food at 55–75 °C during experiments.
Incubator	Any lab supplier	25°C, 50% humidity, 12H light/dark cycle
Pipettes and Pipette tips (1; 1000μL)	Any lab supplier	Used to make fly aspirators, load food into assay.
Computer with Bonsai and flyPAD analysis software from the supplier	www.bonsai-rx.org	Bonsai for data acquisition, www.flypad.rocks provided software for data analysis.

Setting up and procurement of the flyPAD Units and optoPAD LED modules

TIMING: 1-2 months prior to start of experiment

This section describes acquisition, initial setup, and pre-experimental validation of the flyPAD/optoPAD system. Early setup is critical to allow sufficient time for hardware troubleshooting, software configuration, and pilot testing prior to data collection.

1. Procurement

    a. Obtain flyPAD units, sensor boards, control electronics, optogenetic LED modules, power supplies, and all required cables from the manufacturer.

    b. Confirm with manufacturer that purchased units are compatible with both open-loop and closed-loop optogenetic configurations if desired.

    c. Ensure that replacement units, covers, screws, and cables are available prior to beginning experiments.

2. Initial Hardware setup

    a. Place the flyPAD unit on a vibration-minimized, level benchtop in a temperature- and humidity-controlled behavioral testing room.

    b. Connect flyPAD sensor boards and optogenetic control modules to the acquisition computer according to manufacturer specifications.

    c. Secure units evenly to sensor boards, ensuring that all contact points are flush and screws are tightened uniformly to prevent signal drift.

3. Software installation and configuration

    a. Install Bonsai (behavioral acquisition software) and required flyPAD/optoPAD packages on the desired laboratory computer.

       NOTE: The correct version of Bonsai to install is provided by the manufacturer.

    b. Verify correct COM port assignments for flyPAD sensors and optogenetic controllers (port numbers may vary between systems).

    c. Load open-loop and closed-loop Bonsai workflows and confirm that all nodes execute without errors.

4. Pre-experimental calibration and validation

    a. Run a baseline channel test with empty arenas to confirm stable capacitance signals across all channels (Figure 2).

    b. Load agar-only food (1% agar in distilled water) into test wells and initiate a short test recording to verify that all channels show dynamic, non-flat traces (see Figure 2 and Figure 3).

i.   Click on "Start"

ii.   Double click on "flyPAD sensors data".

iii.   Verify all lines are in a zig-zag fashion.

iv.   If lines are straight, seek troubleshooting.

5. Temperature and Humidity

    a.   Maintain testing room temperature at 22–25°C and relative humidity between 40-55%, as environmental variability can affect both feeding behavior and sensor sensitivity.

    b.   Conduct experiments in darkness or under controlled lighting conditions to prevent unintended visual stimulation.

CRITICAL: Allocate sufficient time for pilot experiments to confirm reliable feeding engagement, sensor stability, and accurate synchronization of optogenetic stimulation before collecting experimental data.

Cooking, dispensing, and storing standard food for husbandry

TIMING: 3-4 weeks prior to experiment; 1-2 hours.

1. Cooking fly food for husbandry (Base volume: 1L). Scale as necessary.

    a.   Add 1L of distilled water and 200g of Nutri-Fly® to a large pot.

    b.   Place pot on an electric stove top, set on high heat (simmer setting, approximately between 85-96□. Stir continuously to avoid food sticking to the bottom of the pot.

    c.   Once the mixture reaches boiling point, put the burner to a higher heat (approximately 100□), and continuously stir for an additional 10-15 minutes.

    d.   Take off the food from the heat. Turn off the heat, and allow the food to cool to 70□ with a continuous stir.

    e.   Add propionic acid (4.8 mL) and Tegosept (10 mL) as preservatives. Stir food to evenly distribute ingredients.

2. Dispensing

    a.   It is recommended to use a rapid dispensing system (see equipment below)

i. For husbandry vials, fill to ~2-3cm depth.

ii. For collection vials (for collecting flies prior to beginning assay) fill to ~1-2cm.

3. Storage

   a. Cover newly created food vials with cheese cloth and allow food to solidify overnight (6-8H).

   b. Food can then be stored in 4☐ for approximately 1 month.

      NOTE: Do NOT insert vial plugs during cooling.

      NOTE: Cover tightly with plastic film and cover in a plastic bag.

Retinal Food for husbandry and optoPAD

TIMING: ~2 weeks prior to experiment

1. Cooking Retinal food (optogenetic experiments)

   a. Prepare retinal stock solution

      i. Dissolve all-trans retinal in ethanol to make a 100mM stock solution.

   b. Follow the standard food preparation steps outlined above.

2. Addition of Retinal to standard food

   a. Once food reaches a temperature of 60☐, turn off all of the lights in the laboratory. Dim ambient lighting may be used.

   b. Add the appropriate amount of retinal.

      i. Husbandry vials: 1 mL retinal stock solution per 500 mL food.

      ii. Collection vials: 1 mL retinal stock solution per 250 mL food.

3. Repeat standard food protocol for cooling, dispensing, and storage.

CRITICAL 1: Place retinal food into a light-sensitive plastic bag to prevent degradation.

CRITICAL 2: You may prepare 0.4mM retinal-containing food for husbandry (1:500 ratio). This is recommended if flies will be collected from these vials and immediately used in behavioral experiments. For collection vials, you may use 0.2mM retinal

288 containing food (1:250 ratio), and flies should remain on this food for at least 3 days prior
289 to the experiment.
290
291 **Preparing Experimental Flies**
292
293 **TIMING:** ~2 weeks prior to experiment
294
295 1. Maintain newly acquired stocks (e.g., from the Bloomington Stock Center) in
296 husbandry vials as described above in a 25☐ incubator.
297
298 a. Temperatures between 18☐ to 27☐ can be used to slow or speed up
299 development, respectively.
300
301 b. Fly incubators should be set to allow for a 12H dark/light cycle.
302
303 c. Humidity should be set to 50-55%.
304
305 2. Transfer flies to a new vial with food every 3-5 days, to prevent overcrowding and to
306 easily identify virgin progeny (if desired).
307
308 3. Flies in the assay may be:
309
310 a. Wild-type strains (*Canton-S, w1118*, etc.) or
311
312 b. Progeny from Gal4/UAS crosses, including crosses for optogenetic
313 experiments (see protocol below).
314
315 4. Ideal sample collection conditions:
316
317 a. 20-25 flies per group is often sufficient sample size for most applications of
318 the flyPAD/optoPAD assay.
319
320 i. Run a power analysis if desired to confirm adequate sample size.
321
322 b. Collected flies can be kept in collection vials prior to beginning the
323 experiment.
324
325 c. For most experiments, flies should be between 3-7 days old.
326
327 **CRITICAL:** Do NOT use CO_2 within 24h of beginning the flyPAD experiment, as this
328 can influence fly behavior. Instead, use a fly aspirator to transfer flies into the
329 appropriate arenas.
330
331
332
Preparing Experimental Food for flyPAD and/or optoPAD Assays

TIMING: 1 week prior to experiment; 15 minutes.

The flyPAD/optoPAD assay allows the experimenter to choose any two food types for
presentation to the fly. Each food type used in our lab-including its specific ingredients,
preparation instructions, and storage conditions-is described individually below. Sucrose-
based food is provided as the primary example. Any deviations from this procedure (steps
omitted or added) for other food types are indicated in Table 3.

1. Preparation of sucrose food

a. To a microwave-safe beaker, add 10 mL of hot distilled water (up to 95□),
0.70g sucrose (200mM) and 0.07g Agar (0.7%). For consistency, place the
mixture on the mug warmer on 55-75□, stirring every 15s until all
components are completely dissolved.

b. Pour sucrose food into a 5 mL Eppendorf tube, filled approximately half-way
(approximately 2.5 mL).

c. This will create 4 tubes of food, which is good for 8 experiments.

d. Ensure food solidifies within 1 hour at RT.

2. Food can be stored at RT for approximately 2 weeks or reheated twice.

Table 3. Experimental flyPAD/OptoPad Food Recipes

Food Type	Ingredients	Assay
Yeast and Sucrose	10 mL distilled water 7% (0.70g) yeast 200 mM sucrose (.70g) 0.76% (0.76g) Agar	Consumption (open-loop)
Yeast (protein)	10 mL distilled water 7% (0.70 g) yeast 0.7% (0.07 g) agar	Choice (open-loop) Learning (closed loop)
Sucrose (sugar)	10mL distilled water 200mM sucrose (.70 g) 0.7% (.07 g) agar	Choice (open-loop) Learning (closed loop)

364 **Innovation**

The flyPAD/optoPAD feeding assay provides several advantages over existing feeding paradigms such as Con-Ex (blue dye), café, and FLIC (Deshpande et al., 2014; Ro et al., 2014; Itskov et al., 2014; Ja et al., 2007). flyPAD/optoPAD combines high-resolution, capacitance-based detection of feeding on solid food with closed-loop, millisecond-precision optogenetic control, enabling neural circuit perturbations that occur strictly during active feeding bouts. Unlike Con-Ex, which provides accurate but low-resolution measures of total consumption, or FLIC, which offers high-resolution detection of liquid-food interactions, the flyPAD/optoPAD systems allows researchers to causally link neuronal activity to microstructural feeding decisions in real time on solid food.

This integration of precise behavioral readout and temporally locked circuit manipulation makes the flyPAD/optoPAD system uniquely suited for dissecting how specific neurons drive feeding preference, consummatory microstructure, and the moment-to-moment dynamics of ingestive behavior.

Limitations

While the flyPAD/optoPAD offers several advantages, there are also some limitations worth noting. First, the assay only allows for the choice between two different food sources, which may not be representative and engage the full behavioral repertoire. Second, the chambers are only in theory designed to study the behavior of individual flies; it is not practical to use the device to study feeding behavior in group dynamics. Third, the assay is mainly optimized for studying short-term feeding behavior within an hour. As such, it is critical the experimenter considers the importance of circadian rhythms, which are known to influence overall consumption patterns in *Drosophila* (Xu, Zheng & Sehgal, 2008)

Institutional Permissions

There are no specific institutional permissions required, other than standard bio-safety training as directed by local and federal requirements.

Table 4: Key Resources

Reagent Type	Reagent or Resource	Source	Identifier
Chemicals	Agar, Pure, Powder	Thermo Scientific	400402500
Chemicals	D-Sucrose	Fisher BioReagents	BP220-1
Chemicals	Yeast Flakes	Genesee Scientific	Cat#62-106
Chemicals	Quinine Hydrochloride Dihydrate	Sigma-Aldrich	22630
Chemicals	All trans-retinal	Sigma-Aldrich	R2500-1G
Chemicals	All trans-retinal	Spectrum Chemical	R3041-1GM

		Mfg. Co.	
Chemicals	Propionic Acid	Fisher Scientific	A258-500
Chemicals	Tegosept	Genesee Scientific	Cat#20-258
Chemicals	dH2O		
Drosophila medium	Nutri-Fly Pre-Mixed Fly Food	Genesee Scientific	66123
Equipment	flyPAD and optoPAD highthroughput feeding assay	Itskov *et. al,* 2014	https://flypad.rocks/
Equipment	Insect Rearing Chamber	Darwin Chambers Company	IN023
Software, algorithms	MATLAB R2024 software	The Mathworks, Inc.	https://www.mathworks.com
Software, algorithms	Bonsai	Manufacturer	www.bonsai-rx.org
Software, algorithms	BioRender	BioRender	https://biorender.com
Drosophila research	Flystaff Drosophila Vials	Genesee Scientific	32-113
Drosophila research	Droso-Plugs	Genesee Scientific	902064
Glassware	Beaker 250 mL	Fisher Scientific	50-199-7455
tube	5 mL Eppendorf Tubes	Genesee Scientific	24285
Eppendorf tips	Pipette Tips	Genesee Scientific	24121R
Eppendorf pipette	Eppendorf Research plus 1-channel	Fisher Scientific	TI13690034
Drosophila research	Cheese Cloth	Genesee Scientific	53-100
General consumable	Plastic Wrap	Fisher Scientific	15-610
General consumable	Kim Wipes	Genesee Scientific	88115
General consumable	Scissors	Amazon	amazon.com
General consumable	Mouse pads	Amazon	amazon.com
General consumable	Aluminum Foil	Fisher Scientific	01-213-102
Equipment	Computer	Any standard supplier	
Equipment	Microwave	Any standard supplier	

Equipment	Mug Warmer	Amazon	amazon.com

Preparing data files prior to assay

1. Establishing a directory for saved data

   a. Open the "Bonsai" software on the laboratory PC that is connected to the flyPAD unit(s).

   b. Choose the relevant protocol you wish to run (provided from manufacturer).

      i. Ex.: "OpenLoopSimple"

   c. Double click on "matrix-writer"

      i. A panel should open up to the right on the screen.

   d. Where it says "path", select the ellipsis.

   e. Select the directory where you want your file to be saved.

   NOTE: It is recommended to save each experiment in its own folder.

2. Naming the File (see Figure 4)

   a. Files will be named according to experimental conditions (number of groups and number of flies per group).

      i. For example, C01_01_08_C02_09_16

      ii. This would suggest that you have two groups (C01 and C02).

      iii. Each one of those groups will utilize 4 chambers (01_08 refers to EACH flyPAD channels, so 01 and 02 combined make up the first chamber).

3. Preparing a text/log file for analysis

   a. In the experimental folder, make a .txt file

   b. Each group will be labeled using the following convention:

      i. Events.ConditionLabel{1} = 'Group 1 information'

      ii. Events.ConditionLabel{2} = 'Group 2 information'

c. Each substrate (food) will be labeled with the following convention:

        i.   Events.SubstrateLabel{1} = 'food pipetted on the LEFT in each chamber'

        ii.   Events.SubstrateLabel{2}= 'food pipetted on the RIGHT in each chamber'

CRITICAL 1:. How you label the data file indicates where each group is located and what conditions they are exposed to for the analysis. However, files can be renamed after the experiment is run.

CRITICAL 2: Be sure to remember and log which food of interest is pipetted where and keep this consistent *within* each experiment. You may alternate food sides *between* experiments.

Behavior Protocol: flyPAD total consumption or two-choice

For the flyPAD, experimenters may run an "open-loop" formatted experiment. In this type of experiment, the fly is allowed to openly feed between two choice points. This type of experiment is ideal for studying:

a. Total consumption: baseline feeding behavior on a standard food (yeast/sucrose).
b. Choice: baseline choice feeding behavior between two different types of food.

For example, you may wish to determine the total amount of food a particular fly line consumes over the 50 minute testing period (Figure 5), or if the fly prefers protein vs sucrose (Figure 6)

1. Pre-Experiment PC preparation

    a. Turn on the PC that is associated with your flyPAD rig.

    b. Open up the "Bonsai" software.

    c. Select your protocol of choice, in this case choose an open-loop protocol.

    d. Create a directory and .txt file as directed above.

2. Warming food for the assay

    a. Place 5 mL tubes of sucrose and protein and on a microwave-safe tray, with the lids open.

b.   Fill a beaker (200 mL) with ~150 mL of water.

c.   Microwave the tubes and the beaker for approximately 1 minute.

    i.   While microwaving check every 15-20 seconds.

    ii.   Sucrose and protein will be done when they are in liquid form and warm to the touch.

d.   Place tin foil over the water-filled beaker and push the tubes with your food through to the other side. Store tubes in warm water (see Figure 1).

    i.   Place the beaker with tubes on the mug warmer at 55 □ during experiments.

    ii.   This is to keep the food warm and liquified while you are working with it.

CRITICAL: You do NOT want the food to solidify too early in the paradigm, as it makes it difficult for the flies to consume. Ensure that the food is indeed warm, in a liquid state, and the water remains warm to the touch while working with it.

3.   Loading food into flyPAD channels (Substrate A; Left Channel)

a.   Set a 10μL pipette to 10μL.

b.   Load desired food into the LEFT channel within each arena.

    i.   Each channel doesn't get a precise amount of food; your goal here is to provide a base layer of the food.

    ii.   We recommend ~3.33μL into each (3 chambers per pipette full).

CRITICAL: While pipetting the food, you should NOT be getting any food on the brass section of the arena, as this will interfere with accuracy of results and require more troubleshooting (see below).

4.   Loading food into flyPAD channels (Substrate B; Right Channel, see Figure 5).

a.   Repeat step 3 for your alternate (two-choice assay) or same (consumption assay) substrate for the right channel.

NOTE: If running a total-consumption experiment, you should load the same substrate in BOTH channels (Figure 6).

CRITICAL 1: The sucrose food is transparent, so avoiding the brass becomes more

difficult. As such, you may put a 1:200 ratio of blue dye into the sucrose food if necessary. Our lab has determined no significant differences with fly feeding between sucrose or sucrose with dye containing foods (data unpublished but available upon request).

CRITICAL 2: Please ensure you are changing your pipette tip after different foods to prevent cross-contamination.

5.   Add an additional layer of food of each substrate to create a dome shape.

     a.   Pipette 1.25μL of substrate A and B over each (respectively).

     b.   See Figure 7 for an illustrative example.

6. Pre-Experimental Validation

     a.   See step 4 of the setting up and procurement of the flyPAD Units and optoPAD LED modules section.

     b.   Once you confirm all capacitance lines are moving and wavy, select "stop" on the bonsai software.

          i.   If any are flat, wipe up excess food using a lint-free cloth and re-check.

CRITICAL: If an entire row is horizontal, this indicates a likely connection issue with that particular choice equipment and the rig. Please see troubleshooting section below.

7. Load flies into each arena using a fly aspirator.

     a.   Ensure you load flies as per your experimental conditions.

     b.   To load flies:

          i.   Partially open each arena.

          ii.   Tilt your fly aspirator such that the fly is pushed toward the back of the arena.

          iii.   Immediately close chamber shut when flies are loaded.

8. Starting the experiment

     a.   Hit "start" on the bonsai software to initiate the experiment.

     b.   Set a lab timer for 50 minutes.

c. Ensure the chambers are exposed to minimal lighting (darkness preferable)

d. If you wish to run multiple experiments in the same day:

i.   Place the protein and sucrose food onto the mug warmer, set to 55□.

ii. Alternate sides of presented food between experiments as a control.

9. Ending the experiment

a. After 50 minutes, press "stop" in Bonsai to end the experiment.

b. Re-collect the flies for additional tests or discard them based on institutional

parameters.

10. Clean-up and resetting for another experiment

a. To clean, use a sterile 10μL pipette tip and insert it into each choice hole to

push the food out.

b. Use lightly damp lint-free tissue paper with distilled water to wipe up the

residual food. You should also do this on the upside-down portion of the

arenas and cover.

c. Once complete, ensure the arenas are dry using and clear of all food residue.

OptoPAD Protocols (open-loop: total consumption and two-choice)

This section provides detailed methodology on performing a total consumption or choice

experiment, where an experimenter wishes to determine the flies' baseline consumption,

or preference after neuronal activation or inhibition of food consumption between two

different food types. This type of experiment is ideal for studying:

a. How activation or inhibition of specific cells impacts feeding behavior in real-time.

For example, you may wish to determine whether your particular fly line has a preference

for protein or sucrose when a subset of neurons or glia cells are activated or inhibited

with the LED module.

TIMING: 15 minute preparation, plus 50 minute experiment, with 15 minutes of

cleanup/reset.

CRITICAL: Ensure flies have been raised on retinal food (1:500) or placed in a

collection retinal vial (1:250) for 3 days prior to beginning the experiment.

1. Repeat steps 1 (PC preparation) through 6 (pre-experimental validation) above from the "open-loop choice assay" protocol.

   a. Use either the total consumption or two-choice formats as desired.

2. When loading flies for optoPAD:

   a. You should include a non-LED control, or a subset of flies not exposed to light. See Figure 4 for an example experimental schematic.

   b. Example of groups:

      i. Non-LED - variable A level 1

      ii. Non-LED - variable A level 2

      iii. LED- variable A level 1

      iv. LED- variable A level 2

3. Situating the optogenetic modules for stimulation

   a. Place the provided optogenetic module(s) over each desired LED exposed arena.

   b. Turn the LED electronic module "ON". You can set the voltage based on what percent LED stimulation you would like.
      i. EX. 3V: 50%

   c. Cover the non-LED groups using mousepads or similar to avoid any light saturation.

CRITICAL SAFETY WARNING: Do NOT look directly into the LED lights as this may cause visual impairment.

4. Repeat steps 8 (starting the experiment) through 10 (cleaning up) from the above flyPAD total consumption/two-choice protocol.

OptoPAD Closed-Loop Assay

Another major type of experiment routinely performed on the optoPAD is a closed-loop experiment. In a closed-loop experiment, the fly's own behavior is used as an input to modulate system output. For example, you may perform an experiment whereby every time a fly takes a sip of a particular substrate, there is consequential LED delivery, and real-time circuit manipulation. Usually, synchronization between light and food consumption is achieved on the left side: the LED light turns on when the fly drinks from

the left, and remains off on the right, and vice versa.

As opposed to open-loop, the closed-loop behavior is best to test:

a. If flies can learn to avoid or prefer a particular food option.
b. If neural activity alters decision-making at the moment of choice.

1. Pre-Experiment PC Preparation for "Left ON" experiment

    a. Open the Bonsai software.

    b. Select the OptoPAD program, "Octopad_B2_4_NewVersion_Blinking _v2".

    c. When saving a directory:

        i. Create a new folder named "closedloop_date".

        ii. In this folder, create two subfolder files, one "LEFT ON" and the second "RIGHT ON."

        iii. Use naming conventions for conditions and flies per arena as described above.

2. Setting the LED to only trigger in the left channel

    a. Identify the node in Bonsai corresponding to the feeding-event stream from the left channel.

        i. Configure the workflow so that feeding events from the left channel are routed to the LED.

        ii. This will cause feeding interactions in the left channel to trigger LED activation.

    b. Identify the node corresponding to the feeding-event stream from the right channel.

        i. Configure the workflow so that feeding events from the right channel are recorded only and not connected to the LED digital output mode.

3. Repeat steps 2 (warming food) through 10 (cleaning up and resetting above).

CRITICAL 1: Wells should likely contain the same tastants, but this depends on experimental hypotheses and the research question.

4. Run "Right ON" experiments by following steps 1-3 above, except switching left and

722   right channel instructions.
723

724   **NOTE:** It is recommended to run both "Left ON" and "Right ON" experiments for each
725   condition you plan to test, and averaging results to account for any intrinsic side bias.
726

727   **Data Processing**
728

729   Data will be processed using data analysis software provided by the manufacturer. Below, we
730   will describe a workflow using MATLAB, but the same data handling and analytical concepts
731   can be applied to other software. We will start from processing the raw data and conclude with
732   recommended statistical applications.
733

734   1. Preparing the directory for data analytical software.
735

736   a. Open the desired directory on the PC where you saved your data.
737

738   b. There will be two data files and your .txt file.
739

740   i.   The larger file (~58MB) is your data file from the 50 min experiment,
741   and the one you need for analysis.
742

743   ii. Ensure it is named appropriately as described above.
744

745   iii. The smaller file (1-5MB) is your pre-experimental validation file. This
746   data file needs to be deleted and/or removed from this directory.
747

748   2. Data Analytics
749

750   a. Open your desired analytical software.
751

752   b. Use manufacturer recommended analytical tools for each experimental
753   protocol above.
754

755   i.   Analytical tools for various analytical software are provided by the
756   manufacturer.
757

758   c. Data will be saved in the original directory where the experiment was
759   originally saved.
760

761   i.   Navigate to the experimental folder(s) you wish to analyze.
762

763   ii. Open the "mrep" folder.
764

765   iii. Open "DatainExcelFormat".
766

767   3. Output metrics from the assay are described in Table 5.

768
769 **Table 5. Data Output and Interpretation**

Output	Description	Interpretation/Use
Spill Quantity	Fraction of time of the experimental recording that the signal was saturated due to contact with the top electrode	Reflects food contacting electrodes due to improper loading or flies spilling the food during feeding.
Removed Non Eaters	Number (or proportion) of flies excluded because they did not meet a minimum feeding threshold	Flies that never fed or fed below a defined cutoff (2 or less activity bouts on both electrodes during the whole duration of the experiment).
Sip Durations	Duration (ms) of a single feeding contact	Proxy for ingestion strength or palatability. Together with Inter Sip Interval define the frequency of feeding (feeding rhythm)
Inter Sip Intervals (ISI)	Time between consecutive sips within a feeding episode	Short ISI = sustained feeding; long ISI = disengagement.
Number of Sips	Total count of feeding events per fly	Proxy for total intake
Activity Bout Number	Number of episodes of active exploration of food sources	Feeding initiation frequency
Activity Bout Duration	Median duration of active exploration of food sources	Sustained engagement with food
Feeding Bursts Number	Number of feeding bursts per fly. A feeding burst is a sub-behavior within an activity bout, characterized as rapid sequences of sips with very short ISIs	Micro-pattern of ingestion
Feeding Bursts Duration	Duration of each feeding burst	How intensely flies feed within bouts
Activity Bout IBI	Time between the end of one bout and the start of the next	Reflects satiety, motivation, or learning
Feeding Burst IBI	Pause between feeding bursts (within a bout).	Short feeding burst IBI's may be interpreted as increased motivation for a particular food source.

770
771 **Data Analysis: flyPAD or optoPAD (two-choice example)**

772
773 Below are specific instructions for analyzing a flyPAD or optoPAD experiment where a fly was
774 presented with a choice between either protein or sucrose. Please note that these are mock data

775 used for instructional purposes and not actual data from any real experiments. In addition to
776 looking at raw number of sips, below is a recommendation for creating a preference index.
777
778 **Preference Index**
779
780 A preference index, as opposed to looking at raw number of sips, has several advantages. Due to
781 several extraneous variables not limited to but including fly genetics, time of day, vial
782 conditions, and/or housing parameters, there may be baseline differences in feeding behavior. A
783 preference index normalizes the data to allow more interpretable comparisons between
784 experiments.
785
786 A preference index will take the form of: $\frac{(NumberSipsA - NumberSipsB)}{NumberSipsA + NumberSipsB}$
787
788 Such that:
789
790   1.0= Full preference for Substrate A
791   0.0 = No preference
792   -1.0 = Full preference for Substrate B
793
794   1. Calculating the Preference Index
795
796   a. Open the "Number of Sips" panel within the excel document.
797
798   b. For each group, apply the preference index formula above between your first and
799   second substrate (Table 7).
800
801   i. In the formula, substrate A is typically the substrate of interest.
802
803 **Table 6. Number of Sips Example**

Group 1, yeast	Group 2, yeast	Group 1, sucrose	Group 2, sucrose
456	345	283	383
673	309	118	258
173	286	51	489
430	76	386	226

804
805
806
807   2. Example for calculating a protein preference index:
808
809   a. Subtract each sucrose value from the yeast value.
810
811   b. Divide this by the addition of the total number of sips for both yeast and sucrose.

Table 7. Calculating a protein preference index

Group 1, protein (yeast)	Group 1, sucrose	Formula	Protein Preference Index Value
456	283	(456-283) / (456 +283)	0.23
673	118	…	0.70
173	51	…	0.55
430	386	…	0.04

3. Statistical Applications

    a. Analysis Pipeline 1: Comparing the protein preference index values from group 1 and group 2 to examine group level differences in preference.

        i. Independent samples t-test (one group, two-levels).

            - Ex. Male vs female flies on protein preference.

        ii. Two-way ANOVA (two groups, two-levels)

            - Ex. Sex (male vs female) and starvation condition (fed vs starved) on protein preference.

    b. Analysis Pipeline 2: Comparing the preference value vs 0 to examine the absolute preference of each group.

        i. One-sample t-test against 0.

            - Ex. Despite group 1 (M = 0.10) having a significantly increased protein preference compared to group 2 (M = 0.0), both groups have absolute no preference (not significantly different from 0.0).

4. Other potentially useful calculated output metrics are featured in Table 8.

Table 8: Calculated metrics from raw output, their description, and interpretation

Metric	Calculation	Description	Interpretation
Preference Index	(SipA-SipB) / (SipA + SipB)	Relative feeding choice between two	1.0 = A preference 0.0 = No preference

		substrates	-1.0 = B preference
Vigor	(SipA,B) / Duration Sip A,B	Measure of feeding intensity (tempo)	Higher values indicate stronger engagement with the substrate, higher urgency
Average sips per activity bout	(SipA,B) /ActivityBoutsA,B	Measure of feeding continuity	Higher values indicate longer, sustained engagement with the substrate

flyPAD and optoPAD Troubleshooting

1. Flies are not engaging with the assay or show very low feeding activity.

   a. Food has dried or solidified before or during the assay.

      i. Ensure assay food remains warm and liquid during loading and forms a dome-like shape in each well.

      ii. Low feeding motivation→ Increase dry starvation to 2–4 h prior to assay.

   b. Flies were exposed to CO_2 within 24 h of the experiment.

      i. Avoid CO_2 anesthesia within 24 h; transfer flies using a fly aspirator.

   c. Room temperature or humidity is outside the optimal range.

      i. Perform assays at 22–25 °C and ~40–60% relative humidity.

2. The food dries or runs out during the assay.

   a. Low ambient humidity

      i. Ensure consistent (40-50%) room humidity.

   b. Extended preparation and loading food time before experiment start.

      i. Practice sufficiently to ensure this time is minimized. We recommend 10 minutes from loading the food to starting the experiment at maximum.

   c. Certain food substrates may dry faster than others.

i. Load food in a consistent order and alternate sides between experiments.

d. Experimental time is too long.

i. The assay is optimally 30-50 minutes in length, only.

3. Flat/horizontal channels while starting assay.

a. Food is touching the brass electrodes.

i. Remove excess food with a piece of lab tissue.

ii. Repeat the pre-test until all lines are wavy/zig-zag.

4. Entire row of channels is inactive.

a. Loose or improperly connected flyPAD arena.

i. Ensure the connection to the COM port is secure.

5. The LED lights are not functioning properly.

a. LED module is not properly seated or there are faulty connections.

i. Reposition the module to ensure LEDs are securely attached.

b. Voltage box turned off or not properly set.

i. Verify voltage output (3V or as desired) prior to starting the experiment.

6. Flies showing intrinsic bias for one probe/channel in a choice assay or closed-loop assay.

a. Check if the chamber/arena is leveled. Adjust the length of screws at corners to adjust the level.

b. Check if light is not evenly shed on the chamber in regular assay and if there is light bleed through in optogenetic assay.

c. If bias persists, switch substances position in alternate trials and average the preference index of two back-to-back trials.

7. Issues with Data analysis, error messages.

a. Files are incorrectly formatted or labeled (log file or bonsai file), Verify file
naming matches the arena layout and experimental conditions.

b. Ensure the log file contains the correct groups and substrates.

c. Ensure the "test" file used to troubleshoot the capacitance detectors is deleted
from the directory.

d. Ensure your length of experiment time entered matches how long the behavior
was recorded.

Expected Outcomes

Feeding behavior in *Drosophila melanogaster* is a sensitive behavioral output that at baseline is
dependent on sex, mating status, social conditioning, age, and genotype (e.g., Liu et al., 2024).
Table 9 is a general table that outlines expected feeding outcomes across multiple variables.

965 **Table 9: Expected Outcomes**

Variable	Expected outcome on feeding behavior	References
Age (young vs old)	Total feeding generally	Galenza et al., 2016

	declines with age. Sucrose diet increases lifespan.	Wong et al., 2009
Sex (male vs female)	Females (mated) generally consume more total food, and more protein, than males.	Camus et al., 2018 Vargas et al., 2010
Female mating status (virgin or mated)	Mated females generally have a higher preference for protein, and eat more, compared to virgins.	Camus et al., 2018 Ribeiro & Dickson 2010
Male mating status	Mixed results. Virgin males generally have an increased protein intake, which declines after ejaculation, and then is restored.	Liu et al., 2024 Camus et al., 2018
Nutritional History/Diet	Starvation or dietary restriction, and macronutrient specific diets affects total feeding,choice behavior, fecunditiy, and fitness.	Liu et al., 2024 Itskov & Rubeiro, 2013 Lee et al., 2008
Temperature	At lower temperatures, *Drosophila melanogaster* prefers plant foods as opposed to yeast.	Brankatschk et al. 2018
Genetic Background	Common control lines (*Canton-S, W118, Oregon-R)* show baseline differences in feeding behavior.	Camus et al. 2018 Wong et al. 2009 Lee et al. 2008

Acknowledgements

We would like to thank Tarandeep Singh Dadyala, Sumaya Noor Smithy, and Laura White for helpful discussions and feedback on experimental design and protocol optimization. We would also like to thank the manufacturer, Pavel Itskov, for providing technical details when we were writing the manuscript. This work is supported by a Maximizing Investigators' Research Award to L.S. (National Institute of General Medical Sciences R35GM147504).

Author Contributions

All authors were involved with conceptualizing, planning, and reviewing all versions of the manuscript. N.J.C: Primary writer and editor for all versions of the manuscript.
M.E.: Figures, resource tables, reviewing all versions of the manuscript.

980 I.T.S.: Figures, resource tables, writing- closed-loop protocol, reviewing all versions of the
981 manuscript. L.S.: Supervision, funding acquisition, writing – reviewed, edited, and provided
982 feedback on all versions of the manuscript. All authors have read and approved the final version
983 of the manuscript.
984
985 **Declaration of Interests**
986
987 The authors have no conflict of interests to disclose.
988
989 **Data Availability**
990
991 All data provided in this manuscript are mock datasets for educational and instructional
992 purposes. For inquiries regarding the data analysis software and associated tools used with the
993 flyPAD/optoPAD system, readers should consult the manufacturer documentation provided with
994 the equipment, or contact the corresponding authors for further information.
1021 **Figure Legends**
1022
Figure 1. Required materials for the flyPAD assay. (A) A fly aspirator is used to transfer a
single fly into a flyPAD chamber. The aspirator is constructed from plastic tubing, cheese cloth,
and a pipette tip. (B) A 10μL pipette and tips are used to load the appropriate food media into

each flyPAD channel (white arrow). Approximately 3μL of media is loaded per channel. (C)
Food media are stored in 5 mL Eppendorf tubes prior to heating. They are then heated and
maintained in a hot water bath during the loading process. (D) A standard mug warmer is used to
keep the food media in a liquid state between experiments, reducing the need for repeated
reheating. (E) The flyPAD unit (solid red rectangle) is used for experimentation. One fly is
loaded into each flyPAD chamber (dashed white rectangle). The transparent top panel of each
chamber can slide open to allow loading of food media and flies during assay preparation.
1033
Figure 2. Key features of the Bonsai setup and assay initiation. Shown is the typical
appearance of a Bonsai workflow file. Critical parameters that should be edited prior to assay
initiation are indicated by dashed boxes. MatrixWriter" specifies where the Bonsai output file
will be saved once the assay is stopped, and the "Path" field can be modified to direct the file to
the appropriate storage location. "flyPADSensorsData" should be opened during media loading
to confirm that all channels are functioning properly. After these settings have been configured
for the experiment, the "Start" button can be clicked to initiate the assay.
1041
Figure 3. Example of flyPAD sensor data. During loading of food media into the flyPAD
channels, sensor data should be monitored to verify proper loading. Each panel corresponds to an
individual flyPAD channel within a chamber and is numbered accordingly. After successful
media loading, sensors should display a characteristic zig-zag pattern, indicating that sensor
function has not been compromised by the media. In contrast, a flat signal (dashed box) suggests
that media has spilled onto the sensor, which can interfere with subsequent data analysis. To
avoid this issue, channels should be cleaned thoroughly and media loaded carefully. Once all
sensors exhibit the zig-zag pattern, media loading is complete and the assay can be initiated.
1050
Figure 4. Example of Fly/optoPAD choice assay setup and corresponding log file. (A)
Schematic of a flyPad experiment with four experimental conditions looking at two different
genotypes (Genotype A or B) and two illumination states (LED on or off). Each flyPAD channel
contains one of two food substrates (e.g. yeast on the right, sucrose on the left) with one fly
loaded per chamber. Continuous LED illumination (e.g. optogenetic stimulation) is applied to the
top row of chambers, while the bottom row remains unilluminated. Each genotype is tested under
both LED conditions (example shown: n=4 per condition). Condition labels (C0#) correspond to
entries in the Bonsai and log files. The numeric ranges (e.g. 01_08) indicate the channel numbers
assigned to each condition. Channels are numbered from left to right, starting with 01 at the
bottom left. (B) Example log file corresponding to the experimental setup in (A) to be used as a
reference for downstream MATLAB analysis. Events.ConditionLabel {#} specifies the genotype
and LED condition for each group. This may also include additional factors such as: fly sex,
mating status, starvation condition, etc. The number of condition labels in the log file must match
those defined in the Bonsai file. Events.SubstrateLabel {1} and {2} indicate the substrates
loaded on the left and right channels of each chamber, respectively. If only one substrate is
tested, only SubstrateLabel {1} is required.
1067
1068 **Figure 5. General procedure for open-loop flyPAD assay to measure total consumption.** (A)
1069 Food media is liquified by microwaving for ~30s in 10s intervals. The media should not reach a
1070 boiling state. The media is then maintained in a hot water bath to prevent solidification prior to
1071 loading. (B) Approximately 3μL of food media is pipetted into each flyPAD channel, taking on a

dome-like shape. For total consumption measurements, identical media is loaded into both channels. (C) A single fly is loaded into each flyPAD chamber using a fly aspirator through a small opening in the transparent top panel. The chamber is then fully sealed, and the assay can begin.

Figure 6. General procedure for open loop flyPAD assay to measure two-choice. (A) Each food media is liquified by microwaving for ~30s in 10s intervals. The media should not reach a boiling state. The media are then maintained in a hot water bath to prevent solidification prior to loading. (B) Approximately 2-3μL of food media is pipetted into each flyPAD channel, taking on a dome-like shape. For choice experiments, one media type should be consistently loaded on the left channel, and the other media on the right channel. (C) A single fly is loaded into each flyPAD chamber using a fly aspirator through a small opening in the transparent top panel. The chamber is then fully sealed, and the assay can begin.

Figure 7. Proper loading of food media into assay chambers. Using a micropipette, dispense ~2-3μL of liquified food media into the center of each channel. In longer assays spanning 50 minutes, the media should form a raised, dome-like meniscus (A), which enables effective interaction with the food by the fly. In shorter assays spanning 30 minutes or less, food can take on a flat shape within the channel, without spilling onto the sensor (B). If the media spreads into a flat layer that spills onto the electrode (C), the data will not be viable for any assay. If this occurs, wipe the sensor with a lab tissue and add ~1μL of additional media to the top until a domed shape is achieved.

Figure 8. Closed-loop experiment pairing optogenetic neural manipulation with a specific food source. In closed-loop experiments, the LED is paired with the electrode associated with one designated food type. (A) When the LED is paired with the left electrode, interactions with the right food source do not trigger illumination. However, when the fly interacts with the left food source, the LED is activated to induce optogenetic neural manipulation. (B) Conversely, when the LED is paired with the right electrode, interactions with the right food source activate the LED to induce optogenetic neural manipulation, while interactions with the left food source do not trigger illumination.

References

Brankatschk, M., Gutmann, T., Knittelfelder, O., Palladini, A., Prince, E., Grzybek, M., Brankatschk, B., Shevchenko, A., Coskun, Ü. and Eaton, S. (2018). A Temperature-Dependent

Switch in Feeding Preference Improves Drosophila Development and Survival in the Cold.
Developmental Cell, 47(2), pp.257–259. doi:https://doi.org/10.1016/j.devcel.2018.10.010.

Brüning, J.C. and Henning Fenselau (2023). Integrative neurocircuits that control metabolism
and food intake. *Science*, 381(6665). doi:https://doi.org/10.1126/science.abl7398.

Camus, M.F., Huang, C.-C., Reuter, M. and Fowler, K. (2018). Dietary choices are influenced by
genotype, mating status, and sex in Drosophila melanogaster. *Ecology and Evolution*, 8(11),
pp.5385–5393. doi:https://doi.org/10.1002/ece3.4055.

Catalani, E., Fanelli, G., Silvestri, F., Cherubini, A., Del Quondam, S., Bongiorni, S., Taddei,
A.R., Ceci, M., De Palma, C., Perrotta, C., Rinalducci, S., Prantera, G. and Cervia, D. (2021).
Nutraceutical Strategy to Counteract Eye Neurodegeneration and Oxidative Stress in Drosophila
melanogaster Fed with High-Sugar Diet. *Antioxidants*, 10(8), p.1197.
doi:https://doi.org/10.3390/antiox10081197.

Çoban, B., Poppinga, H., Rachad, E.Y., Geurten, B., Vasmer, D., Rodriguez Jimenez, F.J.,
Gadgil, Y., Deimel, S.H., Alyagor, I., Schuldiner, O., Grunwald Kadow, I.C., Riemensperger,
T.D., Widmann, A. and Fiala, A. (2024). The caloric value of food intake structurally adjusts a
neuronal mushroom body circuit mediating olfactory learning in *Drosophila. Learning &*
memory (Cold Spring Harbor, N.Y.), [online] 31(5), p.a053997.
doi:https://doi.org/10.1101/lm.053997.124.

Deshpande, S.A., Carvalho, G.B., Amador, A., Phillips, A.M., Hoxha, S., Lizotte, K.J. and Ja,
W.W. (2014). Quantifying Drosophila food intake: comparative analysis of current methodology.
Nature methods, [online] 11(5), pp.535–540. doi:https://doi.org/10.1038/nmeth.2899.

Galenza, A., Hutchinson, J., Campbell, S.D., Hazes, B. and Foley, E. (2016). Glucose modulates
Drosophila longevity and immunity independent of the microbiota. *Biology Open*, [online] 5(2),
pp.165–173. doi:https://doi.org/10.1242/bio.015016.

He, L., Wu, B., Shi, J., Du, J. and Zhao, Z. (2023). Regulation of feeding and energy
homeostasis by clock-mediated Gart in Drosophila. *Cell Reports*, [online] 42(8), p.112912.
doi:https://doi.org/10.1016/j.celrep.2023.112912.

Itskov, P.M., Moreira, J.-M., Vinnik, E., Lopes, G., Safarik, S., Dickinson, M.H. and Ribeiro, C.
(2014). Automated monitoring and quantitative analysis of feeding behaviour in Drosophila.
Nature Communications, 5(1). doi:https://doi.org/10.1038/ncomms5560.

Itskov, P.M. and Ribeiro, C. (2013). The Dilemmas of the Gourmet Fly: The Molecular and
Neuronal Mechanisms of Feeding and Nutrient Decision Making in Drosophila. *Frontiers in*
Neuroscience, 7. doi:https://doi.org/10.3389/fnins.2013.00012.

Ja, W.W., Carvalho, G.B., Mak, E.M., de la Rosa, N.N., Fang, A.Y., Liong, J.C., Brummel, T.
and Benzer, S. (2007). Prandiology of Drosophila and the CAFE assay. *Proceedings of the*

National Academy of Sciences, 104(20), pp.8253–8256.
doi:https://doi.org/10.1073/pnas.0702726104.

Kelly, K.P., Alassaf, M., Sullivan, C.E., Brent, A.E., Goldberg, Z.H., Poling, M.E., Dubrulle, J.
and Rajan, A. (2022). Fat body phospholipid state dictates hunger-driven feeding behavior. *eLife*,
[online] 11, p.e80282. doi:https://doi.org/10.7554/eLife.80282.

Lee, K.P., Simpson, S.J., Clissold, F.J., Brooks, R., Ballard, J.W.O., Taylor, P.W., Soran, N. and
Raubenheimer, D. (2008). Lifespan and reproduction in Drosophila: New insights from
nutritional geometry. *Proceedings of the National Academy of Sciences*, [online] 105(7),
pp.2498–2503. doi:https://doi.org/10.1073/pnas.0710787105.

Li, W., Wang, Z., Syed, S., Lyu, C., Lincoln, S., O'Neil, J., Nguyen, A.D., Feng, I. and Young,
M.W. (2021). Chronic social isolation signals starvation and reduces sleep in Drosophila.
Nature, 597(7875), pp.239–244. doi:https://doi.org/10.1038/s41586-021-03837-0.

Liu, C., Tian, N., Chang, P. and Zhang, W. (2024). Mating reconciles fitness and fecundity by
switching diet preference in flies. *Nature Communications*, [online] 15(1), pp.9912–9912.
doi:https://doi.org/10.1038/s41467-024-54369-w.

Moreira, J.-M., Itskov, P.M., Goldschmidt, D., Baltazar, C., Steck, K., Tastekin, I., Walker, S.J.
and Ribeiro, C. (2019). optoPAD, a closed-loop optogenetics system to study the circuit basis of
feeding behaviors. *eLife*, [online] 8, p.e43924. doi:https://doi.org/10.7554/eLife.43924.

Münch, D., Goldschmidt, D. and Ribeiro, C. (2022). The neuronal logic of how internal states
control food choice. *Nature*, 607(7920), pp.747–755. doi:https://doi.org/10.1038/s41586-022-
04909-5.

Pool, A.-H., Kvello, P., Mann, K., Cheung, Samantha K., Gordon, Michael D., Wang, L. and
Scott, K. (2014). Four GABAergic Interneurons Impose Feeding Restraint in Drosophila.
Neuron, 83(1), pp.164–177. doi:https://doi.org/10.1016/j.neuron.2014.05.006.

Pool, A.-H. and Scott, K. (2014). Feeding regulation in Drosophila. *Current Opinion in
Neurobiology*, 29, pp.57–63. doi:https://doi.org/10.1016/j.conb.2014.05.008.

Ribeiro, C. and Dickson, B.J. (2010). Sex Peptide Receptor and Neuronal TOR/S6K Signaling
Modulate Nutrient Balancing in Drosophila. *Current Biology*, 20(11), pp.1000–1005.
doi:https://doi.org/10.1016/j.cub.2010.03.061.

Ro, J., Pak, G., Malec, P.A., Lyu, Y., Allison, D.B., Kennedy, R.T. and Pletcher, S.D. (2016).
Serotonin signaling mediates protein valuation and aging. *eLife*, 5.
doi:https://doi.org/10.7554/elife.16843.

Rulifson, E.J. (2002). Ablation of Insulin-Producing Neurons in Flies: Growth and Diabetic
Phenotypes. *Science*, 296(5570), pp.1118–1120. doi:https://doi.org/10.1126/science.1070058.

Strilbytska, O., Semaniuk, U., Bubalo, V., Storey, K.B. and Lushchak, O. (2022). Dietary Choice
Reshapes Metabolism in Drosophila by Affecting Consumption of Macronutrients.
Biomolecules, 12(9), p.1201. doi:https://doi.org/10.3390/biom12091201.

Vargas, M.A., Luo, N., Yamaguchi, A. and Kapahi, P. (2010). A Role for S6 Kinase and
Serotonin in Postmating Dietary Switch and Balance of Nutrients in D. melanogaster. *Current*
Biology, [online] 20(11), pp.1006–1011. doi:https://doi.org/10.1016/j.cub.2010.04.009.

Wong, R., Piper, M.D.W., Wertheim, B. and Partridge, L. (2009). Quantification of Food Intake
in Drosophila. *PLoS ONE*, 4(6), p.e6063. doi:https://doi.org/10.1371/journal.pone.0006063.

Xu, K., Zheng, X. and Sehgal, A. (2008). Regulation of Feeding and Metabolism by Neuronal
and Peripheral Clocks in Drosophila. *Cell Metabolism*, 8(4), pp.289–300.
doi:https://doi.org/10.1016/j.cmet.2008.09.006.

1198

A.

C03_17_24 - *Genotype A, LED*   C04_25_32 - *Genotype B, LED*

C01_01_08 - *Genotype A, No LED*   C02_09_16 - *Genotype B, No LED*

B.

```
Events.ConditionLabel{1}= 'Genotype A, No LED';
Events.ConditionLabel{2}= 'Genotype B, No LED';
Events.ConditionLabel{3}= 'Genotype A, LED';
Events.ConditionLabel{4}= 'Genotype B, LED';
Events.SubstrateLabel{1}= 'Substrate A';
Events.SubstrateLabel{2}= 'Substrate B';
```

A.

30 sec

B.

7% yeast
food

C.

Fly aspirator

A.

30 sec

B.

7% yeast food

200 mM sucrose food

C.

Fly aspirator

A. Dome shape within sensor ✓ Sufficient for 50-min assay

B. Flat shape within sensor ✓ Sufficient for 30-min or less assay

C. Flat and spilled onto electrode ✗ Not viable for any assay

A. Left ON

B. Right ON