Validation and optimisation of wearable accelerometer data pre-processing for digital measure implementation and development

Joss Langford[1,2*], Jia Ying Chua[2], Ian Long[3], Alan C. Williams[2], Melvyn Hillsdon[1]

[1] Department of Public Health and Sport Sciences, University of Exeter, Exeter, United Kingdom

[2] ActivInsights Limited, Kimbolton, Cambridge, United Kingdom

[3] Stoat Group Limited, Devon, United Kingdom

* Corresponding author

E-mail: jl1515@exeter.ac.uk (JL)

Abstract

The increasing use of accelerometers as digital health technologies in clinical trials and clinical care is driving the need for data processing to meet medical standards. The aim of this study was to create and test a modular pipeline for the pre-processing of high-resolution accelerometry that assures the quality, transparency and traceability of digital measures from sensor-level data. The objective is for the pipeline to be a foundational layer in the development, implementation and comparison of measures.

The study developed the open GENEAcore package to meet the requirements of regulators, verifying the engineering implementation and analytically validating outputs against reference datasets. Early stages included the optimisation of calibration and non-wear detection. Data-driven detection of behavioural transitions was then validated to give direct bout outputs without the need to identify rules for epoch aggregation and interruptions. The utility for measure development was shown by comparing two algorithms for the characterisation of activity intensity in both the epoch and bout paradigms.

Non-wear was detected with a balanced accuracy of 92.3% and the commonly used 13mg acceleration standard deviation threshold was empirically validated for the first time. The detection of transitions proved reliable with 99% detected, on average, within 2 seconds of their occurrence to give a mean expected event duration of 68.6s from a log-normal distribution. The different activity intensity algorithms were more than 99% concordant during movement but their outputs diverged in low movement conditions. Importantly, variable duration bouts created 31% higher daily activity durations compared to 1-second epochs.

This evaluation of pre-processing steps has confirmed the attention to detail required to create robust and reproducible results for later clinical validation where small changes in an algorithm or its implementation may have clinically meaningful consequences.

Introduction

Overview

Public health and physical behaviour research have been significantly advanced by the advent of wearable raw data accelerometers [1]. Increasingly, these sensor-level measurement systems are used in commercial research to evaluate medical products as digital health technologies [2] and to guide clinical care [3]. In contrast to consumer-grade devices with embedded algorithms, the original measurement outputs are preserved, allowing full verification of the sensor engineering [4] as well as futureproofing for new algorithms.

Wearable sensors make it possible to collect high-resolution information about an individual's physiology, behaviour and daily routines over long periods with low burden to participants and healthcare professionals. Unlike traditional episodic measures (for example, from clinic visits), these data are gathered continuously in real life. They can be more sensitive, timely, and representative of someone's complete health and behaviour [5]. As well as reducing measurement error (misclassification), these objectively recorded data eliminate the recall and desirability response biases that come with relying on people's memory, subjective judgement or self-report [6]. These biases, particularly recall [7] and desirability [8], can lead to over- or under-estimates of time in active behaviours that in turn attenuates associations with health outcomes [9].

However, challenges remain in processing the large volumes of data from objective measurement into digital health measures that are meaningful to both patients and clinicians [10]. To be trusted, digital measures need to be accurate, reliable and clearly associated with health outcomes in the clinical evidence [11]. They also need to be produced through a process that is transparent and repeatable with an understanding of any potential risk to safety [12]. Regulators also insist on full traceability of the intended use throughout the development, implementation and operation of the system [13].

As raw data technologies move from research settings into clinical applications, there is

growing pressure to make the way we handle data both more efficient and more trustworthy.

Clear and consistent pre-processing, where raw sensor data are prepared before further

analysis, is then an essential foundation [14]. Standardised pre-processing strengthens

interoperability between devices and facilitates reproducibility of algorithm implementation. It

lowers costs, speeds up development, and lays the groundwork for regulatory approval, which

in turn supports clinical adoption and reimbursement.

In this work, we describe and test a pre-processing framework for wearable

accelerometer measurements (Fig 1).

Fig 1. Generic sensor data processing pipeline. A representation of a generic sensor data

processing pipeline showing the scope and aims of this work in the context of classified physical

behaviours, digital measures and the level of inference required for interpretation.

The framework uses open-source software in an integrated test environment to extend

verification from the original device measurements through to the digital measures output from

the system. It is a fast, flexible layer for the implementation, assessment and analytical

validation of digital measures.

The scope of this work covers the first 3 stages of Fig 1, transforming raw sensor data

into contiguous periods of pre-processed data, ready for classification into physical behaviours

(for example, sleep or walking). Our objective in this early processing is to keep assumptions to

a minimum, and to document them carefully when they are required. We want to create a

transparent, reproducible and data-driven process to give the strongest possible foundation for

more complex later stages. While both heuristic and machine learning approaches are essential

tools for analysing digital measures in the context of health outcomes, the acceleration

measurements directly represent real-world behaviours and standard units (SI) can often be

preserved for reproducibility and interoperability.

The overall aim of the study is to define an open, modular pipeline (*GENEAcore*) for the

pre-processing of high-resolution accelerometry that can assure the quality and traceability of

raw, sensor-level data as they are successively refined into digital measures. The process must

be transparent and efficiently deliver essential summary information concerning data source

integrity, interoperability and quality. This manuscript will document the verification of the

engineering implementation and the analytical validation of outputs against reference datasets.

Pre-processing pipeline

There are several examples of processing pipelines for accelerometry data openly

available for use by researchers. Some rely on heuristic assumptions (GGIR [15]) or large

machine learning datasets (OxWearables [16]) while others are focused on integrating data

from multiple devices (NiMBaLWear [17], SKDH [18]). They are fully integrated packages,

designed to calculate complex measures and summary reports using a range of technical

approaches. Yet, they are all configured to output final digital measures rather than to assist the

comparison and development of measures in a verified software environment which can satisfy

the needs of clinical regulators. They do not support bout characterisation from raw data and

can have opaque rules for managing imputation and interruptions. All have the common pre-

processing elements of calibration, non-wear detection and basic measure characterisation

which have been simplified into the *GENEAcore* package.

Returning to the first stage shown in Fig 1, the measurement period information

contains all essential metadata including directions for correct file reading, unique data source

identifiers to link records, calibration of the data to maximise sensitivity, determination of data

completeness and the assessment of non-wear. The accurate determination of wear time is

critical to understanding both participant adherence and data loss due to operational

constraints [19]. A review of 14 non-wear detection algorithms has provided a consolidated

approach with good accuracy and time resolution (S1 Table). In this study, we demonstrate the

traceability of file handling, the performance of the non-wear detection approach and formally

validate a widely used threshold for non-movement.

The second step in the pipeline is the decomposition of continuous time series data into

a contiguous series of time windows. This splits the sensor-level data into manageable chunks

for characterisation, analysis and visualisation. Fixed-duration epochs are traditionally used as

they were simple to implement on early wearables and are convenient for researchers [20].

However, there is no intrinsic relationship between the events and the physical behaviours

under investigation because the timing is arbitrary. The selection of the epoch duration is also

difficult; too short and there is not enough data to characterise a behaviour effectively, too long

and they contain multiple behaviours. Changes in estimates of time spent at various activity

levels with different epoch lengths are well reported [21, 22] acknowledging interactions with

activity level cutpoints [23] and the subsequent impacts on public health guideline adherence

[24] and associations with related outcomes [20, 25]. Variable duration events [26] based on

transitions [27] identified from the variability of the data itself gives an alternative approach.

This method has the advantage of ensuring that different, consecutive behaviours are not

summarised into a single time window - improving accuracy and reducing confusion in later

classification stages.

The concept of a bout was originally used in public health guidelines based on a

minimum duration of physical activity that was linked to health benefits [28]. Increasingly,

bouts are used to analyse all types of physical behaviours (for example, movement patterns

[29]) with variable minimum durations [30, 31] and different bout criteria [32]. When working

with epochs, bouts need to be reconstructed with, often, complicated rules for epoch inclusion

and allowances for interruptions. Events provide a more natural transformation to behavioural

bouts but may identify more interruptions than long epochs depending on the minimum event

duration specified. In this study, we determine optimal parameters for data-driven event

detection and investigate the resulting physical behaviour event durations. We discuss how

event approaches may help to resolve the challenges with epoch length selection and bout rules.

In the third step of Fig 1, the sensor-level measurements are processed to characterise

the epoch or event. This involves the application of mathematical operations to each of the

individual sensor measurements to describe the data in some way. Often, several features

(ideally, independent) will be extracted for a multi-dimensional perspective. There are many

standard, open algorithms that can be applied [33], and the pipeline flexibly supports the

verified implementation of novel, raw data functions. In this study, we will show the utility of

our framework by comparing two commonly used techniques for estimating activity intensity in

both the epoch and event paradigms.

The pre-processing pipeline is designed to establish consistent inputs for the remaining

2 steps beyond the scope of this work. Digital measures may be based directly on the data

characteristics at this stage (for example, mean activity intensity) or rely on further

manipulation and modelling to infer physical behaviours that would be recognisable to an

observer (for example, sleep or walking). Ultimately, these measures are often aggregated into

summaries of daily living activities over multiple days to represent the lifestyle of the

participant (for example, total active duration).

Without pre-processing as a distinct and independent layer, the validation of a digital

measure becomes inseparable from its base assumptions and calculation processes. A digital

measure then cannot be compared independently of the package that created it. The risk is that

we become locked into a plurification of end-to-end processing pipelines, such as those

mentioned earlier, for every digital measure or have different interpretations of an individual's

physical behaviours depending on the pre-processing used.

Methods

The study used triaxial acceleration and temperature data from the GENEActiv Original

(ActivInsights Ltd, Kimbolton, United Kingdom) in the raw output '.bin' format that contains all

measurements and recording information. These were processed in the open-source R package,

GENEAcore (https://cran.r-project.org/package=GENEAcore), hosted by CRAN (The R

Foundation, Vienna, Austria). The work was completed by a cross-functional team of academic,

industry and specialist software personnel.

Engineering verification of the software was completed through unit and integration

testing with design controls (as described in ISO62304[34]) and algorithm code inspection by

relevant experts in the team. Development for the package was undertaken across Windows

and macOS utilising R Studio and its associated R package development libraries *dev tools*,

usethis and *testthat*. Source code control, package building and continuous integration testing

were performed using GitHub and GitHub Actions. Unit testing utilised independently calculated

values or reference results from previously published calculations and packages providing 90%

code coverage by file. Builds were tested across a range of platforms including Windows, macOS

and Linux. Package build and final release test approval is completed by automated build

routines in clean, independent environments. The measure outputs were validated against the

GGIR and *GENEAread* R packages (also hosted by CRAN) to ensure equivalence.

Verification testing was performed using multi-day, wrist-worn datasets that included

68 participants and 69 unique devices from 100 files. Analytical validation testing was

completed with objective and observed criterion measure data for 36 participants from 65 files

(Table1). The datasets were selected from available resources to include a wide range of ages

and a reasonable gender balance.

Table 1. Verification and validation datasets.

Dataset	Measurement period & data	Number of files	Mean age	Male %
Verification testing				
Randomly selected participants from the HeLP study [35]	9 days @ 85.7Hz	33	9.75 yrs	49%

Randomly selected participants from the SafeHeart study [36]	17 days @ 50Hz 30 days @ 20Hz	17 17	64 yrs	81%
Sequential recording periods from the lead author of this study	15 days @ 100Hz	33	49.5 yrs	100%
Analytical validation				
Laboratory-generated multi-orientation datasets	12 hours @ 100Hz 12 hours @ 10Hz	1 1	n/a	n/a
Nights with simultaneous sleep polysomnography [37]	1 night @ 85.7Hz	55	44.9 yrs	61%
Office-collected non-wear following an observer-controlled protocol	8 hours @ 100Hz	8	33.9 yrs	25%

Descriptions of the data and participants used in the verification testing and analytical

validation.

Ethical approval for the HeLP study was obtained from the Peninsula College of

Medicine and Dentistry in March 2012 (reference number 12/03/140). The Institutional

Review Board and Regional Ethics Committees of Copenhagen University Hospital

Rigshospitalet and Amsterdam University Medical Centre approved the SafeHeart study

protocol. Permission has been granted by the first author for the use of personal data for this

study. The polysomnography studies were approved by the University College London ethics

committee and the NRES Committee North East Sunderland ethics committees. ActivInsights

have authorised the use of the office-collected non-wear data and confirmed it was collected

with consent under appropriate safeguards.

All data processing and statistical tests are processed in R version 4.4.2 [38].

Measurement period information

204   The measurement period information is initially extracted from a bin file by the

205   *GENEAcore* package, which then implements an optimised file read procedure to extract 1-

206   second downsampled time series data for calibration, non-wear detection and transition

207   detection.

Calibration

A rolling standard deviation is calculated from the 1-second downsampled accelerations

over a defined window of 120 seconds, with a threshold of 13mg applied to the 3 axes mean

identifying measurements of non-movement. These selected measurements are iteratively fitted

to a sphere to estimate auto-calibration figures from real world data [39]. The values selected

for all processing parameters are shown in S2 Table with references.

All multi-day files are auto-calibrated using this method with the number of iterations

required and the improvements in calibration, as calculated by residual error, are recorded. The

laboratory-generated multi-orientation datasets both with 53 periods of non-movement each of

about 20 minutes over 12 hours are used to evaluate the correct convergence of the auto-

calibration. Once the calibration was completed, the scale and offset elements are systematically

perturbed and the residual error calculated (+/- 1% scale and +/- 50% offset).

Non-wear detection

Many approaches to the detection of non-wear have been published since the advent of

raw data acceleration measurement and a summary of algorithms is shown in S1 Table. These

are normally for single devices although dual device systems are known [40]. The methods use

features derived from acceleration across 3 axes in varying combinations with temperature and,

occasionally, time of day.

While signal properties can change between device and sensor models, the standard

deviation of acceleration appears to offer a consistent level of detection performance robust to

poor calibration [41]. Low pass signal filtering, including epochs and running means, allow

absolute values of acceleration to be used [42, 43] and posture changes to be determined [44,

45]. However, using filtered signals in subsequent analysis reduces dynamic range [46]. Some

publications note that the z axis has different noise characteristics and temperature

dependencies compared to the other axes [47, 48]. This can be explained by the construction of

the sensor with the z axis being through the plane of the silicon layers as opposed to in-plane for

the x and y axes.

The temperature sensors within devices can be used with absolute thresholds [43, 47,

49, 50], but have complex signal characteristics over time. While the thermal lag due to the mass

of the wearable would suggest a first order response, the initial rate of change is suppressed by

the internal hysteresis of the sensor and quantization of signal output [51]. Publications report

seeing maximum changes in temperature on device removal of 2-3 °C after 3-5 minutes [43, 52,

53] and a decoupling from body temperature after 16 minutes [49]. These findings help to

define an optimal time window for detecting device removal using temperature changes.

Earlier algorithms all use fixed length epochs with varying overlaps but, more recently,

event-based approaches have been deployed with transition detection of the acceleration [43,

54] and temperature signals [52, 53]. Event-based approaches have the advantage of precise

output time resolutions without needing to use very short epoch lengths which are prone to

classification error [55].

Several algorithms use 'boarding rules' to update the assessment of non-wear events

and neighbouring events once an initial classification has been made. These can be based on the

duration of the non-wear events [44, 56], the duration of the gaps between non-wear events

[54] or the signal characteristics of the event [57]. In addition, these approaches can be

combined with sleep determination [44, 57], recognising this as the most likely behavioural

confusion. Different rule sets are applied during night periods or when proximal events are

determined to be sleep.

The increased performance levels of newer approaches (resolution, accuracy and

minimum detection periods) have come at the expense of processing complexity. The

GENEAcore non-wear algorithm aims to simplify and reduce processing times without

compromising accuracy. Importantly for interoperability and modularity, the detected non-

wear periods are stored in the metadata separately from the time series measurements. It uses

approaches, features and parameter values that have been successfully and consistently used in

the prior art described above (S1 Table and S2 Table). It is an event-based approach with

candidate events identified when the previously derived, 2-minute rolling average of 1-second

downsampled acceleration standard deviation drops below 13mg [58]. Boarding rules are used

on candidate events to combine events that are separated by less than 5 minutes. These events

are classified as non-wear when any of these rules are satisfied:

• Temperature drop due to device removal – more than 0.7 °C per minute drop in the first

4 minutes of non-movement with less than 2 posture changes; or

• Sustained non-movement without posture change – any single event that is greater than

2 hours in duration; or

• Long-running non-movement – continuous non-movement periods greater than 12

hours in duration irrespective of the number of posture changes.

The resulting outputs have a minimum detectable non-wear event duration of 5 minutes

(timespan of temperature drop rule plus rolling average lag) with a time resolution of 1 second.

These are stored in the measurement period information.

Here, we test the non-wear algorithm with wrist-worn data against both the observer

diaries of controlled non-wear and sleep clinic participation. The office non-wear and sleep

polysomnography criterion data are processed to create estimates for non-wear for each

second. The estimates are then matched with observer records to ascertain true positive, true

negative, false positive and false negative rates which are combined into balanced accuracy. A

sensitivity analysis investigated the impact of changes in the acceleration standard deviation

threshold on the classification accuracy.

Epochs and events

286   Epochs and events are specified by their integer second start time and duration. Epochs

287   start from the beginning of the first minute of the day. The transition detection for event start

288   times utilises the 1-second downsampled time series from the measurement period information

289   and the resulting transition time points are stored there.

Transition detection

GENEAcore uses Pruned Exact Linear Time (PELT) changepoint detection [59] from the

Changepoint R package to identify transitions in the mean and variance of downsampled

acceleration. Each axis is assessed separately with the resulting timestamps combined while

amalgamating transitions within 5 seconds of each other (minimum event duration).

The PELT method uses a penalty value to control sensitivity to changes in the data and,

therefore, the average number of transitions detected. Accelerometer studies in animals have

previously used sample entropy to select penalty values that maximise the information density

of the event summary [60]. Here, we calculated the root mean square error (RMSE) of the event

summary output compared to the 1-second epoch output on each axis for different penalty

values on a day randomly selected from each of the 100 verification files. Inflection points in the

relationship between the average number of events per day and RMSE represent the point

where a stable representation of the original signal has been reached.

Once suitable penalty values are selected for each axis, the 2 multi-orientation

laboratory datasets are used to test the method. The timings of true transitions are compared to

those determined by the PELT algorithm. The impact of varying the minimum event duration on

the average number of events per day are then calculated. Additionally, we investigate the

distribution of event durations.

Characterisation

At this point in the processing, the measurement period information contains sensor

calibration parameters, non-wear periods and event transitions. The raw, sensor-level

measurements can now be processed as events or epochs with the selected algorithms to

characterise the data. Any further processing to classify the events or epochs (for example, as

sleep or high intensity activity) is a further step beyond the scope of pre-processing (Fig 1).

The architecture of the *GENEAcore* package allows different methods to be executed

within a consistent processing environment. Here, we compare two activity intensity measures

and different approaches to aggregation.

The earliest raw acceleration summary algorithm [61] took the sum of the gravity-

subtracted vector magnitudes in an epoch (SVM_{gs}) using the absolute value of any negative

numbers. The resulting measure can be difficult to interpret as it is linearly dependent on both

the sample frequency and epoch length; the mean of the absolute gravity subtracted

acceleration (AGSA) addresses this drawback. Euclidean norm minus one (ENMO) was later

proposed [56] and differs only by setting negative numbers from the subtraction of gravity to

zero. Both AGSA and ENMO are commonly expressed in milli-units of gravity, m*g*, where *g* is

equal to 9.81 m/s^2.

AGSA and ENMO are calculated for all time points in the verification dataset using 1-

second epochs as this is recognised to give the most accurate assessment of instantaneous

intensity [62, 63]. The percentage overlap of each above and below the 13m*g* standard

deviation movement threshold were calculated to understand their ability to discriminate

effectively in low movement conditions. AGSA and ENMO are then aggregated by day to

investigate the concordance [64, 65] of assessed active duration using a previously defined

AGSA threshold of 62.5m*g* [61, 66] and the modelled ENMO equivalent.

Despite the absence of an intensity criterion measure, it is still possible to illustrate how

an independent pre-processing stage supports future research. Event outputs are also

calculated for each day in the verification dataset as well as the epoch outputs. Both are then

aggregated by day to investigate the differences between variable duration event and fixed

duration 1-second epochs aggregation on estimates of daily active duration.

Results

The *GENEAcore* pre-processing pipeline was configured to output 3 data objects: the

measurement period information, 1-second epochs and variable duration events. Once

completed, the measurement period information contained calibration values, non-wear

periods and detected transitions separated from both the original data source and output time-

series data. These objects were traceably linked by a unique source identifier derived from

source file itself, rather than any manually controlled file name.

Calibration

During auto-calibration of the verification datasets, the mean number of iterations

required to meet the incremental fit improvement threshold (see S2 Table) was 51.3 (31.6 SD).

The mean improvement from factory calibration was 9.5mg resulting in a mean residual sphere

fit error of 6.7mg (0.48mg SD).

All calibration perturbations of the laboratory-generated multi-orientation datasets

resulted in an increased residual sphere fit error, confirming convergence. The mean increase

for a 1% change in scale was 1.1mg, and 1.2mg for a 50% change in offset.

The repeatability of the auto-calibration was evaluated on the 2 devices worn multiple

times by the lead author. The mean standard deviation of calibration gain and offset between

devices was 19.5mg and 31.5mg, while the within device mean standard deviation was 1.0mg

and 2.1mg, respectively.

Non-wear detection

The sleep polysomnography (PSG) and office non-wear criterion data were processed to

create estimates for non-wear for each second. The office non-wear contained 157 observed

non-wear events with a mean duration of 1268 seconds (min. 178, max. 3629). All assessments

of non-wear during polysomnography measurements were assumed to be false positives. Table

2 shows the resulting sensitivity, specificity and balanced accuracy alongside the base

observation rates as described in the methods.

Table 2. Non-wear detection performance.

	True positives	True negatives	False positives	False negatives	Sensitivity	Specificity	Balanced accuracy
Sleep PSG	0	1.8×10^6	906	0	—	99.95%	—
Office non-wear	66564	118240	2326	11918	84.8%	98.1%	91.4%
Combined	66564	2.0×10^6	3232	11918	84.8%	99.8%	92.3%

Results of non-wear detection algorithm testing in different wear scenarios.

The acceleration standard deviation threshold of value was varied between 6mg and

20mg to test whether the widely used 13mg threshold [58] for identifying non-movement was

optimal for this sensor and non-wear algorithm in these wear scenarios. The changes in

detection performance are shown in Fig 2, where a marginal increase in balanced accuracy from

92.3% to 92.6% can be seen at an acceleration standard deviation threshold of 12mg.

Fig 2. Non-wear detection sensitivity analysis. Changes in non-wear detection performance

as measured with balanced accuracy for different values of mean acceleration standard

deviation threshold.

Transition detection

Optimal penalty values for PELT changepoint detection were found to be 18, 25 and 16

for the x, y and z axes respectively. When validated against the 2 multi-orientation datasets, only

1 of 104 transitions were not correctly identified and mean minimum distance to transitions

was 1.65 seconds.

For each axis individually, the target number of events per day, based on the RMSE

inflection points, was approximately 500. When the detected transitions for each axis in the

verification dataset were combined with a minimum event duration of 5 seconds, the mean

number of events per complete day (n=1,296) was 1,116 (221 SD). S1 Fig shows plots of RMSE,

events per day and penalty value with inflection points for each axis and verification dataset

sub-group.

The minimum permitted event duration when combining transitions was varied from 3

to 60 seconds with plots of resulting transitions per day for each verification dataset sub-group

shown in Fig 3.

Fig 3. Impact of minimum event duration on events per day. The variation of the number of

daily events with minimum event duration in the 3 verification population groups.

The distribution of event durations approximated closely to the log normal (μ = 3.9,

σ=1.1) giving an expected event duration, E(X), of 68.6 seconds.

Data summaries based on variable length events provided a compression ratio of 87.9

compared to a 1-second fixed epoch output, while preserving a 1 second resolution for changes

in the underlying data.

Characterisation

The relationship between AGSA and ENMO is shown in Fig 4 for every second in the verification dataset (130,484,716 data points). The red group has all time classed as moving (mean standard deviation of acceleration above 13mg) and the blue group contains the remaining time classed as non-movement. Above 1g of mean acceleration, AGSA and ENMO give near identical results. In the lower acceleration ranges of the moving group, ENMO consistently under-reports compared to AGSA. In the non-movement group, the asymmetric treatment of negative values in the ENMO algorithm is apparent, with values of mean acceleration often reported below 1mg despite the 6-7mg noise floor of the sensor. The overlap of AGSA values in the moving and non-movement groups was 6.0% while the ENMO overlap was 10.5%.

Fig 4. Comparison of activity intensity calculated by AGSA and ENMO. Density plot of 1s epochs summarised by AGSA and ENMO for the verification dataset split into groups of non-movement (blue) and movement (red).

The typical sedentary to active threshold for AGSA of 62.5mg [61, 66] is shown vertically in Fig 4. A linear regression of the movement group (r^2 = 0.957, slope = 0.872, intercept = -20.5 mg) establishes an equivalent threshold for ENMO of 34.1mg, shown horizontally.

When these sedentary to active thresholds are applied daily over 1,296 complete days, the concordance of the resulting daily active duration was found to be greater than 0.99. Correlation and Bland-Altman plots are shown in S2 Fig. ENMO has bias of +1.25 minutes per day compared to AGSA.

The *GENEAcore* pre-processing pipeline allows us to contrast event and epoch segmentation. The correlation of active duration per day determined with variable duration events and fixed duration 1-second epochs is shown in S3 Fig with an accompanying Bland-Altman regression plot. The concordance was 0.831 (95% CI, 0.821–0.842), however the

432   Pearson correlation coefficient was 0.985 (95% CI, 0.983–0.986). The regression model of the

433   differences shows the event-derived active durations per day to be consistently 31.5% (95% CI,

434   30.8 - 32.2) higher than those from 1-second epochs.

Discussion

Our aim in this study was to define and test a simplified, modular pipeline for the pre-

processing of high-resolution accelerometry – assuring the quality of raw, sensor-level data as

they are successively refined into digital measures.

The handling of data objects during the study demonstrated full traceability from input

data sources to the output digital measures with a record of all processing steps in the

measurement period information. The modularity of implementation and reporting provides

the required transparency, while also supporting easy application of new or improved

techniques. Integrated testing and change control will assure regulators of the suitability for

health applications.

Results recorded for the calibration stage confirm the success of earlier studies [39],

showing residual error comparable to the noise floor of the sensor itself. Furthermore, the

perturbation tests demonstrate correct convergence of the optimised method and repeatability

tests confirm that the calibration of these sensors is static over time. The detection of

transitions using changepoint analysis proved reliable with 99% of transitions detected, on

average, within 2 seconds of their occurrence.

This is the first study to empirically test the 13mg acceleration standard deviation

threshold commonly used in the literature since 2014 [58]. This was achieved by calculating the

balanced accuracy of non-wear detection through a range of threshold values. The results

indicate a slight reduction of the threshold could be beneficial for this specific device, however

preserving current practice affords some leeway for noisier sensors, with a minimal reduction

in efficacy.

The non-wear algorithm components, all based on existing know-how, are independent

of body location and show no variation between adults and children or male / female

participants [49]. We consider that the approaches and results have high generalisability to

wear position and population. The criterion datasets used to test the non-wear algorithm were

selected to challenge in situations where performance can be marginal: sleep measured by PSG

and short periods in a lived environment controlled by an observer. Of the previous non-wear

validation studies, only 2 used observers [44, 57] and 1 used a separate device [54], all others

relied on participant diaries or expert rating of data. The high overall specificity is excellent for

sleep measurement where false positives for non-wear during the night are problematic for

clinicians. The sensitivity to short off-wrist periods is reasonable; it will be higher and

comparable to the other published methods for longer periods of non-wear where the

participant is not in proximity with the device. The detection of non-wear during transport will

remain difficult, however this is normally accompanied by long, still periods where accuracy is

high.

The detailed comparison of 2 commonly used measures of activity intensity (AGSA and

ENMO) reveal that attention must be given through every stage of digital measure selection and

embodiment for any health application. A small change in the algorithm has major

consequences. The sedentary AGSA threshold of 62.5mg [61, 66] was translated to an

equivalent ENMO threshold of 34.1mg; a value which is exactly in line with other studies across

a range of populations [67, 68]. Therefore, for higher movement levels above the sedentary

thresholds, AGSA and ENMO are substitutable for measuring daily active times using the linear

regression described. More care is required at lower movement levels, when analysing

individual events and when using activity intensity as a continuous measure. Below the

sedentary threshold, ENMO quickly becomes non-linear with reduced movement, outputting

results inconsistent with the measurement noise of the sensor. There is more headroom for

AGSA below the sedentary threshold until sensitivity is lost in a more linear fashion.

A particular contribution of this work is the comparison of epoch and event pre-

processing made possible by the *GENEAcore* pipeline. Variable duration events created higher

assessments of daily activity duration compared to those calculated from 1-second epochs. The

disparity in outputs was an extra 19 minutes for every additional hour of activity. The

magnitude of this disparity will vary with the epoch length selected for comparison. The

threshold used in this analysis to identify active behaviours encompasses light, moderate and

vigorous intensity physical activity. Longer epochs are consistently reported to reduce

estimates of both vigorous and sedentary time [22, 23, 24, 25, 69], pushing all estimates

towards the mid-intensity bands. So, the disparities in reported active time compared to events

will decrease with longer epochs. However, in the epoch paradigm, these longer epochs will

under-report vigorous active times, whereas variable duration events will still capture short

bursts of higher intensity physical activity. This ability for event-based processing to adapt to

the latent structures of the data provides a solution to suggestions that different epoch lengths

are more effective at detecting behaviours of different intensity [20, 21].

Data-driven estimates of the typical durations of human movement behaviours are not

widely available. The log normal distribution found here, with an expected event duration of

just over a minute, is similar to that of other land-based mammals [60]. A minimum event

duration of 5 seconds preserved the complexity of daily data, allows short bouts of activity to be

recorded [69, 70] and gives a finite time for the transitions themselves [27]. These findings

confirm the existing conventions of epoch length selection of between 1 and 60 seconds. The

compromises associated with epoch length choices outlined above are resolved by variable

duration encoding. The event-based paradigm delivers the timestamp accuracy of short epochs

and efficient data compression of longer epochs with high classification confidence. In addition,

the relative timings of the transitions themselves expose underlying patterns of physical

behaviours.

The uncertainties associated with the selection of appropriate epoch length are

compounded once epochs are grouped to represent bouts of physical behaviour. The *GENEAcore*

pre-processing pipeline specifies event transitions prior to characterisation and classification –

allowing an explicit translation from data-driven events to bouts without further processing. In

contrast, existing epoch-only pipelines must apply additional rules with parameters to identify

bout transitions, epoch aggregation and permissible interruptions. These rules require

substantial verification and analytical validation for them to be used repeatably and with

confidence.

At first glance, event approaches appear more complicated than the use of traditional

epochs. The variable durations pose methodological problems in parameter selection,

visualisation, aggregation and bridging of reporting boundaries (for example, across days).

However, these obstacles are inherent to bout representations and data-driven event transition

detection simplifies the later stages of process validation. Lastly, events defined in the early

steps of pre-processing supply event duration as a valuable, additional feature for later

classification stages.

Our study has limitations in the small sizes of datasets used for the analytical validation

steps, absence of processing time comparisons between existing packages and the lack of a

criterion measure for activity intensity. Future work can use other larger, previously collected

data to further investigate comparative accuracies of AGSA vs. ENMO activity intensities and

epoch vs. event pre-processing. Ideally, this work would use continuous measures of activity

intensity rather than cutpoint-derived energy expenditure classes to avoid threshold effects. A

better understanding of the impact of the rules for bout creation from epochs on the reporting

of physical activity acquisition (durations, volumes and patterns) at different epoch lengths is

needed. Data-driven events would provide a stable benchmark for this analysis. Study strengths

include the use of a cross-functional team to create a maintainable, engineered tool with

backward compatibility for existing data. Deliberately constraining scope to the pre-processing

pipeline alone enhanced the depth and focus of results. The thorough documentation of

assumptions will support the consistent implementation of existing measures identified as

important to public health [71, 72] as well as clinical research.

Conclusions

Our focus on verification and analytical validation of accelerometer pre-processing has exposed detailed engineering considerations that must underpin reliable digital measures. The rush towards results with health associations often leaves little room for these important topics.

The meticulous evaluation of the simplest steps of pre-processing has confirmed the attention to detail that is required to create robust and reproducible results in later clinical validation. The integrity of data pathways is preserved through standardised approaches and rigorous implementation.

We have shown this optimised framework to be a capable building block of future research, clinical care and public digital health systems. It can deliver the basic digital measures needed to understand physical behaviours over multiple days with minimum assumptions.

Acknowledgements

The authors acknowledge the contributions of the study leaders and participants that provided data for the verification and analytical validation datasets, particularly: Katrina Wyatt and Jennifer Lloyd of the HeLP study; Diana My Frodi and Maarten Kolk of the SafeHeart study; and Vincent van Hees, Sarah Charman and Kirstie Anderson for the Newcastle polysomnography and accelerometer data. The authors would also like to thank Brad Metcalf of University of Exeter for his statistical support. For the purpose of open access, the authors have applied a Creative Commons Attribution (CC BY) licence to any Author Accepted Manuscript version arising from this submission.

References

1. Liu F, Wanigatunga AA, Schrack JA. Assessment of Physical Activity in Adults Using Wrist Accelerometers. Epidemiologic Reviews. 2021 Jul 2;43(1).

2. U.S. Food and Drug Administration. Digital Health Technologies for Remote Data Acquisition in Clinical Investigations. U.S. Food and Drug Administration; 2023.

3. Canali S, Schiaffonati V, Aliverti A. Challenges and Recommendations for Wearable Devices in Digital health: Data quality, interoperability, Health equity, Fairness. PLOS Digital Health. 2022 Oct 13;1(10).

4. Goldsack JC, Coravos A, Bakker JP, Bent B, Dowling AV, Fitzer-Attas C, et al. Verification, analytical validation, and clinical validation (V3): the foundation of determining fit-for-purpose for Biometric Monitoring Technologies (BioMeTs). npj Digital Medicine. 2020 Apr 14;3(1).

5. Guan G, Merav Mofaz, Qian G, Patalon T, Erez Shmueli, Yamin D, et al. Higher sensitivity monitoring of reactions to COVID-19 vaccination using smartwatches. npj Digital Medicine. 2022 Sep 9;5(1):140–0.

6. Aunger J, Wagnild J. Objective and subjective measurement of sedentary behavior in human adults: A toolkit. American Journal of Human Biology. 2020 Dec 5;34(1).

7. Prince SA, Cardilli L, Reed JL, Saunders TJ, Kite C, Douillette K, et al. A comparison of self-reported and device measured sedentary behaviour in adults: a systematic review and meta-analysis. International Journal of Behavioral Nutrition and Physical Activity. 2020 Mar 4;17(1).

8. Tully MA, Panter J, Ogilvie D. Individual Characteristics Associated with Mismatches between Self-Reported and Accelerometer-Measured Physical Activity. PLoS ONE. 2014 Jun 11;9(6):e99636.

9.   Guo W, Key TJ, Reeves GK. Accelerometer compared with questionnaire measures of
physical activity in relation to body size and composition: a large cross-sectional
analysis of UK Biobank. BMJ Open. 2019 Jan;9(1):e024206.

10. International Organization for Standardization. IEC 62366-1:2015. Application of
usability engineering to medical devices. Geneva: ISO; 2015.

11. European Commission. Clinical Evaluation: A Guide for Manufacturers and Notified
Bodies under Directives 93/42/EEC and 90/385/EEC. MEDDEV 2.7/1 revision 4.
Brussels: EC; 2016.

12. International Organization for Standardization. ISO 13485:2016. Medical devices —
Quality management systems — Requirements for regulatory purposes. Geneva: ISO;
2016.

13. International Organization for Standardization. Geneva: ISO 14971:2019. Medical
devices — Application of risk management to medical devices. Geneva: ISO; 2019.

14. Dibben GO, Santillan C, Brage S, Buman M, Duncan E, Granat MH, et al. Reporting of
Accelerometry in Health Research: A Scoping Review of Current Guidance. Scandinavian
Journal of Medicine and Science in Sports. 2025 Oct 1;35(10):e70143–3.

15. Migueles JH, Rowlands AV, Huber F, Sabia S, van Hees VT. GGIR: A Research Community–
Driven Open Source R Package for Generating Physical Activity and Sleep Outcomes
From Multi-Day Raw Accelerometer Data. Journal for the Measurement of Physical
Behaviour. 2019 Sep 1;2(3):188–96.

16. Doherty A, Chan S, Yuan H, Walmsley R. accelerometer: A Python Toolkit for Extracting
Physical Activity and Behavior Metrics from Wearable Sensor Data. Zenodo 14515076.
2020.

17. Beyer KB, Weber KS, Cornish BF, Vert A, Thai V, Godkin FE, et al. NiMBaLWear analytics
pipeline for wearable sensors: a modular, open-source platform for evaluating multiple
domains of health and behaviour. BMC Digital Health. 2024 Feb 8;2(1).

18. Lin W, Karahanoglu FI, Demanuele C, Khan S, Cai X, Santamaria M, et al. SciKit digital

health package for accelerometry-measured physical activity: comparisons to existing

solutions and investigations of age effects in healthy adults. Frontiers in Digital Health.

2023 Nov 27;5.

19. Frodi DM, Kolk MZH, Langford J, Knops R, Tan HL, Andersen TO, et al. Long-term

adherence to a wearable for continuous behavioural activity measuring in the SafeHeart

implantable cardioverter defibrillator population. European Heart Journal - Digital

Health. 2024 Aug 1;5(5):622–32.

20. Gabriel K, McClain JJ, Schmid KK, Storti KL, High RR, Underwood DA, et al. Issues in

accelerometer methodology: the role of epoch length on estimates of physical activity

and relationships with health outcomes in overweight, post-menopausal women.

International Journal of Behavioral Nutrition and Physical Activity. 2010;7(1):53.

21. Altenburg TM, Wang X, van Ekris E, Andersen LB, Møller NC, Wedderkopp N, et al. The

consequences of using different epoch lengths on the classification of accelerometer

based sedentary behaviour and physical activity. PLOS ONE. 2021 Jul

15;16(7):e0254721.

22. Lee JH, Kwon S, Lee S, Jang SY, Jo W, Jin J, et al. The influence of accelerometer epoch

length on physical activity output in adolescent athletes. Journal of exercise

rehabilitation. 2023 Dec 26;19(6):370–4.

23. Burchartz A, Kolb S, Klos L, Schmidt S, von Haaren-Mack B, Niessner C, et al. How specific

combinations of epoch length, non-wear time and cut-points influence physical activity.

German Journal of Exercise and Sport Research. 2024 Jun;54:169-178.

24. Haddadj R, Verdot C, Salanave B, Deschamps V, Vanhelst J. Impact of varying

accelerometer epoch length on physical activity patterns in adults: Considerations for

public health. PloS one. 2024;19(12):e0316176.

25. Aadland E, Nilsen AKO. Accelerometer epoch length influence associations for physical

activity intensities with body mass index and locomotor skills in young children. Journal

of Sports Sciences. 2022 Jun 25;40(14):1568–77.

26. Granat MH. Event-based analysis of free-living behaviour. Physiological Measurement.

2012 Oct 31;33(11):1785–800.

27. Twaites J, Everson R, Langford J, Hillsdon M. Transition Detection for Automatic

Segmentation of Wrist-Worn Acceleration Data: A Comparison of New and Existing

Methods. Journal for the Measurement of Physical Behaviour. 2019;1–10.

28. Pate RR. Physical activity and public health. A recommendation from the Centers for

Disease Control and Prevention and the American College of Sports Medicine. JAMA: The

Journal of the American Medical Association. 1995 Feb 1;273(5):402–7.

29. Kobeissi E, Dempsey P, Alshallal A, Rennie KL, Brage S, Wijndaele K, et al. A Systematic

Review and Taxonomy of Metrics and Methods Used to Describe Temporal Movement

Behavior Patterns Measured Using Wearable Devices. Journal for the Measurement of

Physical Behaviour. 2026 Jan 1;9(1).

30. Ahmadi M, Hamer M, Jason, Murphy M, Sanders JP, Doherty A, et al. Brief bouts of

device-measured intermittent lifestyle physical activity and its association with major

adverse cardiovascular events and mortality in people who do not exercise: a

prospective cohort study. The Lancet Public Health. 2023 Oct 1;8(10):e800–10.

31. Ma T, Sirard J, Yang L, Li Y, Tsang S, Fu A. Revisiting the concept of bout: associations of

moderate-to-vigorous physical activity sessions and non-sessions with mortality.

International Journal of Behavioral Nutrition and Physical Activity. 2024 Jul 29;21(1).

32. Eke H, Bonn SE, Trolle Lagerros Y. Wrist-worn accelerometers: Influence of decisions

during data collection and processing: A cross-sectional study. Health Science Reports.

2024 Jan;7(1).

33. Liang YT, Wang C, Hsiao CK. Data Analytics in Physical Activity Studies With

Accelerometers: Scoping Review. Journal of Medical Internet Research. 2024 Sep

11;26:e59497–7.

34. International Organization for Standardization. IEC 62304:2006. Medical device

software — Software life cycle processes. Geneva: ISO; 2006.

35. Wyatt KM, Lloyd JJ, Abraham C, Creanor S, Dean S, Densham E, et al. The Healthy

Lifestyles Programme (HeLP), a novel school-based intervention to prevent obesity in

school children: study protocol for a randomised controlled trial. Trials. 2013;14(1):95.

36. Frodi DM, Kolk MZH, Langford J, Andersen TO, Knops RE, Tan HL, et al. Rationale and

design of the SafeHeart study: Development and testing of a mHealth tool for the

prediction of arrhythmic events and implantable cardioverter-defibrillator therapy.

Cardiovascular Digital Health Journal. 2021 Dec;2(6):S11–20.

37. van Hees VT, Sabia S, Anderson KN, Denton SJ, Oliver J, Catt M, et al. A Novel, Open

Access Method to Assess Sleep Duration Using a Wrist-Worn Accelerometer. PLOS ONE.

2015 Nov 16;10(11):e0142533.

38. R Core Team. R: The R Project for Statistical Computing. R-project.org. 2025.

39. van Hees VT, Fang Z, Langford J, Assah F, Mohammad A, da Silva ICM, et al.

Autocalibration of accelerometer data for free-living physical activity assessment using

local gravity and temperature: an evaluation on four continents. Journal of Applied

Physiology. 2014 Oct 1;117(7):738–44.

40. Syed S, Morseth B, Hopstock LA, Horsch A. Evaluating the performance of raw and epoch

non-wear algorithms using multiple accelerometers and electrocardiogram recordings.

Scientific Reports. 2020 Apr 3;10(1).

41. van Hees VT, Renström F, Wright A, Gradmark A, Catt M, Chen KY, et al. Estimation of

Daily Energy Expenditure in Pregnant and Non-Pregnant Women Using a Wrist-Worn

Tri-Axial Accelerometer. PLoS ONE. 2011 Jul 29;6(7):e22922.

42. Rahimi-Eichi H, Coombs III G, Vidal Bustamante CM, Onnela JP, Baker JT, Buckner RL.

Open-source Longitudinal Sleep Analysis From Accelerometer Data (DPSleep):

Algorithm Development and Validation. JMIR mHealth and uHealth. 2021 Oct

6;9(10):e29849.

43. Vert A, Weber KS, Thai V, Turner E, Beyer KB, Cornish BF, et al. Detecting accelerometer

non-wear periods using change in acceleration combined with rate-of-change in

temperature. BMC Medical Research Methodology. 2022 May 20;22(1).

44. Ahmadi MN, Nathan N, Sutherland R, Wolfenden L, Trost SG. Non-wear or sleep?

Evaluation of five non-wear detection algorithms for raw accelerometer data. Journal of

Sports Sciences. 2019 Dec 11;38(4):399–404.

45. Sundararajan K, Georgievska S, te Lindert BHW, Gehrman PR, Ramautar J, Mazzotti DR,

et al. Sleep classification from wrist-worn accelerometer data using random forests.

Scientific Reports. 2021 Jan 8;11(1).

46. Lettink A, van Wieringen WN, Altenburg TM, Chinapaw MJM, van Hees VT. Inter-Brand, -

Dynamic Range, and -Sampling Rate Comparability of Raw Accelerometer Data as Used

in Physical Behavior Research. Journal for the Measurement of Physical Behaviour. 2024

Jan 1;7(1).

47. Skovgaard EL, Roswall MA, Pedersen NH, Larsen KT, Grøntved A, Brønd JC.

Generalizability and performance of methods to detect non-wear with free-living

accelerometer recordings. Scientific reports. 2023 Spring;13(1):2496.

48. Analog Devices. ADXL345 Data Sheet. Analog Devices; 2022.

49. Zhou SM, Hill RA, Morgan K, Stratton G, Gravenor MB, Bijlsma G, et al. Classification of

accelerometer wear and non-wear events in seconds for monitoring free-living physical

activity. BMJ Open. 2015 May 11;5(5):e007447–7.

50. Rasmussen MGB, Pedersen J, Olesen LG, Brage S, Klakk H, Kristensen PL, et al. Short-

term efficacy of reducing screen media use on physical activity, sleep, and physiological

713  stress in families with children aged 4–14: study protocol for the SCREENS randomized

714  controlled trial. BMC Public Health. 2020 Mar 23;20(1).

715 51. Texas Instruments. TMP1075 Data Sheet. Texas Instruments; 2024.

716 52. Pagnamenta S, Grønvik KB, Aminian K, Vereijken B, Paraschiv-Ionescu A. Putting

717  Temperature into the Equation: Development and Validation of Algorithms to

718  Distinguish Non-Wearing from Inactivity and Sleep in Wearable Sensors. Sensors. 2022

719  Feb 1;22(3):1117–7.

720 53. Barakat A, Manga A, Sheikh A, McWilliams R, Rowlands AV, Singh H. Feasibility of Using

721  a GENEActiv Accelerometer with Triaxial Acceleration and Temperature Sensors to

722  Monitor Adherence to Shoulder Sling Wear Following Surgery. Sensors. 2024 Jan

723  29;24(3):880.

724 54. Syed S, Morseth B, Hopstock LA, Horsch A. A novel algorithm to detect non-wear time

725  from raw accelerometer data using deep convolutional neural networks. Scientific

726  Reports. 2021 Apr 23;11(1).

727 55. Ahmadi MN, Pavey TG, Trost SG. Machine Learning Models for Classifying Physical

728  Activity in Free-Living Preschool Children. Sensors. 2020 Aug 5;20(16):4364.

729 56. van Hees VT, Gorzelniak L, Dean León EC, Eder M, Pias M, Taherian S, et al. Separating

730  Movement and Gravity Components in an Acceleration Signal and Implications for the

731  Assessment of Human Daily Physical Activity. PLoS ONE. 2013 Apr 23;8(4):e61691.

732 57. Barouni A, Ottenbacher J, Schneider J, Feige B, Riemann D, Herlan A, et al. Ambulatory

733  sleep scoring using accelerometers—distinguishing between nonwear and sleep/wake

734  states. PeerJ. 2020 Jan 2;8:e8284–4.

735 58. da Silva IC, van Hees VT, Ramires VV, Knuth AG, Bielemann RM, Ekelund U, et al. Physical

736  activity levels in three Brazilian birth cohorts as assessed with raw triaxial wrist

737  accelerometry. International Journal of Epidemiology. 2014 Oct 30;43(6):1959–68.

59. Killick R, Fearnhead P, Eckley IA. Optimal Detection of Changepoints With a Linear
Computational Cost. Journal of the American Statistical Association. 2012 Oct
17;107(500):1590–8.

60. Langford J, Fawcett T, Price E, Bradley D, Wilson A, Croft D. Bio-logging reveals heritable
patterns of natural behaviours in sheep. Acta IMEKO. 2024 Mar 25;13(1):1–6.

61. Esliger DW, Rowlands AV, Hurst TL, Catt M, Murray P, Eston RG. Validation of the GENEA
Accelerometer. Medicine & Science in Sports & Exercise. 2011 Jun;43(6):1085–93.

62. Rowlands AV, Powell SM, Humphries R, Eston RG. The effect of accelerometer epoch on
physical activity output measures. Journal of Exercise Science & Fitness. 2006 Jan 1;4(1).

63. Fabre N, Lhuisset L, Bernal C, Bois J. Effect of epoch length on intensity classification and
on accuracy of measurement under controlled conditions on treadmill: Towards a better
understanding of accelerometer measurement. PLOS ONE. 2020 Jan 24;15(1):e0227740.

64. Lin LI. A concordance correlation coefficient to evaluate reproducibility. Biometrics.
1989 Mar 1;45(1):255–68.

65. Lin LI. A note on the concordance correlation coefficient. Biometrics. 2000 Mar; 56(1):
324– 325.

66. Fraysse F, Post D, Eston R, Kasai D, Rowlands AV, Parfitt G. Physical Activity Intensity
Cut-Points for Wrist-Worn GENEActiv in Older Adults. Frontiers in Sports and Active
Living. 2021 Jan 15;2.

67. Boddy LM, Noonan RJ, Rowlands AV, Hurter L, Knowles ZR, Fairclough SJ. The
backwards comparability of wrist worn GENEActiv and waist worn ActiGraph
accelerometer estimates of sedentary time in children. Journal of Science and Medicine
in Sport. 2019 Jul;22(7):814–20.

68. Matthews CE, Saint-Maurice P, Freeman JR, Hayes HA, Shreves AH, Doherty A, et al.
Performance Evaluation of Algorithms to Estimate Daily Sedentary Time Using Wrist-
Worn Sensors in Free-Living Adults. Journal for the Measurement of Physical Behaviour.
2025;8(1).

69. Sanders T, Cliff DP, Lonsdale C. Measuring Adolescent Boys' Physical Activity: Bout
Length and the Influence of Accelerometer Epoch Length. PLoS ONE. 2014 Mar
18;9(3):e92040.

70. Skinner AM, Rowlands AV, Dimitris Vlachopoulos, Barker AR, Janz KF, Moore SA. The
Influence of Accelerometer Epoch Length on Associations of Physical Activity Intensity
and Volume with Bone Outcomes. Journal of sports science and medicine. 2023 Mar
1;117–32.

71. World Health Organization. Measuring physical activity in adults using wearable
technologies: report of a scientific meeting, 10–12 June 2025, Montreal, Canada. Geneva:
WHO; 2025.

72. Beauchamp MK, D'Amore C, Raina P, McIlroy W, Adesina N, Ahmadi M, et al. Establishing
global standards on wearable technology for measuring mobility in ageing populations:
an international consensus exercise. Age and Ageing. 2026 Jan;55(1).
778

Supporting information

S1 Table. Non-wear detection methods. Descriptions of different methods for the detection
of non-wear including: the general method, features used for detection, the minimum detectable
non-wear duration, the time resolution of non-wear periods, the criterion used in their
validation and the wear location.

S2 Table. Processing parameters. The definitions, values and the supporting references for
the parameters used in the pre-processing operations.

S1 Fig. Changepoint penalty value assessments. The variation of the number of daily events
and the RMSE of intensity with changepoint penalty value for each axis of acceleration in the 3
verification population groups.

S2 Fig. Comparing daily active duration using AGSA and ENMO. (a) Correlation plot of
active hours per day as assessed by AGSA and ENMO, (b) Bland-Altman plot of the means and

differences of AGSA and ENMO active hours per day with mean bias and 95% SD confidence

intervals.

S3 Fig. Comparing daily active duration using event and epoch aggregates of AGSA. (a)

Correlation plot of active hours per day as assessed by AGSA 1s epoch and variable duration

events, (b) Bland-Altman regression plot modelling the differences of AGSA 1s epoch and

variable duration event active hours by event-assessed duration per day with mean bias and

95% SEE confidence intervals.

Figure 1

Figure 2

Figure 3

Figure 4