Article

Chitozyme: First Peroxidase-like Activity of Chitosan for Multiplexed Visual Detection of H₂O₂, Glucose and Lactate on Paper-based Device

Syed Rahin Ahmed¹, Xuan Weng¹ and Suresh Neethirajan^{1,*}

¹BioNano Laboratory, School of Engineering, University of Guelph, Guelph, ON N1G 2W1

* Correspondence: sneethir@uoguelph.ca; Tel.: +1-519-824-4120

Academic Editor: name

Received: date; Accepted: date; Published: date

Abstract: Visual read-out diagnostics tools are promising candidates for field applicable medical devices. Current colorimetric biosensors require introduction of natural enzymes or nanozymes, which has some serious drawbacks for practical applications. Chitosan, a natural polymer, provides safe and efficient compound in medical and pharmaceutical technology. Herein, we report on a simple, cost-efficient, field-portable, environmental friendly and ultra-sensitive multiplex detection platform based on peroxidase-like activity of chitosan in the presence of 3,3',5,5'-Tetramethylbenzidine (TMBZ) and H₂O₂. This straight forward signal amplification strategy was successfully applied to detect H₂O₂, glucose and lactate with the limit of detection (LOD) of 2.64 pM, 0.104 μM and 2.8 nM respectively, represents the lowest LOD of H₂O₂, glucose and lactate with visual read-out. The chitosan-based assay performance was also retained in complex biological media for glucose and lactate detection. Furthermore, the proposed assay was successfully demonstrated as a paper-based colorimetric biosensor. Most importantly, the simplicity, biocompatibility and sensitivity of the proposed assay will open new doors for instrument free naked eye visual detection of H₂O₂, glucose and lactate detection.

Keywords: Chitozyme; Hydrogen peroxide; paper biosensor; lactate; glucose

1. Introduction

Instrument free point of care diagnostic tools has high potential to provide a simple, cheap and easy-accessible device for medical applications. A key example of such device is colorimetric read out strategy that enable the development of low-cost diagnostics to detect analytes visually [1-5]. Over the last few decades, conventional enzymes i.e., horseradish peroxidase (HRP), alkaline phosphate (ALP) have been used for visual color amplification. However, their denaturation behaviour at high temperature, pH stability and high cost limits their use in real life applications.² Recently, researchers are attempting to replace natural enzymes with nanomaterials (known as nanozymes) to overcome the aforementioned problems; and few nanomaterials delivered as promising candidates of this class for various applications [6-8]. Up-to-date, although much efforts have been given on nanomaterials based artificial enzyme developments, their aggregation problems in different chemical environments; monodispersibility; uniform size preparation and toxicological implications are major concerns in the practical applications for in-vitro or in-vivo detection of analytes.

Introducing a new diagnostic tool that is easy-to-operate, containing non-toxic chemicals and ability to detect extremely small analyte concentrations with enhanced superior visual color would enable the development of low-cost commercially viable technologies for medical applications. In view of above discussions, here we propose to amplify the visual detection using chitosan as a peroxidase-like enzyme. We proposed the term of our new approach to be “chitozymes”.

Chitosan has received lots of research interest because of its structural and broader range of chemical and physical attributes in fundamental science, applied research, and industrial biotechnology. For example, chitosan can be used as an advanced biofabrication material; surface adaptation of cell/protein-integrating biological systems; and enzyme surface immobilization in preparation of medical sensing devices [9,10]. Chitosan has benefits of natural availability as well as biocompatibility, brings it as a potential candidate to the creation of chitosan-based analytical tools for point-of-care clinical, biotechnological, and environmental analysis. Recently, the integration of chitosan with nanomaterials was reported as a peroxidase-like activity for visual sensing application [11,12]. However, up-to-now, no reports on bare chitosan as a peroxidase-like enzymatic activity has been published yet. Hence, here for the first time, we present peroxidase-like activity of chitosan for rapid ultrasensitive detection of H_2O_2 , glucose and lactate. This approach will diminish the requirements of others material to integrate with chitosan for sensing applications as well as reduce the cost, labour and multi-step preparative process.

Positively charged gold nanoparticles (Au NPs) exhibits higher peroxidase activity in comparison to negatively charged Au NPs [13]. The surface charge of nanoparticles plays a crucial role in influencing the peroxidase-like activity. Hence, we targeted three positively charged chemicals i.e., Cetyl trimethylammonium bromide (CTAB), Poly-L-Lysine (PLL) and chitosan to check their enzymatic activity towards TMBZ/ H_2O_2 . Our results confirm that chitosan demonstrated stronger peroxidase-like activity in comparison to other two chemicals, and allowed us to develop a portable multiplex system for simultaneous detection of glucose, lactate and H_2O_2 .

Chitosan could catalyse the oxidation of the peroxidase substrate TMBZ by H_2O_2 to develop a blue color in aqueous solution, which could introduce us a new way of visual detection of H_2O_2 . A sensitive and accurate determination of H_2O_2 is much needed because of its numerous application in food and pharmaceutical industries, and environmental analysis [14]. In addition, H_2O_2 is produced by the oxidation of glucose due to catalysis by glucose oxidase (GOD) and the oxidation of lactic acid catalysed by lactate oxidase (LOD). The proposed sensing mechanism was further extended to quantitatively analyse glucose and lactic acid in samples. The determination of glucose concentration at lower limit is essential for medical diagnostic applications such as for diabetic patients, as well as in industrial applications [15]. The quantitative measurement of lactic acid concentration in body fluids has importance in clinical assessment to determine the patients with diabetic coma, bacterial infections and others medical symptoms [16]. Lactate concentration plays a key parameter in healthcare, food industries and for assessing patient health conditions like hemorrhage, respiratory failure, hepatic disease, sepsis and tissue hypoxia [17].

The creative idea to demonstrate the chitosan-based biosensing assay on a paper-based device towards detecting H_2O_2 , glucose and lactate would facilitate the pathway for obtaining reliable healthcare data under non-laboratory conditions at low-cost. Recently, paper-based analytical tools has gained attractive alternative to highly sophisticated instrumentation for numerous applications. Paper-based device can be easily printed, coated and impregnated. The cellulose composition in paper is compatible for proteins and biomolecules; it is as well as environmentally compatible, easily disposable and accessible almost everywhere. In particular, paper-based biosensor is highly suitable to point-of-need monitoring in less industrialized countries [18]. Testing of target analytes would simply involve dropping the solutions (blood or serum or sweat or saliva) on chitosan/TMBZ treated paper. This concept can promptly inform even nonprofessional users with clear and unambiguous visual color against target analytes.

2. Materials and Methods

2.1. Materials

Chitosan, poly-L-lysine (PLL), cetyltrimethylammonium bromide (CTAB), 3,3',5,5'-tetramethylbenzidine (TMBZ), hydrogen peroxide (H_2O_2), sulfuric acid (H_2SO_4), glucose,

94 sucrose, fructose, lactose, galactose, glucose oxidase (GOD from *Aspergillus Niger*), L-Ascorbic acid,
95 L-(+)-Lactic acid, lactic oxidase (LOD from *Aerococcus Viridans*) and nunc-Immuno 96-well plates
96 were purchased from Sigma-Aldrich (St. Louis, MO, USA). Maltose and acetic acid were received
97 from Fisher Scientific Company (New Jersey, USA). All experiments were performed using highly
98 pure deionized (DI) water ($>18\text{ M}\Omega\cdot\text{cm}$).

99 2.2. Preparation of water-soluble chitosan

100 Chitosan solution was prepared by dissolving 2 mg/mL chitosan in 0.1 M glacial acetic acid at
101 60°C for 30 min. Once the initial turbid colored solution turned to clear one, stopped the heating and
102 cool down at room temperature. Prepared solution was stored at room temperature for further
103 experiments.

104 2.3. Hydrogen peroxide (H_2O_2) detection using GOD and chitosan

105 Various concentrated solutions of H_2O_2 were prepared using PBS buffer prior to the sensing
106 experiments. Then, 100 μL of each concentrated H_2O_2 solutions were placed in 96 well plate and
107 subsequently, 100 μL of chitosan/TMBZ- H_2O_2 mixture solution were added in the above solution.
108 After 30 min, the ultraviolet-visible (UV-vis) spectrum of developed color was recorded using a
109 Cytation 5 spectrophotometer (BioTek Instruments, Inc., Ontario, Canada).

110 2.4. Glucose detection using GOD and chitosan

111 The experiment for glucose detection was performed as follows: (a) 50 μL of GOD (1 mg mL^{-1})
112 and 50 μL of different concentrated glucose in PBS buffer solution (pH 7.5) were incubated in 96 well
113 plates at room temperature for 30 min; (b) Then, 100 μL of chitosan/TMBZ- H_2O_2 mixture solution
114 were added in the above solution, and (c) the mixed solution was incubated at room temperature for
115 30 min and then the ultraviolet-visible (UV-vis) spectrum of developed color was recorded.

116 2.5. Lactic acid detection using LOD and chitosan

117 Lactic acid detection was performed as follows: (a) 50 μL of LOD (1 mg mL^{-1}) and 50 μL of
118 different concentrated lactic acid in PBS buffer solution (pH 7.5) were incubated in 96 well plate at
119 room temperature for 30 min; (b) Then, 100 μL of chitosan/TMBZ- H_2O_2 mixture solution were added
120 in the above solution, and (c) the mixed solution was incubated at room temperature for 30 min and
121 then the ultraviolet-visible (UV-vis) spectrum of developed color was recorded.

122 2.6. Preparation of paper-based device

123 Xerox printer (colorQube 8580, Japan) was used to fabricate the paper-based sensing device.
124 Here, Whatman filter paper (GE Healthcare UK limited, Buckinghamshire, UK) was used due to its
125 hydrophilic and biocompatible behavior as well as low cost. The fabrication process includes three
126 steps:

127 1) The desired geometry i.e., Circle, square and rhombus-like shaped was drawn in white
128 background for H_2O_2 , glucose and lactic acid detection respectively by using the software Adobe
129 Illustrator.

130 2) Printed the wax-paper with circle, square and rhombus-like shaped on the surface by using
131 the wax printer.

132 3) Melted the wax-printed paper on a hot plate at 175°C for 40 s. The wax covered area will be
133 hydrophobic, while the area without wax will be hydrophilic (sensing area).

134 2.7. Detection of H_2O_2 , glucose and lactic acid on paper-based device

135 Printed wax-paper was used to detect H_2O_2 , glucose and lactate. Firstly, 5 μL of chitosan
136 solution (2 mg/mL) was dropped on sensing area (hydrophilic area) for 5 min to dry-up. Then,
137 various concentrated target solutions (5 μL) was added on the surface, and color was developed

138 within 10 min. ImageJ software was used to process the pixel value of the detection analyte from the
139 colored image.

140 3. Results and Discussion

141 Firstly, the enzymatic activity of chitosan (1 mg/mL, 20 μ L), CTAB (1 mg/mL, 20 μ L) & PLL (1%,
142 20 μ L) were examined with 1 mL of mixture solution of peroxidase substrate TMBZ (5 mM) and
143 H_2O_2 (10 mM) separately for 5 min (Fig. 1A). A strong blue color was developed for chitosan whereas
144 the blue color developed with CTAB was not so significant in comparison to chitosan. No enzymatic
145 change of color was observed for PLL during the reaction time (Fig. 1B). The absorbance peak of
146 resulting blue solution was located at 660 nm due to the oxidation of TMBZ.

Figure 1. The catalytic activity of chitosan, CTAB and PLL: (A) schematic presentation of the sensing experiment; (B) naked eye image of the experimental results.

150 The viability of the sensing assay was performed with different kinds of reaction mixtures i.e. a)
151 chitosan/TMBZ- H_2O_2 ; b) chitosan/TMBZ; c) chitosan/ H_2O_2 ; d) TMBZ- H_2O_2 /CH₃COOH and e)
152 chitosan/CH₃COOH. As shown in Figure 2A, it was clearly observed with the naked eye that a deep
153 blue colored solution was developed by the reaction of chitosan and TMBZ- H_2O_2 complex (Fig.
154 2A-a). However, none of the other mixtures showed characteristic blue color during reaction (Fig.
155 2A- b, c,d,e). Here, the viability test of acetic acid (CH₃COOH) was confirmed since it was used to
156 dissolve chitosan. Spectroscopic study of colored solutions revealed a significant enhanced spectrum
157 peak centered at 660 nm compared to other solutions (Fig. 2B). The control experiments confirmed
158 that the glacial acetic acid did not affect the peroxidase-like activity of chitosan, and it is specific only
159 in the presence of TMBZ- H_2O_2 complex.

Figure 2. The viability of the developed sensing assay: (A) visual color of the experimental results with different mixtures; (B) Absorbance of the corresponding mixture solutions.

The optimal conditions for catalytic activity of chitosan was examined with several parameters, such as concentration, pH, TMBZ concentration, H_2O_2 concentration and reaction time. These parameters may significantly affect the enzymatic activities of chitosan. The concentration of chitosan has significant effect on enzymatic reaction. Experimental studies on different concentrated chitosan showed that the intense blue color was observed with concentration of 2 mg/mL (Fig. S1). The pH effect on the enzymatic activity of chitosan was studied in the range of pH 1–12. Results revealed that the catalytic activity of chitosan increased with increasing the pH up to 4, then sharply decreased with increasing pH values (Fig. S2A). The reaction time to get maximum catalytic activity was investigated from 0 to 20 min (Fig. S2B). The optical density of blue color generated through catalytic activity of chitosan was observed to be a maximum within 9 min, after that, it's became steady. The optimum concentration of H_2O_2 and TMBZ were also checked at 0 to 20 mM, and 1 to 10 mM respectively. The highest peroxidase-like activities of chitosan were observed with the concentration of 10 mM and 5 mM of H_2O_2 and TMBZ respectively (Fig. S2C and S2D). All the reactions were carried out at room temperature which is a convenient for sensing applications.

The enzymatic activity of chitosan depends on H_2O_2 concentration in the presence of TMBZ. The proposed method was applied to detect H_2O_2 , and the optical density of experimental results increased with increasing H_2O_2 concentration (Fig. 3A). The calibration graph of the optical density at 660 nm versus H_2O_2 concentration showed linearity in the range of 10 pM to 1 mM with a detection limit of 2.64 pM (Fig. 3A).

In addition, the catalytic reaction of chitosan is integrated with the reaction between glucose and GOD for quantitative measurement of glucose concentration (Fig. 4A). The visual detection of glucose was completed by two steps reaction process i.e., at first, glucose and GOD were reacted in PBS (pH 7.4) at room temperature for 10 min which produced H_2O_2 ; then the chitosan was added to detect H_2O_2 visually. Fig. 4B showed the linear relationship of the absorbance (660 nm) versus glucose concentration in the range of 0.2–10 μ M with the limit of detection of 0.104 μ M, which is lower than any paper reported so far to our best of knowledge. The glucose level for diabetic patient when fasting glucose is around 4–7 mM. Thus, our proposed glucose sensor is much sensitive than the practical needs. Several control experiments were also performed to check the specificity of the proposed method for glucose detection using sucrose, fructose, lactose, galactose and maltose. The results were investigated and no characteristic peaks were obtained for control experiments (Fig. S3). Hence, the visual detection method proposed for glucose was highly sensitive and selective. The naked eye image of glucose detection system is shown in Figure S4. Besides the analysis of glucose in phosphate buffer, additional experiments were conducted to ability of detecting glucose in complex biological matrix, such as blood. Results revealed that chitozymes based visual analysis is robust and selective enough to detect glucose in complex media with limit of detection of 0.23 μ M, which is promising for practical applications (Fig. S5).

Figure 3. Experimental results of H₂O₂ detection: (A) Absorbance spectra of sensing H₂O₂; (B) Calibration curve of absorbance Vs H₂O₂ concentration.

Figure 4. Experimental results of the glucose detection based on the developed Chitozyme biosensing assay: (A) Absorbance spectra of the glucose sensing experiment; (B) Calibration curve of the absorbance Vs glucose concentration.

The analytical reliability of the colorimetric measurements for the glucose assay was compared to the signal recorded in a commercial glucometer (Contour next blood glucose Meter, Japan). The analysis on the glucometer was carried out by the addition of a small drop of target solution in blood media on a disposable test strip. The glucose level appears on the meter display is in mM unit. As shown in Figure S6, commercial glucometer has the capability to detect glucose concentration with the range of 4-10 mM; whereas the proposed method showed its sensitivity up to μM level.

Furthermore, similar two step assay was conducted for lactic acid detection using LOD to produce H₂O₂ at the initial step (Fig. 5A). The absorbance of sensing results was linearly correlated with lactic acid in the range of 10 mM to 10 nM with the limit of detection of 2.8 nM (Fig. 5B). While lactic acid analysis was performed with commercial milk, detection limit was obtained at 5.3 nM (Fig. S7). Thus, the developed chitozymes based analysis of lactic acid would be a promising candidate for practical applications in complex media. A comparison study with others techniques was also performed based on recently published papers as shown in Table S1& S2. Superiority in terms of sensitivity of the present study in comparison to other techniques was clearly observed here.

Figure 5. Experimental results of the lactic acid detection based on the developed chitozyme biosensing assay: (A) Absorbance spectra of the lactic acid sensing (Inset: visual detection image); (B) Calibration curve of the absorbance vs lactic acid concentration.

Proposed chitozymes based colorimetric detection method was further demonstrated with paper-based sensor to make it more flexible in real life applications. Paper based disposable devices has been investigated extensively recently due to low-cost and field-portable in remote area applications [19]. Here, paper device was prepared by wax printer with desired geometry. Circle, square and rhombus-like shaped (diameter of 8 mm each) was drawn and printed for H₂O₂, glucose and lactic acid detection respectively. The bottom part of the paper devices was covered with wax to prevent sample leaking.

Figure 6. Experimental results of the chitozyme based paper-based biosensor: (A) Visual detection of H₂O₂, glucose and lactic acid; (B) Calibration curve of color intensity vs H₂O₂ concentration (C means control experiment with ascorbic acid & glucose); (C) Calibration curve of color intensity vs glucose concentration (C means control experiments with fructose & sucrose); (D) Calibration curve of color intensity vs lactic acid concentration (C means control experiments with ascorbic acid & glucose).

The desired detection zone on printed wax paper device was initially modified with chitosan solution (5 uL) and allowed to dry at room temperature for 10 min. Then, the detection zone was spotted with different concentrated H₂O₂, mixture containing different concentrated glucose and

GOx; mixture containing different concentrated lactic acid and LOD prepared in PBS buffer solution (pH 7.5) separately. Visual image of detections on paper-based device is shown in Figure 6A. The linear concentration range for H₂O₂ detection was 1 mM to 100 nM with the limit of detection of 64.3 nM (Fig. 6B); for glucose detection was 10 μM to 0.6 μM with the limit of detection of 0.46 μM (Fig. 6C) and for lactic acid, the detection was 10 mM to 10 nM with limit of detection 7.9 nM (Fig. 6D).

4. Conclusions

In conclusion, an user and environment friendly colorimetric read-out assay was introduced based on peroxidase-like enzymatic activity of chitosan for the first time. This straight forward signal amplification strategy was successfully applied to detect H₂O₂, glucose and lactic acid with the limit of detection of 2.64 pM, 0.104 μM and 2.8 nM respectively which was the lowest value so far reported to the best of our knowledge. Proposed assay showed its superiority and practicability in detecting the target analytes in the complex media. Chitosan-based bioassay was successfully demonstrated on the paper-based device to make a cheaper and easier medical tool. Because chitosan is a natural and biocompatible compound, this approach would open a new window in colorimetric bioassay and catalytic chemistry. The near future will be an exciting period in the chitosan-based visual biosensor research area.

Supplementary Materials: The following are available online at www.mdpi.com/link, Figure S1-S7, Table S1 and Table S2.

Acknowledgments: The authors sincerely thank the Natural Sciences and Engineering Research Council of Canada (400705) for funding this study.

Author Contributions: SN conceived the study. SRA conducted experiments and collected data. XW helped with the fabrication of paper based biosensing device. SRA wrote the manuscript with XW. All authors read and approved the manuscript.

Conflicts of Interest: The authors declare no conflict of interest.

References

1. Besant JD, Das J, Burgess IB, et al. Ultrasensitive visual read-out of nucleic acids using electrocatalytic fluid displacement. *Nat Commun.* 2015; 6: 6978.
2. Xia F, Zuo X, Yang R, et al. Colorimetric detection of DNA, small molecules, proteins, and ions using unmodified gold nanoparticles and conjugated polyelectrolytes. *Proc Natl Acad Sci USA.* 2010; 107: 10837–10841.
3. Rica RDL, Stevens MM. Plasmonic ELISA for the ultrasensitive detection of disease biomarkers with the naked eye. *Nat Nanotechnol.* 2012; 7: 821–824.
4. Gao Z, Hou L, Xu M, et al. Enhanced Colorimetric Immunoassay Accompanying with Enzyme Cascade Amplification Strategy for Ultrasensitive Detection of Low-Abundance Protein. *Sci Rep.* 2014; 4: 3966.
5. Lin Y, Ren J, Qu X. Catalytically Active Nanomaterials: A Promising Candidate for Artificial Enzymes. *Acc Chem Res.* 2014; 47: 1097–1105.
6. Ahmed SR, Kim J, Tran VT, et al. *In situ* self-assembly of gold nanoparticles on hydrophilic and hydrophobic substrates for influenza virus-sensing platform. *Sci Rep.* 2017; 7: 44495.
7. Wei H, Wang E. Nanomaterials with enzyme-like characteristics (nanozymes): next-generation artificial enzymes. *Chem Soc Rev.* 2013; 42: 6060–6093.
8. Wang X, Hu Y, Wei H. Nanozymes in bionanotechnology: from sensing to therapeutics and beyond. *Inorg Chem Front.* 2016; 3: 41–60.
9. Kumar MNVR, Muzzarelli RAA, Muzzarelli C, et al. Chitosan Chemistry and Pharmaceutical Perspectives. *Chem Rev.* 2004; 104: 6017–6084.
10. Suginta W, Khunkaewla P, Schulte A. Electrochemical Biosensor Applications of Polysaccharides Chitin and Chitosan. *Chem Rev.* 2013; 113: 5458–5479.
11. Jiang H, Chen Z, Cao H, et al. Peroxidase-like activity of chitosan stabilized silver nanoparticles for visual and colorimetric detection of glucose. *Analyst.* 2012; 137: 5560–5564.

- 290 12. Chang H, Lv J, Zhang H, et al. Qiao, Photoresponsive colorimetric immunoassay based on chitosan
291 modified AgI/TiO₂ heterojunction for highly sensitive chloramphenicol detection. *Biosens Bioelectron.*
292 2017; 87: 579–586.
- 293 13. Jv Y, Li B, Cao R. Positively-charged gold nanoparticles as peroxidase mimic and their application
294 in hydrogen peroxide and glucose detection. *Chem Commun.* 2010; 46: 8017–8019.
- 295 14. Chen W, Cai S, Ren Q, et al. Recent advances in electrochemical sensing for hydrogen peroxide: a review.
296 *Analyst.* 2012; 137: 49-58.
- 297 15. Heller A, Feldman B. Electrochemical Glucose Sensors and Their Applications in Diabetes Management.
298 *Chem Rev.* 2008; 108: 2482–2505.
- 299 16. Pundir CS, Narwal V, Batra B. Determination of lactic acid with special emphasis on biosensing methods:
300 A review. *Biosens Bioelectron.* 2016; 86: 777–790.
- 301 17. Rathee K, Dhull V, Dhull R, et al. Biosensors based on electrochemical lactate detection: A comprehensive
302 review. *Biochem Biophys Rep.* 2016; 5: 35–54.
- 303 18. Busa LSA, Mohammadi S, Maeki M, et al. Advances in Microfluidic Paper-Based Analytical Devices for
304 Food and Water Analysis. *Micromachines.* 2016; 7: 86.
- 305 19. López-Marzoa AM, Merkoçi A. Paper-based sensors and assays: a success of the engineering design and
306 the convergence of knowledge areas. *Lab Chip.* 2016; 16: 3150–3176.