A direct comparison of four high risk human papilloma virus tests versus the cobas test for detecting cervical intraepithelial neoplasia and cervical cancer

Peng Xue,^{a, b,*} Li-Li Gao,^c Jian Yin,^a Li-Li Han,^c Jing Zhao,^{a, b} Li Li,^a Samuel Seery,^d
Xue-Yan Han,^b Ting-Yuan Li,^a Yu Jiang,^b Jie Shen,^{c,#} and Wen Chen^{a,#}

^aDepartment of Epidemiology, National Cancer Center/Cancer Hospital, Chinese
Academy of Medical Sciences and Peking Union Medical College, Beijing 100021, PRC

^bSchool of Public Health, Chinese Academy of Medical Sciences and Peking Union
Medical College, Beijing 100730, PRC

^cDepartment of Women's Health Care, Beijing Obstetrics and Gynecology Hospital,
Capital Medical University Beijing Maternal and Child Health Care Hospital, Beijing
100026, PRC

^dSchool of Humanities and Social Sciences, Chinese Academy of Medical Sciences and
Peking Union Medical College, Beijing, 100730, PRC

Peng Xue, Li-Li Gao and Jian Yin contributed equally to this work.

Running title: Comparison clinical performances of four HR-HPV tests

[#]Address corresponding to Jie Shen, shenjielujun96@126.com; Wen Chen,
chenwen@cicams.ac.cn.

^{*}Present address: Peng Xue, Department of Epidemiology, National Cancer
Center/Cancer Hospital, Chinese Academy of Medical Sciences and Peking Union
Medical College, 17 South Pan Jia Yuan Lane, Beijing 100021.

ABSTRACT

This study is to evaluate performances and genotyping capabilities of four human papilloma virus (HR-HPV) tests based on real-time polymerase chain reaction (PCR) technology platforms compared with the cobas test. Discordant results were further analyzed using INNO-LiPA HPV genotyping test, the gold standard laboratory test to determine presence and type of HPV infection. Over 200 samples from Hospital patients were collected and analyzed using five HR-HPV tests. Women with positive test results were referred directly to colposcopy. If a positive result was returned, biopsies were administered for pathological classification. Clinical performances and genotyping capabilities between the four HR-HPV and cobas tests were compared and contrasted. High levels of agreement were observed, though all HR-HPV tests presented discrepancies compared with the cobas test. Cervical intraepithelial neoplasia Grade 2 or higher lesions (CIN2+) was set as the threshold, and all five tests performed with equally high sensitivity. Lower levels of specificity were observed across all five tests. Results suggest the four HR-HPV tests analyzed are as effective as the cobas test in genotyping capacities and diagnosing CIN. Therefore, these test kits should be used for HPV screening, especially in developing nations because they are cost effective and reliable. Minor discrepancies between tests are generally unavoidable though this may add complexity to the clinical decision-making process. As such, we recommend that efforts be made to standardize HPV genotyping tests as well as to optimize clinical sensitivity and specificity. Focusing on these issues will drive the development of HPV detection techniques, therefore save lives.

KEY WORDS HR-HPV, cervical screening, Real-time PCR, cobas HPV.

INTRODUCTION

Cervical cancer is associated with a substantial burden of disease and is the cause of a substantial number of deaths among women in developing countries (1). Emerging wealth evidence confirms that persistent infection with high risk human papillomavirus (HR-HPV) is associated with more than 99% of all cervical cancers (2, 3). In particular, HPV 16 and 18 are known to be the most common HPV types, leading to an estimated 70% of all cervical cancers (4, 5).

Accordingly, it has been proposed that HPV detection methods as a alone screening tool for the detection of high-grade cervical intraepithelial neoplasia and cervical cancer (6). The Hybrid Capture 2 (HC2) test is based on a signal-amplified hybridization method and has received approval from the Food and Drug Administration (FDA) yet (7-9), this test can only determine whether HPV infection is present, it neither determines specific genotypes, such as HPV 16 and 18, nor is it capable of identifying between single and multiple HPV infections.

Recently, various tests based on real time polymerase chain reaction (PCR) have been developed. Compared with traditional HC2 tests, the real-time PCR method has several the advantages, such as; convenience of use, high throughput, less time and lower cost. The most frequently administered HPV genotyping test in trials is the cobas HPV test (Roche Systems Inc., Branchburg, NJ, USA). The advantage of the cobas test is that it is a fully automated real-time PCR DNA amplification test which has been approved for screening by the FDA in 2014 (10). The cobas test is initially developed with the clinical cut-off values (11) and this enhances the level of specificity thereby maximizing the predictive value of oncogenic risk of CIN2+. Unfortunately, the cobas test is not flawless, requiring access to highly specialized, bulky instrumentation for sample pretreatment and detection purposes. The cobas HPV system weighs over 150 kg and is 166 cm wide (12). Each detection cost is also comparatively high at \$35+ or more per test. These issues make the cobas HPV test impractical and inhibitive for developing nations with less

developed infrastructures.

Recently in China, numerous commercially available HR-HPV tests based on real-time PCR have been available in hospitals and laboratories. Unfortunately, prior to 2015 China did not regulate HPV testing around clinical sensitivity or specificity thresholds, and only recently have researchers actually investigated the performances of commonly used HR-HPV tests. Therefore, it would seem necessary to conduct a more comprehensive investigation into the most accessible HPV test kits in order to promote best screening practice for health services in developing countries.

This study focused on four widely used HR-HPV test kits i.e. Tellgen, HybriBio, Liferiver and Sansure, all of which are based on real-time PCR technology. Over 200 samples were collected to appraise and compare the efficacy of each test against the cobas HPV test for detecting HR-HPV DNA. Any discrepant genotyping outcomes were compared with HPV genotyping using INNO-LiPA HPV test in order to determine levels of agreement with the gold standard laboratory diagnostic tool. Levels of sensitivity and specificity HR-HPV test used for diagnosing cervical intraepithelial neoplasia (CIN) and cervical cancer were then analyzed and contrasted. The overarching aim was to determine whether the more cost effective and more easily administered HR-HPV tests can be used for national screening campaigns in developing countries, like China.

MATERIALS AND METHODS

Study design and patients

In order to evaluate genotyping capacities and clinical performance of the four HR-HPV tests, we collected a total of 214 cytology samples with cervical lesions results from December 2016 to April 2017. Samples were originally taken from 214 women aged 23 to 65 years whom had visited Peking University First Hospital, for routine examination. Cytology samples collected were transferred into PreservCyt solution (Hologic Inc., Bedford, MA) and then stored at 4°C for testing.

HPV tests were performed using cobas (Roche Molecular Systems Inc., Roche, Shanghai, China), Tellgen (Nucleic Acid Detection Kit for HPV and 16/18 genotyping, Tellgen, Shanghai, China), HybriBio (14 HR-HPV with 16/18 Genotyping Real-time PCR Kit, HybriBio, Guangdong, China), Liferiver (HPV Genotyping Real time PCR Kit, Liferiver, Shanghai, China), and Sansure (HR-HPV DNA Fluorescence Diagnostic Kit, Sansure, Hunan, China) sequentially across specimens. Any discrepant genotyping outcomes between tests were compared using INNO-LiPA HPV test (INNO-LiPA HPV Genotyping Extra, Innogenetics, Belgium).

Women with positive test results were referred directly to colposcopy. If colposcopy returned a positive test result, four-quadrant biopsies were taken. If the colposcopy was unable to detect lesions, a random biopsy was obtained at the squamocolumnar junction in that quadrant at 2, 4, 8, or 10 o'clock.

Informed consent was requested and consequently approved by all participants in this study. This study was formally approved by the institutional review boards of the Cancer Hospital, Chinese Academy of Medical Sciences (NO.12-72/606) and National Health and Family Planning Commission of the People's Republic of China (No.2015071).

Real-time PCR HPV testing

A sample of 1 mL of liquid cytology was separated for investigation using five real-time PCR HPV tests i.e. cobas, Tellgen, HybriBio, Liferiver and Sansure all of which are based on TaqMan technology and reportedly can detect 14 HR-HPV genotypes (HPV16, 18, 31, 33, 35, 39, 45, 51, 52, 56, 58, 59, 66 and 68). Besides the 14 HPV types, Liferiver and Sansure can also detect HPV82.

The cobas HPV test has a differentiating feature; the cobas 4800 system is a highly automated instrument for DNA extraction using Roche HPV DNA kit, PCR amplification on the cobas x480 instrument and detection on the cobas z480 Analyzer. The remaining four HR-HPV tests i.e. Tellgen, HybriBio, Liferiver and Sansure perform part of manual DNA extraction using related HPV DNA kits and PCR amplification with a mixture of

multiple probes and detection on the ABI 7500 or SLAN-96P automated analyzer. The experimental conditions for the five HR-HPV tests follow the guidelines provided within the associated protocols. During each run, both positive and negative controls were included to ensure proper PCR responses were not subjected to carry over contamination. The resulting fluorescence from the reaction is then measured to determine whether HPV is present in the sample.

INNO-LiPA HPV test

The INNO-LiPA HPV test is based on reverse line hybridization using SPF10 primers (13). Part of the L1 region of the HPV genome is amplified by multiplex PCR, and includes biotin-labeled primer which is denatured and hybridized to strip. It can identify 28 HPV types containing 15 HR-HPV (16, 18, 31, 33, 35, 39, 45, 51, 52, 56, 58, 59, 68, 73 and 82), 3 probable HR-HPV (26, 53, 66) and 10 low-risk HPV (HPV6, 11, 26, 40, 43, 44, 54, 69, 70, 71 and 74). Results are interpreted using direct-vision method or utilizing the analytical software, LIRAS for LiPA HPV.

Statistical analysis

SAS 9.2 software (SAS Institute Inc., Cary, NC) was used for statistical analysis. The agreement rates and corresponding Kappa coefficients with 95% confidence intervals (CIs) were calculated to estimate the level of agreement between the four HR-HPV tests and the cobas HPV test. The Median score and Mann-Whitney U tests were calculated *p* values for the median four HR-HPV and cobas tests Cycle threshold (Ct) values for concordant vs. discordant positive specimens. Chi square test was used to compare the HPV positive rates, sensitivity, specificity, positive predictive value (PPV) and negative predictive value (NPV). *P* values less than 0.05 (two-sided) were considered statistically significant.

RESULTS

Overall HPV DNA positivity with the five HR-HPV tests

Of the 214 cytology samples, 4 cases were considered invalid due to lack of remnant DNA and were thereby excluded. The remaining 210 samples were included for analysis. Table 1 displays the HPV DNA positive results with histopathologic grading. Data demonstrate an overall positive correlation of HPV DNA with the histopathologic grading ($p < 0.0001$), except for the positive rate of HPV 18 due to the relative small number of HPV18 type cases within the sample. In 210 cytology samples, overall HPV positive rates ranged from 50.0% to 53.3%. None of the included tests performed less well in overall HPV positive rate ($p = 0.964$).

Agreement among tests

Table 2 displays independent levels of agreement for each of the four HR-HPV tests compared with the cobas test. The four HR-HPV tests, compared with the cobas test demonstrated a high level of agreement with 95.24%, 95.71%, 95.24% and 94.76% of all samples ($\kappa = 0.905, 0.914, 0.905$ and 0.895 , respectively). In cases infected with HPV 16, levels of agreement in the four HR-HPV tests against those analyzed using the cobas HPV test were 96.19%, 98.10%, 98.10% and 95.24% ($\kappa = 0.908, 0.953, 0.954$ and 0.885 , respectively). In cases infected with HPV 18, agreement with the cobas HPV test were 99.52%, 98.10%, 98.57% and 98.57% ($\kappa = 0.939, 0.740, 0.835$ and 0.835 , respectively). HR-HPVs other than HPV types 16 and 18 were also analyzed, and again performed with equally high levels of agreement compared with those of the cobas HPV test with 90.48%, 91.90%, 89.52% and 87.62% ($\kappa = 0.783, 0.813, 0.768$ and 0.726 , respectively). 22 discrepancies were eventually resolved using the INNO-LiPA HPV test (see Table 4). 10 cases were negative while LiPA system identified 12 additional positive cases. Histopathologic analysis revealed 6 cases were CIN2+ and 16 were < CIN2. Median Ct values of concordance and discordance in the four HR-HPV tests and cobas test results are presented in Table 3. All concordant cases between the four HR-HPV as well as the cobas HPV tests were significantly lower median cobas Ct values compared to those discordant cases ($p < 0.001$).

Clinical performance of the five HR-HPV tests for detection of CIN2+

Clinical performances of each of the HR-HPV tests as well as the cobas HPV test to a reference standard of CIN2+ were analyzed (see Table 5). Data revealed that all tests were similarly sensitive with 94.59%, 94.59%, 94.59%, 95.59% and 93.24%, respectively. However, each of the tests was significantly less specific. HR-HPV tests i.e. Tellgen, HybriBio, Liferiver, Sansure and the cobas HPV test resulted in 72.79%, 73.53%, 71.32%, 69.86% and 73.53%, specificity. Overall, there was no significant difference in either sensitivity ($p = 0.971$) or specificity ($p = 0.953$) across these HR-HPV tests for detecting CIN2+.

DISCUSSION

This study focused on four readily available and widely used HR-HPV test kits i.e. the Tellgen, HybriBio, Liferiver, and the Sansure, all of which are based on real-time PCR technology. Clinical performance and genotyping capabilities were compared with the cobas HPV test. Discrepancies were further compared with HPV genotyping using INNO-LiPA HPV test, the gold standard laboratory diagnostic test. Since approval by the U.S. FDA and validation by numerous studies (14, 15), the cobas HPV test is as effective as the Hybrid Capture 2 HPV test which has become a gold standard screening tool for evaluating the efficacy of the newly developed HPV methods (16). In addition, INNO-LiPA has been widely used in clinical trials focusing on HPV vaccine research for the identification of specific sequences in the L1 region of the HPV genome of 28 HPV types (17).

Considering that the four HR-HPV tests; the Tellgen, HybriBio, Liferiver, and Sansure have become widely available within hospitals and laboratories in China, there were few studies which compare and then verify test performances simultaneously. In this study, by using CIN2+ as a threshold and reference standard, it became possible to compare levels of sensitivity and specificity of all four HR-HPV tests which ultimately performed very

similar when compared with the cobas HPV test. Similar HPV-DNA detection rates were also discovered compared with both the cobas HPV test and the INNO-LiPA HPV test. This study ultimately found that these HR-HPV tests and cervical histopathology positively correlate. All HPV tests detected HPV from CIN2+ samples in approximately 90.3% to 100.0% of cases therefore this analysis demonstrates that these cheaper, simpler HR-HPV tests perform equally at detecting HPV infection in cervical lesions.

All four HR-HPV tests demonstrated high levels of agreement with the cobas HPV test for detecting HR-HPV DNA. However, discrepancies detected by these HR-HPV tests in 22 (10.5%) of the 210 samples, thus genotyping using INNO-LiPA was performed to explore potential causes. The cobas-negative/four HR-HPV tests-positive samples resulted in approximately 70% of the all discrepancies, the four HR-HPV-positive results found in half of the cases were not confirmed by LiPA, whereas, the LiPA detection test detected the presence of different low-risk genotypes, thus representing false-positive results though the four HR-HPV tests. These differences may lead to over-referral to colposcopy and potentially false prognostic stress in healthy women.

On the other hand, 7 samples were cobas-positive/four HR-HPV-negative; 5 samples returned positive for HR-HPV using the LiPA system, with 3 cases bearing a CIN2+, which were later identified as HPV 33 and 16. It should be noted that LiPA does not have an established clinical cut off value and it was expected that the positive sample below the critical cut off point would be detected using these HR-HPV tests. Genotype-specific results detected using these HR-HPV tests showed that samples with discordant genotyping results had Ct values significantly closer to the test limits of detection which was consistent with genotyping samples. This indicates that these samples may have contained a lower viral load based on high Ct values and are more likely to produce discordant results across the tests (18). These minor discrepancies were considered unavoidable due to different cut-off values and a lack of standardization. No test is absolutely sensitive or specific and we therefore suggest that cut-off values for HR-HPV

tests need to be adjusted for optimization in order to reduce false-positive results and without sacrificing sensitivity in detecting HR-HPV.

This study had some limitations that should be addressed. First of all, this is a preliminary study with a relatively small sample ($n = 210$). Those results led to a high-quality appraisal of four commonly used HR-HPV tests however due to the small sample size our recommendations remain tentative. Additional research with a large sample size is required to verify these findings. Secondly, the total HR-HPV concordance rates were perhaps overestimated due to the fact that detection methods identify HPV types as a pool rather than each HPV Genotype. Further research is required to analyse test capabilities for each HPV genotype. A more powerful form of sequencing generation could also be used, which may provide more definitive HPV genotype information. Lastly, manual operation of the four HR-HPV tests were time consuming and labor intensive and were therefore more likely to suffer potential sample cross-contamination. We suggest the manufacturer should use an automated DNA extraction system if conditions are available in order to minimize potential cross-contamination as well as to reduce hands-on time.

Ultimately, this study demonstrates that the four included HR-HPV tests have strong levels of agreement and similar clinical performance when compared with the cobas HPV test. Therefore, nations with less developed healthcare systems should consider using one of these test kits for HPV and for national cervical cancer screening. Slight discrepancies between tests are generally unavoidable. It is therefore recommended that HPV genotyping tests should be optimized and standardized around test sensitivity and specificity. This will enhance HPV detection techniques and reduce the burden of cervical cancer in developing regions and countries. Further research with larger samples and comparison with various HPV detection methods is of course required.

ACKNOWLEDGMENTS

This study was supported by National Natural Science Foundation of China

(NO.71373166)

The authors have declared that no competing interest exists.

REFERENCES

1. Shi JF, Canfell K, Lew JB, Qiao YL. 2012. The burden of cervical cancer in China: synthesis of the evidence. *Int J Cancer* 130:641-52.10.1002/ijc.26042
2. Smith JS, Lindsay L, Hoots B, Keys J, Franceschi S, Winer R, Clifford GM. 2007. Human papillomavirus type distribution in invasive cervical cancer and high-grade cervical lesions: a meta-analysis update. *Int J Cancer* 121:621-32.10.1002/ijc.22527
3. Sundstrom K, Eloranta S, Sparen P, Arnheim Dahlstrom L, Gunnell A, Lindgren A, Palmgren J, Ploner A, Sanjeevi CB, Melbye M, Dillner J, Adami HO, Ylitalo N. 2010. Prospective study of human papillomavirus (HPV) types, HPV persistence, and risk of squamous cell carcinoma of the cervix. *Cancer Epidemiol Biomarkers Prev* 19:2469-78.10.1158/1055-9965.EPI-10-0424
4. Kjaer SK, Frederiksen K, Munk C, Iftner T. 2010. Long-term absolute risk of cervical intraepithelial neoplasia grade 3 or worse following human papillomavirus infection: role of persistence. *J Natl Cancer Inst* 102:1478-88.10.1093/jnci/djq356
5. Munoz N, Bosch FX, Castellsague X, Diaz M, de Sanjose S, Hammouda D, Shah KV, Meijer CJ. 2004. Against which human papillomavirus types shall we vaccinate and screen? The international perspective. *Int J Cancer* 111:278-85.10.1002/ijc.20244
6. Alemany L, Saunier M, Alvarado-Cabrero I, Quiros B, Salmeron J, Shin HR, Pirog EC, Guimera N, Hernandez-Suarez G, Felix A, Clavero O, Lloveras B, Kasamatsu E, Goodman MT, Hernandez BY, Laco J, Tinoco L, Geraets DT, Lynch CF, Mandys V, Poljak M, Jach R, Verge J, Clavel C, Ndiaye C, Klaustermeier J, Cubilla A, Castellsague X, Bravo IG, Pawlita M, Quint WG, Munoz N, Bosch FX, de Sanjose S, Group HVS. 2015. Human papillomavirus DNA prevalence and type distribution in anal carcinomas worldwide. *Int J Cancer* 136:98-107.10.1002/ijc.28963
7. Giorgi-Rossi P, Franceschi S, Ronco G. 2012. HPV prevalence and accuracy of HPV testing to detect high-grade cervical intraepithelial neoplasia. *Int J Cancer* 130:1387-94.10.1002/ijc.26147

298 8. Li J, Zhang D, Zhang Y, Wang X, Lin Y, Hu L. 2011. Prevalence and genotype distribution of human
299 papillomavirus in women with cervical cancer or high-grade precancerous lesions in Chengdu, western
300 China. *Int J Gynaecol Obstet* 112:131-4. [10.1016/j.ijgo.2010.08.010](https://doi.org/10.1016/j.ijgo.2010.08.010)

301 9. Lorincz AT. 1996. Hybrid Capture method for detection of human papillomavirus DNA in clinical
302 specimens: a tool for clinical management of equivocal Pap smears and for population screening. *J Obstet*
303 *Gynaecol Res* 22:629-36, <https://www.ncbi.nlm.nih.gov/pubmed/9037955>

304 10. Cui M, Chan N, Liu M, Thai K, Malaczynska J, Singh I, Zhang D, Ye F. 2014. Clinical performance of
305 Roche Cobas 4800 HPV Test. *J Clin Microbiol* 52:2210-1. [10.1128/JCM.00883-14](https://doi.org/10.1128/JCM.00883-14)

306 11. Wright TC, Jr., Stoler MH, Behrens CM, Apple R, Derion T, Wright TL. 2012. The ATHENA human
307 papillomavirus study: design, methods, and baseline results. *Am J Obstet Gynecol* 206:46 e1-46
308 e11. [10.1016/j.ajog.2011.07.024](https://doi.org/10.1016/j.ajog.2011.07.024)

309 12. Shah SS, Senapati S, Klacsmann F, Miller DL, Johnson JJ, Chang HC, Stack MS. 2016. Current
310 Technologies and Recent Developments for Screening of HPV-Associated Cervical and Oropharyngeal
311 Cancers. *Cancers (Basel)* 8. [10.3390/cancers8090085](https://doi.org/10.3390/cancers8090085)

312 13. Kleter B, van Doorn LJ, Schrauwen L, Molijn A, Sastrowijoto S, ter Schegget J, Lindeman J, ter
313 Harmsel B, Burger M, Quint W. 1999. Development and clinical evaluation of a highly sensitive
314 PCR-reverse hybridization line probe assay for detection and identification of anogenital human
315 papillomavirus. *J Clin Microbiol* 37:2508-17, <https://www.ncbi.nlm.nih.gov/pubmed/10405393>

316 14. Heideman DA, Hesselink AT, Berkhof J, van Kemenade F, Melchers WJ, Daalmeijer NF, Verkuijten M,
317 Meijer CJ, Snijders PJ. 2011. Clinical validation of the cobas 4800 HPV test for cervical screening
318 purposes. *J Clin Microbiol* 49:3983-5. [10.1128/JCM.05552-11](https://doi.org/10.1128/JCM.05552-11)

319 15. Stoler MH, Wright TC, Jr., Sharma A, Apple R, Gutekunst K, Wright TL, Group AHS. 2011. High-risk
320 human papillomavirus testing in women with ASC-US cytology: results from the ATHENA HPV study. *Am*
321 *J Clin Pathol* 135:468-75. [10.1309/AJCPZ5JY6FCVNMOT](https://doi.org/10.1309/AJCPZ5JY6FCVNMOT)

322 16. Chen W, Yu LL, Wang H, Fu CJ, Chen F, Cao YQ, Kang LN, Zhang X, Zhao FH, Geng L, Yu L. 2012.
323 [Evaluation of cobas 4800 high-risk HPV test as a tool in cervical cancer screening and cytology triage].
324 *Zhonghua Zhong Liu Za Zhi* 34:543-8. [10.3760/cma.j.issn.0253-3766.2012.07.014](https://doi.org/10.3760/cma.j.issn.0253-3766.2012.07.014)

325 17. Mollers M, King AJ, Knol MJ, Scherpenisse M, Meijer CJ, van der Klis FR, de Melker HE. 2015.
326 Effectiveness of human papillomavirus vaccine against incident and persistent infections among young
327 girls: Results from a longitudinal Dutch cohort study. *Vaccine* 33:2678-83.10.1016/j.vaccine.2015.04.016
328 18. Cubie HA, Cuschieri K. 2013. Understanding HPV tests and their appropriate applications.
329 *Cytopathology* 24:289-308.10.1111/cyt.12083
330
331
332
333
334
335
336
337
338
339
340
341
342
343
344
345
346
347
348
349
350
351
352
353
354
355
356
357
358

Table 1 Positivity rates of the five HPV tests according to histopathology classification, n (%)

HPV test	HPV type	Normal (n = 121)	CIN1 (n = 15)	CIN2 (n = 39)	CIN3 (n = 31)	SCC (n = 4)	ALL (n = 210)	<i>P</i> value
Tellgen	HR-HPV	23 (19.01)	14 (93.33)	37 (94.87)	29 (93.55)	4 (100.0)	107 (50.95)	<0.0001
	HPV16	13 (10.74)	3 (20.00)	20 (51.28)	20 (64.51)	4 (100.0)	60 (28.57)	<0.0001
	HPV18	2 (1.65)	2 (13.33)	3 (7.69)	1 (3.22)	0 (0)	8 (3.81)	0.0765
	HPV others	16 (13.22)	12 (80.00)	29 (74.35)	13 (41.94)	0 (0)	70 (33.33)	<0.0001
HybriBio	HR-HPV	24 (19.83)	12 (80.00)	37 (94.87)	29 (93.55)	4 (100.0)	106 (50.48)	<0.0001
	HPV16	13 (10.74)	2 (13.33)	19 (48.72)	20 (64.51)	4 (100.0)	58 (27.62)	<0.0001
	HPV18	2 (1.65)	2 (13.33)	2 (5.13)	1 (3.22)	0 (0)	7 (3.33)	0.1014
	HPV others	20 (15.53)	11 (73.33)	26 (66.67)	10 (32.26)	0 (0)	67 (31.90)	<0.0001
Liferiver	HR-HPV	25 (20.66)	14 (93.33)	38 (97.44)	28 (90.32)	4 (100.0)	109 (51.90)	<0.0001
	HPV16	14 (11.57)	3 (20.00)	19 (48.72)	20 (64.51)	4 (100.0)	60 (28.57)	<0.0001
	HPV18	3 (2.48)	2 (13.33)	3 (7.69)	2 (6.45)	0 (0)	10 (4.76)	0.1646

	HPV others	22 (18.18)	13 (86.67)	30 (76.92)	13 (41.94)	0 (0)	78 (37.14)	<0.0001
Sansure	HR-HPV	29 (23.97)	12 (80.00)	37 (94.87)	30 (96.77)	4 (100)	112 (53.33)	<0.0001
	HPV16	15 (12.40)	2 (13.33)	18 (46.15)	21 (67.74)	4 (100)	60 (28.57)	<0.0001
	HPV18	3 (2.48)	3 (20.00)	3 (7.69)	1 (3.22)	0 (0)	10 (4.76)	0.0517
	HPV others	25 (20.66)	11 (73.33)	30 (76.92)	12 (38.71)	0 (0)	78 (37.14)	<0.0001
cobas	HR-HPV	25 (20.66)	11 (73.33)	36 (92.31)	29 (93.55)	4 (100.0)	105 (50.00)	<0.0001
	HPV16	13 (10.74)	4 (26.67)	20 (51.28)	21 (67.74)	4 (100.0)	62 (29.52)	<0.0001
	HPV18	2 (1.65)	2 (13.33)	4 (10.26)	1 (3.22)	0 (0)	9 (4.28)	0.0433
	HPV others	19 (15.70)	10 (66.67)	25 (64.10)	11 (35.48)	1 (25.00)	66 (31.43)	<0.0001

360 Abbreviations: CIN1, cervical intraepithelial neoplasia grade1; CIN2, cervical intraepithelial neoplasia grade 2; CIN3, cervical intraepithelial
361 neoplasia grade 3; SCC, squamous cell carcinoma; HR-HPV, high-risk human papillomavirus.

362

363

364

365

366

367

368 **Table 2** Concordance rates between the results of the four HPV tests and the cobas test

		cobas		% Agreement (95%CI)			Kappa (95%CI)	<i>p</i> value
				Overall	Positive	Negative		
Tellgen	Positive	Positive (all)	Negative					
		101	4	95.24	94.39	96.12	0.905	0.5271
	Negative	6	99	(91.46 - 97.39)	(88.30 - 97.40)	(90.44 - 98.48)	(0.847 - 0.962)	
		Positive for HPV16	Negative for HPV16					
	Positive	57	3	96.19	91.94	97.97	0.908	0.4795
	Negative	5	145	(92.66 - 98.06)	(82.47 - 96.51)	(94.21 - 99.31)	(0.845 - 0.970)	
	Positive	Positive for HPV18	Negative for HPV18					0.3173
		8	0	99.52	88.89	100.00	0.939	
	Negative	1	201	(97.35 - 99.92)	(56.05 - 98.01)	(98.12-100)	(0.819 - 1.000)	
		Positive for others	Negative for others					
	Positive	58	12	90.48	87.88	91.67	0.783	0.3711
	Negative	8	132	(85.75 - 93.75)	(77.86 - 93.73)	(86.00 - 95.17)	(0.692 - 0.873)	
		Positive (all)	Negative					

HybriBio	Positive	101	5	95.71	96.19	95.24	0.914	0.7389
	Negative	4	100	(92.06 - 99.73)	(90.61 - 98.51)	(89.33 - 97.95)	(0.860 - 0.969)	
		Positive for HPV16	Negative for HPV16					
	Positive	58	0	98.10	93.55	100.00	0.953	0.0455
	Negative	4	148	(95.21 - 99.26)	(84.55 - 97.46)	(97.47 - 100)	(0.908 - 0.999)	
		Positive for HPV18	Negative for HPV18					
	Positive	6	1	98.10	66.67	99.50	0.740	0.3173
	Negative	3	200	(95.21 - 99.26)	(35.42 - 87.94)	(97.24 - 99.91)	(0.496 - 0.985)	
Liferiver		Positive for others	Negative for others					
	Positive	58	9	91.90	87.88	93.75	0.813	0.8084
	Negative	8	135	(87.42 - 94.88)	(77.86 - 93.73)	(88.55 - 96.68)	(0.728 - 0.898)	
		Positive (all)	Negative					
	Positive	102	7	95.24	97.14	93.33	0.905	0.2059
	Negative	3	98	(91.46 - 97.39)	(91.93 - 99.02)	(86.87 - 96.73)	(0.847 - 0.962)	
		Positive for HPV16	Negative for HPV16					
	Positive	59	1	98.10	95.16	99.32	0.954	0.3173
Negative	3	147	(95.21 - 99.26)	(86.71 - 98.34)	(96.27 - 99.88)	(0.909 - 0.999)		
		Positive for HPV18	Negative for HPV18					
	Positive	8	2	98.57	88.89	99.00	0.835	0.5637
	Negative	1	199	(95.88 - 99.51)	(56.05 - 98.01)	(96.45 - 99.73)	(0.651 - 1.000)	
		Positive for	Negative for					

		others	others					
	Positive	61	17	89.52	92.42	88.19	0.768	0.0105
	Negative	5	127	(84.65 - 92.98)	(83.46 - 96.72)	(81.91 - 92.50)	(0.677 - 0.859)	
		Positive (all)	Negative					
Sansure	Positive	103	9	94.76	98.10	91.43	0.895	0.0348
	Negative	2	96	(90.87 - 97.05)	(93.32 - 99.48)	(84.51 - 95.43)	(0.835 - 0.955)	
		Positive for HPV16	Negative for HPV16					
	Positive	56	4	95.24	90.32	97.30	0.885	0.5271
	Negative	6	144	(91.46 - 97.39)	(80.45 - 95.49)	(93.26 - 98.94)	(0.815 - 0.954)	
		Positive for HPV18	Negative for HPV18					
	Positive	8	2	98.57	88.89	99.00	0.835	0.5637
	Negative	1	199	(95.88 - 99.51)	(56.50 - 98.01)	(96.45 - 99.73)	(0.651 - 1.000)	
		Positive for others	Negative for others					
	Positive	59	19	87.62	89.39	86.81	0.726	0.0189
	Negative	7	125	(82.48 - 91.41)	(79.69 - 94.77)	(80.31 - 91.39)	(0.629 - 0.824)	

369
370
371
372
373
374
375
376

377

378

379

380

381 **Table 3.** Ct values for concordant and discordant four HPV tests and the cobas test results.

	Ct (all) Median	Ct (HPV16) Median	Ct (HPV18) Median	Ct (HPV others) Median
Tellgen+/cobas+	(20.03, 27.80)	(18.42, 27.90)	(19.44, 28.25)	(20.75, 27.55)
Tellgen-/cobas+	(0, 38.75)	(0, 36.90)	(0, 35.10)	(0, 38.90)
Tellgen+/cobas-	(25.42, 0)	(26.30, 0)	NA	(24.71, 0)
HybriBio+/cobas+	(28.96, 27.75)	(27.06, 28.15)	(35.83, 27.30)	(30.00, 27.55)
HybriBio-/cobas+	(0, 37.80)	(0, 35.90)	(0, 35.1)	(0, 38.78)
HybriBio+/cobas-	(29.41, 0)	NA	(37.42, 0)	(28.54, 0)
Liferiver+/cobas+	(24.24, 27.95)	(23.47, 28.40)	(22.84, 28.25)	(25.05, 27.78)
Liferiver-/cobas+	(0, 38.80)	(0, 36.9)	(0, 35.10)	(0, 39.00)
Liferiver+/cobas-	(34.84, 0)	(37.48, 0)	(37.09, 0)	(34.32, 0)
Sansure+/cobas+	(27.83, 27.80)	(27.76, 27.80)	(30.69, 28.25)	(27.73, 27.60)
Sansure-/cobas+	(0, 38.75)	(0, 38.70)	(0, 35.10)	(0, 38.80)
Sansure+/cobas-	(36.03, 0)	(37.83, 0)	(35.10, 0)	(34.22, 0)

382 Abbreviations: Ct: Cycle threshold; Tellgen Cutoff: 30; HybriBio Cutoff: 36; Liferiver Cutoff: 38; Sansure Cutoff: 39; cobas Cutoff: 40.5; NA:

383 not available;

384

385 **Table4** Discordant results between the four HPV tests and cobas HPV test for HR-HPV compared to histopathology and

386 INNO-LiPA

Subject ID	Result by: histopathology	Tellgen	HybriBio	Liferiver	Sansure	cobas	INNO-LiPA	Test result interpretation
A004	Normal	<u>POS</u>	NEG	<u>POS</u>	<u>POS</u>	NEG	NGE	False positive (Tellgen, Liferiver and Sansure)
A008	CIN2	<u>NEG</u>	<u>NEG</u>	POS	<u>NEG</u>	<u>NEG</u>	HPV82	False negative (Sansure) or untargeted ^a
A009	CIN3	POS	POS	<u>NEG</u>	POS	POS	HPV33	False negative (Liferiver)
A012	Normal	POS	POS	POS	<u>NEG</u>	<u>NEG</u>	HPV39,66	False negative (Sansure and cobas)
A020	CIN3	POS	<u>NEG</u>	POS	POS	POS	HPV16	False negative (HybriBio)
A040	CIN2	POS	POS	POS	POS	<u>NEG</u>	HPV52	False negative (cobas)
A042	CIN3	<u>NEG</u>	POS	POS	POS	POS	HPV16	False negative (Tellgen)
A054	Normal	NEG	NEG	<u>POS</u>	NEG	NEG	NEG	False positive (Liferiver)
A064	CIN3	POS	POS	<u>NEG</u>	POS	<u>NEG</u>	HPV53	False negative (Liferiver and cobas) or untargeted
A075	Normal	<u>POS</u>	NEG	NEG	NEG	NEG	NEG	False positive (Tellgen)
A100	Normal	POS	POS	POS	POS	<u>NEG</u>	HPV52	False negative (cobas)
A112	Normal	NEG	NEG	NEG	NEG	<u>POS</u>	NEG	False positive (cobas)
A127	Normal	NEG	NEG	NEG	<u>POS</u>	NEG	NEG	False positive (Sansure)
A128	Normal	NEG	NEG	<u>POS</u>	NEG	NEG	NEG	False positive (Liferiver)
A131	Normal	NEG	NEG	NEG	<u>POS</u>	NEG	NEG	False positive (Sansure)
A132	Normal	NEG	NEG	NEG	<u>POS</u>	NEG	NEG	False positive (Sansure)
A133	Normal	NEG	NEG	NEG	<u>POS</u>	NEG	NEG	False positive (Sansure)
A187	Normal	NEG	NEG	NEG	NEG	<u>POS</u>	NEG	False positive (cobas)

A192	Normal	POS	<u>NEG</u>	POS	POS	POS	HPV16,56	False negative (HybriBio)
A197	Normal	NEG	NEG	NEG	<u>POS</u>	NEG	HPV43	False negative (cobas)
A202	Normal	<u>NEG</u>	POS	POS	POS	POS	HPV18,31,52,59,74	False negative (Tellgen)
A209	Normal	<u>NEG</u>	POS	<u>NEG</u>	<u>NEG</u>	<u>NEG</u>	HPV18,51	False negative (Tellgen, Liferive, Sansure and cobas)

387 Abbreviations: POS: Positive; NEG: Negative; CIN2: cervical intraepithelial neoplasia grade 2; CIN3: cervical intraepithelial neoplasia grade 3;

388 a: Lesion caused by HPV genotypes not targeted by cobas or the four HR-HPV tests; Underline: discordant results with LiPA test.

389

390

391

392

393

394

395

396

397

398

399

400

401 **Table 5** Clinical performance of these HR-HPV tests for detection of CIN2+ in women with positive HPV results.

Endpoint	HPV test	Sensitivity (95%CI)	Specificity (95%CI)	PPV (95%CI)	NPV (95%CI)
CIN2+ (n = 74)	Tellgen	94.59 (86.91 - 97.88)	72.79 (64.77 - 79.57)	65.42 (56.02 - 73.76)	96.12 (90.44 - 98.48)
	HybriBio	94.59 (86.91 - 97.88)	73.53 (65.54 - 80.22)	66.04 (56.60 - 74.35)	96.15 (90.53 - 98.49)
	Liferiver	94.59 (86.91 - 97.88)	71.32 (63.22 - 78.26)	64.22 (54.88 - 72.59)	96.04 (90.26 - 98.45)
	Sansure	95.95 (88.75 - 98.61)	69.86 (61.68 - 76.93)	63.39 (54.17 - 71.73)	96.94 (91.38 - 98.95)
	cobas	93.24 (85.14 - 97.08)	73.53 (65.54 - 80.22)	65.71 (56.23 - 74.09)	95.24 (89.33 - 97.95)

402 Abbreviations: CIN2+: cervical intraepithelial neoplasia grade 2 or higher; CI: confidence interval; PPV: positive predictive value; NPV:
403 negative predictive value.

404