Restored tactile sensation improves neuroprosthetic arm control

Sharlene N Flesher^{1,2,3,4}, John E Downey^{1,2,5}, Jeffrey M Weiss⁶, Christopher L Hughes^{1,2}, Angelica J Herrera^{1,2}, Elizabeth C Tyler-Kabara^{1,6,7,8}, Michael L Boninger^{1,6,8,9}, Jennifer L Collinger^{1,2,6,9*}, Robert A Gaunt^{1,2,6*}

¹ Department of Bioengineering, University of Pittsburgh, Pittsburgh, PA

² Center for the Neural Basis of Cognition, Pittsburgh, PA

³ Department of Neurosurgery, Stanford University, Palo Alto, CA

⁴ Department of Electrical Engineering, Stanford University, Palo Alto, CA

⁵ Department of Organismal Biology, University of Chicago, Chicago, IL

⁶ Department of Physical Medicine and Rehabilitation, University of Pittsburgh, Pittsburgh, PA

⁷ Department of Neurological Surgery, University of Pittsburgh, Pittsburgh, PA

⁸ McGowan Institute for Regenerative Medicine, University of Pittsburgh, Pittsburgh, PA

⁹ Department of Veterans Affairs, Pittsburgh, PA

*These authors contributed equally to this work.

Summary paragraph: The sense of touch is critical for skillful hand control¹⁻³, but is largely missing for people who use prosthetic devices. Instead, prosthesis users rely heavily on visual feedback, even though state transitions that are necessary to skillfully interact with objects, such as object contact, are relayed more precisely through tactile feedback⁴⁻⁶. Here we show that restoring tactile sensory feedback, through intracortical microstimulation of the somatosensory cortex⁷, enables a person with a bidirectional intracortical brain-computer interface to improve their performance on functional object transport tasks completed with a neurally-controlled prosthetic limb. The participant had full visual feedback and had practiced the task for approximately two years prior to these experiments. Nevertheless, successful trial times on a commonly used clinical upper limb assessment task were reduced from a median time of 20.9 s (13.1 - 40.5 s interquartile range) to 10.2 s (5.4 - 18.1 s interquartile range) when vision was supplemented with microstimulation-evoked cutaneous percepts that were referred to different fingers and were graded in intensity based on real-time prosthesis contact forces. Faster completion times were primarily due to a reduction in the amount of time spent attempting to grasp objects. These results demonstrate the importance of tactile sensations in upper-limb control and the utility of creating bidirectional brain-computer interfaces to restore this stream of information using intracortical microstimulation.

32 We use our hands to interact with our environment, often by exploring and manipulating objects.
 33 Without tactile somatosensory feedback, even simple manipulation tasks become clumsy and slow¹⁻³.
 34 Outside of investigational settings, this source of feedback is rarely provided for prosthetic devices⁸, and
 35 in the context of human brain-computer interfaces (BCIs), has only recently become possible^{7,9-11}. These
 36 studies have begun to describe the perceptual characteristics of cortical stimulation, however, the
 37 potential benefits of a bidirectional BCI on function have remained unexplored. This is despite the fact
 38 that the need for somatosensory feedback in BCIs has long been suggested as the next step towards
 39 complete upper-limb restoration¹²⁻¹⁴ and cited by amputees as a desired feature¹⁵⁻¹⁷. Here we show that
 40 a bidirectional BCI (Fig. 1) that provides these tactile percepts improves performance in functional object
 41 transport tasks using a BCI-controlled robotic arm. The percepts were driven in real-time by sensors in a
 42 prosthetic hand (Fig. 1c,d), evoked through intracortical microstimulation (ICMS) of area 1 of
 43 somatosensory cortex (S1) and experienced by a participant as originating from his own palm and
 44 fingers.

45 We used two tasks to evaluate performance: an object transfer task (Fig. 1f) and a modified version of
 46 the Action Research Arm Test (ARAT)¹⁸ (Fig. 1g). Both tasks were completed using the Modular Prosthetic
 47 Limb (MPL)¹⁹. The robotic arm was controlled using neural activity recorded from two 88-channel
 48 microelectrode arrays implanted chronically in primary motor cortex (M1) (Fig. 1b) of a human
 49 participant with tetraplegia resulting from a cervical spinal cord injury. Five degrees-of-freedom (DoF),
 50 consisting of 3D endpoint translation, pronation/supination of the wrist, and hand grasp aperture (Fig.
 51 1a)—with the hand in a power grasp conformation—were continuously and simultaneously controlled by
 52 the participant during all tasks (Fig. 1e). Tactile feedback was delivered in the first four experimental
 53 sessions by ICMS through two 32-channel microelectrode arrays implanted in area 1 of S1 (Fig. 1b).
 54 Stimulation pulses were delivered at 100 pulses per second and pulse amplitude was modulated linearly
 55 by the reaction torques measured at the metacarpophalangeal joint of the fingers on the MPL (Fig. 1d).
 56 Pulse trains were delivered to electrodes which, when stimulated, evoked percepts on corresponding
 57 fingers (Fig. 1c).

58 We first tested the effect of providing ICMS-induced tactile feedback on functional performance using
 59 an object transfer task that was familiar to the participant. The goal was to transport a compliant object

60 across the workspace (Fig. 1f) as many times as possible in two minutes (Supplemental Video 1). We
 61 compared the number of transfers completed during four sessions with ICMS to four sessions without
 62 ICMS. Each session consisted of five two-minute trials. Across a total of 20 trials with ICMS, 352 transfers
 63 were completed compared to 315 transfers in the 20 trials without ICMS (Table 1). The number of
 64 transfers increased from 15.8 ± 3.8 transfers per trial to 17.8 ± 2.4 transfers per trial with ICMS, though
 65 this difference was not statistically significant ($t_{38} = -2.02$, $P = 0.050$, t -test). However, we observed
 66 qualitative improvements during the task that led us to examine the data in more detail.

67 The object transfer task can be broken up into grasp, transport and release phases. We defined these
 68 phases using the physical location of the MPL hand. The transport zone consisted of a region 22.5 cm
 69 wide and centered on the starting location of the hand at the beginning of a trial. The grasp zone was
 70 located to the left side of the transport zone, while the release zone was located to the right (Fig. 1f).
 71 We first examined the amount of time spent in each movement zone per transfer. We found that the
 72 time spent in the grasp zone decreased from 3.3 ± 1.2 s per transfer without ICMS to 2.3 ± 0.4 s per
 73 transfer with ICMS ($t_{38} = 3.3$, $P = 0.002$, t -test, Fig. 2a) while time spent in the release zone decreased
 74 from 2.8 ± 1.0 s per transfer without ICMS to 2.3 ± 0.5 s per transfer with ICMS ($t_{38} = 2.0$, $P = 0.048$, t -
 75 test, Fig. 2a). Time spent in the transport zone per transfer was no different with or without ICMS ($2.1 \pm$
 76 0.6 s without ICMS, 2.3 ± 0.3 s with ICMS, $t_{38} = -1.3$, $P = 0.206$, t -test, Fig. 2a). To uncover the reason
 77 behind the lower grasp times with ICMS, we examined the total distance travelled while the MPL was in
 78 the grasp zone. We found that there was significantly more movement in the grasp zone in trials without
 79 ICMS compared to trials with ICMS (44.2 ± 13.1 cm/transfer without ICMS, 32.4 ± 5.9 cm/transfer with
 80 ICMS, $t_{38} = 3.7$, $P = 0.0007$, t -test, Fig. 2b). This suggests that in trials without ICMS, the additional time
 81 was used to move the hand into an ideal configuration to grasp the object. This effect is further
 82 illustrated by comparing the spatial distributions of time spent across the workspace per transfer (Fig.
 83 2c). With ICMS-evoked sensations, the participant spent less time in the immediate vicinity of the object.

84 We then compared performance on a modified version of the ARAT¹⁸, which is a clinically validated test
 85 of unilateral upper-limb function and one that has been used previously to assess arm control
 86 performance in BCI systems^{20,21}. We placed different objects on the left side of the workspace, one at a
 87 time, and asked the participant to grasp the object and place it on a raised platform on the right side of

the table as quickly as possible (Fig. 1g and Supplemental Videos 2-4). A score of three was awarded if the task was completed in under five seconds, a score of two was awarded if the task was completed in under two minutes and a score of one was awarded if the object was touched but the task was not completed in two minutes. A score of zero was awarded otherwise. Each of the nine objects were attempted three times, for a total of 27 trials per ARAT session. The final score was the sum of the best score of the three attempts for each object.

Prior to these experiments, the participant had performed 23 ARAT sessions over a period of 23 months using several different control schemes, including four preliminary sessions with ICMS-driven tactile feedback (Fig. 3a). These four exploratory sessions included ICMS, but did not have consistent mapping between finger torque feedback and stimulation parameters. Further, these sessions were intermixed with sessions without ICMS rather than being performed consecutively with fixed parameters as in our final experimental design. Over these 23 sessions, performance had plateaued, with a median ARAT score of 18 and an interquartile range (IQR) of 16.25 – 19 (Fig. 3a). We then began collecting data to compare the effect of ICMS on ARAT performance. In the first block of four sequential sessions—which included ICMS, enabling our participant to feel tactile sensations perceived as originating from his own hand when the robotic hand grasped an object—his ARAT score increased significantly to a median of 21 and a range of 20 – 21 ($U = 5$, $P = 0.005$, Wilcoxon rank-sum test, Table 1, Fig. 3a). Performance with ICMS was also significantly better than the four subsequent matched control sessions without ICMS in which he achieved a median ARAT score of 17 with a range of 16 – 19 ($U = 0$, $P = 0.029$, Wilcoxon rank-sum test, Fig. 3a). ARAT scores in these control sessions were no different than the 23 historical sessions ($U = 39$, $P = 0.65$, Wilcoxon rank-sum test, Fig. 3a). Individual session scores are shown in Table 1. Despite the significantly improved scores in sessions with ICMS, there was no change in the total number of trials that were successfully completed ($U = 7$, $P = 0.83$, Wilcoxon rank-sum, Table 1). Therefore, the improved ARAT scores occurred as a result of completing individual trials more quickly. In the ARAT scoring system, successfully transferring an object in less than five seconds, and achieving a score of three, is considered normal, unimpaired performance¹⁸. In the absence of tactile sensations evoked by ICMS, a score of three was achieved only once during the 4 sessions (108 trials). When tactile sensations were provided, a score of three was attained 15 times during the 108 trials.

Overall, we found that trials were consistently completed much more quickly when ICMS feedback was delivered (Fig. 3b, Supplemental Video 2); 14% of the trials with ICMS-evoked tactile feedback were completed more quickly than the fastest trial without ICMS. In fact, discounting the single trial that was completed in less than five seconds without ICMS, 25% of the trials with ICMS were completed more quickly than trials without ICMS (Fig. 3b). Successfully completed trial lengths decreased from a median time of 20.9 s (13.1 - 40.5 s IQR) to 10.2 s (5.4 - 18.1 s IQR) when tactile feedback was provided ($U = 1676$, $P < 0.0001$, Wilcoxon rank-sum test, Table 1, Fig. 3b and Supplemental Video 3). These faster completion times were the cause of the 3.5-point improvement in the ARAT score that occurred when ICMS was provided and can be interpreted as meaning that ICMS-induced tactile sensations allowed 3.5 more objects, out of 9 possible, to be transported to the platform in a normal time (< 5 seconds). The improved times were not due to differences in the commanded velocities. While the distributions of translation velocity commands measured at each time step were statistically different between conditions ($D = 0.02$, $P < 0.0001$, 2-sample Kolmogorov-Smirnov test, Extended Data Fig. 1), the velocities were functionally equivalent. The median translation velocity was 16.7 cm/s (11.5 – 23.2 cm/s IQR) with ICMS and 16.4 cm/s (11.4 – 22.6 cm/s IQR) without ICMS. Similar results were observed for wrist rotation and grasp velocities (Extended Data Fig. 1).

The ARAT task can be broadly divided into reach, grasp, and transport phases (Supplemental Video 4). We separated the trials into these three sequential task phases: (1) reach, consisting of movement onset to first object contact; (2) grasp, consisting of first object contact to successful object liftoff; and (3) transport, consisting of object liftoff to object release. The median time spent reaching decreased from 2.1 s (1.5 – 3.5 s IQR) without ICMS to 1.5 s (1.2 – 2.3 s IQR) when ICMS was provided, representing a 27.8% improvement ($n = 78$ without ICMS and $n = 85$ with ICMS, $U = 2204$, $P = 0.0002$, Wilcoxon rank-sum test, Fig. 3c). Likewise, the median time spent transporting the object decreased from 2.9 s (2.0 – 4.0 s IQR) to 2.1 s (1.8 – 3.0 s IQR), representing a 22.3% improvement ($n = 78$ without ICMS and $n = 85$ with ICMS, $U = 2366.5$, $P = 0.002$, Wilcoxon rank-sum test, Fig. 3c). Most impressively, the amount of time spent attempting to grasp the object decreased from 13.8 s (7.2 – 35.4 s IQR) without ICMS to 5.8 s (1.9 – 13.5 s IQR) with ICMS, resulting in a 44.7% improvement in performance ($n = 78$ without ICMS and $n = 85$ with ICMS, $U = 1819.5$, $P < 0.0001$, Wilcoxon rank-sum test, Fig. 3c). We speculated that, much like in the object transfer task, the participant spent less time attempting to grasp the objects during

145 trials with ICMS-evoked tactile percepts because the percepts increased his certainty about object
146 contact timing and his confidence that he had successfully grasped the object. Why the amount of time
147 spent in the other two phases decreased is less clear. Since object contact and contact force cannot be
148 felt without ICMS, he may have taken longer positioning the hand to improve the amount of information
149 about object interaction he could extract visually, thus increasing the amount of time spent reaching.
150 For the transport phase, the participant may have been less confident about his grasp stability, causing
151 him to move more slowly during transport to avoid dropping the object.

152 By design, the objects in the ARAT task vary in size, shape, weight and, therefore, the overall difficulty in
153 grasping them. As a result of the significant time spent practicing this task, the participant had classified
154 the nine ARAT objects as being either easy (5 cm cube, 7.5 cm cube and sphere) or difficult (2.5 cm cube,
155 10 cm cube, rock, small cylinder, large cylinder and water pouring) to complete. All of the objects that
156 were rated as easy, as well as the 10 cm cube and large cylinder, were completed more quickly with
157 ICMS than without ICMS (Fig. 3d, Extended Data Table 1). Including ICMS did not significantly improve
158 performance with the rock, small cylinder or water pouring task although the median completion time
159 did go down for all of the objects. Therefore, other factors, such as the controllable degrees of freedom
160 or kinematic constraints in the robotic arm, may have limited performance on these objects. However,
161 for those objects that could be completed more easily, adding ICMS feedback further improved
162 performance.

163 Prior to conducting the functional tasks each session, BCI decoder performance was tested in the
164 absence of ICMS-evoked tactile feedback using a random target sequence task²². This task explicitly
165 measured how well the participant could independently control each DoF by moving to specific locations
166 in the 5 DoF workspace. On the days when ICMS-evoked tactile feedback was not provided, sequence
167 task performance was slightly higher, achieving a median score of 100% on all four days compared to a
168 median of 95% (range 90-100%) on the days where ICMS was delivered during the functional tasks (12
169 scores per condition, $U = 40.5$, $P = 0.025$, Wilcoxon rank-sum test, median scores for individual sessions
170 in Table 1). This suggests that decoder performance itself, and thus the participant's ability to control
171 the robotic arm, did not favor the days on which ICMS was provided.

In many bidirectional upper-limb prosthetics studies where amputees receive restored sensory feedback through electrical stimulation of the peripheral nerves, the effect of artificial sensations on performance are measured without visual or auditory feedback^{12,23–25}. Our approach differed from these studies in that our aim was to investigate the effect of providing artificial somatosensory feedback on tasks that were already possible with existing sensory modalities, namely vision. Here, we demonstrated that in highly-practiced tasks where normal visual feedback was available, adding artificial tactile feedback through ICMS enabled a person with spinal cord injury using a BCI to significantly improve their task scores, primarily by spending less time attempting to grasp the objects (Fig. 2a,c, 3b,c).

As with any single-subject study, it is uncertain whether these findings will generalize to future experiments. However, there are several reasons to believe that these results accurately represent the potential of restoring somatosensory percepts using ICMS. First, using the same fundamental neural decoding and control methods, we have demonstrated that two participants achieved similar scores on functional tasks with vision alone^{20,22} and that these scores were only exceeded when ICMS-evoked tactile feedback was provided (Fig. 3a). This suggests that without artificial tactile feedback, control is impaired, much as it is when tactile sensations are absent in people with otherwise normal motor control capabilities^{3,26}. Second, we found that performance improvements were driven primarily by reductions in the time taken to successfully grasp an object. State transitions, such as object contact⁵ during the grasp phase, are uniquely encoded by tactile feedback in the intact nervous system. That the percepts signaled these state transitions with high temporal accuracy, and enabled him to grasp objects more quickly, suggests that ICMS delivered to area 1 of S1 can improve task performance in a way that is congruent to the way natural cutaneous feedback improves grasp performance. Finally, when ICMS-induced percepts were provided, performance improved significantly, and when they were removed, performance returned to pre-ICMS levels (Fig. 3a). Therefore, these observations suggest that the observed improvements were primarily due to the addition of reliable sensory information, rather than the result of additional practice. This immediate performance improvement also suggests that ICMS in S1 was not akin to sensory substitution cues that could have been provided by electrical or mechanical stimulation of intact skin or audio or visual cues, as the relationship between these cues and behavior must be learned²⁷.

Ultimately, ICMS-induced tactile percepts improved task performance to levels not previously observed, decreased the time spent grasping in ways that were analogous to the role of natural tactile sensations during grasp state transitions, and do not appear to be the result of practice, suggesting that including naturalistic somatosensory feedback, like that induced with ICMS, could have a major impact on the future development and performance of dexterous prosthetic limb systems.

Methods

Implantation and electrode arrays

This study was conducted under an Investigational Device Exemption from the U.S. Food and Drug Administration and is registered at ClinicalTrials.gov (NCT01894802). The study was approved by the Institutional Review Boards at the University of Pittsburgh and the Space and Naval Warfare Systems Center Pacific. Informed consent was obtained before any study procedures were conducted.

A 28-year-old male participant with tetraplegia due to a C5 motor/C6 sensory ASIA B spinal cord injury was implanted with two sets of microelectrode arrays (Blackrock Microsystems, Inc., Salt Lake City, Utah, Fig. 1b). Two intracortical microelectrode arrays with 88 wired channels (10x10 array, 1.5 mm length platinum electrodes) were implanted in the hand and arm region of M1 in order to decode movement intent. Two additional microelectrode arrays with 32 wired channels were implanted in area 1 of S1 (6x10 array, 1.5 mm length and coated with a sputtered iridium oxide film) in order to evoke sensations in the fingers of the right hand when stimulated⁷. The study sessions described here took place between 717 and 738 days after the arrays were implanted.

Neural Recording

Voltage recordings from each electrode were band-pass filtered between 0.3 Hz and 7.5 kHz and digitized at 30,000 samples per second using a NeuroPort signal processor (Blackrock Microsystems, Inc., Salt Lake City, Utah). Electrical artifacts induced by microstimulation were rejected using a combination of digital signal blanking and filtering. During each stimulus pulse the recorded signals were blanked using a sample-and-hold circuit. The signals were then high-pass filtered using a 750 Hz first-order Butterworth filter that minimized the effect of additional transient discontinuities in the signal, enabling fast settling of the wideband signal to baseline. A spike threshold was set at -4.5 times the root-mean-square of this high-pass filtered signal. Any transient threshold crossings that occurred in the sample

immediately after the blanking period were rejected in software. Using this approach, we were able to record single unit activity within 740 μ s of the end of a stimulus pulse²⁸.

Motor decoding

To investigate the ability of the participant to use ICMS-evoked tactile percepts during continuous control of a prosthesis, we first created a mapping between population-level neural firing rates recorded in M1 and desired arm movements. A 5 DoF decoder was used in this study, comprising translation of the endpoint in 3D space, wrist pronation and supination, and flexion and extension of all fingers and the thumb, with the thumb always opposite the fingers. All 5 DoFs were controlled simultaneously. A 5 DoF control scheme was chosen as it provided a balance between fast training times and a sufficient degree of dexterity to grasp the different objects used in these experiments.

To train the decoder, the participant observed a virtual version of the Modular Prosthetic Limb (MPL)¹⁹ moving in a 3D environment, as has been described previously²⁰. In this task, the participant was asked to observe and imagine performing the motions of the MPL as the hand was first translated, then oriented, and finally commanded to grasp targets that were randomly presented throughout the workspace using a combination of virtual objects and auditory cues. After observing the completion of 27 trials, which took approximately 7 minutes, an optimal linear estimator decoder was derived using an encoding model that relates neural firing rates to arm kinematics. The encoding model was:

$$f = b_0 + b_x v_x + b_y v_y + b_z v_z + b_\theta v_\theta + b_g v_g \quad (\text{Equation 1})$$

where f is the square-root transformed firing rate of a recorded unit, v is a kinematic velocity, and b is a regression coefficient for a given velocity dimension. The dimensions shown in Equation 1 are x , y , and z translation, wrist rotation (θ), and grasp (g). The b coefficients were calculated using linear regression²⁹. Decoder weights were then calculated using indirect optimal linear estimation (Fig. 1e)³⁰.

The participant used the decoder trained from observation data to repeat the training task, however the computer constrained the decoded movement velocities to those that were on the ideal path³¹. Once this task was completed, a new decoder was trained using the data from the second training set. During task performance, all firing rates were scaled, prior to being decoded, by dividing them by the ratio between the population firing rate during the most recent 300 ms and the population firing rate during

decoder calibration. This method of scaling firing rates prior to decoding was developed to compensate for a correlated increase in firing rate across the recorded population that we observe when the prosthetic hand approaches objects³². This scaling allowed the participant to better stabilize the hand near objects in order to grasp them. Ultimately, this velocity decoder was then used, without computer assistance—that is the decoders and prosthetic arm control systems were naïve to the goal—to complete the tasks used to evaluate performance.

Decoder performance was evaluated using the physical MPL in a sequence task, where the goal was to acquire instructed combinations of hand endpoint position, wrist orientation and grasp posture^{20,22}. A total of 3 sets of 10 trials were performed with the robotic limb without computer assistance to establish the baseline decoder performance accuracy in the absence of objects and ICMS. A trial was considered successful if the participant was able to place the robotic hand within a position target that was 5 cm in diameter, orient the wrist to within ± 0.25 radians and control the grasp aperture to be at least 80% of the way to maximum flexion or extension of the digits being used.

Intracortical microstimulation

Stimulation pulse trains consisted of cathodal phase first, current-controlled, charge-balanced pulses delivered at a rate of 100 pulses per second. The cathodal phase was 200 μ s long, the anodal phase was 400 μ s long, and the amplitude of the anodal phase was set to half the amplitude of the cathodal phase. The phases were separated by a 100- μ s interphase period. Detailed descriptions of sensory percepts evoked via ICMS of S1 have previously been reported⁷. Briefly, ICMS elicited percepts that were described by the participant as originating from the bases of the 2nd through 5th digits and up to the distal interphalangeal joint of the index finger. We selected the electrodes used to provide ICMS-evoked tactile percepts prior to the experiments and focused on electrodes that elicited easily detectable percepts with a clear projected location. One electrode, with a projected field in the proximal interphalangeal joint of the index finger, was mapped to the output of the torque sensor located at the index finger metacarpal phalangeal joint of the MPL. Four electrodes with projected fields in either the middle, ring or little finger were mapped to the torque sensor output from the middle finger of the MPL (Fig. 1c). Together, the projected fields from the selected electrodes spanned the index, middle, ring and little fingers.

For tasks with ICMS, torque sensors located in the motors controlling the MPL fingers provided the signal that was used to modulate ICMS pulse train amplitude according to the follow equation:

$$A_t = \left(\frac{T_t - T_{min}}{T_{max} - T_{min}} \right) * (A_{max} - A_{min}) + A_{min} \quad \text{(Equation 2)}$$

where A_t refers to the commanded pulse train amplitude at time step t , A_{min} and A_{max} refer to the electrode-specific range of stimulus amplitudes, and T represents the torque sensor data that was being used to relay grasp force. We also set values for the minimum and maximum torque readings, T_{min} and T_{max} , respectively, that corresponded to the minimum and maximum stimulation amplitudes. The selected torque thresholds were 0.1 Nm and 0.5 Nm, which corresponded approximately to light touch and strong grasp, respectively. These values were linearly mapped to stimulus amplitudes that ranged from 14 to 64 μ A in increments of 4 or 6 μ A (Fig. 1d). New torque values were sampled every 20 ms and used to update the pulse train amplitude in real time.

Functional task descriptions and scoring metrics

We used two different paradigms to quantify the effects of providing ICMS on the participant's ability to complete functionally relevant tasks. Both the object transfer task and Action Research Arm Test (ARAT) have been successfully performed with vision as the only source of feedback^{20,22}. Here we directly compared performance with and without ICMS-evoked tactile percepts while vision was always present.

For the object transfer task, we asked the participant to reach to and grasp a cylindrical object (16 cm tall and 4.3 cm in diameter) with a weighted base placed on the left side of the table, lift the object off of the table, carry it to the target area on the right, and release the object (Fig. 1f). Two boundaries were marked on the table that defined a 22.5 cm region where the object was not allowed to touch the table (red area in Fig. 1f). If the object touched the table between these boundaries, the task could be continued by moving the object back to the left side of the table and continuing. Once the object was placed on the right side of the table, an experimenter returned the object to the start position and the participant repeated the process as many times as possible in two minutes (Supplemental Video 1). Performance on this task was measured as the number of times the object was successfully moved across the table in two minutes. This task was always completed prior to the ARAT task.

308 We also conducted a modified version of the Action Research Arm Test (ARAT)^{18,33}, which consisted of
 309 moving eight different objects from the left side of a table to a raised platform located on the right side
 310 (Fig. 1g). These objects were selected from the suite of objects that are part of the standard ARAT task³³
 311 and included four cubes (2.5 cm, 5 cm, 7.5 cm and 10 cm along each edge), a 7.5 cm diameter ball, a
 312 rock, and two cylinders (2.5 cm and 1 cm in diameter and 16 cm tall). Additional objects from the ARAT
 313 task were too small to be grasped by the MPL. The target platform was 34 x 20.5 cm and was elevated
 314 12 cm off the table surface. The objects started approximately 70 cm away from the target platform. A
 315 ninth object from the original ARAT task was also included in which a cup filled with small pieces of paper
 316 and plastic, as a proxy for water, was placed at the right side of the workspace, and an empty cup was
 317 placed 20 cm to the left of it. The participant's task was to empty the "water" from the cup on the right
 318 into the empty cup on the left and replace the originally grasped cup back on the table in an upright
 319 position. This task was considered a success if any "water" landed in the target cup and if the original
 320 cup was placed upright on the table.

321 In all cases, the participant was instructed to complete the task as quickly as possible. The participant
 322 had a maximum of two minutes per attempt, and three attempts per object. Each attempt at transferring
 323 the objects was considered a trial. Trials were timed by experimenters from movement onset to the
 324 object being successfully placed on the target platform. Each trial was scored on a 3-point system in
 325 which a score of zero was awarded if the object was never touched, a score of one was awarded if the
 326 object was touched but the participant was unable to complete the task, a score of two was awarded if
 327 the task was completed in less than two minutes but more than five seconds, and a score of three was
 328 awarded if the task was completed in under five seconds. The best score from the three attempts for
 329 each object was added together to create a single score for the test. Therefore, for the task with nine
 330 objects, a perfect score was 27.

331 The score, which is the validated metric of the ARAT task, fails to take into account other aspects of
 332 performance, such as the total number of completed attempts per object and the actual completion
 333 time. Therefore, we recorded video of all trials and measured the time spent reaching for, grasping, and
 334 transporting the object. All task phase calculations were done offline, marking individual video frames
 335 that spanned each event. Reaching was defined as the time from movement onset until the first object

contact. Grasping was defined as the period between object contact and successful object liftoff from the table. The transport phase spanned object liftoff until object release.

We tested the two feedback conditions in a block-design over the course of these experiments. For the first four sessions, ICMS feedback was delivered to five electrodes. Each experiment day, three blocks of the sequence task, five blocks of the object transfer task, and one ARAT session were completed. For the next four consecutive sessions, the same testing protocol was followed, but ICMS was not delivered.

Statistical analysis

Statistical analyses were performed in MATLAB (The MathWorks). Data that were not normally distributed, as determined using Lilliefors test ($\alpha = 0.05$), are reported as medians and interquartile ranges (IQR) and the Wilcoxon rank-sum test was used to assess significance for differences in the median unless otherwise stated. The Mann-Whitney U test statistic is reported for all Wilcoxon rank-sum tests. Normally-distributed data, as determined using Lilliefors test ($\alpha = 0.05$), are reported as mean $\pm$ standard deviation and a two-tailed Student's t-test was used to assess significance for differences in the mean. Specific statistical tests are noted in the text. All object transfer data have $n = 20$ trials per feedback condition.

Data availability

Data supporting these findings as well as software routines to analyze these data are available from the corresponding author upon reasonable request.

354 **End Notes**

355 **Supplementary Information** is available in the online version of this paper.

356 **Acknowledgements** We thank N. Copeland for his continuing and extraordinary commitment to this
 357 study as well as insightful discussions with the study team; Debbie Harrington (Physical Medicine and
 358 Rehabilitation) for regulatory management of the study; Ahmed Jorge for help with data collection; Peter
 359 Gibson and Ben Clarkson for video data processing; the University of Pittsburgh Clinical and Translational
 360 Science Institute and the Office of Investigator-Sponsored Investigational New Drugs and Investigational
 361 Device Exemption support for assistance with protocol development and regulatory reporting and
 362 compliance; the volunteer members of the Data Safety and Monitoring Board for their continued
 363 monitoring of this study; H. Jourdan (Department of Physical Medicine and Rehabilitation) for financial
 364 and organizational support; and Blackrock Microsystems (Salt Lake City, UT, USA), especially Robert
 365 Franklin, for technical support related to this project. This material is based upon work supported by the
 366 Defense Advanced Research Projects Agency (DARPA) and Space and Naval Warfare Systems Center
 367 Pacific (SSC Pacific) under Contract No. N66001-16-C-4051 and the Revolutionizing Prosthetics program
 368 (Contract No. N66001-10-C-4056). S.N.F. was supported by an NSF Graduate Research Fellowship under
 369 grant number DGE-1247842. The views, opinions, and/or findings contained in this article are those of
 370 the authors and should not be interpreted as representing the official views or policies of the
 371 Department of Veterans Affairs, Department of Defense, or US Government.

372 **Author Contributions** S.N.F., J.E.D., J.L.C., and R.A.G. designed the study. S.N.F., J.E.D., J.M.W., C.L.H.,
 373 A.J.H., J.L.C., and R.A.G. conducted the experiments. S.N.F. analyzed the data. All authors contributed to
 374 the interpretation of the results. S.N.F. wrote the paper with R.A.G. and J.L.C., and all authors provided
 375 critical review, edits, and approval for the final manuscript.

376 **Author Information** The authors declare that they have no competing interests. Correspondence and
 377 requests for materials should be addressed to R.A.G. (rag53@pitt.edu) and J.L.C. (collinger@pitt.edu).

Table 1: Performance metrics for each task per experiment day. The total number of object transfers is the sum of all five 2-minute trials per day. ARAT scores were computed as the sum of the best score per object, with a maximum score of 27. Each of the nine objects was attempted 3 times, so that the maximum number of trials attempted per session was 27. The total median and IQR trial time for successful ARAT trials was calculated by pooling trial times across all four sessions per feedback condition and calculating the median and IQR from the aggregate distribution.

	Session	Object Transfer (transfers per day)	ARAT Score (out of 27)	ARAT Trials Completed (out of 27)	Median and IQR trial time for Successful ARAT Trials (s)	Median Sequence Task % Correct
With ICMS Feedback	1	97	21	19	11.9 (6.6 – 27.7)	90
	2	74	21	22	12.0 (5.6 – 38.9)	90
	3	93	21	21	8.8 (6.0 – 17.2)	100
	4	88	20	19	8.1 (4.6 – 11.9)	100
	Total	352	83	81	10.2 (5.4 – 18.1)	
Without ICMS Feedback	1	88	19	23	14.0 (11.1 – 30.9)	100
	2	55	16	19	27.6 (18.8 – 37.2)	100
	3	74	17	23	18.7 (12.3 – 41.7)	100
	4	98	17	13	40.5 (15.5 – 48.4)	100
	Total	315	69	78	20.9 (13.1 – 40.5)

385

386 **Fig. 1:** Overview of bidirectional BCI system components, operation and tasks. **a**, The study participant
387 used a bidirectional intracortical BCI to control a robotic prosthesis in real time. The arm was positioned
388 near the participant to provide clear visual feedback, but physical contact was not possible. The
389 participant controlled the prosthesis in five dimensions, illustrated by the dark blue arrows (3D
390 translation, wrist rotation and grasp). **b**, Four microelectrode arrays were implanted in the left
391 hemisphere. Arrays in primary motor cortex (blue) recorded signals which were used to control the
392 modular prosthetic limb. Arrays in somatosensory cortex (red) delivered stimulation pulses, which
393 artificially activated neurons, resulting in sensory percepts referred to the hand. **c**, Stimulation of the

electrode arrays in the somatosensory cortex evoked percepts from the base of the fingers. Colored grids represent individual electrodes on the two microelectrode array and the locations on the hand where stimulation through each electrode evoked a percept (index finger = purple, middle finger = blue, ring finger = green, little finger = orange)⁷. Torque sensors in the robot fingers were used to drive selected electrodes in the somatosensory cortex with matching somatotopic fields (e.g. index finger torque sensor controlled electrodes evoking percepts in the index finger). **d**, The torque measured at the base of the fingers increased as more force was applied to the objects. Stimulation current amplitude was modulated by torque using a linear transformation. **e**, Threshold crossing events were detected from the multichannel neural recordings in the motor cortex. Each row represents an individual electrode and each mark represents a threshold crossing event. Using an optimal linear estimation decoding scheme, endpoint velocity (v_x , v_y , v_z) as well as wrist pro/supination velocity (v_θ) and grasp velocity (v_g) were simultaneously and continuously decoded. **f**, Overhead view of the object transfer task showing the grasp (gray area), transport (red area) and release (green area) zones. The cylindrical object was placed in the grasp zone by the experimenter, was grasped using the prosthesis, moved over the transport zone and placed in the release zone. This process was repeated as many times as possible in two minutes. **g**, Overhead view of the Action Research Arm Test (ARAT) showing the object presentation position (green dot) and the raised platform target (green box). Different objects (not all objects shown) were positioned at a standard location, grasped and then placed on the platform as quickly as possible. For all tasks, the arm was under full control of the user from the start to the end of a trial.

Fig. 2: Object transfer performance. **a**, Amount of time spent in each task zone, per transfer, by feedback condition ($n = 20$ trials per feedback condition). Data for all trials are shown with red lines indicating the mean value and the whiskers indicating one standard deviation. The amount of time spent in the grasp and release zones decreased significantly with ICMS feedback ($*P = 0.002$ and 0.048 , t -test, respectively), but the amount of time in the transport zone per transfer was not affected. **b**, Distribution of average path lengths in the grasp zone per trial for the two feedback conditions, computed as the total path length divided by the number of transfers. The longer path lengths ($*P = 0.0007$, t -test) without ICMS suggest that the extra time spent in the grasp zone was to adjust the endpoint position, rather than to hold the robot still while attempting to issue a grasp command. **c**, Spatial map of the amount of time spent in each location in the workspace per transfer. Each individual square represents a 1×1 cm region. Without stimulation, there was substantially more time spent near the object in the grasp zone as shown by the increase in the number of locations colored yellow in the grasp zone. Red lines indicate zone boundaries. Color indicates the amount of time spent in each location per transfer.

Fig. 3: Effect of ICMS on ARAT task performance. **a**, Comparison of ARAT scores before experiment onset, which spanned a range of controlled degrees of freedom and occasionally employed ICMS feedback (blue dots), to data from the current experiment with ICMS feedback (blue) and without (gray). ARAT scores with ICMS feedback were significantly higher than historic performance (* $P = 0.005$, Wilcoxon rank-sum test) as well as control tests (* $P = 0.029$, Wilcoxon rank-sum test) conducted without ICMS. Red lines indicate the median score per paradigm. **b**, Cumulative distribution of individual trial times, including failed trials, shown on a log-normalized axis. Trials for all four sessions for each feedback condition were combined to compute the empirical cumulative distribution. The horizontal red line and connected vertical red lines indicate the times at which 50% of all attempted trials were completed for each condition. Vertical dashed lines connected to red X's indicate when 50% of successful trials were completed. Shading indicates the 95% confidence bounds, calculated with Greenwood's formula. **c**, Amount of time spent in each phase of the ARAT task. Red lines are medians, box outlines are

interquartile ranges, and whiskers are the range of the data excluding outliers which are shown as red '+' symbols. All task phases were faster when ICMS feedback was provided (*P < 0.01, Wilcoxon rank-sum test). For this analysis we included trials containing a successful reach, grasp and transport phase. Water pouring trials were not included as the transport phase is not defined. n = 78 trials for all phases without ICMS feedback and n = 85 trials for all phases with ICMS feedback. **d**, Effect of ICMS feedback on completion times for individual objects. Gray dots indicate trial times without ICMS feedback while blue dots are individual trial times with ICMS. Median trial times are marked for each object/feedback paradigm with an X. Medians for each object are connected with a red line for visualization. Trial times were significantly lower for five of the nine objects when ICMS feedback was provided (*P < 0.05, Wilcoxon rank-sum test).

450 Extended Data

451

452 **Extended Data Fig. 1:** Distribution of commanded robot velocities for each timestep during all ARAT
453 trials with (blue) and without ICMS (gray) **a**, Commanded 3D translation velocity. The distribution of
454 commanded translation velocities were different for trials with and without ICMS ($D = 0.02$, $P < 0.001$,
455 2-sample Kolmogorov-Smirnov test). **b**, Commanded wrist rotation velocity. The distributions of
456 commanded wrist rotation velocities were different for trials with and without ICMS ($D = 0.055$, $P <$
457 0.0001 , Kolmogorov-Smirnov test). The median wrist rotation velocity was -0.22 rad/s ($-0.74 - 0.26$ rad/s
458 IQR) with ICMS and -0.13 rad/s ($-0.61 - 0.33$ rad/s IQR) without ICMS. **c**, Commanded grasp velocity. The

459 distributions of commanded grasp velocities were different for trials with and without ICMS ($D = 0.058$,
460 $P < 0.0001$, Kolmogorov-Smirnov test). The median grasp velocity was 0.074 a.u. (-0.571 – 0.680 a.u. IQR)
461 with ICMS and -0.001 a.u. (-0.763 – 0.711 a.u. IQR) without ICMS.

462 **Extended Data Table 1:** Successful ARAT trial times by object. All successful water pouring attempts are
 463 listed as there were not enough successfully completed trials to calculate the median and IQR.

Object	Median (s)	IQR (s)	n	Mann-Whitney U Statistic	Significance (p-value, Wilcoxon rank-sum test)
10 cm cube – no ICMS	46.6	24.2 – 80.3	8	13	0.027
10 cm cube – ICMS	13.1	9.8 – 18.6	9		
2.5 cm cube – no ICMS	44.5	32.8 – 62.4	10	30	0.408
2.5 cm cube – ICMS	32.0	15.1 – 60.1	8		
5 cm cube – no ICMS	13.2	10.3 – 29.0	8	17	0.043
5 cm cube – ICMS	6.8	4.3 – 11.5	10		
7.5 cm cube – no ICMS	27.6	13.7 – 38.9	9	15	0.010
7.5 cm cube – ICMS	10.2	6.0 – 13.3	11		
Sphere – no ICMS	12.3	10.9 – 17.8	11	22.5	0.024
Sphere – ICMS	5.9	4.4 – 12.3	10		
Rock – no ICMS	23.9	18.7 – 40.1	9	29	0.541
Rock – ICMS	21.2	6.3 – 52.2	8		
Large Cylinder – no ICMS	14.4	11.2 – 18.3	12	27.5	0.019
Large Cylinder – ICMS	6.6	4.5 – 9.2	11		
Small Cylinder – no ICMS	27.0	15.4 – 32.3	10	29.5	0.078
Small Cylinder – ICMS	9.5	5.7 – 23.2	11		
Water – no ICMS (all times)	76			n/a	
Water – ICMS (all times)	24, 43.9, 48.1				

464

465 **Supplementary Information**

466 **Supplemental Video 1:** Object transfer example trials with and without ICMS feedback. In the full trial,
467 the task lasts for two minutes. The first minute from a trial with the median number of transfers for each
468 feedback condition is used to illustrate performance.

469 **Supplemental Video 2:** Fastest ARAT trials for each object and feedback condition.

470 **Supplemental Video 3:** ARAT trials for the median completion time for each object and feedback
471 condition. In cases where there were an even number of completed trials, the faster trial is shown in the
472 video.

473 **Supplemental Video 4:** Example ARAT trial with ICMS feedback, annotated to indicate task state
474 transitions and illustrate ICMS delivery.

475 References

- 476 1. Jenmalm, P., Dahlstedt, S. & Johansson, R. S. Visual and tactile information about object-
477 curvature control fingertip forces and grasp kinematics in human dexterous manipulation. *J.*
478 *Neurophysiol.* **84**, 2984–2997 (2000).
- 479 2. Monzée, J., Lamarre, Y. & Smith, A. M. The effects of digital anesthesia on force control using a
480 precision grip. *J. Neurophysiol.* **89**, 672–683 (2003).
- 481 3. Rothwell, J. C. *et al.* Manual motor performance in a deafferented man. *Brain* **105**, 515–542
482 (1982).
- 483 4. Flanagan, J. R., Bowman, M. C. & Johansson, R. S. Control strategies in object manipulation tasks.
484 *Curr. Opin. Neurobiol.* **16**, 650–659 (2006).
- 485 5. Johansson, R. S. & Flanagan, J. R. Coding and use of tactile signals from the fingertips in object
486 manipulation tasks. *Nat Rev Neurosci* 345–359 (2009).
- 487 6. Augurelle, A.-S., Smith, A. M., Lejeune, T. & Thonnard, J.-L. Importance of cutaneous feedback in
488 maintaining a secure grip during manipulation of hand-held objects. *J. Neurophysiol.* **89**, 665–
489 671 (2003).
- 490 7. Flesher, S. N. *et al.* Intracortical microstimulation of human somatosensory cortex. *Sci. Transl.*
491 *Med.* **8**, 361ra141 (2016).
- 492 8. Graczyk, E. L., Resnik, L., Schiefer, M. A., Schmitt, M. S. & Tyler, D. J. Home Use of a Neural-
493 connected Sensory Prosthesis Provides the Functional and Psychosocial Experience of Having a
494 Hand Again. *Sci. Rep.* **8**, 9866 (2018).
- 495 9. Salas, M. A. *et al.* Proprioceptive and cutaneous sensations in humans elicited by intracortical
496 microstimulation. *Elife* **7**, e32904 (2018).
- 497 10. Collins, K. L. *et al.* Ownership of an artificial limb induced by electrical brain stimulation. *Proc.*
498 *Natl. Acad. Sci.* **114**, 166–171 (2016).
- 499 11. Hiremath, S. V. *et al.* Human perception of electrical stimulation on the surface of
500 somatosensory cortex. *PLoS One* **12**, e0176020 (2017).
- 501 12. Schiefer, M., Tan, D., Sidek, S. M. & Tyler, D. J. Sensory feedback by peripheral nerve stimulation
502 improves task performance in individuals with upper limb loss using a myoelectric prosthesis. *J.*
503 *Neural Eng.* **13**, 016001 (2015).
- 504 13. Schwartz, A. B., Cui, X. T., Weber, D. J. J. & Moran, D. W. Brain-controlled interfaces: movement
505 restoration with neural prosthetics. *Neuron* **52**, 205–220 (2006).
- 506 14. Fagg, A. H. *et al.* Biomimetic brain machine interfaces for the control of movement. *J. Neurosci.*
507 **27**, 11842–11846 (2007).
- 508 15. Davidson, J. A survey of the satisfaction of upper limb amputees with their prostheses, their
509 lifestyles, and their abilities. *J. Hand Ther.* **15**, 62–70 (2002).
- 510 16. Wijk, U. & Carlsson, I. Forearm amputees' views of prosthesis use and sensory feedback. *J. Hand*
511 *Ther.* **28**, 269–278 (2015).
- 512 17. Biddiss, E., Beaton, D. & Chau, T. Consumer design priorities for upper limb prosthetics. *Disabil.*
513 *Rehabil. Assist. Technol.* **2**, 346–357 (2007).
- 514 18. Lyle, R. C. A performance test for assessment of upper limb function in physical rehabilitation
515 treatment and research. *Int. J. Rehabil. Res.* **4**, 483–492 (1981).
- 516 19. Johannes, M. S., Bigelow, J. D. & Burck, J. M. An overview of the developmental process for the
517 modular prosthetic limb. *Johns Hopkins APL Tech. Dig.* **30**, 207–216 (2011).

20. Wodlinger, B. *et al.* Ten-dimensional anthropomorphic arm control in a human brain–machine interface: difficulties, solutions, and limitations. *J. Neural Eng.* **12**, 016011 (2015).
21. Bouton, C. E. *et al.* Restoring cortical control of functional movement in a human with quadriplegia. *Nature* **533**, 247–250 (2016).
22. Collinger, J. L. *et al.* High-performance neuroprosthetic control by an individual with tetraplegia. *Lancet* **381**, 557–564 (2012).
23. Raspopovic, S. *et al.* Restoring natural sensory feedback in real-time bidirectional hand prostheses. *Sci. Transl. Med.* **6**, 222ra19 (2014).
24. Graczyk, E. L. *et al.* The neural basis of perceived intensity in natural and artificial touch. *Sci. Transl. Med.* **8**, 362ra142 (2016).
25. Wendelken, S. *et al.* Restoration of motor control and proprioceptive and cutaneous sensation in humans with prior upper-limb amputation via multiple Utah Slanted Electrode Arrays (USEAs) implanted in residual peripheral arm nerves. *J. Neuroeng. Rehabil.* **14**, 121 (2017).
26. Johansson, R. S., Hger, C. & Bäckström, L. Somatosensory control of precision grip during unpredictable pulling loads. III. Impairments during digital anesthesia. *Exp. Brain Res.* **89**, 204–213 (1992).
27. Maidenbaum, S., Abboud, S. & Amedi, A. Sensory substitution: closing the gap between basic research and widespread practical visual rehabilitation. *Neurosci. Biobehav. Rev.* **41**, 3–15 (2014).
28. Weiss, J. M., Flesher, S. N., Franklin, R., Collinger, J. L. & Gaunt, R. A. Artifact-free recordings in human bidirectional brain–computer interfaces. *J. Neural Eng.* **16**, 016002 (2019).
29. Marquardt, D. W. Generalized Inverses, Ridge Regression, Biased Linear Estimation, and Nonlinear Estimation. *Technometrics* **12**, 591 (1970).
30. Wang, W., Chan, S. S., Heldman, D. A. & Moran, D. W. Motor cortical representation of position and velocity during reaching. *J. Neurophysiol.* **97**, 4258–4270 (2007).
31. Velliste, M., Perel, S., Spalding, M. C., Whitford, A. S. & Schwartz, A. B. Cortical control of a prosthetic arm for self-feeding. *Nature* **453**, 1098–1101 (2008).
32. Downey, J. E. *et al.* Motor cortical activity changes during neuroprosthetic-controlled object interaction. *Sci. Rep.* **7**, 16947 (2017).
33. Yozbatiran, N., Der-Yeghiaian, L. & Cramer, S. C. A standardized approach to performing the action research arm test. *Neurorehabil. Neural Repair* **22**, 78–90 (2008).