Sensor localization using magnetic dipole-like coils: A method for highly accurate co-registration in on-scalp MEG

Christoph Pfeiffer^{a,*}, Silvia Ruffieux^a, Lau M. Andersen^b, Alexei Kalabukhov^a,
Dag Winkler^a, Robert Oostenveld^d, Daniel Lundqvist^b, Justin
F. Schneiderman^c

^a*Department of Microtechnology and Nanoscience - MC2, Chalmers University of Technology, Gothenburg, Sweden*

^b*NatMEG, Department of Clinical Neuroscience, The Karolinska Institute, Stockholm, Sweden*

^c*MedTech West and the Institute of Neuroscience and Physiology, Sahlgrenska Academy, University of Gothenburg, Gothenburg, Sweden*

^d*Donders Institute for Brain, Cognition and Behaviour, Radboud University, Nijmegen, Netherlands*

Abstract

Source modelling in magnetoencephalography (MEG) requires precise co-registration of the sensor array and the anatomical structure of the measured individual's head. In conventional MEG, positions and orientations of the sensors relative to each other are fixed and known beforehand, requiring only localization of the head relative to the sensor array. Since the sensors in on-scalp MEG are positioned on the scalp, locations of the individual sensors depend on the subject's head shape and size. The positions and orientations of on-scalp sensors must therefore be measured at every recording. This can be achieved by inverting conventional head localization, localizing the sensors relative to the head - rather than the other way around.

In this study we present a practical method for localizing sensors using magnetic dipole-like coils attached to the subject's head. We implement and evaluate the method in a set of on-scalp MEG recordings using a 7-channel on-scalp MEG system based on high critical temperature superconducting quantum interfer-

*Corresponding author

Email address: christoph.pfeiffer@chalmers.se (Christoph Pfeiffer)

ence devices (high- T_c SQUIDs). The method provides accurate estimates of individual sensor positions and orientations with short averaging time (≤ 2 mm and < 3 degrees, respectively, with 1-second averaging), enabling continuous sensor localization. Calibrating and jointly localizing the sensor array can further improve the localization accuracy (< 1 mm and < 2.5 degrees, respectively, with 1-second coil recordings).

We demonstrate source localization of on-scalp recorded somatosensory evoked activity based on co-registration with our method. Equivalent current dipole fits of the evoked responses corresponded well (within 5.3 mm) with those based on a commercial, whole-head MEG system.

Keywords: Magnetoencephalography (MEG), On-scalp MEG, Co-registration, Sensor localization, Magnetic dipole, coil, High- T_c SQUID.

1. Introduction

On-scalp magnetoencephalography (MEG) has been shown in simulations to provide distinct advantages over traditional, low- T_c SQUID-based MEG. At closer proximity to the head –and thus to the neural sources– on-scalp MEG should be able to measure weaker signals as well as capture higher spatial frequencies compared to conventional MEG [1, 2]. In addition to smaller standoff, on-scalp MEG sensors - primarily optically pumped magnetometers (OPMs) and high- T_c SQUIDs - allow flexible sensing of the head; that is, the sensors can be moved (individually or in small units containing a few sensors) relative to each other in order for the sensor array to fit the head of individual subjects [3]. This is especially beneficial for studies on children, whose heads are significantly smaller than the one-size-fits-all helmets in most commercial MEG systems [4].

In general, translating MEG (sensor-level) signals to neural (source-level) activity requires co-registration of functional and structural data. An important step in this process is the reliable determination of the measurement/sensor locations relative to the subject’s head during the recording. In conventional MEG systems this is achieved by placing a set of small magnetic coils on the

subject's head and digitizing their positions with respect to landmarks (e.g., fiducials) on the head. Energizing the coils at different times and/or frequencies and detecting the distribution of the magnetic fields they generate (with the MEG system) allows accurate localization of the coils relative to the MEG sensor array [5, 6]. In order to localize the coils in such a way, the positions and orientations of the sensors relative to each other have to be known. This presents an issue when using flexible sensor arrays in on-scalp MEG. Because the sensors in such a system would be at least partially independently positioned, the sensors' relative positions and orientations vary from subject to subject, and from session to session. Instead of a one-time calibration as used with rigid, whole-head sensor arrays, it is necessary to determine the sensor locations for each MEG recording session.

Measuring all the sensor positions and locations in a full-head array manually would be very time consuming and cumbersome, especially in arrays with high channel count. We have therefore developed and simulated the efficacy of a method for localizing independent MEG sensors with an array of small, magnetic dipole-like coils attached to the subject's head [7]. Herein, we present the implementation of this sensor localization method in MEG recordings with a 7-channel high- T_c SQUID-based on-scalp MEG system. We furthermore validate its utility by using in source localization of somatosensory evoked fields.

2. Methods

2.1. Sensor localization

For an array of on-scalp MEG sensors recording a set of magnetic dipole-like coils (e.g., head position indicator, HPI, coils), the signal generated at the k th magnetometer by the j th magnetic dipole whose moment is $\vec{m}_j$ can be defined as

$$\begin{aligned} S_{k,j} &= \frac{\mu_0}{4\pi} \left(\frac{3\vec{r}_{j,k}(\vec{m}_j \cdot \vec{r}_{j,k})}{|\vec{r}_{j,k}|^5} - \frac{\vec{m}_j}{|\vec{r}_{j,k}|^3} \right) \cdot \vec{n}_k \\ &= L_m(\vec{r}_{j,k}) \vec{m}_j \cdot \vec{n}_k \end{aligned} \quad (1)$$

where L_m is the lead field, $\vec{r}_{j,k} = \vec{r}_j - \vec{r}_k$ a vector defining the location of the dipole j relative to sensor k , $\vec{n}_k = |n_k|\hat{n}_k$ a vector combining the orientation ($\hat{n}_k$) and sensitivity ($|n_k|$) of sensor k , and $\vec{m}_j$ the magnetic moment of dipole j .

The position and orientation of a magnetic dipole is fit to recorded data $S_{k,j}^{rec}$ by finding the dipole location that minimizes the residual variance between the data and the calculated signals.

$$\arg \min_{\vec{r}_j, \vec{m}_j} \left(\frac{\sum_k (S_{k,j}^{rec} - L_m(\vec{r}_{j,k}) \vec{m}_j \cdot \vec{n}_k)^2}{\sum_k S_{j,k}^{rec2}} \right). \quad (2)$$

As described in [7], the standard coil localization procedure can be adapted to determine the position and orientation of an individual MEG sensor with respect to an array of coils by simply swapping the roles of magnetometers and dipoles:

$$\arg \min_{\vec{r}_k, \vec{n}_k} \left(\frac{\sum_j (S_{k,j}^{rec} - L_m(\vec{r}_{k,j}) \vec{n}_k \cdot \vec{m}_j)^2}{\sum_j S_{j,k}^{rec2}} \right). \quad (3)$$

The on-scalp MEG system used here employs seven sensors that are fixed relative to each other in a single cryostat [8]. When multiple sensors are fixed relative to each other it is, in principle, possible to improve their localization by taking into account the array's geometry [7]. Instead of solving eq. 3 for each sensor individually, the array can be combined into a single localization routine, wherein a single rigid transformation (rotation and translation) is applied to the whole sensor array. The number of parameters to be estimated is thus reduced by a factor of 7 compared to localizing the sensors individually. In this case, eq. 3 is replaced by:

$$\arg \min_{T, R} \left(\sum_k \frac{\sum_j (S_{k,j}^{rec} - L_m(\vec{r}'_{k,j}) \vec{n}'_k \cdot \vec{m}_j)^2}{\sum_j S_{j,k}^{rec2}} \right) \quad (4)$$

where T and R describe the 3-dimensional translation and rotation applied to the entire array, $\vec{r}'_{k,j} = (R\vec{r}_k + T) - \vec{r}_j$ is the location of the rigidly transformed position of sensor k relative to dipole j , and $\vec{n}'_k = R\vec{n}_k$ the rigidly transformed sensitivity vector.

48 To reduce the impact that noisy sensors can have on the localization accu-
49 racy, the sensors can be weighted according to their signal-to-noise ratio when
50 summing the residual variances in eq. 4.

$$\arg \min_{T,R} \left(\sum_k w_k \frac{\sum_j (S_{k,j}^{rec} - L_m(\vec{r}'_{k,j}) \vec{n}'_k \cdot \vec{m}_j)^2}{\sum_j S_{j,k}^{rec2}} \right) \quad (5)$$

51 where $w_k = \frac{SNR_k}{\sum_k SNR_k}$ is the weight applied to the k-th sensor.

52 2.2. Measurement setup

53 The sensor localizations described here were performed as part of a set of
54 MEG recordings at the National MEG Facility (NatMEG) at the Karolinska
55 Institutet in Stockholm, Sweden. The main aim of the recordings was to com-
56 pare and contrast recordings with a 7-channel high- T_c SQUID-based on-scalp
57 system [8] to recordings with a commercial, whole-head system - in this case,
58 a 306-channel Elekta TRIUX system (Elekta Neuromag Oy). Several different
59 experimental paradigms were recorded in five neurotypical subjects (4 male and
60 1 female, ages 30-49). For each session the same paradigm was first recorded
61 on a subject with the commercial MEG system, followed by the on-scalp MEG
62 recording. All experiments were approved by the Swedish Ethical Review Au-
63 thority (EPN 2018-571-31-1) and conducted in compliance with national legis-
64 lation and the code of ethical principles defined in the Declaration of Helsinki.
65 All participants gave informed consent.

66 Ten dipole-like head position indicator (HPI) coils of the TRIUX system were
67 used both in the head localization as part of the conventional MEG recordings
68 and in the sensor localization as part of the on-scalp recordings. The coils were
69 driven at frequencies from 537 to 987 in steps of 50 Hz. The frequencies were
70 chosen relatively high in order to spectrally separate them from neural activity
71 (including high frequency components up to 500 Hz). The frequency steps are
72 chosen such that potential intermittent-frequency artefacts would coincide with
73 the power line harmonics (50 Hz in Sweden), which are filtered as part of the
74 standard preprocessing and therefore do not require any additional treatment.

75 The recordings were divided into blocks of stimulations with the coils energized
76 for 10 to 30 seconds before and after each block. This was done as a cautionary
77 measure to prevent potential artifacts from the coils to corrupt the MEG record-
78 ings. Recording before and after each stimulation block also allowed monitoring
79 if/how the head moved.

80 The subjects were recorded seated with their heads comfortably stabilized
81 using vacuum pillows (without being completely immobilized). To further min-
82 imize head movements during the coil recordings, the subjects were instructed
83 to keep their head still. For each paradigm (in some cases two paradigms with
84 similar neural activation) a coarse region of interest was determined prior to the
85 recording session based on knowledge about the expected activity and/or previ-
86 ous recordings on the same subject using the same or a similar paradigm. The
87 coils were then distributed closely around the region of interest, while maintain-
88 ing sufficient room for placement of the cryostat. In order to minimize relative
89 movements between coils, nine coils were fixed to small plastic plates (three coils
90 per plate) that were roughly shaped to fit to the subject's head. The tenth coil
91 was then fixed to the head individually. Figure 1 shows a set of coils arranged
92 around a region of interest on an EEG cap on one of the subject's head. The
93 red tags mark the different target locations for the on-scalp system. The coils,
94 head shape and target location tags were digitized using a AC electromagnetic
95 tracking system Polhemus Fastrak (Polhemus, Colchester, VT 05446, USA).

96 At the beginning of each recording the subject was recorded in the TRIUX
97 system. These recordings were used to localize the underlying neural activity
98 and project the resulting neuromagnetic fields onto the scalp surface. Such field
99 maps were used to guide the placement of the cryostat (i.e., the red markers in
100 Fig. 1) for each experimental paradigm and subject [9, 10]. More importantly
101 for localizing the sensors, the whole-head recordings were used to determine
102 the positions, orientations, and magnetic moments of the coils relative to each
103 other and to the head via traditional head localization [5]. HPI coil locations
104 and orientations obtained thus were used for the ensuing on-scalp recordings.
105 Only coil locations where the goodness of fit exceeded 0.98 were used in the

Figure 1: Photograph showing HPI coils attached to a subjects head. Three triplets of coils (each attached to a rectangular plastic holder) can be seen surrounding a region of interest marked by red tags that indicate measurement locations. Inset: a plastic holder with three HPI coils attached.

106 sensor localization.

107 The sensor fits were performed in MATLAB R2015a (Mathworks, Natick,
 108 MA, USA) using the FieldTrip toolbox [11]. The coil amplitudes were extracted
 109 from the data via multitaper frequency transform using Slepian tapers and used
 110 in a linear grid search to provide a starting point for the non-linear fit. Finally,
 111 the sensor locations were fitted to the extracted coil amplitudes by solving eq.
 112 3 using unconstrained optimization (quasi-newton algorithm) with the starting
 113 point obtained from the grid search.

When fitting the sensors jointly, the known layout of the sensor array is rigidly aligned to the individually fitted sensor locations using an iterative closest points (ICP) algorithm that was modified to minimize distances between corresponding point pairs (that is, points corresponding to the same sensor) rather than closest points. The resulting transformed sensor array then serves as starting point for a non-linear fit.

2.3. Evaluation

Defining the performance of the sensor localization is not straightforward in a realistic measurement setup, like the one we present here, wherein the "ground truth" (i.e., the true sensor locations relative to the head) is not known with arbitrary precision. Generally, the accuracy of the fitted locations are affected by a combination of random errors (e.g., due to sensor noise), systematic errors (resulting from, e.g., errors in the coil positions) and variations in the true location (resulting from head movements).

2.3.1. Random errors

Assuming head movements are negligible during a single (30 second) recording, we estimate the effects of random errors. We split each 30-second coil recording into multiple shorter segments, each of which was independently used to localize the sensors. Variations in an individual sensor's location over segments were then used to provide an estimate of the sensor localization accuracy. To this end, we define $MD(\vec{r}_{k,i}) = \|\vec{r}_k - \vec{r}_{k,i}\|$ as the euclidean distance of the i -th segment's fitted position $\vec{r}_{k,i}$ from the mean location $\vec{r}_k$ over all such segments. Describing the spread of the sensor locations around the mean MD provides an estimate of random errors - and thus the location accuracy. Similarly, we define $aMD(\hat{n}_{k,i}) = 2 \arcsin(\|\bar{n}_k - \hat{n}_{k,i}\|/2)$ as an estimate of the angular accuracy (i.e., the segment-by-segment angular deviation of the corresponding sensor orientations from the mean orientation over segments $\bar{n}_k = \frac{1}{N} \sum \hat{n}_{k,i}$).

141 2.3.2. Systematic errors

142 One limitation to these metrics is that they do not provide information about
 143 systematic errors that would result in a shift in the mean position. Furthermore,
 144 despite subjects' efforts to minimize head movement during coil recordings, the
 145 possibility of small movements cannot be excluded - the subjects heads were
 146 comfortably stabilized with vacuum pillows, but not immobilized. These issues
 147 can be dealt with by taking advantage of the fact that the sensors are housed in
 148 a common cryostat, i.e., fixed relative to each other. The distances between the
 149 (true) sensor locations are thus constant and independent of head movements.
 150 Localization errors can therefore also be estimated by comparing the distances
 151 between the fitted sensor locations with those from the known layout of the
 152 sensor array. To this end, we estimate a relative localization accuracy as the
 153 average deviation of the distances between the estimated sensor locations from
 154 the distances derived from the known layout:

$$\Delta XD(\vec{r}_{k,i}) = \frac{1}{N-1} \sum_{l=1}^N (||\vec{r}_{k,i} - \vec{r}_{l,i}|| - ||\vec{r}_k^* - \vec{r}_l^*||) \quad (6)$$

where $\vec{r}_l$ and $\vec{r}_k$ denote the positions of the localized sensors l and k, $\vec{r}_l^*$ and $\vec{r}_k^*$ their respective positions according to the reference (e.g., the system design), and N=7 the number of sensors. The sum is divided by N-1 because the term for l=k is always zero. This metric is only useful for evaluating individual sensor fits because distances between sensors are constant and determined by the sensor array when jointly localizing the sensors (because the positions are a result of rigidly rotating and translating the sensor array). Analogously, we can estimate the relative localization accuracy with respect to the orientation as the average deviation of the angles between the estimated sensor orientations from the angles between the reference sensor orientations:

$$\Delta XA(\vec{n}_{k,i}) = \frac{2}{N-1} \sum_{l=1}^N (\arcsin(\frac{||\vec{n}_{k,i} - \vec{n}_{l,i}||}{2}) - \arcsin(\frac{||\vec{n}_k^* - \vec{n}_l^*||}{2})) \quad (7)$$

155 where $\vec{n}_l$ and $\vec{n}_k$ denote the orientations of the localized sensors l and k and $\vec{n}_l^*$
 156 and $\vec{n}_k^*$ their orientations according to the reference (e.g., the system design).

157 2.3.3. Head movements

158 Localizing sensors from shorter coil recordings/segments is favourable when
 159 trying to detect - and compensate for - head movements as it enables estima-
 160 tion of recording positions with higher temporal resolution. This is how head
 161 movements are conventionally detected/tracked: the sensor locations with re-
 162 spect to the head are estimated at multiple time instances and compared to the
 163 initial position. In order for us to investigate how the accuracy of the sensor
 164 localization depends on the time the coil signals are recorded, MD , aMD and
 165 ΔXD were computed for different segment lengths t_{trial} between 1 and 10 sec-
 166 onds. For each segment length, the 30 seconds coil recording was split into $n =$
 167 $30/t_{trial}$ consecutive trials.

168 2.3.4. Source localization

169 Finally, we tested the usefulness of our sensor localization procedure in
 170 localizing neural activity. The MEG experiments included recordings of so-
 171 matosensory evoked fields (SEFs). Using our sensor localization method for
 172 co-registration of the on-scalp data, source localization of the N20m-component
 173 was performed and compared to source localization using the conventional MEG
 174 data recorded with the TRIUX system. Because of the small coverage of the
 175 on-scalp system we recorded at four separate locations (aimed to capture the
 176 dipolar field pattern of the N20m-component) and combined the resulting data.
 177 One sensor was excluded due to excessive noise, resulting in 24 individual sensor
 178 locations. The same experimental paradigm - electric stimulation (below motor
 179 threshold) of the median nerve with 360 ms inter-stimulus interval and 1 000
 180 repetitions - as well as preprocessing - bandpass filter between 5 and 200 Hz
 181 with bandstop filters applied at 50 Hz and harmonics, 50 ms pre- to 200 ms
 182 post-stimulus epochs, baseline correction (-50 to 0 ms baseline window), and
 183 time-locked averaging - was used for the recordings with both systems. For
 184 comparability, only the magnetometers were used for the dipole fit with the
 185 TRIUX system.

Figure 2: Spectrum of the measured magnetic fields showing peaks at the coil signal frequencies.

3. Results

The Fourier spectrum of a coil recording is shown in Fig. 2. Clear peaks with a signal-to-noise ratio (SNR) on the order of $\sim 10^2$ are visible at the coil frequencies. An example of a sensor localization based on an 10-second trial can be seen in Fig. 3. In this case, the fitted sensor positions and orientations match well with the design of the sensor array (all pairs being within 0.5 mm and 2 degrees of the design) [8].

In some recordings, individual sensors trapped flux, which led to a strong increase in noise ($\sim 10\times$ higher white noise and a shift in the $1/f$ -like noise knee from 10 to 500-1000 Hz). Localization of these noisy sensors was severely degraded - with errors on the order of centimeters. However, with such high noise data from these sensors was not useful for the MEG recordings and the sensor localization therefore inconsequential.

Fig. 4-a shows the mean euclidean distances of the fitted sensor locations

Figure 3: Example of individually fitted sensor locations and orientations (red). Magnetic dipoles from the coils are shown in green.

from the mean locations $MD(\vec{r}_{k,i})$ as a function of the duration of the coil recording segments t_{trial} used for the localizations. As expected, a clear correlation between the localization accuracy and the length of the coil recordings can be observed. With the exception of channel 1 (which exhibited high noise in the recording) all channels reach $MD < 1$ mm even with just 1-second recordings of the coil signals (channel 1 with four seconds or more). The mean angular deviations from the mean fitted sensor orientations $aMD(\hat{n}_{k,i})$ - seen in Fig. 4-b - show a similar trend versus coil recording time. The orientation fits deviate from the mean by less than 3 degrees with one second of coil signal recording.

Fig. 5-a shows the mean differences of the distances between the fitted sen-

Figure 4: Sensor localization accuracy. a) Mean distance from the mean location $MD(\vec{r}_{k,i})$ as a function of the segment length. b) Mean angular deviation from the mean orientation $aMD(\vec{n}_{k,i})$ for different segment lengths. Error bars indicate one standard deviation.

sors from the distances between sensors in a reference array, $\Delta XD(\vec{r}_k)$. In this case, we used the design of the system as the reference and again present results for different lengths of coil recording segments t_{trial} . On average all channels differ by less than 1 mm from the design already with 1-second coil recordings. With increasing t_{trial} , the mean $\Delta XD(\vec{r}_k)$ converge to values $< \pm 0.4$ mm. These can be assumed to stem from a combination of systematic errors and small deviations between the actual sensor array and the design. As before, the decrease of the standard deviation (i.e., the segment-by-segment spread) with longer coil recording time indicates a decrease in random localization errors. The mean differences of the angles between the fitted sensors from the angles between the sensors in the design of the system $\Delta XA(\vec{n}_k)$, seen in Fig.

Figure 5: Pairwise sensor localization accuracy, with the cryostat design as the reference. a) Mean difference in distance to the other sensors $\Delta XD(\vec{r}_k)$. b) Mean difference in angle to the other sensors $\Delta XA(\vec{n}_k)$.

221 5-a, show a similar decrease in standard deviation with increasing coil recording
 222 time. With 1 second coil recordings all channels differ by ~ 2 degrees or less
 223 from the design of the system.

224 Using short segments, it is possible to continuously monitor the sensor lo-
 225 cations in order to detect movements of the head relative to the sensors. Head
 226 movements manifest themselves as a shift and/or rotation of the whole sensor
 227 array between segments. An example of a head movement captured with 2-
 228 second coil recordings can be seen in figure 6. In this case, the subject's head
 229 moved approximately 2 mm upwards during a stimulus session.

230 Distances from the mean location $MD(\vec{r}_{k,i})$ as well as angular deviations
 231 from the mean orientation $aMD((\vec{n})_{k,i})$ when localizing the sensors jointly are

Figure 6: Successive sensor localizations (red and blue) showing head movement (~ 2 mm) between coil recordings.

shown in Fig. 7. The joint localizations were performed on the same data used to individually localize the sensors in Fig. 4. Both $MD(\vec{r}_{k,i})$ and $aMD(\vec{n}_{k,i})$ show a similar trend as when localizing the sensors individually. Compared to the individual localization, the noisier sensors show significant improvement (especially in $MD(\vec{r}_{k,i})$) while the lower noise sensors worsen. However, the spread in location and orientation around the mean decreases in general, indicating an overall improvement in localization accuracy. This is especially pronounced in case of the location: with one second of data, all sensors exhibit $MD(\vec{r}_{k,i}) < 1$ mm and $aMD(\vec{n}_{k,i}) < 2.5$ degrees (compared to ≤ 2 mm and < 3 degrees, respectively, when localizing them individually). The joint localizations shown here were performed using the sensor positions obtained via individually localizing the sensors with a 10-second coil recording to define the sensor array.

Weighting the sensors according to SNR to reduce the impact of noisy sensors (here, e.g., Ch1) did not result in an improvement in accuracy. In fact, the

Figure 7: Joint sensor localization accuracy using the sensor locations obtained from 10-second coil recording individual localization as rigid sensor array. a) Mean distance from the mean location $MD(\vec{r}_{k,i})$ as a function of the segment length. b) Mean angular deviation from the mean orientation $aMD(\vec{n}_{k,i})$ for different segment lengths. Error bars indicate one standard deviation. For reference, we include the mean of the corresponding deviations that were obtained when localizing the sensors individually as dotted lines.

246 average accuracy for long coil recordings when localizing SNR-weighted sensors
 247 was worse compared to localizing equally weighted sensors.

248 Dipole fits of the N20m-component recorded on-scalp and conventionally
 249 can be seen in Fig. 8. The two dipoles are 5.3 mm apart, which is within the
 250 localization accuracy of conventional whole-head MEG systems [12, 13].

Figure 8: Dipole fits of N20m component based on on-scalp (red) and conventional (blue) MEG recording. The on-scalp dipole fit was performed using individually localized sensor positions estimated with our method.

251 4. Discussion

252 With ≤ 2 mm and < 3 degrees for 1-second coil recordings, the sensor local-
 253 ization method described here reaches significantly higher accuracy than what
 254 has been suggested as required for on-scalp MEG (< 4 mm and < 10 degrees,
 255 according to [14]).

256 An advantage of our method is that it allows for continuous co-registration
 257 in parallel with the MEG recording. Movements of the subject's head during
 258 the MEG recording can thus be detected and accounted for, similarly to con-
 259 tinuous head localization used in commercial whole-head MEG systems. The
 260 measurements shown here were a first practical attempt of using the method
 261 described theoretically in [7]. However, as the experimental session was per-
 262 formed in parallel with other on-scalp MEG experiments, we erred on the side
 263 of caution by turning the coils off during stimulations (in order to avoid the
 264 possibility that they would generate artifacts that might compromise the MEG
 265 recordings). While it remains to be experimentally verified, it is likely that our

method can be used during a stimulus or other experimental protocol because the coil recordings showed no interference at frequencies below 500 Hz (see Fig. 2). Furthermore, in cases where neural signals of interest coincide with the coil frequencies, it is trivial to change the coil frequencies to avoid potential interference (if the neural frequencies of interest are known). The upper limit for the coil frequencies is strictly set by the Nyquist frequency (half of the sampling frequency, in this case $5 \text{ kHz}/2 = 2.5 \text{ kHz}$) and generally should be kept well below any low-pass filters used by the data acquisition system (e.g., anti-aliasing filters, in our case 1 600 Hz).

Taking advantage of the fixed geometry of the sensor array to jointly localize the sensors proved useful. The increased accuracy at shorter segment lengths is especially important for continuous sensor localization. Furthermore, by using individually localized sensor positions from a longer coil recording to define the array geometry, the method is not limited to systems where the sensor array is rigid. For systems consisting of multiple individually positionable sensors [15, 16, 17] or units containing a few sensors [18], one can calibrate the sensor array at the start of a recording by carefully recording the coil signals for a longer duration of time (while minimizing head movement) and localizing the sensors individually. The calibrated array can then be used for fast, joint sensor localization. This, of course, assumes that the sensors are fixed with respect to one another for the duration of the recording.

Localized sensor positions and orientations were used to fit an equivalent current dipole to somatosensory evoked activity recorded sequentially at multiple locations. The estimated dipole position from the on-scalp recording was $\sim 4 \text{ mm}$ from that which was estimated from the conventional MEG recording. This lies well within the 8-11 mm variability seen between different commercial, whole-head MEG systems [13]. Considering the differences in sampling between on-scalp and conventional MEG, it is also possible that the on-scalp system is differently sensitive to neural activity, as compared conventional MEG. Previous works by our group with a high- T_c SQUID [10] as well as by Zetter et al. [14] with OPMs also report differences between the N20m-components detected with

297 on-scalp and conventional MEG systems.

298 The measurements reported here were part of a series of benchmarking
299 recordings to compare an on-scalp MEG system [8] to a commercial, whole-
300 head MEG system. It was therefore possible to use full-head recordings of the
301 coil array on the subject's head in order to reliably estimate the positions and
302 orientations of the dipolar coils. This is, however, not a viable solution for on-
303 scalp systems in general. The coil orientations should instead be inferred from
304 other measurements. Flat coils with markers to digitize the orientation as part
305 of the head-digitization would be able to solve this issue in the future.

306 5. Conclusion

307 We have presented a method for localizing MEG sensors with the help of
308 magnetic dipole-like coils (introduced in [7]) and implemented it in a set of
309 on-scalp MEG recordings using a 7-channel, high- T_c SQUID-based system [8].
310 The method provided high accuracy estimates of the sensor positions and ori-
311 entations with short averaging time (≤ 2 mm and < 3 degrees respectively
312 with 1-second coil recordings). It enables continuous estimation of the posi-
313 tions of sensors with respect to a subject's head (i.e., head localization) with
314 good temporal resolution. Calibrating and jointly localizing the sensor array
315 can furthermore improve the localization accuracy (< 1 mm and < 2.5 degrees
316 respectively with 1-second coil recordings). We demonstrate the efficacy of the
317 method by using it in localization of neural activity.

318 Acknowledgments

319 Data for this study was collected at NatMEG, the National infrastructure for
320 Magnetoencephalography, Karolinska Institutet, Sweden. The NatMEG facility
321 is supported by the Knut & Alice Wallenberg foundation (2011-0207). This
322 work was financially supported by the Knut and Alice Wallenberg foundation
323 (KAW 2014.0102), the Swedish Research Council (2017-00680), the Swedish

Childhood Cancer Foundation (MT2014-0007), and Tillväxtverket via the European Regional Development Fund (20201637).

References

- [1] E. Boto, R. Bowtell, P. Krüger, T. M. Fromhold, P. G. Morris, S. S. Meyer, G. R. Barnes, M. J. Brookes, On the potential of a new generation of magnetometers for MEG: a beamformer simulation study, *PLOS ONE* 11 (8) (2016) e0157655.
- [2] J. Iivanainen, M. Stenroos, L. Parkkonen, Measuring MEG closer to the brain: Performance of on-scalp sensor arrays, *NeuroImage* 147 (2017) 542–553.
- [3] J. F. Schneiderman, S. Ruffieux, C. Pfeiffer, B. Riaz, On-scalp meg. in: Supek s., aine c. (eds) *magnetoencephalography*, Springer (2019) 1–23.
- [4] B. Riaz, C. Pfeiffer, J. F. Schneiderman, Evaluation of realistic layouts for next generation on-scalp MEG: spatial information density maps, *Scientific Reports* 7 (1) (2017) 6974.
- [5] S. Ern, L. Narici, V. Pizzella, G. Romani, The positioning problem in biomagnetic measurements: A solution for arrays of superconducting sensors, *IEEE Transactions on Magnetics* 23 (2) (1987) 1319–1322.
- [6] K. Uutela, S. Taulu, M. Hmlinen, Detecting and correcting for head movements in neuromagnetic measurements, *NeuroImage* 14 (6) (2001) 1424–1431.
- [7] C. Pfeiffer, L. M. Andersen, D. Lundqvist, M. Hmlinen, J. F. Schneiderman, R. Oostenveld, Localizing on-scalp MEG sensors using an array of magnetic dipole coils, *PLOS ONE* 13 (5) (2018) e0191111.
- [8] C. Pfeiffer, S. Ruffieux, L. Jnsson, M. L. Chukharkin, A. Kalaboukhov, M. Xie, D. Winkler, J. F. Schneiderman, A 7-channel high-Tc SQUID-based on-scalp MEG system, *bioRxiv* (2019) 534107.

- 351 [9] M. Xie, J. F. Schneiderman, M. L. Chukharkin, A. Kalabukhov, B. Riaz,
352 D. Lundqvist, S. Whitmarsh, M. Hämäläinen, V. Jousmäki, R. Oostenveld,
353 et al., Benchmarking for on-scalp MEG sensors, *IEEE Transactions on*
354 *Biomedical Engineering* 64 (6) (2017) 1270–1276.
- 355 [10] L. M. Andersen, R. Oostenveld, C. Pfeiffer, S. Ruffieux, V. Jousmäki,
356 M. Hämäläinen, J. F. Schneiderman, D. Lundqvist, Similarities and dif-
357 ferences between on-scalp and conventional in-helmet magnetoencephalog-
358 raphy recordings, *PLOS ONE* 12 (7) (2017) e0178602.
- 359 [11] R. Oostenveld, P. Fries, E. Maris, J.-M. Schoffelen, Fieldtrip: Open source
360 software for advanced analysis of MEG, EEG, and invasive electrophys-
361 iological data, *Computational Intelligence and Neuroscience* 2011 (2011)
362 1–9.
- 363 [12] R. Leahy, J. Mosher, M. Spencer, M. Huang, J. Lewine, A study of dipole
364 localization accuracy for MEG and EEG using a human skull phantom,
365 *Electroencephalography and clinical neurophysiology* 107 (2) (1998) 159–
366 173.
- 367 [13] T. Bardouille, L. Power, M. Lalancette, R. Bishop, S. Beyea, M. J. Tay-
368 lor, B. T. Dunkley, Variability and bias between magnetoencephalography
369 systems in non-invasive localization of the primary somatosensory cortex,
370 *Clinical neurology and neurosurgery* 171 (2018) 63–69.
- 371 [14] R. Zetter, J. Iivanainen, M. Stenroos, L. Parkkonen, Requirements for
372 coregistration accuracy in on-scalp MEG, *Brain topography* 31 (6) (2018)
373 931–948.
- 374 [15] O. Alem, R. Mhaskar, R. Jiménez-Martínez, D. Sheng, J. LeBlanc,
375 L. Trahms, T. Sander, J. Kitching, S. Knappe, Magnetic field imaging with
376 microfabricated optically-pumped magnetometers, *Optics Express* 25 (7)
377 (2017) 7849–7858.

- 378 [16] J. Iivanainen, R. Zetter, M. Grön, K. Hakkarainen, L. Parkkonen, On-
379 scalp MEG system utilizing an actively shielded array of optically-pumped
380 magnetometers, *NeuroImage* 194 (2019) 244 – 258.
- 381 [17] E. Boto, N. Holmes, J. Leggett, G. Roberts, V. Shah, S. S. Meyer, L. D.
382 Muñoz, K. J. Mullinger, T. M. Tierney, S. Bestmann, et al., Moving magne-
383 toencephalography towards real-world applications with a wearable system,
384 *Nature* 555 (7698) (2018) 657.
- 385 [18] A. Borna, T. R. Carter, J. D. Goldberg, A. P. Colombo, Y.-Y. Jau, C. Berry,
386 J. McKay, J. Stephen, M. Weisend, P. D. Schwindt, A 20-channel mag-
387 netoencephalography system based on optically pumped magnetometers,
388 *Physics in Medicine & Biology* 62 (23) (2017) 8909.