CLOSED-LOOP NEUROSCIENCE OF BRAIN RHYTHMS: OPTIMIZING REAL-TIME QUANTIFICATION OF NARROW-BAND SIGNALS TO EXPEDITE FEEDBACK DELIVERY

A PREPRINT

Nikolai Smetanin

Center for Bioelectric Interfaces
Higher School of Economics
Moscow, Russia, 101000
n.m.smetanin@gmail.com

Anastasia Belinskaya

Center for Bioelectric Interfaces
Higher School of Economics
Moscow, Russia, 101000
belinskaya.anastasy@gmail.com

Mikhail Lebedev

Center for Bioelectric Interfaces
Higher School of Economics
Moscow, Russia, 101000
mikhail.lebedev@gmail.com

Alexei Ossadtchi*

Center for Bioelectric Interfaces
Higher School of Economics
Moscow, Russia, 101000
ossadtchi@gmail.com

November 29, 2019

Keywords Closed-loop neuroscience, neurofeedback, brain rhythm, envelope, instantaneous phase, feedback latency, optimal filtering, Hilbert transform

ABSTRACT

Closed-loop Neuroscience is based on the experimental approach where the ongoing brain activity is recorded, processed, and passed back to the brain as sensory feedback or direct stimulation of neural circuits. The artificial closed loops constructed with this approach expand the traditional stimulus-response experimentation. As such, closed-loop Neuroscience provides insights on the function of loops existing in the brain and the ways the flow of neural information could be modified to treat neurological conditions.

Neural oscillations, or brain rhythms, are a class of neural activities that have been extensively studied and also utilized in brain rhythm-contingent (BRC) paradigms that incorporate closed loops. In these implementations, instantaneous power and phase of neural oscillations form the signal that is fed back to the brain.

Here we addressed the problem of feedback delay in BRC paradigms. In many BRC systems, it is critical to keep the delay short. Long delays could render the intended modification of neural activity impossible because the stimulus is delivered after the targeted neural pattern has already completed. Yet, the processing time needed to extract oscillatory components from the broad-band neural signals can significantly exceed the period of oscillations, which puts a demand for algorithms that could minimize the delay.

We used EEG data collected in human subjects to systematically investigate the performance of a range of signal processing methods in the context of minimizing delay in BRC systems. We proposed a family of techniques based on the least-squares filter design – a transparent and simple approach, as it required a single parameter to adjust the accuracy versus latency trade-off. Our algorithm performed on par or better than the state-of-the-art techniques currently used for the estimation of rhythm envelope and phase in closed-loop EEG paradigms.

1 Introduction

Investigations of neural oscillations have a long history and continue to be an area of intensive research, particularly when such neuroimaging techniques are used as noninvasive electroencephalography (EEG) and magnetoencephalography (MEG), and invasive electrocorticography (ECoG) and stereo EEG (sEEG).

A plethora of experimental paradigms and relevant analysis methods have been developed for dealing with specific types of neuronal oscillations, including the methods for their induction and suppression [1, 2]. These paradigms fall in one of two categories. In the first category of studies, changes in neural oscillations are investigated that are induced by a variety of stimuli; the stimuli are presented without a consideration of the ongoing brain activity. In the second category [3], a closed-loop design is implemented where the stimuli are selected based on the characteristics of the ongoing brain activity.

Below, we describe the studies from the second category where the closed loop is formed from neural oscillatory activity. We refer to this experimental approach as brain rhythm contingent (BRC) paradigm.

1.1 Brain rhythm contingent paradigms

As shown in Figure 1A, the BRC paradigm operates through three steps: data acquisition, data processing, and stimulus generation. During the data acquisition step, brain activity is measured, usually with multiple spatially distributed electromagnetic sensors, and streamed to a computer. During the data processing step, a computer routine handles the multichannel data in real-time to extract the parameters of oscillatory activity, typically amplitude and phase. Lastly, during the stimulus generation step, these parameters are converted into a feedback delivered to the brain either directly, using stimulation applied to the nervous tissue (also called neuromodulation), or through natural senses: vision, hearing or touch. The feedback could act in a subtle way by modulating the parameters of an already ongoing stimulation (direct or through natural senses) or by contributing to the algorithm that selects a stimulus from a set of discrete possibilities.

Figure 1: Schematics of the BRC paradigm. A. A diagram depicting signal flow in a closed-loop system. B. Signal processing pipeline. C. The sources of delays mounting to the total latency of the BRC system. Technical and fundamental sources of the delay are marked in blue and red, respectively.

Many implementations of BRC paradigm have been developed, which allow implementation of a variety of goal-directed behaviors dependent on a closed-loop design [3]. The most distinct paradigms are: neurofeedback (NFB), brain computer interface (BCI), closed-loop brain stimulation, and brain state-contingent stimulus delivery.

The NFB is a form of biofeedback that enables subjects with the capacity to monitor and control their own brain activity [4]. With NFB, subjects gain access to the neural signals of different brain structures, and learn to modulate them in specific ways [5], [6], [4], [7]. NFB approach is used as a therapy for neurological disorders [8], [9] and as cognitive enhancement therapy [10]. Steps of NFB operation include extraction of neural features of interest, their transformation using case-specific algorithms, and generation of sensory feedback delivered to the subject.

BCIs operate very much like NFB, with neural activity being recorded, processed, and directed to an external device that provides some sort of feedback to the subject. Yet, the emphasis here is not on the feedback per se but on the brain control of the device, which serves some useful purpose, for example a computer cursor or a prosthetic limb [11]. Clinically relevant BCIs are intended for functional restoration and rehabilitation of patients with neurological disabilities. For example, an EEG-based, motor-imagery BCI that controls a hand exoskeleton aids in rehabilitation of stroke patients [12], [13] by facilitating Hebbian plasticity that occurs owing to synchronization of cortical modulations with proprioceptive feedback caused by exoskeleton movements [14].

In BRC paradigms with *closed-loop brain stimulation*, parameters of neural oscillations, most often instantaneous amplitude and phase, affect the characteristics of stimulation directly applied to the brain [15]. Electromagnetic devices are commonly used to deliver the stimulation, which can be invasive, such as deep brain stimulation (DBS) [16], or noninvasive, such as transcranial brain stimulation (TBS) [17]. Thus, EEG oscillatory patterns have been used to control such types of TBS as transcranial magnetic stimulation (TMS) [18] and transcranial alternating current stimulation (tACS)[19]. BRC paradigms with stimulation through normal senses can be considered a type of closed-loop brain stimulation, as well, for example, stimulation with continuously flashing visual stimuli[20].

BRC paradigms have gained popularity, where cognitive processes are investigated under experimental conditions with *brain rhythm-contingent stimulus delivery*. This is because the brain handles cognitive tasks differently depending on the brain oscillatory patterns taking place just before the task onset [21, 22, 23, 24, 25, 26, 27]. Accordingly, it is of interest to create an experimental paradigm, where the task starts only after a particular oscillatory pattern is detected. Such brain rhythm-contingent stimulus delivery can drastically reduce experimental time and minimize participants' fatigue, particularly when the neural patterns of interest are represented by short-lived bursts of activity [28] or desired brain states[29].

1.2 Signal processing and latency in BRC paradigms

Despite the conceptual differences between the four applications of the BRC paradigm described above, their signal processing pipelines are quite similar and can be summarized as shown in Figure 1B.

In this paper, we consider the BRC paradigms based on EEG and MEG recordings. The major advantage of these imaging techniques over those that employ metabolic (e.g., positron-emission tomography) or blood oxygenation level-dependent (e.g., functional magnetic resonance imaging) measurements is their unsurpassed millisecond-scale temporal resolution. This fine temporal resolution allows exploring fine rhythmic structure of brain activity and, in principle, enables nearly instantaneous interaction with the brain circuits. Yet, as explained in detail below, time-lags of both technical and fundamental nature occur during the online extraction of oscillatory parameters from the ongoing brain activity. These lags vary from one implementation to another and can be significantly reduced with the use of appropriate signal processing methods.

For a BRC paradigm to be efficient, it needs to be temporally specific, that is feedback should be issued when it can affect the targeted neuronal activity. Temporal specificity of a BRC system is characterized by an overall delay between the onset of the neural event of interest and the time when the participant receives the stimulus corresponding to this event (Figure 1C. This delay, called the overall BRC system latency, incorporates time-lags related to different factors, some of them technical (i.e, software and hardware delays) and some fundamental (i.e. required to collect a sufficient amount of neural data).

The technical time-lags typically do not exceed 100 ms; this time is needed for hardware communication and low-level software processing. This delay can be reduced by specific hardware and software solutions. Based on our experience, it is feasible to reduce the technical delay to 20-30 ms.

The fundamental lag cannot be reduced that easily because it is composed of the time needed to collect a snapshot of neural oscillatory data that could be then quantified with an appropriate algorithm like band-pass filtering followed by the extraction of information about the instantaneous power or phase of the narrow-band process. If suboptimal

approaches are used for extracting power and/or phase of rhythmic neural components, the fundamental delay amounts to approximately 0.5 s, causing undesirable effects in BRC implementations.

The adverse effect of feedback delays has been reported in many NFB and the BCI studies, where participants learned to properly modulate their own brain activity, with or without a strategy recommended by the experimenter [30]. A similar problem with long feedback delays has been known from the studies on evidence-based learning. Thus, back in 1948, Grice showed that learning to discriminate complex visual patterns drastically depended on the feedback signal latency [31]. Impaired performance with delayed has been also demonstrated in the studies on motor learning e.g for the prism adaptation task [32]. According to Rahmandad et al. [27], behavioral learning is impaired when feedback delay is unknown. Moreover, elongated feedback delay decreases the sense of agency during BCI control [33], the finding also corroborated by a simulation study [34] showing that feedback delay and temporal blur adversely influence the automatic (strategy free) learning.

Temporal specificity is also an an important consideration for the experimental settings with closed-loop brain stimulation and brain state triggered stimulus delivery. Indeed, the key requirement for these methods are the accurate estimation of instantaneous oscillatory features and the timely delivery of the stimuli to efficiently interfere with the oscillatory neural patterns.

1.3 Low-latency method for envelope and phase detection

In the present study, we explore several approaches aimed at reducing the delay between neuronal events and the corresponding feedback in the BRC paradigm. We propose a simple method to directly control the delay. Our algorithm is based on the adaptive optimization of the complex-valued finite impulse response (FIR) filter weights to yield the desired reduction of delay. The proposed approach is capable of extracting the instantaneous power and phase of neural oscillations with the shorter latency and higher accuracy as compared to other relevant techniques. With this approach, the user can explicitly specify the desired delay and assess the corresponding accuracy of envelope estimation. These features of the algorithm allow for flexible control of the latency-accuracy trade-off in BRC applications. Strikingly, our approach can work even with negative delays, paving way for achieving predictive feedback.

2 Methods

Our basic assumption is that the measured neural activity $x[n]$ is a sum of the narrow-band signal $s[n]$ (targeted neural activity of a BRC paradigm) and background colored broad-band noise $\eta[n]$.

$$x[n] = s[n] + \eta[n] \quad (1)$$

The targeted neural activity $s[n]$ can be represented as the real part of analytic signal $y[n]$ [35] :

$$y[n] = s[n] + js_h[n] = a[n]e^{j\phi[n]} \quad (2)$$

where $s_h[n]$ is the imaginary part of the analytic signal, often called “second quadrature” of the original signal $s[n]$, $a[n]$ is the envelope, reflecting the instantaneous power of the narrow band process at time stamp n , and $\phi[n]$ is its instantaneous phase at the time stamp n . Importantly, once the estimate of the complex valued analytic signal $y[n]$ is available that corresponds to signal $s[n]$, the envelope $a[n]$ and the phase $\phi[n]$ of $s[n]$ can be computed as

$$a[n] = \sqrt{\Re\{y[n]\}^2 + \Im\{y[n]\}^2}, \phi[n] = \text{atan}\left(\frac{\Im\{y[n]\}}{\Re\{y[n]\}}\right) \quad (3)$$

Figure 2A shows that, with these expressions, extraction of the ground-truth signal is not restricted to causal operations. This is because the discrete Fourier transform (DFT) and narrow-band Hilbert transform in the frequency domain can be applied to the entire batch of data, followed by the conversion of the real and imaginary components of the analytic signal into the ground-truth values of the instantaneous envelope $a[n]$ and phase $\phi[n]$.

In addition to the non-causal approach illustrated in Figure 2, the other panels of Figure 2 show a graphical summary of the available causal techniques that can be applied in real time to compute the envelope and instantaneous phase of a narrow-band component extracted from a broad-band signal.

2.1 Existing methods

For didactic purposes, we start with the most basic approaches.

Figure 2: Methods for narrow-band signal envelope estimation. A. Ideal non-causal system for ground truth signal extraction. B. Envelope detector based on rectification of the band-filtered signal. C. Sliding window narrow-band Hilbert transform-based method. D. cFIR family filters that perform finite impulse response causal filter approximation of ideal non-causal systems. D'. Filter design for cFIR family filters.

Rectification and smoothing of the band-filtered signal (*rect*) is the conceptually most straightforward way to estimate the envelope $a[n]$. In this method, low-pass filtering is performed applied to the rectified narrow-band filtered signal. This approach can be mathematically expressed as:

$$a_r[n] = h_{lp}[n] * |h_{bp}[n] * x[n]| \quad (4)$$

where $*$ is the convolution operator, h_{bp} is the impulse response of the band-pass filter, $|\cdot|$ denotes the absolute value (i.e., the rectification step), and h_{lp} is the impulse response of the low-pass filter that performs smoothing of the rectified signal. Without loss of generality, we can assume that both h_{bp} and h_{lp} are linear phase FIRs designed by a Hanning window method with the number of taps N_{bp} and N_{lp} , correspondingly. The cutoff frequency for the low-pass filter f_c is set to correspond to one-half the expected bandwidth of the narrow-band process, i.e. $f_c = (f_2 - f_1)/2$. FIR filters have a linear phase and therefore the total delay D and the number of taps in the individual filters N_{bp} and N_{lp} have the following interrelationship:

$$D = \frac{N_{lp} - 1}{2} + \frac{N_{bp} - 1}{2} \quad (5)$$

In order to ensure that the maximum performance is achieved for a given group delay value D , we used grid search over variables N_{bp} and D in our comparative analysis. The parameter N_{lp} was determined from the formula 5. Since N_{bp} and N_{lp} are positive, this method can estimate the envelope values only with a positive delay, which corresponds to the count of signal samples taken from the past.

Sliding window narrow-band Hilbert transform (*hilb*) is the second most commonly used method that is based on the use of the analytic signal $y[n]$ computed using Hilbert transform [35]. There are various implementations of this approach. In the current work we resorted to the use of the windowed DFT. DFT is calculated on each window of length N_t which is zero-padded to the length of N_f samples. Next, the coefficients corresponding to the negative frequency values and those in the positive frequency semi-axis that fall outside the band of interest are zeroed out. The DFT coefficients within the band of interest are doubled. Then, the inverse DFT is performed and $N_t - D$ -th element of the resultant complex valued sequence is used as an estimate of the analytic signal with delay D . This way, two operations are performed simultaneously: band-pass filtering and extracting the analytic signal that is then used to estimate the envelope. This algorithm is illustrated in Figure 2.

In matrix representation and using temporal embedding to form vector, $\mathbf{x}[n]$ this can be written as:

$$y_h[n] = \frac{2}{N_f} \cdot \mathbf{w}_{N_t-D}^* \mathbf{W}_{\Delta f} \mathbf{x}[n], \quad (6)$$

where vector $\mathbf{x}[n]$ contains the last N_t samples of $x[n]$, i.e. $\mathbf{x}[n] = [x[n], x[n-1], \dots, x[n-N_t+1]]$, $\mathbf{W}_{\Delta f}$ is the N_f -by- N_t modified DFT matrix with zeros on the k -th row for k outside of $[f_1 N_f / f_s, f_2 N_f / f_s]$ range corresponding to the physical band of interest $\Delta f = [f_1, f_2]$ and $\mathbf{w}_{N_t-D}$ is the $(N_t - D)$ -th row of the DFT matrix.

The parameters to be optimized for this method are window length N_t and zero-padded length of the signal N_f which is used to perform the DFT. The overall delay of this method is explicitly determined by parameter D .

Sliding window Hilbert transform with AR prediction of the narrow-band filtered signal (*ffiltar*) was proposed by Chen et al. [36] and applied practically [18]. In this method, the sliding window vector $\mathbf{x}[n]$ containing N_a last samples is forward-backward band-pass filtered. Then, flanker N_e samples are truncated to eliminate edge artifacts, and $2N_b$ samples are forward predicted by using an AR model fitted to $N_a - N_e$ samples. Finally, Hilbert transformation of prediction is used to estimate the analytic signal value in the middle of the predicted range. In the original work, this value was used to determine the current phase and time for stimulation. Here we used this technique as a benchmark for our and other methods being compared but only at the processing latency $D = 0$, that is the latency this approach was originally designed for.

Using matrix notation, we can formalize this method as follows.

$$y_p[n] = \frac{2}{N_f} \cdot \mathbf{w}_{N_a}^* \mathbf{W}_+ \mathcal{P}_{AR(p)} \{\tilde{\mathbf{x}}[n]\} \quad (7)$$

where $\tilde{\mathbf{x}}[n]$ contains forward-backward filtered last N_a samples of the $x[n]$, $\mathcal{P}_{AR(p)}$ denotes AR model based prediction operation and adds $2N_e$ predicted samples by using a p -th order AR model, $\mathbf{W}_+$ is the N_f -by- $(N_a + N_e)$ modified DFT matrix with zeros on the rows corresponding to the negative frequencies and $\mathbf{w}_{N_a}$ is the (N_a) -th row of the DFT matrix. As the narrow-band filter for the forward-backward filtering part, we use Butterworth filter of the order k as suggested in [36].

One of the disadvantages of this approach is that it has multiple parameters that need to be tuned to achieve the optimal performance. To attain the best performance in this study, we searched over the parameter grid composed of the following variables: AR order p , number of edge samples N_e , and Butterworth filter of the order k .

2.2 Proposed method

In order to build the analytic signal $y[n]$ that corresponds to the narrow-band signal $s[n]$ extracted from the noisy broadband measurements $x[n]$, one can apply the ideal narrow-band Hilbert transform filter [37, 38]. The complex-valued frequency response of this combined filter can be defined as:

$$H_D(e^{jw}) = \begin{cases} 2e^{-jwD}, & w \in [w_c - \delta w, w_c + \delta w] \subseteq [0, \pi] \\ 0, & \text{otherwise} \end{cases} \quad (8)$$

where $\delta w = 2\pi\delta f$ is half of the pass band width, and D is the group delay measured in samples. Strictly speaking, for any finite delay D this filter is non-causal and cannot be applied in real time. To reconstruct the analytic signal causally, one can find a causal complex-valued finite impulse response (cFIR) filter $\mathbf{b} = \{b[n]\}$ of length N_t that approximates the ideal complex valued frequency response $H_D(e^{jw})$ [39]. This filter can be then applied in real-time as $y_c[n] = x[n] * b[n]$, the procedure that incurs a fixed processing delay of D samples.

Causal complex valued FIR $b[n]$ can be found by solving the least squares optimization problem. Various definitions of the cost functions lead to different filters.

Frequency domain least squares (cFIR) is the first and most straightforward approach (denoted F-cFIR). The least squares filter design strategy consists in finding the complex valued vector of the cFIR filter weights $\mathbf{b}$ of length N_t that minimizes the L_2 norm between the cFIR filter frequency response obtained by the DFT and the discrete appropriately sampled version $\mathbf{h}_D$ of an ideal response H_D in the frequency domain. To increase the frequency resolution, we use truncated N_f -samples DFT-matrix $\mathbf{W}$ with dimension $N_f \times N_t$ ($N_f \geq N_t$). The use of transform matrix-based formulation of the DFT in this case is equivalent to the DFT of N_t samples long vector zero-padded to length N_f . Taking in account the fact that due to orthogonality $\mathbf{W}^H \mathbf{W} = N_f \mathbf{I}$ the solution $\mathbf{b}_F$ of the normal equation for the optimization problem (2.2) can be found by a simple inverse DFT (9) of the desired complex valued characteristics of the narrow-band Hilbert transformer.

$$\mathbf{b}_{cFIR} = \arg \min_{\mathbf{b}} \|\mathbf{W}\mathbf{b} - \mathbf{h}_D\|_{L_2}$$

$$\mathbf{b}_{cFIR} = \frac{1}{N_f} \mathbf{W}^H \mathbf{h}_D \quad (9)$$

The last formula 9 is equivalent to expression 6 but can be used with negative delays, $D \leq 0$. This simple method, however, does not take into account the second order frequency domain statistics of the target signal and could be further improved. Note that the cFIR approach with parameters N_t , N_f and $D \geq 0$ matches the sliding window narrow-band Hilbert transform approach with the same parameters N_t , N_f and D .

Frequency domain weighted least squares (wcFIR) is the method that follows optimal filter design ideas, where power spectral density of the input signal $x[n]$ are used as weights. We thus formulate the weighted frequency domain least squares design technique (denoted wcFIR) via optimization problem (2.2) whose solution can be found by solving the normal equations (10):

$$\mathbf{b}_{wcFIR} = \arg \min_{\mathbf{b}} \|\mathbf{X}(\mathbf{W}\mathbf{b} - \mathbf{h}_D)\|_{L_2}$$

$$\mathbf{b}_{wcFIR} = (\mathbf{W}^H \mathbf{X}^T \mathbf{X} \mathbf{W})^{-1} \mathbf{W}^H \mathbf{X}^T \mathbf{X} \mathbf{h}_D \quad (10)$$

where $\mathbf{X}$ is the diagonal matrix formed from the square roots of the power spectral density magnitudes of $x[n]$. The temporal dimension of $\mathbf{W}$ is set according to the specified delay D by cropping $\mathbf{W}$ from the full-blown twiddle-factor matrix. Therefore only $\mathbf{W}^H \mathbf{W} = N_f \mathbf{I}$ holds true while $\mathbf{W} \mathbf{W}^H \neq N_f \mathbf{I}$. This way, at the optimum $\|\mathbf{W}\mathbf{b} - \mathbf{h}_D\| \neq 0$ and therefore $\mathbf{b}$ is just an approximation of the ideal filter and can be computed even for negative D .

Panels D and D' of Figure 2 illustrate these two approaches. The delay D corresponds to the slope of the phase response within the pass-band and theoretically can be set to an arbitrary value. Then, the optimization procedure aims at finding such complex vector of cFIR filter coefficients $\mathbf{b}$ that both ideal phase response and an ideal amplitude response are approximated sufficiently and accurately. We do so in the frequency domain by analytically solving the least squares or the weighted least squares problems.

Conceptually, having in mind the two tasks of optimal envelope and instantaneous phase estimation we could have formulated the two separate optimization problems and used two different sets of weights implementing two different band-pass complex-valued filters delivering optimal accuracy in estimation of envelope and phase approximation with the specified delay. In this case, however, due to non-linearity of the target functional, we would have to perform an iterative optimization in order to find the optimal FIR filter weights vector $\mathbf{b}$.

Time domain least squares is the last approach from this family (denoted tcFIR) that is based on minimization of the squared distance in the time domain between the complex delayed ground truth signal $y[n - D]$ and the filtered signal $x[n] * b[n]$:

$$\mathbf{b}_{tcFIR} = \arg \min_{\mathbf{b}} \|x[n] * b[n] - y[n - D]\|_{L_2} \quad (11)$$

In this case, the ground truth signal $y[n-D]$ is obtained non-causally from the training samples via an ideal zero-phase Hilbert transformer (8). According to Parseval's theorem, this approach is equivalent to the wcFIR approach. However in contrast to the frequency domain formulation, it allows for implementation of recursive schemes for solving (11) and therefore may potentially account for non-stationarity in the data. One of the most straightforward approaches is to use recursive least squares (RLS) [40], to update filter coefficients on the fly.

2.3 Methods comparison

Data and preprocessing We compared the described methods using resting state EEG data recorded from 10 subjects during NFB training sessions. EEG recordings were performed using 32 AgCl electrodes placed according to a 10-20-system with the ground electrode at AFz position and reference electrodes on both ears. The impedance for all electrodes was kept below 10 KOhm. The signal was sampled at 500 Hz using the NVX-136 amplifier (Medical Computer Systems Ltd.) and bandpass-filtered in 0.5-70 Hz band. These *preprocessing* filters incurred overall delay of no more than 10 milliseconds in the bandwidth of interest (8-12Hz).

For each subject, we used 2 minutes of resting state recordings. The first minute of the data – for training or the parametric grid search and the second minute is used for testing the performance. To eliminate eye artifacts we performed independent component analysis (ICA) on the training data, identified eye-movement related components by means of the mutual information spectrum [41] and removed from the data several components exhibiting the highest mutual information with either of the two frontal channels Fp1 and Fp2.

For the following analysis only parietal P4 channel of the cleaned data is used as feedback signal $x[n]$ in (1).

Individual alpha band We determine individual alpha range by the following procedure: estimate power spectrum using Welch method with 2 second 50%-overlap boxcar window, find the frequency f_0 with maximal SNR in the 8-12Hz range, define individual band as $[f_1, f_2]$ interval, where $f_1 = f_0 - 2$ Hz and $f_2 = f_0 + 2$ Hz.

Ground truth signal As the ground truth signal $s[n]$ in (1), we used non-causally computed analytic signal obtained by zeroing out the DFT coefficients corresponding to the frequencies of the individual alpha band $[f_1, f_2]$ followed by the Hilbert transform. Once we have the analytic signal, we can estimate both envelope and instantaneous phase without any additional delay.

Performance metrics To measure the accuracy of the envelope and instantaneous phase estimates obtained with the described techniques as a function of the group delay D we used the following metrics. All metrics are based on the ground truth envelope and instantaneous phase information extracted from the ground truth signal $s[n]$ extracted non-causally from the real EEG data as described above and then shifted to match the specific delay value D of causal processing. We will denote the shifted ground truth envelope and phase sequences as $a[n-D]$ and $\phi[n-D]$ correspondingly.

For performance assessment of envelope estimation methods, we calculated the Pearson correlation coefficient between estimated envelope $\hat{a}[n]$ obtained causally by each of the methods and the appropriately shifted ground truth envelope sequence $a[n-D]$:

$$r_a = \frac{\sum_{n \in \mathcal{N}_a} (a[n-D] - m_a)(\hat{a}[n] - m_{\hat{a}})}{\sqrt{\sum_{n \in \mathcal{N}_a} (a[n-D] - m_a)^2} \sqrt{\sum_{n \in \mathcal{N}_a} (\hat{a}[n] - m_{\hat{a}})^2}} \quad (12)$$

where $\mathcal{N}_a = D..N-1$ is the set of time indices with $N = 30000$, m_a and $m_{\hat{a}}$ are sample means averages of $a[n]$ and $\hat{a}[n]$ over set $\mathcal{N}_a$

To asses performance of instantaneous phase estimation we used bias b_ϕ , absolute bias $|b_\phi|$ and the standard deviation σ_ϕ of the delayed ground truth phase $\phi[n-D]$ at the time moments when predicted $\hat{\phi}[n]$ phase crosses 0. These metrics reflect the bias, absolute bias, and the variance in determining zero-crossing moments (negative-to-positive direction) of the delayed signal $s[n-D]$.

$$b_\phi = \frac{1}{|\mathcal{N}_\phi|} \sum_{n \in \mathcal{N}_\phi} \phi[n-D] \quad (13)$$

Method	Param.	Grid
<i>rect</i>	N_{bp}	0, 5, ..., 100
<i>cfir/hilb</i>	N_t	250, 500, 1000
<i>wcfir</i>	N_t	250, 500, 1000
<i>acfir</i>	N_t	250, 500, 1000
	N_u	25, 50
	μ	0.7, 0.8, 0.9
<i>ffiltar</i>	p	10, 25, 50
	N_e	10, 25, 50

Table 1: Grid search space for each method.

$$\sigma_\phi = \sqrt{\frac{1}{|\mathcal{N}_\phi| - 1} \sum_{n \in \mathcal{N}_\phi} (\phi[n - D] - b_\phi)^2} \quad (14)$$

where $\mathcal{N}_\phi = \{n : n \in \mathcal{N}_a, \text{sign}(\hat{\phi}[n]) > \text{sign}(\hat{\phi}[n - 1])\}$ is the set of time moments when $\hat{\phi}[n]$ crosses 0.

Grid search procedure To ensure that the compared methods operate optimally, for each of them, we defined grid search space as described in table 1 and, as described below, looked for the combination of parameters that ensured the best performance for each of the techniques. Here we use the following short method names: *rect* for envelope detector based on rectification of the band-filtered signal, *cfir* for the frequency domain least squares designed cFIR, *wcfir* for the frequency domain weighted least squares designed cFIR and *acfir* for RLS based cFIR filter update scheme. Note that sliding window narrow-band Hilbert transform approach (*hilb*) exactly matches the *cfir* on the range of positive delays.

For each combination of parameters and fixed delay D , we computed the metrics defined above on the training set separately for each subject. Note that for *rect* and *hilb* no negative delay D is possible and for *ffiltar* we used only zero delay ($D = 0$) as this technique was formulated and used in the closed-loop experiments in this specific condition. Frequency band and weights for the *wcfir* approach were computed based on the same training data. We then used optimal values of parameters for each method corresponding to the maximum of r_a , minimum for $|b_\phi|$ and minimum of σ_ϕ values observed on the training set, and estimated the same performance metrics r_a , b_ϕ , $|b_\phi|$ and σ_ϕ on the test data.

3 Results

Figure 3 shows the performance comparison results averaged over the data for ten subjects for the methods explored in this study. To ensure the best performance for each of the techniques, we used training data segments to tune each method's parameters using the grid search procedure described in the Methods section. For each method in panels A, B, C and D, we show the performance metrics r_a , σ_ϕ , b_ϕ , $|b_\phi|$ reflecting envelope correlation accuracy (panel A), phase estimate variance (panel B), phase estimate bias (panel C) and phase estimate absolute bias (panel D) and described by equations 12, 14 and 13 as observed on the test segments of the data with the optimal set of parameters identified over the independent training segments. For each such metric, the curves display their mean values averaged over 10 subjects as a function of the incurred delay. The error bars indicate the 95% confidence intervals obtained with 1,000 bootstrap iterations.

As expected and according to Figure 3.A, the envelope estimation accuracy quantified by the correlation coefficient r_a deteriorates as the processing lag D decreases. The smaller the processing lag, the weaker is the correlation between the non-causally obtained ground-truth envelope and the envelope estimated in real-time by each of the methods. Noticeably, for all positive delays *rect* approach (gray line) has the worst performance that makes it practically unusable for latency values below 150 ms. Methods *cfir* (blue line), *wcfir* (red line) and *acfir* (light blue line) exhibit comparable accuracy. The approach described in [36] (*ffiltar*, black dot) that entails AR modelling of data and forward-backward filtering of the extended data chunk yields at zero latency the accuracy comparable to that delivered by the proposed *cfir* family of methods but requires specification of parameters describing AR model order, filter and extension window length and is significantly more intense computationally.

Panels B, C and D of Figure 3 show the results for real-time estimation of instantaneous phase as a function of time-lag D . For the positive delay values, the bias b_ϕ remains practically negligible and does not exceed 5° for all

considered methods. Noteworthy, the *ffiltar* approach has been formulated for zero-latency and its phase estimation bias corresponding to $D = 0$ hovers around zero. However, as panel D demonstrates, the absolute bias of the phase estimate obtained by this technique is comparable to that delivered by the *cfir* family of methods and therefore close to zero values of the bias reflects symmetric around zero distribution of the bias observed in the 10 datasets explored. The methods from the *cfir* family exhibit a significant growth of phase estimation standard deviation (SD) as lag D decreases. Yet, at zero lag, the variance of the estimate delivered by the proposed here family of techniques appears to be significantly below that of the state of the art method introduced earlier in [36] and successfully used in the number of closed-loop stimulation studies. Here we reported the averaged observations made on the basis of the data recorded from 10 subjects as described in Section 2.3. The comparative performance of the explored methods may line-up differently for each of the datasets depending on the individual alpha-range SNR and other unaccounted factors.

This issue is clarified in Figure 4 that illustrates for individual subjects the same three performance measures as shown in Figure 3 but only for $D = 0$. Separate consideration for each subject allowed us to explore the performance as a function of SNR, which varied across individuals.

Figure 3: Four performance metrics vs incurred processing delay. A - dependence of the correlation coefficient r_a on delay D for the *cfir* family, for rectification based technique and AR-based extrapolation approach(*ffiltar*). B - phase estimation bias b_ϕ as a function of processing delay D . C - phase estimation variance σ_ϕ as a function of processing delay. D - phase estimation absolute bias $|b_\phi|$ vs. D

Figure 4 shows how the performance of the explored methods for $D = 0$ as a function of the SNR observed in each of the individual subjects. Metrics of the performance are the same as in Figure 3. For all methods envelope correlation r_a grows with SNR, panel A. Phase estimate standard deviation improves with SNR and appears to be consistently lower for the *cfir* family of methods. Phase estimate bias b_ϕ shows positive skew for the low SNR value. The absolute bias for *ffiltar* technique appears to be independent of the SNR. For high SNR values starting with 1.0 the absolute bias value obtained by all methods from the *cfir* family appears to be consistently lower than that of the *ffiltar*.

Figure 4: Four performance metrics vs. alpha rhythm SNR. A - envelope correlation coefficient r_a , B - phase estimation standard deviation (SD) σ_ϕ and C,D - phase estimation bias b_ϕ and absolute bias $|b_\phi|$. Each dot corresponds to a dataset and is positioned on the x-axis according to the P4 alpha-rhythm SNR.

In some versions of the closed-loop paradigms e.g. [24], a full-blown envelope reconstruction is not required. Instead, of interest is the discrete detection of time moments with high instantaneous band power. The transition into the zone of high instantaneous band power values may serve as a feedback or a trigger for either stimulus presentation or direct brain stimulation act. Suppose we want to perform detection of the time when the instantaneous rhythm power exceeds the 95% threshold. As shown in the left panel of Figure 5.A for the binary classification case the moments when the envelope falls into the top area above the dashed line are labeled as *High* and the rest of the moments are labeled as *Low*. The graph on the right panel of Figure 5.A shows balanced accuracy score (class recall average) for such binary detection task as a function of allowed processing delay parameter D . The analysis is done for one subject with median SNR selected from the pull of 10 subjects. We can observe that the best performance in the binary classification task is

achieved by the weighted *cfir* method with zero processing latency ($D = 0$). Similar results for ternary classification of the three-state problem are shown in Figure 5.B. Just like in the binary case, the moments when the envelope falls into the top area corresponding to the 5% of the largest envelope values are labeled as *High* and additionally, label *Low* is assigned to the time instances when the envelope takes on values from the lowest 5%. The rest of the time moments are labeled as *Medium*. Interestingly, the *cfir* family of methods delivers the best performance for zero processing delay $D = 0$. In the binary classification scenario, we can achieve about 75% of balanced accuracy. The rectification-based approach at best provides just above 60% accuracy and that peaks in 100-150 ms processing lag range. The results are qualitatively the same for the ternary classification case.

A. High/Other

B. High/Middle/Low

Figure 5: Discrete paradigm accuracy for one subject with median SNR. A) - binary classification task. The goal is to detect the time instances when alpha envelope is in the upper 5% quantile of its values. B) - ternary classification task to distinguish lower and upper 5% quantiles of the envelope values from the mid-range values falling into 5%-85% range.

Finally, to explore the morphology of the alpha-burst events in the *High/Other* classification task described above in Figure 6, we averaged the ground truth envelope around moments when the decoder crossed its own threshold. This computation was performed for *rect* and *wcfir* approaches for predefined delay parameters from [300, 100, 0, -100] ms set (for *rect* only positive values were used). Also, we computed averaged envelope across a set of randomly picked

time moments (denoted as *rand*) and across moments when the ground truth envelope crossed the *High* threshold (denoted as *ideal*), which can not be done causally, see Figure 6.d

Figure 6: Average envelope computed using real-time detected markers of *High* events for different pre-specified processing delay values.

4 Discussion

The standard techniques for estimating instantaneous power of EEG rhythms, such as the methods based on the rectification of narrow-band filtered signal and the STFT based algorithms, incur significant delays, hindering the performance of BRC systems. Such delays, combined with the lags of acquisition hardware and the time required for stimulus presentation, result in significant lags between the actual brain activity and the signal used to control the experimental flow in BRC paradigms and/or utilized as a feedback to the subject. In the closed-loop stimulation paradigms, this would mean that the timing of the stimulating pulse can not be accurately aligned to the desired feature of the oscillatory brain activity. In NFB setting or settings requiring an explicit feedback signal that reflects subject's performance, these standard approaches close the loop more than 300 ms past the targeted neural event [42]. Such delays may be especially harmful when the targeted brain rhythm patterns can be described as discrete events of a limited duration [7], [43],[44], where the feedback can arrive after the event has completed. Such low temporal specificity of the feedback signal hinders learning, especially in the automatic learning scenarios [30].

Here we systematically explored a series of methods for minimizing latency in RBC systems. We distinguished a family of best-performing techniques that are based on the least-squares filter design. These methods allow for a simpler and more transparent control over the accuracy-versus-latency trade-off compared to the other existing approaches.

Our results confirm that the proposed methodology based on least-squares filter design noticeably reduces latency of EEG envelope and phase estimation. This procedure is simple; yet its performance pars or exceeds that of the more complex approaches, such as the one based on the use of the AR-model [36], the most ubiquitous method for closed-loop studies [45], [18],[46],[47]. With our method, users can specify the desired delay and achieve the best possible envelope estimation accuracy possible with a linear method. As evident from Figure 3, the spectral density weighted cFIR technique allowed us to generate zero-latency feedback that accurately tracked the instantaneous power profile of the EEG-rhythm. The performance of our method on a typical segment of data can be appreciated from Figure 7 that shows true values and estimates of the envelope and phase obtained with cFIR method for various user specified lags.

We see this work as a systematic effort aimed at building a zero- or even negative-latency feedback systems that will allow transferring the predictive control methodology successfully exercised in technical systems to the tasks where the brain is the controlled object [48]. As shown in 5, *wcFIR* approach allows for correct forward prediction 100 ms ahead of rhythmic activity bursts, with AUC exceeding 70%. This illustration of the successful predictive behavior suggests that the proposed family of simple approaches together with the necessary hardware optimization will open

Figure 7: Envelope and phase estimates obtained by cFIR method (blue line) for different delay values D , each row corresponds to delays from -50ms to 200ms. Red line denotes ground-truth signal calculated by the ideal non-causal filter, gray dashed line denotes the same ground-truth signal but shifted by the corresponding delay D to facilitate comparison

up a way for the implementation of predictive control that enables a more efficient interaction with the functioning brain.

While our method reduces feedback latency, this reduction comes at a cost of less accurate envelope estimation. Deterioration of performance is especially sizeable in when the SNR is low and therefore, for the latency-reduction algorithm to be efficient, care should be taken to improve the SNR with such methods as spatial filtering of multi-electrode recordings.

The optimal latency-accuracy trade-off is the issue that needs to be addressed for each particular application and each particular subject. As shown in Figure 3, the methods outlined in this work allow the users to smoothly control this trade-off and choose the optimal operational point for each specific application.

As mentioned above, to achieve the true predictive scenario, though, the improvements need to be made not only of the signal processing algorithms but also of the hardware employed for signal acquisition, as well as the low-level software that handles EEG data transfer from the acquisition device to the computer memory buffer. To this end, it is worth considering specialized systems based on the FPGA programmable devices that eliminate the uncertain processing delays present in computer operating systems not designed to operate in real-time.

In the context of neurofeedback, additional consideration should be given to the physiological aspects of the sensory modality used to deliver the feedback signal. For instance, it is known that visual inputs, although very informative [33], are processed slower compared to tactile inputs and therefore tactile feedback could be a better option for predictive control.

The signal processing approaches presented here could be advanced by employing more sophisticated decision rules capable of extracting the hidden structure from the data. Thus, convolutional neural networks [49],[50] and novel recursive architectures hold a significant promise to further improve the accuracy of real-time zero-lag envelope and phase estimation.

Acknowledgments

This work is supported by the Center for Bioelectric Interfaces NRU HSE, RF Government grant, ag. No.14.641.31.0003. We also thank Dr. Maria Volodina for the careful proofread of the final version of this manuscript.

References

- [1] Buzsáki G. *Rhythms of The Brain*. Oxford Univ. Press, 2009.
- [2] Paul L. Nunez and Ramesh Srinivasan. *Electric Fields of the Brain: The neurophysics of EEG*. Oxford University Press, January 2006.
- [3] Müller-Dahlhaus F Ziemann U. Zrenner C, Belardinelli P. Closed-loop neuroscience and non-invasive brain stimulation: A tale of two loops. *Front Cell Neurosci.*, 2016.
- [4] Sitaram R., Ros T., Stoeckel L., Haller S., Scharnowski F., Lewis-Peacock J., Weiskopf N., Blefari M., Rana M., Oblak E., Birbaumer N., and Sulzer J. Closed-loop brain training: the science of neurofeedback. *Nature Reviews Neuroscience*, 2016.
- [5] Stermann M., MacDonald L., and Stone R. K. Operant control of the EEG alpha rhythm and some of its reported effects on consciousness. *Altered states of consciousness*, 1969.
- [6] Stermann M., MacDonald L., and Stone R. K. Biofeedback training of the sensorimotor electroencephalogram rhythm in man: effects on epilepsy. *Epilepsia*, 1974.
- [7] Ossadtchi A., Shamaeva T., Okorokova E., Moiseeva V., and Lebedev M. Neurofeedback learning modifies the incidence rate of alpha spindles, but not their duration and amplitude. *Scientific Reports*, 2017.
- [8] Lofthouse N., Hendren R., Hurt E., Arnold L. E., and E. Butter. A review of complementary and alternative treatments for autism spectrum disorders. *Autism research and treatment*, 2012.
- [9] Coben R., Linden M., and Myers T.E. Neurofeedback for autistic spectrum disorder: a review of the literature. *Applied Psychophysiology and Biofeedback*, 2010.
- [10] Zoefel B., Huster R. J., and Herrmann C. S. Neurofeedback training of the upper alpha frequency band in EEG improves cognitive performance. *Neuroimage*, 2011.
- [11] Wolpaw J. et al. *Brain-computer interfaces: principles and practice*. Oxford Univ. Press, 2012.
- [12] Frolov A., Kozlovskaya I., Biryukova E., and Bobrov P. Robotic devices in poststroke rehabilitation. *Zhurnal Vysshei Nervnoi Deyatelnosti Imeni I.P. Pavlova*, 2017.
- [13] Ang K. K., Chua K. S., Phua K. S., Wang C., Chin Z. Y., Kuah C. W., Low W., and Guan C. A. Randomized controlled trial of EEG-based motor imagery brain-computer interface robotic rehabilitation for stroke. *Clinical EEG and neuroscience*, 2015.
- [14] Moritz Grosse-Wentrup, Donatella Mattia, and Karim Oweiss. Using brain-computer interfaces to induce neural plasticity and restore function. *Journal of Neural Engineering*, 8(2):025004, mar 2011.
- [15] Ahmed El Hady. *Closed Loop Neuroscience*. Elsevier, 2016.
- [16] Chao-Hung Kuo, Gabrielle A. White-Dzuro, and Andrew L. Ko. Approaches to closed-loop deep brain stimulation for movement disorders. *Neurosurgical Focus*, 45(2):E2, aug 2018.
- [17] Til Ole Bergmann, Anke Karabanov, Gesa Hartwigsen, Axel Thielscher, and Hartwig Roman Siebner. Combining non-invasive transcranial brain stimulation with neuroimaging and electrophysiology: Current approaches and future perspectives. *NeuroImage*, 140:4–19, oct 2016.
- [18] Schaworonkow N. and Zrenner C. Triesch J, Ziemann U. EEG-triggered TMS reveals stronger brain state-dependent modulation of motor evoked potentials at weaker stimulation intensities. *Brain Stimul.*, 12(1):110–118, 2019.
- [19] Caroline Lustenberger, Michael R. Boyle, Sankaraleengam Alagapan, Juliann M. Mellin, Bradley V. Vaughn, and Flavio Fröhlich. Feedback-controlled transcranial alternating current stimulation reveals a functional role of sleep spindles in motor memory consolidation. *Current Biology*, 26(16):2127–2136, aug 2016.
- [20] Gan Huang, Jia Liu, Linling Li, Li Zhang, Yixuan Zeng, Lijie Ren, Shiqing Ye, and Zhiguo Zhang. A novel training-free externally-regulated neurofeedback (ER-NF) system using phase-guided visual stimulation for alpha modulation. *NeuroImage*, 189:688–699, April 2019.

- 449 [21] M. S. Nokia and J. Wikgren. Effects of hippocampal state-contingent trial presentation on hippocampus-
450 dependent nonspatial classical conditioning and extinction. *Journal of Neuroscience*, 34(17):6003–6010, April
451 2014.
- 452 [22] Ali Mazaheri, Martine R. van Schouwenburg, Andrew Dimitrijevic, Damiaan Denys, Roshan Cools, and Ole
453 Jensen. Region-specific modulations in oscillatory alpha activity serve to facilitate processing in the visual and
454 auditory modalities. *NeuroImage*, 87:356–362, February 2014.
- 455 [23] Johanna M. Zumer, René Scheeringa, Jan-Mathijs Schoffelen, David G. Norris, and Ole Jensen. Occipital al-
456 pha activity during stimulus processing gates the information flow to object-selective cortex. *PLoS Biology*,
457 12(10):e1001965, October 2014.
- 458 [24] Jason Samaha, Luca Iemi, and Bradley R. Postle. Prestimulus alpha-band power biases visual discrimination
459 confidence, but not accuracy. *Consciousness and Cognition*, 54:47–55, September 2017.
- 460 [25] Mathis Kaiser, Daniel Senkowski, Niko A. Busch, Johanna Balz, and Julian Keil. Single trial prestimulus oscil-
461 lations predict perception of the sound-induced flash illusion. *Scientific Reports*, 9(1), April 2019.
- 462 [26] Simon Hanslmayr, Alp Aslan, Tobias Staudigl, Wolfgang Klimesch, Christoph S. Herrmann, and Karl-Heinz
463 Bäuml. Prestimulus oscillations predict visual perception performance between and within subjects. *NeuroImage*,
464 37(4):1465–1473, October 2007.
- 465 [27] Rahmandad H., Repenning N., and Sterman J. Effects of feedback delay on learning. *System Dynamics Review*,
466 2009.
- 467 [28] Freek van Ede, Andrew J. Quinn, Mark W. Woolrich, and Anna C. Nobre. Neural oscillations: Sustained rhythms
468 or transient burst-events? *Trends in Neurosciences*, 41(7):415–417, July 2018.
- 469 [29] Mark Lawrence Andermann, Jaakko Kauramäki, Tapio Palomäki, Christopher I. Moore, Riitta Hari, Iiro P.
470 Jääskeläinen, and Mikko Sams. Brain state-triggered stimulus delivery: An efficient tool for probing ongoing
471 brain activity. *Open journal of neuroscience*, 2, 2012.
- 472 [30] Lacroix J.M. *Mechanisms of Biofeedback Control*. Springer, Boston, MA, USA, 1986.
- 473 [31] Grice G. R. The relation of secondary reinforcement to delayed reward in visual discrimination learning. *Journal*
474 *of experimental psychology*, 1948.
- 475 [32] S Kitazawa, T Kohno, and T Uka. Effects of delayed visual information on the rate and amount of prism
476 adaptation in the human. *The Journal of Neuroscience*, 15(11):7644–7652, November 1995.
- 477 [33] Evans N., Gale S., Schurger A., and Blanke O. Visual feedback dominates the sense of agency for brain-machine
478 actions. *PLoS ONE*, 2015.
- 479 [34] Oblak E.F., Lewis-Peacock J.A., and Sulzer J.S. Self-regulation strategy, feedback timing and hemodynamic
480 properties modulate learning in a simulated fMRI neurofeedback environment. *PLoS Comput Biol*, 2017.
- 481 [35] Oppenheim A.V. and Schaffer R.W. *Discrete-Time Signal Processing. 3rd Edition*. Prentice Hall, 2010.
- 482 [36] Chen L.L., Madhavan R., Rapoport B.I., and Anderson W.S. Real-time brain oscillation detection and phase-
483 locked stimulation using autoregressive spectral estimation and time-series forward prediction. *IEEE Trans*
484 *Biomed Eng*, 60:753–762, 2013.
- 485 [37] Andreas Bruns. Fourier-, Hilbert- and wavelet-based signal analysis: are they really different approaches? *Jour-*
486 *nal of Neuroscience Methods*, 137(2):321–332, August 2004.
- 487 [38] Bradley Voytek, Andrew S Kayser, David Badre, David Fegen, Edward F Chang, Nathan E Crone, Josef Parvizi,
488 Robert T Knight, and Mark D’Esposito. Oscillatory dynamics coordinating human frontal networks in support
489 of goal maintenance. *Nature Neuroscience*, 18(9):1318–1324, July 2015.
- 490 [39] Wei Rong Lee, Lou Caccetta, Kok Lay Teo, and Volker Rehbock. Optimal design of complex FIR filters with
491 arbitrary magnitude and group delay responses. *IEEE Transactions on Signal Processing*, 2006.
- 492 [40] Simon Haykin. *Adaptive Filter Theory*. Prentice Hall, 2002.

- 493 [41] Alexei Ossadtchi, Platon Pronko, Sylvain Baillet, Mark E. Pflieger, and Tatiana Stroganova. Mutual informa-
494 tion spectrum for selection of event-related spatial components. application to eloquent motor cortex mapping.
495 *Frontiers in Neuroinformatics*, 7, 2014.
- 496 [42] Leslie H. Sherlin, Martijn Arns, Joel Lubar, Hartmut Heinrich, Cynthia Kerson, Ute Strehl, and M. Barry Ster-
497 man. Neurofeedback and basic learning theory: Implications for research and practice. *Journal of Neurotherapy*,
498 15(4):292–304, October 2011.
- 499 [43] Hyeyoung Shin, Robert Law, Shawn Tsutsui, Christopher I Moore, and Stephanie R Jones. The rate of transient
500 beta frequency events predicts behavior across tasks and species. *Elife*, 6:e29086, 2017.
- 501 [44] Maxwell A. Sherman, Shane Lee, Robert Law, Saskia Haegens, Catherine A. Thorn, Matti S. Hämäläinen,
502 Christopher I. Moore, and Stephanie R. Jones. Neural mechanisms of transient neocortical beta rhythms: Con-
503 verging evidence from humans, computational modeling, monkeys, and mice. *Proceedings of the National*
504 *Academy of Sciences*, 113(33):E4885–E4894, 2016.
- 505 [45] Christoph Zrenner, Debora Desideri, Paolo Belardinelli, and Ulf Ziemann. Real-time EEG-defined excitability
506 states determine efficacy of TMS-induced plasticity in human motor cortex. *Brain Stimulation*, 11(2):374–389,
507 March 2018.
- 508 [46] Natalie Schaworonkow, Pedro Caldana Gordon, Paolo Belardinelli, Ulf Ziemann, Til Ole Bergmann, and
509 Christoph Zrenner. μ -rhythm extracted with personalized eeg filters correlates with corticospinal excitability
510 in real-time phase-triggered eeg-tms. *Frontiers in Neuroscience*, 12:954, 2018.
- 511 [47] Jianguang Ni, Thomas Wunderle, Christopher Murphy Lewis, Robert Desimone, Ilka Diester, and Pascal Fries.
512 Gamma-rhythmic gain modulation. *Neuron*, 92(1):240–251, October 2016.
- 513 [48] Yang-Yu Liu and Albert-László Barabási. Control principles of complex systems. *Rev. Mod. Phys.*, 88, Sep
514 2016.
- 515 [49] Schirrmester R.T., Springenberg J.T., Fiederer L., Glasstetter M., Eggenberger K., Tangermann M., Hutter F.,
516 Burgard W., and Ball T. SelDeep learning with convolutional neural networks for EEG decoding and visualiza-
517 tion. *Hum Brain Mapp*, 2017.
- 518 [50] Halme HL Parkkonen L. Zubarev I, Zetter R. Adaptive neural network classifier for decoding meg signals.
519 *Neuroimage*, 11(197):425–434, August 2019.