Light Sheet Fluorescence Microscopy as a New Method for Unbiased Three-Dimensional Analysis of Vascular Injury

Nicholas E. Buglak^{1,2,3,4}, Jennifer Lucitti⁵, Pablo Ariel⁶, Sophie Maiocchi^{1,2,4}, Francis J. Miller Jr.^{5,7}, and Edward S. M. Bahnson^{*1,2,3,4,8}

¹Department of Surgery, Division of Vascular Surgery, University of North Carolina at Chapel Hill, NC 27599.

²Center for Nanotechnology in Drug Delivery, University of North Carolina at Chapel Hill, NC 27599.

³Curriculum in Toxicology & Environmental Medicine, University of North Carolina at Chapel Hill, NC 27599.

⁴McAllister Heart Institute, University of North Carolina at Chapel Hill, NC 27599

⁵Department of Medicine, Duke University, Durham, NC 27708.

⁶Microscopy Services Laboratory, Department of Pathology and Laboratory Medicine. University of North Carolina at Chapel Hill, NC 27599.

⁷Department of Medicine, Veterans Administration Medical Center, Durham NC 27705.

⁸Department of Cell Biology & Physiology. University of North Carolina at Chapel Hill, NC 27599.

*Corresponding Author:

Edward S. M. Bahnson

edward_bahnson@med.unc.edu

Tel: (919)-843-0842

Short Title: 3D Unbiased Analysis of Preclinical Vasculopathies

Keywords: Light sheet fluorescence microscopy, restenosis, neointimal hyperplasia, vessel remodeling, vascular disease

Abstract

Aims. Assessment of preclinical models of vascular disease are paramount in the successful translation of novel treatments. The results of these models have traditionally relied on 2-D histological methodologies. Light sheet fluorescence microscopy (LSFM) is an imaging platform that allows for 3-D visualization of whole organs and tissues. In this study, we describe an improved methodological approach utilizing LSFM for imaging of preclinical vascular injury models while minimizing analysis bias.

Methods and Results. The rat carotid artery segmental pressure-controlled balloon injury and mouse carotid artery ligation injury were performed. Arteries were harvested and processed for LSFM imaging and 3-D analysis, as well as for 2-D area histological analysis. Artery processing for LSFM imaging did not induce vessel shrinkage or expansion, and was reversible by rehydrating the artery, allowing for subsequent sectioning and histological staining *a posteriori*. By generating a volumetric visualization along the length of the arteries, LSFM imaging provided different analysis modalities including volumetric, area, and radial parameters. Thus, LSFM-imaged arteries provided more precise measurements compared to classic histological analysis. Furthermore, LSFM provided additional information as compared to 2-D analysis in demonstrating remodeling of the arterial media in regions of hyperplasia and periadventitial neovascularization around the ligated mouse artery.

Conclusions. LSFM provides a novel and robust 3-D imaging platform for visualizing and quantifying arterial injury in preclinical models. When compared with classic histology, LSFM outperformed traditional methods in precision and quantitative capabilities. LSFM allows for more comprehensive quantitation as compared to traditional histological methodologies, while minimizing user bias associated with area analysis of alternating, 2-D histological artery cross-sections.

49 **Translational Perspective**

50 A more reproducible and robust quantitation of vascular pathology in preclinical models is
 51 necessary to accelerate translational discovery. Current methodology to assess vascular disease
 52 has significant limitations. The methodology described herein employs a modern imaging
 53 modality, light sheet fluorescence microscopy (LSFM), to improve assessment of established
 54 preclinical vascular injury models. LSFM provides more comprehensive and precise analysis
 55 capabilities than classical histological approaches. Hence, LSFM applied to vascular research
 56 has the potential to drive new basic discoveries, and ultimately translation of novel therapies.

I. Introduction

Cardiovascular disease (CVD) is the leading cause of death and disability in the world^{1, 2}. Atherosclerosis is the major underlying cause of most CVD. Severe, symptomatic atherosclerosis is treated by percutaneous or surgical revascularization, and the long-term success of both approaches is limited by arterial restenosis³. Despite the advances in revascularization procedures, restenosis rates remain unacceptably high³. Hence, animal models are essential to understand the pathophysiology of restenosis and to discover and optimize new drugs to improve surgical outcomes. Restenosis occurs secondary to the development of neointimal hyperplasia and vessel remodeling⁴. Several preclinical injury models exist for studying restenosis⁵ including the rat carotid artery balloon injury model and the mouse carotid ligation model.

Classically, histological methods are used to quantify neointimal hyperplasia in preclinical models⁶. This classic approach involves sectioning the artery into two-dimensional (2-D) cross-sections and staining a selection of sections for analysis. A major limitation with this approach is that it assumes predictable and homogenous arterial injury for each animal and between animals. This suggests that a subset of microns-thick cross-sections will represent the lesion as a whole. However, arterial injury is not homogenous⁷, which is also observable using the methodology described herein. Furthermore, the pathology of neointimal hyperplasia is defined, but no consensus exists on how to properly quantify it. The literature varies between using three to ten cross-sections at variable regions throughout the artery for quantification⁸⁻²³, which has the potential to introduce bias based on which and how many slices are analyzed. Moreover, classic histology limits our ability to observe and measure vessel remodeling, the other major component of restenosis, because only a few cross-sections are used to inform all of the morphological changes in the artery.

Light sheet fluorescence microscopy (LSFM) allows fast imaging of large (several mms in diameter), fixed, transparent tissue, by illuminating samples with a sheet of laser light from the side, and collecting the resulting fluorescence from the excited plane at a 90° angle²⁴. This

technique generates optical sections of the sample, is significantly faster, and causes less photo damage compared with other commonly used optically sectioning techniques such as laser scanning confocal and multiphoton microscopy²⁵. This powerful tool has been used in a wide range of applications from imaging intracellular dynamics in a single cell, to developmental dynamics in entire organisms²⁶. Indeed, LSFM has unraveled new mechanisms for megakaryocyte function and platelet migration from the bone marrow^{27, 28}. Herein we describe a methodology using LSFM to image preclinical restenosis murine models in three-dimensions (3-D). We quantify changes in neointimal hyperplasia and arterial media volumes, alongside vessel remodeling, in response to arterial ligation or endothelial denudation. Additionally, we benchmark this novel methodology against classic histological analysis. To image the arteries, they must be made transparent by homogenizing their refractive index. Out of the many clearing methods available²⁹, we selected iDISCO+³⁰ for its excellent clearing, speed, low cost, and—critically—optimization for immunostaining in large tissue samples. We believe utilizing LSFM for analysis of preclinical restenosis models will improve rigor and reproducibility associated with these models and may identify novel mechanisms of this pathology. The overall aim of this manuscript is to provide a framework for analyzing preclinical restenosis models.

II. Methods

All rat procedures described here have been approved by the Institutional Animal Care and Use Committee (IACUC) of the University of North Carolina at Chapel Hill. All mouse procedures were approved by the Durham Veterans Affairs Medical Center (DVAMC) IACUC (2013-002). All procedures conform to the Guide for the Care and Use of Laboratory Animals published by the United States National Institutes of Health.

Reagents and Materials

Acetic Acid (ARK2183-1L, Sigma-Aldrich St. Louis, MO). Dibenzyl ether (DBE) (33630-1L, Sigma-Aldrich). Dichloromethane (DCM) (BDH1113-4LG, VWR, Radnor, PA). EDTA (BDH9232, VWR). Methanol (A4525K, Thermo-Fisher Scientific, Waltham, MA). Microcentrifuge

tubes (05-408-129, Thermo-Fisher Scientific). Paraformaldehyde (158127; Sigma-Aldrich). Phosphate buffered saline (PBS) (20-134; Apex BioResearch Products). Sodium azide (NaN₃) (S2002, Sigma-Aldrich). Tris Base (T8600, US Biological Sciences, Salem, MA). Tween-20 (P1379, Sigma-Aldrich).

Rat Carotid Artery Balloon Injury

The rat carotid balloon injury model was first described by Clowes et al⁴. Since then it has been extensively described³¹. Shears et al. first described the modified, pressure-controlled segmental injury used in this study³². Adult male rats (n = 7 for LSM, n = 10 for histology) aged 12-14 weeks were used for surgery. Rats were anesthetized with inhaled isoflurane (0.5–2%). Atropine (0.1 mg/kg) was administered subcutaneously (SC) to reduce airway secretions and Carprofen (5 mg/Kg) was administered SC for pain management at the time of surgery. After a sterile prep and midline neck incision, the left common, internal, and external carotid arteries were dissected and the internal and common carotid arteries were temporarily occluded using micro serrefine clamps (18055-05, Fine Science Tools, Foster City, CA). A No. 2 French Fogarty balloon catheter (Edwards Lifesciences, Irvine, CA) was inserted through an arteriotomy in the external carotid artery, and advanced into the common carotid artery. The balloon was inflated to 5 atmospheres of pressure for 5 min to create a uniform injury. After 5 minutes, the balloon was deflated and removed, the external carotid artery was ligated, and blood flow was restored. The incision site was sutured and rats were housed in a temporary cage under a heat lamp until consciousness was restored, then moved to their home cage. Rats were euthanized by anesthesia overdose followed by bilateral thoracotomy 2 weeks after injury.

Rat Carotid Artery Harvest for LSM

Both carotid arteries were harvested after in situ perfusion-fixation with 200 mL of 1X PBS followed by cold 4% paraformaldehyde. Arteries were harvested so that a portion of the bifurcation was included, as this helps with the sample orientation during imaging. Arteries were placed in

4% paraformaldehyde at 4°C overnight then moved to a 1.5 mL microcentrifuge tube with PBS + 0.02% NaN₃ for 24 hours.

Mouse Carotid Ligation

C57BL/6 mice were crossed with NOD mice (The Jackson Laboratory, Bar Harbor, ME) and used in this study. Mice (n = 3-6 for LSM, n = 12 for histology) aged 12-14 weeks were anesthetized with inhaled isoflurane (1%-3%). Neck fur was removed with a depilatory agent, the field prepared with a 10% povidone-iodine solution, an incision was made slightly left of midline, and the left carotid artery just proximal to the bifurcation was exposed. Novafil™ polybutester (7-0) was used to ligate the carotid artery and close the incision. Mice were administered acetaminophen in drinking water (35mg/ml) for 4 days. Mice were euthanized 4 weeks after surgery under continuous isoflurane anesthesia (1.5%-3%) by cardiac puncture exsanguination with a 23g needle.

Mouse Carotid Artery Harvest for LSM

Four weeks after ligation, carotid arteries were harvested after in situ perfusion-fixation with 2-4% paraformaldehyde (PFA) in PBS. Arteries were fixed in 2-4% PFA at 4°C overnight and then moved to PBS + 0.02% NaN₃ for 24 hours.

Carotid Artery Staining and Processing for LSM

Workflow is outlined in **Figure 1**. Arteries were blocked in IHC-Tek diluent overnight in a 37°C incubator. Rat arteries were probed with primary rabbit anti-CD31 (ab28364, Abcam) antibody diluted 1:500 in IHC-Tek diluent for 3 days at 37°C, washed four times over 24 hours in 1X PBS + 0.2% Tween-20 (PTwH wash buffer) at room temperature (RT), and probed with goat anti-rabbit Alexa Fluor 647 IgG secondary antibody (A-21245, Thermo-Fisher Scientific) diluted 1:500 in IHC-Tek diluent for 2 days at 37°C. Mouse arteries were stained following the same protocol with rabbit anti-CD31 (ab28364, Abcam) and probed with donkey anti-rabbit Alexa Fluor 568 IgG secondary antibody (A10042, Thermo-Fisher Scientific). For multiplex analysis mouse arteries were stained for CD31 in tandem with CD68 or VE-Cadherin. For the CD68 samples:

1:500 dilution of rabbit anti-CD31 and mouse anti-CD68 (MCA341GA, Bio-Rad, Hercules, CA) followed by a 1:500 dilution of Alexa Fluor 647 and goat anti-mouse Alexa Fluor 555 (A-21425, Thermo-Fisher Scientific) as described above. For the VE-Cadherin samples: 1:500 dilution of rabbit anti-CD31 and 1:300 dilution of goat anti-VE-Cadherin (AF-1002, R&D, Minneapolis, MN) followed by a 1:500 dilution of Alexa Fluor 568 and donkey anti-goat Alexa Fluor 633 (A-21082, Thermo-Fisher Scientific). Arteries were again washed in PTwH wash buffer over 24 hours at RT. After this wash, arteries were stored in PBS + 0.02% NaN₃ in the dark at 4°C up to three weeks.

Next arteries were embedded in 1% agarose (20-102GP, Apex Bioresarch Products, San Diego, CA) prepared in 1X TAE buffer. Agarose embedding was performed in a petri dish so that the final height of the agarose-artery block was 5 mm. To ensure success of the agarose embedding: 1) low-melt agarose should not be used as it is not compatible with iDISCO+; 2) prior to embedding, arteries should be warmed for over 1 hour in a 37°C incubator; 3) agarose temperature should be monitored using a thermometer and arteries should only be added once the agarose temperature is at 40°C to minimize heat damage. The final dimensions of the agarose-artery block that allow for proper sample mounting within the LSM are 5 mm x 5 mm x 10-15 mm (W x H x L).

Arteries were then dehydrated in methanol (20% MeOH → 40% → 60% → 80% → 100% X2) for 1 hour at RT per dilution and then kept overnight at RT. Arteries were cleared the next day in glass scintillation vials (986492, Wheaton, Millville, NJ) on a plate shaker at RT. Arteries were exposed to a 33% methanol - 66% DCM solution for 3 hours, then 100% DCM for 15 minutes followed by a second round of 100% DCM for 15 minutes. Lastly, arteries were cleared in 100% DBE overnight. Arteries can be stored in the DBE indefinitely until imaging.

Light Sheet Fluorescence Microscopy Imaging

Imaging was performed using a LaVision BioTec Ultramicroscope II³³ equipped with zoom body optics, an Andor Neo sCMOS camera, an Olympus MVPLAPO 2X/0.5 objective and a 5.7 mm working distance corrected dipping cap (total magnifications ranging from 1.3X to 12.6X with

variable zoom). Cleared agarose-artery blocks were mounted in a sample holder and submerged in a 100% DBE reservoir. Arteries were imaged at 1.6X mag (0.8X zoom), using the three light sheet configuration from a single side, with the horizontal focus centered in the middle of the field of view, an NA of 0.043 (beam waist at horizontal focus = 17 μm), and a light sheet width of 60-100% (adjusted depending on artery length to ensure even illumination in the Y axis). Pixel size was 3.9 μm and spacing of Z slices was 5 μm . Up to three channels were imaged per species: autofluorescence with 488 nm laser excitation and a Chroma ET525/50m emission filter was used for both species; mouse CD31-AF568 or CD68-AF555 stained channel imaged using 561 nm laser excitation and a Chroma ET600/50m emission filter; rat CD31-AF647 or mouse VE-cadherin-AF633 stained channel imaged using 647 nm laser excitation and a Chroma ET690/50m emission filter. The chromatic correction module on the instrument was used to ensure both channels were in focus. To reduce bleaching, while setting up the parameters for image acquisition the sample was viewed with low laser power and long exposures; this was inverted when acquiring data to maximize acquisition speed.

Analysis Terminology

Internal elastic lamina (IEL) corresponds to the inner-most lamina, facing the lumen. External elastic lamina (EEL) corresponds to the outer-most lamina. As described below, an analysis tool is used to obtain the radius from the artery center to the edge of the lumen, touching the intima, or to either lamina. The change in radius (Δr) corresponds to the difference between the radius to the IEL surface edge and the radius to the lumen surface edge. This Δr is used to identify the region of hyperplasia through an automated, unbiased approach.

Image Rendering and Analysis

Workflow is outlined in **Figure 2**. LSMF-saved TIFF files (ImInspector Pro v5.1.350) were converted to an Imaris (Bitplane, Oxford Instruments) file using Imaris File Converter software (v9.3.1). Acquired LSMF images were first analyzed using the two-dimensional slice function and snapshots were obtained of cross-sections in the x-y, x-z, and y-z axis. The file was cropped to

only include the artery to reduce file size and expedite all further steps. The Surfaces tool created three separate surfaces representing the artery lumen, IEL, and EEL. Each surface was created by semi-automatically tracing the appropriate landmark using the 647- or 561-channel for the CD31-stained intima, which would represent the lumen volume, and the 488-channel for both elastic laminae. Fluorescence captured in the 488-channel from the IEL was used as the spatial landmark for generating the IEL surface, while fluorescence from the EEL in the 488-channel was used for the EEL surface. For the rat model, contours were created using the isoline mode at 5% density along the artery every 15 μm in the x-z plane starting at the bifurcation until the end of the sample. For the mouse model, 15-20% density was used with traces every 10 μm . For some optical slices the semi-automatic tracing failed to identify the landmark. In these cases, a manual trace was performed. The starting slice position at the bifurcation and final slice position were noted for the first rendered surface, to ensure that all three rendered surfaces visualized the same arterial section. This was first performed on the healthy artery for each animal, and then on the injured artery.

Using the IEL surface, a filament was generated through the center of the artery. A Distance Transformation was performed inside the IEL surface to generate a heat-map along the artery from the artery center to the surface edge. This heat-map is presented as a new channel whose intensity corresponds to the distance in μm , from the surface edge. This was replicated for the other surfaces, and channels were named "Distance to 'respective surface'". Next, the Filament tool, using the Line style with auto-center and auto-depth activated, was used on the IEL heat-map channel to generate a filament through the center of the heat-map, corresponding to the artery center. Then the generated filament was converted using a Filament to Spots XTension (provided by Matthew Gastinger, Bitplane), which automatically placed an individual spot roughly every 4 μm along the artery. Using these Spots, the radius along the artery from the center of the artery to the surface edge of the lumen, IEL, and EEL was obtained. The average intensity at each spot (which corresponds to the radius, in μm) and the spot position were exported to

Microsoft Excel. Cumulative distance along the filament (i.e. the center of the artery) was calculated using a Pythagorean formula for Euclidean distance between adjacent spots. Up to 50 spots from the lateral edges of the filament were removed because they measured the distance to the lateral region of each surface, which is inconsistent with the rest of the spot measurements along the filament. A plot was generated of the radius to each surface along the artery. The change in radius (Δr = radius to IEL surface edge – radius to lumen surface edge) was obtained to identify regions of hyperplasia. To control for potential sample processing artifacts and to provide baseline parameters, we used the contralateral, uninjured carotid artery as a reference per animal. In the healthy artery the intima is a single monolayer of endothelial cells, thus the Δr centers close to zero. Using the average Δr for the uninjured artery, we estimate the 99% confidence interval (CI) and define any Δr above this threshold as indicative of neointimal hyperplasia in the animal's contralateral injured artery. In the injured artery, the spot positions above the 99% CI were noted. Using the identified starting and ending spot positions, all three surfaces were cropped in Imaris, which generated three new surfaces exclusively in the region with neointimal hyperplasia. Using the newly cropped surfaces, the volume of each respective surface was exported to Excel and the intima-to-media (I:M) ratio ((IEL volume - lumen volume) / (EEL volume – IEL volume)) was obtained.

To visualize the hyperplasia and media (as shown in the right panel of **Figure 2** or in **Figure 3F**) we generated new surfaces using a combination of the Imaris Surface tool and masking of the distance transformation channels. We defined the media volume as the space between the IEL and EEL surfaces, and the hyperplasia as the space between the lumen and the IEL surface.

Carotid Artery Histological Processing and Histology

Rat or mouse carotid arteries were harvested after in situ perfusion-fixation with PBS and cold 2% paraformaldehyde. Arteries were placed in 2% paraformaldehyde for 1h at 4°C followed by 30% sucrose overnight at 4°C. Arteries were snap-frozen in O.C.T. (4583; Tissue-Tek,

Torrance, CA) and stored at -80°C. 5 µm sections were cut throughout the entire common carotid artery for staining.

After imaging for LSM, the agarose-embedded artery was rehydrated in PBS + 0.02% NaN₃ for 3 days in a 40°C water bath. Artery was manually extracted and soaked overnight in 30% sucrose followed by cryosectioning and staining.

Arteries were hematoxylin & eosin (H&E) stained for morphological analysis and the LSM-processed, rehydrated artery cross-sections were also stained for immunofluorescence. Arterial cross-sections on room temperature slides were fixed with ice-cold acetone for 5 min then permeabilized with 0.3% Triton X-100 (X100; Sigma-Aldrich) for probing with 1:100 primary anti-alpha-smooth muscle actin (α-SMA) antibody (Rabbit, ab5694, Abcam, Cambridge, UK) diluted in IHC-Tek diluent (1W-1000; IHC World, Woodstock, MD) followed by 1:500 dilution of goat anti-rabbit Alexa Fluor 555 IgG secondary antibody (A-21429, Thermo-Fisher Scientific) and 1:500 DAPI diluted in PBS. Slides were rinsed in distilled water and coverslips were attached using ProLong Gold antifade reagent (P36930, Invitrogen). Slides were imaged on the day after staining using a Zeiss Axio Imager.A2 microscope (Oberkochen, Germany). Samples were illuminated by the X-Cite 120 LED Boost (Excelitas Technologies, Waltham, MA) laser with DAPI (385 nm), GFP (470 nm) or Cy3 (560 nm) excitation filters. Images were taken on a 5X objective (Zeiss Plan Apochromat 5x.0,16 440620 (2)) for quantification of rat arteries or a 10X objective (Zeiss Plan Neofluar 10x/0.30 440330) for mouse arteries. 20X objective (Zeiss Plan Apochromat 20x/0.8 ∞/0.17) images of the rat arteries were included for supplemental figures. Images were captured using an AxioCam HRm (Carl Zeiss MicroImaging GmbH, 426511-9901-000 10-33 VDC 5W) camera.

Histological and Optical Slice Analysis

All images were quantified using ImageJ software (<https://imagej.nih.gov/ij/>). Intima-to-media (I:M) ratio per animal was determined by dividing the area of neointimal hyperplasia (IEL-lumen area) by the medial area (EEL-IEL area) per optical or histological slice per animal. 6-10

discontinuous arterial cross-sections were analyzed for the injured rat carotid arteries comparison to LSM volume and optical slice I:M ratio. 3-4 discontinuous arterial cross-sections were analyzed for the healthy rat or mouse carotid artery arterial shrinkage analysis.

Virtual snapshots for the mouse or rat were taken at a constant zoom for each animal per species. Up to 10 virtual snapshots of the injured rat artery in the x-z axis were taken every 300 μm for analysis around the peak slice of injury. Area and perimeter values were obtained of the external elastic lamina, internal elastic lamina (IEL), and lumen. Up to 10 virtual snapshots of mouse arteries were analyzed every 100-200 μm along the carotid artery. Then 3-10 slices, as defined in the figures, were averaged per animal.

Statistics

For all multiple group analyses, a one-way ANOVA with Tukey's correction was performed in Origin software. Two group analysis was performed using a student's t-test, and paired-group t-test when appropriate, in Origin software. One group comparison to a reference value was analyzed with a one sample t-test. Precision was assessed using the coefficient of variation (CV%) for the 3-D volume to 2-D areal analysis. Rat arterial remodeling was analyzed with SAS statistical software (SAS Institute). The partial correlation coefficient (ρ) between neointimal thickness and vessel radius was estimated using a mixed model³⁴, controlling for subject variation and in repeated measures along the length of the artery. The confidence interval and p-value for the mixed model correlation estimator was determined with a normal approximation using the delta method³⁵. SAS Code provided in the online supplement.

III. Results

The aim of this manuscript is to validate LSM for analysis of preclinical restenosis models. The combination of iDISCO+ with LSM is ideally suited to image injured arteries, generating datasets to measure multiple aspects of the arterial injury response efficiently while minimizing bias. This technique generates optical sections of the transparent sample that can then be processed for analysis (**Videos 1-3** –see legend for hyperlink to videos–). A comparison

of the workflow for histological and LSM processing is provided in **Figure 1**, while the analysis workflow utilizing Bitplane Imaris software is depicted in **Figure 2**.

Imaging of injured murine arteries. First, we set out to image and analyze the rat carotid balloon injury model. LSM imaging of injured and contralateral uninjured arteries clearly allows for identification of the three layers of the arterial anatomy: intima, media (including the individual lamellae) and adventitia (**Fig 3 B-E**). A close-up volume rendering and x-y plane optical slice of the healthy (**Fig. 3 B**) and injured (**Fig. 3 C**) rat arteries shows a clear CD31⁺ endothelial layer and clear morphological differences on the luminal side between the uninjured and injured arteries. To identify the area of neointimal hyperplasia we present two longitudinal views of the rat injured artery from the bifurcation through 5 mm along the common carotid (**Fig. 3 D**). A surface rendering of the hyperplastic region is presented, delineated by the red box (**Fig. 3 D**). This rendering is performed for the hyperplastic region of all arteries to measure vessel stenosis, and an overlay of this rendering onto the whole artery is shown in **Video 2** (<https://youtu.be/LoAlApmV22o>). To compare the 2D images obtained through LSM with those obtained through H&E staining and histological imaging, **Figure 3 E** presents a transverse (x-z) cross-section of the artery in **Figure 3 D**, side-by-side with an H&E-stained cross-section of the same sample after it was rehydrated, cryo-sectioned, and stained. **Figure 3 F** depicts a surface rendering of the hyperplastic lesion and the media of the injured rat carotid artery, the volumes of these regions are used to analyze stenosis through volumetric I:M ratio. To show the intra- and inter-subject variability of the model, representative slices in the longitudinal x-y and y-z planes and determination of the region of hyperplasia of each rat artery used for analysis are shown in **Sup. Figure 1**.

Even though the rat carotid balloon injury is a commonly used model, the availability of many genetically modified mouse strains makes mouse models of arterial injury very important tools in vascular biology. Hence, it was of interest to determine if our methodological approach is suitable for arteries that are ten times smaller than those of the rat. Therefore, we assessed the

ability of our imaging and analysis methodology in the mouse carotid ligation model (**Fig. 3 G-J**). We compare the results obtained through LSFM with those obtained by histological imaging of H&E-stained sections. We show representative H&E cross-sections from three ligated mouse arteries processed exclusively for histology (**Fig. 3 G**) next to longitudinal (x-y) and transverse (x-z) slice views of a representative ligated mouse artery imaged through LSFM (**Fig. 3 H**). Notably, the transverse cross-sections corresponding to the different regions of the artery highlight the variability in neointimal hyperplasia development within a single artery. The intra- and inter-subject variability can be observed in **Sup. Fig. 2**, which shows representative slices of each mouse artery used for analysis. Specific to the mouse ligation model, we observed a CD31⁺ periadventitial neovascular plexus around the ligation and along the injured artery (**Fig. 3 I**). This CD31⁺ plexus was not present in the contralateral uninjured artery (**Fig. 3 J**). This neovascularization was consistent for all ligated arteries (**Sup. Fig. 2**), including arteries with low (<0.1) I:M ratios, and presents a phenomenon that has not been described in rodents by classic histological approaches.

Inflammation is a critical characteristic of the arterial injury response, and macrophages are known to be present at the site of injury. Moreover, inflammatory macrophages can drive angiogenesis³⁶. Therefore, it was of interest to assess the presence of macrophages in injured arteries while assessing the multiplexing capabilities of our methodology. To do this, some arteries were stained for CD68, a commonly used macrophage marker, together with CD31, prior to LSFM imaging. **Figure 4** shows presence of the neovascular plexus and macrophages in the adventitia of the injured carotids evidenced by the CD31⁺ staining and CD68⁺ puncta. To further assess the multiplexing capabilities of the technique, some uninjured arteries were stained for VE-cadherin, an endothelial cell adhesion marker, and CD31. **Sup. Figure 3** shows optical slices and optical volumes of uninjured mouse carotids showing the traditional VE-cadherin⁺ cobblestone structure of intact endothelium and the CD31⁺ cells. These experiments show that

multiplexing in mouse arteries is possible using this methodology, further expanding the capabilities of LSMF in preclinical arterial injury models.

Quantification of arterial stenosis. Next, we compared the quantitative capabilities of our methodology with those of traditional histology. We used the intima-to-media (I:M) ratio for both injury models to benchmark the LSMF volumetric analysis approach to classic histology. To do this, we compared the volumetric I:M ratio to the H&E I:M ratio of a separate cohort of age-matched animals. The results (**Fig. 5 A-B**) showed comparable average I:M ratios with no significant difference for both rat (LSMF = 0.77, H&E = 0.62) and mouse (LSMF = 0.79, H&E = 0.63), thus validating our methodology. However, the coefficient of variation (CV%) of the estimates was lower for the LSMF volumetric approach compared to traditional H&E histology (Rat = 28% vs 41%, Mouse = 79% vs 127%; **Sup. Table 1**) suggesting increased precision. Additionally, LSMF imaging allows for optical slicing of the artery as thin as 3.2 μm along the x-z axis. Thus, per animal model, we took equidistant optical slices for each injured artery and analyzed the slices using the classic 2-D area trace. We evaluated the utility of analyzing 3, 5, and 10 optical slices per animal, as these are within the commonly reported number of cross-sections analyzed in the literature (**Table 1**). We then compared these values to the volumetric approach and observed no significant difference between the volumetric and classical analyses (**Sup. Fig. 4**). However, for the rat, the more optical slices used, the more precise the I:M estimate, as illustrated by the CV% (**Sup. Table 1**). For the mouse ligation model, any slice number used for analysis had a CV% similar to the volume approach, and almost half of that for the H&E-analyzed samples (optical slice = 65-69% vs. H&E = 127%, **Sup. Table 1**). The lower precision of the estimates in the mouse carotid ligation model can be explained by the higher intra- and inter-subject variability of the model (**Sup. Fig. 1 and 2**) Our results suggest that analysis using volumetric renderings produces similar estimates to classical histological approaches with greater precision, while allowing for optical sectioning.

Importantly, the use of the Filament analysis tool allowed for unbiased identification of the hyperplastic region (**Fig. 2**). **Figure 5 C** shows a representative plot of the IEL radius overlaid with the change in radius (Δr = IEL radius – lumen radius). In the healthy artery the lumen and IEL radii are practically identical, thus the Δr centers close to zero. Meanwhile, in the injured artery there is a dramatic increase in Δr .

Quantification of remodeling. The ability to look at the whole artery led us to assess the quantification of vessel remodeling after arterial injury. Notably, the arterial filament allowed us to measure positive vessel remodeling within the hyperplastic region in the rat model. We compared the individual spots corresponding to the hyperplastic region along the filament to the non-hyperplastic region of the injured artery and to each rat's respective contralateral healthy artery. Our results (**Fig. 5 D-E**) show a significant increase in IEL radius within the hyperplastic region ($p=0.008$ to non-hyperplastic regions within the same vessel, $p=0.01$ to healthy). Furthermore, intimal thickness (T) correlated positively with vessel radius (R) ($p_{TR}=0.58$; 95% CI 0.38-0.80; $p<0.0001$) (**Sup. Table 2**) indicating compensatory remodeling in the identified region of neointimal hyperplasia. Additionally, constrictive remodeling was evidenced by an increase in medial thickness in injured vessels. The medial thickness of the hyperplastic segment was on average 8 μm thicker per rat compared to their contralateral healthy artery ($p=0.01$) (**Fig. 5 F**), but was not significantly different from the thickness of nearby uninjured arterial segments ($p=0.512$) (data not shown).

Validation of morphological parameters after LSFM processing. Even though the iDISCO+ clearing method we used has been shown to preserve morphology and size of cleared samples³⁰, it was important to verify that there was no arterial shrinkage or expansion in our experimental approach. Therefore, we compared the luminal surface area and perimeter, or EEL perimeter, between LSFM-processed and H&E-processed age-matched healthy arteries for both the rat (**Fig. 6 A-C**) and mouse models (**Fig. 6 D-F**). As noted in **Figure 1**, LSFM processing is reversible by removing the artery from the agarose block and rehydrating it. **Figure 7** shows representative α -

smooth muscle actin immunofluorescent staining and H&E staining of a rehydrated artery. This verifies that arterial samples processed for LSFM imaging can be used to analyze other biological parameters *a posteriori*. To assess if the rehydration process affects arterial size, we compared the I:M ratio and percent occlusion obtained by the optical volume to the optical slice or rehydrated H&E slice analysis for a single rat artery. Equally spaced optical sections along the region of hyperplasia of one rat were analyzed for the I:M ratio, percent occlusion, and perimeter. After rehydration, cryosectioning, and H&E staining, equally spaced sections were quantified and compared to the results from the optical slices. No significant differences were found for any of the parameters assessed (**Sup. Fig. 5**). However, the precision of the optical slice measurement was still higher as evidenced by lower CV% (38% for the percent occlusion estimate using optical slices vs 52% using rehydrated and H&E-stained slides).

IV. Discussion

We describe a platform for analyzing vascular injury models in both mouse and rat arteries in 3-D. While the maximum resolution of the LSFM used herein is lower than that of alternative imaging techniques, it can precisely measure key parameters of arterial injury in both rat and mouse injury models without the need for physical sectioning. Further, the reconstructed LSFM images in 3-D allow for several approaches to injury quantification of the same sample. As such, this study shows that LSFM can be used to analyze neointimal hyperplasia formation and arterial remodeling after injury and outperforms traditional histology methods in quantitative capabilities and precision.

The advantages of utilizing 3D datasets and volumetric analysis compared to classical 2D analysis have been studied by Gorelashvili *et al.* in the bone marrow²⁷. They showed that LSFM followed by 3D volumetric analysis of cells in the bone marrow reduces analysis bias and provides additional information to classically used techniques. Here we draw analogous conclusions regarding the power of this methodology when used to analyze vascular injury and subsequent stenosis. In the rat model we were able to obtain more precise estimates of stenosis and more

detailed quantitative information about remodeling compared to histology. We compared vessel diameter and medial thickness in areas of hyperplasia to the contralateral artery and regions without hyperplasia within the same artery. Additionally, we were able to use a repeated measures design to establish a correlation between intimal thickness and vessel diameter. In the mouse model, we obtained more precise estimates of stenosis compared to histology, and we found periadventitial neovascularization of injured arteries.

Classic histological methods for analyzing neointimal hyperplasia development provide an incomplete picture as a select few cross-sections are used to inform the full extent of pathological changes that occur in the vessel. The classic approach introduces user bias in slide selection for analysis and is not standardized throughout the scientific literature (**Table 1**). Furthermore, an extensive literature review of preclinical cardiovascular studies revealed that methodological sources of bias compound the preexisting limitations of interpreting animal data in modelling human disease³⁷. Thus, poor translation of preclinical models is in part due to the lack of rigor and reproducibility in preclinical study design and analysis. LSFM markedly improves the rigor of vascular disease assessment and is likely to improve reproducibility of preclinical findings. This is primarily because LSFM reduces user bias by imaging the entire intact artery, which can then be visualized in 3-D using Bitplane Imaris. Importantly, we show that tissue processing for LSFM imaging and analysis, and subsequent rehydration and cryosectioning, does not induce arterial shrinkage or expansion. Thus, the necessary parameters for quantifying vessel stenosis remain unaltered by processing. The obtained I:M ratio is comparable to that obtained by classic histology, but is more precise as evidenced by the lower CV%. Furthermore, the Filament tool in Bitplane Imaris allows for unbiased identification of hyperplastic regions, and detection of compensatory remodeling and medial thickening throughout the artery. It is important to note that even though sample processing time is longer than traditional processing for histology (**Fig. 1**), most of it is not active researcher time. Intravenous injection labeling for CD31 is a possible alternative strategy to expedite the staining protocol. This approach has been shown to effectively

label endothelial cells in the heart³⁸ as well as in the kidney³⁹. Additionally, it is also worth noting that there is a learning curve to become proficient with Bitplane Imaris analysis software. However, once enough proficiency is achieved, analysis time is not significantly longer than traditional histology. Importantly, unlike lengthy cryosectioning, analysis time allows for direct interaction with the data in a more intellectually engaging manner.

To our knowledge, the observation of the neovascular plexus around the adventitia of ligated mouse artery has not been previously reported. Kwon et al. showed that the adventitial vasa vasorum of pig coronaries increased after balloon injury. The neovascularization 28 days after injury was proportional to the degree of stenosis⁴⁰. No such finding had been reported for murine models. The periadventitial CD68⁺ regions in injured arteries close to the neovascular plexus suggest that inflammation might contribute to the angiogenic process. However, the process leading to the development of the periadventitial neovascular plexus in the injured arteries grants further research that is beyond the scope of this work. Nonetheless, our observation exemplifies the potential LSFM has in driving biological discoveries in murine models of vascular injury.

Compared to confocal microscopy, LSFM has significantly reduced photobleaching. That said, photobleaching can still be a problem if, during imaging, a single optical section is exposed to laser light for a very long time. In that case, the resulting photobleaching generates a darker x-y plane in the sample, which shows up as a “break”; this makes semi-automatic tracing on vessels (**Fig. 2**) more difficult. This problem can be circumvented by minimizing exposure during sample setup and, post hoc, by manually tracing contours in the vessels if semi-automatic tracing does not work. Biological studies often involve the need to study different molecular markers at once. Hence, multiplexing capabilities are an important feature of powerful techniques. Multiplexing using LSFM has been successfully used to image kidney organoids⁴¹, the mouse brain⁴², and human neural aggregates⁴³. Here we show the labeling of two different molecular markers at once to exemplify the multiplexing capabilities of the technique in studying vascular injury. We show

concurrent staining for CD68 and CD31, and for VE-cadherin and CD31. The resolution of this technique is enough to show the characteristic morphology of the endothelium as evidenced by the cobblestone pattern of VE-cadherin staining. However, when comparing the resolution of this form of LSFM to confocal microscopy, LSFM does not allow visualization of subcellular details; being able to rehydrate the tissue and do high-resolution confocal microscopy on sections allows multiple levels of detail to be probed, from the tissue-wide to the subcellular. The number of potential markers probed can be extended even further by staining individual sections with different combinations of antibodies. Thus, this method is flexible enough to allow exploration of varying levels of detail in the sample with many different molecular markers.

Potential application extends beyond preclinical rodent models as human samples have been imaged successfully with LSFM, including developing urogenital organs⁴⁴, the vascular network of the gingiva⁴⁵, and skin biopsies⁴⁶. Importantly in the cardiovascular field, Merz *et al.* successfully imaged a sample of human atrium at single-cell resolution³⁸. One consideration during imaging of human tissue is the sample size that can be successfully cleared and fit into the sample chamber. However, the human left anterior descending (LAD) coronary is well within the size that can be successfully imaged. The diameter of a healthy proximal human LAD is 3.7 mm in diameter⁴⁷ and has a wall thickness of 1.0 mm⁴⁸. Hence, we believe that the cardiovascular research potential of this methodology goes beyond preclinical murine models of vascular injury.

In conclusion, we describe a new tool to improve the analysis of preclinical vascular models. LSFM provides increased flexibility in analytical approaches and better precision in injury quantification over traditional histological analysis. Overall, this new workflow based on LSFM and volumetric analysis minimizes bias, it is versatile, more comprehensive than classical histological approaches, and will allow for a deeper understanding of preclinical vasculopathy models. Additionally, we believe this methodology can be applied to the study of other rodent vascular disease models, such as atherosclerosis, hypertension, and aneurysms.

V. Funding

This work and ESMB were supported by the National Institutes of Health, National Center for Advancing Translational Sciences, UNC Clinical and Translational Science Award-K12 Scholars Program [KL2TR002490] and the National Heart, Lung, and Blood Institute [K01HL145354]. NB was supported by the American Heart Association Predoctoral Fellowship [20PRE35120321]. FJM is supported by the Office of Research and Development, Department of Veterans Affairs [2I01BX001729] and the National Institutes of Health [HL130039]. The UNC Microscopy Services Laboratory, Department of Pathology and Laboratory Medicine, is supported in part by [P30 CA016086] Cancer Center Core Support Grant to the UNC Lineberger Comprehensive Cancer Center. Research reported in this publication was supported in part by the North Carolina Biotech Center Institutional Support Grant [2016-IDG-1016]. The UNC Neuroscience Center Microscopy Core Facility is supported, in part, by funding from the National Institute of Neurological Disorders and Stroke at the National Institutes of Health Neuroscience Center Support Grant [P30 NS045892] and the Eunice Kennedy Shriver National Institute of Child Health and Human Development at the National Institutes of Health Intellectual and Developmental Disabilities Research Center Support Grant [U54 HD079124]. The content of this manuscript is new and solely the responsibility of the authors and does not necessarily represent the official views of the granting agencies.

VI. Acknowledgments

The authors thank Dr. Hussein Kassam for assisting N.E.B. with rat surgeries and Ms. Danielle Berlin for proof reading the manuscript. Microscopy was performed at the UNC Microscopy Services Laboratory. Image analysis was performed at the UNC Neuroscience Center Microscopy Core Facility.

VII. Author Contributions

E.S.M.B. conceived the idea and directed the work. P.A., E.S.M.B., and N.E.B. established the processing and imaging protocol. N.E.B. drafted the initial manuscript and performed the rat surgeries, sample acquisition, and image analysis. J.L. and F.J.M. performed the mouse

548 surgeries, sample acquisition, and analysis. S.M. performed all histology for the rat arteries. All
549 authors contributed to writing and editing the manuscript.
550 **Conflict of Interest:** none declared.

References

1. Association AH. Cardiovascular Disease: A Costly Burden for America, Projections Through 2035. American Heart Association CVD Burden Report, 2017.
2. Roth GA, Johnson C, Abajobir A, Abd-Allah F, Abera SF, Abyu G, Ahmed M, Aksut B, Alam T, Alam K, Alla F, Alvis-Guzman N, Amrock S, Ansari H, Arnlov J, Asayesh H, Atey TM, Avila-Burgos L, Awasthi A, Banerjee A, Barac A, Barnighausen T, Barregard L, Bedi N, Belay Ketema E, Bennett D, Berhe G, Bhutta Z, Bitew S, Carapetis J, Carrero JJ, Malta DC, Castaneda-Orjuela CA, Castillo-Rivas J, Catala-Lopez F, Choi JY, Christensen H, Cirillo M, Cooper L, Jr., Criqui M, Cundiff D, Damasceno A, Dandona L, Dandona R, Davletov K, Dharmaratne S, Dorairaj P, Dubey M, Ehrenkranz R, El Sayed Zaki M, Faraon EJA, Esteghamati A, Farid T, Farvid M, Feigin V, Ding EL, Fowkes G, Gebrehiwot T, Gillum R, Gold A, Gona P, Gupta R, Habtewold TD, Hafezi-Nejad N, Hailu T, Hailu GB, Hankey G, Hassen HY, Abate KH, Havmoeller R, Hay SI, Horino M, Hotez PJ, Jacobsen K, James S, Javanbakht M, Jeemon P, John D, Jonas J, Kalkonde Y, Karimkhani C, Kasaeian A, Khader Y, Khan A, Khang YH, Khera S, Khoja AT, Khubchandani J, Kim D, Kolte D, Kosen S, Krohn KJ, Kumar GA, Kwan GF, Lal DK, Larsson A, Linn S, Lopez A, Lotufo PA, El Razek HMA, Malekzadeh R, Mazidi M, Meier T, Meles KG, Mensah G, Meretoja A, Mezgebe H, Miller T, Mirrakhimov E, Mohammed S, Moran AE, Musa KI, Narula J, Neal B, Ngalesoni F, Nguyen G, Obermeyer CM, Owolabi M, Patton G, Pedro J, Qato D, Qorbani M, Rahimi K, Rai RK, Rawaf S, Ribeiro A, Safiri S, Salomon JA, Santos I, Santric Milicevic M, Sartorius B, Schutte A, Sepanlou S, Shaikh MA, Shin MJ, Shishehbor M, Shore H, Silva DAS, Sobngwi E, Stranges S, Swaminathan S, Tabares-Seisdedos R, Tadele Atnafu N, Tesfay F, Thakur JS, Thrift A, Topor-Madry R, Truelsen T, Tyrovolas S, Ukwaja KN, Uthman O, Vasankari T, Vlassov V, Vollset SE, Wakayo T, Watkins D, Weintraub R, Werdecker A, Westerman R, Wiysonge CS, Wolfe C, Workicho A, Xu G, Yano Y, Yip P, Yonemoto N, Younis M, Yu C, Vos T, Naghavi M, Murray C. Global, Regional, and National Burden of Cardiovascular Diseases for 10 Causes, 1990 to 2015. *J Am Coll Cardiol* 2017;**70**:1-25.
3. Weintraub WS. The pathophysiology and burden of restenosis. *Am J Cardiol* 2007;**100**:3K-9K.
4. Clowes AW, Reidy MA, Clowes MM. Mechanisms of stenosis after arterial injury. *Lab Invest* 1983;**49**:208-215.
5. Jackson CL. Animal models of restenosis. *Trends Cardiovasc Med* 1994;**4**:122-130.
6. Tulis DA. Histological and morphometric analyses for rat carotid balloon injury model. *Methods Mol Med* 2007;**139**:31-66.
7. Mitsutake Y, Reifart J, Pyun WB, Lyons JK, Deuse T, Schrepfer S, Ikeno F. Differences in Vascular Response between Balloon Overstretch and Stent Overexpansion in Nonatherosclerotic Porcine Coronary Arteries. *Comp Med* 2017;**67**:350-355.
8. E Y, He N, Fan H. The effect of short-term intra-arterial delivery of paclitaxel on neointimal hyperplasia and the local thrombotic environment after angioplasty. *Cardiovasc Intervent Radiol* 2013;**36**:1086-1096.
9. Cao T, Zhang L, Yao LL, Zheng F, Wang L, Yang JY, Guo LY, Li XY, Yan YW, Pan YM, Jiang M, Chen L, Tang JM, Chen SY, Wang JN. S100B promotes injury-induced vascular remodeling through modulating smooth muscle phenotype. *Biochimica et Biophysica Acta - Molecular Basis of Disease* 2017;**1863**:2772-2782.
10. Madigan M, Entabi F, Zuckerbraun B, Loughran P, Tzeng E. Delayed inhaled carbon monoxide mediates the regression of established neointimal lesions. *J Vasc Surg* 2015;**61**:1026-1033.
11. Bahnson ES, Kassam HA, Moyer TJ, Jiang W, Morgan CE, Vercammen JM, Jiang Q, Flynn ME, Stupp SI, Kibbe MR. Targeted Nitric Oxide Delivery by Supramolecular Nanofibers for the Prevention of Restenosis After Arterial Injury. *Antioxidants & Redox Signaling* 2016;**24**:401-418.

- 597 12. Meng Z, Gao P, Chen L, Peng J, Huang J, Wu M, Chen K, Zhou Z. Artificial Zinc-Finger Transcription
598 Factor of A20 Suppresses Restenosis in Sprague Dawley Rats after Carotid Injury via the PPARalpha
599 Pathway. *Molecular Therapy - Nucleic Acids* 2017;**8**:123-131.
- 600 13. Yang J, Fan Z, Yang J, Ding J, Yang C, Chen L. MicroRNA-24 Attenuates Neointimal Hyperplasia in
601 the Diabetic Rat Carotid Artery Injury Model by Inhibiting Wnt4 Signaling Pathway. *Int J Mol Sci*
602 2016;**17**.
- 603 14. Chen J, Zhang J, Yang J, Xu L, Hu Q, Xu C, Yang S, Jiang H. Histone demethylase KDM3a, a novel
604 regulator of vascular smooth muscle cells, controls vascular neointimal hyperplasia in diabetic
605 rats. *Atherosclerosis* 2017;**257**:152-163.
- 606 15. Yeh YH, Chang SH, Chen SY, Wen CJ, Wei FC, Tang R, Achilefu S, Wun TC, Chen WJ. Bolus injections
607 of novel thrombogenic site-targeted fusion proteins comprising annexin-V and Kunitz protease
608 inhibitors attenuate intimal hyperplasia after balloon angioplasty. *Int J Cardiol* 2017;**240**:339-346.
- 609 16. Buglak NE, Jiang W, Bahnson ESM. Cinnamic aldehyde inhibits vascular smooth muscle cell
610 proliferation and neointimal hyperplasia in Zucker Diabetic Fatty rats. *Redox Biology* 2018;**19**:166-
611 178.
- 612 17. Lin JS, Wang CJ, Li WT. Photodynamic therapy of balloon-injured rat carotid arteries using
613 indocyanine green. *Lasers Med Sci* 2018;**33**:1123-1130.
- 614 18. He M, Wang C, Sun JH, Liu Y, Wang H, Zhao JS, Li YF, Chang H, Hou JM, Song JN, Li AY, Ji ES.
615 Roscovitine attenuates intimal hyperplasia via inhibiting NF-kappaB and STAT3 activation induced
616 by TNF-alpha in vascular smooth muscle cells. *Biochem Pharmacol* 2017;**137**:51-60.
- 617 19. Zhang Y, Li YH, Liu C, Nie CJ, Zhang XH, Zheng CY, Jiang W, Yin WN, Ren MH, Jin YX, Liu SF, Zheng
618 B, Wen JK. miR-29a regulates vascular neointimal hyperplasia by targeting YY1. *Cell Prolif* 2017;**50**.
- 619 20. Makino Y, Miyahara T, Nitta J, Miyahara K, Seo A, Kimura M, Suhara M, Akai A, Akagi D, Yamamoto
620 K, Hoshina K. Proresolving Lipid Mediators Resolvin D1 and Protectin D1 Isomer Attenuate
621 Neointimal Hyperplasia in the Rat Carotid Artery Balloon Injury Model. *J Surg Res* 2019;**233**:104-
622 110.
- 623 21. Li YQ, Wang JY, Qian ZQ, Li YL, Li WN, Gao Y, Yang DL. Osthole inhibits intimal hyperplasia by
624 regulating the NF-kappaB and TGF-beta1/Smad2 signalling pathways in the rat carotid artery after
625 balloon injury. *Eur J Pharmacol* 2017;**811**:232-239.
- 626 22. Zhang WX, Tai GJ, Li XX, Xu M. Inhibition of neointima hyperplasia by the combined therapy of
627 linagliptin and metformin via AMPK/Nox4 signaling in diabetic rats. *Free Radic Biol Med*
628 2019;**143**:153-163.
- 629 23. Xiong W, Luo Y, Wu L, Liu F, Liu H, Li J, Liao B, Dong S. Chemerin Stimulates Vascular Smooth
630 Muscle Cell Proliferation and Carotid Neointimal Hyperplasia by Activating Mitogen-Activated
631 Protein Kinase Signaling. *PLoS One* 2016;**11**:e0165305.
- 632 24. Power RM, Huisken J. A guide to light-sheet fluorescence microscopy for multiscale imaging.
633 *Nature Methods* 2017;**14**:360-373.
- 634 25. Chatterjee K, Pratiwi FW, Wu FCM, Chen P, Chen BC. Recent Progress in Light Sheet Microscopy
635 for Biological Applications. *Applied Spectroscopy* 2018;**72**:1137-1169.
- 636 26. Chen BC, Legant WR, Wang K, Shao L, Milkie DE, Davidson MW, Janetopoulos C, Wu XS, Hammer
637 JA, 3rd, Liu Z, English BP, Mimori-Kiyosue Y, Romero DP, Ritter AT, Lippincott-Schwartz J, Fritz-
638 Laylin L, Mullins RD, Mitchell DM, Bembenek JN, Reymann AC, Bohme R, Grill SW, Wang JT,
639 Seydoux G, Tulu US, Kiehart DP, Betzig E. Lattice light-sheet microscopy: imaging molecules to
640 embryos at high spatiotemporal resolution. *Science* 2014;**346**:1257998.
- 641 27. Gorelashvili MG, Angay O, Hemmen K, Klaus V, Stegner D, Heinze KG. Megakaryocyte volume
642 modulates bone marrow niche properties and cell migration dynamics. *Haematologica* 2019.
- 643 28. Stegner D, vanEeuwijk JMM, Angay O, Gorelashvili MG, Semeniak D, Pinnecker J, Schmithausen P,
644 Meyer I, Friedrich M, Dutting S, Brede C, Beilhack A, Schulze H, Nieswandt B, Heinze KG.

- Thrombopoiesis is spatially regulated by the bone marrow vasculature. *Nature communications* 2017;**8**:127.
29. Richardson DS, Lichtman JW. Clarifying Tissue Clearing. *Cell* 2015;**162**:246-257.
30. Renier N, Adams EL, Kirst C, Wu Z, Azevedo R, Kohl J, Autry AE, Kadiri L, Umadevi Venkataraju K, Zhou Y, Wang VX, Tang CY, Olsen O, Dulac C, Osten P, Tessier-Lavigne M. Mapping of Brain Activity by Automated Volume Analysis of Immediate Early Genes. *Cell* 2016;**165**:1789-1802.
31. Tulis DA. Rat carotid artery balloon injury model. *Methods Mol Med* 2007;**139**:1-30.
32. Shears LL, 2nd, Kibbe MR, Murdock AD, Billiar TR, Lizonova A, Kovesdi I, Watkins SC, Tzeng E. Efficient inhibition of intimal hyperplasia by adenovirus-mediated inducible nitric oxide synthase gene transfer to rats and pigs in vivo. *J Am Coll Surg* 1998;**187**:295-306.
33. Ariel P. UltraMicroscope II - A User Guide. Chapel Hill (NC), 2018.
34. Hamlett A, Ryan L, Wolfinger R. On the use of PROC MIXED to Estimate Correlation in the Presence of Repeated Measures. SUGI 29 Proceedings, 2004.
35. Lehmann EL. Elements of Large-Sample Theory. Springer, 1999.
36. Corliss BA, Azimi MS, Munson JM, Peirce SM, Murfee WL. Macrophages: An Inflammatory Link Between Angiogenesis and Lymphangiogenesis. *Microcirculation* 2016;**23**:95-121.
37. Ramirez FD, Motazedian P, Jung RG, Di Santo P, MacDonald ZD, Moreland R, Simard T, Clancy AA, Russo JJ, Welch VA, Wells GA, Hibbert B. Methodological Rigor in Preclinical Cardiovascular Studies: Targets to Enhance Reproducibility and Promote Research Translation. *Circ Res* 2017;**120**:1916-1926.
38. Merz SF, Korste S, Bornemann L, Michel L, Stock P, Squire A, Soun C, Engel DR, Detzer J, Lorchner H, Hermann DM, Kamler M, Klode J, Hendgen-Cotta UB, Rassaf T, Gunzer M, Totzeck M. Contemporaneous 3D characterization of acute and chronic myocardial I/R injury and response. *Nature communications* 2019;**10**:2312.
39. Klingberg A, Hasenberg A, Ludwig-Portugall I, Medyukhina A, Mann L, Brenzel A, Engel DR, Figge MT, Kurts C, Gunzer M. Fully Automated Evaluation of Total Glomerular Number and Capillary Tuft Size in Nephritic Kidneys Using Lightsheet Microscopy. *J Am Soc Nephrol* 2017;**28**:452-459.
40. Kwon HM, Sangiorgi G, Ritman EL, Lerman A, McKenna C, Virmani R, Edwards WD, Holmes DR, Schwartz RS. Adventitial vasa vasorum in balloon-injured coronary arteries: visualization and quantitation by a microscopic three-dimensional computed tomography technique. *J Am Coll Cardiol* 1998;**32**:2072-2079.
41. Held M, Santeramo I, Wilm B, Murray P, Levy R. Ex vivo live cell tracking in kidney organoids using light sheet fluorescence microscopy. *PloS one* 2018;**13**:e0199918.
42. Kim SY, Cho JH, Murray E, Bakh N, Choi H, Ohn K, Ruelas L, Hubbert A, McCue M, Vassallo SL, Keller PJ, Chung K. Stochastic electrotransport selectively enhances the transport of highly electromobile molecules. *Proc Natl Acad Sci U S A* 2015;**112**:E6274-6283.
43. Gualda EJ, Simao D, Pinto C, Alves PM, Brito C. Imaging of human differentiated 3D neural aggregates using light sheet fluorescence microscopy. *Front Cell Neurosci* 2014;**8**:221.
44. Isaacson D, McCreedy D, Calvert M, Shen J, Sinclair A, Cao M, Li Y, McDevitt T, Cunha G, Baskin L. Imaging the developing human external and internal urogenital organs with light sheet fluorescence microscopy. *Differentiation* 2020;**111**:12-21.
45. Azaripour A, Lagerweij T, Scharfbillig C, Jadcak AE, Swaan BV, Molenaar M, Waal RV, Kielbassa K, Tigchelaar W, Picavet DI, Jonker A, Hendriks EML, Hira VVV, Khurshed M, Noorden C. Three-dimensional histochemistry and imaging of human gingiva. *Sci Rep* 2018;**8**:1647.
46. Abadie S, Jarret C, Colombelli J, Chaput B, David A, Grolleau JL, Bedos P, Lobjois V, Descargues P, Rouquette J. 3D imaging of cleared human skin biopsies using light-sheet microscopy: A new way to visualize in-depth skin structure. *Skin Res Technol* 2018;**24**:294-303.

692 47. Dodge JT, Jr., Brown BG, Bolson EL, Dodge HT. Lumen diameter of normal human coronary
693 arteries. Influence of age, sex, anatomic variation, and left ventricular hypertrophy or dilation.
694 *Circulation* 1992;**86**:232-246.
695 48. Perry R, Joseph MX, Chew DP, Aylward PE, De Pasquale CG. Coronary artery wall thickness of the
696 left anterior descending artery using high resolution transthoracic echocardiography--normal
697 range of values. *Echocardiography* 2013;**30**:759-764.
698
699

Figure 1

Figure 2

Figure 3

Figure 4

Figure 5

Figure 6

Figure 7

Tables

Table 1. Neointimal hyperplasia analysis in rat balloon injury models according to published method sections.

Number of Analyzed Sections Directly Described	
Analysis Format	Reference
10 evenly-spaced, discontinuous sections	8, 9
8 total sections	10
6 to 10 evenly-spaced, discontinuous sections	11, 12
5 total sections	13, 14
3 evenly-spaced, discontinuous sections	15, 16
Detailed Analysis Descriptions	
Analysis Format	Reference
"3 pictures per section," number of sections per rat not listed	17
"6 random, noncontiguous microscopic fields per section [for] 3 sections"	18
"6 random, noncontiguous microscopic fields per section [for] 6 sections"	19
"For three sections, the common carotid artery was divided into five equal parts, from which one of the three central samples was selected for evaluation"	20
"The sections (4- μ m thickness, 350 μ m apart) [...] from the middle portion of the balloon-treated segment were analysed"	21
none stated	22, 23

Figure Legends

Figure 1. Sample processing workflow comparison for H&E or LSFM analysis. Arteries processed for LSFM imaging are able to be rehydrated and cryosectioned for H&E analysis.

Figure 2. Analysis workflow in Bitplane Imaris. Rat carotid artery used as the representative sample. Video 2 overlays the three surfaces with the fluorescent volumetric rendering (<https://youtu.be/LoAlApmV22o>).

Figure 3. Demonstration of LSFM imaging for rat and mouse injury models. **(A)** Vessel orientation and figure image colors, unless otherwise stated. **(B)** Volumetric rendering and slice view of healthy rat artery (L = lumen). **(C)** Representative balloon-injured rat artery. **(D)** Representative slice views of an injured rat artery (arrowheads = hyperplastic region). White box indicates region of the surface rendering of the internal elastic lamina (IEL) and lumen volumes. **(E)** Representative cross-section in x-z axis from arrowheads in (D) with H&E-stained image of the same artery after rehydration and cryosectioning. **(F)** Cross-section of surface rendering of the

representative rat (from (D) and (E)) arterial media, hyperplasia, and both surfaces overlaid. **(G)** Representative H&E-stained cross-sections of ligated mouse artery processed exclusively for histology. **(H)** Longitudinal and cross-sectional slice rendering of representative ligated mouse carotid artery (white arrow = ligation site at bifurcation) **(I)** Representative volume rendering of ligated mouse carotid artery depicting the periadventitial neovascular plexus (white arrow = ligation site at bifurcation, arrowheads = neo-vascular plexus). **(J)** Slice and volumetric rendering of representative healthy mouse artery.

Figure 4. Macrophage presence in the adventitia of injured mouse carotids. Uninjured and ligated mouse carotid arteries stained for CD31 and CD68 show presence of CD68⁺ puncta only in the ligated arteries around the ligation site and periadventitially next to the CD31⁺ neovascular plexus (n = 3, white = CD31, magenta = CD68, green = autofluorescence). **(A)** 3D rendering healthy mouse carotid arteries. **(B)** 3D rendering healthy mouse carotid arteries.

Figure 5. Demonstration of LSFM quantification. **(A)** Rat intima-to-media (I:M) ratio using optical 3-D volume compared to a separate cohort of injured rat arteries harvested directly for histological processing (LSFM n = 7, H&E n = 10, n.s. = no significance, student's t-test). **(B)** Mouse I:M ratio using optical volume compared to a separate cohort of ligated mouse arteries harvested directly for histological processing (LSFM n = 6, H&E n = 12). **(C)** Plot of injured rat artery IEL radius and change (Δ) in radius between the lumen and the IEL along the artery center (dataset shown in Fig. 3 D). 99% confidence interval (CI) defines the threshold at which the Δ Radius is indicative of neointimal hyperplasia formation. 99% CI determined using the average Δ Radius along the contralateral healthy rat artery (Sup. Fig. 1). **(D-E)** Average IEL radii per rat along their hyperplastic region compared to the non-hyperplastic region of the injured artery or to their contralateral healthy artery ([#]p<0.05, paired t-test). **(F)** Plot of medial thickness comparing the hyperplastic region of each injured artery to the respective contralateral healthy control artery.

Figure 6. Artery processing for LSFM imaging does not cause arterial shrinkage. Comparison between healthy, uninjured arteries processed only for LSFM or only for hematoxylin & eosin

(H&E) analysis. **(A)** Representative cross-section of healthy rat arteries. **(B)** Average luminal surface area per healthy rat artery (n = 7, n.s. = no significance, student's t-test). **(C)** Average luminal and external elastic lamina (EEL) perimeter per healthy rat artery. **(D)** Representative cross-section of healthy mouse arteries. **(E)** Average luminal surface area per healthy mouse artery (LSFM n = 6, H&E n = 8). **(F)** Average luminal and EEL perimeter per healthy mouse artery.

Figure 7. Histological staining of rehydrated, LSFM-processed rat carotid artery. **(A)** Immunofluorescent staining for alpha-smooth muscle (α SM) actin. **(B)** H&E staining of three representative arterial cross-sections.

Video 1. Injured rat artery slices in x-z and x-y planes followed by volume reconstruction (green = autofluorescence, white = CD31). To see video visit: <https://youtu.be/32oVvm0l3zq>

Video 2. Injured rat artery slice view in x-y plane followed by volume reconstruction and surface generation of the hyperplastic region. Surface coloring matches that used in Figure 2 (green = autofluorescence, white = CD31, yellow = lumen, blue = internal elastic lamina, pink = external elastic lamina). To see video visit: <https://youtu.be/LoAlApmV22o>

Video 3. Healthy rat artery slices in x-z and x-y planes followed by volume reconstruction (green = autofluorescence, white = CD31). To see video visit: <https://youtu.be/fPvTRZzmUy4>