Interaction of intraocular pressure and ganglion cell function in open angle glaucoma

Authors and affiliations

Al-Nosairy K. O.¹, Van den Bosch J.J.O.N.^{1,2}, Pennisi V.¹, Thieme H.¹, Mansouri K.^{3,4}, Choritz L.¹, Hoffmann M. B.^{1,5}

¹Department of Ophthalmology, Otto-von-Guericke University, Leipziger Str. 44, 39120 Magdeburg, Germany

²Implandata Ophthalmic Products GmbH, Hannover, Germany

³Glaucoma Research Center, Montchoisi Clinic, Swiss Vision Network, Lausanne, Switzerland

⁴Department of Ophthalmology, University of Colorado, Denver, US

⁵Center for Behavioral Brain Sciences, Magdeburg, Germany

Correspondent details:

Michael B. Hoffmann

University Eye Clinic, Visual Processing Laboratory, Leipziger Str. 44, 39120 Magdeburg, Germany

Phone: +49 391 67 13585

Fax: +49 391 67 13570

Email: michael.hoffmann@med.ovgu.de

Abstract

Purpose: To test the feasibility of simultaneous steady-state pattern electroretinogram (PERG) and intraocular pressure (IOP) measurements with an IOP sensor and to test a model for IOP manipulation during lateral decubitus positioning (LDP) and its impact on the PERG.

Design: A prospective, observational study.

Methods: 15 healthy controls and 15 treated glaucoma patients participated in the study. 8 patients had an intraocular IOP sensor (eyemate-IO[®], Implants of Ophthalmic Products GmbH) in the right eye (GLA_{IMP}) and 7 had no sensor and with glaucoma in the left eye. (1) We tested the feasibility of simultaneous IOP and PERG recordings by comparing PERGs with and without simultaneous IOP-read out in GLA_{IMP}. (2) All participants were positioned in the following order: sitting1 (S1), right LDP (LD_R), sitting2 (S2), left LDP (LD_L) and sitting3 (S3). For each position, PERG amplitudes and IOP were determined with rebound tonometry (Icare[®] TA01i) in all participants without the IOP sensor.

Results: Electromagnetic intrusions of IOP sensor readout onto steady-state PERG-recordings had, due to different frequency ranges, no relevant effect on PERG amplitudes. IOP and PERG measures were affected by LDP, e.g., IOP was increased during LD_R vs S1 in the lower eyes of GLA_{IMP} and controls ($P < 0.001$ and $P < 0.05$, respectively) and PERG amplitude was decreased ($P < 0.05$ and $P < 0.01$, respectively).

Conclusions: During LDP, IOP and PERG measurements changed more in the lower eye. IOP changes induced by LDP may be a model for studying the interaction of IOP and ganglion cell function.

Introduction

While glaucoma is a global leading cause of irreversible blindness^{1–3}, its pathogenesis is not yet fully understood. One approach to fill this gap is deciphering the mechanisms via which intraocular pressure (IOP), a well-known risk for glaucoma, impacts on ganglion cell function. An attractive maneuver to elucidate these mechanisms is the manipulation of IOP, e.g. by utilizing posture-induced IOP-changes. And to manifest its impact on retinal function, e.g. with non-invasive electrophysiology.

An important tool for the assessment of retinal ganglion cell function is the steady state pattern electroretinogram (ssPERG)^{4,5}. Consequently, the PERG is of value for the detection and investigation of glaucoma^{6,7}. Previous studies indicate that PERG-measurements during IOP manipulations may be of diagnostic importance in glaucoma and have a predictive role for future conversion of glaucoma suspects. Specifically, posture-induced IOP changes offer a straightforward option to manipulate IOP. For example, supine posture with -10° head down tilt in a subpopulation of glaucoma suspects was found to be associated with an increase in IOP and a decrease of PERG amplitude^{8,9}. An increase in IOP was also reported during lateral decubitus posture (LDP) for both participants with healthy vision^{10–12} and for glaucoma patients^{13–15}. Such approaches would greatly benefit from simultaneous IOP and PERG measurements, which was up-to-date technically not feasible. Recently developed telemetric IOP monitoring devices have the potential to fill this gap, as they open the possibility of simultaneous IOP- and electrophysiological recordings.

Telemetric IOP sensors were first proposed in 1967¹⁶ but were not tested in clinical trials. At present, two technologies are commercially available for continual IOP monitoring: a contact lens sensor (Sensimed Triggerfish)^{17–19} and an implantable intraocular pressure sensor (eyemate-IO sensor)^{20–23}. With the latter approach, a wireless ring shaped sensor is placed in the ciliary sulcus during co-implantation of an intraocular lens during cataract surgery²¹. Studies have shown the eyebate-IO sensor to be safe, well tolerated and of good functionality to continuously measure IOP^{21,23–25}. In principle, it should enable simultaneous PERG and IOP readings.

This study is the first to perform simultaneous electroretinographical recordings and continual IOP measurements in glaucoma patients. We assessed the feasibility of simultaneous PERG and IOP measurements and demonstrated the relation of ganglion cell responses to IOP changes in LDP, with the lower eye, i.e. also termed 'dependent eye'¹²), being most affected during LDP.

Methods

Participants

This prospective observational study was conducted in University Eye Clinic of Otto-von-Guericke University of Magdeburg. The participants gave their written consent to participate in the study. The procedures followed the tenets of the declaration of Helsinki and the protocol was approved by the ethical committee of the Otto-von-Guericke University of Magdeburg, Germany. All participants, three groups as detailed below, underwent complete ophthalmic examinations and best corrected visual acuity testing using the Early Treatment for Diabetic Retinopathy Study's (ETDRS) chart²⁶ and visual field testing using the Swedish Interactive Threshold Algorithm 24-2 protocol (SITA) of Humphrey Field Analyzer 3 (Carl Zeiss Meditec AG, Jena, Germany). Clinical and demographic data for all participants are shown in Table 1.

Exclusion criteria were any systemic diseases, ocular diseases or surgeries that might affect electrophysiology recordings except cataract surgery and, in the glaucoma group, glaucoma surgery or incipient cataract that did not decrease BCVA < 0.8²⁷ and refractive error exceeding -6 or +3 D or astigmatism ≥ 2 D.

GLA_{IMP} group. Eight patients (age range [years]: 62-77 years; mean ± SE: 71.1 ± 1.9 years) with open angle glaucoma who had previously been implanted with a telemetric IOP sensor in the right eye (eyemate-IO®, Implants of Ophthalmic Products GmbH, Hannover, Germany), were enrolled in the study. Five of 8 subjects participated in experiment 1 to test the influence of eyebate-IO sensor electromagnetic radiation on PERG recordings in both eyes. All 8 participants participated in experiment 2 to determine the LDP effect on PERG amplitudes. Intraocular pressure and PERG measurements were included only for the right eye as not all left eyes had glaucoma. The eyebate-IO sensor had previously (at least 3

years prior to the present measurements) been implanted in glaucomatous patients as part of a preceding study²⁴.

GLA_{LE}-group. The left eyes of 7 patients (age range [years]: 34-78; mean age $\pm$ SE: 51.4 ± 5.6) with open angle glaucoma (POAG) were included and were compared to controls' left eyes (CON_{LE} as defined below). Not all right eyes of this group had glaucoma; right eyes were, therefore, not included in the analysis.

All OAG eyes either had glaucoma hemifield test that was outside normal limits (Humphrey Field Analyzer 3, as detailed below), a typical visual field glaucomatous damage manifested as a cluster of 3 or more non-edge points all depressed on the pattern deviation plot $< 5\%$ and one of which depressed $< 1\%$ or abnormal corrected pattern standard deviation $< 5\%$ on Humphrey Swedish interactive threshold algorithm 24-2 (SITA fast)²⁸ and a glaucomatous appearance of the optic disc, i.e. with a general enlargement of the cupping defined as vertical cup-to-disc ratio ≥ 0.7 , retinal fiber layer defect or a local notching of the rim and an open anterior chamber angle. All glaucoma patients were under IOP-lowering treatment.

Normal controls (CON_{RE}/CON_{LE}). Fifteen subjects (age range [years]: 33-65; mean age $\pm$ SE: 52.4 ± 3.8) with normal visual acuity (see Table 1) without ocular diseases were included in the study. Eight were included in experiment 1; eleven were enrolled in experiment 2. In experiment 2, controls' right eyes (CON_{RE}) were compared to the right eye of GLA_{IMP} group whereas left eyes of controls (CON_{LE}) were compared to GLA_{LE} group.

IOP sensor (eyemate-IO sensor) and external reader device (Mesograph)

The eyemate-IO sensor is a Micro-Electro-Mechanical System Application-Specific Integrated Circuit (MEMS ASIC) comprising pressure sensitive sensor cells, temperature sensors, analog-to-digital converters, and telemetry. The ASIC is bonded to a gold-made circular micro-coil antenna, both parts are silicon-encapsulated²⁹. For readout of the sensor, typically an external reader device (Mesograph) is held in front of the eye with < 5 cm distance to power the implant. Upon button-press on the device, it emits a radio frequency field at 13.56 MHz, which establishes an electromagnetic link towards the implant. When sufficient power is available to the ASIC, a pressure reading is being performed and digitized data is transferred back to the external reader device. The eyemate-IO sensor can

obtain IOP measurements comprising around 10 individual samples per second^{21,29}. As an alternative to holding the external reader device manually in front of the eye, a reader can be connected to a coiled circular antenna that is positioned and attached around the eye with a plaster (Figure 1). This way, a continual IOP readout can be performed³⁰ at an average sampling rate of 9.2 Hz and recorded with a computer via a USB connection. We applied this continuous readout mode in our study.

Visual stimuli, procedure and recordings

The EP2000 evoked potential system was used for stimulation, recording and analysis of steady-state PERGs³¹ following the international society for clinical electrophysiology of vision (ISCEV) standards for PERG-recordings³². The stimuli were presented binocularly at a frame rate of 75 Hz on a monochrome monitor (MDG403, Philips; P45 phosphor) in a dimly-lit room. The participants maintained fixation at the center of the monitor using a fixation cross (1° diameter) which was replaced by a 200 ms-duration digit randomly appearing every 5-20 seconds. Contrast-inverting (15 Hz) checkerboard patterns (visual field: 25° x 25°; mean luminance: 45 cd/m²; contrast: 98%) with two check sizes, 0.8° and 15°, were presented for stimulation at a viewing distance of 57 cm. Two PERG blocks were recorded for each sitting position and 4 PERG blocks were recorded for each LDP. Repeated blocks were averaged offline using Igor (IGOR Pro, WaveMetrics, Portland). Each PERG block contained 8 stimulus cycles of 10 trials per checksize (80 sweeps of 1.066 s trial duration). Signals exceeding ± 90 μ V were considered artifacts originating from eye movements or blinking, rejected and recollected. The pupils were not dilated.

Five silver-cup skin electrodes (9 mm diameter; Silver EEG Cup Electrodes, Natus Manufacturing Limited, Ireland) filled with conductive paste (Ten20, WEAVER and Company, USA) were used as recording and reference electrodes and applied after cleaning the skin with a cleaning paste (skinPure, NIHON KODEN Corporation, Tokyo, Japan) to keep skin conductance below 5 kOhm. Binocular PERGs were recorded using active skin electrodes placed about 5 mm below the eyelids vertically aligned with the pupil in the primary gaze^{6,33}. The reference electrode was attached to the temple ipsilateral to the corresponding eye. Similarly, the ground electrode was placed at Fpz^{34,35}. The signal was amplified by 50 k (Grass Model 12, Astro-

Med, Inc., West Warwick, RI, USA), band-pass filtered 1-100 Hz and digitized at 1 kHz with 12-bit resolution by a G4 Power Macintosh computer running the EP2000 Evoked Potentials System³¹.

Experiment 1 – Effect of eyemate-IO sensor read-out on PERG recordings.

Since the eyemate IO-sensor functions employ electromagnetic radiation coupling with the antenna, we tested for possible interferences and effects of eyemate-IO sensor on the PERG recordings. To confirm any changes detected in patients, we additionally simulated IOP readout in controls, which required, due to the read-out procedure, the reader to detect a sensor, which was therefore placed externally next to the eye, i.e. without IOP-functionality. For this purpose, the eyemate-IO sensor was fixed below the right lateral third of the lower eyelid with the antenna placed in front of it to allow electromagnetic coupling to the external reader. Participants in experiment 1 underwent four blocks of PERG binocular recordings, two with read-out switch on, i.e. IO-Reader_{ON}, and two with read-out switched off, i.e. IO-Reader_{OFF}, in a counterbalanced sequence ('A-B-B-A'-scheme). The two PERG blocks (2 x 80 sweeps) per condition and same readout status were averaged. To assess the raw electrophysiological recordings prior to averaging in EP2000, the recordings were acquired using the PowerLab recording system (Model M880, ADInstrument Pty Ltd, Australia).

Experiment 2 – Effect of lateral decubitus posture (LDP) on PERG.

PERGs were recorded binocularly while the participants were positioned in the following sequence: Sitting (S1), right LDP (LD_R), sitting (S2), left LDP (LD_L) and sitting (S3). Five minutes after taking each position, the IOP was measured with an Icare tonometer (Icare® TA01i Tonometer, Helsinki, Finland) for the control- and the GLA_{LE}-group or simultaneously during PERG recording with eyemate-IO sensor for the GLA_{IMP}-group. Pillows were used to support the head during lateral decubitus and to assure that the head was parallel to the ground. Care was taken that the pillow did not compress the lower eye during LDP. The IOP was measured in the center of the cornea and started always with the right eye followed by the left eye; the IOP value was an automatic average of one set comprising 6 measurements. As a measure to reduce the probability of IOP erroneous readings (i.e tilting of the

device and misalignment), we asked the patient to always look straight and kept the orientation of Icare TA01i parallel to the ground in both sitting position and LDP. Also, the use of Icare TA01i is clinically robust and appeared insensitive to lateral and angular deviations during measurements³⁶.

Analysis and statistics

ssPERG were Fourier-analyzed and the response amplitude at stimulation frequency, i.e. 15 Hz was determined and corrected for the noise estimate, i.e. an average of the neighboring frequencies below and above 15 Hz^{37,38}.

The non-averaged traces acquired during experiment 1 were digitized with a PowerLab recording system (Model M880, ADInstrument Pty Ltd, Australia) and exported to Igor (IGOR Pro, WaveMetrics, Portland) for subsequent analyses and the determination of intrusion frequency during IOP readout from the eyemate-IO sensor. ssPERG amplitudes were transferred into IGOR sheet containing glaucoma and controls data and graphic representation of each amplitude check size in both eyes and both reader-out states were plotted. Similarly, ssPERG amplitudes were plotted in Igor for controls and glaucoma_{IMP} patients for different postures. To test the effect of LDP on IOP and ganglion cell function, these measures were compared for S1 vs LD_R and S2 vs LD_L with paired T-tests using SPSS 24 (statistical Package for the Social Sciences, IBM). P values were corrected using the Bonferroni-Holm correction³⁹ for multiple comparisons.

Results

Experiment 1 – Feasibility of simultaneous eyemate-IO sensor read-out and ssPERG recordings

A qualitative overview over the ssPERG recordings with (IO-Reader_{ON}) and without simultaneous eyemate-IO sensor read-out (IO-Reader_{OFF}) is given in Figure 2 for a participant with the eyemate-IO sensor implanted. In Figure 2A non-averaged recording traces for IO-Reader_{ON} and IO-Reader_{OFF} are depicted. It is evident that there is a periodic high amplitude intrusion into the recordings only for IO-Reader_{ON}, which was in the order of 80 μ V (peak-to-peak) and at a frequency centred around 9.2 Hz, i.e. the known approximate readout frequency (see Methods). A set of

averaged ssPERGs from the same participant is given in Figure 2B for both check sizes, for both conditions (IO-Reader_{ON} and IO-Reader_{OFF}), and both eyes (with and without sensor implant for the right and left eye, respectively). Significant responses ($P < 0.01$) are evident at the stimulation frequency (15Hz) for all ssPERGs obtained, as underlined by the peak in the spectrum at 15 Hz. For the IO-Reader_{ON} condition, a second peak is observed at around 9 Hz, i.e. corresponding to the read-out frequency (see Methods and Results above). It is also evident that there is an overspill of this intrusion to neighboring frequencies³⁷. This can be attributed to slight variations in the read-out frequency and in an ssPERG averaging that is not locked to the read-out. These intrusions are expected to have minimal effect on the amplitude obtained at the stimulation frequency of 15 Hz, as it is not a directly neighboring frequency, and as noise estimates allow for the subtraction of the noise from the actual response amplitude³⁸.

To test this directly, we performed a quantitative comparison of ssPERG-amplitudes for IO-Reader_{ON/OFF} in a total of 5 participants with an eyemate-IO implant and in an additional 8 controls without the implant, but use of the IO-Reader (see Methods). For this purpose we analyzed the effect of the conditions IO-Reader_{ON/OFF} on ssPERG-amplitudes for both check sizes, i.e. 0.8° and 15° as depicted in Figure 3. On average only small trends were observed, which did not reach significance with T-test statistical testing for neither eye nor checksizes as shown in Table 2 for the right eyes of GLA_{IMP} group and of controls with attached eyemate-IO sensor/reader antenna. Taken together, these results indicate a lack of relevant impact of the eyemate-IO sensor on ssPERG recordings.

Experiment 2 – Influence of postural change on IOP

We investigated the effect of posture (lateral decubitus vs. sitting) on IOP and PERG amplitudes as depicted in Figure 4. We were particularly interested in these effects in the glaucomatous eyes of the participants with simultaneous IOP and PERG-readout, i.e. the right eye of the GLA_{IMP} group (Figure 4 A). For GLA_{IMP}, the IOP increased for LD_R vs S1 by 5.1 mmHg ($P < 0.001$), and for LD_L vs S2 by 2.7 mmHg ($P = 0.01$; Table 3). In CON_{RE}, the IOP increased by 1.5 mmHg for LD_R vs S1 ($P < 0.05$) and for LD_L vs S2 ($P < 0.05$). In summary, effects of posture on IOP were evident for both groups and particularly pronounced for LD_R in GLA_{IMP} group, i.e. IOP increase in the lower eye.

The above findings might be due either to the recorded right eye being the lower in LD_R, alternatively it could be a sequential effect, as LD_R was the 2nd and LD_L the 4th condition. The former is clearly supported by Figure 4B, where the simultaneously recorded data from the left eye are depicted. Here, the IOP reductions were significant in controls for the lower eye during LDP. In CON_{LE}, there was a significant increase of IOP during LD_L vs S2 by 3.5 mmHg ($P < 0.01$), while a trend for increased IOP for LD_R vs S1 did not reach significance (Figure 4B, Table 4). As the left eye of the GLA_{IMP}-group was not glaucomatous for all subjects, we confirmed this effect in the GLA_{LE}-group with glaucomatous left eyes (Figure 4 C and Table 3). In GLA_{LE}, the statistical power of our design was not sufficient to resolve a significant difference of IOPs of left eye between sitting and LDP.

Experiment 2 – Influence of postural change on PERG

For GLA_{IMP} right eye (Figure 4A) 0.8° and 15° PERG-amplitudes decreased [by 0.24 μ V ($P = 0.001$) and 0.16 μ V ($P = 0.02$), respectively] for LD_R vs S1, but not for LD_L vs S2. Similarly, for CON_{RE}, 0.8° and 15° PERG amplitudes decreased for LD_R vs S1 by 0.31 μ V ($P = 0.001$) and by 0.14 μ V ($P = 0.005$), respectively, but not for LD_L vs S2 (Table 5). Similar effects, but as expected for LD_L vs S2, were evident for CON_{LE} as a significant amplitude decrease by 0.29 μ V ($P < 0.001$) and 0.11 μ V ($P = 0.010$) for 0.8° and 15° PERG-amplitudes, respectively, and for GLA_{LE} (Figure 4 C) as a significant amplitude decrease, which reached significance only for 0.8° ($P < 0.001$; Table 6).

Taken together, LDP induced an IOP increase and PERG decrease predominantly in the lower eye.

Discussion

In the present study we demonstrated that ssPERG recordings are feasible during simultaneous continuous IOP measurements with the eyemate-IO sensor. This opens novel directions to investigate the relationship of IOP and ganglion-cell function. Our findings corroborate the hypothesis that LDP-induced IOP increase in the lower eye is accompanied by reduced ganglion response as determined with ssPERG, both in control and glaucoma eyes.

We observed in the LDP a significantly higher IOP increase in the lower eye compared to the reference condition, i.e. sitting, in both controls and GLA_{IMP}. This is in accordance with previous studies, that demonstrated LDP induced IOP increase of the lower eye in comparison to either sitting and/or supine postures in either healthy subjects^{10,11,40} or untreated glaucoma patients^{14,41}. The GLA_{IMP}-group showed higher IOP fluctuations and rise than healthy subjects during postural change from sitting to LDP, which is in line with other studies that reported that higher LDP- or supine-induced IOP changes were associated with more progression or asymmetrical visual field defects of untreated and treated glaucoma patients^{13,14,42–45}; however, others did not report a relationship of LDP-induced changes and glaucoma progression or laterality^{15,46}. Our GLA_{LE}-group showed no significant IOP change during LDP vs sitting. It should be noted, however, that in the present study the IOP measurements were performed with different devices in the GLA_{LE}- and control group compared to the GLA_{IMP}-group. Further, a study comparing the LDP-induced IOP changes in open and closed angle glaucoma and healthy controls did not report significant differences in IOP positional changes between the groups⁴⁷. It is difficult to compare our findings to the aforementioned studies because of different methodologies used.

We observed a reversible reduction of ganglion cell responses as reflected by the ssPERG for LDP, especially in the lower eye. This is supported by a previous report on the effect of recumbent posture the ssPERG after 8 minutes of head-down body tilt (HDT) of -10° in a subpopulation of glaucoma suspects and early manifest glaucoma patients^{8,9}. It was also shown in a longitudinal study that glaucoma suspects at higher risk for glaucoma conversion had more susceptibility to moderate HDT changes⁹. In contrast, to the authors' knowledge, there is no prior study characterizing electrophysiological changes during LDP. As LDP does, in contrast to supine posture, not require adapted stimulation set-ups, future studies using LDP might therefore open the possibility of a widespread use of the PERG-recordings during posture changes.

Pathophysiology of IOP change in lateral decubitus

It was hypothesized that higher IOP in the lower eye during recumbent posture might be due to gravity and subsequent episcleral venous pressure changes⁴⁸. In fact, biomechanical effects at the optic nerve head caused by choroidal vascular

congestion and altered IOP and cerebrospinal fluid pressure gradient and increased IOP may contribute to PERG changes^{49,50}. However, ocular perfusion pressure was not measured in the present study. It was also shown that artificial IOP elevation decreased the PERG signal in controls and ocular hypertension subjects^{51,52}.

Practical considerations and potential applications

We provide proof-of-concept for telemetric IOP measurement during steady-state PERG recordings. The robustness of the PERG recordings is strongly related to the frequency-based amplitude measurements as this allows for the dissociation of the intrusions around 9.2 Hz from the stimulation related response at 15 Hz. As a matter of course, for a closer spacing of stimulation and readout frequencies (or their harmonics) than in the present study read-out intrusions might influence PERG measurements. As another consequence of the requirement of a frequency-related analysis to avoid reader intrusions, transient PERG recordings cannot be performed with simultaneous readout, unless other measures are taken, e.g. substantially longer averaging. Since we demonstrated LDP induced-IOP and -ganglion cell changes, it is of great promise to study continuous IOP variations during sleep position and in glaucoma patients and studying the feasibility of implementing LDP as a provocative test in glaucoma suspects in which those at risk of glaucoma conversions may show more postural induced IOP and ganglion cell responsiveness changes.

Limitations and potential confounds

Potential confounds related to electrode placement, participant positioning and optics as well as other limitations were addressed. Artificial PERG changes induced by misalignment of the glasses or by electrode displacement during LDP were addressed by careful positioning. Checking the effect of astigmatism, i.e. the dependence of PERG amplitude changes on astigmatism, we did not find any significant correlations. Small sample size, due to the rarity of patients with an IOP-sensor implant, was another limiting factor in our study. Further, it should be noted the current observations may not reflect the real changes of IOP during sleep due to various physiological and environmental conditions that might additionally influence IOP.

Conclusion

In this study, we demonstrate that continuous IOP monitoring with the eyemate-IO sensor can be used for simultaneous IOP measurements and ssPERG recordings. We report reduced ssPERG in the lower eye during lateral decubitus positioning and thus demonstrate the relation of IOP changes on retinal ganglion cell function. This opens the possibility to perform investigations to scrutinize the relationship of IOP and ganglion-cell function as tested in the present study. Further studies with bigger sample size are required to detail the observed effects and their relationship to glaucoma asymmetry and progression.

Acknowledgements

This work was supported by European Union's Horizon 2020 research and innovation programme under the Marie Skłodowska-Curie grant agreement (No. 675033) to Michael B. Hoffmann and Lars Choritz.

Conflict of Interest Statement – None

References

1. Quigley HA, Broman AT. The number of people with glaucoma worldwide in 2010 and 2020. *Br J Ophthalmol*. 2006;90(3):262-267. doi:10.1136/bjo.2005.081224
2. Pascolini D, Mariotti SP. Global estimates of visual impairment: 2010. *Br J Ophthalmol*. 2012;96(5):614-618. doi:10.1136/bjophthalmol-2011-300539
3. Tham Y-C, Li X, Wong TY, Quigley HA, Aung T, Cheng C-Y. Global Prevalence of Glaucoma and Projections of Glaucoma Burden through 2040: A Systematic Review and Meta-Analysis. *Ophthalmology*. 2014;121(11):2081-2090. doi:10.1016/j.ophtha.2014.05.013
4. Bach M, Hoffmann M. The origin of the Pattern Electroretinogram. In: Heckenlively JR, Arden GB, eds. *Principles and Practice of Clinical Electrophysiology of Vision*. 2nd ed. MIT Press; 2006:185-196.
5. Bach M, Cuno A-K, Hoffmann MB. Retinal conduction speed analysis reveals different origins of the P50 and N95 components of the (multifocal) pattern electroretinogram. *Exp Eye Res*. 2018;169:48-53. doi:10.1016/j.exer.2018.01.021
6. Porciatti V, Ventura LM. Normative data for a user-friendly paradigm for pattern electroretinogram recording. *Ophthalmology*. 2004;111(1):161-168. doi:10.1016/j.ophtha.2003.04.007
7. Bach M, Hoffmann MB. Update on the Pattern Electroretinogram in Glaucoma. *Optom Vis Sci*. 2008;85(6):386. doi:10.1097/OPX.0b013e318177ebf3
8. Ventura LM, Golubev I, Lee W, et al. Head-down Posture Induces Perg Alterations in Early Glaucoma. *J Glaucoma*. 2013;22(3):255-264. doi:10.1097/IJG.0b013e318232973b
9. Porciatti V, Feuer WJ, Monsalve P, et al. Head-down Posture in Glaucoma Suspects Induces Changes in Iop, Systemic Pressure, and Perg That Predict Future Loss of Optic Nerve Tissue. *J Glaucoma*. 2017;26(5):459-465. doi:10.1097/IJG.0000000000000648
10. Lee JY, Yoo C, Jung JH, Hwang YH, Kim YY. The effect of lateral decubitus position on intraocular pressure in healthy young subjects. *Acta Ophthalmol (Copenh)*. 2012;90(1):e68-72. doi:10.1111/j.1755-3768.2011.02208.x
11. Malihi M, Sit AJ. Effect of Head and Body Position on Intraocular Pressure. *Ophthalmology*. 2012;119(5):987-991. doi:10.1016/j.ophtha.2011.11.024
12. Seo H, Yoo C, Lee T-E, Lin S, Kim YY. Head position and intraocular pressure in the lateral decubitus position. *Optom Vis Sci Off Publ Am Acad Optom*. 2015;92(1):95-101. doi:10.1097/OPX.0000000000000432

13. Kim KN, Jeoung JW, Park KH, Lee DS, Kim DM. Effect of lateral decubitus position on intraocular pressure in glaucoma patients with asymmetric visual field loss. *Ophthalmology*. 2013;120(4):731-735. doi:10.1016/j.ophtha.2012.09.021
14. Lee JY, Yoo C, Kim YY. The effect of lateral decubitus position on intraocular pressure in patients with untreated open-angle glaucoma. *Am J Ophthalmol*. 2013;155(2):329-335.e2. doi:10.1016/j.ajo.2012.08.003
15. Lee T-E, Yoo C, Lin SC, Kim YY. Effect of Different Head Positions in Lateral Decubitus Posture on Intraocular Pressure in Treated Patients With Open-Angle Glaucoma. *Am J Ophthalmol*. 2015;160(5):929-936.e4. doi:10.1016/j.ajo.2015.07.030
16. Collins CC. Miniature passive pressure transensor for implanting in the eye. *IEEE Trans Biomed Eng*. 1967;14(2):74-83.
17. Leonardi M, Leuenberger P, Bertrand D, Bertsch A, Renaud P. First Steps toward Noninvasive Intraocular Pressure Monitoring with a Sensing Contact Lens. *Invest Ophthalmol Vis Sci*. 2004;45(9):3113-3117. doi:10.1167/iovs.04-0015
18. Aptel F, Musson C, Zhou T, Lesoin A, Chiquet C. 24-hour Intraocular Pressure Rhythm in Patients With Untreated Primary Open Angle Glaucoma and Effects of Selective Laser Trabeculoplasty. *J Glaucoma*. 2017;26(3):272-277. doi:10.1097/IJG.0000000000000604
19. Tojo N, Abe S, Ishida M, Yagou T, Hayashi A. The Fluctuation of Intraocular Pressure Measured by a Contact Lens Sensor in Normal-tension Glaucoma Patients and Nonglaucoma Subjects. *J Glaucoma*. 2017;26(3):195-200. doi:10.1097/IJG.0000000000000517
20. Melki S, Todani A, Cherfan G. An Implantable Intraocular Pressure Transducer: Initial Safety Outcomes. *JAMA Ophthalmol*. 2014;132(10):1221-1225. doi:10.1001/jamaophthalmol.2014.1739
21. Koutsonas A, Walter P, Roessler G, Plange N. Implantation of a Novel Telemetric Intraocular Pressure Sensor in Patients With Glaucoma (ARGOS Study): 1-Year Results. *Invest Ophthalmol Vis Sci*. 2015;56(2):1063-1069. doi:10.1167/iovs.14-14925
22. Koutsonas A, Walter P, Plange N. Selbsttonometrie mit einem telemetrischen, intraokularen Drucksensor bei Patienten mit Glaukom. *Klin Monatsblätter Für Augenheilkd*. 2016;233(6):743-748. doi:10.1055/s-0041-106191
23. Koutsonas A, Walter P, Roessler G, Plange N. Long-term follow-up after implantation of a telemetric intraocular pressure sensor in patients with glaucoma: a safety report. *Clin Experiment Ophthalmol*. 2018;46(5):473-479. doi:10.1111/ceo.13100
24. Choritz L, Mansouri K, van den Bosch J, et al. Telemetric Measurement of Intraocular Pressure via an Implantable Pressure Sensor—12-Month Results from the ARGOS-02 Trial. *Am J Ophthalmol*. September 2019. doi:10.1016/j.ajo.2019.09.011
25. Enders P, Hall J, Bornhauser M, et al. Telemetric Intraocular Pressure Monitoring after Boston Keratoprosthesis Surgery. *Ophthalmology*. 2019;126(2):322-324. doi:10.1016/j.ophtha.2018.09.028

26. Ferris FL, Kassoff A, Bresnick GH, Bailey I. New visual acuity charts for clinical research. *Am J Ophthalmol*. 1982;94(1):91-96.
27. Bach M, Mathieu M. Different effect of dioptric defocus vs. light scatter on the Pattern Electroretinogram (PERG). *Doc Ophthalmol*. 2004;108(1):99-106. doi:10.1023/B:DOOP.0000018415.00285.56
28. Anderson DR, Patella VM. *Automated Static Perimetry*. Mosby; 1999.
29. Todani A, Behlau I, Fava MA, et al. Intraocular Pressure Measurement by Radio Wave Telemetry. *Invest Ophthalmol Vis Sci*. 2011;52(13):9573-9580. doi:10.1167/iovs.11-7878
30. Koutsonas A, Walter P, Kuerten D, Plange N. Automated, Noncontact Intraocular Pressure Home Monitoring after Implantation of a Novel Telemetric Intraocular Pressure Sensor in Patients with Glaucoma: A Feasibility Study. *BioMed Res Int*. 2018;2018. doi:10.1155/2018/4024198
31. Bach M. Bach – Freiburg Evoked Potentials. <http://localhost:4000/ep2000/>. Accessed January 14, 2019.
32. Bach M, Brigell MG, Hawlina M, et al. ISCEV standard for clinical pattern electroretinography (PERG): 2012 update. *Doc Ophthalmol*. 2013;126(1):1-7. doi:10.1007/s10633-012-9353-y
33. Bach M, Ramharter-Sereinig A. Pattern electroretinogram to detect glaucoma: comparing the PERGLA and the PERG Ratio protocols. *Doc Ophthalmol Adv Ophthalmol*. 2013;127(3):227-238. doi:10.1007/s10633-013-9412-z
34. Atienza M, Cantero JL, Dominguez-Marin E. The time course of neural changes underlying auditory perceptual learning. *Learn Mem Cold Spring Harb N*. 2002;9(3):138-150. doi:10.1101/lm.46502
35. Guideline 5: Guidelines for Standard Electrode Position Nomenclature. *J Clin Neurophysiol*. 2006;23(2):107-110.
36. Beasley IG, Laughton DS, Coldrick BJ, Drew TE, Sallah M, Davies LN. Does Rebound Tonometry Probe Misalignment Modify Intraocular Pressure Measurements in Human Eyes? *J Ophthalmol*. 2013;2013. doi:10.1155/2013/791084
37. Bach M, Meigen T. Do's and don'ts in Fourier analysis of steady-state potentials. *Doc Ophthalmol*. 1999;99(1):69-82. doi:10.1023/A:1002648202420
38. Meigen T, Bach M. On the statistical significance of electrophysiological steady-state responses. *Doc Ophthalmol*. 1999;98(3):207-232. doi:10.1023/A:1002097208337
39. Holm S. A Simple Sequentially Rejective Multiple Test Procedure. *Scand J Stat*. 1979;6(2):65-70.
40. Lee T-E, Yoo C, Kim YY. Effects of different sleeping postures on intraocular pressure and ocular perfusion pressure in healthy young subjects. *Ophthalmology*. 2013;120(8):1565-1570. doi:10.1016/j.ophtha.2013.01.011

41. Piven I, Glovinsky Y. Intraocular pressure curves of untreated glaucoma suspects and glaucoma patients in sitting and lateral decubitus positions using the goldmann applanation tonometer. *J Glaucoma*. 2014;23(8):541-546. doi:10.1097/IJG.0000000000000084
42. Hirooka K, Shiraga F. Relationship between postural change of the intraocular pressure and visual field loss in primary open-angle glaucoma. *J Glaucoma*. 2003;12(4):379-382.
43. Kiuchi T, Motoyama Y, Oshika T. Relationship of progression of visual field damage to postural changes in intraocular pressure in patients with normal-tension glaucoma. *Ophthalmology*. 2006;113(12):2150-2155. doi:10.1016/j.opthta.2006.06.014
44. Kiuchi T, Motoyama Y, Oshika T. Postural response of intraocular pressure and visual field damage in patients with untreated normal-tension glaucoma. *J Glaucoma*. 2010;19(3):191-193. doi:10.1097/IJG.0b013e3181a98f30
45. Park J-H, Yoo C, Yoo E, Kim YY. Intraocular Pressure Elevation during Lateral Body Posture in Side-sleeping Glaucoma Patients. *Optom Vis Sci Off Publ Am Acad Optom*. 2019;96(1):62-70. doi:10.1097/OPX.0000000000001322
46. Kaplowitz K, Blizzard S, Blizzard DJ, et al. Time Spent in Lateral Sleep Position and Asymmetry in Glaucoma. *Invest Ophthalmol Vis Sci*. 2015;56(6):3869-3874. doi:10.1167/iovs.14-16079
47. Sawada A, Yamamoto T. Posture-Induced Intraocular Pressure Changes in Eyes with Open-Angle Glaucoma, Primary Angle Closure with or without Glaucoma Medications, and Control Eyes. *Invest Ophthalmol Vis Sci*. 2012;53(12):7631-7635. doi:10.1167/iovs.12-10454
48. Friberg TR, Sanborn G, Weinreb RN. Intraocular and episcleral venous pressure increase during inverted posture. *Am J Ophthalmol*. 1987;103(4):523-526. doi:10.1016/s0002-9394(14)74275-8
49. Longo A, Geiser MH, Riva CE. Posture changes and subfoveal choroidal blood flow. *Invest Ophthalmol Vis Sci*. 2004;45(2):546-551. doi:10.1167/iovs.03-0757
50. Downs JC, Roberts MD, Burgoyne CF. Mechanical environment of the optic nerve head in glaucoma. *Optom Vis Sci Off Publ Am Acad Optom*. 2008;85(6):425-435. doi:10.1097/OPX.0b013e31817841cb
51. Kremmer S, Tolksdorf-Kremmer A, Stodtmeister R. [Simultaneous registration of VECF and pattern ERG during artificially raised intraocular pressure]. *Ophthalmol J Int Ophthalmol Int J Ophthalmol Z Augenheilkd*. 1995;209(5):233-241. doi:10.1159/000310622
52. Colotto A, Falsini B, Salgarello T, Buzzonetti L, Cermola S, Porrello G. Transiently raised intraocular pressure reveals pattern electroretinogram losses in ocular hypertension. *Invest Ophthalmol Vis Sci*. 1996;37(13):2663-2670.

Figure 1

Figure 2

Figure 3

Figure 4

Figure legends

Figure 1. Recording setup for a GLA_{IMP} patient with binocular electrodes for PERG recordings and the antenna around the right eye for simultaneous continuous read-out of IOP responses in Experiments 1 & 2. (Permission obtained from patient)

Figure 2. Effect of IO-readout on PERG recordings. (A) Raw PERG recording trace for eyemate-IO sensor on/off states. For IO-Reader_{ON} (top trace) the applied electromagnetic field induces a noise intrusion of around 9 Hz, i.e. at the read-out frequency (see Methods). For IO-Reader_{OFF} (bottom trace) the noise intrusion is absent. (B) ssPERGs and frequency spectra for IO-Reader_{ON/OFF} -states for both eyes for two check sizes (0.8° and 15°). Noise intrusions are reduced due to non-phase locked averaging, but still evident for both the eye with eyemate-IO sensor and -Reader and for the fellow eye. As a consequence, a response at around 9.2 Hz is evident in the frequency spectra for IO-Reader_{ON}, in addition to the stimulus evoked response at 15 Hz.

Figure 3.

PERG amplitudes for 0.8° and 15° checksizes and IO-Reader_{ON} and IO-Reader_{OFF} in GLA_{IMP} (n=5) and controls (n=8). No significant differences between IO-Reader_{ON} and IO-Reader_{OFF} were evident. Average data $\pm$ SEM (red) and single subject effects are depicted. GLA_{IMP}: Glaucoma patients with eyemate-IO implant.

Figure 4. Mean $\pm$ SEM IOP (top row) and Mean $\pm$ SEM ssPERG data (0.8° checksize: middle row; 15° checksize: bottom row) for different postures (S1= sitting 1; LD_R= right lateral decubitus; S2= sitting 2; LD_L= left lateral decubitus; S3= sitting 3). T-test were performed for control right eyes (CON_{RE}) vs GLA_{IMP}-group' right eyes and for control left eyes (CON_{LE}) (A and B) vs left eyes GLA_{LE}-group (C) and for conditions S1 vs LD_R and S2 vs LD_L. * P value: < 0.05; ** P value < 0.01; *** P value < 0.001.

Table 1. Summary of participant characteristics.

Participant	Sex	Age [years]	Decimal visual acuity		Visual field – MD [dB]	
			OD	OS	OD	OS
GS01 _{1,2}	M	70	0.4	0.6	-4.00	-1.26
GS02 _{1,2}	F	76	1.0	0.7	-6.97	-3.17
GS03 _{1,2}	M	68	1.0	1.0	1.11	.66
GS04 _{1,2}	M	72	1.0	1.0	-3.25	-3.08
GS05 _{1,2}	F	62	0.8	0.8	-.99	-3.67
GS06 ₂	M	77	0.9	1.0	-3.72	-2.01
GS07 ₂	M	67	1.0	1.0	-3.29	-4.11
GS08 ₂	F	77	1.0	1.0	-1.16	.16
C01 _{1,2}	F	60	1.25	1.25	-1.34	-0.64
C02 ₁	M	65	0.8	1.0	-0.7	0.7
C03 _{1,2}	F	57	1.25	1.25	-1.10	0.46
C04 _{1,2}	M	68	1.6	1.6	-1.0	0.68
C05 ₁	M	56	1.25	1.25	-0.10	0.87
C06 ₁	F	57	1.0	1.0	-0.64	-0.22
C07 ₁	M	52	1.0	1.0	-0.85	0.78
C08 _{1,2}	M	63	1.0	1.0	0.57	1.02
C09 ₂	F	33	1.25	1.25	-0.7	-0.7
C10 ₂	F	34	1.25	1.25	-0.72	-1.08
C11 ₂	F	33	1.25	1.00	-0.5	-0.68
C12 ₂	M	58	1.25	1.25	-0.98	0.51
C13 ₂	F	56	1.00	1.00	-1.02	-0.29
C14 ₂	F	54	1.00	1.25	1.27	-1.66
C15 ₂	F	65	1.00	1.00	0.56	-0.79
G01 ₂	M	53	1.60	1.25	-.30	-9.67
G02 ₂	M	34	1.00	1.00	-1.29	-2.46
G03 ₂	F	35	1.00	1.00	-2.93	-.24
G04 ₂	M	78	.80	.63	.49	-16.95
G05 ₂	F	59	1.00	1.00	-1.05	-5.03
G06 ₂	F	51	1.00	1.00	-2.86	-2.03
G07 ₂	F	50	1.00	.80	-1.63	-2.92
GS= Glaucoma patients with eyemate-IO sensor; C= Control participants; G= Glaucoma patients without sensors						
₁ Participants in experiment 1; ₂ Participants in experiment 2						

Table 2. PERG amplitudes comparisons between the IO-Reader_{Off/On} states in right eye of controls with attached eyemate-IO sensor and right eye of glaucoma patients with eyemate-IO implant.

Check	IO-Reader	Controls with attached eyemate-IO-sensor (N=8)				GLA _{IMP} (N=5)			
		M. amplitude ± SE	M. diff. ± SE	t(7)	P	M. amplitude ± SE	M. diff. ± SE	t(4)	P
0.8°	Off	.86 ± 0.10	-.003 ± 0.02	-1.16	.88	.40 ± 0.07	-.01 ± 0.02	-3.37	.73
	On	.86 ± 0.11				.41 ± 0.08			
15°	Off	.77 ± 0.04	-.03 ± 0.08	-0.81	.45	.57 ± 0.07	-.05 ± 0.04	-1.15	.31
	On	.80 ± 0.04				.62 ± 0.07			

GLA_{IMP}= eyemate-IO sensor patients' right eye, M.= mean; diff.= difference; SE= standard error of mean;

Table 3. Intraocular pressure of right eye during different body postures

Posture	CON _{RE} (N=11)				GLA _{IMP} (N=8)			
	Mean IOP ± SE	M. diff. ± SE	t(10)	P	Mean IOP ± SE	M. diff. ± SE	t(7)	P
S1	15.36 ± 1.04	-1.55 ± 0.58	-2.68	0.023	15.98 ± 1.79	-5.14 ± 0.57	-8.96	< .001
LD _R	16.91 ± 0.89				21.12 ± 1.57			
S2	14.00 ± 1.10	-1.45 ± 0.62	-2.33	0.04	16.97 ± 1.72	-2.71 ± 0.78	-3.47	0.01
LD _L	15.45 ± 1.23				19.68 ± 1.91			

CON_{RE}= Controls' right eye, GLA_{IMP}= eyemate-IO sensor patients' right eye, M. diff.= mean difference; SE= standard error of mean; IOP= intraocular pressure; S1& S2= sitting 1& 2; LD_R = right lateral decubitus posture; LD_L = left lateral decubitus posture

Table 4. Intraocular pressure of left eye during different body postures

Posture	CON _{LE} (N=11)				GLA _{LE} (N=7)			
	Mean IOP ± SE	M. diff. ± SE	t(10)	P	Mean IOP ± SE	M. diff. ± SE	t(6)	P
S1	15.09 ± 0.80	-0.91 ± 0.59	-1.53	0.16	18.00 ± 1.65	-0.71 ± 0.89	-0.8	0.45
LD _R	16.00 ± 1.10				18.71 ± 1.95			
S2	13.73 ± 1.03	-3.45 ± 0.88	-3.94	.003	17.86 ± 1.50	-0.86 ± 0.59	-1.44	0.20
LD _L	17.18 ± 1.40				18.71 ± 1.63			

CON_{LE}= Controls' left eye, GLA_{LE}= Left eye glaucoma patients, M.= mean; diff.= difference; SE= standard error of mean; IOP= intraocular pressure; S1& S2= sitting 1& 2; LD_R = right lateral decubitus posture; LD_L = left lateral decubitus posture

Table 5. PERG amplitude changes of right eye during different body postures

Check	Posture	CON _{RE} (N=11)				GLA _{IMP} (N=8)			
		M. Amplitude ± SE	M. diff. ± SE	t(10)	P	M. Amplitude ± SE	M. diff. ± SE	t(7)	P
0.8°	S1	0.10 ± 0.10	0.31 ± 0.07	4.58	0.001	0.49 ± 0.03	0.24 ± 0.04	5.78	0.001
	LD _R	0.69 ± 0.06				0.25 ± 0.04			
15°	S1	0.84 ± 0.04	0.14 ± 0.04	3.59	0.005	0.63 ± 0.04	0.16 ± 0.06	2.87	0.02
	LD _R	0.70 ± 0.04				0.47 ± 0.05			
0.8°	S2	0.93 ± 0.08	0.06 ± 0.04	1.67	0.12	0.33 ± 0.04	-0.01 ± 0.04	-0.13	0.89
	LD _L	0.87 ± 0.08				0.34 ± 0.06			
15°	S2	0.81 ± 0.05	0.03 ± 0.04	0.74	0.47	0.55 ± 0.05	0.02 ± 0.06	0.29	0.78
	LD _L	0.78 ± 0.05				0.53 ± 0.03			

CON_{RE}= Controls' right eye, GLA_{IMP}= eyemate-IO sensor patients' right eye, M.= mean; diff.= difference; SE= standard error of mean; S1& S2= sitting 1& 2; LD_R = right lateral decubitus posture; LD_L = left lateral decubitus posture

Table 6. PERG amplitude changes of left eye during different body postures

Check	Posture	CON _{LE} (N=11)				GLA _{LE} (N=7)			
		M. Amp.± SE	M. diff. ± SE	t(10)	P	M. Amp.± SE	M. diff. ± SE	t(6)	P
0.8°	S1	1.04 ± 0.09				0.82 ± 0.20			
	LD _R	0.97 ± 0.07	0.07 ± 0.05	1.318	0.217	0.81 ± 0.17	0.01 ± .09	0.12	0.91
15°	S1	0.82 ± 0.05				0.85 ± 0.13			
	LD _R	0.82 ± 0.04	-0.002 ± .04	-0.065	0.949	0.75 ± 0.07	0.10 ± .07	1.48	0.19
0.8°	S2	1.01 ± 0.07				0.80 ± 0.17			
	LD _L	0.71 ± 0.06	0.29 ± 0.06	5.167	< .001	0.60 ± 0.18	0.20 ± .03	7.31	< .001
15°	S2	0.83 ± 0.05				0.75 ± 0.13			
	LD _L	0.71 ± 0.04	0.11 ± 0.04	3.180	0.010	0.63 ± 0.08	0.11 ± .09	1.31	0.24

CON_{LE}= Controls' left eye, GLA_{LE}= Left eye glaucoma patients, M.= mean; Amp.= amplitude; diff.= difference; SE= standard error of mean; S1& S2= sitting 1& 2; LD_R = right lateral decubitus posture; LD_L = left lateral decubitus posture