1

2 **An ultrasensitive, rapid, and portable coronavirus SARS-CoV-2 sequence detection**

3 **method based on CRISPR-Cas12**

4 Curti Lucia¹, Pereyra-Bonnet Federico^{1,2} and Gimenez Carla Alejandra^{2,*}

5 ¹INPA-National Scientific and Technical Research Council (CONICET)- Argentina-

6 University of Buenos Aires, Argentina

7 ²CASPR Biotech, San Francisco California, USA

8

9 *Corresponding author: Gimenez Carla Alejandra, BSc, PhD

10 CASPR Biotech, 479 Jessie, San Francisco, CA 94103, USA

11 E-mail: carla@caspr.bio

12 Telephone: +54 9-(11)-57344671

13

14 **Short running title:** Coronavirus sequence detection by CRISPR-Cas12

15

16 **Keywords:** CRISPR-Cas12, emerging virus, SARS-CoV2, COVID-19, 2019-nCoV,

17 diagnosis.

18

19 **Abstract**

20

21 Severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) has received global
22 attention due to the recent outbreak in China. In this work, we report a CRISPR-Cas12 based
23 diagnostic tool to detect synthetic SARS-CoV-2 RNA sequences in a proof-of-principle
24 evaluation. The test proved to be sensitive, rapid, and potentially portable. These key traits of
25 the CRISPR method are critical for virus detection in regions that lack resources to use the
26 currently available methods.

27

28 **Introduction**

29

30 The rapid spread of the novel coronavirus is clearly a major concern for countries across the
31 world. This virus is a single-stranded RNA virus with high sequence identity to SARS-CoV
32 and has, therefore, been named SARS-CoV-2. This coronavirus variant is capable of
33 transmission from person to person, which makes it a dangerous virus that can cause a
34 pandemic.

35

36 Several SARS-CoV-2 detection assays have been reported to be currently under
37 development. The WHO website provides information on several virus detection protocols
38 that have been used in different countries such as China, Germany, Japan, and the US, among
39 others (1). All these are real-time reverse transcription PCR (rRT-PCR) based assays, and
40 despite their established efficiency, require highly specialized personnel and expensive
41 equipment for implementation. In such a backdrop, any development toward ultrasensitive,
42 cheaper, and portable diagnostic tests for the assessment of suspected cases, regardless of the

43 presence of qualified personnel or sophisticated equipment for virus detection, could help
44 advance the diagnosis of COVID-19.

45

46 CRISPR is a biotechnological technique well-known for its use in gene editing. Notably,
47 CRISPR has been recently used for the *in vitro* detection of nucleic acids, thereby emerging
48 as a powerful and precise tool for molecular diagnosis (2-4). Within the CRISPR-Cas effector
49 family, Cas12 is a RNA-guided DNase belonging to the class II type V-A system that induces
50 indiscriminate single-stranded DNA (ssDNA) collateral cleavage after target recognition.
51 This leads to the degradation of ssDNA reporters that, emit a fluorescence signal on cleavage
52 or alternatively, could be detected on a paper strip (by lateral flow) in a portable manner (5).
53 Therefore, CRISPR-Cas12 based tools possess the potential to emerge as an *in situ* diagnostic
54 tool for rapid detection of the SARS-CoV-2 virus. In this work, we employed CRISPR-
55 Cas12a and its unspecific collateral ssDNase activity to generate a fast, accurate, and
56 portable SARS-CoV-2 sequence detection method.

57

58 **Materials and methods**

59

60 For the detection assays, we included synthetic RNA fragments of SARS-CoV-2
61 corresponding to the *RdRp*, *ORF1b* and *ORF1ab* genes, using the WH-human1 sequence
62 (GenBank MN908947) as a reference. Briefly, the SARS-CoV-2 fragments were synthesized
63 as complementary DNA oligonucleotides and treated with fill-in PCR (NEBNext® High-
64 Fidelity 2X PCR Master Mix) to generate the DNA templates. These DNA templates were
65 transcribed into RNA using an *in vitro* transcription (IVT) Kit (Ambion, Invitrogen) under
66 the control of a T7 promoter.

67

68 A target amplification step was performed using the TwistAmp® Basic recombinase
69 polymerase amplification (RPA) kit (TwistDx, Cambridge, United Kingdom) and RT was
70 carried out by the addition of 5 µL of synthetic SARS-CoV-2-RNA input, 2.5 µL of M-
71 MuLV reverse transcriptase (NEB), and 1 µL of murine RNase inhibitor (NEB) in a 50 µL
72 final reaction volume. Reactions were run for 30 min at 42°C. The RT-RPA was carried out
73 in one step, with the same pair of primers for both reactions.

74

75 To generate the CRISPR-detection complex, we mixed 75 nM of the commercially available
76 LbCas12a endonuclease (NEB) with the same amount of single guide RNA (sgRNA)
77 synthesized in-house by hybridization of DNA oligonucleotides followed by fill-in PCR and
78 IVT. The complexation reaction was carried out in a solution containing 1X NEBuffer 2.1
79 (NEB) (composed of 100 mM NaCl, 50 mM Tris-HCl, 10 mM MgCl₂, 1 mM DTT, pH 7.9)
80 at room temperature for 10 min.

81

82 For plate reader-based assays, ssDNA reporters labeled with FAM were included in the
83 detection mix; while for portable detection with paper strips, ssDNA reporters labeled with
84 biotin were included. The reaction was initiated by diluting the LbCas12a complexes to a
85 final concentration of 37 nM LbCas12a : 37 nM sgRNA in a solution containing 1X
86 NEBuffer 2.1 (NEB) and 1 µM custom ssDNA reporter substrates in a 40 µL reaction
87 volume. A 2 µL aliquot of RT-RPA input was used. Reactions were incubated for up to 90
88 min at 37°C. Fluorescence measurements were acquired at 10 min intervals (λ_{ex} : 485 nm and
89 λ_{em} : 535 nm) in a SpectraMax M2 fluorescence plate reader (Molecular Devices) operated in
90 the 384-well microplate format. For paper-based measurements, we used the Milenia
91 HybriDetect 1 (TwistDx) lateral flow system, as per the manufacturer's instructions. Finally,
92 we simulated a clinical sample by adding synthetic SARS-CoV-2 RNA fragments (10^5

93 copies/ μ L) to saliva samples from a healthy donor and measuring the results on the test strips.
 94 The saliva was treated with heat and chemicals before using 2 μ L as input for RT-RPA (5).
 95
 96 Details about the target sequences, sgRNA, primers for RT-RPA, and ssDNA reporters are
 97 listed in the accompanying table 1. The general scheme of target synthesis, amplification,
 98 detection with CRISPR, and visualization of the result are shown in the figure 1.
 99

101 **Figure 1.** CRISPR-based detection method for novel coronavirus. General scheme of
 102 CRISPR detection procedure. In the two strategies all the process takes less than 60 min.

103

104 **Table. List of DNA oligonucleotides and ssDNA reporters used in this work**

Name	Sequence	Source
nCoV-RdRp-sgRNA R	ATGTGTGGCGGTTCACTATATGTatctacacttagtagaaatta	This paper
nCoV-RdRp-Template PCR fill-in F1	CCCTAATACGACTCACTATAGGCAAGTATTGAGTGAAATGGTCATGTGTGGCGGTTCACT	GenBank MN908947
nCoV-RdRp-Template PCR fill-in R2	GCAGTTGTGGCATCTCCTGATGAGGTTCCACCTGGTTTAACATATAGTGAACCGCCACAC	GenBank MN908947
nCoV-RdRp-Template PCR fill-in F3	AGATGCCACAACCTGCTTATGCTAATAGTGTTTTTAACATTGTCAAGCTGTC	GenBank MN908947
nCoV-RdRp-Primer RT-RPA F	GTGARATGGTCATGTGTGGCGG	Charité, Berlin, Germany
nCoV-RdRp-Primer RT-RPA R	CARATGTTAAASACACTATTAGCATA	Charité, Berlin, Germany
nCoV-ORF1b-sgRNA R	TGGTTGCTTTGTAGTTACCTGTatctacacttagtagaaatta	This paper
nCoV-ORF1b- Template PCR fill-in F1	CCCTAATACGACTCACTATAGGCAATGGGGTTTTACAGGTAACCTACAAAGCAACCATGA	GenBank MN908947
nCoV-ORF1b- Template PCR fill-in R2	TCACAACTAGCTACATGTGCATTACCATGGACTTGACAATACAGATCATGGTTGCTTTGT	GenBank MN908947
nCoV-ORF1b- Template PCR fill-in F3	AGCTAGTTGTGATGCAATCATGACTAGGTGTCTAGCTGTCCACGAGTGCTTTGTTAAGCG	GenBank MN908947
nCoV-ORF1b-Primer RT-RPA F	TGGGGYTTTACRGGTAACCT	Hong Kong University
nCoV-ORF1b-Primer RT-RPA R	AACRCGCTTAACAAAGCACTC	Hong Kong University
nCoV-ORF1ab-sgRNA R	GTCTGTACCGTCTGCGGTATGTGatctacacttagtagaaatta	This paper
nCoV-ORF1ab- Template PCR fill-in F1	CCCTAATACGACTCACTATAGGCCCTGTGGGTTTTACACTTAAAAACACAGTCTGTACCG	GenBank MN908947
nCoV-ORF1ab- Template PCR fill-in R2	GGAGTTGATCACAACCTACAGCCATAACCTTTCCACATACCGCAGACGGTACAGACTGTGT	GenBank MN908947
nCoV-ORF1ab- Template PCR fill-in F3	GTTGTGATCAACTCCGCGAACCCATGCTTCAGTCAGCTGATGCACAATCGTT	GenBank MN908947
nCoV-ORF1ab-Primer RT-RPA F	CCCTGTGGGTTTTACACTTAA	China CDC
nCoV-ORF1ab-Primer RT-RPA R	ACGATTGTGCATCAGCTGA	China CDC
T7-Cas12scaffold F	CCCTAATACGACTCACTATAGGTAATTTCTACTAAGTGTAGAT	This paper
ssDNA biotin	/56-FAM/TTATT/3Bio	Chen <i>et al.</i> , 2018
ssDNA FAM	/56-FAM/TTATT/3IABkFQ	Chen <i>et al.</i> , 2018

105

106

107 **Results and Discussion**

108

109 The best results were obtained in attempts to detect the *ORF1ab* region with primers
110 suggested by the China CDC (6). Based on visualization of these results in a fluorescence
111 spectrophotometer, the limit of detection (LOD) for *ORF1ab* coronavirus sequences was
112 estimated to be up to 10 copies/ μ L (figure 2 panel A of the accompanying figure), which is 4

orders of magnitude lower than the viral load found for the patient reported in Berlin (10^5 copies/ μ L) (7).

When assays were performed using a paper strip, we obtained a similar LOD, demonstrating that the detection system could be rendered portable without loss in sensitivity (figure 2 panel B of the accompanying figure).

Figure 2. A) Assessment of the detection limit of synthetic SARS-CoV-2 by the CRISPR-based assay, inferred from fluorescence values after 30 min. Background subtracted fluorescence represents sample minus control fluorescence without target. Detection of synthetic SARS-CoV-2 RNA using commercially available paper strips (lateral-flow) in buffer (B) or in saliva (C). The optimization in saliva continues in process to obtain more intense signal on paper strip. Microsoft Excel 2016 was used to arrange the data for analysis. GraphPad Prism v.8.1.2 (San Diego, USA) was used to plot the graph. In all cases, data represent mean $\pm$ SD (n=3).

Finally, we simulated a clinical sample by adding synthetic SARS-CoV-2 RNA fragments to saliva samples collected from a healthy person. There are advantages using saliva for the diagnosis of COVID-19. Saliva specimens can be provided by the patient easily without any invasive procedures (8). Positive results for this artificial saliva-sample were obtained using the paper strip-based assay (figure 2 panel C). These latest results indicate that the CRISPR detection system reported here was not inhibited by naturally occurring molecules in the saliva and can reach real saliva specimens concentration (8), therefore, emerging to be a promising method got the rapid and portable detection of clinical cases of COVID-19.

Conclusions

The current rRT-PCR-based COVID-19 diagnostic approaches have been shown to be very efficient and accurate for virus detection. However, the virus can spread to regions where the equipment required to perform rRT-PCR is not available. We demonstrated that the CRISPR-Cas12 based detection method is characterized by a LOD value lower than the minimum levels needed to presently detect the virus in clinical samples. The main advantage of the CRISPR-diagnostic method reported here is its portability and low cost (USD 1-2/reaction) (2). A possible criticism for this study is that the method has not been tested using patient samples. This is because clinical cases have not yet been reported in our region (South America). Nevertheless, our work represents a proof-of-principle study of the usefulness of the CRISPR-Cas12 detection technology that could be deployed worldwide without the requirement of sophisticated instrumentation.

Note

While we were preparing this paper, another two protocols for SARS-CoV-2 detection using CRISPR diagnostics (SHERLOCK, v.20200214; DETECTER v.20200215) was published, confirming the potential of the technique.

157

158 **Disclosure statement**

159 CG and PBF are shareholders in CASPR Biotech.

160

161 **Funding**

162 CASPR Biotech is a privately funded company. CL, PBF are supported by the National

163 Scientific and Technical Research Council (CONICET) of Argentina.

164

165 **Acknowledgments**

166 The authors thanks Guillermo Repizo from IBR-CONICET for his critical suggestion on this
167 manuscript.

168

169 **References**

- 170 1. [https://www.who.int/emergencies/diseases/novel-coronavirus-2019/technical-](https://www.who.int/emergencies/diseases/novel-coronavirus-2019/technical-guidance/laboratory-guidance)
171 [guidance/laboratory-guidance](https://www.who.int/emergencies/diseases/novel-coronavirus-2019/technical-guidance/laboratory-guidance).
- 172 2. Gootenberg JS, Abudayyeh OO, Lee JW, et al. Nucleic acid detection with CRISPR-
173 Cas13a/C2c2.2017 Science.;356(6336):438–442.
- 174 3. Chen JS, Ma E, Harrington LB, et al. CRISPR-Cas12a target binding unleashes
175 indiscriminate single-stranded DNase activity.2018. Science.;360(6387):436–439.
- 176 4. Li SY, Cheng QX, Wang JM, et al. CRISPR-Cas12a-assisted nucleic acid detection.
177 2018. Cell Discov.;4:20.
- 178 5. Myhrvold C, Freije CA, Gootenberg JS, et al. Field-deployable viral diagnostics using
179 CRISPR-Cas13. 2018. Science;360(6387):444-448.
- 180 6. Zhu N, Zhang D, Wang W, et al. A novel coronavirus from patients with pneumonia
181 in China, 2019. N Engl J Med. DOI: 10.1056/NEJMoa2001017.

Coronavirus sequence detection by CRISPR-Cas12

- 182 7. Camilla Rothe, Mirjam Schunk. Transmission of 2019-nCoV Infection from an
183 Asymptomatic Contact in Germany, 2019. N Engl J Med. DOI:
184 10.1056/NEJMc2001468.
- 185 8. Kelvin Kai-Wang To, Owen Tak-Yin Tsang, Cyril Chik-Yan Yip, et al. Consistent
186 Detection of 2019 Novel Coronavirus in Saliva. 2020. Clin Infect Dis, pii:ciaa149.
187