Resource optimization in COVID-19 diagnosis

Taniwaki, S.A^{1*}; Silva, S.O.S¹; Santana-Clavijo, N.F¹; Conselheiro, J. A^{1,2}; Barone, G.T^{1,2}; Menezes, A.A.R²; Pereira, E.S²; Brandão, P.E¹

¹ Laboratory of Applied Molecular Biology and Serology, Department of Preventive Veterinary Medicine and Animal Health, School of Veterinary Medicine, University of São Paulo (USP), São Paulo – SP, Brazil.

² Laboratory of Diagnostics of Zoonosis and Vector-Borne Diseases, Zoonosis Surveillance Division, Health Surveillance Coordination, Municipal Health Secretariat, São Paulo – SP, Brazil.

*Corresponding author: Sueli A. Taniwaki. Laboratory of Applied Molecular Biology and Serology, Department of Preventive Veterinary Medicine and Animal Health, School of Veterinary Medicine, University of São Paulo (USP), Av. Prof. Dr. Orlando M. Paiva, n. 97, CEP 05508-270, São Paulo-SP, Brazil. staniwaki@usp.br

Abstract

The emergence and rapid dissemination worldwide of a novel *Coronavirus* (SARS-CoV-2) results in decrease of swabs availability for clinical samples collection, as well as, reagents for RT-qPCR diagnostic kits considered a confirmatory test for COVID-19 infection. This scenario, showed the requirement of improve de diagnostic capacity, so the aim of this study were to verify the possibility of reducing the reaction volume of RT-qPCR and to test cotton swabs as alternative for sample collection. RT-qPCR volumes and RNA sample concentration were optimized without affecting the sensitivity of assays, using both probe-based and intercalation dyes methods. Although rayon swabs showed better performance, cotton swabs could be used as alternative type for clinical sample collection. COVID-19 laboratory diagnosis is important to isolate and restrict the dissemination of virus, so seek for alternatives to decrease the coast of assays improve the control of disease.

Key-words: SARS-CoV-2, COVID-19, RT-qPCR, cotton swabs, reaction volume

1. Introduction

On December 31, 2019, the World Health Organization (WHO) was notified about cases of severe pneumonia in Wuhan, China (WHO, 2020a). The etiological agent was identified as a coronavirus (*Coronaviridae: Betacoronavirus: Sarbecovirus: Severe acute respiratory syndrome-related coronavirus*) named as SARS-CoV-2 (GORBALENYA et al., 2020; LU et al., 2020a) and the disease denominated as COVID-19 (*Coronavirus disease 2019*) (WHO, 2020b). Confirmed cases are currently peaking in Latin America and, for instance, on June 13th 2020, more than 850,000 cases had been confirmed in Brazil (available in <https://covid.saude.gov.br/>).

The diagnostic test recommended for SARS-CoV-2 is a Real-Time Polymerase Chain Reaction (RT-qPCR), which can detect active infection by detection of viral genome. Collection of only nasopharyngeal samples (1 swab used for both nostrils) or combined with oropharyngeal (1 swab) sample, should be made with synthetic fibers swabs as more reliable due to a lesser interference with the downstream reactions (CDC, 2020a). Protocols with primers and probes designed at the Charité Virology Institute, Berlin (CORMAN et al., 2020) or at the Centers of Disease Control and Prevention (CDC, 2020b; LU et al., 2020b) became one of the standard procedures for SARS-CoV-2 detection.

With the increasing demand for RT-qPCR tests for both patients and population screening, there is an ongoing shortage of both RT-qPCR reagents and rayon swabs; besides, the relatively high costs of these consumables for poor countries pose a barrier for a wider use of testing. This manuscript reports on the optimization of the Charité and the CDC RT-qPCR protocols for SARS-CoV-2 detection regarding concentration and volumes of reagents for both probe and intercalant agent-based platforms, as well as on the substitution of rayon swabs for cotton swabs for sample collection.

2. Materials and Methods and Results

2.1 Probe protocol: optimization of reaction volume and sample concentration

An inactivated SARS-CoV-2 isolate (kindly provided by Prof. Edison L. Durigon, Institute of Biomedical Sciences, University of São Paulo, Brazil) was used for total RNA extraction with QIAamp[®] Viral RNA Mini kit (QIAGEN).

Total RNA concentrations of 10 and 20% for final reactions of 20, 15 and 10 μ L, were tested using AgPath-ID One Step RT-PCR kit (Applied Biosystems[™]) and StepOne[™] Real-Time PCR Systems (Applied Biosystems[™]), with primers and probes for the RdRp and E genes (CORMAN et al., 2020). Amplification conditions were 48 °C/20 min (reverse transcription), 95 °C/10 min (for the activation of the DNA polymerase) followed by 45 cycles at 95 °C/ 10 seconds and 60 °C/ 30 seconds.

Though the different total RNA concentrations resulted in 1 Cq difference, no difference was found for the three final reaction volumes (Table 1).

Table 1 – Cq mean for probe-based RT qPCR for the E and RdRp genes of SARS-CoV-2 with a range of RNA concentration and final reaction volumes.

E gene			RdRp gene		
Final volume (μ L)	RNA (μ L)	Cq mean	Final volume (μ L)	RNA (μ L)	Cq mean
20	4	12,31	20	4	14,60
	2	13,41		2	-
15	3	12,32	15	3	14,58
	1,5	13,39		1,5	15,77
10	2	12,47	10	2	15,00
	1	13,51		1	15,58

Performance of E and RdRp genes of SARS-CoV-2 RT-qPCRs, based on final reaction volume of 10 μ L with 2 μ L of RNA (Table 1), were verified with a relative standard curve built with 10^{-2} to 10^{-8} dilutions of positive RNA control. Respectively for

E and RdRp genes, detection were observed until 10^{-6} (Cq mean= 33.71, E= 91.2% and $r^2= 0.998$) and 10^{-5} dilutions (Cq mean= 31.98, E= 96.82% and $r^2= 0.998$). This lower sensitivity for RdRp gene is possibly a consequence of a lower transcription of the corresponding ORF due to a 3-to-5' attenuation (IRIGOYEN et al., 2016). So, the results showed that the final volume does not interfere with the sensitivity of this assay.

The kits GoTaq[®] probe 1-Step RT-qPCR (Promega) and SuperScript III Platinum One-Step RT-PCR System (ThermoFisher Scientific) were also tested for the final reaction volumes of 15 and 10 μ L, as per manufacturer's instructions, in a 7500 Real-Time PCR Systems (Applied Biosystems[™]), and the same results were found when compared to the AgPath-ID One Step RT-PCR kit (Applied Biosystems[™]).

For the CDC protocol (LU et al., 2020b), 2019-nCoV CDC-qualified Probe and Primer for SARS-CoV-2 kit (Biosearch[™] Technologies), which includes primers and probes for SARS-CoV-2 N gene (N1 and N2) and Human RNase P (endogen control), were tested with final reaction volume of 20 and 10 μ L with 25% concentration of total RNA, using AgPath-ID One Step RT-PCR (Applied Biosystems[™]) and StepOne[™] Real-Time PCR Systems (Applied Biosystems[™]). Same conditions and relative standard curve (10^{-2} to 10^{-8}) were used with E gene reaction (CORMAN et al., 2020) to possibly the comparison between assays performances.

All reactions resulted in adequate efficiencies and r^2 values (Table 2) and detected the targets up to the 10^{-8} dilution (Table 2), though the N2 reaction resulted in a more intense fluorescence for the higher dilutions, allowing an easier visualization of amplification. The difference between the sensitivities of this last test with the previously described for the Charité RdRp and E genes protocol is possibly due to different positive control aliquots used.

Table 2 –Performance for probe-based RT-qPCR assays for N (N1 and N2) and E genes of SARS-CoV-2.

Dilution	Cq mean				
	E	N1		N2	
	10 µL	10 µL	20 µL	10 µL	20 µL
10 ⁻²	18.65	18.59	18.49	18.90	18.35
10 ⁻³	21.71	21.82	21.87	22.35	21.83
10 ⁻⁴	25.10	24.84	25.15	25.34	25.27
10 ⁻⁵	28.12	27.92	28.01	28.52	28.33
10 ⁻⁶	31.32	31.30	31.32	32.35	31.66
10 ⁻⁷	34.40	34.03	34.31	35.09	35.18
10 ⁻⁸	37.79	36.96	38.39	38.26	38.38
Efficiency (%)	106.3	111.0	102.7	103.3	99.7
r²	0.993	0.999	0.998	0.998	0.999

The limit of detection (LOD) was calculated for each assay using 14 replicates of each dilution (1x10⁻⁷, 2x10⁻⁸ and 1x10⁻⁸). The N2 reaction showed slight better results with LOD of 2x10⁻⁸ dilution, while E and N1 assays had LOD of 1x10⁻⁷ dilution (Table 3).

Table 3 – Limit of detection for probe-based RT-qPCR assays for N (N1 and N2) and E genes of SARS-CoV-2.

	1x10 ⁻⁸			2x10 ⁻⁸			1x10 ⁻⁷		
	E	N1	N2	E	N1	N2	E	N1	N2
Positive	5	4	9	6	6	11	14	11	13
percentage	35,71	28,57	64,29	42,9	42,9	78,6	100,0	78,6	92,9
Ct mean	39,06	36,97	37,66	38,81	37,04	37,57	36,37	36,03	36,19
lower Ct	36,74	36,16	36,35	37,19	36,01	35,45	34,12	34,96	35,03
higher Ct	41,53	37,90	38,20	40,64	37,31	39,21	38,21	38,60	38,07
SD	1,94	0,90	0,66	1,46	0,51	1,10	1,09	1,13	0,98

2.2 Intercalating dyes: optimization of primers concentration

For the intercalating dyes system, primers targeting human RNase P, as per the CDC protocol (LU et al., 2020b), were used at 50, 100, 150, 200 and 250 nM each. For E and RdRp genes from the Charité protocol (CORMAN et al., 2020), primers were tested at 100 and 200 nM and 100, 200, 300 and 400 nM, respectively.

Reactions were carried out with *Power SYBR[®] Green RNA to Ct 1-Step kit* (Applied Biosystems[™]) using a StepOne[™] and 7500 Real-Time PCR Systems (Applied Biosystems[™]) equipment, using 3 or 2 µL of RNA for 15 and 10 µL final volume reactions, respectively. Cycling conditions were 48 °C/20 min (reverse transcription), 95 °C/10 min (for the activation of the DNA polymerase) followed by 45 cycles at 95 °C/10 seconds and 60 °C/30 seconds and melt curve analyses (95 °C/15 secs, 60 °C/15 secs and 95 °C/15 secs).

For the RNase P, the 100 nM primers reaction was found as more efficient (E=101.8%; $r^2=0.991$), with a specific peak at 81 °C in the melt curve; at higher primers concentrations, primers dimers were detected in melt curve analysis and at 50 nM there was a loss in sensitivity.

E gene reaction was found as sensitive as the probe-based reaction for this target, being able to detect up to 10^{-6} dilution of positive RNA control with primers at both 100 and 200 nM, with a unique melting peak at 77 °C. At 200 nM primers concentration, though, a higher efficiency was found (E= 99.3 % and $r^2=0.998$), as well as an earlier target detection for the 10^{-6} dilution (Cq mean= 32.53 and 30.07, respectively, for 100 and 200 nM).

The RdRp reaction showed insufficient performance, as, with primers at 200 nM there was a loss in sensitivity, while, at 400 nM the sensitivity was equal to the probe-

based reaction (detection up to 10^{-5} dilution), but efficiency could not be evaluated due to primer dimers formation visualized in melt curve analysis.

2.3 Comparison between cotton and rayon swabs

An inactivated SARS-CoV-2 isolate was 2-fold diluted in sterile saline from 2^{-1} to 2^{-9} in 100 μ L, in duplicate. For each respective dilution, one rayon (Inlab Diagnóstica, Brazil) and one cotton (Labor, Brazil) swab, both with plastic stalks, were dipped into the tube (one swab/tube), being all the content absorbed by the swabs. Undiluted SARS-CoV-2 (2^0) was also included, as well as sterile saline (negative control).

Next, each swab was immersed in a conical tube containing 3 mL sterile saline and kept at 2-8 °C for 72 hours, in order to mirror the clinical flow of clinical samples. All swabs were finally submitted to RNA extraction and tested with the probe-based E gene RT-qPCR protocol.

Rayon swabs showed more consistency results, with detection through all dilutions from 2^0 to 2^{-9} linearly (Table 4), while cotton swabs failed to detect in two dilutions (2^{-5} and 2^{-6}). CDC does not recommend the use of calcium arginate swabs and swabs with wooden shafts, because these material types may contain PCR inhibitors, and indicate only use of synthetic fibers swabs for collection of clinical samples (CDC, 2020a). However, some study showed good performance of cotton swabs for detection of Human norovirus and rhinovirus by RT-qPCR (WARIS et al., 2013; LEE et al., 2018). Both cotton and rayon swabs have the same chemical structure, with O-H groups which form hydrogen bonds with nucleic acids, and showed optimal absorption capacity (above 100 μ L), but low extraction and recovery efficiency (BRUIJNS, 2018). Although the amplification have been failed in some dilutions, in lower and higher

dilution detection was possible, thus cotton swabs could be an alternative to rayon swabs for clinical sample collection.

Tabela 4 – Probe-based RT-qPCR to the E gene of serial dilutions of SARS-CoV-2 sampled with cotton and rayon swabs.

Dilution	Cq mean		Cq difference
	Cotton	Rayon	
2 ⁰	27.98	26.93	1.05
2 ⁻¹	30.78	27.29	3.49
2 ⁻²	29.82	27.57	2.25
2 ⁻³	33.46	27.90	5.55
2 ⁻⁴	35.06	28.47	6.59
2 ⁻⁵	*	30.11	NA
2 ⁻⁶	*	30.82	NA
2 ⁻⁷	33.06	31.96	1.10
2 ⁻⁸	39.96	34.35	5.61
2 ⁻⁹	37.45	37.10	NA

*: not detected; NA= does not apply

3. Conclusions

The consequences of the ongoing shortage of rayon swabs and RT-qPCR reagents on the sampling and testing of COVID-19 suspected persons will surely have a negative impact on the control of the disease due to a lower sampling and a higher sub notification rate. Here, we demonstrated that the decrease of reaction volume not interferes in sensitivity of assays, allowing the diagnostic capacity increase, and a choice of the type of swab must be based in the local availability and the urgency of the sampling and testing.

4. Consolidated final protocols

Once no loss of sensitivity was demonstrated using 10 μ L final volume reactions, the following protocols are suggested for probe-based and intercalating dyes RT-qPCRs for detection of SARS-CoV-2.

Amplification conditions were optimized for both probe and intercalating dye reactions and for different kits and primers sets.

Stage	Step	Temperature	Time
Holding	Reverse transcription	48 °C	20 min
Holding	Polymerase activation	95 °C	10 min
Cycling 45x	Denaturing	95 °C	10 s
	Annealing/Extension	60 °C	30 s
Melting curve*	Denaturing	95 °C	15 s
	Annealing	60 °C	15 s
	Denaturing	95 °C	15 s

*applies only for the intercalating-dye based reactions

- **For probes-based reactions**

SARS-CoV-2 E gene		
Reagent	Final concentration	Volume/sample (μ L)
AgPath-ID Buffer (2x)	1x	5.00
Primer E_Sarbeco_F1 (10 μ M)	400 nM	0.40
Primer E_Sarbeco_R2 (10 μ M)	400 nM	0.40
Probe E_Sarbeco_P1 (10 μ M)	200 nM	0.20
RT mix (25x)	1x	0.40
H2O	to 10 μ L	1.10
Total		7.50
RNA		2.50

AgPath-ID One Step RT-PCR (Applied Biosystems™)

*GoTaq probe 1-Step RT-qPCR kit (Promega) and SuperScript III Platinum One-Step RT-PCR System (ThermoFisher Scientific) were also tested

SARS-CoV2 N1 and N2 and Human RNase P		
Reagent	Final concentration	Volume/sample (μL)
AgPath-ID Buffer (2x)	1x	5.00
Primer/probe mix (50x)	1x	0.75
RT mix (25x)	1x	0.40
H2O	to 10 μL	1.35
Total		7.50
RNA		2.50

AgPath-ID One Step RT-PCR (Applied Biosystems™)

- For intercalating dyes

SARS-CoV-2 E gene		
Reagent	Final concentration	Volume/sample (μL)
SYBR mix	1x	5.00
Primer E_Sarbeco_F1 (10 μM)	200 nM	0.20
Primer E_Sarbeco_R2 (10 μM)	200 nM	0.20
RT mix (125x)	1x	0.08
H2O	to 10 μL	2.02
Total		7.50
RNA		2.50

Power SYBR® Green RNA to Ct 1-Step kit (Applied Biosystems™)

Human RNase P		
Reagent	Final concentration	Volume/sample (μL)
SYBR mix	1x	5.00
Primer RP-F (10 μM)	100 nM	0.10
Primer RP-R (10 μM)	100 nM	0.10
RT mix (125x)	1x	0.08
H2O	to 10 μL	2.72
Total		8.00
RNA		2.00

Power SYBR® Green RNA to Ct 1-Step kit (Applied Biosystems™)

5. Acknowledgments

This work was funded by FAPESP grant number 2017/24907-6, CNPq (Brazilian National Board for Scientific and Technological Development grant number 307291/2017-0 and CAPES (Coordenação de Aperfeiçoamento de Pessoal de Nível Superior, Brasil - Finance Code 001). Authors were grateful to Prof. Edison L. Durigon and his team of Laboratory of Clinical and Molecular Virology, Institute of Biomedical Sciences, University of São Paulo, Brazil, for provided the inactivated SARS-CoV-2 isolate. And we thank Dr. Caroline Cotrim Aires from Laboratory of Diagnostics of Zoonosis and Vector-Borne Diseases, and Dr. Solange Maria de Saboia e Silva from Health Surveillance Coordination for the institutional support and incentive to the development of the study.

6. References:

BRUIJNS, B. B.; TIGGELAAR, R. M.; GARDENIERS, H. The Extraction and Recovery Efficiency of Pure DNA for Different Types of Swabs. **Journal of Forensic Sciences**, v. 63, n. 5, p. 13–18, 2018.

CDC. Centers for Disease Control and Prevention. Interim Guidelines for Collecting, Handling, and Testing Clinical Specimens for COVID-19. p. 1–5, 2020a. Available in: <<https://www.cdc.gov/coronavirus/2019-ncov/lab/guidelines-clinical-specimens.html>>.

CDC. Centers for Disease Control and Prevention. 2019-Novel Coronavirus (2019-nCoV) Real-time RT-PCR Primer and Probe Information, 2020b. Available in: <<https://www.cdc.gov/coronavirus/2019-ncov/lab/rt-pcr-panel-primer-probes.html>>.

CORMAN, V. M.; LANDT, O.; KAISER, M.; MOLENKAMP, R.; MEIJER, A.; CHU, D. K. W.; BLEICKER, T.; BRÜNINK, S.; SCHNEIDER, J.; SCHMIDT, M. L.; MULDER, D. G. J. C.; HAAGMANS, B. L.; VAN DER VEER, B.; VAN DEN BRINK, S.; WIJSMAN, L.; GODERSKI, G.; ROMETTE, J. L.; ELLIS, J.; ZAMBON, M.; PEIRIS, M.; GOOSSENS, H.; REUSKEN, C.; KOOPMANS, M. P. G.; DROSTEN, C. Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR. **Eurosurveillance**, v. 25, n. 3, p. 1–8, 2020.

GORBALENYA, A. E.; BAKER, S. C.; BARIC, R. S.; DE GROOT, R. J.; DROSTEN, C.; GULYAEVA, A. A.; HAAGMANS, B. L.; LAUBER, C.; LEONTOVICH, A. M.; NEUMAN, B. W.; PENZAR, D.; PERLMAN, S.; POON, L. L. M.; SAMBORSKIY, D. V.; SIDOROV, I. A.; SOLA, I.; ZIEBUHR, J. The species Severe acute respiratory

syndrome-related coronavirus: classifying 2019-nCoV and naming it SARS-CoV-2. **Nature Microbiology**, v. 5, n. 4, p. 536–544, 2020.

IRIGOYEN, N.; FIRTH, A. E.; JONES, J. D.; CHUNG, B. Y. W.; SIDDELL, S. G.; BRIERLEY, I. High-Resolution Analysis of Coronavirus Gene Expression by RNA Sequencing and Ribosome Profiling. **PLoS Pathogens**, v. 12, n. 2, 2016.

LEE, C.; PARK, S.; CHO, K.; EUN, J.; SUNGHEE, Y.; GWANGPYO, L. Comparison of Swab Sampling Methods for Norovirus Recovery on Surfaces. **Food and Environmental Virology**, v. 10, n. 4, p. 378–385, 2018.

LU, R.; ZHAO, X.; LI, J.; NIU, P.; YANG, B.; WU, H.; WANG, W.; SONG, H.; HUANG, B.; ZHU, N.; BI, Y.; MA, X.; ZHAN, F.; WANG, L.; HU, T.; ZHOU, H.; HU, Z.; ZHOU, W.; ZHAO, L.; CHEN, J.; MENG, Y.; WANG, J.; LIN, Y.; YUAN, J.; XIE, Z.; MA, J.; LIU, W. J.; WANG, D.; XU, W.; HOLMES, E. C.; GAO, G. F.; WU, G.; CHEN, W.; SHI, W.; TAN, W. Genomic characterisation and epidemiology of 2019 novel coronavirus: implications for virus origins and receptor binding. **The Lancet**, v. 395, n. 10224, p. 565–574, 2020a.

LU X; WANG L; SAKTHIVEL SK; WHITAKER B; MURRAY J; KAMILI S; et al. US CDC real-time reverse transcription PCR panel for detection of severe acute respiratory syndrome coronavirus 2. **Emerging Infectious Diseases**, v.26, n. 8. 2020b Available in: <<https://doi.org/10.3201/eid2608.201246>>

WARIS, M.; ÖSTERBACK, R.; LAHTI, E.; VUORINEN, T.; RUUSKANEN, O.; PELTOLA, V. Comparison of sampling methods for the detection of human rhinovirus RNA. **Journal of Clinical Virology**, v. 58, n. 1, p. 200–204, 2013. Available in: <<http://dx.doi.org/10.1016/j.jcv.2013.06.002>>.

WHO. World Health Organization. Pneumonia of unknown cause–China. January, 2020a Available in: <<https://www.who.int/csr/don/05-january-2020-pneumonia-of-unknown-cause-china/en/>>

WHO. World Health Organization. Novel coronavirus (2019-nCoV). February 2020b. Available in: <<https://www.who.int/docs/default-source/coronaviruse/situation-reports/20200211-sitrep-22-ncov.pdf>>.