3T v 7T MRI for imaging subcortex

3 versus 7 Tesla Magnetic Resonance Imaging for parcellations of subcortical brain structures

Bethany R Isaacs^{1,2}, Martijn J Mulder^{1,3}, Josephine M Groot¹, Nikita van Berendonk¹, Nicky Lute^{1,4},
Pierre-Louis Bazin^{1,5}, Birte U Forstmann^{1*}, Anneke Alkemade^{1*}

¹. University of Amsterdam, Integrative Model-based Cognitive Neuroscience Research Unit,
Amsterdam, The Netherlands

². Maastricht University Medical Centre, Department of Experimental Neurosurgery, Maastricht, The
Netherlands

³. University of Utrecht, Psychology and Social Sciences, Utrecht, The Netherlands

⁴. Vrije University, Clinical Neuropsychology, Amsterdam, The Netherlands

⁵. Max Planck Institute for Human, Cognitive and Brain Sciences, Leipzig, Germany

* Contributed equally

Corresponding author:

Anneke Alkemade, PhD
Integrative Model-based Cognitive Neuroscience
(IMCN) research unit
University of Amsterdam
Nieuwe Achtergracht 129B | Room G3.08
Postbus 15926 | 1001 NK Amsterdam
jmalkemade@gmail.com

3T v 7T MRI for imaging subcortex

Abstract

7 Tesla (7T) magnetic resonance imaging holds great promise for improved visualization of the human brain for clinical purposes. To assess whether 7T is superior regarding localization procedures of small brain structures, we compared manual parcellations of the red nucleus, subthalamic nucleus, substantia nigra, globus pallidus interna and externa, created on a commonly used clinical anisotropic clinical 3T with an optimized isotropic (o)3T and standard 7T scan. While the clinical 3T MRI scan did not allow delineation of an anatomical plausible structure due to its limited spatial resolution, optimized 3T and 7T parcellations could be directly compared. We found that 7T outperformed the o3T MRI as reflected by higher Dice scores, which were used as a measurement of interrater agreement for manual parcellations on quantitative susceptibility maps. This increase in agreement was associated with higher contrast to noise ratios for smaller structures, but not for the larger globus pallidus segments. Additionally, control-analyses were performed to account for potential biases in manual parcellations by assessing semi-automatic parcellations. These results did not confirm increased Dice scores for 7T, although they revealed a higher consistency for structure volumes for 7T compared to optimized 3T, potentially indicative of more agreement with increased field strength. Together these results indicate the importance of the use of isotropic voxels with both 3T and 7T, and the potential for clinical applications.

3T v 7T MRI for imaging subcortex

Introduction

The availability of 7 Tesla (T) Magnetic Resonance Imaging (MRI) scanners has rapidly increased in recent years (Forstmann, Isaacs, and Temel 2017; Keuken et al. 2018; Ladd et al. 2018). The theoretical benefits of anatomical 7T MRI over lower field strengths can be attributed to the increased spatial resolution, contrast- and signal-to-noise ratios (CNR and SNR, respectively), which collectively result in higher quality imaging within feasible time frames [4,5]. Improved visibility of pathological alterations on 7T has been reported in the literature for brain tumors [6], epilepsy [7], multiple sclerosis [8], stroke [9], and neurodegenerative diseases [10]. However, to what extent increased visibility afforded by 7T has the potential to improve clinical outcomes regarding invasive neuro interventions remains unknown.

A promising clinical application of 7T MRI is the target visualization of structures for deep brain stimulation (DBS) surgery (Forstmann et al., 2017; Horn, 2019). DBS procedures target structures within the subcortex, which is comprised of a large number of small, iron and calcium-rich nuclei that are located in close proximity to one another [2]. Relatedly, alterations in biometals such as iron in human tissue are commonly observed in pathological processes, for instance, the occurrence of dopaminergic neurodegeneration of the substantia nigra (SN) in Parkinson's disease (PD). Such changes in the chemical composition can cause disease specific structural alterations in shape, volume and location [12–14]. Moreover, the neurophysical properties of both physiological and aberrant accumulation of biometals can be exploited to increase the visibility of structural boundaries with both ultra-high field (UHF) MRI and tailored post-processing techniques, such as quantitative susceptibility mapping (QSM) (Deistung et al. 2013; Neumann et al. 2011; Schweser et al. 2012; Wang et al. 2017).

The main DBS targets for PD and dystonic disorders are the globus pallidus interna (GPi) and subthalamic nucleus (STN) [20–23]. Identification of the STN benefits from visualization of the border of the SN, which has also been targeted for epilepsy (Loddenkemper, et al., 2001) Also the parcellation of the GPi benefits from visualizing the boundary with the external segment of the GP (GPe), the stimulation of which has been shown to modulate functional connectivity in Huntington's disease patients [25]. Additionally, the red nucleus (RN) is often used as a landmark for identification and orientation of the surrounding nuclei [26].

Conventional MRI can fail to capture the detailed local neuroanatomy due a weaker field strength. These limitations can be directly translated into a clinical setting with regards to the accuracy of MRI based targeting protocols for DBS implantations. DBS of the STN has been related to a number of psychiatric, cognitive, and emotional disturbances (Temel, Blokland, Steinbusch, & Visser-Vandewalle, 2005). Moreover, a fraction of patients will fail to respond to stimulation and or maintain their parkinsonian symptoms, and may require the removal or reimplantation of their DBS leads (Beric et al. 2002; Temel et al. 2005). These failures to appropriately respond to neurointervention can partially be

3T v 7T MRI for imaging subcortex

attributed to suboptimal placement of the DBS lead as a consequence of both inaccurate visualization of the target and reliance on landmark identification [29].

Clinical MRI often includes parallel imaging (PI) techniques to reduce acquisition time which is associated with an SNR penalty. This is warranted for both practical reasons, to improve image contrast, as well as clinical reasons, as patients with movement disorders cannot be scanned for extended periods of time. PI reconstructions result in spatially varying noise amplification, which is reflected in the g-factor. However, PI can result in both g-factor penalties and longitudinal magnetization saturation, which can produce anatomically inaccurate and distorted images [2,30]. Thus, anisotropic voxel sizes are commonly employed in order to maintain a higher SNR in-plane, however this negatively affects the creation of accurate three-dimensional representation of DBS nuclei. Alternatively, 7T can provide higher isotropic spatial resolution and SNR within a shorter time period [31].

Additionally, DBS surgeries commonly incorporate intra-operative micro electrode recordings and behavioral testing in awake patients to confirm optimal lead placement [32–34]. This is a time-consuming procedure and distressing for the patient. The higher spatial accuracy that 7T MRI offers could contribute to more accurate surgical targeting and clinical efficacy. Additionally, it can reduce the length of the surgery and the requirement for reimplantation, while ultimately contributing to the abolishment of the need for awake testing during surgery and dramatically improving patient comfort (Forstmann et al. 2017; Lyons et al. 2004).

In this work, we investigate the potential of 7T for improved targeting with a quantitative comparison of 3T with 7T MRI scans. We acquired two sets of 3T data; one representative of the resolution of clinical 3T (c3T) MRI typically used for DBS targeting, as well as an optimized set of 3T (o3T) and a set of 7T data from the same participants. We would like to clarify that we did not attempt to run the same optimized protocol at 3T and 7T, since this would have increased scan times for the o3T further, resulting in a scan sequence which would not have any practical use in a clinical setting. Important differences between the obtained contrasts are the anatomical contrast and the voxel size and shape. Direct quantitative comparisons were drawn from both manual and semi-automated parcellations the GPe, GPi, RN, SN, and STN. Given the iron rich nature of these deep brain structures, and our previous studies indicating that for such structures QSM outperforms T2*-weighted images we used QSM contrasts for parcellations (Alkemade, et al. 2017). Additionally, a semi-automated parcellation approach was employed to parcellate the GP, RN, SN and STN, in order to identify potential biases occurring with manual parcellations and whether accuracy increases with field strength.

3T v 7T MRI for imaging subcortex

Methods

Participants

10 healthy participants (male = 2, female = 8, mean age = 25.9 y, S.D age = 5.8 y), healthy as assessed by self-reports, were scanned at the Spinoza Centre for Neuroimaging in Amsterdam, The Netherlands, on a Philips 7T and 3T Achieva MRI system, with a 32-channel head array coil. The research was approved in writing by the LAB Ethics Review Board of the Faculty of Social and Behavioral Sciences, the local Ethical Committee of the Department of Psychology at the University of Amsterdam (ERB number 2016-DP-6897). All participants provided written informed consent prior to the scanning.

Data acquisition

c3 Tesla

Whole-brain T1-weighted images obtained with a 3D Turbo Field Echo (TFE) sequence with 1mm isotropic voxel sizes, field of view (FOV) = 240 x 188, 220 slices, echo time (TE) = 3.7 ms, repetition time (TR) = 8.2 ms, TFE factor = 142, TFE shots = 118, SENSE_{PA} = 2.5, acquisition time (TA) = 04:42 mins, obtained in the transverse plane. Whole-brain T2-weighted images obtained with a Turbo Spin Echo sequence (TSE) with 0.45 x 0.45 x 2mm voxel sizes, FOV = 230 x 182, 48 slices, TE = 80 ms. TR = 3000 ms, TSE factor = 15, TSE shots = 150, TA = 06:12 mins, obtained in the transverse plane. Total acquisition time was 10 minutes and 54 seconds.

o3 Tesla

Whole-brain T1-weighted images were obtained with a 3D Fast Field Echo (FFE) sequence with 1mm isotropic voxel sizes, FOV = 240 x 188, 220 slices, TE = 3.7ms, TR = 8.2ms, TFE factor = 142, TFE shots = 293, TA = 11:38ms in the transverse plane (no acceleration factor). Whole brain T2-weighted images were acquired with 3D Fast Field Echo (FFE) sequence with voxel sizes 1mm isotropic, TE_{1,2,3,4,5} = [4.1ms, 9.8ms, 13.85, 19.55, 23.60], TR = 46ms, echo spacing (ES) = 9.75ms, FA = 20, FOV = 240 x 188, 140 slices, SENSE_{PA} = 2, TA = 10:08mins. The main difference between the clinical and optimized 3T scans is the voxel size. Two separate scans were collected with o3T, with a total acquisition time of 21 minutes and 46 seconds. We would like to note that we were unable to match the o3T spatial resolution with that of the 7T due to specific absorption rate (SAR) limitations.

7 Tesla

For 7T, one scan incorporating both T1 and T2* contrasts was obtained using a MP2RAGEME (magnetization-prepared rapid gradient echo multi-echo) sequence [37]. The MP2RAGEME is an extension of the MP2RAGE sequence by [38] and consists of two rapid gradient echo (GRE_{1,2}) images that are acquired in sagittal plane after a 180° degrees inversion pulse and excitation pulses with inversion times TI_{1,2} = [670ms, 3675.4ms]. A multi-echo readout was added to the second inversion at four echo times (TE₁ = 3ms, TE_{2,1-4} = 3, 11.5, 19, 28.5ms). Other scan parameters include flip angles FA_{1,2} = [4°, 4°]; TR_{GRE1,2} = [6.2ms, 31ms]; bandwidth = 404.9 MHz; TR_{MP2RAGE} = 6778ms; acceleration factor SENSE_{PA} = 2; FOV = 205 x 205 x 164mm; acquired voxel size = 0.70 x 0.7 x 0.7mm; acquisition

3T v 7T MRI for imaging subcortex

matrix was 292 x 290; reconstructed voxel size = 0.64 x 0.64 x 0.7mm; turbo factor (TFE) = 150 resulting in 176 shots; Total acquisition time = 19.53min.

Maps

T1 and T2*-maps for o3T and 7T MRI scans were created using the following procedure: T1-maps were computed using a look-up table [38]. T2*-maps were computed by least-squares fitting of the exponential signal decay over the multi-echo images of the second inversion. For QSM, the 3T data underwent more extensive clipping at the frontal and sinus regions as compared to the 7T MRI data. This was required since the algorithm is sensitive to non-local artefacts, which are more prominent in these regions on o3T MRI scans. For QSM, phase maps were pre-processed using iHARPERELLA (integrated phase unwrapping and background phase removal using the Laplacian) of which the QSM images were computed using LSQR [39,40]. Scans were reoriented and skull information was removed as described previously (Forstmann et al., 2014). The c3T MRI sequence did not allow the calculation of quantitative T1 and T2* maps or QSM images.

Parcellation methods

Manual parcellation

Inspection of the c3T scans revealed that despite the high in-plane resolution, which allowed the identification of the structures of interest in the axial plane, we were unable to create a biologically plausible 3-dimensional reconstruction of the structures of interest due to the anisotropic nature of the voxel sizes. We therefore decided not to pursue further analyses of the c3T MRI scans. Multi-echo data was not acquired, and therefore it was not possible to reconstruct QSM images for parcellations.

For o3T and 7T images, manual parcellations were performed in individual space using the QSM images for the GPe/i, RN, SN, and STN by two independent trained researchers. Given the level of familiarity of these raters with MRI data, we concluded that blinding for the scan sequence was impossible. T1-maps and/or MP2RAGE images were used for additional anatomical orientation and identification of landmarks such as the ventricles, pons and corpus callosum. T2*-maps were also used where appropriate. See Supporting Information 1 for the approach used for manual parcellations. Raters were blind to each other's parcellations, and inter-rater agreement was determined by the Dice correlation coefficient (see statistical methods).

Semi-automated parcellation: Multimodal Image Segmentation Tool (MIST)

Semi-automated parcellation was performed for the combined GPe/i, RN, SN and STN with FSL's Multimodal Image Segmentation Tool (MIST) (Visser, Douaud, Jenkinson, et al. 2016; Visser, Keuken, Forstmann, et al. 2016). QSM-maps and T1-images were used as input for MIST. MIST output parcellations were compared across field strength (o3T vs 7T), as well as across parcellation method (manual vs. semi-automated) in order to assess for potential biases in manual parcellations such as order or practice effects.

3T v 7T MRI for imaging subcortex

The o3T brain extracted T1-weighted and QSM maps were co-registered via a multi-step process, where first whole brain T2*-maps were registered to the corresponding T1-weighted images using FLIRT (as implemented in FSL version 6.0.1) with 6 degrees of freedom, nearest neighbor interpolation and mutual information cost function. This transformation was then applied to the QSM-maps, extrapolated from the fifth echo of the T2* sequence, also with 6 degrees of freedom, mutual information cost function and instead a sinc interpolation. The same transforms were applied to the manual parcellations to allow for direct comparisons with MIST outputs. All registrations were visually inspected for misalignments by comparing the following landmarks: ventricles, pons, and corpus callosum.

The 7T MP2RAGEME sequence allowed the calculation of all contrasts from a single sequence and thus in the same space, not requiring any registration steps. The MP2RAGE was used as the whole-brain anatomical reference image and the fourth echo of the second inversion was used for the T2* image. Resampling was achieved with Nibabel (version 2.3), with second order spline interpolation, and constant mode parameter. Where appropriate, the header information was copied from the fourth echo of the second inversion to the MP2RAGE. Images were resampled as MIST only handles (near) isotropic voxel sizes. Images were resampled to 0.8mm.

Dice coefficients

Dice coefficients were assessed to determine interrater reliability (Dice, 1945). Dice scores were compared between o3T and 7T images to test the directed hypothesis that 7T images result in higher interrater agreement as compared to o3T images. The Dice coefficient was calculated as follows:

$$Dice = \frac{2 \times |m1 \cap m2|}{|m1| + |m2|}$$

Where $|m_i|$ is the size of mask i and $|m_1 \cap m_2|$ is the size of the conjunct mask of mask 1 and 2. A conjunct mask of a set of masks M only includes voxels included by both raters (Dice, 1945).

Volume calculations

For manual parcellations, all volume calculations were performed using the conjunct volume of the individual raters, as described previously. Calculations for manual Dice factors were calculated in the space in which the parcellations were performed (Alkemade et al., 2017; Keuken et al., 2013). Masks from the MIST output were compared with manually parcellated conjunction masks resampled to 0.8mm for the 7T data, and the masks that were registered from T2* to T1 for the o3T MRI data.

Anatomical distance

The anatomical distance between the centers of mass of the individual structures in the left and right hemisphere was assessed, providing a measure for changes in the perceived location of the individual structures across field strengths.

Distances were calculated as follows:

$$Distance(l, r) = \sum_{i=1}^n (l_i - r_i)^2$$

3T v 7T MRI for imaging subcortex

l and r correspond to the left and right hemisphere. The square root of the sum is obtained by adding the power of the left x , y and z coordinates of the center of mass of the individual structures l and r [46].

Contrast to noise ratios

Contrast-to-noise ratios (CNRs) of the QSM images were calculated to assess differences in visibility of the anatomical structure under investigation. Intensities of non-zero voxels were extracted using the *segmentation_statistics* implemented in Nighres [47]. The CNR was calculated as follows:

$$CNR = \frac{S_I - S_O}{\sigma_O}$$

S_I is the signal inside the mask, represented by the mean value of all the voxels in the conjunct mask. S_O is the signal outside the mask, calculated as the mean value of all voxels that directly border the outside of the disjunct mask (all voxels scored inside the mask of a single rater). σ_O is the standard deviation of the set of QSM intensities in these voxels. This approach was adopted to ensure that voxels outside the mask were not part of the separate individual masks.

Statistical Methods

All statistical analyses were conducted within a Bayesian framework (Table 1) using the BayesFactor toolbox (Morey and Wagenmakers 2014) in R (R Core Team, 2016), interpreted in light of the assumptions proposed by [50] and adapted by [51]. Bayesian approaches draw their distribution from the observed data, rather than requiring a Gaussian distribution, which can be difficult to meet in both small and or clinical datasets when using a frequentist perspective. Bayesian statistics sample from the data itself, the priors for each group are calculated independently, in which smaller groups will receive greater uncertainty compared to a larger group for comparison, with no trade off with statistical power [52]. Therefore, a limited number of subjects should not negatively impact our results [53,54]. Each test is performed independently of the others so we assume multiple comparisons are not a confounder in the present study [55–57]. We incorporated a within subjects' approach, and for all analyses data was collapsed across hemisphere.

3T v 7T MRI for imaging subcortex

Table 1. Bayes Factor Interpretation

<i>Bayes Factor</i>		<i>Interpretation</i>
BF_{10} <i>Evidence for H^1</i>	>100	Decisive evidence
	30-100	Very strong evidence
	10-30	Strong evidence
	3-10	Substantial evidence
	1-3	Anecdotal evidence
	1	No evidence
BF_{01} <i>Evidence for H_0</i>	0.33-1	Anecdotal evidence
	0.33-0.10	Substantial evidence
	0.10-0.03	Strong evidence
	0.03-0.01	Very strong evidence
	<0.01	Decisive evidence

H1 = experimental hypothesis, and H0 = null hypothesis.

Manual o3T v 7T

For manual parcellations, we hypothesized that Dice scores and CNRs are higher for 7T compared to o3T MRI scans, as assessed with one-tailed paired samples t-tests per structure. For each one-tailed test, two models were obtained. The first model (M_1) tested for a positive effect in support of our hypotheses, and a second model (M_2), tested for a negative effect in which 7T is either no different or is outperformed by the o3T MRI data. The preferred model was then reported along with the model comparisons.

We had no hypotheses on the direction of potential changes in volumes or Anatomical distances across field strengths, and therefore conducted two-tailed paired samples t-tests. These analyses provided a single model testing for a difference either way, compared to the null hypothesis. Where appropriate, we calculated the reciprocal to determine the evidence supporting the null-hypothesis.

Manual v semi-automated

Similarly, when assessing manual and semi-automated parcellations within field strength (manual o3T v MIST 3T and manual 7T v MIST 7T), two-tailed paired samples t-tests were conducted for CNRs, volumes, and Anatomical distances, which we did not expect to differ.

3T v 7T MRI for imaging subcortex

The Dice score for the MIST output parcellations is comprised of a conjunction mask including only the voxels selected by both the MIST parcellation and the resampled manual conjunction mask. Therefore, Dice scores were not directly tested across parcellation methods.

Semi-automated o3T v 7T

o3T and 7T MIST parcellation Dice scores and CNRs were compared with a one-tailed paired samples t-test, under the assumption that both Dice scores and CNRs would be higher for 7T than for o3T, indicating that 7T is subject to fewer biases than o3T. Volumes and Anatomical distances were again assessed with two-tailed paired samples t-tests.

<< Table 1. Bayes Factor Interpretation >>

Data sharing and accessibility statement

All anonymized data and analysis scripts are available from <https://osf.io/4nrku/>, under the terms of the Creative Commons Attribution License and complies with the rules of the General Data Protection Regulation (EU) 2016/679.

3T v 7T MRI for imaging subcortex

Results

The MR contrasts are illustrated in figure 1. QSM contrasts obtained from 3T and 7T sequences allowed for manual parcellation of the brain structures under investigation, resulting in biologically plausible 3D reconstructions (see figure 2). As previously mentioned, the 3T images provided excellent in-plane resolution, though did not reasonably allow for anatomically accurate reconstructions due to the anisotropic voxel sizes. Therefore, no formal analyses were pursued for the clinical scans. All results have been averaged across hemisphere, and presented with a margin of error of <0.1%. See table 2 for the results of the manual parcellations, and table 3 and figure 3 for MIST parcellations.

Table 2: Comparison of manual parcellations across optimized 3T and 7T MRI

Structure	Dice scores			Conjunct Volumes			Anatomical Distance			QSM CNRs		
	<i>o3T</i>	<i>7T</i>	<i>BF</i>	<i>o3T</i>	<i>7T</i>	<i>BF</i>	<i>o3T</i>	<i>7T</i>	<i>BF</i>	<i>o3T</i>	<i>7T</i>	<i>BF</i>
GPe	0.77	0.81	M_1 $BF_{10}=4.11$	1047	999.52	$BF_{01}=0.58$	37.35	36.72	$BF_{01}=0.86$	1.22	1.09	M_2 $BF_{10}=5.43$
	(0.04)	(0.05)	(Mc $BF_{10}=34.26$)	(127.46)	(136.57)		(1.67)	(1.22)		(0.16)	(0.14)	(Mc $BF_{10}=46.65$)
GPi	0.71	0.76	M_1 $BF_{10}=4.22$	420.25	415.09	$BF_{01}=0.31$	31.35	31.95	$BF_{01}=0.89$	0.81	0.75	M_2 $BF_{01}=0.66$
	(0.06)	(0.05)	(Mc $BF_{10}=35.13$)	(85.71)	(95.59)		(2.00)	(1.40)		(0.16)	(0.15)	(Mc $BF_{01}=3.62$)
RN	0.80	0.85	M_1 $BF_{10}=7.20$	233.40	227.44	$BF_{01}=0.47$	8.73	8.71	$BF_{01}=0.32$	1.92	2.05	M_1 $BF_{10}=1.64$
	(0.06)	(0.03)	(Mc $BF_{10}=63.73$)	(45.44)	(32.59)		(0.41)	(0.43)		(0.51)	(0.42)	(Mc $BF_{10}=11.76$)
SN	0.78	0.81	M_1 $BF_{10}=1.06$	434.45	401.39	$BF_{01}=0.35$	16.49	16.45	$BF_{01}=0.31$	1.62	1.78	M_1 $BF_{10}=1.20$
	(0.07)	(0.03)	(Mc $BF_{10}=6.89$)	(75.49)	(91.94)		(0.83)	(0.48)		(0.17)	(0.32)	(Mc $BF_{10}=7.87$)
STN	0.69	0.74	M_1 $BF_{10}=1.85$	88.90	87.35	$BF_{01}=0.51$	17.93	17.71	$BF_{01}=0.34$	1.04	1.33	M_1 $BF_{10}=61.75$
	(0.06)	(0.06)	(Mc $BF_{10}=13.59$)	(15.12)	(23.20)		(1.34)	(1.47)		(0.26)	(0.32)	(Mc $BF_{10}=629.61$)

Table 2: Dice scores, conjunct volumes, Anatomical distances, and QSM CNRs are averaged across hemisphere and presented as mean values and standard deviations for *o3T* and *7T* MRI contrasts. BF_{10} indicates evidence for the alternative, and BF_{01} refers to evidence for the null hypothesis. Dice scores and QSM CNRs were compared using a Bayesian one-tailed paired samples *t*-test, where BF_{10} assumes that in both cases *7T* is higher than *o3T* (model 1 (M_1)), and BF_{01} assumes either no difference or a decrease in *7T* compared to *o3T* (model 2 (M_2)). For one-tailed *t*-tests, only the *BF* for the winning model is noted, and the likelihood ratio is calculated between the winning and losing models and is noted by 'Mc' (standing for model comparisons). Conjunct volumes and Anatomical distances were compared between *3T* and *7T* with two-tailed paired-samples *t*-tests. *o3T* = optimized 3T, QSM = quantitative susceptibility mapping, CNR = contrast to noise ratio, GPe = globus pallidus externa, GPi = globus pallidus interna, RN = red nucleus, SN = substantia nigra, STN = subthalamic nucleus.

3T v 7T MRI for imaging subcortex

<< Figure 1. Contrasts >>

Figure 1. Contrasts

A single subjects clinical 3T (c3T) (T1- and T2- weighted), optimized 3T (o3T) (T1w, T2* and QSM images) and 7T (MP2RAGE, T2* maps and QSM images) in the coronal and axial planes. Brightness and contrast levels were chosen to best visualize the basal ganglia.

<< Figure 2. Voxel sizes >>

Figure 2: Voxel sizes

Example of a single subjects clinical 3T (c3T) T2 weighted, optimized 3T (o3T) and 7T T2* maps in the coronal, sagittal and axial planes. The RN, SN and STN are highlighted to exemplify the difficulty in identification of the nuclei in the coronal and sagittal planes for the c3T compared to the o3T and 7T due to the anisotropic voxel sizes, making 3D parcellations impossible. T2 weighted images and T2* maps are presented as they show the iron rich RN, SN and STN as hypointense structures. This was done since the c3T scan did not allow for QSM calculations, which would result in a hyperintense contrast of these brain nuclei.

Manual parcellations: o3T v 7T

Dice Scores

Dice scores for the GPe ($BF_{10} = 4.11$), GPi ($BF_{10} = 4.22$) and RN ($BF_{10} = 7.20$) all reported substantial evidence in favor of 7T parcellations having a higher Dice than o3T. Additionally, these models were 33, 54 and 64 times (respectively) more likely than either no difference, or o3T having a higher Dice than 7T (referred to in the following sections as the alternative). For the SN ($BF_{10} = 1.06$) and STN ($BF_{01} = 1.85$), only anecdotal evidence was found in favor of 7T over o3T, which were 7 and 13 times more likely than the alternative, respectively. All winning models noted here were at least moderately more likely than the second model.

Volumes

When assessing for differences in volumes across field strength per structure, we found consistent anecdotal evidence for no difference for the GPe ($BF_{01} = 0.58$), GPi ($BF_{01} = 0.31$), RN ($BF_{01} = 0.47$), SN ($BF_{01} = 0.35$) and STN ($BF_{01} = 0.51$). Substantial evidence was found for differences in volumes per rater for o3T parcellations for the RN ($BF_{10} = 3.89$), and SN ($BF_{10} = 6.72$), and at 7T, strong evidence was found for the SN ($BF_{10} = 29.87$). All other structures showed either anecdotal or no evidence for differences across raters. Surprisingly, the GPe, GPi, SN, and STN showed higher standard deviations at 7T than o3T. Additionally, Pearson's Rho correlation indicated that for o3T, Dice scores correlated with volumes for the GPe ($r = 0.49$), GPi ($r = 0.76$), RN ($r = 0.45$), SN ($r = 0.14$) and STN ($r = 0.61$), and at 7T for the GPe ($r = 0.61$), GPi ($r = 0.86$), RN ($r = 0.41$), SN ($r = 0.80$) and STN ($r = 0.42$). This is indicative of a bias where larger structures have a higher Dice score.

3T v 7T MRI for imaging subcortex

Anatomical Distance

When assessing for differences in distances across field strengths per structure, we found consistent evidence for no differences for the GPe ($BF_{01} = 0.86$, *anecdotal*), GPi ($BF_{01} = 0.89$, *substantial*), RN ($BF_{01} = 0.32$, *substantial*), SN ($BF_{01} = 0.31$, *substantial*), and STN ($BF_{01} = 0.34$, *substantial*).

QSM CNRs

When assessing for differences in QSM CNRs for manual parcellations across field strength per structure, we found very strong evidence for higher CNRs for the STN for 7T than o3T ($BF_{10} = 61.75$), which was 630 times, and decisively more likely than no difference, or higher CNRs at o3T. However, the RN ($BF_{10} = 1.64$) and SN ($BF_{10} = 1.20$) showed only anecdotal evidence for increased CNRs at 7T than 3T, which are 12 and 8 times more likely than no differences or higher CNRs at o3T, respectively. For the GPi ($BF_{10} = 0.66$), no evidence was found either way and for the GPe ($BF_{10} = 5.43$), substantial evidence was found for increased CNR at o3T than 7T, which was 47 times more likely than higher CNRs at 7T.

MIST parcellations

Dice Scores

Dice scores were calculated per field strength, per structure with a one-sided paired samples t-test for manual and semi-automated parcellations. For the GPe/i ($BF_{10} = 631.44$), we found decisive evidence that 7T Dice scores were higher than o3T, which was 22501 times more likely than no difference, or higher Dice scores at o3T (referred to as the alternative). For the RN ($BF_{10} = 9.15$), we found substantial evidence that 7T Dice scores were higher than o3T, which is 83 times more likely than the alternative. For the SN ($BF_{01} = 0.61$), we found anecdotal evidence for the alternative, with increased Dice scores at o3T than 7T which is 3 times more likely than the initial hypothesis that 7T Dice scores are higher than o3T. For the STN ($BF_{10} = 1.04$), only anecdotal evidence was found for higher Dice scores at 7T than o3T which was 7 times more likely than the alternative (see figure 3).

Volumes

Two-tailed paired samples t-tests were conducted to assess differences in the volume of manual parcellations compared to the MIST output parcellations per field strength. For o3T, we found that the GPe/i ($BF_{10} = 456.18$, *decisive*) and STN ($BF_{10} = 11.96$, *strong*) were smaller with MIST than when manually parcellated. Whereas the o3T SN was larger for MIST than for manual parcellations ($BF_{10} = 4.10$, *substantial*). For the o3T RN ($BF_{01} = 0.14$, *substantial*) we have evidence for no difference. For 7T, we found evidence for no difference between manual and MIST parcellations for the RN ($BF_{01} = 0.33$, *substantial*), SN ($BF_{01} = 0.74$, *anecdotal*), and STN ($BF_{01} = 0.31$, *substantial*). However, the GPe/i showed very strong evidence for increased volumes for manual parcellations than MIST ($BF_{10} = 36.11$). Next, the volumes of MIST parcellations were compared across o3T and 7T. Again, for the RN ($BF_{01} = 0.33$, *substantial*), SN ($BF_{01} = 0.31$, *substantial*), and STN ($BF_{01} = 0.88$, *anecdotal*), no differences in volumes were found. Finally, for the GPe/i, we found very strong evidence for increased volumes at 7T than o3T for MIST parcellations ($BF_{10} = 69.08$).

3T v 7T MRI for imaging subcortex

408
409
410
411
412
413
414
415
416
417
418
419
420
421
422
423
424
425
426
427
428
429
430
431
432
433
434
435
436
437
438
439
440
441

Anatomical Distances

Two-tailed paired samples t-tests were conducted to assess for differences in the anatomical distance of manual parcellations with the MIST output parcellations per field strength. For o3T, the GPe/i ($BF_{01} = 0.43$), RN ($BF_{01} = 0.34$), SN ($BF_{01} = 0.40$), and STN ($BF_{01} = 0.51$) all showed anecdotal evidence for no difference. For 7T, anecdotal evidence for no difference was found for the GPe/i ($BF_{01} = 0.37$), RN (0.37) and STN (0.60), and for the SN we found anecdotal evidence for an increase in distance for MIST parcellations compared to manual ($BF_{10} = 1.33$).

QSM CNRs

Two-tailed paired samples t-tests were conducted to assess differences in the CNR of manual parcellations with the MIST output parcellations per field strength. For o3T, the GPe/i ($BF_{10} = 138.13$, *decisive*) and the SN ($BF_{10} = 3.58$, *substantial*) showed evidence for increased CNRs with manual over MIST parcellations. The o3T RN ($BF_{01} = 0.32$, *substantial*) and STN ($BF_{01} = 0.66$, *anecdotal*) showed evidence for no difference in CNRs. For 7T, the GPe/i ($BF_{10} = 1.42$) showed anecdotal evidence for higher CNRs for manual parcellations than MIST, the RN ($BF_{01} = 0.30$) showed substantial evidence for no difference, and the SN ($BF_{10} = 3.54$, *substantial*) and STN ($BF_{01} = 29.89$, *strong*) evidence for increased CNRs for manual parcellations than MIST. Next, CNRs of MIST parcellations were compared across o3T and 7T. The GPe/i ($BF_{01} = 0.70$), RN ($BF_{01} = 0.38$) and STN ($BF_{01} = 0.61$) all showed anecdotal evidence for no difference in CNR across field strength. The SN ($BF_{10} = 6.02$) showed substantial evidence for increased CNR at o3T than 7T, and was 51 times more likely than 7T CNRs being higher than o3T. Each winning model showed substantial evidence that it was more likely than the alternative.

<< Figure 3. MIST parcellations >>

Figure 3. MIST parcellations

Outline of masks for each structure manually parcellated optimized 3T (o3T) (left) and 7T (right) from a single subject. Manual masks are shown in opaque purple for o3T and yellow for 7T. MIST masks are shown as an orange outline for o3T and pink outline for 7T. Structures are shown in the coronal, sagittal, and axial planes. QSM contrasts were used for all parcellations. QSM = quantitative susceptibility mapping, GPe/i = combined globus pallidus externa and interna, RN = red nucleus, SN = substantia nigra, STN = subthalamic nucleus.

Table 3a: Comparison of manual and semi-automated within and across field strength (Dice scores and volumes)

Structure	Dice Scores			Volumes						
	<i>o3T*</i>	<i>7T*</i>	<i>3T v 7T MIST **</i> <i>BF</i>	<i>o3T</i> <i>Manual</i>	<i>o3T</i> <i>MIST</i>	<i>o3T MIST v</i> <i>Manual BF</i>	<i>7T</i> <i>Manual</i>	<i>7T MIST</i>	<i>7T MIST v</i> <i>7T Manual</i> <i>BF</i>	<i>o3T v 7T</i> <i>MIST BF</i>
GPe/i	0.74 (0.02)	0.82 (0.03)	<i>M</i> ₁ BF ₁₀ =631.44 (<i>Mc</i> BF ₁₀ =22501)	1466.20 (161.59)	1171.91 (72.41)	BF ₁₀ =456.18	1467.49 (193.23)	1258.50 (80.45)	BF ₁₀ =36.11	BF ₁₀ =69.08
RN	0.81 (0.06)	0.85 (0.04)	<i>M</i> ₁ BF ₁₀ =9.15 (<i>Mc</i> BF ₁₀ =82.93)	233.40 (45.45)	236.55 (54.65)	BF ₀₁ =0.14	236.26 (31.16)	239.77 (39.16)	BF ₀₁ =0.33	BF ₀₁ =0.33
SN	0.78 (0.02)	0.75 (0.12)	<i>M</i> ₂ BF ₀₁ =0.61 (<i>Mc</i> BF ₁₀ =3.17)	434.55 (475.36)	503.60 (101.65)	BF ₁₀ =4.10	445.47 (58.41)	497.64 (127.78)	BF ₀₁ =0.74	BF ₀₁ =0.31
STN	0.67 (0.08)	0.70 (0.07)	<i>M</i> ₁ BF ₁₀ =1.04 (<i>Mc</i> BF ₁₀ =6.61)	88.95 (15.15)	75.25 (16.49)	BF ₁₀ =11.96	90.01 (15.69)	91.06 (24.15)	BF ₀₁ =0.31	BF ₀₁ =0.88

* overlap between manual and semi-automated parcellations, ** preferred model

Table 3b: Comparison of manual and semi-automated within and across field strength (Anatomical distance and QSM CNRs)

Structure	Anatomical Distance							QSM CNRs						
	o3T Manual	o3T MIST	o3T MIST v Manual BF	7T Manual	7T MIST	7T MIST v Manual BF	o3T v 7T MIST BF	o3T Manual	o3T MIST	o3T MIST v Manual BF	7T Manual	7T MIST	7T MIST v Manual BF	o3T v 7T MIST** BF
GPe/i	35.61 (1.70)	35.32 (1.38)	BF ₀₁ =0.43	35.32 (1.16)	35.45 (1.32)	BF ₀₁ =0.37	BF ₀₁ =0.42	1.39 (0.17)	1.21 (0.13)	BF ₁₀ =138.13	1.36 (0.17)	1.25 (0.20)	BF ₁₀ =1.42	M ₁ BF ₀₁ = 0.70 (Mc BF ₁₀ =3.88)
RN	8.73 (0.40)	8.67 (0.51)	BF ₀₁ =0.34	8.77 (0.44)	8.91 (0.69)	BF ₀₁ =0.37	BF ₁₀ =1.22	1.89 (0.50)	1.86 (0.47)	BF ₀₁ =0.32	1.89 (0.34)	1.89 (0.36)	BF ₀₁ =0.30	M ₁ BF ₀₁ = 0.38 (Mc BF ₁₀ =1.48)
SN	16.48 (0.82)	16.34 (0.74)	BF ₀₁ =0.40	16.44 (0.45)	17.10 (1.27)	BF ₁₀ =1.33	BF ₁₀ =2.03	1.91 (0.18)	1.80 (0.15)	BF ₁₀ =3.58	1.94 (0.26)	1.50 (0.35)	BF ₁₀ =3.54	M ₁ BF ₁₀ = 6.02 (Mc BF ₁₀ =52.39)
STN	17.93 (1.33)	17.55 (0.83)	BF ₀₁ =0.51	17.72 (1.44)	17.98 (1.31)	BF ₀₁ =0.60	BF ₀₁ =0.60	1.05 (0.15)	1.11 (0.27)	BF ₀₁ =0.66	1.36 (0.19)	1.19 (0.20)	BF ₁₀ =29.8 9	M ₁ BF ₀₁ = 0.61 (Mc BF ₁₀ =3.17)

** preferred model

Table 3a and 3b: Dice scores, conjunct volumes, Anatomical distances, and QSM CNRs are averaged across hemispheres and presented as mean values and standard deviations for o3T and 7T MRI contrasts. BF_{10} indicates evidence for the alternative, and BF_{01} refers to evidence for the null hypothesis. Dice scores of the agreement between manual and MIST parcellations are compared across o3T and 7T with a one-tailed paired samples t-tests where BF_{10} assumes that 7T is higher than o3T (model 1 (M_1)), and BF_{01} assumes no difference or a decrease in 7T compared to o3T (model 2 (M_2)). The volume, Anatomical distance and QSM CNRs are calculated for manual and MIST parcellations, and are compared both within o3T and 7T, as well as across field strength for MIST parcellations only. Volumes and Anatomical distances are compared with a two-tailed paired samples t-test. CNRs are additionally compared with a one-tailed paired samples t-tests wherein each case the BF_{10} assumes that 7T CNRs are higher than o3T (model 1 (M_1)), and BF_{01} assumes no difference or a decrease in 7T compared to o3T (model 2 (M_2)). For one-tailed paired samples t-tests, only the BF for the winning model is noted, and the likelihood ratio is calculated between the winning and losing models and is noted by Mc = model comparisons). QSM = quantitative susceptibility mapping, CNR = contrast to noise ratio, GPe/i = combined globus pallidus externa and interna, RN = red nucleus, SN = substantia nigra, STN = subthalamic nucleus.

3T v 7T MRI for imaging subcortex

Discussion

We set out to investigate whether a clinically feasible 7T sequence can outperform optimized and clinically feasible 3T MRI protocols for the visualization of a selection of subcortical DBS or landmarks including the GPe, GPi, RN, SN, and STN. Our main findings can be summarized as follows: 1) clinical 3T MRI did not allow for accurate manual 3D parcellation of subcortical nuclei primarily due to anisotropic voxel sizes. 2) 7T outperformed optimized 3T MRI protocol for manual parcellations for larger structures (GPe, GPi and RN); 3) very strong evidence for increased QSM CNR at 7T was found for the STN when manually parcellated; 4) When using MIST for semi-automatic parcellations, Dice scores did not indicate that 7T outperformed optimized 3T. It is important to note that technically it would be possible to match the 3T scan sequence more closely to the 7T sequence. However, the resulting protocol would not meet the criteria for implementation in clinical practice due to SAR and time constraints. It is thus important to note that the improved performance of the 7T protocol cannot be exclusively attributed to differences in field strengths, but also to the constraints imposed by clinical practice.

Clinical MRI

While the 3T scans allowed us to discern all targets in the axial plane, it was not possible to create a biologically plausible 3D reconstruction due to the anisotropic voxels employed. Arguably, identification for DBS targeting is sufficient due to the increased SNR along the axial plane, with a fast scan time of around 4 minutes [58]. However, anisotropic voxels are severely limited in their ability to portray an accurate 3D representation of the target structure and are prone to partial voluming effects [59]. Therefore, quantitative analysis of the 3T were not performed, as biologically plausible 3D renderings could not be obtained. We have included the scan parameters and parcellations in a qualitative manner to illustrate the variation in target visibility across scans (see figure 2) [60–62]

Manual parcellations

For manual parcellations, we found varying evidence in support of higher inter-rater agreement at 7T compared to the 3T MRI for the GPe/i, RN, SN and STN, which suggests that larger nuclei, and to a lesser extent, smaller nuclei have a higher visibility at increased field strengths. Additionally, the volumes were smaller for all structures at 7T, which could be explained by smaller voxels and increased SNR, which counters the effects of partial voluming observed both with lower field strengths and larger voxel sizes [59]. Interestingly, the CNR values were lower for 7T than 3T for the GPe/i. CNRs for the RN, SN, and STN were slightly higher at 7T than 3T, though did not result in a higher agreement between raters. This discrepancy may be explained by a raters bias with regards to prior information about size, shape and surrounding anatomy [63]. Moreover, while QSM is a quantitative measure, a strict consensus on the susceptibility values of specific structures is still lacking [64]. Additionally, smaller voxel sizes provide more precise information regarding the size and shape of structures. However, the number of voxels required to define a structure will increase, which may not directly improve the accuracy of manual parcellation. For example the STN Dice scores showed only a minor

3T v 7T MRI for imaging subcortex

increase at 7T compared to o3T, despite there being evidence for increased CNR at 7T. Moreover, higher resolution allows for more freedom in where to place structural boundaries. Whereas larger voxels make labeling more reproducible, though not necessarily more anatomically accurate.

Semi-automated parcellations

For semi-automated parcellations, two sets of comparisons were computed. First, we set out to compare within field strength the differences between manual and MIST parcellations to assess for biases occurring with manual parcellations. Second, we compared MIST parcellations across field strength, as we initially did for the manual parcellations, to determine whether semi-automated protocols benefit from 7T MRI. Of note, the o3T MRI data was pre-registered to a common 1mm isotropic space to allow for computations using MIST. Similarly, to maintain the high spatial resolution but ensure compatibility, the 7T images were resampled to 0.8 mm isotropic.

Volumes for the GPe/i were very different across segmentation method. This inconsistency could be explained by the fact that manual raters parcellated the GPe and GPi separately, but were combined in MIST parcellating the GP as a single structure. Moreover, the MIST prior for the GP includes the medial medullary lamina which we did not include in our manual parcellations. Dice scores were calculated for the manual and MIST parcellations per field strength, and then compared across o3T and 7T. The RN and STN Dice scores were higher for 7T than o3T across parcellation method. Interestingly the SN Dice scores were higher for o3T than 7T. o3T MIST parcellations had smaller volumes for the GPe/i, RN and STN, and larger for the SN compared to manual parcellations. For 7T, MIST volumes were more consistent with manual parcellations. MIST parcellated volumes did not differ across field strength for the RN, SN, and STN, however the GPe/i did, which suggests larger structures may be more accurately parcellated with 7T than o3T. Generally, CNRs were higher for manual than for MIST parcellations at o3T, apart from the STN which had a higher CNR for MIST. Similarly, for 7T all CNRs were higher or equal for manual parcellations than for MIST. We found no difference in CNRs across field strength for the MIST parcellations of the palladium, RN and STN.

Interestingly, the MIST SN had higher a CNR for o3T than 7T. These findings suggest that overall, the semi-automated parcellation procedures that were applied do not appear to rely as heavily on CNR as manual parcellations and are therefore not subject to the same biases as manual parcellation. However, the SN and STN at o3T may be an exception. It may be that for smaller nuclei, semi-automated methods using lower field strengths or images with larger voxel sizes rely more on CNR for identification of structural boundaries, whereas higher field strengths or images with submillimeter resolution instead rely on the spatial information.

Applications

It is important to consider the relevance of these findings in light of neurosurgical applications. Previous work has shown that while the visualization of the STN at 7T shows increased SNR, target localization is not necessarily improved (van Laar et al., 2016; [66,67]). Our findings indicate that optimization of 3T

3T v 7T MRI for imaging subcortex

MRI scans through the use of isotropic voxels and QSM do indeed allow for more accurate visualization of the surgical target, a result that is within reach for clinical application without the need for the investment in new hardware [68,69]. We would like to note that despite improved anatomical orientation, individual variation in the internal structure of the STN may continue to require awake testing of patients during surgery to obtain the desired clinical effect. Additionally, we have shown that an o3T scan can be obtained in a timeframe that is sensible within clinical practice and can account for age related increases in pathological iron deposition by using multiple and increasing echo times without superseding SAR limitations. This is a particularly important finding given the limitations of both c3T and 7T imaging, which include proneness to increased geometric distortions which reduce spatial accuracy and increase artefacts, B1 field inhomogeneity, power deposition, and altered specific absorption rates [2,70]. Moreover, patient-related contraindications such as metal and or electronic implants, prostheses and foreign bodies, vascular or renal disorders, weight and claustrophobia can limit the potential patient population able to undergo a 7T MRI [3,71,72]. Thus, while our results indicate that 7T is to an extent superior to 3T, o3T could provide a more clinically viable option.

Considerations

The cohort tested in this study consists of young healthy participants, and it is well known that older participants and PD patients have increased iron content in basal structures (Wang et al., 2016). Since the effects on QSM increase with age and disease, we may underestimate the clinical relevance of these findings (Alkemade et al., 2017). Moreover, the o3T consists of two separate scans, whereas the 7T acquisition includes a multi contrast scan obtained within a single session. A multi contrast scan at lower fields would have resulted in an increased scanning time, and therefore be arguably more difficult for scanning with patient populations, especially those with movement disorders. Additionally, a direct comparison between the 3T and 7T data would require co-registration to the same space involving resampling of the data. Since the outcomes of such a comparison could differ substantially dependent on the registration approach chosen, we decided not to perform such analyses.

Conclusions

Given the limited availability and compatibility restrictions in the patient population of 7T MRI systems for clinical application, our results have merit for more short-term improvement of clinical neuroimaging procedures for surgical purposes. Finally, the use of isotropic voxels is of great importance in these efforts, and we call for caution in the application of anisotropic voxels.

Acknowledgments

This research was supported by a Vidi and Vici grant from the Dutch Organization for Scientific Research (BUF), and an NWO-STW grant from the Dutch Organization for Scientific Research (BUF, MJM, AA).

3T v 7T MRI for imaging subcortex

References

1. Forstmann BU, Isaacs BR, Temel Y. Ultra High Field MRI-Guided Deep Brain Stimulation. Trends Biotechnol. 2017; doi:10.1016/j.tibtech.2017.06.010
2. Keuken MC, Isaacs BR, Trampel R, van der Zwaag W, Forstmann BU. Visualizing the Human Subcortex Using Ultra-high Field Magnetic Resonance Imaging. Brain Topogr. 2018; doi:10.1007/s10548-018-0638-7
3. Ladd ME, Bachert P, Meyerspeer M, Moser E, Nagel AM, Norris DG, et al. Pros and cons of ultra-high-field MRI/MRS for human application. Prog Nucl Magn Reson Spectrosc. The Authors; 2018;109: 1–50. doi:10.1016/j.pnmrs.2018.06.001
4. Kraff O, Fischer A, Nagel AM, Mönninghoff C, Ladd ME. MRI at 7 Tesla and above: Demonstrated and potential capabilities. J Magn Reson Imaging. 2015;41: 13–33. doi:10.1002/jmri.24573
5. Tractnig S, Springer E, Bogner W, Hangel G, Strasser B, Dymerska B, et al. Key clinical benefits of neuroimaging at 7 T. Neuroimage. 2018;168: 477–489. doi:10.1016/j.neuroimage.2016.11.031
6. Peerlings J, Compter I, Janssen F, Wiggins CJ, Postma AA, Mottaghy FM, et al. Characterizing geometrical accuracy in clinically optimised 7T and 3T magnetic resonance images for high-precision radiation treatment of brain tumours. Phys Imaging Radiat Oncol. Elsevier; 2019;9: 35–42. doi:10.1016/j.phro.2018.12.001
7. De Ciantis A, Barba C, Tassi L, Cosottini M, Tosetti M, Costagli M, et al. 7T MRI in focal epilepsy with unrevealing conventional field strength imaging. Epilepsia. 2016;57: 445–454. doi:10.1111/epi.13313
8. Inglese M, Fleysher L, Oesingmann N, Petracca M. Clinical applications of ultra-high field magnetic resonance imaging in multiple sclerosis. Expert Rev Neurother. Taylor & Francis; 2018;18: 221–230. doi:10.1080/14737175.2018.1433033
9. Madai VI, von Samson-Himmelstjerna FC, Bauer M, Stengl KL, Mutke MA, Tovar-Martinez E, et al. Ultrahigh-field mri in human ischemic stroke - a 7 Tesla study. PLoS One. 2012;7. doi:10.1371/journal.pone.0037631
10. McKiernan EF, O'Brien JT. 7T MRI for neurodegenerative dementias in vivo: A systematic review of the literature. J Neurol Neurosurg Psychiatry. 2017;88: 564–574. doi:10.1136/jnnp-2016-315022
11. Horn A. The impact of modern-day neuroimaging on the field of deep brain stimulation. Curr Opin Neurol. 2019;32: 1. doi:10.1097/wco.0000000000000679
12. Aquino D, Bizzi A, Grisoli M, Garavaglia B, Bruzzone MG, Nardocci N, et al. Age-related Iron Deposition in the Basal Ganglia: Quantitative Analysis in Healthy Subjects. Radiology. 2009;252: 165–172. doi:10.1148/radiol.2522081399
13. Keuken MC, Bazin PL, Backhouse K, Beekhuizen S, Himmer L, Kandola A, et al. Effects of aging on T1 , T2* , and QSM MRI values in the subcortex. Brain Struct Funct. 2017;222: 2487–2505. doi:10.1007/s00429-016-1352-4
14. Zwirner J, Möbius D, Bechmann I, Arendt T, Hoffmann K-T, Jäger C, et al. Subthalamic

3T v 7T MRI for imaging subcortex

- nucleus volumes are highly consistent but decrease age-dependently-a combined magnetic resonance imaging and stereology approach in humans. *Hum Brain Mapp.* John Wiley & Sons, Ltd; 2017;38: 909–922. doi:10.1002/hbm.23427
15. Deistung A, Schäfer A, Schweser F, Biedermann U, Turner R, Reichenbach JR. Toward in vivo histology: A comparison of quantitative susceptibility mapping (QSM) with magnitude-, phase-, and R 2* -imaging at ultra-high magnetic field strength. *Neuroimage.* 2013;65: 299–314. doi:10.1016/j.neuroimage.2012.09.055
16. Neumann J, Schäfer A, Bowtell R, Turner R, Wharton S, Forstmann BU, et al. Direct visualization of the subthalamic nucleus and its iron distribution using high-resolution susceptibility mapping. *Hum Brain Mapp.* Wiley Subscription Services, Inc., A Wiley Company; 2011;33: 2831–2842. doi:10.1002/hbm.21404
17. Schweser F, Sommer K, Deistung A, Reichenbach JR. Quantitative susceptibility mapping for investigating subtle susceptibility variations in the human brain. *Neuroimage.* 2012;62: 2083–2100. doi:10.1016/j.neuroimage.2012.05.067
18. Schweser F, Sommer K, Deistung A, Reichenbach JR, Schäfer A, Forstmann BU, et al. Direct visualization of the subthalamic nucleus and its iron distribution using high-resolution susceptibility mapping. *Neuroimage.* 2012;46: 2831–2842. doi:10.1002/jmri.25693
19. Wang Y, Spincemaille P, Liu Z, Dimov A, Deh K, Li J, et al. Clinical quantitative susceptibility mapping (QSM): Biometal imaging and its emerging roles in patient care. *J Magn Reson Imaging.* 2017;46: 951–971. doi:10.1002/jmri.25693
20. Kleiner-Fisman G, Herzog J, Fisman DN, Tamma F, Lyons KE, Pahwa R, et al. Subthalamic nucleus deep brain stimulation: summary and meta-analysis of outcomes. *Mov Disord.* 2006; S290-304. doi:10.1002/mds.20962
21. Odekerken VJJ, Boel JA, Schmand BA, De Haan RJ, Figee M, Van Den Munckhof P, et al. GPi vs STN deep brain stimulation for Parkinson disease. *Neurology.* 2016;86: 755–761. doi:10.1212/WNL.0000000000002401
22. Odekerken VJJ, van Laar T, Staal MJ, Mosch A, Hoffmann CFE, Nijssen PCG, et al. Subthalamic nucleus versus globus pallidus bilateral deep brain stimulation for advanced Parkinson's disease (NSTAPS study): A randomised controlled trial. *Lancet Neurol.* 2013;12: 37–44. doi:10.1016/S1474-4422(12)70264-8
23. Vidailhet M, Vercueil L, Houeto J-L, Krystkowiak P, Benabid A-L, Cornu P, et al. Bilateral Deep-Brain Stimulation of the Globus Pallidus in Primary Generalized Dystonia. *N Engl J Med.* 2005;352: 459–467. doi:10.1056/nejmoa042187
24. Loddenkemper, T., Pan, A., Neme, S., Baker, K. B., Rezai, A. R., Dinner, D. S., ... & Lüders, H O. Deep Brain Stimulation in Epilepsy. *Deep Brain Stimulation in Neurological and Psychiatric Disorders.* 2008. pp. 561–569. doi:10.1007/978-1-59745-360-8_28
25. Ligot N, Krystkowiak P, Simonin C, Goldman S, Peigneux P, Van Naemen J, et al. External globus pallidus stimulation modulates brain connectivity in Huntington's disease. *J Cereb Blood Flow Metab.* 2011;31: 41–46. doi:10.1038/jcbfm.2010.186
26. Pollo C, Meuli R, Maeder P, Vingerhoets F, Ghika J, Villemure JG. Subthalamic Nucleus Deep

3T v 7T MRI for imaging subcortex

- Brain Stimulation for Parkinson's Disease: Magnetic Resonance Imaging Targeting Using Visible Anatomical Landmarks. *Stereotact Funct Neurosurg.* 2003;80: 76–81. doi:10.1159/000075163
27. Beric A, Kelly PJ, Rezai A, Sterio D, Mogilner A, Zonenshayn M, et al. Complications of deep brain stimulation surgery. *Stereotactic and Functional Neurosurgery.* 2002. pp. 73–78. doi:10.1159/000064600
28. Temel Y, Blokland A, Steinbusch HWM, Visser-Vandewalle V. The functional role of the subthalamic nucleus in cognitive and limbic circuits. *Prog Neurobiol.* 2005;76: 393–413. doi:10.1016/j.pneurobio.2005.09.005
29. Low HL, Ismail MN bin M, Taqvi A, Deeb J, Fuller C, Misbahuddin A. Comparison of posterior subthalamic area deep brain stimulation for tremor using conventional landmarks versus directly targeting the dentatorubrothalamic tract with tractography. *Clin Neurol Neurosurg.* Elsevier; 2019;185: 105466. doi:10.1016/j.clineuro.2019.105466
30. Wiggins GC, Polimeni JR, Potthast A, Schmitt M, Alagappan V, Wald LL. 96-Channel receive-only head coil for 3 Tesla: Design optimization and evaluation. *Magn Reson Med.* 2009;62: 754–762. doi:10.1002/mrm.22028
31. Tona KD, van Osch MJP, Nieuwenhuis S, Keuken MC. Quantifying the contrast of the human locus coeruleus in vivo at 7 Tesla MRI. *PLoS One.* 2019;14. doi:10.1371/journal.pone.0209842
32. Hutchison WD, Allan RJ, Opitz H, Levy R, Dostrovsky JO, Lang AE, et al. Neurophysiological identification of the subthalamic nucleus in surgery for Parkinson's disease. *Ann Neurol.* John Wiley & Sons, Ltd; 1998;44: 622–628. doi:10.1002/ana.410440407
33. Limousin P, Pollak P, Benazzouz A, Hoffmann D, Le Bas J-F, Perret JE, et al. Effect on parkinsonian signs and symptoms of bilateral subthalamic nucleus stimulation. *Lancet.* Elsevier; 1995;345: 91–95. doi:10.1016/S0140-6736(95)90062-4
34. Limousin P, Krack P, Pollak P, Benazzouz A, Ardouin C, Hoffmann D, et al. Electrical Stimulation of the Subthalamic Nucleus in Advanced Parkinson's Disease. *N Engl J Med.* Massachusetts Medical Society ; 1998;339: 1105–1111. doi:10.1056/NEJM199810153391603
35. Lyons KE, Wilkinson SB, Overman J, Pahwa R. Surgical and hardware complications of subthalamic stimulation: A series of 160 procedures. *Neurology.* 2004. pp. 612–616. doi:10.1212/01.WNL.0000134650.91974.1A
36. Alkemade A, de Hollander G, Keuken MC, Schäfer A, Ott DVM, Schwarz J, et al. Comparison of T2*-weighted and QSM contrasts in Parkinson's disease to visualize the STN with MRI. Jiang Q, editor. *PLoS One.* Public Library of Science; 2017;12: e0176130. doi:10.1371/journal.pone.0176130
37. Caan MWA, Bazin P, Marques JP, Hollander G, Dumoulin S, Zwaag W. MP2RAGEME: T₁, T₂^{*}, and QSM mapping in one sequence at 7 tesla. *Hum Brain Mapp.* John Wiley & Sons, Ltd; 2018; hbm.24490. doi:10.1002/hbm.24490
38. Marques JP, Kober T, Krueger G, Van Der Zwaag W, Van De Moortele P-F, Gruetter R. MP2RAGE, a self bias-field corrected sequence for improved segmentation and T₁-mapping

3T v 7T MRI for imaging subcortex

- at high field. *Neuroimage*. 2009;49: 1271–1281. doi:10.1016/j.neuroimage.2009.10.002
39. Li W, Wang N, Yu F, Han H, Cao W, Romero R, et al. A method for estimating and removing streaking artifacts in quantitative susceptibility mapping. *Neuroimage*. 2015;108: 111–122. doi:10.1016/j.neuroimage.2014.12.043
40. Li W, Avram A V, Wu B, Xiao X, Liu C. Integrated Laplacian-based phase unwrapping and background phase removal for quantitative susceptibility mapping. *NMR Biomed*. NIH Public Access; 2014;27: 219–27. doi:10.1002/nbm.3056
41. Forstmann BU, Keuken MC, Schafer A, Bazin P, Alkemade A, Turner R. Multi-modal ultra-high resolution structural 7-Tesla MRI data repository. *Sci Data*. 2014;1: 140050.
42. Visser E, Keuken MC, Douaud G, Gaura V, Bachoud-Levi A-C, Remy P, et al. Automatic segmentation of the striatum and globus pallidus using MIST: Multimodal Image Segmentation Tool. *Neuroimage*. 2016;125: 479–497. doi:10.1016/j.neuroimage.2015.10.013
43. Visser E, Keuken MC, Forstmann BU, Jenkinson M. Automated segmentation of the substantia nigra, subthalamic nucleus and red nucleus in 7T data at young and old age. *Neuroimage*. 2016;139: 324–336. doi:10.1016/j.neuroimage.2016.06.039
44. Dice LR. Measurements of Amount of Ecologic Association between Species. *Ecology*. Ecological Society of America; 1945;26: 297–302. doi:10.2307/1932409
45. Keuken MC, Bazin PL, Schafer A, Neumann J, Turner R, Forstmann BU. Ultra-high 7T MRI of structural age-related changes of the subthalamic nucleus. *J Neurosci*. 2013;33: 4896–4900. doi:10.1523/JNEUROSCI.3241-12.2013
46. Ledez D. Euclidean distance mapping: geological applications. *Terra Nostra (Proc IAMG)*. 2002;4: 25–30.
47. Huntenburg JM, Steele CJ, Bazin P-L. Nighres: processing tools for high-resolution neuroimaging. *Gigascience*. Oxford Academic; 2018;7. doi:10.1093/gigascience/giy082
48. Morey RD, Wagenmakers EJ. Simple relation between Bayesian order-restricted and point-null hypothesis tests. *Stat Probab Lett*. North-Holland; 2014;92: 121–124. doi:10.1016/j.spl.2014.05.010
49. Team RC. R : A language and Environment for Statistical Computing. *J Comput Graph Stat*. 2015;5: 299–314.
50. Jeffreys H. *Theory of probability*. Oxford, UK: Oxford University Press. 1961;
51. Wetzels R, Matzke D, Lee MD, Rouder JN, Iverson GJ, Wagenmakers E-J. *Statistical Evidence in Experimental Psychology*. *Perspect Psychol Sci*. SAGE PublicationsSage CA: Los Angeles, CA; 2011;6: 291–298. doi:10.1177/1745691611406923
52. Isaacs BR, Trutti AC, Pelzer E, Tittgemeyer M, Temel Y, Forstmann BU, et al. Cortico-basal white matter alterations occurring in Parkinson's disease. *PLoS One*. Public Library of Science; 2019;14. doi:10.1371/journal.pone.0214343
53. Akaike H. Likelihood and the Bayes procedure. Springer, New York, NY; 1998. pp. 309–332. doi:10.1007/978-1-4612-1694-0_24
54. Rouder JN, Morey RD, Speckman PL, Province JM. Default Bayes factors for ANOVA designs. *J Math Psychol*. Academic Press; 2012;56: 356–374. doi:10.1016/j.jmp.2012.08.001

3T v 7T MRI for imaging subcortex

55. Evans NJ, Wagenmakers E-J. Theoretically meaningful models can answer clinically relevant questions. *Brain*. 2019;142: 1170–1172. doi:10.1093/brain/awz096
56. Berry DA, Hochberg Y. Bayesian perspectives on multiple comparisons. *J Stat Plan Inference*. 1999;82: 215–227. doi:10.1016/S0378-3758(99)00044-0
57. Perlmuter JS. Assessment of parkinson disease manifestations. *Current Protocols in Neuroscience*. 2009. doi:10.1002/0471142301.ns1001s49
58. Ranjan M, Boutet A, Xu DS, Lozano CS, Kumar R, Fasano A, et al. Subthalamic Nucleus Visualization on Routine Clinical Preoperative MRI Scans: A Retrospective Study of Clinical and Image Characteristics Predicting Its Visualization. *Stereotact Funct Neurosurg*. Karger Publishers; 2018;96: 120–126. doi:10.1159/000488397
59. Mulder MJ, Keuken MC, Bazin P-L, Alkemade A, Forstmann BU. Size and shape matter: The impact of voxel geometry on the identification of small nuclei. Bergsland N, editor. *PLoS One*. Public Library of Science; 2019;14: e0215382. doi:10.1371/journal.pone.0215382
60. Springer E, Dymerska B, Cardoso PL, Robinson SD, Weisstanner C, Wiest R, et al. Comparison of Routine Brain Imaging at 3 T and 7 T. *Invest Radiol*. 2016; doi:10.1097/RLI.0000000000000256
61. Knowles BR, Friedrich F, Fischer C, Paech D, Ladd ME. Beyond T2 and 3T: New MRI techniques for clinicians. *Clin Transl Radiat Oncol*. 2019; doi:10.1016/j.ctro.2019.04.009
62. Polders DL, Leemans A, Luijten PR, Hoogduin H. Uncertainty estimations for quantitative in vivo MRI T1 mapping. *J Magn Reson*. 2012;224: 53–60. doi:10.1016/j.jmr.2012.08.017
63. Despotović I, Goossens B, Philips W. MRI Segmentation of the Human Brain: Challenges, Methods, and Applications. *Comput Math Methods Med*. 2015; doi:10.1155/2015/450341
64. Fang J, Bao L, Li X, van Zijl PCM, Chen Z. Background field removal for susceptibility mapping of human brain with large susceptibility variations. *Magn Reson Med*. John Wiley and Sons Inc.; 2019;81: 2025–2037. doi:10.1002/mrm.27492
65. van Laar PJ, Oterdoom DLM, ter Horst GJ, van Hulzen ALJ, de Graaf EKL, Hoogduin H, et al. Surgical Accuracy of 3-Tesla Versus 7-Tesla Magnetic Resonance Imaging in Deep Brain Stimulation for Parkinson Disease. *World Neurosurg*. 2016;93: 410–412. doi:10.1016/j.wneu.2016.06.084
66. Hamel W, Köppen JA, Alesch F, Antonini A, Barcia JA, Bergman H, et al. Targeting of the Subthalamic Nucleus for Deep Brain Stimulation: A Survey Among Parkinson Disease Specialists. *World Neurosurg*. Elsevier; 2017;99: 41–46. doi:10.1016/J.WNEU.2016.11.012
67. Bot M, Verhagen O, Caan M, Potters W, Dilai J, Odekerken V, et al. Defining the dorsal STN border using 7.0-Tesla MRI: a comparison to microelectrode recordings and lower field strength MRI. *Brain Stimul*. Karger Publishers; 2019;12: 587. doi:10.1016/j.brs.2018.12.952
68. Wu Y, Ragin AB, Du H, Sidharthan S, Dunkle EE, Koktzoglou I, et al. Sub-millimeter isotropic MRI for segmentation of subcortical brain regions and brain visualization. *J Magn Reson Imaging*. 2010;31: 980–986. doi:10.1002/jmri.22120
69. Milchenko M, Norris SA, Poston K, Campbell MC, Ushe M, Perlmuter JS, et al. 7T MRI subthalamic nucleus atlas for use with 3T MRI. *J Med Imaging*. 2018;5: 1.

3T v 7T MRI for imaging subcortex

792 doi:10.1117/1.jmi.5.1.015002

793 70. van Osch MJP, Webb AG. Safety of Ultra-High Field MRI: What are the Specific Risks?

794 Current Radiology Reports. 2014. doi:10.1007/s40134-014-0061-0

795 71. Van Der Kolk AG, Hendrikse J, Zwanenburg JJM, Visser F, Luijten PR. Clinical applications of

796 7 T MRI in the brain. European Journal of Radiology. 2013. pp. 708–718.

797 doi:10.1016/j.ejrad.2011.07.007

798 72. Moser E, Stahlberg F, Ladd ME, Trattnig S. 7-T MR-from research to clinical applications?

799 NMR in Biomedicine. 2012. pp. 695–716. doi:10.1002/nbm.1794

800 73. Wang JY, Zhuang QQ, Zhu LB, Zhu H, Li T, Li R, et al. Meta-analysis of brain iron levels of

801 Parkinson's disease patients determined by postmortem and MRI measurements. Sci Rep.

802 Nature Publishing Group; 2016;6. doi:10.1038/srep36669

803

Figure 1

Figure 2

Figure 3