Reverse-transcription recombinase-aided amplification assay for H5 subtype avian influenza virus

Suchun Wang^{#1}, Yang Li^{#1}, Nan Jiang^{1,2}, Fuyou Zhang^{1,3}, Qingye Zhuang¹, Guangyu Hou¹, Lijian Jiang^{1,4}, Jianmin Yu¹, Xiaohui Yu¹, Hualei Liu¹, Chenglong Zhao^{1,5}, Liping Yuan¹, Baoxu Huang¹, Kaicheng Wang^{1*}

1. China Animal Health and Epidemiology Center, Qingdao, Shandong Province, China
2. Shennong Agricultural Industry Group, Kunming, Yunnan 650000, China
3. Shandong agricultural university, Taian, China
4. Guangdong Haid Institute of Animal Husbandry & Veterinary, Guangzhou, Guangdong, 511400, China
5. Yanbian University, Agricultural College, Yanji, Jilin 133002, China

*Correspondence author: China Animal Health and Epidemiology Center, 369 Nanjing Road, Qingdao, Shandong, China

E-mail address: wangkaicheng@cahec.cn

#: These authors contributed equally to this article.

Abstract: The H5 subtype Avian Influenza Virus has caused huge economic losses to the poultry industry and is a threat to human health. A rapid and simple test is needed to confirm infection in suspected cases during disease outbreaks. In this study, we developed a reverse-transcription recombinase-aided amplification assay for the detection of H5 subtype avian influenza virus. Assays were performed at a single temperature (39°C), and the results were obtained within 20 min. The assay showed no cross-detection with Newcastle disease virus or infectious bronchitis virus. The analytical sensitivity was 10³ RNA copies per reaction at a 95% confidence interval according to probit regression analysis, with 100% specificity. Compared with published reverse-transcription quantitative real-time polymerase chain reaction assays, the κ value of the reverse transcription recombinase-aided amplification assay in 365 avian clinical samples was 0.970 ($p < 0.001$). The sensitivity for avian clinical sample detection was 94.44% (95%CI, 70.63% - 99.71%), and the specificity was 100% (95%CI, 98.64% - 100%). These results indicated that our reverse-transcription recombinase-aided amplification assay may be a valuable tool for detecting H5 subtype avian influenza virus.

KEYWORDS: avian influenza, H5 subtype, reverse-transcription recombinase-aided amplification assay

INTRODUCTION:

Avian influenza virus (AIV) is a negative-sense RNA virus that belongs to the orthomyxoviridae family [1]. The AIV genome is composed of eight distinct RNA

segments encoding at least 10 proteins which coordinate functions, components and structure of the virus[2,3]. AIV can be classified into highly pathogenic avian influenza virus (HPAIV) and low pathogenic avian influenza virus (LPAIV) based on pathogenicity in chickens. Since the first H5N1 HPAIV was detected in 1996, these viruses have been prevalent among poultry in Asia, Europe, and Africa. The viruses constantly undergo genetic drift and shift that permanently threatens poultry industry and human health[4].

All LPAIV and HPAIV infections of subtypes H5 in poultry are notifiable to the World Organization for Animal Health (O.I.E.). Determination of the type of AIV is of utmost importance for the diagnosis of these infections. This can be achieved biologically by determination of the intravenous pathogenicity index (IVPI) in experimentally inoculated chickens or molecularly by nucleotide sequence analysis of the site encoding the AIV. Since animal experiment facilities or expensive equipment are required for either pathway, solutions for alternative techniques have been sought in the past. These included restriction enzyme cleavage patterns, probe hybridization and real time RT-PCR (RT-qPCR) approaches [5-13]. Based on the widespread availability of RT-qPCR technology in diagnostic laboratories and its recent favorable use in pathotyping of H5 subtype HPAIV of the goose/Guangdong (gs/GD) lineage.

Recently, rapid isothermal amplification techniques have been developed, such as loop-mediated isothermal amplification (LAMP) [14], recombinase polymerase amplification (RPA) [15], recombinase assisted amplification (RAA) [16] and strand displacement amplification (SDA) [16], and used for large-scale testing. Among these rapid nucleic acid detection methods, reverse transcription recombinase-mediated isothermal amplification (RT-RAA) is a rapid thermostatic nucleic acid amplification technology that utilizes a recombinant enzyme obtained from bacteria or fungi. At normal temperature, the recombinant enzyme can tightly bind to the primer DNA to form a polymer of enzymes and primers. When the primer searches the template DNA for a complementary sequence that perfectly matches it, with the help of a single-stranded DNA binding protein, open the double-stranded structure of the template DNA. Under the action of DNA polymerase, a new complementary DNA strand is formed, and the amplification product grows exponentially. This technology has the characteristics of high sensitivity, stronger specificity and reliability.

In this study, in order to efficiently detect H5 subtype AIV, a RT-RAA assay was designed and its analytical specificity and sensitivity were used to evaluate. Our study suggest that RT-RAA meets the need of field test, presents a rapid and sensitive detection method that can be used as an alternative to animal inoculation or nucleotide sequencing.

Materials and methods

Ethics statement

This study was conducted according to the animal welfare guidelines of the World Organization for Animal Health [17] and was approved by the Animal Welfare Committee of the China Animal Health and Epidemiology Center (CAHEC). CAHEC has permission to engage in activities of highly pathogenic avian influenza virus. The

swab samples were collected for this study after being granted permission by multiple relevant parties, including the Ministry of Agriculture and Rural Affairs of China, the China Animal Health and Epidemiology Center, the relevant veterinary sections of the provincial and county governments, and the relevant farm owners.

Samples and extraction of viral nucleic acids

All the AIV used in this study were isolated and identified in the National Avian Influenza Professional Laboratory in China Animal Health and Epidemiology Center and were stored at -80°C. The Newcastle disease virus (NDV) and infectious bronchitis virus (IBV) were all maintained at our lab. The samples were centrifuged at 12,000× g for 10 min and the supernatant from each sample was used for RNA extraction on the QIAxtractor platform using a QIAcube HT kit (Qiagen, Hilden) according to the manufacturer's instructions. The extracted RNA were stored at -80°C for subsequent tests.

Preparation of plasmid standard

A H5 subtype AIV plasmid standard was developed using reference strain A/duck/Yunnan/5310/2006(H5N1) (GenBank accession number CY030889). A 1776 bp fragment of the whole hemagglutinin (HA) gene of H5 subtype AIV was cloned into the pUC57 vector to quantify DNA copy number. The recombinant plasmid DNA were prepared with a SanPrep Column Plasmid Mini-Preps Kit (Sangon) and quantified using a Thermo Scientific Multiskan GO Microplate Photometer (Thermo Fisher Scientific). The DNA copy number was calculated according to the following formula: DNA copy number = (copy number/μl) = $[6.02 \times 10^{23} \times \text{plasmid concentration (ng/μl)} \times 10^{-9}] / [\text{DNA length} \times 660]$. The sensitivity of the RT-RAA was evaluated by real-time fluorescence detection, using a dilution series of standard recombinant plasmids ranging from 10⁶–10¹ DNA copies per reaction.

Design of H5 RT-RAA primers and exo-probes

To detect H5 subtype AIV, a total of 4636 available HA gene segments of H5 subtype AIV obtained from GenBank database were aligned, which contained the HA gene sequence of all currently circulating branches of H5 subtype AIV, and highly conserved regions were subsequently identified with Molecular Evolutionary Genetics Analysis (MEGA) software 6.0 [18, 19] for the design the gene-specific primers and probes. Primers were designed using OLIGO 7 software [20] and showed no major non-specific sequence similarities by BLAST analysis. Three H5 forward primers and five reverse primers were designed to select the best primers and probes in combination. The appropriate primers and probe used in this study were shown in Table 1. The 30th base at the 5' end of the probe was labeled with the FAM light-emitting group. The 30th base was connected to the abasic site Tetrahydrofuran (THF). The 31st base was labeled with the BHQ1 quenching group, and the 3' end was modified by C3-spacer blocking. All the primers and probes were synthesized by Sangon Biotech.

TABLE 1 Primer and probe sequences used for RT-RAA, RT-qPCR and

conventional PCR assays.

Primer	Sequence (5'-3')	Size (bp)	Gene	Source
H5- F	CAGTTTGAGGCYGTGGAAGGGAATTTAAYAA	32	HA	This study
H5- R	CTTGTCRTAAAGGTTCTTGACATTTGAGTCAT	32	HA	This study
H5- P	CTAGATGTCTGGACTTATAATGCTGAACT/i6FAMdT/ /THF/ /iBHQ1dT/GGTTCTCATGGAAAAT[C3-spacer]	47	HA	This study
H5+1456	ACGTATGACTATCCACAATACTCAG	25	HA	(Spackman E, 2002)
H5-1685	AGACCAGCTACCATGATTGC	20	HA	(Spackman E, 2002)
H5+1637	FAM-TCAACAGTGGCGAGTTCCTAGCA-TAMRA	24	HA	(Spackman E, 2002)

RT-RAA for detection of H5 subtype AIV

The appropriate primers and exo-probes were screened by H5 RAA assay and verified by H5 RT-RAA assay. According to the manufacturer's instructions, the RT-RAA reaction was performed with an RT exo kit in 50 µl reaction mixture including all the necessary enzymes and reagents for RT and DNA amplification in lyophilized pellets (Jiangsu Qitian Bio-Tech Co. Ltd.). The reaction mixture contained the following: 2 µl RNA template, 25 µl rehydration buffer, 15.7 µl ddH₂O, 2.5 µl of magnesium acetate, 2.1 µl of each primer (10 µM) and 0.6 µl target-specific RT-RAA exo-probe. For amplification, the tubes were then transferred to a tube holder in an RT-RAA fluorescence detection device (QT-RAA-F7200; Jiangsu Qitian Bio-Tech Co. Ltd.) set at 39°C for 20 min. Each run included nuclease-free water as a negative control.

Specificity, sensitivity and reproducibility of RT-RAA

Using the RNA of H5 subtype AIV as the template, a total of 7 groups of RNA with different concentrations were established for nucleic acid amplification under the optimal RT-RAA conditions. Then the selected appropriate primers and external probes were verified by H5 RT-RAA analysis. And the sensitivity was determined using serially diluted mixture of HA plasmids (each plasmid was 10⁷ copies/µL -10¹ copies/µL) as quantitative standards. Then take 2 µL as the reaction template, and perform RT-RAA amplification according to the aforementioned loading method, with eight replicates for each dilution.

The specificity of the RT-RAA assay for H5 subtype AIV was evaluated using four H5-positive AIVs, 10 other subtype AIVs (H1N2, H3N2, H4N2, H6N2, H7N3, H7N9, H9N2, H10N7, H11N9), two NDVs and two IBVs. These viruses are the main respiratory viruses affecting birds and were previously identified by our lab. The details of all the viruses tested are listed in Table 2.

TABLE 2 Samples tested in the study.

Sample	Virus	HA subtype(clade)	H5 RT-RAA assay	H5 RT-qPCR assay
--------	-------	-------------------	-----------------	------------------

K144	AIV	H5N1(2.3.2.1)	+	+
QD1	AIV	H5N2(7)	+	+
G2324	AIV	H5N6(2.3.4.4)	+	+
G2084	AIV	H5N6(2.3.4.4)	+	+
Q221	AIV	H1N2	-	-
X1330	AIV	H3N2	-	-
P174	AIV	H4N2	-	-
A1267	AIV	H6N2		
H7N3	AIV	H7N3	-	-
1605	AIV	H7N9	-	-
X169	AIV	H9N2	-	-
H9	AIV	H9N2	-	-
T55	AIV	H10N2	-	-
S82	AIV	H11N2	-	-
ND	NDV	/	-	-
JS1816	NDV	/	-	-
M41	IBV	/	-	-
H52	IBV	/	-	-

Detection and evaluation of clinical samples by H5 RT-RAA

About 365 Oral-pharyngeal and cloacal swab samples were collected from live poultry market, which were immediately placed into 1mL antibiotic-containing PBS as described above and then stored at -80°C until total nucleic acids were extracted with viral RNA/DNA extraction kit above. The avian clinical samples collected were evaluated the performance of the RT-RAA assay and compared it with the performance of published RT-qPCR assays as described previously for H5 subtype AIV. The primers and probe for the RT-qPCR assays are listed in Table 1[21, 22]. In addition, the positive clinical samples were confirmed by sequence analysis to verify the positive results in conventional PCR.

Statistical analysis

To determine the RT-RAA detection limit, a probit analysis was performed at a confidence interval of 95%, and the kappa and p values of RT-qPCR and RT-RAA were calculated. In addition, we calculated the sensitivity and specificity of RT-qPCR and RT-RAA for detection in clinical samples of poultry. All statistical analyses were performed in SPSS 21.0 (IBM).

RESULTS

Analytical sensitivity of RT-RAA

The sequence of the appropriate H5 RT-RAA primers and exo-probe are listed in Table 1, which has the best amplification efficiency under the same reaction conditions. The detection results of RT-RAA sensitivity assay are shown in Figure 1. The primer and probe combinations designed by the present invention have a RNA concentration of 10^7 copies / μ L, 10^6 copies/ μ L, 10^5 copies/ μ L, and 10^4 copies/ μ L, 10^3

copies / μ L, a fluorescence amplification curve appears. Therefore, the detection limit of H5 RT-RAA assay was 10^3 RNA copies/ μ L(Fig 1).

Fig 1 Analytical sensitivity of the H5 RT-RAA assay. A dilution range from 10^7 to 10^1 copies per reaction of H5 subtype AIV RNA molecular was, respectively, used to evaluate the detection limit of H5 RT-RAA assay, negative represents negative control.

1—7: 10^7 copies / μ L— 10^1 copies / μ L; 8: negative control.

Analytical specificity of RT-RAA

The results showed that the test group corresponding to the H5 subtype AIV RNA template showed normal fluorescence detection curves, and the other virus test groups and negative control groups did not show amplification curves. Thus, the RT-RAA assay did not cross-react with other subtype AIVs, NDVs and IBVVs and demonstrated high specificity for the detection of H5 subtype AIV(Fig 2).

Fig 2 Analytical specificity of the H5 RT-RAA assay. Detection signals were recorded by real-time fluorescence RT-RAA with four samples including H5 subtype AIVs (H5N1, H5N2 and H5N6), while no signals were detected from the fourteen samples including other subtype AIVs, NDVs, IBVs and negative controls.

Evaluation of the RT-RAA for clinical samples

The 365 conserved avian clinical samples collected from live-poultry markets were tested by RT-RAA assay and compared with RT-qPCR. A threshold cycle (CT) value of 36 was used as the cut-off for a positive result in RT-qPCR. RT-qPCR detected 18 of the 365 samples as positive (4.93%, 18/365), while RT-RAA correctly identified and differentiated 17 positive samples, with a sensitivity of 94.44% (95%CI, 70.63% - 99.71%) and 100% specificity (95%CI, 98.64% - 100%) (Table 3). The κ value for RT-RAA and RT-qPCR was 0.970 ($p < 0.001$). All the detected positive clinical samples were verified as H5 subtype AIV positive by conventional PCR and sequence analysis.

TABLE 3 Detection of H5 subtype AIV in avian clinical samples.

		RT-RAA		Total	Kappa (κ)	P-value of kappa	Sensitivity% (95%CI)	Specificity % (95%CI)
		Positive	Negative					
RT-qPCR	Positive	17	1	18	0.970	<0.001	94.44 (70.63~99.71)	100 (98.64~100)
	Negative	0	347	347				
Total		17	348	365				

DISCUSSION

The frequent antigen shift and antigen drift of AIV increase the difficulty of AIV detection. Among all the AIV subtypes, highly pathogenic H5 AIV often leads to high

morbidity and mortality in poultry. Nowadays, the H5 virus are widely prevalent and of significant concern to the poultry industry and public health in China. Therefore, early detection of AIV and H5 subtype is very necessary in the surveillance and control of AIV outbreaks. Until now, there were many methods used in AIV detection, such as gold immunochromatographic assay [23], microarray [24], immunosensor [25], immune-fluorescence [26] and enzyme linked immunosorbant assay[27]. However, these methods require complex and costly devices and are difficult to perform in the field [28]. So far, only detection methods based on nucleic acid sequence-based amplification [29], RT-LAMP, and RT-RPA [30] have been used for the rapid detection of the AIV in the field. Moreover, as the mutation rate of the H5 subtype AIV accelerates, the previous detection methods may not meet the actual detection needs, some of them fail to subtype AIV while others cannot be applied in early diagnosis owing to inadequate sensitivity [31,33], so a rapid diagnostic method capable of detecting all popular strains is needed.

RAA is a novel isothermal amplification and detection assay requiring only 20 min to complete, compared with approximately 1 and 3 hr, respectively, for RT-qPCR and conventional PCR. In PCR, thermocycling is required for double-stranded DNA separation, primer binding and amplification. There should be no more than 30 amplification cycles for pathogen detection in PCR. RAA could produce a positive signal in as little as 4 min, at half the cost of RT-qPCR [34]. Recombinase-aided amplification can be carried out using a portable device with no complicated processes, while the instruments for RT-qPCR and conventional PCR are much more expensive. Recombinase-aided amplification can also use reverse transcriptase and a fluorescent probe system to detect RNA amplicons in real time [35]. Until now, RAA has only been used for detecting human pathogens, including Salmonella, respiratory syncytial virus [36], coxsackievirus [34], hepatitis B virus [37] and Schistosoma japonicum-specific gene fragments[38] and there have been no previous reports of the use of RAA for detecting H5 subtype AIV[39].

In this study, an H5 RT-RAA assay was created, which indicating the potential value of our method in application of early detection and rapid diagnosis of the H5 subtype AIV. In the detection of experimental and clinical samples, this method showed higher sensitivity along with high efficiency. The results showed that the RT-qPCR detected 18 of the 365 samples as positive (4.93%, 18/365), while RT-RAA correctly identified and differentiated 17 positive samples, with a sensitivity of 94.44% (95%CI, 70.63% - 99.71%) and 100% specificity (95%CI, 98.64% - 100%) (Table 3). The κ value for RT-RAA and RT-qPCR was 0.970 ($p < 0.001$). Moreover, the clinical swab samples detected as positive in RT-RAA were also certified as positive by PCR detection and sequencing. According to the primer/exo-probe design method and the sample detection results, the RT-RAA method was considered to be suitable for detecting almost all H5 subtypes. To the best of our knowledge, this is the first RT-RAA method for the detection of multiple H5 clades in AIV.

In conclusion, the RT-RAA method established in our study can quickly and accurately identify H5 subtype AIV, including the current epidemic strains, which meets the need for H5 subtype AIV testing, It is expected that this RT-RAA for

detection of emerging H5 subtype AIV rapidly will be applied in surveillance of clinical samples in field experiments and provide a powerful and valuable tool for the control of H5 subtype AIV.

Acknowledgement

This work was supported by the National Key Research and Development Program of China (2017YFC120050)

Author Contributions

Conceptualization: Suchun Wang, Yang Li, Kai-Cheng Wang, Hualei Liu and Baoxu Huang.

Data Curation: Suchun Wang, Kai-Cheng Wang.

Formal Analysis: Suchun Wang, Yang Li.

Funding Acquisition: KaiCheng Wang, Baoxu Huang.

Investigation: Suchun Wang, Yang Li, Nan Jiang, Fuyou Zhang, Qingye Zhuang, Guangyu Hou.

Methodology: Suchun Wang, Yang Li, KaiCheng Wang.

Project Administration: KaiCheng Wang.

Resources: Guangyu Hou, Qingye Zhuang, Lijian Jiang, Jianmin Yu, Chenglong Zhao, Xiaohui Yu and Liping Yuan.

Supervision: KaiCheng Wang.

validation: Suchun Wang, Yang Li, Nan Jiang, Fuyou Zhang.

Visualization: Suchun Wang, Yang Li.

Writing – Original Draft Preparation: Suchun Wang, Yang Li.

Writing – Review & Editing: Suchun Wang, Yang Li, Kai-Cheng Wang, Hualei Liu, Baoxu Huang.

REFERENCES:

1. Walker G. T., Fraiser M. S., Schram J. L., Little M. C., Nadeau J. G., & Malinowski D. P. (1992). Strand displacement amplification – an isothermal, in vitro DNA amplification technique. *Nucleic Acids Research*, 20, 1691 – 1696. <https://doi.org/10.1093/nar/20.7.1691>.
2. Dadonaite B., Gilbertson B., Knight M. L., Trifkovic S., Rockman S., Laederach A., et al. (2019). The structure of the influenza A virus genome. *Nature Microbiology*, 4(11), 1781 – 1789. <https://doi.org/10.1038/s41564-019-0513-7>.
3. Nakajima K. (1997). Influenza virus genome structure and encoded proteins. *Nihon Rinsho*, 55, 2542 – 2546.
4. Peiris J. S. M., de Jong M. D., Guan Y. (2007). Avian influenza virus (H5N1): a threat to human health. *Clin. Microbiol. Rev*, 20, 243 – 267. <https://doi.org/10.1128/cmr.00037-06>.
5. Short K. R., Richard M., Verhagen J. H., van Riel D., Schrauwen E. J. A., van den Brand J. M. A., et al. (2015). One health, multiple challenges: The

- inter-species transmission of influenza A virus. *One Health*, 1, 1 – 13. <https://doi.org/10.1016/j.onehlt.2015.03.001>.
6. Peiris J., Yu W., Leung C., Cheung C., Ng W., Nicholls J., et al. (2004). Re-emergence of fatal human influenza A subtype H5N1 disease. *Lancet*, 363, 617-619. [https://doi.org/10.1016/S0140-6736\(04\)15595-5](https://doi.org/10.1016/S0140-6736(04)15595-5).
7. Lai S., Qin Y., Cowling B. J., Ren X., Wardrop N. A., Gilbert M., et al. (2016). Global epidemiology of avian influenza A H5N1 virus infection in humans, 1997 – 2015: a systematic review of individual case data. *Lancet Infect. Dis*, 16(7), e108 – e118. [https://doi.org/10.1016/S1473-3099\(16\)00153-5](https://doi.org/10.1016/S1473-3099(16)00153-5).
8. Bai H., Wang R. H., Hargis B., Lu H. G., Li Y. B. (2012). A SPR aptasensor for detection of avian influenza virus H5N1. *Sensors*, 12, 12506 – 12518. <https://doi.org/10.3390/s120912506>.
9. Xie Z. X., Pang Y. S., Liu J. B., Deng X. W., Tang X. F., et al. (2006). A multiplex RT-PCR for detection of type A influenza virus and differentiation of avian H5, H7, and H9 hemagglutinin subtypes. *Mol Cell Probes*, 20: 245 – 249. <https://doi.org/10.1016/j.mcp.2006.01.003>.
10. Yuen K. Y., Chan P. K. S., Peiris M., Tsang D. N. C., Que T. L., et al. (1998). Clinical features and rapid viral diagnosis of human disease associated with avian influenza A H5N1 virus. *Lancet*, 351(9101), 467 – 471. [https://doi.org/10.1016/S0140-6736\(98\)01182-9](https://doi.org/10.1016/S0140-6736(98)01182-9).
11. Velumani S., Du Q., Fenner B. J., Prabakaran M., Wee L. C., et al. (2008). Development of an antigen-capture ELISA for detection of H7 subtype avian influenza from experimentally infected chickens. *J Virol Methods*, 147(2), 219 – 225. <https://doi.org/10.1016/j.jviromet.2007.09.004>.
12. Qi X., Li X. H., Rider P., Fan W. X., Gu H. W., et al. (2009). Molecular Characterization of Highly Pathogenic H5N1 Avian Influenza A Viruses Isolated from Raccoon Dogs in China. *PLoS One*, 4(3), e4682. <https://doi.org/10.1371/journal.pone.0004682>.
13. Lee C. W., Suarez D. L. (2004). Application of real-time RT-PCR for the quantitation and competitive replication study of H5 and H7 subtype avian influenza virus. *J Virol Methods*, 119(2), 151 – 158. <https://doi.org/10.1016/j.jviromet.2004.03.014>.
14. Notomi T., Okayama H., Masubuchi H., Yonekawa T., Watanabe K., Amino N., & Hase T. (2000). Loop-mediated isothermal amplification of DNA. *Nucleic Acids Research*, 28, E63. <https://doi.org/10.1093/nar/28.12.e63>.
15. Lutz S., Weber P., Focke M., Faltin B., Hoffmann J., Muller C., et al. (2010). Microfluidic lab-on-a-foil for nucleic acid analysis based on isothermal recombinase polymerase amplification (RPA). *Lab on a Chip*, 10, 887 – 893. <https://doi.org/10.1039/b921140c>.
16. Zhang X., Guo L., Ma R., Cong L., Wu Z., Wei Y., et al. (2017). Rapid detection of Salmonella with Recombinase Aided Amplification. *Journal of Microbiol Methods*, 139, 202 – 204. <https://doi.org/10.1016/j.mimet.2017.06.011>.
17. Edgar R. C. (2004). MUSCLE: Multiple sequence alignment with high accuracy

- and high throughput. *Nucleic Acids Research*, 32, 1792 – 1797. <https://doi.org/10.1093/nar/gkh340>.
18. Hall B. G. (2013). Building phylogenetic trees from molecular data with MEGA. *Molecular Biology and Evolution*, 30, 1229 – 1235. <https://doi.org/10.1093/molbev/mst012>.
19. Tamura K., Stecher G., Peterson D., Filipski A., & Kumar S. (2013). MEGA6: Molecular evolutionary genetics analysis version 6.0. *Molecular Biology and Evolution*, 30, 2725 – 2729. <https://doi.org/10.1093/molbev/mst197>.
20. Rychlik W. (2007). OLIGO 7 primer analysis software. *Methods in Molecular Biology*, 402, 35 – 60. <https://doi.org/10.3732/apps.1200524>.
21. Hoffmann E., Stech J., Guan Y., Webster R. G., Perez D. R. (2001). Universal primer set for the full-length amplification of all influenza A viruses. *Arch Virol*, 146(12):2275-89. <https://doi.org/10.1007/s007050170002>.
22. Bao H., Wang X., Tao Q., Cai D., Wang F., & Chen H. (2009). Development of one step RT-PCR technique for detection of H7 subtype avian influenza. *Sheng Wu Gong Cheng Xue Bao*, 25, 1658 – 1663. <https://doi.org/10.3969/j.issn.1008-0589.2013.02.15>.
23. Peng D., Hu S., Hua Y., Xiao Y., Li Z., Wang X., et al. (2007). Comparison of a new gold-immunochromatographic assay for the detection of antibodies against avian influenza virus with hemagglutination inhibition and agar gel immunodiffusion assays. *Veterinary Immunology & Immunopathology*, 117(1 – 2), 17 – 25. <https://doi.org/10.1016/j.vetimm.2007.01.022>.
24. Gall A., Hoffmann B., Harder T., Grund C., Höper D., Beer M. (2009). Design and validation of a microarray for detection, hemagglutinin subtyping, and pathotyping of avian influenza viruses. *Journal of Clinical Microbiology*, 47(2), 327 – 34. <https://doi.org/10.1128/JCM.01330-08>.
25. Wang R., Wang Y., Lassiter K., Li Y., Hargis B., Tung S., et al. (2009). Interdigitated array microelectrode based impedance immunosensor for detection of avian influenza virus H5N1. *Talanta*, 79(2), 159 – 64. <https://doi.org/10.1016/j.talanta.2009.03.017>.
26. Zhao G. Y., Shu-Chun W. U., Wang L. N., Chen G. B., Cui S. J. (2008). A Disposable Amperometric Enzyme Immunosensor for Rapid Detection of Avian Influenza Virus. *Chinese Journal of Animal & Veterinary Sciences*, 39(10), 1442 – 8. <https://doi.org/10.1080/00207540801918588>.
27. Song J. L., Zhang W. D., Wang J. P., Zuo-Sheng L. I., Feng Z. L., Yuan-Yuan H. U., et al. (2008). Development of type and subtype-specific immunofluorescence techniques for detection of avian influenza virus. *Journal of Yunnan University*, 30(5), 526 – 30. <https://doi.org/10.3724/SP.J.1005.2008.01083>.
28. Boyle D. S., Lehman D. A., Lillis L., Peterson D., Singhal M., Armes N., et al. (2013). Rapid detection of HIV-1 Proviral DNA for early infant diagnosis using recombinase polymerase amplification. *mBio*, 4(2), e00135-13. <https://doi.org/10.1128/mBio.00135-13>.
29. Collins R. A., Ko L. S., Fung K. Y., Chan K. Y., Xing J., Lau L. T. (2003). Rapid and sensitive detection of avian influenza virus subtype H7 using NASBA.

- Biochemical and Biophysical Research Communications, 300, 507 – 515. [https://doi.org/10.1016/S0006-291X\(02\)02896-6](https://doi.org/10.1016/S0006-291X(02)02896-6).
30. Abd El Wahed A., Weidmann M., & Hufert F. T. (2015). Diagnostics-in-a-Suitcase: Development of a portable and rapid assay for the detection of the emerging avian influenza A (H7N9) virus. *Journal of Clinical Virology*, 69, 16 – 21. <https://doi.org/10.1016/j.jcv.2015.05.004>.
 31. Jin M., Chen H.(2004). Development of Enzyme-Linked Immunosorbent Assay with Nucleoprotein as Antigen for Detection of Antibodies to Avian Influenza Virus. *Avian Diseases*, 48(4), 870 – 8. <https://doi.org/10.1637/7226-062204r>.
 32. Snyder D. B. (1986). Latest developments in the enzyme-linked immunosorbent assay (ELISA). *Avian Diseases*, 30(1), 19 – 23. <https://doi.org/10.2307/1590607>.
 33. Zhou E. M., Cantin M. F. (1998). Evaluation of a competitive ELISA for detection of antibodies against avian influenza virus nucleoprotein. *Avian Diseases*, 42(3), 517 – 22. <https://doi.org/10.2307/1592678>.
 34. Yan T. F., Li X. N., Wang L., Chen C., Duan S. X., Qi J. J., et al. (2018). Development of a reverse transcription recombinase-aided amplification assay for the detection of coxsackievirus A10 and coxsackievirus A6 RNA. *Archives of Virology*, 163, 1455 – 1461. <https://doi.org/10.1007/s00705-018-3734-9>.
 35. Zhang X., Guo L., Ma R., Cong L., Wu Z., Wei Y., et al. (2017). Rapid detection of Salmonella with Recombinase Aided Amplification. *Journal of Microbiol Methods*, 139, 202 – 204. <https://doi.org/10.1016/j.mimet.2017.06.011>.
 36. Qi J., Li X., Zhang Y., Shen X., Song G., Pan J., et al. (2019). Development of a duplex reverse transcription recombinase-aided amplification assay for respiratory syncytial virus incorporating an internal control. *Archives of Virology*, 164, 1843 – 1850. <https://doi.org/10.1007/s00705-019-04230-z>.
 37. Shen X. X., Qiu F. Z., She L. P., Yan T. F., Zhao M. C., Qi J. J., et al. (2019). A rapid and sensitive recombinase aided amplification assay to detect hepatitis B virus without DNA extraction. *BMC Infectious Diseases*, 19, 229. <https://doi.org/10.1186/s12879-019-3814-9>.
 38. Song Z., Ting L., Kun Y., Wei L., Jian-Feng Z., Li-Chuan G., et al. (2018). Establishment of a recombinase-aided isothermal amplification technique to detect Schistosoma japonicum specific gene fragments. *Zhongguo Xue Xi Chong Bing Fang Zhi Za Zhi*, 30, 273 – 277. <https://doi.org/10.16250/j.32.1374.2018120>.
 39. Spackman E., Senne D. A., Myers T. J., et al. (2002). Development of a Real-Time Reverse Transcriptase PCR Assay for Type A Influenza Virus and the Avian H5 and H7 Hemagglutinin Subtypes[J]. *Journal of Clinical Microbiology*, 40(9), 3256-3260. <https://doi.org/10.1128/JCM.40.9.3256-3260.2002>.