Diffuse optical reconstructions of fNIRS data using Maximum Entropy on the Mean

Zhengchen Cai^{a,*}, Alexis Machado^b, Rasheda Arman Chowdhury^c, Amanda Spilkin^a, Thomas Vincent^d, Ümit Aydin^{a,e}, Giovanni Pellegrino^{b,f}, Jean-Marc Lina^{g,h}, Christophe Grova^{a,b,h,i}

^a*Department of Physics and PERFORM Centre, Concordia University, Montréal, Québec, Canada*

^b*Neurology and Neurosurgery Department, Montreal Neurological Institute (MNI), McGill University, Montréal, Québec, Canada*

^c*CHU Sainte-Justine Research Centre, Montréal, Québec, Canada*

^d*Centre de médecine préventive et d'activité physique, Montreal Heart Institute, Montreal Heart Institute, Montréal, Canada, Montréal, Québec, Canada*

^e*MRC Social, Genetic and Developmental Psychiatry Centre, Institute of Psychiatry, Psychology and Neuroscience, King's College London, PO80, London, UK*

^f*San Camillo Hospital IRCCS, Venice, Italy*

^g*École de technologie supérieure de l'Université du Québec, Canada*

^h*Centre de Recherches Mathématiques, Université de Montréal, Québec, Canada*

ⁱ*Multimodal Functional Imaging Lab, Biomedical Engineering Department, McGill University, Montréal, Québec, Canada*

Abstract

Functional near-infrared spectroscopy (fNIRS) measures the hemoglobin concentration changes associated with neuronal activity. Diffuse optical tomography (DOT) consists of reconstructing the optical density changes measured from scalp channels to the oxy-/deoxy-hemoglobin (i.e., HbO/HbR) concentration changes within the cortical regions. In the present study, we adapted a nonlinear source localization method developed and validated in the con-

*Corresponding author at: Concordia University, Department of Physics, Loyola Science Complex, 7141 Sherbrooke Street West Montréal, Québec, Canada, H4B 1R6. Tel: +1 514 848 2424 ext.3393.

Email address: zhengchen.cai@mail.concordia.ca (Zhengchen Cai)

text of Electro- and Magneto-Encephalography (EEG/MEG): the Maximum Entropy on the Mean (MEM), to solve the inverse problem of DOT reconstruction. We first introduced depth weighting strategy within the MEM framework for DOT reconstruction to avoid biasing the reconstruction results of DOT towards superficial regions. We also proposed a new initialization of the MEM model improving the temporal accuracy of the original MEM framework. To evaluate MEM performance and compare with widely used depth weighted Minimum Norm Estimate (MNE) inverse solution, we applied a realistic simulation scheme which contained 4000 simulations generated by 250 different seeds at different locations and 4 spatial extents ranging from 3 to 40cm^2 along the cortical surface. Our results showed that overall MEM provided more accurate DOT reconstructions than MNE. Moreover, we found that MEM was remained particularly robust in low signal-to-noise ratio (SNR) conditions. The proposed method was further illustrated by comparing to functional Magnetic Resonance Imaging (fMRI) activation maps, on real data involving finger tapping tasks with two different montages. The results showed that MEM provided more accurate HbO and HbR reconstructions in spatial agreement with the main fMRI cluster, when compared to MNE.

Keywords: fNIRS, Diffuse Optical Tomography (DOT), Maximum Entropy on the Mean (MEM), Minimum Norm Estimation (MNE), Depth weighting, Personalized Optimal Montage

Highlights

- We introduced a new fNIRS reconstruction method - Maximum Entropy on the Mean.
- We implemented depth weighting strategy within the MEM framework.
- We improved the temporal accuracy of the original MEM reconstruction.
- Performances of MEM and MNE were evaluated with realistic simulations and real data.
- MEM provided more accurate and robust reconstructions than MNE.

1. Introduction

Functional Near-infrared spectroscopy (fNIRS) is a non-invasive functional neuroimaging modality. It detects changes in oxy-/deoxy-hemoglobin (i.e., HbO/HbR) concentration within head tissues through the measurement of near-infrared light absorption using sources and detectors placed on the surface of the head (Scholkmann et al., 2014; Yücel et al., 2021). In continuous wave fNIRS, the conventional way to transform variations in optical density to HbO/HbR concentration changes at the level of each source-detector channel, is to apply the modified Beer Lambert Law (mBLL) (Delpy et al., 1988). This model assumes homogeneous concentration changes within the detecting region, i.e., ignoring the partial volume effects which indicates the absorption of light within the illuminated regions varies locally. This assumption reduces quantitative accuracy of HbO/HbR concentration changes

14 when dealing with focal hemodynamic changes (Boas et al., 2001; Strangman
15 et al., 2003).

16 In order to handle these important quantification biases associated with
17 sensor level based analysis, diffuse optical tomography (DOT) has been pro-
18 posed to reconstruct, from sensor level measures of the optical density, the
19 fluctuations of HbO/HbR concentrations within the brain (Arridge, 1999).
20 This technique not only provides better spatial localization accuracy and
21 resolution of the underlying hemodynamic responses (Boas et al., 2004a;
22 Joseph et al., 2006), but also avoids partial volume effect in classical mBLL,
23 hence achieves better quantitative estimation of HbO/HbR concentration
24 changes (Boas et al., 2001; Strangman et al., 2003). DOT has been applied
25 to reconstruct hemodynamic responses in sensory and motor cortex during
26 median-nerve stimulation (Dehghani et al., 2009; Hughes et al., 2004) and
27 finger tapping (Boas et al., 2004a; Yamashita et al., 2016); to conduct visual
28 cortex retinotopic mapping (Zeff et al., 2007; White and Culver, 2010; Egge-
29 brecht et al., 2012) and to simultaneous image hemodynamic responses over
30 the motor and visual cortex (White et al., 2009).

31 To formalize DOT reconstruction, one needs to solve two main problems.
32 The first one is the forward problem which estimates a forward model or sen-
33 sitivity matrix that maps local absorption changes within the brain to varia-
34 tions of optical density changes measured by each channel (Boas et al., 2002).
35 The second problem is the inverse problem which aims at reconstructing the
36 fluctuations of hemodynamic activity within the brain from scalp measure-
37 ments (Arridge, 2011). The forward problem can be solved by generating a
38 subject specific anatomical model, describing accurately propagation of light

within the head. Such anatomical model is obtained by segmenting anatomical Magnetic Resonance Imaging (MRI) data, typically into five tissues (i.e., scalp, skull, cerebro-spinal fluid (CSF), white matter and gray matter), before initializing absorption and scattering coefficients values for each tissue type and for each wavelength (Fang, 2010; Machado et al., 2018). Solving the inverse problem relies on solving an ill-posed problem which does not provide a unique solution, unless specific additional constraints are added. The most widely used inverse method in DOT is a linear approach based on Minimum Norm Estimate (MNE) originally proposed for solving the inverse problem of Magnetoencephalography (MEG) and Electroencephalography (EEG) source localization (Hämäläinen and Ilmoniemi, 1994). It minimizes the L_2 norm of the reconstruction error along with Tikhonov regularization (Boas et al., 2004b; Zeff et al., 2007; Dehghani et al., 2009; Eggebrecht et al., 2012, 2014; Tremblay et al., 2018). Other strategies to solve DOT inverse problem have also been considered, such as sparse regularization using the L_1 norm (Süzen et al., 2010; Okawa et al., 2011; Kavuri et al., 2012; Prakash et al., 2014; Tremblay et al., 2018) and Expectation Maximization (EM) algorithm (Cao et al., 2007). A non-linear method based on hierarchical Bayesian model for which inference is obtained through an iterative process (Shimokawa et al., 2012, 2013) has been proposed and applied on finger tapping experiments in (Yamashita et al., 2016).

Maximum Entropy on the Mean (MEM) framework was first proposed by Amblard et al., 2004 and then applied and carefully evaluated by our group in the context of EEG/MEG source imaging (Grova et al., 2006; Chowdhury et al., 2013). The MEM framework was specifically designed and evaluated

64 for its ability to recover spatially extended generators (Heers et al., 2016;
65 Pellegrino et al., 2016; Chowdhury et al., 2016; Grova et al., 2016). We
66 recently demonstrated its excellent performances when dealing with focal
67 sources (Hedrich et al., 2017) and when applied on clinical epilepsy data
68 (Chowdhury et al., 2018; Pellegrino et al., 2020). In addition to its unique
69 ability to recover the spatial extent of the underlying generators, we also
70 demonstrated MEM’s excellent accuracy in low SNR conditions, with the
71 ability to limit the influence of distant spurious sources (Chowdhury et al.,
72 2016; Hedrich et al., 2017; Heers et al., 2016; Pellegrino et al., 2020; von
73 Ellenrieder et al., 2016; Aydin et al., 2020).

74 We believe that these important aspects should be carefully considered
75 in the context of fNIRS reconstruction. The first one is the ability to ac-
76 curately recover the spatial extent of the underlying hemodynamic activity
77 for both focal and extended generators. The second one is to provide robust
78 reconstruction results when data SNR decreases, especially when considering
79 the fact that it is challenging to maintain a good intra-subject consistence
80 using continuous-wave fNIRS due to its relatively low SNR (Chen et al.,
81 2020). Therefore, our main objective was to adapt the MEM framework
82 for fNIRS reconstruction and carefully evaluate its performance. Moreover,
83 fNIRS reconstruction results tends to be biased towards more superficial re-
84 gions, because the light sensitivity profile decreases exponentially with the
85 depth of the generators (Strangman et al., 2013). To overcome this bias, we
86 implemented and evaluated a depth weighted variant of the MEM framework.

87 The article is organized as follows. The methodology of depth weighted
88 MEM for DOT is first presented. Then, we described our validation frame-

work using realistic simulations and associated validation metrics. fNIRS reconstruction using MEM was compared with widely used depth weighted Minimum Norm Estimate (MNE) inverse solution. Finally, illustrations of the methods on finger tapping fNIRS data set acquired with two different montages from 6 healthy subjects are provided and compared with functional Magnetic Resonance Imaging (fMRI) results.

2. Material and Methods

2.1. fNIRS reconstruction

To perform fNIRS reconstructions, the relationship between measured optical density changes on the scalp and wavelength specific absorption changes within head tissue is usually expressed using the following linear model (Arridge, 1999):

$$Y = AX + e \quad (1)$$

where Y is a matrix ($p \times t$) which represents the wavelength specific measurement of optical density changes in p channels at t time samples. X ($q \times t$) represents the unknown wavelength specific absorption changes in q locations along the cortex at time t . A ($p \times q$) is called the light sensitivity matrix which is actually the forward model relating absorption changes in the head to optical density changes measured in each channel. Finally, e ($p \times t$) models the additive measurement noise. Solving the fNIRS tomographic reconstruction problem consists in solving an inverse problem which can be seen as the estimation of matrix X (i.e. the amplitude for each location q at time t). However, this problem is ill-posed and admits an infinite number of possible

112 solutions. Therefore, solving the DOT inverse problem requires adding ad-
113 ditional prior information or regularization constraints to identify a unique
114 solution.

115 In DOT studies, anatomical constraints can be considered by defining the
116 reconstruction solution space (i.e. where q is located) within the gray matter
117 volume (Boas and Dale, 2005) or along the cortical surface (Huppert et al.,
118 2017; Machado et al., 2021). In EEG and MEG source localization studies
119 (Dale and Sereno, 1993; Grova et al., 2006; Chowdhury et al., 2013), it also
120 is common to constrain the reconstruction along the cortical surface. In this
121 study, the reconstruction space was considered as the mid surface defined as
122 the middle layer between gray matter/pial and gray/white matter interfaces
123 (Fischl et al., 2002).

124 2.2. Minimum Norm Estimation (MNE)

125 Minimum norm estimation is one of the most widely used reconstruction
126 methods in DOT (Zeff et al., 2007; Dehghani et al., 2009; White et al., 2009;
127 White and Culver, 2010; Eggebrecht et al., 2012, 2014; Yamashita et al.,
128 2016). Such estimation can be expressed using a Bayesian formulation which
129 solves the inverse problem by estimating the posterior distribution $P(X|Y) =$
130 $\frac{P(Y|X)P(X)}{P(Y)}$ (i.e. the probability distribution of parameter X conditioned on
131 data Y). A solution can be computed by imposing Gaussian distribution
132 priors on the generators X ($P(X) = N(0, \Sigma_s^{-1})$) and the noise e ($P(e) =$
133 $N(0, \Sigma_d^{-1})$). Σ_d is the inverse of the noise covariance which could be estimated
134 from baseline recordings. Σ_s is the inverse of the source covariance which is
135 assumed to be an identity matrix in conventional MNE.

136 The Maximum a Posteriori (MAP) estimator of the posterior distribution

137 $P(X|Y)$ can be obtained using maximum likelihood estimation:

$$\begin{aligned} \hat{X}_{MNE} &= \operatorname{argmin} (||Y - AX||_{\Sigma_d}^2 + \lambda ||X||_{\Sigma_s}^2) \\ &= (A^T \Sigma_d A + \lambda \Sigma_s)^{-1} A^T \Sigma_d Y \end{aligned} \quad (2)$$

139 where $\hat{X}_{MNE}$ is the reconstructed absorption changes along the cortical sur-
140 face. λ is a hyperparameter to regularize the inversion using the priori mini-
141 mum norm constraint $||X||_{\Sigma_s}^2$. In this study, we applied the standard L-Curve
142 method to estimate this λ as suggested in (Hansen, 2000).

143 2.3. Depth weighted MNE

144 Standard MNE solutions assumes $\Sigma_s = I$, which then tends to bias the
145 inverse solution towards the generators exhibiting large sensitivity in the for-
146 ward model, therefore the most superficial ones (Fuchs et al., 1999). When
147 compared to EEG-MEG source localization, such bias is even more pro-
148 nounced in fNIRS since within the forward model light sensitivity values
149 decrease exponentially with the depth (Strangman et al., 2013). This bias
150 can be compensated by scaling the source covariance matrix such that the
151 variances are equalized (van der Sluis, 1969; Fuchs et al., 1999). In the con-
152 text of DOT, depth weighted MNE has been proposed by Culver et al., 2003
153 as an approach to compensate this effect and applied in different studies (Zeff
154 et al., 2007; Dehghani et al., 2009; White et al., 2009; Eggebrecht et al., 2012,
155 2014). In practice, depth weighting can be formulated differently, here we
156 consider a generalized expression for the implementation of depth weighted
157 MNE as proposed in Lin et al., 2006. It consists in initializing the source
158 covariance matrix as $\Sigma_s^{-1/2} = \Lambda$, resulting in a so called depth weighted MNE

159 solution, described as follows:

$$\begin{aligned}
 \hat{X}_{dMNE} &= \operatorname{argmin} (||Y - AX||_{\Sigma_d}^2 + \lambda ||X||_{\Sigma_s}^2) \\
 &= (A^T \Sigma_d A + \lambda (\Lambda \Lambda^t)^{-1})^{-1} A^T \Sigma_d Y \\
 \operatorname{diag}(\Lambda) &= \frac{1}{\operatorname{diag}((A^T \Sigma_d A))^\omega}
 \end{aligned}
 \tag{3}$$

161 Depth weighted MNE solution takes into account the forward model A for
 162 each position in the brain and therefore penalizes most superficial regions
 163 exhibiting larger amplitude in A , by enhancing the contribution to deeper
 164 regions. ω is a weighting parameter tuning the amount of depth compen-
 165 sation to be applied. The larger is ω , the more depth compensation is con-
 166 sidered. $\omega = 0$ would therefore refer to no depth compensation and an
 167 identity source covariance model. $\omega = 0.5$ refers to standard depth weight-
 168 ing approach mentioned above. In the present study, we carefully evalu-
 169 ated the impact of this parameter on DOT accuracy with a set of ω values
 170 (i.e. $\omega = 0, 0.1, 0.3, 0.5, 0.7$ and 0.9).

171 2.4. Maximum Entropy on the Mean (MEM) for fNIRS 3D reconstruction

172 2.4.1. MEM framework

173 The main contribution of this study is the first adaptation and evalua-
 174 tion of MEM method (Amblard et al., 2004; Grova et al., 2006; Chowdhury
 175 et al., 2013) to perform DOT reconstructions in fNIRS. Within the MEM
 176 framework, the intensity of x , i.e. amplitude of X at each location q in Eq.1,
 177 is considered as a random variable, described by the following probability
 178 distribution $dp(x) = p(x)dx$. The Kullback-Leibler divergence or ν -entropy
 179 of $dp(x)$ relative to a prior distribution $d\nu(x)$ is defined as,

$$S_v(dp(x)) = - \int_x \log \left(\frac{dp(x)}{d\nu(x)} \right) dp(x) = - \int_x f(x) \log(f(x)) d\nu(x) \tag{4}$$

181 where $f(x)$ is the ν -density of $dp(x)$ defined as $dp(x) = f(x)d\nu(x)$. Following
 182 a Bayesian approach to introduce the data fit, we denote C_m as the set of
 183 probability distributions on x that explains the data on average:

$$184 \quad Y - [A|I_q] \begin{bmatrix} E_{dp}[x] \\ e \end{bmatrix} = 0, \quad dp \in C_m \quad (5)$$

185 where Y represents the measured optical density changes, $E_{dp}[x] = \int x dp(x)$
 186 represents the statistical expectation of x under the probability distribution
 187 dp , and I_q is an identity matrix of $(q \times q)$ dimension. Therefore, within the
 188 MEM framework, a unique solution of $dp(x)$ could be obtained,

$$189 \quad dp^*(x) = \operatorname{argmax}_{dp(x) \in C_m} (S_v(dp(x))) \quad (6)$$

190 The solution of $dp^*(x)$ can be solved by maximizing the ν -entropy which
 191 is a convex function. It is equivalent to minimize an unconstrained concave
 192 Lagrangian function i.e., $L(dp(x), \kappa, \lambda)$, along with two Lagrangian constraint
 193 parameters, i.e., κ and λ . It is finally equivalent to maximize a cost function
 194 $D(\lambda)$ which is described as,

$$195 \quad D(\lambda) = \lambda^T Y - F_v(A^T \lambda) - \frac{1}{2} \lambda^T \Sigma_d^{-1} (\Sigma_d^{-1})^T \lambda \quad (7)$$

196 where Σ_d^{-1} is the noise covariance matrix. F_v represents the free energy
 197 associated with reference $d\nu(x)$. It is important to mention that $D(\lambda)$ is now
 198 an optimization problem within a space of dimension equal to the number of
 199 sensors. Therefore, if we estimate $\lambda^* = \operatorname{argmax}_{\lambda} D(\lambda)$, the unique solution
 200 of MEM framework is then obtained from the gradient of the free energy.

$$201 \quad \hat{X}_{MEM} = \nabla_{\xi} F_v^*(\xi)|_{\xi=A^T \lambda^*} \quad (8)$$

202 For further details on MEM implementation and theory we refer the reader
 203 to (Amblard et al., 2004; Grova et al., 2006; Chowdhury et al., 2013).

204 2.4.2. Construction of the prior distribution for MEM estimation

205 To define the prior distribution $d\nu(x)$ mentioned above, we assumed that
206 brain activity can be depicted by a set of K non-overlapping and independent
207 cortical parcels. Then the reference distribution $d\nu(x)$ can be modeled as,

$$208 \quad d\nu(x) = \prod_{k=1}^K [(1 - \alpha_k)\delta(x_k) + \alpha_k N(\mu_k, \Sigma_k)] dx_k, \quad 0 < \alpha_k < 1 \quad (9)$$

209 Each cortical parcel k is characterized by an activation state, defined by the
210 hidden variable S_k , describing if the parcel is active or not. Therefore we
211 denote α_k as the probability of k^{th} parcel to be active, i.e., $Prob(S_k = 1)$. δ_k
212 is a Dirac function that allows to “switch off” the parcel when considered as
213 inactive (i.e., $S_k = 0$). $N(\mu_k, \Sigma_k)$ is a Gaussian distribution, describing the
214 distribution of absorptions changes within the k^{th} parcel, when the parcel
215 is considered as active ($S_k = 1$). This prior model, which is specific to our
216 MEM inference, offers a unique opportunity to switch off some parcels of the
217 model, resulting in accurate spatial reconstructions of the underlying activity
218 patterns with their spatial extent, as carefully studied and compared with
219 other Bayesian methods in Chowdhury et al., 2013.

220 The spatial clustering of the cortical surface into K non-overlapping par-
221 cel was obtained using a data driven parcellization (DDP) technique (La-
222 palme et al., 2006). DDP consisted in first applying a projection method, the
223 multivariate source prelocalization (MSP) technique (Mattout et al., 2005),
224 estimating a probability like coefficient (MSP score) between 0 and 1 for each
225 vertex of the cortical mesh, characterizing its contribution to the data. DDP
226 is then obtained by using a region growing algorithm, along the tessellated
227 cortical surface, starting from local MSP maxima. Once the parcellization is

done, the prior distribution $d\nu(x)$ is then a joint distribution expressed as the multiplication of individual distribution of each parcel in Eq.9 assuming statistical independence between parcels,

$$d\nu(x) = d\nu_1(q_1)d\nu_2(q_2)...d\nu_k(q_k)...d\nu_K(q_K) \quad (10)$$

where $d\nu(x)$ is the joint probability distribution of the prior, $d\nu_k(q_k)$ is the individual distribution of the parcel k described as Eq.9.

To initialize the prior in Eq.9, μ_k which is the mean of the Gaussian distribution, $N(\mu_k, \Sigma_k)$, was set to zero. Σ_k at each time point t , i.e. $\Sigma_k(t)$, was defined by Eq.11 according to (Chowdhury et al., 2013),

$$\begin{aligned} \Sigma_k(t) &= \eta(t)W_k(\sigma)^TW_k(\sigma) \\ \eta(t) &= 0.05\frac{1}{\mathcal{P}_k}\sum_{i \in \mathcal{P}_k}\hat{X}_{MNE}^2(i, t) \end{aligned} \quad (11)$$

where $W_k(\sigma)$ is a spatial smoothness matrix, defined by (Friston et al., 2008), which controls the local spatial smoothness within the parcel according to the geodesic surface neighborhood order. Same value of $\sigma = 0.6$ was used as in (Chowdhury et al., 2013). $\eta(t)$ was defined as 5% of the averaged energy of MNE solution within each parcel $\mathcal{P}_k$ at time t . Finally, we can substitute this initialization into Eq.9 to construct the prior distribution $d\nu(x)$, and then obtain the MEM solution using Eq.8.

It is worth mentioning that we did not use MNE solution as the prior of μ_k in Eq.9 at all, which was actually initialized to 0 in our framework. We only used 5% of the averaged energy of MNE solution, over the parcel k , to set the prior for covariance Σ_k . The posterior estimation of parameter μ_k was estimated from the Bayesian framework by conditioning with data.

Moreover, the prior of MEM framework is a mixture of activation probability α_k and a Gaussian distribution (see Eq.9), in which the prior for α_k was informed by a spatio-temporal extension of the MSP score (see Chowdhury et al., 2013 for further details). These aspects completely differentiate MEM from approaches that iteratively update reconstruction results initialized by a MNE solution.

2.4.3. Depth weighted MEM

In addition to adapting MEM for fNIRS reconstruction, we also implemented for the first time, depth weighting within the MEM framework. Two depth weighting parameters, ω_1 and ω_2 , were involved in this process. ω_1 was used to apply depth weighting on the source covariance matrix Σ_k of each parcel k in Eq.11. ω_2 was applied to solve the depth weighted MNE, as described in Eq.3, before using those prior to initialize the source covariance model within each parcel of the MEM model. Therefore, the standard MNE solution $\hat{X}_{MNE}(i, t)$ in Eq.11 was replaced by the depth weighted version of MNE solution $\hat{X}_{dMNE}(i, t)$ described by Eq.3. Consequently, the depth weighted version of $\Sigma_k(t)$ is now defined as,

$$\begin{aligned} \Sigma_k(t)_{dw} &= \Lambda_{\mathcal{P}_k} \eta(t)_{dw} W_k(\sigma)^T W_k(\sigma) \\ \eta(t)_{dw} &= 0.05 \frac{1}{\mathcal{P}_k} \sum_{i \in \mathcal{P}_k} \hat{X}_{dMNE}^2(i, t) \end{aligned} \quad (12)$$

where $\Lambda_{\mathcal{P}_k}$ is the depth weighting matrix for each parcel k , in which ω_1 was involved to construct this scaling matrix as described in Eq.3. This initialization followed the logic that depth weighting is in fact achieved by scaling the source covariance matrix. The other depth weighting parameter, ω_2 , was

considered when solving $\hat{X}_{dMNE}(i, t)$, therefore avoiding biasing the initialization of the source covariance with a standard MNE solution.

To comprehensively compare MEM and MNE and also to investigate the behavior of depth weighting, we first evaluated the reconstruction performance of MNE with different ω_2 (i.e. step of 0.1 from 0 to 0.9). Then two of these values (i.e. $\omega_2 = 0.3$ and 0.5) were selected for the comparison with MEM since they performed better than the others. Note that the following expressions of depth weighted MEM will be denoted as $\text{MEM}(\omega_1, \omega_2)$ to represent the different depth weighting strategies.

2.4.4. Accuracy of temporal dynamics

The last contribution of this study was to improve the temporal accuracy of MEM solutions. In classical MEM approach (Chowdhury et al., 2013), $\hat{X}_{MNE}(i, t)$ in Eq.12 was globally normalized by dividing by $\max_{i \in \Omega, t \in T}(\hat{X}_{MNE}(i, t))$, where Ω represents all the possible locations along the cortical surface and T is the whole time segment. Therefore, the constructed prior along the time actually contained the temporal scaled dynamics from MNE solution. To remove this effect, we performed local normalization for $\hat{X}_{dMNE}(i, t)$ at each time instance t , i.e., by dividing by $\max_{i \in \Omega}(\hat{X}_{dMNE}(i, t))$. This new feature would preserve the spatial information provided by prior distribution, while allowing MEM to estimate the temporal dynamics only from the data.

2.5. Validation of fNIRS reconstruction methods

We evaluated the performance of the two fNIRS reconstruction methods (i.e., MEM and MNE), first within a fully controlled environment involving the use of realistic simulations of fNIRS data, followed by evaluations on real

296 data acquired with a well controlled finger tapping paradigm. Two different
297 fNIRS montages were considered in those two proposed evaluations.

298 **Montage 1:** A full Double Density (DD) montage (see Fig.1) which is
299 a widely used fNIRS montage, was considered given that it allows sufficient
300 dense spatial coverage of fNIRS channel to allow local DOT (Kawaguchi
301 et al., 2007). One healthy subject underwent fNIRS acquisitions with this
302 DD montage, involving the two following sessions,

- 303 • A 10 minutes resting state session was acquired to add realistic physi-
304 ology noise to be considered in our realistic simulations. The subject was
305 seating on a comfortable armchair and instructed to keep the eyes open
306 and to remain awake. The optodes of the full DD montage (i.e. 8 sources
307 and 10 detectors resulting in 50 fNIRS channels) are presented in Fig.1e.
308 The montage composed of 6 second-order distance channels(1.5cm), 24
309 third-order channels(3cm) and 12 fourth-order channels with 3.35cm
310 distance. In addition, we also added one proximity detector paired for
311 each source to construct close distance channels (0.7cm) in order to
312 measure superficial signals within extra-cerebral tissues. To place the
313 montage with respect to the region of interest, the center of the mon-
314 tage was aligned with the center of the right "hand knob" area, which
315 controls the left hand movement (Raffin et al., 2015), projected on the
316 scalp surface and then each optodes were projected on the scalp surface
317 (see Fig.1d).
- 318 • The subject was asked to sequentially tap the left thumb against the
319 other digits around 2Hz, therefore the main elicited hemodynamic re-
320 sponse was indeed expected over the right hand knob area. The finger

321 tapping paradigm consisted in 10 blocks of 30s tapping task and each
322 of them was followed by a 30 to 35s resting period. The beginning/end
323 of each block was informed by an auditory cue.

Fig.1. fNIRS measurement montage 1 and the anatomical model considered for DOT forward model estimation. (a) Anatomical 3D MRI segmented in five tissues, namely, scalp (green), skull (brown), CSF (light green), gray matter (purple) and white matter (black). (b) Optical fluence of one optode calculated through Monte Carlo simulation of Photons within this head model, using MCXLab. (c) Sensitivity profile of the whole montage in volume space. (d) Sensitivity profile, i.e. the summation of sensitivity map of all channels, along the cortical surface. Green dots represent detectors, including one proximity detector 0.7cm for each source, and red dots represent sources. (e) double-density montage 1 considered for this acquisition. There were 50 channels in total, 12 of 3.8cm (black), 24 of 3cm (blue), 6 of 1.5cm (yellow) and 8 of close distance (0.7cm) channels.

324 **Montage 2:** A personalized optimal montage (see Fig.8) following the
 325 methodology we previously reported in Machado et al., 2018. First, the
 326 hand knob within right primary motor cortex was drawn manually along the
 327 cortical surface and defined as a target region of interest (ROI) using the
 328 Brainstorm software (Tadel et al., 2011). Then we applied optimal montage
 329 estimation (Machado et al., 2014, 2018) in order to estimate personalized
 330 montages, built to maximize a priori fNIRS sensitivity and spatial overlap
 331 between channels with respect to the target ROI. To ensure good spatial
 332 overlap between channels for local 3D reconstruction, we constructed person-
 333 alized optimal montages composed of 3 sources and 15 detectors (see Fig.7b).
 334 The source-detector distance was set to vary from $2cm$ to $4.5cm$ and each
 335 source was constrained such that it has to create channels with at least 13
 336 detectors. Finally, we also manually added 1 proximity channel, located at
 337 the center of the 3 sources. Five subjects underwent fNIRS acquisitions with
 338 personalized optimal montage during a similar finger tapping task as the one
 339 for montage 1, in which 20 blocks were acquired by alternating a task (period
 340 of $10s$) and a resting state period ranging from $30s$ to $60s$.

341 All 6 subjects have signed written informed consent forms for this study
 342 which was approved by the Central Committee of Research Ethics of the Min-
 343 ister of Health and Social Services Research Ethics Board, Québec, Canada.

344 *2.5.1. MRI and fMRI Data acquisitions*

345 Anatomical MRI data were acquired on those 6 healthy subjects (25 ± 6
 346 years old, right-handed) and were considered to generate realistic anatomical
 347 head models. MRI data were acquired in a GE 3T scanner at the PERFORM
 348 Center of Concordia University, Montréal, Canada. T1-weighted anatomical

images were acquired using the 3D BRAVO sequence ($1 \times 1 \times 1 \text{ mm}^3$, 192 axial slices, 256×256 matrix), whereas T2-weighted anatomical images were acquired using the 3D Cube T2 sequence ($1 \times 1 \times 1 \text{ mm}^3$ voxels, 168 sagittal slices, 256×256 matrix).

Participants also underwent functional MRI acquisition (without fNIRS) while performing the same finger opposition tasks considered in fNIRS. fMRI acquisition consisted in a gradient echo EPI sequence ($3.7 \times 3.7 \times 3.7 \text{ mm}^3$ voxels, 32 axial slices, $\text{TE} = 25 \text{ ms}$, $\text{TR} = 2,000 \text{ ms}$). fMRI Z-maps were generated by standard first-level fMRI analysis using FEAT from FSL v6.0.0 software (Smith et al., 2004; Jenkinson et al., 2012).

2.5.2. fNIRS Data acquisition

fNIRS acquisitions were conducted at the PERFORM Center of Concordia University using a Brainsight fNIRS device (Rogue Research Inc., Montréal, Canada), equipped with 16 dual wavelength sources (685 nm and 830 nm), 32 detectors and 16 proximity detectors (for short distance channels). All montages (i.e., double density and optimal montages) were installed to cover the right motor cortex. Knowing a priori the exact positions of fNIRS channels estimated on the anatomical MRI of each participant, we then used a 3D neuronavigation system (Brainsight TMS navigation system, Rogue Research Inc.) to guide the installation of the sensors on the scalp. Finally, every sensor was glued on the scalp using a clinical adhesive, collo-dion, to prevent motion and ensure good contact to the scalp (Yücel et al., 2014; Machado et al., 2018).

372 2.5.3. *fNIRS forward model estimation*

373 T1 and T2 weighted anatomical images were processed using FreeSurfer
374 V6.0 (Fischl et al., 2002) and Brain Extraction Tool2 (BET2) (Smith et al.,
375 2004) in FMRIB Software Library (FSL) to segment the head into 5 tissues
376 (i.e. scalp, skull, Cerebrospinal fluid (CSF), gray matter and white matter
377 see Fig.1a).

378 Same optical coefficients used in (Yücel et al., 2014; Machado et al., 2018)
379 for the two wavelengths considered during our fNIRS acquisition, $685nm$
380 and $830nm$, were assigned to each tissue type mentioned above. Fluences
381 of light for each optode (see Fig.1b) was estimated by Monte Carlo simula-
382 tions with 10^8 photons using MCXLAB developed by Fang and Boas, 2009;
383 Yu et al., 2018 (<http://mcx.sourceforge.net/cgi-bin/index.cgi>). Sen-
384 sitivity values were then computed using the adjoint formulation and were
385 normalized by the Rytov approximation (Arridge, 1999).

386 For each source-detector pair of our montages, the corresponding light
387 sensitivity map was first estimated in a volume space, and then further con-
388 strained to the 3D mask of gray matter tissue (see Fig.1c), as suggested in
389 Boas and Dale, 2005. Then, these sensitivity values within the gray mat-
390 ter volume were projected along the cortical surface (see Fig.1d and Fig.7c)
391 using the Voronoi based method proposed by (Grova et al., 2006). We con-
392 sidered the mid-surface from FreeSurfer as the cortical surface. This surface
393 was downsampled to 25,000 vertices. This volume to surface interpolation
394 method has the ability to preserve sulco-gyral morphology (Grova et al.,
395 2006). After the interpolation, the sensitivity value of each vertex of the
396 surface mesh represents the mean sensitivity of the corresponding volumetric

397 Voronoi cell (i.e., a set of voxels that have closest distances to a certain vertex
398 than to all other vertices).

399 *2.5.4. fNIRS data preprocessing*

400 Using the coefficient of variation of the fNIRS data, channels exhibiting a
401 standard deviation larger than 8% of the signal mean were rejected (Schmitz
402 et al., 2005; Schneider et al., 2011; Eggebrecht et al., 2012; Piper et al., 2014).
403 Superficial physiological fluctuations were regressed out at each channel using
404 the average of all proximity channels' (0.7cm) signals (Zeff et al., 2007). All
405 channels were then band-pass filtered between 0.01Hz and 0.1Hz using a 3rd
406 order Butterworth filter. Changes in optical density (i.e., ΔOD) were calcu-
407 lated using the conversion to log-ratio. Finally, ΔOD of finger tapping data
408 were block averaged around the task onsets. Note that since sensors were
409 glued with collodion, we observed very minimal motion during the acquisi-
410 tions. Real background signal considered to generate realistic simulations
411 also underwent the same preprocessing.

412 *2.5.5. Realistic Simulations of fNIRS Data*

413 We first considered realistic simulations of fNIRS data to evaluate DOT
414 methods within a fully controlled environment. To do so, theoretical task-
415 induced HbO/HbR concentration changes were simulated within cortical sur-
416 face regions with a variety of locations, areas and depths. Corresponding
417 optical density changes in the channel space were then computed by apply-
418 ing the corresponding fNIRS forward model, before adding real resting state
419 fNIRS baseline signal as realistic physiological noise at different signal to
420 noise ratio (SNR) levels.

As presented in Fig.2a, we defined three sets of evenly distributed seeds within the field of view of DOT reconstruction. The locations were selected with respect to the depth relative to the skull, namely we simulated 100 "superficial seeds", 100 "middle seeds" and 50 "deep seeds". The cortical regions in which we simulated an hemodynamic response were generated by region growing around those seeds, along the cortical surface. To simulate generators with different spatial extents (denoted here as Se), we considered four levels of neighborhood orders, growing geodesically along the cortical surface, resulting in spatial extents ranging from $Se = 3, 5, 7, 9$ (corresponding areas of 3 to 40 cm^2). For simplification, these cortical regions within which an hemodynamic response was simulated will be denoted as "generators" in this paper. For each vertex within a "generator", a canonical Hemodynamic Response Function (HRF) was convoluted with a simulated experimental paradigm which consisted in one block of 20s task surrounded by 60s pre/post- baseline period (Fig.2b). Simulated HbO/HbR fluctuations within the theoretical generator (Fig.2c) were then converted to the corresponding absorption changes of two wavelengths (i.e., 685nm and 830nm). After applying the forward model matrix A in Eq.1, we estimated the simulated, noise free, task induced ΔOD in all channels.

Fig.2. Workflow describing our proposed realistic fNIRS simulation framework.(a) 100 Superficial seeds (black dots), 100 Middle seeds (red dots), 50 Deep seeds (blue dots) with spatial extent of $Se = 3, 5, 7, 9$ neighbourhood order within the field of view. (b) Convolution of a canonical HRF model with an experimental block paradigm (60s before and 50s after the onset). (c) Simulated theoretical HbO/HbR fluctuations along the cortical surface within the corresponding generator. (d) Realistic simulations obtained by applying the fNIRS forward model and addition of the average of 10 trials of real fNIRS background measurements at 830nm. Time course of ΔOD of all channels with SNR of 5, 3, 2 and 1 respectively are presented

440 ΔOD of real resting state data were then used to add realistic fluctuations
 441 (noise) to these simulated signals. Over the 10min of recording, we randomly
 442 selected 10 baseline epochs of 120s each, free from any motion artifact by

visual inspection. To mimic a standard fNIRS block average response, realistic simulations were obtained by adding the average of these 10 real baseline epochs to the theoretical noise-free simulated ΔOD , at five SNR levels (i.e. SNR = 5, 3, 2, 1). SNR was calculated through the following equation,

$$SNR_{\lambda} = \frac{\max(abs(\Delta OD_{\lambda}[0, t_1]))}{mean(std(\Delta OD_{\lambda}[-t_0, 0]))} \quad (13)$$

where $\Delta OD_{\lambda}[0, t_1]$ is the optical density changes of a certain wavelength λ in all channels during the period from 0s to $t_1 = 60s$. $std(\Delta OD_{\lambda}[-t_0, 0])$ is the standard deviation of ΔOD_{λ} during baseline period along all channels. Simulated trials for each of four different SNR levels are illustrated in Fig.2d. A total number of 4000 realistic simulations were considered for this evaluation study, i.e., 250 (seeds) $\times$ 4 (spatial extents) $\times$ 4 (SNR levels). Note that resting state fNIRS baseline signal was preprocessed before adding to the simulated signals.

2.5.6. Validation metric

Following the validation metrics described in (Grova et al., 2006; Chowdhury et al., 2013, 2016; Hedrich et al., 2017), we applied 4 quantitative metrics to access the spatial and temporal accuracy of fNIRS 3D reconstructions. Further details on the computation of those four validation metrics are reported in Supplementary material S1.

- **Area Under the Receiver Operating Characteristic (ROC) curve (AUC)** was used to assess general reconstruction accuracy considering both sensitivity and specificity. AUC score was estimated as the area under the ROC curve, which was obtained by plotting sensi-

466 tivity as a function of (1- specificity). AUC ranges from 0 to 1, the
467 higher it is the more accurate the reconstruction is.

468 • **Minimum geodesic distance (Dmin)** measuring the geodesic dis-
469 tance in millimeters, following the circumvolutions of the cortical sur-
470 face, from the vertex that exhibited maximum of reconstructed activity
471 to the border of the ground truth. Low Dmin values indicate better
472 accuracy in estimating the location of the generator.

473 • **Spatial Dispersion (SD)** assessed the spatial spread of the estimated
474 generator distribution and the localization error. It is expressed in mil-
475 limeters. A reconstructed map with either large spatial spread around
476 the ground truth or large localization error would result in large SD
477 values.

478 • **Shape error(SE)** evaluated the temporal accuracy of the reconstruc-
479 tion. It was calculated as the root mean square of the difference between
480 the normalized reconstructed time course and the normalized ground
481 truth time course. Low SE values indicate high temporal accuracy of
482 the reconstruction.

483 *2.6. Statistics*

484 Throughout all of the quantitative evaluations among different methods
485 involving different depth weighting factors ω in the results section, Wilcoxon
486 signed rank test was applied to test the significance of the paired differences
487 between each comparison. For each statistical test, we reported the median
488 value of paired differences, together with its p-value (Bonferroni corrected).

We are only showing results at 830nm for simulations, since the ones from 690nm under the same SNR level would have provided similar reconstructed spatiotemporal maps except for the reversed amplitudes. However, reconstruction results on real data indeed involved both wavelengths.

3. Results

3.1. Evaluation of MEM v.s. MNE using realistic simulations

We first investigated the effects of depth weighting factor ω_2 selection for depth weighted MNE. To do so, we evaluated spatial and temporal performances of DOT reconstruction for a set of ω_2 (step of 0.1 from 0 to 0.9). Based on those results reported in the Supplementary material S2 and Fig.S1, we decided to consider that most accurate fNIRS reconstructions were obtained when considering $\omega_2 = 0.3$ and 0.5 for depth weighted MNE. Therefore only those two values were further considered for comparison with MEM reconstructions.

Comparison of the performance of MEM and MNE on superficial realistic simulations are presented in Table.1 and Fig.3, for 4 levels of spatial extent ($Se = 3, 5, 7, 9$), using boxplot distribution of the 4 validation metrics. We evaluated 3 depth weighted implementations of MEM, namely, MEM($\omega_1 = 0.3, \omega_2 = 0.3$), MEM($0.3, 0.5$) and MEM($0.5, 0.5$), as well as 2 depth weighted implementations of MNE, namely, MNE(0.3) and MNE(0.5).

For spatial accuracy, results evaluated using Dmin, we obtained median Dmin values of $0mm$ for all methods, indicating the peak of the reconstructed map, was indeed accurately localized inside the simulated generator. It is worth mentioning that MEM($0.5, 0.5$) provided few Dmin values larger than

513 $0mm$ in $Se = 3$ and $Se = 5$ cases, which consisted of superficial and focal
 514 generators. Since MEM accurately estimated the spatial extent, more depth
 515 weighting considered for MEM(0.5,0.5) could results in focal and deeper
 516 reconstruction, hence resulting in non-zero D_{min} values. On the other hand,
 517 MNE would over-estimate the size of the underlying generators, therefore
 518 resulting in $0mm$ D_{min} , but larger SD values in similar conditions.

519 When considering the general reconstruction accuracy using AUC, for
 520 focal generators such as $Se = 3$ and 5, we found significant larger AUC (see
 521 Table.1) for MEM(0.3,0.3) and MEM(0.3,0.5) when compared to the most
 522 accurate version of MNE, i.e., MNE(0.3). When considering more extended
 523 generators, i.e., $Se = 7$ and 9, MEM(0.3,0.5) and MEM(0.5,0.5) achieved
 524 significantly larger AUC than MNE(0.3). However, the AUC of MNE(0.5)
 525 was significantly larger than MEM(0.3,0.3) for $Se = 7$ as well as significantly
 526 larger than MEM(0.3,0.5) and MEM(0.5,0.5) for $Se = 9$.

527 In terms of spatial extent of the estimated generator distribution and
 528 the localization error, MEM provided significantly smaller SD values among
 529 all the comparisons. Finally, for temporal accuracy of the reconstruction
 530 represented by SE, MNE provided significantly lower values, but with a small
 531 difference (e.g., 0.01 or 0.02, see results on real data as a reference of this
 532 effect size), than MEM among all comparisons when $Se = 3, 5$.

Fig.3. Evaluation of the performances of MEM and MNE using realistic simulations involving superficial seeds for different spatial extent ($Se = 3, 5, 7, 9$). Boxplot representation of the distribution of four validation metrics for three depth weighted strategies of MEM and two depth weighted strategies of MNE, namely: MEM(0.3, 0.3) in blue, MEM(0.3, 0.5) in green, MEM(0.5, 0.5) in red, MNE(0.3) in magenta and MNE(0.5) in black. Results were obtained after DOT reconstruction of 830nm ΔOD .

Superficial Seeds		Se = 3		Se = 5		Se = 7		Se = 9	
		MNE (0.3)	MNE (0.5)	MNE (0.3)	MNE (0.5)	MNE (0.3)	MNE (0.5)	MNE (0.3)	MNE (0.5)
AUC	MEM (0.3, 0.3)	0.02*	0.06**	0.04**	0.03**	0.01	-0.04**	-0.01	-0.10**
	MEM (0.3, 0.5)	0.02*	0.05**	0.05**	0.04**	0.05**	0.00	0.05**	-0.03**
	MEM (0.5, 0.5)	-0.01	0.03	0.00	0.01	0.04**	0.00	0.07**	-0.01
Dmin	MEM (0.3, 0.3)	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00
	MEM (0.3, 0.5)	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00
	MEM (0.5, 0.5)	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00
SD	MEM (0.3, 0.3)	-4.26**	-8.31**	-4.48**	-7.63**	-4.16**	-6.43**	-3.85**	-6.28**
	MEM (0.3, 0.5)	-3.78**	-8.23**	-4.11**	-7.11**	-3.86**	-6.30**	-3.71**	-6.24**
	MEM (0.5, 0.5)	-1.64**	-5.56**	-2.60**	-4.97**	-2.90**	-4.79**	-2.84**	-5.01**
SE	MEM (0.3, 0.3)	0.02**	0.02**	0.01**	0.01**	0.01	0.01	0.01	0.01
	MEM (0.3, 0.5)	0.02**	0.02**	0.01**	0.01**	0.01	0.01	0.00	0.00
	MEM (0.5, 0.5)	0.02**	0.02**	0.01**	0.01**	0.00	0.00	0.00	0.00

Table1. Wilcoxon signed rank test results of reconstruction performance comparison of MEM and MNE in superficial seeds case. Median values of paired difference are presented in the table. p values were corrected for multiple comparisons using Bonferroni correction, * indicates $p < 0.01$ and ** represents $p < 0.001$. Median of the paired difference of each validation metrics is color coded as follows: green: MEM is significantly better than MNE, red: MNE is significantly better than MEM and gray: non-significance.

Similar comparison between MEM and MNE were conducted respectively for middle seed simulated generators and deep seed simulated generators. Results were overall reporting similar trends when comparing MEM and MNE methods for middle and deep seeds, and as expected more depth weighting resulted in more accurate reconstructions (described in details in supplementary material, Fig.S2 and Table.S1 for middle seeds, Fig.S3 and Table.S2 for deep seeds).

Fig.4. Comparisons of the reconstruction maps using MEM and MNE in realistic simulations. Three theoretical regions with spatial extent $Se = 5$ (11 cm^2) were selected near the hand knob at different depth. The first column presents the locations and the size of the generator along the cortical surface. (a) Superficial seed case with reconstructed maps reconstructed using all MEM and MNE implementations considered in this study. (b) Middle seed case with reconstructed maps reconstructed using all MEM and MNE implementations considered in this study. (c) Deep seed case with reconstructed maps reconstructed using all MEM and MNE implementations considered in this study. 20% inflated and zoomed maps are presented on the left corner of each figure. 100% inflated right hemisphere are presented on the right side. All the maps were normalized by their own global maximum and no threshold was applied.

540 To further illustrate the performance of MEM and MNE as a function of
541 the depth of the generator, we are presenting some reconstruction results in
542 Fig.4. Three generators with a spatial extent of $Se = 5$, were selected for this

illustration. They were all located around the right "hand knob" area, and were generated from a superficial, middle and deep seed respectively. The first column in Fig.4 shows the location and the size of the simulated generator, considered as our ground truth. The generator constructed from the superficial seed only covered the corresponding gyrus, whereas the generators constructed from the middle seed, included parts of the sulcus and the gyrus. Finally, when considering the deep seed, the simulated generator covered both walls of the sulcus, extended just a little on both gyri. For superficial case, MEM(0.3, 0.3) and MEM(0.3, 0.5) provided similar performances in term of visual evaluation of the results and quantitative evaluations ($AUC = 0.96$, $D_{min} = 0mm$, $SD = 1.94mm, 2.15mm$, $SE = 0.03$). On the other hand, for the same simulations, MNE(0.3) and MNE(0.5) resulted in less accurate reconstructions, spreading too much around the true generator, as confirmed by validation metric, exhibiting notably large SD values ($AUC = 0.86, 0.89$, $D_{min} = 0mm$, $SD = 9.84mm, 14.63mm$, $SE = 0.02$). When considering the simulation obtained with the middle seed, MEM(0.3, 0.5) retrieved accurately the gyrus part of the generator but missed the sulcus component, since less depth compensation was considered. When increasing depth sensitivity, MEM(0.5, 0.5) clearly outperformed all other methods, by retrieving both the gyrus and sulcus aspects of the generator, resulting in the largest $AUC = 0.98$ and the lowest $SD = 2.93mm$. MNE(0.3) was not able to recover the deepest aspects of the generator, but also exhibited a large spread outside the ground truth area as suggested by a large $SD = 9.69mm$. MNE(0.5) was able to retrieve the main generator, but also exhibited a large spatial spread of $SD = 10.16mm$. When considering the generators obtained from the

568 deep seed, MNE(0.3) only reconstructed part of gyrus, missing completely
569 the main sulcus aspect of the generator, resulting in low AUC of 0.57 and
570 large SD of $10.34mm$. MEM(0.3, 0.5) was not able to recover the deepest
571 aspects of the sulcus, but reconstructed accurately the sulci walls, resulting
572 in an AUC of 0.89 and a SD of $2.71mm$. MEM(0.5, 0.5) recovered the deep
573 simulated generator very accurately, as demonstrated by the excellent scores
574 ($AUC = 0.97$, $SD = 2.11mm$) when compared to MNE(0.5). For those three
575 simulations, all methods recovered the underlying time course of the activity
576 with similar accuracy (i.e., similar SE values). In supplementary material,
577 we added Video.1, illustrating the behavior of all the simulations and all
578 methods, following the same layout provided in Fig.4.

579 Note that for this quantitative evaluation of fNIRS reconstruction meth-
580 ods using realistic simulation framework, we considered fNIRS data at only
581 one wavelength ($830nm$). Using single wavelength in the context simulation
582 based evaluation is a common procedure in DOT literature (Zhan et al., 2012;
583 Dehghani et al., 2009; White and Culver, 2010; Okawa et al., 2011; Trem-
584 blay et al., 2018; Shimokawa et al., 2012, 2013), since we may expect overall
585 similar performances for $685nm$ wavelength under the same SNR level.

586 3.2. *Effects of depth weighting on the reconstructed generator as a function* 587 *of the depth and size of the simulated generators*

588 To summarize the effects of depth weighting in 3D fNIRS reconstructions,
589 we further investigated the validation metrics, AUC, SD and SE, as a function
590 of depth and size of the simulated generators. Dmin was not included due
591 to the fact that we did not find clear differences among methods throughout
592 all simulation parameters from previous results. In the top row of Fig.5,

250 generators created from all 250 seeds with a spatial extent of $Se = 5$ were selected to demonstrate the performance of different versions of depth weighting as a function of the average depth of the generator. Whereas in the bottom row of Fig.5, we considered 400 generators constructed from all 100 superficial seeds with 4 different spatial extents of $Se = 3, 5, 7, 9$, to illustrate the performance of different versions of depth weighting as a function of the size of the generator. According to AUC, depth weighting was indeed necessary for all methods when the generator moved to deeper regions ($> 2cm$) as well as when the size was larger than $20cm^2$. Moreover, any version of MEM always exhibited clearly less false positives, as indicated by lower SD values, than all of MNE versions, whatever was the depth or the size of the underlying generator. We found no clear trend and difference of temporal accuracy among methods when reconstructing generators of different depths and sizes.

3.3. Robustness of 3D reconstructions to the noise level

All previous investigations were obtained from simulations obtained with a SNR of 5, in this section we compared the effect of the SNR level in Fig.6, on depth weighted versions of MNE and MEM, for superficial seeds only and generators of spatial extent $Se = 5$. We only compared MEM(0.3, 0.5) and MNE(0.5) considering the observation from previous results that these two methods were overall exhibiting best performances in this condition. Regarding Dmin, paired differences were not significant but MNE exhibited more Dmin values above $0mm$ than MEM at all SNR levels, suggesting that MNE often missed the main generators while MEM was more accurate in reconstructing the maximum of activity within the simulated generator. Regard-

ing AUC, MEM(0.3, 0.5) exhibited values higher than 0.8 at all SNR levels, whereas MNE(0.5) failed to recover accurately the generator for $SNR = 1$. Besides, in Table.2, we found that difference of AUC between MEM and MNE increased when SNR level decreased, suggesting the good robustness of MEM when decreasing the SNR level. The difference of SD also increased when SNR levels decreased. Indeed, MEM exhibited stable SD values among most SNR levels (except $SNR = 1$), whereas for MNE SD values were highly influenced by the SNR level. Finally, for both methods, decreasing SNR levels resulted in less accurate time course estimation (SE increased), slightly more for MEM when compared to MNE.

Fig.5. Effects of depth weighting on the depth and size of the simulated generators. First row demonstrates the validation matrices, AUC, SD and SE, as a function of depth of generators. We selected 250 generators created from all 250 seeds with a spatial extent of $SD = 5$. Depth was calculated by the average of minimum Euclidean distance from each vertex, within each generator, to the head surface. Second row demonstrates the validation matrices, AUC, SD and SE, as a function of size of generators. Involving 400 generators which constructed from 100 superficial seeds with 4 different spatial extend of $Se = 3, 5, 7, 9$. Line fittings were performed via a 4 knots spline function to estimate the smoothed trend and the shade areas represent 95% confident interval. Color coded points represent the values of validation matrices of all involved generators.

Fig.6. Evaluation of the performances of MEM and MNE at four different SNR levels. Boxplot representation of the distribution of four validation metrics for MEM(0.3,0.5) and MNE(0.5) involving superficial seeds with spatial extent $Se = 5$. SNR levels ($SNR = 1, 2, 3, 5$) are represented using different colors.

$Se = 5$ ($\sim 11 \text{ cm}^2$)		SNR = 1	SNR = 2	SNR = 3	SNR = 5
		MNE (0.5)	MNE (0.5)	MNE (0.5)	MNE (0.5)
AUC	MEM (0.3, 0.5)	0.14**	0.07**	0.05**	0.04**
Dmin	MEM (0.3, 0.5)	0.00	0.00	0.00	0.00
SD	MEM (0.3, 0.5)	-17.63**	-12.40**	-9.22**	-7.11**
SE	MEM (0.3, 0.5)	0.05**	0.03**	0.02**	0.01**

Table.2. Reconstruction performance comparison of MEM and MNE with different SNR levels. Median of paired difference of validation metric (i.e. AUC, Dmin, SD and SE) values of $Se = 5$ are presented in the table following the SNR increase from 1 to 5. ** indicates corrected $p < 0.001$.

628 3.4. Evaluation of MEM and MNE on real fNIRS data

629 For all finger tapping fNIRS data considered in our evaluations, two
630 wavelength (i.e., 685nm and 830nm) were reconstructed first and then con-
631 verted to HbO/HbR concentration changes along cortical surface using spe-
632 cific absorption coefficients. All the processes from fNIRS preprocessing to
633 3D reconstruction were completed in Brainstorm (Tadel et al., 2011) us-
634 ing the NIRSTORM plugin developed by our team ([https://github.com/](https://github.com/Nirstorm)
635 **Nirstorm**). For full double density montage (montage 1), reconstructed HbR
636 amplitudes were reversed to positive phase and normalized to their own
637 global maximum, to facilitate comparisons. In Fig.7.a, we showed the re-
638 constructed HbR maps at the peak of the time course (i.e., 31s) for MEM
639 and MNE by considering the 4 depth weighted versions, previously evalu-
640 ated, i.e., MEM(0.3, 0.3), MEM(0.3, 0.5), MNE(0.3) and MNE(0.5). The
641 two depth weighted versions of MEM clearly localized well the "hand knob"
642 region, while exhibiting very little false positives in its surrounding. On the
643 other hand, both depth weighted version of MNE clearly overestimated the
644 size of the hand knob region and were also exhibiting some distant possibly
645 spurious activity. The fMRI Z-map obtained during the corresponding fMRI
646 task is presented on Fig.7.b, after projection of the volume Z-map on the
647 cortical surface. Fig.7.c showed the time courses within the region of inter-
648 est representing the "hand knob". Each curve represents the reconstructed
649 time course of one vertex of the hand knob region and the amplitude were
650 normalized by the peak value within the whole region.

Fig.7. Application of MEM versus MNE reconstruction of HbR during a finger tapping task on one healthy subject. (a) Reconstructed maps of HbR (e.g. 20% inflation on the left and 100% inflation on the right side.) from MEM and MNE with different depth compensations. Each map was normalized by its own global maximum. (b) fMRI Z-map results projected along the cortical surface. (c) Reconstructed time courses of HbR within the hand knob region from MEM and MNE. Note that the hand knob region, represented by the black profile, was also matched well with the mean cluster of fMRI activation map on primary motor cortex. No statistical threshold was applied on fNIRS reconstructions.

Fig.8. Personalized fNIRS montage and comparisons between MEM and MNE reconstructions with respect to fMRI Z-map at individual level. a) the region of interest defined as the hand knob, b) optimal montage targeting the ROI consisting 3 sources (red) and 15 detectors (green) and one proximity (in the center of sources not shown), c) normalized sensitivity profile of the optimal montage which calculated as the sum of all channels sensitivity along the cortical surface, d) optimal montage glued on the scalp of the one subject, using collodion. fMRI Z-map of each subject during finger tapping task (threshold with $Z > 3.1$, Bonferroni corrected), black profile represents the main cluster along M1 and S1. MEM reconstruction maps at the corresponding HbO/HbR peak times, using depth weighted option 0.3, 0.3. MNE reconstruction maps, at the corresponding HbO/HbR peak times, using depth weighted option 0.3. Reconstructed time courses within the black profile, solid lines represent the main time courses and the shade areas represent standard deviation within the region of interest. Reconstructed time courses were normalized by the maximum amplitude, for each method respectively, before averaging.

Results obtained on 5 subjects for acquisition involving personalized optimal fNIRS montage (montage 2) and corresponding fNIRS reconstructions are presented in Fig.8. For every subject, fMRI Z-maps are presented along

the left hemisphere only and thresholded at $Z > 3.1$ ($p < 0.01$, corrected using Gaussian random field theory), The most significant fMRI cluster along M1 and S1 was delineated using a black profile. Reconstruction maps at the corresponding HbO/HbR peaks are then presented. Similar accuracy between MEM and MNE, with good overlap with fMRI results, were found for subjects 4 and 5, while MNE was overestimating the spatial extent of the generator. For subject 1, 2 and 3, MNE exhibited poor spatial correspondence with fMRI results. Averaged reconstructed time courses within the fMRI main cluster region are shown with standard deviation as the error bar. Comparing to simulations results, MEM exhibited overall very similar time course estimations than MNE in all cases. Considering the task duration was 10s, the reconstructed peak timing of HbO/HbR appeared accurately within the range of 10s to 20s.

4. Discussion

4.1. Spatial accuracy of 3D fNIRS reconstruction using MEM

In the present study, we first adapted the MEM framework in the context of 3D fNIRS reconstruction and extensively validated its performance. The spatial performance of reconstructions can be considered in two aspects, 1) correctly localizing the peak of the reconstructed map close enough to the ground truth area, 2) accurately recovering the spatial extent of the generator. According to our comprehensive evaluations of the proposed depth-weighted implementations of MEM and MNE methods, accurate localization was overall not difficult to achieve as suggested by our results using Dmin metric. Almost all methods provided median value of Dmin to be 0mm in all

simulation conditions except for the lowest $SNR = 1$ condition where more localization error was found. On the other hand, recovering the actual spatial extent of the underlying generator is actually the most challenging task in fNIRS reconstruction. When considering the results of MNE on both realistic simulations and real finger tapping tasks, either from visual inspection (Fig.4, Fig.7 and Fig.8) or quantitative evaluation by SD (Fig.3, Table.1 and supplementary section S2), we found that MNE overall reconstructed well the main generator but largely overestimated the size of the underlying generator. MEM was specifically developed, in the context of EEG/MEG source imaging, as a method able to recover the spatial extent of the underlying generators, which has been proved not to be the case for MNE-based approaches (Chowdhury et al., 2013, 2016; Grova et al., 2016; Hedrich et al., 2017; Pellegrino et al., 2020). A recent review (Sohrabpour and He, 2021) in the context of EEG/MEG source imaging has also demonstrated that the Bayesian approach with sparsity constraints is required to accurately estimate the spatial extent. These important properties of MEM was successfully demonstrated in our results on fNIRS reconstructions. These excellent performances were reliable for different sizes and depths of simulated generators, and for real finger tapping fNIRS data as well.

4.2. Importance of depth weighting in 3D fNIRS reconstruction

Biophysics models of light diffusion in living tissue are clearly demonstrating that, at all source-detector separations, light sensitivity decreases exponentially with depth (Strangman et al., 2013). The general solution to grant the ability of depth sensitivity compensation in fNIRS reconstruction is to introduce depth weighting during the reconstruction. In this study, we

investigated the impact of depth weighting effects on fNIRS reconstruction, as a function of the location and the spatial extent of the underlying generators. Our results are showing that when considering little or no depth weighting ($\omega = 0.0$ and 0.1) only most superficial generators along the gyral crown were accurately reconstructed missing the deepest parts, therefore resulting in low AUC values. On the other hand, larger depth weighted values, $\omega = 0.7$ and 0.9 , would bias too much the importance of deep generators and consequently, the most superficial aspects of the underlying generators were not recovered. According to our detailed evaluation on MNE reported in Fig.S1, depth weighted values of $\omega = 0.3$ and 0.5 were considered as good candidates to offer an ideal trade off. As expected, MNE(0.5) reported larger spatial dispersion around the true generator, than MNE(0.3). Depth weighting was also important when recovering more extended generators ($> 20cm^2$, Fig.5), for both MNE and MEM, since those extended generators were actually involving both superficial and deep regions.

4.3. Implementation of depth weighting strategy within the MEM framework

In this study, we are proposing for the first time a depth weighting strategy within the MEM framework, by introducing two parameters: ω_1 acting on scaling the source covariance matrix, and ω_2 tuning the initialization of the reference for MEM. When compared to depth weighted MNE, the MEM framework demonstrated its ability to reconstruct, different depth of focal generators as well as larger size generators, exhibiting excellent accuracy and few false positives (see Fig.5). When considering deeper focal generators ($depth > 2cm$), MEM(0.5,0.5) clearly outperformed all other methods (see AUC and SD values in Fig 5). In summary, for a large range of depths and

spatial extents of the underlying generators, MEM methods exhibited accurate results (large AUC values) and less false positives (lower SD values) when compared to MNE methods.

In practice, we would suggest to consider either $\omega_2 = 0.3$ or 0.5 for the initialization of MEM in all cases and only tune ω_1 . This is due to the fact that MNE(0.3 or 0.5) provided a generally good reconstruction with larger true positive rate in most scenarios, therefore providing MEM an accurate reference model ($d\nu(x)$) to start with. Even when considering the most focal simulated generators ($Se = 3$) case (see Fig.3, Table.1 and Fig.5), MEM($0.3, 0.3$) and MEM($0.3, 0.5$) were actually exhibiting very similar performances. Our proposed suggestion to tune ω_1 and ω_2 parameters was actually further confirmed when considered results obtained from real data. For both montages, MEM($0.3, 0.3$) results in excellent spatial agreement with fMRI Z-maps.

Note that depth weighting was also considered in DOT studies using MNE (Culver et al., 2003; Zeff et al., 2007; Dehghani et al., 2009; White et al., 2009; Eggebrecht et al., 2012, 2014) and a hierarchical Bayesian DOT algorithm (Shimokawa et al., 2012, 2013; Yamashita et al., 2016). A spatially-variant regularization parameter β was added to a diagonal regularization matrix featuring the sensitivity of every generator (forward model), and the value of β was tuned according to the sensitivity value of a certain depth. In practice, this strategy would result in similar depth compensation as ours, but we preferred the depth weighting parameter ω which mapped the amount of compensation from 0 to 1 (as described in Eq.3) for easier interpretation and comparison. This is also a standard procedure introduced in EEG/MEG source localization studies (Fuchs et al., 1999; Lin et al., 2006). Finally, using

the depth weighted MNE solution as the prior is a common consideration in Hierarchical Bayesian framework based fNIRS reconstructions (Shimokawa et al., 2012, 2013; Yamashita et al., 2016)

4.4. Temporal accuracy of 3D fNIRS reconstruction using MEM

Another important contribution of this study was that we improved the temporal accuracy time courses estimated within the MEM framework, resulting in similar temporal accuracy the one obtained with MNE. For instance, the largest significant SE difference between MEM and MNE was only 0.02 for $Se = 3$ and 0.01 for $Se = 5$. Corresponding time course estimations are also reported for MEM and MNE in real data (Fig.7 and Fig.8), suggesting again very similar performances. For instance, SE between MEM and MNE HbO time course was estimated as 0.02 for *Sub05* in Fig.8. Moreover, we found no significant SE differences between MEM and MNE for more extended generators ($Se = 7,9$). These findings are important considering that MNE is just a linear projection therefore the shape of the reconstruction will directly depend on the averaged signal at the channel level. On the other hand, MEM is a nonlinear technique, applied at every time sample, which is not optimized for the estimation of resulting time courses.

4.5. Robustness of fNIRS reconstructions to the noise level

To further investigate the effects of the amount of realistic noise in our reconstructions on both reconstruction methods, we performed the comparisons along 4 different SNR levels, i.e., $SNR = 1, 2, 3, 5$. As shown in Fig.6 and Table.2, we found that MEM was overall more robust than MNE when dealing with simulated signals at lower SNR levels. This is actually a very

important result since when reconstructing HbO/HbR responses, one has to consider at least two ΔOD of two different wavelengths exhibiting different SNR levels. For the simulation results, we reported reconstruction results obtained from $830nm$ data, whereas when considering real data (Fig.7 and Fig.8), we had to convert the reconstruction absorption changes at $685nm$ and $830nm$ into HbO/HbR concentration changes. Therefore, our final results were influenced by the SNR of all involved wavelengths.

fNIRS is inherently sensitive to inter-subject variability (Novi et al., 2020), as also suggested in our application on real data presented in Fig.8. Data from *Sub05* were exhibiting a good SNR level and therefore both MEM and MNE reconstructed accurately the main cluster of the activation, while MNE presented more spatial spread and false positive activation outside the fMRI ROI. When considering subjects for whom we obtained lower SNR data, e.g., *Sub02* and *Sub03*, MEM still recovered an activation map similar to fMRI map. In those cases, MNE not only reported suspicious activation pattern but also incorrectly reconstruct the peak amplitude outside the fMRI ROI. Our results suggesting MEM robustness in low SNR conditions for DOT are actually aligned with similar findings suggested for EEG/MEG source imaging, when considering source localization of single trial data (Chowdhury et al., 2018; Aydin et al., 2020).

4.6. Comprehensive evaluation and comparison of the reconstruction performance using MEM and MNE

To perform a detailed evaluation of our proposed fNIRS reconstructions methods, we developed a fully controlled simulation environment, similar to the one proposed by our team to validate EEG/MEG source localization

802 methods (Chowdhury et al., 2013, 2016; Hedrich et al., 2017). The fNIRS
 803 resting state data, acquired by the same montage (montage1) and under-
 804 went the same preprocessing as conducted for the real data, was added to
 805 the simulated true hemodynamic response for each channel. Indeed such en-
 806 vironment provided us access to a ground truth, which is not possible when
 807 considering real fNIRS data set. Previous studies validated tomography re-
 808 sults (Eggebrecht et al., 2014; Yamashita et al., 2016) by comparing with
 809 fMRI activation map which can indeed be considered as a ground truth, but
 810 only for well controlled and reliable paradigms. Since fMRI also measures
 811 a signal of hemodynamic origin, it is reasonable to check the concordance
 812 between fMRI results and DOT reconstructions. Therefore, as preliminary
 813 illustrations, we also compared our MEM and MNE results to fMRI Z-maps
 814 obtained during finger tapping tasks on 6 healthy participants, suggesting
 815 overall excellent performances of MEM when compared to MNE. Further
 816 quantitative comparison between fMRI and fNIRS 3D reconstruction, was
 817 out of the scope of this paper and will be considered in future studies.

818 4.7. Sampling size of fNIRS reconstructions

819 As opposed to several other fNIRS tomography studies that reconstruct
 820 fNIRS responses within a 3D volume space, here we proposed to use the
 821 mid-cortical surface as anatomical constraint to guide DOT reconstruction.
 822 However, the maximum spatial resolution of our surface based reconstruction
 823 was similar to the volume based one. Indeed, DOT reconstruction within a
 824 volume space usually down-sampled light sensitivity maps to either $2 \times 2 \times$
 825 2 mm^3 (Eggebrecht et al., 2014), $3 \times 3 \times 3 \text{ mm}^3$ (Eggebrecht et al., 2012)
 826 or $4 \times 4 \times 4 \text{ mm}^3$ (Yamashita et al., 2016) matrices, resulting in the down-

827 sampled voxel volume ranging from 8mm^3 to 64mm^3 . In our case, when
 828 projecting from volume space into cortical surface space, a unique set of
 829 voxels were assigned to each vertex along the cortical surface according to
 830 the Voronoi based projection method (Grova et al., 2006). Considering the
 831 mid-surface resolution (i.e., 25,000 vertices) used in this study, the average
 832 volume of a Voronoi cell was 25mm^3 , which falls in the above volume range.
 833 Therefore, we can conclude that both volume-based and surface-based fNIRS
 834 reconstructions as implemented here would result in similar sampling of the
 835 reconstruction space.

836 *4.8. fNIRS montage for 3D reconstructions*

837 In previous reported studies (Zeff et al., 2007; White and Culver, 2010;
 838 Zhan et al., 2012; Eggebrecht et al., 2012, 2014), a high density montage
 839 was considered which was proved to be able to provide high spatial resolu-
 840 tion and robustness to low SNR conditions (White and Culver, 2010). In
 841 the present study, we first considered a full double density montage, as pro-
 842 posed in (Kawaguchi et al., 2007), to generate realistic simulations, and then
 843 analyzed finger tapping results on real data acquired on one subject. Dou-
 844 ble density montages have been involved in several inverse modelling studies
 845 such as (Kawaguchi et al., 2004; Sakakibara et al., 2016; Machado et al.,
 846 2018). We also illustrated, in 5 other subjects, MEM performance when
 847 considering real data set acquired by optimal montages, exhibiting a large
 848 amount of local spatial overlap between channels. In this case, probe design
 849 was optimized to maximize the sensitivity to the hand knob ROI, while also
 850 ensuring sufficient spatial overlap between sensors (e.g., at least 13 detectors
 851 had to construct channels with each of the three sources, and the channel

852 distance was ranging from $2cm$ to $4.5cm$, see Fig.8a). We have previously
 853 demonstrated in Machado et al., 2018 that even if high density montages are
 854 usually considered as a gold standard for DOT reconstruction, personalized
 855 optimal montages (Machado et al., 2014, 2018, 2021) have ability to allow
 856 accurate reconstructions along the cortical surface. Finally, evaluating the
 857 performance of MEM when considering high density fNIRS montage would
 858 be of great interest but was out of the scope of this present study.

859 *4.9. Availability of the proposed MEM framework*

860 Several software packages have been proposed to provide fNIRS recon-
 861 struction pipelines, as for instance NeuroDOT (Eggebrecht et al., 2014, 2019),
 862 AtlasViewer(Aasted et al., 2015) and fNIRS-SPM(Ye et al., 2009). To en-
 863 sure an easy access of our MEM methodology to the fNIRS community, we
 864 developed and released a fNIRS processing toolbox - NIRSTORM (<https://github.com/Nirstorm>), as a plugin of Brainstorm software (Tadel et al.,
 865 2011), which is a renown software package dedicated for EEG/MEG analysis
 866 and source imaging. Our package NIRSTORM offers standard preprocessing,
 867 analysis and visualization as well as more advanced features such as person-
 868 alized optimal montage design, access to forward model estimation using
 870 MCXlab(Fang and Boas, 2009; Yu et al., 2018) and the MNE and MEM
 871 implementations considered in this study.

872 *4.10. Limitations and Perspectives*

873 Previously, Tremblay et al., 2018 had comprehensively compared a variety
 874 of fNIRS reconstruction methods using large number of realistic simulations.
 875 Since introducing MEM was our main goal of this study, we did not consider

such wide range of methodological comparisons. We decided to carefully compare MEM with MNE since MNE remains the main method considered for DOT, and is available in several software packages. As suggested in Tremblay et al., 2018, DOT reconstruction methods based on Tikhonov regularization, such as least square regularization in MNE, usually allow great sensitivity, but performed poorly in term of spatial extent - largely overestimating the size of the underlying generator. On the other hand, L1-based regularization (Süzen et al., 2010; Okawa et al., 2011; Kavuri et al., 2012; Prakash et al., 2014) could achieve more focal solutions with high specificity but much lower sensitivity. As demonstrated in our results, the proposed MEM framework allows reaching good sensitivity and accurate reconstruction of the spatial extent of the underlying generator. Bayesian model averaging (BMA) originally proposed for EEG source imaging by Trujillo-Barreto et al., 2004, also allows accurate DOT reconstructions with less false positives when compared to MNE. Similarly, we carefully compared MEM to Bayesian multiple priors approaches in Chowdhury et al., 2013 in the context of MEG source imaging. Comparing MEM with more advanced DOT reconstruction methods, including also the one proposed by Yamashita et al., 2016, would be of great interest but was out of the scope of this study.

Overall the main advantage of the MEM framework is its flexibility. Since the core structure of the MEM framework is to provide a unique reconstruction map by maximizing the entropy relative to a reference source distribution, one could implement its own reference for specific usage. For instance, as considered in the present study, the reference distribution considered the depth weighting MNE solution and spatial smoothing to inform our prior

901 model for MEM. Note that in this study we applied MEM independently for
 902 the two wavelengths and then calculated HbO/HbR concentration changes
 903 after reconstruction, whereas one could directly solve HbO/HbR concentra-
 904 tion changes along with reconstructions. Such procedure has been suggested
 905 by Li et al., 2004, by incorporating signals from the two wavelength within the
 906 same DOT reconstruction model. In the future, the MEM framework would
 907 allow to easily implement such a fusion model, as suggested by Chowdhury
 908 et al., 2015 in the context of MEG/EEG fusion algorithms. We have shown
 909 that MEM-based EEG/MEG fusion allows higher reliability in the source
 910 imaging results (Chowdhury et al., 2018), we will consider such an approach
 911 to estimate directly HbO/HbR fluctuations from the two wavelengths signals.

912 Finally, considering the main contribution of this study was to intro-
 913 duce the MEM framework for 3D fNIRS reconstruction, we decided to first
 914 carefully evaluate the performance of MEM, using well controlled realistic
 915 simulations. We also included few real data set reconstructions to illustrate
 916 the performance of the MEM reconstruction, whereas quantitative evaluation
 917 of MEM reconstructions on larger database will be considered in our future
 918 investigations.

919 5. Conclusion

920 In this study, we introduced a new fNIRS reconstruction method enti-
 921 tled Maximum Entropy on the Mean (MEM). We first implemented depth
 922 weighting into MEM framework and improved its temporal accuracy. To
 923 carefully validate the method, we applied a large number ($n = 4000$) of real-
 924 istic simulations with various spatial extents and depths. We also evaluated

the robustness of the method when dealing with low SNR signals. The comparison of the proposed method with the widely used depth weighted MNE was performed by applying four different quantification validation metrics. We found that MEM framework provided accurate and robust reconstruction results, relatively stable for a large range of spatial extents, depths and SNRs of the underlying generator. Moreover, we implemented the proposed method into a new fNIRS processing plugin - NIRSTORM in Brainstorm software to provide the access of the method to users for applications, validations and comparisons.

Acknowledgments

This work was supported by the Natural Sciences and Engineering Research Council of Canada Discovery Grant Program (CG and JML) and an operating grant from the Canadian Institutes for Health Research (CIHR MOP 133619 (CG)). fNIRS equipment was acquired using grants from NSERC Research Tools and Instrumentation Program and the Canadian Foundation for Innovation (CG). ZC is funded by the Fonds de recherche du Quebec Sante (FRQS) Doctoral Training Scholarship and the PERFORM Graduate Scholarship in Preventive Health Research. GP is funded by Strauss Canada Foundation.

Data availability

The original raw data supporting the findings of this study are available upon reasonable request to the corresponding authors.

947 **Conflict of interest**

948 The authors declare no potential conflict of interest.

Appendix A. Supplementary material

Supplementary material associated with this article can be found at the end of this manuscript.

References

- Aasted C.M., Yücel M.A., Cooper R.J., Dubb J., Tsuzuki D., Becerra L., Petkov M.P., Borsook D., Dan I., Boas D.A., 2015. Anatomical guidance for functional near-infrared spectroscopy: AtlasViewer tutorial. *Neurophotonics* 2, 020801.
- Amblard C., Lapalme E., Lina J.M., 2004. Biomagnetic Source Detection by Maximum Entropy and Graphical Models. *IEEE Transactions on Biomedical Engineering* 51, 427–442.
- Arridge S.R., 1999. Optical tomography in medical imaging. *Inverse Problems* 15, R41.
- Arridge S.R., 2011. Methods in diffuse optical imaging. *Philosophical Transactions of the Royal Society A: Mathematical, Physical and Engineering Sciences* 369, 4558–4576.
- Aydin Ü., Pellegrino G., Ali O.B.K.b., Abdallah C., Abdallah C., Dubeau F., Lina J.M., Lina J.M., Kobayashi E., Grova C., Grova C., Grova C., Grova C., 2020. Magnetoencephalography resting state connectivity patterns as

- indicatives of surgical outcome in epilepsy patients. *Journal of Neural Engineering* 17, 35007.
- Boas D.A., Brooks D.H., Miller E.L., Dimarzio C.A., Kilmer M., Gaudette R.J., Zhang Q., 2001. Imaging the body with diffuse optical tomography. *IEEE Signal Processing Magazine* 18, 57–75.
- Boas D.A., Chen K., Grebert D., Franceschini M.A., 2004a. Improving the diffuse optical imaging spatial resolution of the cerebral hemodynamic response to brain activation in humans. *Optics Letters* 29, 1506.
- Boas D.A., Culver J.P., Stott J.J., Dunn A.K., 2002. Three dimensional Monte Carlo code for photon migration through complex heterogeneous media including the adult human head. *Optics Express* 10, 159.
- Boas D.A., Dale A.M., 2005. Simulation study of magnetic resonance imaging-guided cortically constrained diffuse optical tomography of human brain function. *Applied Optics* 44, 1957.
- Boas D.A., Dale A.M., Franceschini M.A., 2004b. Diffuse optical imaging of brain activation: Approaches to optimizing image sensitivity, resolution, and accuracy, in: *NeuroImage*, Academic Press. pp. S275–S288.
- Cao N., Nehorai A., Jacobs M., 2007. Image reconstruction for diffuse optical tomography using sparsity regularization and expectation-maximization algorithm. *Opt Express* 15, 13695–13708.
- Chen W.L., Wagner J., Heugel N., Sugar J., Lee Y.W., Conant L., Malloy M., Heffernan J., Quirk B., Zinos A., Beardsley S.A., Prost R., Whelan H.T.,

2020. Functional Near-Infrared Spectroscopy and Its Clinical Application in the Field of Neuroscience: Advances and Future Directions. *Frontiers in Neuroscience* 14, 1–15.
- Chowdhury R.A., Lina J.M., Kobayashi E., Grova C., 2013. MEG Source Localization of Spatially Extended Generators of Epileptic Activity: Comparing Entropic and Hierarchical Bayesian Approaches. *PLoS ONE* 8.
- Chowdhury R.A., Merlet I., Birot G., Kobayashi E., Nica A., Biraben A., Wendling F., Lina J.M., Albera L., Grova C., 2016. Complex patterns of spatially extended generators of epileptic activity: Comparison of source localization methods cMEM and 4-ExSo-MUSIC on high resolution EEG and MEG data. *NeuroImage* 143, 175–195.
- Chowdhury R.A., Pellegrino G., Aydin ., Lina J.M., Dubeau F., Kobayashi E., Grova C., 2018. Reproducibility of EEG-MEG fusion source analysis of interictal spikes: Relevance in presurgical evaluation of epilepsy. *Human Brain Mapping* 39, 880–901.
- Chowdhury R.A., Zerouali Y., Hedrich T., Heers M., Kobayashi E., Lina J.M., Grova C., 2015. MEG/EEG Information Fusion and Electromagnetic Source Imaging: From Theory to Clinical Application in Epilepsy. *Brain Topography* 28, 785–812.
- Culver J.P., Durduran T., Furuya D., Cheung C., Greenberg J.H., Yodh A.G., 2003. Diffuse optical tomography of cerebral blood flow, oxygenation, and metabolism in rat during focal ischemia. *Journal of Cerebral Blood Flow and Metabolism* 23, 911–924.

- Dale A.M., Sereno M.I., 1993. Improved Localizadon of Cortical Activity by Combining EEG and MEG with MRI Cortical Surface Reconstruction: A Linear Approach. *Journal of Cognitive Neuroscience* 5, 162–176.
- Dehghani H., White B.R., Zeff B.W., Tizzard A., Culver J.P., 2009. Depth sensitivity and image reconstruction analysis of dense imaging arrays for mapping brain function with diffuse optical tomography. *Applied Optics* 48, D137.
- Delpy D.T., Cope M., Van Der Zee P., Arridge S., Wray S., Wyatt J., 1988. Estimation of optical pathlength through tissue from direct time of flight measurement. *Physics in Medicine and Biology* 33, 1433–1442.
- Eggebrecht A.T., Ferradal S.L., Robichaux-Viehoever A., Hassanpour M.S., Dehghani H., Snyder A.Z., Hershey T., Culver J.P., 2014. Mapping distributed brain function and networks with diffuse optical tomography. *Nature Photonics* 8, 448–454.
- Eggebrecht A.T., Muccigrosso D., Culver J.P., 2019. NeuroDOT: an extensible Matlab toolbox for streamlined optical brain mapping (Conference Presentation), in: Fantini S., Taroni P., Tromberg B.J., Sevick-Muraca E.M. (Eds.), *Optical Tomography and Spectroscopy of Tissue XIII*, SPIE. p. 51.
- Eggebrecht A.T., White B.R., Ferradal S.L., Chen C., Zhan Y., Snyder A.Z., Dehghani H., Culver J.P., 2012. A quantitative spatial comparison of high-density diffuse optical tomography and fMRI cortical mapping. *NeuroImage* 61, 1120–1128.

- von Ellenrieder N., Pellegrino G., Hedrich T., Gotman J., Lina J.M., Grova C., Kobayashi E., 2016. Detection and Magnetic Source Imaging of Fast Oscillations (40160Hz) Recorded with Magnetoencephalography in Focal Epilepsy Patients. *Brain Topography* 29, 218–231.
- Fang Q., 2010. Mesh-based Monte Carlo method using fast ray-tracing in Plücker coordinates. *Biomedical Optics Express* 1, 165.
- Fang Q., Boas D.A., 2009. Monte Carlo Simulation of Photon Migration in 3D Turbid Media Accelerated by Graphics Processing Units. *Optics Express* 17, 20178.
- Fischl B., Salat D.H., Busa E., Albert M., Dieterich M., Haselgrove C., Van Der Kouwe A., Killiany R., Kennedy D., Klaveness S., Montillo A., Makris N., Rosen B., Dale A.M., 2002. Whole brain segmentation: Automated labeling of neuroanatomical structures in the human brain. *Neuron* 33, 341–355.
- Friston K., Harrison L., Daunizeau J., Kiebel S., Phillips C., Trujillo-Barreto N., Henson R., Flandin G., Mattout J., 2008. Multiple sparse priors for the M/EEG inverse problem. *NeuroImage* 39, 1104–1120.
- Fuchs M., Wagner M., Köhler T., Wischmann H.A., 1999. Linear and non-linear current density reconstructions. *Journal of Clinical Neurophysiology* 16, 267–295.
- Grova C., Aiguabella M., Zelmann R., Lina J.M., Hall J.A., Kobayashi E., 2016. Intracranial EEG potentials estimated from MEG sources: A new

- approach to correlate MEG and iEEG data in epilepsy. *Human Brain Mapping* 37, 1661–1683.
- Grova C., Daunizeau J., Lina J.M., Bénar C.G., Benali H., Gotman J., 2006. Evaluation of EEG localization methods using realistic simulations of interictal spikes. *NeuroImage* 29, 734–753.
- Hämäläinen M.S., Ilmoniemi R.J., 1994. Interpreting magnetic fields of the brain: minimum norm estimates. *Medical & Biological Engineering & Computing* 32, 35–42.
- Hansen P.C., 2000. The L-Curve and its Use in the Numerical Treatment of Inverse Problems. in *Computational Inverse Problems in Electrocardiology*, ed. P. Johnston, *Advances in Computational Bioengineering* 4, 119–142.
- Hedrich T., Pellegrino G., Kobayashi E., Lina J.M., Grova C., 2017. Comparison of the spatial resolution of source imaging techniques in high-density EEG and MEG. *NeuroImage* 157, 531–544.
- Heers M., Chowdhury R.A., Hedrich T., Dubeau F., Hall J.A., Lina J.M., Grova C., Kobayashi E., 2016. Localization Accuracy of Distributed Inverse Solutions for Electric and Magnetic Source Imaging of Interictal Epileptic Discharges in Patients with Focal Epilepsy. *Brain Topography* 29, 162–181.
- Hughes I., Blasiole B., Huss D., Warchol M.E., Rath N.P., Hurle B., Ignatova E., David Dickman J., Thalmann R., Levenson R., Ornitz D.M., 2004. Otopetrin 1 is required for otolith formation in the zebrafish *Danio rerio*. *Developmental Biology* 276, 391–402.

- Huppert T., Barker J., Schmidt B., Walls S., Ghuman A., 2017. Comparison of group-level, source localized activity for simultaneous functional near-infrared spectroscopy-magnetoencephalography and simultaneous fNIRS-fMRI during parametric median nerve stimulation. *Neurophotonics* 4, 015001.
- Jenkinson M., Beckmann C.F., Behrens T.E.J., Woolrich M.W., Smith S.M., 2012. Review FSL. *NeuroImage* 62, 782–790.
- Joseph D.K., Huppert T.J., Franceschini M.A., Boas D.A., 2006. Diffuse optical tomography system to image brain activation with improved spatial resolution and validation with functional magnetic resonance imaging. *Applied Optics* 45, 8142.
- Kavuri V.C., Lin Z.J., Tian F., Liu H., 2012. Sparsity enhanced spatial resolution and depth localization in diffuse optical tomography. *Biomedical Optics Express* 3, 943.
- Kawaguchi H., Hayashi T., Kato T., Okada E., 2004. Theoretical evaluation of accuracy in position and size of brain activity obtained by near-infrared topography. *Physics in Medicine and Biology* 49, 2753–2765.
- Kawaguchi H., Koyama T., Okada E., 2007. Effect of probe arrangement on reproducibility of images by near-infrared topography evaluated by a virtual head phantom. *Applied Optics* 46, 1658.
- Lapalme E., Lina J.M., Mattout J., 2006. Data-driven parceling and entropic inference in MEG. *NeuroImage* 30, 160–171.

- Li A., Zhang Q., Culver J.P., Miller E.L., Boas D.A., 2004. Reconstructing chromosphere concentration images directly by continuous-wave diffuse optical tomography. *Optics Letters* 29, 256.
- Lin F.H., Witzel T., Ahlfors S.P., Stufflebeam S.M., Belliveau J.W., Hämäläinen M.S., 2006. Assessing and improving the spatial accuracy in MEG source localization by depth-weighted minimum-norm estimates. *NeuroImage* 31, 160–171.
- Machado A., Cai Z., Pellegrino G., Marcotte O., Vincent T., Lina J.M., Kobayashi E., Grova C., 2018. Optimal positioning of optodes on the scalp for personalized functional near-infrared spectroscopy investigations. *Journal of Neuroscience Methods* 309, 91–108.
- Machado A., Cai Z., Vincent T., Pellegrino G., Lina J.M., Kobayashi E., Grova C., 2021. Deconvolution of hemodynamic responses along the cortical surface using personalized functional near infrared spectroscopy. *Scientific Reports* , 1–19.
- Machado A., Marcotte O., Lina J.M., Kobayashi E., Grova C., 2014. Optimal optode montage on electroencephalography/functional near-infrared spectroscopy caps dedicated to study epileptic discharges. *Journal of Biomedical Optics* 19, 026010.
- Mattout J., Pélérini-Issac M., Garnero L., Benali H., 2005. Multivariate source prelocalization (MSP): Use of functionally informed basis functions for better conditioning the MEG inverse problem. *NeuroImage* 26, 356–373.

- Novi S.L., Forero E.J., Rubianes Silva J.A.I., de Souza N.G.S., Martins G.G., Quiroga A., Wu S.T., Mesquita R.C., 2020. Integration of Spatial Information Increases Reproducibility in Functional Near-Infrared Spectroscopy. *Frontiers in Neuroscience* 14, 1–12.
- Okawa S., Hoshi Y., Yamada Y., 2011. Improvement of image quality of time-domain diffuse optical tomography with lp sparsity regularization. *Biomedical Optics Express* 2, 3334.
- Pellegrino G., Hedrich T., Chowdhury R., Hall J.A., Lina J.M., Dubeau F., Kobayashi E., Grova C., 2016. Source localization of the seizure onset zone from ictal EEG/MEG data. *Human Brain Mapping* 37, 2528–2546.
- Pellegrino G., Hedrich T., PorrasBettancourt M., Lina J., Aydin ., Hall J., Grova C., Kobayashi E., 2020. Accuracy and spatial properties of distributed magnetic source imaging techniques in the investigation of focal epilepsy patients. *Human Brain Mapping* 41, 3019–3033.
- Piper S.K., Krueger A., Koch S.P., Mehnert J., Habermehl C., Steinbrink J., Obrig H., Schmitz C.H., 2014. A wearable multi-channel fNIRS system for brain imaging in freely moving subjects. *NeuroImage* 85, 64–71.
- Prakash J., Shaw C.B., Manjappa R., Kanhirodan R., Yalavarthy P.K., 2014. Sparse recovery methods hold promise for diffuse optical tomographic image reconstruction. *IEEE Journal on Selected Topics in Quantum Electronics* 20.
- Raffin E., Pellegrino G., Di Lazzaro V., Thielscher A., Siebner H.R., 2015. Bringing transcranial mapping into shape: Sulcus-aligned mapping cap-

- tures motor somatotopy in human primary motor hand area. *NeuroImage* 120, 164–175.
- Sakakibara Y., Kurihara K., Okada E., 2016. Evaluation of improvement of diffuse optical imaging of brain function by high-density probe arrangements and imaging algorithms. *Optical Review* 23, 346–353.
- Schmitz C.H., Klemer D.P., Hardin R., Katz M.S., Pei Y., Graber H.L., Levin M.B., Levina R.D., Franco N.A., Solomon W.B., Barbour R.L., 2005. Design and implementation of dynamic near-infrared optical tomographic imaging instrumentation for simultaneous dual-breast measurements. *Applied Optics* 44, 2140.
- Schneider P., Piper S., Schmitz C.H., Schreiter N.F., Volkwein N., Lüdemann L., Malzahn U., Poellinger A., 2011. Fast 3D near-infrared breast imaging using indocyanine green for detection and characterization of breast lesions. *RoFo Fortschritte auf dem Gebiet der Röntgenstrahlen und der Bildgebenden Verfahren* 183, 956–963.
- Scholkmann F., Kleiser S., Metz A.J., Zimmermann R., Mata Pavia J., Wolf U., Wolf M., 2014. A review on continuous wave functional near-infrared spectroscopy and imaging instrumentation and methodology.
- Shimokawa T., Kosaka T., Yamashita O., Hiroe N., Amita T., Inoue Y., Sato M.a., 2012. Hierarchical Bayesian estimation improves depth accuracy and spatial resolution of diffuse optical tomography. *Optics Express* 20, 20427.
- Shimokawa T., Kosaka T., Yamashita O., Hiroe N., Amita T., Inoue Y., Sato

- M.a., 2013. Extended hierarchical Bayesian diffuse optical tomography for removing scalp artifact. *Biomedical Optics Express* 4, 2411.
- van der Sluis A., 1969. Condition numbers and equilibration of matrices. *Numerische Mathematik* 14, 14–23.
- Smith S.M., Jenkinson M., Woolrich M.W., Beckmann C.F., Behrens T.E., Johansen-Berg H., Bannister P.R., De Luca M., Drobnjak I., Flitney D.E., Niazy R.K., Saunders J., Vickers J., Zhang Y., De Stefano N., Brady J.M., Matthews P.M., 2004. Advances in functional and structural MR image analysis and implementation as FSL, in: *NeuroImage*, Academic Press. pp. S208–S219.
- Sohrabpour A., He B., 2021. Exploring the Extent of Source Imaging: Recent Advances in Noninvasive Electromagnetic Brain Imaging. *Current Opinion in Biomedical Engineering* , 100277.
- Strangman G., Franceschini M.A., Boas D.A., 2003. Factors affecting the accuracy of near-infrared spectroscopy concentration calculations for focal changes in oxygenation parameters. *NeuroImage* 18, 865–879.
- Strangman G.E., Li Z., Zhang Q., 2013. Depth Sensitivity and Source-Detector Separations for Near Infrared Spectroscopy Based on the Colin27 Brain Template. *PLoS ONE* 8.
- Süzen M., Giannoula A., Durduran T., 2010. Compressed sensing in diffuse optical tomography. *Optics Express* 18, 23676.
- Tadel F., Baillet S., Mosher J.C., Pantazis D., Leahy R.M., 2011. Brainstorm:

A user-friendly application for MEG/EEG analysis. Computational Intelligence and Neuroscience 2011, 879716.

Tremblay J., Martínez-Montes E., Vannasing P., Nguyen D.K., Sawan M., Lepore F., Gallagher A., 2018. Comparison of source localization techniques in diffuse optical tomography for fNIRS application using a realistic head model. Biomedical Optics Express 9, 2994.

Trujillo-Barreto N.J., Aubert-Vázquez E., Valdés-Sosa P.A., 2004. Bayesian model averaging in EEG/MEG imaging. NeuroImage 21, 1300–1319.

White B.R., Culver J.P., 2010. Quantitative evaluation of high-density diffuse optical tomography: *in vivo* resolution and mapping performance. Journal of Biomedical Optics 15, 026006.

White B.R., Snyder A.Z., Cohen A.L., Petersen S.E., Raichle M.E., Schlaggar B.L., Culver J.P., 2009. Resting-state functional connectivity in the human brain revealed with diffuse optical tomography. NeuroImage 47, 148–156.

Yamashita O., Shimokawa T., Aisu R., Amita T., Inoue Y., Sato M.a., 2016. Multi-subject and multi-task experimental validation of the hierarchical Bayesian diffuse optical tomography algorithm. NeuroImage 135, 287–299.

Ye J.C., Tak S., Jang K.E., Jung J., Jang J., 2009. NIRS-SPM: Statistical parametric mapping for near-infrared spectroscopy. NeuroImage 44, 428–447.

Yu L., Nina-Paravecino F., Kaeli D., Fang Q., 2018. Scalable and massively parallel Monte Carlo photon transport simulations for heterogeneous computing platforms. Journal of Biomedical Optics 23, 1.

- Yücel M.A., Lühmann A.v., Scholkmann F., Gervain J., Dan I., Ayaz H., Boas D., Cooper R.J., Culver J., Elwell C.E., Eggebrecht A., Franceschini M.A., Grova C., Homae F., Lesage F., Obrig H., Tachtsidis I., Tak S., Tong Y., Torricelli A., Wabnitz H., Wolf M., 2021. Best practices for fNIRS publications. *Neurophotonics* 8, 1–34.
- Yücel M.A., Selb J., Boas D.A., Cash S.S., Cooper R.J., 2014. Reducing motion artifacts for long-term clinical NIRS monitoring using collodion-fixed prism-based optical fibers. *NeuroImage* 85, 192–201.
- Zeff B.W., White B.R., Dehghani H., Schlaggar B.L., Culver J.P., 2007. Retinotopic mapping of adult human visual cortex with high-density diffuse optical tomography. *Proceedings of the National Academy of Sciences* 104, 12169–12174.
- Zhan Y., Eggebrecht A.T., Culver J.P., Dehghani H., 2012. Image Quality Analysis of High-Density Diffuse Optical Tomography Incorporating a Subject-Specific Head Model. *Frontiers in Neuroenergetics* 4, 6.

Supplementary material

S1. Validation metrics

Here is a detailed description of the four validation metrics considered in our evaluation. Except the shape error (SE), other metrics were all calculated at the time instant τ when the simulated ΔOD time course reached its peak value (e.g. 12.2s after onset).

Area Under the Receiver Operating Characteristic (ROC) curve (AUC) was used to assess overall detection accuracy of the reconstruction methods. We used a specific version of AUC that has been proposed in (Grova et al., 2006) in order not to bias results towards false positives. In further details, ROC curves were generated by plotting the sensibility of the detection as a function of 1-specificity, while thresholding the normalized reconstruction map from 0 to 1 with a certain step value. In the context of source reconstruction, especially when the generator is focal, the region of true positive is usually much smaller than the region of true negative, whereas non-biased AUC evaluation would require to sample the same amount of active and inactive generators. To overcome this possible bias, we considered a ROC evaluation using the same number of active and inactive generators that were randomly sampled within two different regions: 1) AUC_{close} : inactive generators were sampled within the immediate spatial neighborhood of the ground truth; and 2) AUC_{far} : inactive generators were sampled within the local maxima of the reconstructed activity located far from the ground truth. The final AUC was then the average of AUC_{close} and AUC_{far} .

Minimum geodesic distance (Dmin) was represented by the geodesic distance, following the circumvolutions of the cortical surface, of the vertex

that exhibited maximum of reconstructed activity to the border of the ‘generator’. It should be 0 when the peak of the reconstruction map was located inside the simulated cortical region.

Spatial Dispersion (SD) assessed the spatial spread of the estimated ‘generator’ distribution and the localization error using Eq.S1. The ideal value (i.e. $SD = 0mm$), was achieved when no activation was reconstructed outside the theoretical ‘generator’. The larger the SD was, the more spatially spread were the reconstructed maps.

$$SD = \sqrt{\frac{\sum_{i=1}^K \left(\min_{j \in \Theta} (D^2(i, j)) \hat{X}^2(i, \tau) \right)}{\sum_{i=1}^K \left(\hat{X}^2(i, \tau) \right)}} \quad (S1)$$

where $\min_{j \in \Theta} (D^2(i, j))$ is the minimum Euclidean distance between the vertex i to the vertex j which is located inside the simulated ‘generator’ (Θ). $\hat{X}^2(i, \tau)$ is the power of the amplitude of reconstructed time course on vertex i at time τ . K is the total number of vertices within the reconstruction field of view.

Shape error(SE) evaluated the temporal accuracy of the reconstruction. Reconstructed time courses within the simulation ‘generator’ were averaged and normalized. The root mean square of the difference between this time course and the normalized theoretical time course was estimated and denoted as SE in Eq.S2 as introduced in (Chowdhury et al., 2013)

$$SE = \sqrt{\frac{1}{T} \sum_t \left(\frac{X_{th}(t)}{\max(|X_{th}(t)|)} - \frac{\text{mean}_{j \in \Theta}(\hat{X}(j, t))}{\max(|\text{mean}_{j \in \Theta}(\hat{X}(j, t))|)} \right)^2} \quad (S2)$$

where T is length of the time course. $X_{th}(t)$ is the theoretical time course of

the simulation. $mean_{j \in \Theta}(\hat{X}(j, t))$ is the averaged mean of the reconstructed time courses within the ‘generator’.

S2. Effects of depth weighting on MNE

We first investigated the effects of depth weighting factor ω_2 selection for depth weighted MNE. To do so, we evaluated spatial and temporal performances of DOT reconstruction. As presented in Fig.S1, we compared depth weighted MNE using depth weighting factors $\omega_2 = 0, 0.1, 0.3, 0.5, 0.7, 0.9$ in superficial seeds case. In general, $\omega_2 = 0.3$ and 0.5 provided overall the most accurate results (i.e. median $AUC > 0.8$ and $Dmin = 0mm$). For focal generators (i.e. $Se = 3, 5$), $\omega_2 = 0.3$ performed better than $\omega_2 = 0.5$ considering it was providing significantly lower SD. However, in extended generators (i.e. $Se = 7, 9$), reconstructions with $\omega_2 = 0.5$ were exhibiting more accurate results, consisting in significantly positive AUC difference (0.05 and 0.08, $p < 0.001$) and significantly positive SD difference (2.24 and 2.06, $p < 0.001$). $\omega_2 = 0$ and 0.1 only provided AUC higher than 0.8 in the case of $Se = 3$, whereas $\omega_2 = 0.7$ and 0.9 failed in all cases and even the median values of $Dmin$ were significantly larger (median values around 2-3 cm) than other cases. Based on these results, we decided to consider only the depth weighting values $\omega_2 = 0.3$ and 0.5 for depth weighting MNE in the comparisons with with MEM reconstructions.

S3. MEM v.s. MNE with realistic simulations involving middle and deep seeds

In Fig.S2 and Table.S1, we are presenting the comparison of MEM and MNE in middle seeds case. First of all, we found that more depth compen-

Fig.S1. Evaluation of the performances of depth weighted MNE for different depth weighting factors $\omega = 0, 0.1, 0.3, 0.5, 0.7, 0.9$. Distribution of validation metrics (AUC, Dmin, SD and SE) are displayed using boxplot representations, for simulations involving superficial seeds only and for spatial extents $Se = 3, 5, 7, 9$.

sation was required to provide good reconstructions in all scenarios. Thus, MEM(0.5, 0.5) was compared to the best of MNE - MNE(0.5). Non-significant AUC and Dmin differences were found between them. However, MEM(0.5, 0.5) provided significant lower SD than MNE(0.5), median value of difference of $SD = -5.33, -4.80, -5.00, -4.95, p < 0.001$ for $Se = 3, 5, 7, 9$ respectively. Fig.S3 and Table.S2 are presenting the comparison of MEM and MNE in the comparison of them in deep seeds case. Similarly, no significant AUC and Dmin differences were found. MEM(0.5, 0.5) provided significant lower SD than MNE(0.5), median value of difference of $SD = -6.39, -6.33, -6.97, -5.52, P < 0.001$ for $Se = 3, 5, 7, 9$ respectively. For temporal performance in these two cases, similar to Fig.3, MNE(0.5) gave significant lower SE (-0.01 or $-0.02, p < 0.001$) than MEM when $Se = 3, 5$ (small difference). No significant different SE was found in $Se = 7, 9$.

Fig.S2. Evaluation of the performances of MEM and MNE using realistic simulations involving middle seeds for different spatial extent ($Se = 3, 5, 7, 9$). Boxplot representation of the distribution of four validation metrics for three depth weighted strategies of MEM and two depth weighted strategies of MNE, namely: MEM(0.3, 0.3) in blue, MEM(0.3, 0.5) in green, MEM(0.5, 0.5) in red, MNE(0.3) in magenta and MNE(0.5) in black. Results were obtained after DOT reconstruction of 830nm ΔOD .

Middle Seeds		Se = 3		Se = 5		Se = 7		Se = 9	
		MNE (0.3)	MNE (0.5)	MNE (0.3)	MNE (0.5)	MNE (0.3)	MNE (0.5)	MNE (0.3)	MNE (0.5)
AUC	MEM (0.3, 0.3)	0.03**	-0.03	0.03**	0.00	0.02	-0.05**	-0.02	-0.10**
	MEM (0.3, 0.5)	0.03**	-0.03	0.05**	0.01	0.05**	-0.01	0.05**	-0.02**
	MEM (0.5, 0.5)	0.06**	0.02	0.07**	0.01	0.07**	-0.01	0.08**	0.00
Dmin	MEM (0.3, 0.3)	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00
	MEM (0.3, 0.5)	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00
	MEM (0.5, 0.5)	0.00	0.00	0.00	0.00	0.00	0.00	0.00	0.00
SD	MEM (0.3, 0.3)	-4.05**	-7.21**	-4.27**	-7.25**	-4.10**	-6.40**	-3.58**	-5.43**
	MEM (0.3, 0.5)	-4.00**	-7.06**	-4.09**	-6.90**	-3.96**	-6.40**	-3.65**	-5.45**
	MEM (0.5, 0.5)	-2.54**	-5.33**	-2.46**	-4.80**	-2.85**	-5.00**	-3.08**	-4.95**
SE	MEM (0.3, 0.3)	0.02**	0.02**	0.01**	0.01**	0.01**	0.01*	0.00	0.01
	MEM (0.3, 0.5)	0.01**	0.01**	0.01**	0.01**	0.00	0.00	0.00	0.00
	MEM (0.5, 0.5)	0.01**	0.01**	0.01**	0.01*	0.00	0.00	0.00	0.00

Table.S1. Wilcoxon signed rank test results of reconstruction performance comparison of MEM and MNE in middle seeds case. Median values of paired difference are presented in the table. p values were corrected for multiple comparisons using Bonferroni correction, * indicates $p < 0.01$ and ** represents $p < 0.001$. Median of the paired difference of each validation metrics is color coded as follows: green: MEM is significantly better than MNE, red: MNE is significantly better than MEM and gray: non-significance.

Fig.S3. Evaluation of the performances of MEM and MNE using realistic simulations involving deep seeds for different spatial extent ($Se = 3, 5, 7, 9$). Boxplot representation of the distribution of four validation metrics for three depth weighted strategies of MEM and two depth weighted strategies of MNE, namely: MEM(0.3, 0.3) in blue, MEM(0.3, 0.5) in green, MEM(0.5, 0.5) in red, MNE(0.3) in magenta and MNE(0.5) in black. Results were obtained after DOT reconstruction of 830nm ΔOD .

Deep Seeds		Se = 3		Se = 5		Se = 7		Se = 9	
		MNE (0.3)	MNE (0.5)	MNE (0.3)	MNE (0.5)	MNE (0.3)	MNE (0.5)	MNE (0.3)	MNE (0.5)
AUC	MEM (0.3, 0.3)	0.08**	-0.20**	0.06**	-0.13**	0.05**	-0.08**	0.00	-0.07**
	MEM (0.3, 0.5)	0.09**	-0.17**	0.08**	-0.08**	0.08**	-0.03*	0.05**	-0.02
	MEM (0.5, 0.5)	0.29**	0.03	0.18**	-0.01	0.13**	-0.01	0.06**	-0.01
Dmin	MEM (0.3, 0.3)	0.00	2.91	0.00	0.00	0.00	0.00	0.00	0.00
	MEM (0.3, 0.5)	0.00	2.16	0.00	0.00	0.00	0.00	0.00	0.00
	MEM (0.5, 0.5)	-3.53	0.00	0.00	0.00	0.00	0.00	0.00	0.00
SD	MEM (0.3, 0.3)	-3.73**	-6.25**	-3.65**	-7.37**	-3.51**	-7.39**	-3.46**	-6.21**
	MEM (0.3, 0.5)	-4.00**	-6.61**	-3.83**	-7.54**	-3.95**	-7.63**	-3.82**	-6.50**
	MEM (0.5, 0.5)	-4.56**	-6.39**	-3.73**	-6.33**	-3.10**	-6.97**	-3.33**	-5.52**
SE	MEM (0.3, 0.3)	0.02**	0.02**	0.02**	0.02**	0.01*	0.01	0.01	0.00
	MEM (0.3, 0.5)	0.01**	0.01**	0.01*	0.01**	0.01	0.01	0.00	0.00
	MEM (0.5, 0.5)	0.01*	0.01**	0.01*	0.01*	0.00	0.01	0.00	0.00

Table.S2. Wilcoxon signed rank test results of reconstruction performance comparison of MEM and MNE in deep seeds case. Median values of paired difference are presented in the table. p values were corrected for multiple comparisons using Bonferroni correction, * indicates $p < 0.01$ and ** represents $p < 0.001$. Median of the paired difference of each validation metrics is color coded as follows: green: MEM is significantly better than MNE, red: MNE is significantly better than MEM and gray: non-significance.