Colocalized, Bidirectional Optogenetic Modulations in Freely Behaving Animals with a Wireless Dual-Color Optoelectronic Probe

Lizhu Li,^{1, #} Lihui Lu,^{2, 3, #} Yuqi Ren,^{2, 3} Guo Tang,¹ Yu Zhao,¹ Xue Cai,¹ Zhao Shi,¹ He
Ding,⁴ Changbo Liu,⁵ Dali Cheng,¹ Yang Xie,¹ Huachun Wang,¹ Xin Fu,⁶ Lan Yin,⁶ Minmin
Luo,^{2, 3, 7, *} Xing Sheng^{1, 7, *}

Affiliations

¹Department of Electronic Engineering, Beijing National Research Center for Information
Science and Technology, Center for Flexible Electronics Technology, Tsinghua University,
Beijing, China

²Chinese Institute for Brain Research, Beijing, China

³National Institute of Biological Sciences and School of Life Sciences, Tsinghua University,
Beijing, China

⁴Beijing Engineering Research Center of Mixed Reality and Advanced Display, School of Op
tics and Photonics, Beijing Institute of Technology, Beijing, China

⁵School of Materials Science and Engineering and Hangzhou Innovation Institute, Beihang
University, Beijing, China

⁶School of Materials Science and Engineering, Tsinghua University, Beijing, China

⁷IDG/McGovern Institute for Brain Research, Tsinghua University, Beijing, China

L. Li and L. Lu contributed equally to this work.

*Correspondence to: luominmin@cibr.ac.cn; xingsheng@tsinghua.edu.cn

ABSTRACT

The precise control of neural activities at both cellular and circuit levels reveals significant impacts on the fundamental neuroscience explorations and medical applications. Optogenetic methods provide efficient cell-specific modulations, but state-of-the-art technologies lack the ability of simultaneous neural activation and inhibition in a same brain region of freely moving animals. Here we report bidirectional neuronal activity manipulation accomplished by a wireless, dual-color optogenetic probe in synergy with the co-expression of two spectrally distinct opsins (ChrimsonR and stGtACR2) in a rodent model. Based on vertically assembled, thin-film microscale light-emitting diodes (micro-LEDs) on flexible substrates, the dual-color probe shows colocalized red and blue emissions and allows chronic *in vivo* operations with desirable biocompatibilities. In addition, we discover that neurons co-expressing the two opsins can be deterministically evoked or silenced under red or blue irradiations. Implanted in behaving mice, the wirelessly controlled dual-color probe interferes with dopaminergic neurons in the ventral tegmental area (VTA), increasing or decreasing dopamine levels with colocalized red and blue stimulations. Such bidirectional regulations further generate rewarding and aversive behaviors of freely moving animals in a place preference test. The technologies established here will create numerous opportunities and profound implications for the brain research.

INTRODUCTION

Understanding the brain functions and treating neurological disorders rely on the continuous development of advanced technologies to interrogate with the complex nervous systems^{1, 2}. Over the past decades, optogenetic methods have been emerging as a powerful toolset for effective and precise neural manipulation, owing to their capability of specific cell targeting^{3, 4}. Until now, a range of genetically encoded, light-sensitive proteins (opsins) with distinct spectral responses have been engineered to regulate flows of different cations and anions⁵⁻¹⁶, triggering various biological responses, from cells and *C. elegans* to rodents and non-human primates¹⁷⁻²⁰. Accompanying the development of opsins, advanced technologies for light delivery into the brain have also been progressing rapidly. Apart from various fiber-based emitters interfacing with external light sources²¹⁻²³, recently developed implantable probes based on thin-film, microscale optoelectronic devices have offered a viable solution to versatile neural modulations in untethered animals, when incorporated with wirelessly operated systems²⁴⁻²⁶. In addition to these achievements, bidirectional neural modulations, for example, by activating or suppressing the same or different populations of cells in behaving animals with light at different wavelengths, are still highly demanded. Such capabilities would enable more precise control of neural activities to advance both neuroscience explorations and disease medications²⁷. Certain efforts have been attempted, including explorations based on experiments by expressing excitatory and inhibitory opsins in separate animals^{28, 29} or distinct cells in different neural regions^{17, 30}. There are few reports on dual-color optogenetic activation and inhibition by co-expressing different opsins in the same cells³¹⁻³³, but the efficacy and utility of such bidirectional modulations could be limited by the potential spectral overlay of the two opsins and the lack of miniaturized light delivery systems. Therefore, we envision that wirelessly operated light sources with polychromatic

emissions, which collaborate with co-expressed opsins, can be implemented for untethered use in freely moving animals and enable previously inaccessible applications.

This study presents synergic optoelectronic and biological strategies to overcome challenges of existing techniques, and realize colocalized, bidirectional neural manipulations in untethered behaving animals. With heterogeneously integrated thin-film, microscale light-emitting diodes (micro-LEDs), a wirelessly operated, implantable dual-color probe allows independently controlled red and blue emissions in the same region. We also discover that the co-expression of two channelrhodopsins (ChrimsonR and stGtACR2) in the same group of neurons enables efficient light-evoked cell depolarization and hyperpolarization with controlled cation and anion flows under red and blue illuminations, respectively. With the wireless, dual-color micro-LED probes implanted into the ventral tegmental area (VTA) of freely behaving mice, preference and aversion tests clearly demonstrate the bidirectional optogenetic excitation and inhibition capability. Furthermore, *in vivo* behavioral results associated with the red and blue stimulations in the VTA are validated by increased or decreased dopamine levels detected in the nucleus accumbens (NAc) in real time.

RESULTS

Wirelessly operated, dual-color micro-LED probes and circuit systems

We design and implement a wireless, dual-color micro-LED probe that generates red and blue emissions in the same brain region for bidirectional optogenetic modulations *in vivo*, as schematically illustrated in Figure 1a. Specifically, with the co-expression of spectrally distinct ion channels, ChrimsonR⁷ and stGtACR2⁹, in the same population of cells, red and blue light regulate the inflow of cations and anions, causing cell depolarization and hyperpolarization, respectively (Figure 1b). Figure 1c displays an exploded view of the dual-color probe, with design and fabrication details provided in the supplemental information and

Figure S1. From bottom to top, the probe structure comprises a copper (Cu) coated polyimide (PI) thin-film substrate, an indium gallium phosphide (InGaP) red LED, a silicon oxide (SiO₂) / titanium oxide (TiO₂) based dielectric filter, an indium gallium nitride (InGaN) blue LED. All the devices (red LEDs, blue LEDs and filters) are formed in thin-film, microscale formats (lateral dimension $\sim 125 \mu\text{m} \times 180 \mu\text{m}$, thickness $\sim 5\text{--}7 \mu\text{m}$) through epitaxial liftoff, and vertically assembled via transfer printing^{34–36} (Figure 1d). The intrinsic optical transparency of active layers in the InGaN blue micro-LED allows emissions from the red micro-LED passing through with minimal losses. Additionally, the designed filter selectively reflects blue light and transmits red light, enhancing the emissive efficiencies of both micro-LEDs.

Flexible PI substrates with vertically stacked micro-LEDs are laser milled and encapsulated with a waterproof coating, forming a needle shaped probe with a dimension (thickness $\sim 120 \mu\text{m}$ and width $\sim 320 \mu\text{m}$) similar to a conventional silica fiber (diameter $\sim 220 \mu\text{m}$) used in optogenetic stimulations. Separate metallization and electrical insulation ensure that red and blue micro-LEDs can be independently lighted up and display spectrally varied illuminations (red, blue or combined) in the same location (Figure 1e and Movie S1). Figure 1f plots measured electroluminescence (EL) spectra. The red and blue micro-LEDs exhibit emission peaks at 630 nm and 480 nm, which are in good accordance with the excitation spectra of ChrimsonR and stGtACR2 (data extracted from references^{7, 8}), respectively. Moreover, the spectral separation between the two opsins, as well as their different operational modes (pulsed red light activation for ChrimsonR, and continuous blue light inhibition for stGtACR2), guarantee precise optogenetic activation and inhibition with minimal optical interferences. More optoelectronic properties of the dual-color micro-LED probe, including current-voltage characteristics, quantum efficiencies, power irradiance and angular-dependent emissive profiles are presented in Figure S2. Under injection currents of

10 mA, the red and the blue micro-LEDs can reach irradiances up to 50 mW/mm² and 200 mW/mm² on the device surface, which are sufficient for optogenetic modulations.

To illustrate the optical performance of the devices, we insert a dual-color probe into a brain phantom (Figure 1g, top). For comparison, established models calculate optical profiles based on the optical properties (absorption and scattering) of brain tissues (Figure 1g, bottom). Both experimental and simulation results reveal that vertically stacked dual-color micro-LEDs show colocalized red and blue emission profiles, ensuring effective activation and inhibition in the same brain region. Power irradiances for blue and red emissions decrease to 10% of the original power at propagating distances of ~400 μm and ~800 μm, respectively. Such a unique design of vertical assembly holds prominent advantages compared to a probe with laterally arranged red and blue micro-LEDs displaying highly misaligned emissions, which could severely limit its capability for bidirectional modulations.

Thermal behaviors of the micro-LED probe are also systematically analyzed, with detailed results presented in the supplemental material and Figure S3. Probes are inserted into the brain phantom and the living mouse brain, with temperature rises measured by an infrared camera and a thermocouple, respectively. Both experiments obtain similar results and are in agreement with the thermal models established based on finite-element analysis. Under a pulsed operation (20-Hz frequency, 10-ms pulse duration, current 0–10 mA), the tissue temperature rises caused by the red micro-LED can be restricted within 1 °C. The operation of the blue micro-LED induces a higher thermal effect, because practical neural inhibition for stGtACR2 requires continuous current injection; nonetheless, the tissue heating can be controlled within 2 °C by limiting the current within 5 mA. Such operation modes for red and blue micro-LEDs will be employed throughout *in vivo* experiments, to mitigate undesired neural responses and possible tissue damage due to overheating.

We also design a customized, miniaturized flexible circuit module wirelessly operate the dual-color micro-LED probe. Controlled by radio frequency (RF) communication at 2.4 GHz, the battery-powered circuit (weight ~1.9 g) can independently address the red and the blue micro-LEDs, by adjusting their pulse frequencies, durations and injection currents for versatile neural modulations *in vivo* (Figure 1h, and details in the supplemental material and Figures S4–S7). Designed programming commands also allow the independent operation of multiple micro-LED probes in synchronized or non-synchronized modes (Movie S2). After subcranially implanting the micro-LED probe into a behaving mouse, the light-weighted circuit mounted onto the mouse head provides the capability for optogenetic control in an untethered manner (Figure 1i).

Bidirectional optogenetic manipulations of neurons *in vitro*, with dual-color stimulations

We first evaluate the feasibility of dual-color, bidirectional optogenetic modulations for neurons co-expressing ChrimsonR and stGtACR2 in brain slices with electrophysiological tests. Mixed vectors of AAV-CAG-DIO-ChrimsonR-mCherry and AAV-EF1a-DIO-stGtACR2-EGFP are injected into the ventral tegmental area (VTA) of *DAT-Cre* mice (Figure 2a). The deployment of the soma-targeted variant stGtACR2 exhibits performance superior to previously reported anion channel GtACR2, by enhancing the membrane trafficking and reducing the axonal excitation^{8,9}. Figure 2b shows the fluorescence images of stained cells 3 weeks after viral injection in the VTA, including 4',6-diamidino-2-phenylindole (DAPI, blue), stGtACR2 (green) and ChrimsonR (red). The merged graph and zoomed in views (Figure 2c) clearly reveal the desirable co-expression of the two opsins in target cells. Statistical results show that ~30% of cells are labeled with stGtACR2 or ChrimsonR, and > 83% of cells co-express both opsins (Figure 2d).

Whole-cell patch recordings capture intracellular signals of these cells under optical stimulations (Figure 2e). Other than using our designed micro-LED probe, here we employ a collimated light spot with coupled red (633 nm) and blue (480 nm) laser sources for *in vitro* experiments, to achieve uniform illuminations for better power calibrations. Figures 2f and 2g validate that pulsed red stimulations (20-Hz frequency, 10-ms pulse duration) induce action potentials (APs or spikes) by depolarizing these cells, and spike activation probability increases at elevated red irradiances. In these same cells, continuous blue stimulations suppress activities by hyperpolarization (Figures 2h–2j). Voltage clamp results illustrate that blue light evokes photocurrent reversion when the membrane potential becomes more positive (Figure 2h), and probability for spike inhibition also increases at elevated blue irradiances (Figures 2i and 2j). Threshold irradiances for complete spike activation and inhibition are around 10 mW/mm² and 1 mW/mm² for red and blue illuminations, respectively. The results are consistent with previous reports on cells expressing only ChrimsonR or stGtACR2^{7,9}.

Notably, we do not only demonstrate that the co-expression of both opsins does not affect the efficacy of cell activity excitation or suppression, but also accomplish bidirectional activation and inhibition in the same cell by instantaneously altered red and blue irradiations (Figure 2k). In accordance with activation spectra of two opsins (Figure 1f), red light does not interact with stGtACR2, which exhibits nearly zero responses at wavelengths longer than 600 nm. On the other hand, ChrimsonR should have responded to blue light, which could raise concern about possible optical crosstalk. Nevertheless, spike activations caused by blue light do not occur in cells co-expressing ChrimsonR and stGtACR2, since cell suppressions by turning on stGtACR2 channels arise and saturate at a much lower blue irradiance. This fact can be confirmed by patching the cells under combined red and blue illuminations (Figures 2l and 2m). Example traces in Figure 2l show that red-light-evoked spikes are

suppressed under simultaneous blue illuminations even with very low irradiances. More traces are reported in Figure S8, and Figure 2m summarizes the spike probability (measured and averaged among $n = 5$ cells) with combined red and blue illuminations. Introducing blue irradiances orders of magnitude lower than red irradiances can effectively suppress cell spikes and cause activity inhibition.

Bidirectional optogenetic modulation of preference and aversion behaviors *in vivo*, with wireless dual-color micro-LED probes

Activities of the dopaminergic neurons in the VTA we manage to manipulate in Figure 2 significantly correlate with the reward and aversion based behaviors^{37, 38, 39}. In Figure 3, we further exploit the capability of bidirectional optogenetic manipulations by implanting dual-color micro-LED probes into behaving animals. After the injection of AAV-CAG-DIO-ChrimsonR-mCherry and AAV-EF1a-DIO-stGtACR2-EGFP vectors, the dual-color micro-LED probe is implanted into the VTA of *DAT-cre* mice (Figure 3a). Figure 3b illustrates the locations of the micro-LEDs in the VTA, with a slice showing the co-expression of ChrimsonR and stGtACR2, combined with the luminescence of the red and the blue micro-LEDs under corresponding optical excitations. To access the devices' biocompatibility *in vivo*, histological examinations of brain slices show that inflammatory reactions of the micro-LED probe are similar to a conventional silica fiber (diameter $\sim 220 \mu\text{m}$) 1 day or 3 weeks post implantation (Figure S9), owing to its thin-film geometry, mechanical flexibility and bio-friendly encapsulation. Furthermore, the parylene-based, waterproof encapsulation maintains the micro-LEDs' chronic and stable operation for more than 200 days after implantation (Figure S10).

In vivo experiments are performed within a real-time place preference / avoidance (RTPP) paradigm. After 14 days of virus and micro-LED probe injection, we place mice in a two-compartment apparatus and monitor their behaviors with a camera during optogenetic modulations (Figure 3c and Figure S11). Figure 3d provides the test protocol, including a baseline pretest (10 mins), red LED stimulation (pulsed wave, 30 mins), a rest phase (30 mins), and blue LED stimulation (continuous wave, 30 mins). Representative heat maps in Figure 3e show the change of real-time behaviors during optogenetic stimulations, for mice expressing different combinations of opsins. Summarized results in Figure 3f clearly show that animals co-expressing ChrimsonR and stGtACR2 present strong preference and avoidance behaviors following red and blue stimulations, respectively.

By contrast, the control groups without photosensitive opsins (expressing only EGFP and mCherry) show little difference of behaviors under optical stimulations. More results are provided in Figures S12 and S13, comparing these animals with other group expressing only ChrimsonR or stGtACR2. When introducing only ChrimsonR, animals exhibit place preference under both red and blue stimulations. For the group with only stGtACR2, blue light effectively evokes place aversion while red light generates little influence on animal behaviors. Figures 3g and 3h present the dynamic behavior change at different time courses. These results are consistent with the spectral characteristics of the opsins as well as the *in vitro* neural responses at the cellular level (Figure 2), elucidating the utility of bidirectional modulations with wireless, dual-color micro-LED probes. In conclusion, the implantable wireless dual-color LED probe combining with two spectrally distinct channelrhodopsins can realize bi-directional control the place preference or aversion in freely behaving mice.

Bidirectional optogenetic regulation of dopamine release *in vivo*, with wireless dual-color micro-LED probes

The effective bidirectional behavior manipulations demonstrated in Figure 3 can be ascribed to the light induced activation and inhibition of dopaminergic neurons in the VTA. We further investigate the effect of optogenetic stimulations on the dopamine release using *in vivo* experiments described in Figure 4a. Similarly, combined opsins are expressed and followed by implanting the dual-color micro-LED probe in the VTA of *DAT-cre* mice. While stimulations are performed in the VTA, dopamine signals are simultaneously monitored in one of its main projection areas, the nucleus accumbens (NAc)^{40, 41, 42}. A genetically encoded indicator GRAB_{DA2m}⁴³ is expressed in the NAc, where the fluorescence signals are recorded with a fiber photometer^{28, 29, 39, 44, 45, 46}. The combination of micro-LED based optogenetic stimulations and fiber based photometric recordings realizes all-optical interrogation of neural circuits. During the surgery, we carefully select the insertion direction of the micro-LED probe, in order to obtain desirable stimulation efficacy in the VTA and minimize the optical crosstalk that could disturb photometric recording in the NAc (Figure S14). Fluorescence images in Figure 4b show the expressions of GRAB_{DA2m} in the NAc as well as ChrimsonR and stGtACR2 in the VTA.

Figure 4c plots temporally resolved dopamine release in respond to stimulations caused by red and blue micro-LEDs in a same mouse. More representative curves are presented in Figure S15, with results summarized in Figure 4d. Illuminating the VTA by red or blue micro-LEDs (with injection currents from 0 to 10 mA) clearly induces increased or decreased dopamine signals in the NAc, which correlates with neuron responses at the cellular level (Figure 2). Experiments operated among animal groups expressing different opsins are also performed, with results compared in Figure 4e. In contrast to the group expressing both opsins (ChrimsonR and stGtACR2), mice with only ChrimsonR show increased dopamine release under both red and blue illuminations, and dopamine suppression is observed under blue illumination for the mice with only stGtACR2. The control group

(with EGFP and mCherry) exhibits no response under optical stimulations. These results evidently explain *in vivo* preference and avoidance behaviors presented in Figure 3, showcasing the capability for precise control of neural activities based on bidirectional, dual-color modulations.

We further investigate neural activities with coincident red and blue LED radiations. Figure 4f plots example traces of dopamine transients, in response to the red and blue micro-LEDs under pulse and continuous operations (6-s duration) with varied currents. Statistical data for accumulative dopamine concentration variations (unit: area under the curve during stimulation, or AUC) are summarized in Figure 4g and Figure S16. Decreasing power of the red LED and increasing the power of the blue LED cause the decline of dopamine release. Additionally, the blue LED current required to suppress the dopamine release is much lower than that of the red LED required for dopamine level rises, which further confirms with the cellular activities reported in Figure 2.

Figure 4h (and more examples in Figure S17) presents dopamine dynamics in response to alternated or combined red and blue LED stimulations, for mice expressing ChrimsonR and stGtACR2. The red illumination triggers the elevated dopamine release, while the use of blue light can effectively and promptly suppress the signal. Similar illumination patterns are implemented on animals expressing different opsins (Figure S18) and results are summarized and compared in Figures 4i–4k. Evidently, red or blue illuminations bidirectionally regulate dopamine signals in the experiment group (with both ChrimsonR and stGtACR2). On the other hand, animals expressing only ChrimsonR can be excited by both red and blue light, and those with only stGtACR2 can only be inhibited by blue light and show no reaction under red light. Finally, the control group (with EGFP and mCherry) does not respond to any optical stimulation patterns. Taken together, these results clearly demonstrate that wireless dual-color LED combined with optogenetics enable

efficient, bidirectional control of neuronal activity in the VTA and DA release of the terminal in the NAc.

DISCUSSION

Optogenetic modulations offer a viable solution for cell-type-specific targeting, but still demand biological and engineering tools for versatile functionalities. In this work, we report that bidirectional optogenetic activation and inhibition in behaving animals can be achieved by co-expressing spectrally distinct excitatory and inhibitory channelrhodopsins, cooperating with a wireless, dual-color micro-LED probe for colocalized red and blue stimulations. With collective results at cellular, circuit and behavioral levels, successful proof-of-concept demonstrations in the VTA-NAc circuitry in freely moving mice clearly establish the utility of such a synergistic methodology.

The unique characteristics of thin-film, microscale optoelectronic devices and their integration strategies implicate prominent advantages over conventional fibers or single color LED sources. In terms of technology developments, more directions can be explored and some preliminary efforts are presented below: 1) Multiple red and blue micro-LEDs can be assembled on the implantable probe via the stacking approach, realizing multi-channel, bidirectional modulations imposed on multiple brain regions (Movie S3). It is noted that the efficacy of preference behavior regulations demonstrated in Figure 3 would also be enhanced by implementing dual channel stimulations bilaterally in the VTA^{47, 48, 49, 50}. 2) The use of miniaturized wireless control circuits allows independently addressing multiple, untethered animal objects, making the interrogations of complex social behaviors more feasible (Movie S4)^{51, 52, 53, 54}. 3) Micro-LEDs with other colors can also be included and target on specific opsins more efficiently. For example, a tri-color probe can be achieved by similarly stacking red, green and blue micro-LEDs in a vertical structure (Movie S5). Independently controlled

micro-LEDs can interact with different opsins that express in the same or different populations of cells. 4) Micro-LED based optogenetics can also be combined with previously reported electrical, optical and chemical sensors as well as microchannel devices for drug administrations^{55, 56, 57, 58}, achieving closed-loop neural activity stimulations and detections. Taken together, we envision a multi-spectral, multi-channel, multi-objective and multi-functional optical neural interface, realizing wireless, real-time, bi-directional, close-loop and simultaneous interrogations of neural signals in living animals.

On the other hand, the efficacy of optogenetic modulations can be further optimized by advanced biological engineering. A recent effort fuses excitatory and inhibitory channelrhodopsins (ChrimsonR and GtACR2) in a single, trafficking-optimized tandem protein (BiPOLES), improving the efficiency of opsin co-expression⁵⁹. In addition to the bidirectional modulation of a same cell population, the dual-color (or even tri-color) micro-LED probe can also selectively activate genetically distinct and spatially intermingled cells expressing spectrally resolved opsins (for example, blue for ChR2 or GtACR2, green for C1V1 or GtACR1 and red for ChrimsonR or halorhodopsin)^{3, 7, 8, 22} and provide prolonged activation of neuronal activity by regulating step-function opsins (for example, SSFO and SOUL)^{11, 13, 60}, which is also of critical importance for understanding the circuit structures and functions. Possible applications would include bidirectional modulations of neural activities by leveraging electrophysiological or neurotransmitter signals for medications of seizures or Parkinson's disease^{27, 61, 62}. Besides central nervous systems, peripheral neural circuits can also be targeted for medical treatments^{63, 64}. In summary, the results presented here provide a viable means for fundamental neuroscience studies and advanced neuroengineering applications.

REFERENCES

1. Won SM, Song E, Reeder JT, Rogers JA. Emerging modalities and implantable technologies for neuromodulation. *Cell* 2020, **181**(1): 115-135.
2. Vázquez-Guardado A, Yang Y, Bandodkar AJ, Rogers JA. Recent advances in neurotechnologies with broad potential for neuroscience research. *Nat Neurosci* 2020, **23**(12): 1522-1536.
3. Boyden ES, Zhang F, Bamberg E, Nagel G, Deisseroth K. Millisecond-timescale, genetically targeted optical control of neural activity. *Nat Neurosci* 2005, **8**(9): 1263-1268.
4. Zhang F, Wang L-P, Brauner M, Liewald JF, Kay K, Watzke N, *et al.* Multimodal fast optical interrogation of neural circuitry. *Nature* 2007, **446**(7136): 633-639.
5. Yizhar O, Fenno LE, Davidson TJ, Mogri M, Deisseroth K. Optogenetics in neural systems. *Neuron* 2011, **71**(1): 9-34.
6. Chow BY, Han X, Dobry AS, Qian X, Chuong AS, Li M, *et al.* High-performance genetically targetable optical neural silencing by light-driven proton pumps. *Nature* 2010, **463**(7277): 98-102.
7. Klapoetke NC, Murata Y, Kim SS, Pulver SR, Birdsey-Benson A, Cho YK, *et al.* Independent optical excitation of distinct neural populations. *Nat Methods* 2014, **11**(3): 338-346.
8. Govorunova EG, Sineshchekov OA, Janz R, Liu X, Spudich JL. Natural light-gated anion channels: A family of microbial rhodopsins for advanced optogenetics. *Science* 2015, **349**(6248): 647.
9. Mahn M, Gibor L, Patil P, Cohen-Kashi Malina K, Oring S, Printz Y, *et al.* High-efficiency optogenetic silencing with soma-targeted anion-conducting channelrhodopsins. *Nat Commun* 2018, **9**(1): 4125.

10. Zhang F, Prigge M, Beyriere F, Tsunoda SP, Mattis J, Yizhar O, *et al.* Red-shifted optogenetic excitation: a tool for fast neural control derived from *Volvox carteri*. *Nat Neurosci* 2008, **11**(6): 631-633.
11. Berndt A, Yizhar O, Gunaydin LA, Hegemann P, Deisseroth K. Bi-stable neural state switches. *Nat Neurosci* 2009, **12**(2): 229-234.
12. Hochbaum DR, Zhao Y, Farhi SL, Klapoetke N, Werley CA, Kapoor V, *et al.* All-optical electrophysiology in mammalian neurons using engineered microbial rhodopsins. *Nat methods* 2014, **11**(8): 825-833.
13. Gong X, Mendoza-Halliday D, Ting JT, Kaiser T, Sun X, Bastos AM, *et al.* An ultra-sensitive step-function opsin for minimally invasive optogenetic stimulation in mice and macaques. *Neuron* 2020, **107**(1): 38-51.
14. Brown J, Behnam R, Coddington L, Tervo DGR, Martin K, Proskurin M, *et al.* Expanding the optogenetics toolkit by topological inversion of rhodopsins. *Cell* 2018, **175**(4): 1131-1140.
15. Chuong AS, Miri ML, Busskamp V, Matthews GAC, Acker LC, Sørensen AT, *et al.* Noninvasive optical inhibition with a red-shifted microbial rhodopsin. *Nat Neurosci* 2014, **17**(8): 1123-1129.
16. Chuong AS, Miri ML, Busskamp V, Matthews GAC, Acker LC, Sørensen AT, *et al.* Noninvasive optical inhibition with a red-shifted microbial rhodopsin. *Nature neuroscience* 2014, **17**(8): 1123-1129.
17. Schild LC, Glauser DA. Dual Color Neural Activation and Behavior Control with Chrimson and CoChR in *Caenorhabditis elegans*. *Genet* 2015, **200**(4): 1029-1034.
18. Maimon BE, Sparks K, Srinivasan S, Zorzos AN, Herr HM. Spectrally distinct channelrhodopsins for two-colour optogenetic peripheral nerve stimulation. *Nat Biomed Eng* 2018, **2**(7): 485-496.

19. Lu Y, Truccolo W, Wagner FB, Vargas-Irwin CE, Ozden I, Zimmermann JB, *et al.*
Optogenetically induced spatiotemporal gamma oscillations and neuronal spiking
activity in primate motor cortex. *J Neurosci* 2015, **113**(10): 3574-3587.
20. Galvan A, Caiola MJ, Albaugh DL. Advances in optogenetic and chemogenetic
methods to study brain circuits in non-human primates. *J Neural Transm* 2018,
125(3): 547-563.
21. Park S, Loke G, Fink Y, Anikeeva P. Flexible fiber-based optoelectronics for neural
interfaces. *Chem Rev* 2019, **48**(6): 1826-1852.
22. Yizhar O, Fenno LE, Prigge M, Schneider F, Davidson TJ, O'Shea DJ, *et al.*
Neocortical excitation/inhibition balance in information processing and social
dysfunction. *Nature* 2011, **477**(7363): 171-178.
23. Kampasi K, English DF, Seymour J, Stark E, McKenzie S, Voroslakos M, *et al.* Dual
color optogenetic control of neural populations using low-noise, multishank
optoelectrodes. *Microsyst Nanoeng* 2018, **4**: 1-16.
24. Kim T-i, McCall JG, Jung YH, Huang X, Siuda ER, Li Y, *et al.* Injectable, Cellular-
Scale Optoelectronics with Applications for Wireless Optogenetics. *Science* 2013,
340(6129): 211.
25. Montgomery KL, Yeh AJ, Ho JS, Tsao V, Mohan Iyer S, Grosenick L, *et al.*
Wirelessly powered, fully internal optogenetics for brain, spinal and peripheral
circuits in mice. *Nat Methods* 2015, **12**(10): 969-974.
26. Shin G, Gomez AM, Al-Hasani R, Jeong YR, Kim J, Xie Z, *et al.* Flexible Near-Field
Wireless Optoelectronics as Subdermal Implants for Broad Applications in
Optogenetics. *Neuron* 2017, **93**(3): 509-521.e503.
27. Wykes RC, Kullmann DM, Pavlov I, Magloire V. Optogenetic approaches to treat
epilepsy. *J Neurosci Methods* 2016, **260**: 215-220.

28. Li Y, Zeng J, Zhang J, Yue C, Zhong W, Liu Z, *et al.* Hypothalamic circuits for predation and evasion. *Neuron* 2018, **97**(4): 911-924.
29. Lu L, Ren Y, Yu T, Liu Z, Wang S, Tan L, *et al.* Control of locomotor speed, arousal, and hippocampal theta rhythms by the nucleus incertus. *Nat Commun* 2020, **11**(1): 1-16.
30. Roseberry TK, Lee AM, Lalive AL, Wilbrecht L, Bonci A, Kreitzer AC. Cell-Type-Specific Control of Brainstem Locomotor Circuits by Basal Ganglia. *Cell* 2016, **164**(3): 526-537.
31. Gradinaru V, Zhang F, Ramakrishnan C, Mattis J, Prakash R, Diester I, *et al.* Molecular and cellular approaches for diversifying and extending optogenetics. *Cell* 2010, **141**(1): 154-165.
32. Stuber GD, Sparta DR, Stamatakis AM, van Leeuwen WA, Hardjoprajitno JE, Cho S, *et al.* Excitatory transmission from the amygdala to nucleus accumbens facilitates reward seeking. *Nature* 2011, **475**(7356): 377-380.
33. Noked O, Levi A, Someck S, Amber-Vitos O, Stark E. Bidirectional optogenetic control of inhibitory neurons in freely-moving mice. *IEEE Transactions on Biomedical Engineering* 2020.
34. Li L, Liu C, Su Y, Bai J, Wu J, Han Y, *et al.* Heterogeneous Integration of Microscale GaN Light-Emitting Diodes and Their Electrical, Optical, and Thermal Characteristics on Flexible Substrates. *Adv Mater Technol* 2018, **3**(1): 1700239.
35. Liu C, Zhang Q, Wang D, Zhao G, Cai X, Li L, *et al.* High Performance, Biocompatible Dielectric Thin-Film Optical Filters Integrated with Flexible Substrates and Microscale Optoelectronic Devices. *Adv Opt Mater* 2018, **6**(15): 1800146.

36. Ding H, Lu L, Shi Z, Wang D, Li L, Li X, *et al.* Microscale optoelectronic infrared-to-visible upconversion devices and their use as injectable light sources. *Proc Natl Acad Sci U S A* 2018, **115**(26): 6632-6637.
37. Lammel S, Lim BK, Malenka RC. Reward and aversion in a heterogeneous midbrain dopamine system. *Neuropharmacology* 2014, **76**: 351-359.
38. Morales M, Margolis EB. Ventral tegmental area: cellular heterogeneity, connectivity and behaviour. *Nat Rev Neurosci* 2017, **18**(2): 73.
39. Cohen JY, Haesler S, Vong L, Lowell BB, Uchida N. Neuron-type-specific signals for reward and punishment in the ventral tegmental area. *Nature* 2012, **482**(7383): 85-88.
40. Al-Hasani R, McCall JG, Shin G, Gomez AM, Schmitz GP, Bernardi JM, *et al.* Distinct Subpopulations of Nucleus Accumbens Dynorphin Neurons Drive Aversion and Reward. *Neuron* 2015, **87**(5): 1063-1077.
41. Sun F, Zeng J, Jing M, Zhou J, Feng J, Owen SF, *et al.* A Genetically Encoded Fluorescent Sensor Enables Rapid and Specific Detection of Dopamine in Flies, Fish, and Mice. *Cell* 2018, **174**(2): 481-496 e419.
42. Yuan L, Dou YN, Sun YG. Topography of Reward and Aversion Encoding in the Mesolimbic Dopaminergic System. *J Neurosci* 2019, **39**(33): 6472-6481.
43. Sun F, Zhou J, Dai B, Qian T, Zeng J, Li X, *et al.* Next-generation GRAB sensors for monitoring dopaminergic activity in vivo. *Nat Methods* 2020, **17**(11): 1156-1166.
44. Gunaydin Lisa A, Grosenick L, Finkelstein Joel C, Kauvar Isaac V, Fenno Lief E, Adhikari A, *et al.* Natural Neural Projection Dynamics Underlying Social Behavior. *Cell* 2014, **157**(7): 1535-1551.
45. Lin R, Liang J, Wang R, Yan T, Zhou Y, Liu Y, *et al.* The Raphe Dopamine System Controls the Expression of Incentive Memory. *Neuron* 2020, **106**(3): 498-514.e498.

- 457 46. Cui G, Jun SB, Jin X, Pham MD, Vogel SS, Lovinger DM, *et al.* Concurrent
458 activation of striatal direct and indirect pathways during action initiation. *Nature*
459 2013, **494**(7436): 238-242.
- 460 47. Yuan L, Liang T-Y, Deng J, Sun Y-G. Dynamics and functional role of dopaminergic
461 neurons in the ventral tegmental area during itch processing. *J Neurosci Res* 2018,
462 **38**(46): 9856-9869.
- 463 48. Li Y, Li C-Y, Xi W, Jin S, Wu Z-H, Jiang P, *et al.* Rostral and caudal ventral
464 tegmental area GABAergic inputs to different dorsal raphe neurons participate in
465 opioid dependence. *Neuron* 2019, **101**(4): 748-761.
- 466 49. Taylor NE, Van Dort CJ, Kenny JD, Pei J, Guidera JA, Vlasov KY, *et al.* Optogenetic
467 activation of dopamine neurons in the ventral tegmental area induces reanimation
468 from general anesthesia. *Proc Natl Acad Sci USA* 2016, **113**(45): 12826-12831.
- 469 50. Chang CY, Gardner M, Di Tillio MG, Schoenbaum G. Optogenetic blockade of
470 dopamine transients prevents learning induced by changes in reward features. *Curr*
471 *Biol* 2017, **27**(22): 3480-3486.
- 472 51. de Chaumont F, Ey E, Torquet N, Lagache T, Dallongeville S, Imbert A, *et al.* Real-
473 time analysis of the behaviour of groups of mice via a depth-sensing camera and
474 machine learning. *Nat Biomed Eng* 2019, **3**(11): 930-942.
- 475 52. Kingsbury L, Huang S, Raam T, Letizia SY, Wei D, Hu RK, *et al.* Cortical
476 representations of conspecific sex shape social behavior. *Neuron* 2020, **107**(5): 941-
477 953.
- 478 53. Kingsbury L, Huang S, Wang J, Gu K, Golshani P, Wu YE, *et al.* Correlated neural
479 activity and encoding of behavior across brains of socially interacting animals. *Cell*
480 2019, **178**(2): 429-446.

54. Hong W, Kim D-W, Anderson DJ. Antagonistic control of social versus repetitive self-grooming behaviors by separable amygdala neuronal subsets. *Cell* 2014, **158**(6): 1348-1361.
55. Chiang C-H, Won SM, Orsborn AL, Yu KJ, Trumpis M, Bent B, *et al.* Development of a neural interface for high-definition, long-term recording in rodents and nonhuman primates. *Sci Transl Med* 2020, **12**(538): eaay4682.
56. Qazi R, Gomez AM, Castro DC, Zou Z, Sim JY, Xiong Y, *et al.* Wireless optofluidic brain probes for chronic neuropharmacology and photostimulation. *Nat Biomed Eng* 2019, **3**(8): 655-669.
57. Lu L, Gutruf P, Xia L, Bhatti DL, Wang X, Vazquez-Guardado A, *et al.* Wireless optoelectronic photometers for monitoring neuronal dynamics in the deep brain. *Proc Natl Acad Sci USA* 2018, **115**(7): E1374-E1383.
58. Liu C, Zhao Y, Cai X, Xie Y, Wang T, Cheng D, *et al.* A wireless, implantable optoelectrochemical probe for optogenetic stimulation and dopamine detection. *Microsyst Nanoeng* 2020, **6**(1): 1-12.
59. Vierock J, Rodriguez-Rozada S, Pieper F, Dieter A, Bergs A, Zeitzschel N, *et al.* BiPOLES: a tool for bidirectional dual-color optogenetic control of neurons. *BioRxiv* 2020. <https://doi.org/10.1101/2020.07.15.204347>
60. Selimbeyoglu A, Kim CK, Inoue M, Lee SY, Hong ASO, Kauvar I, *et al.* Modulation of prefrontal cortex excitation/inhibition balance rescues social behavior in CNTNAP2-deficient mice. *SCI TRANSL MED* 2017, **9**(401): eaah6733.
61. Chen S, Weitemier AZ, Zeng X, He L, Wang X, Tao Y, *et al.* Near-infrared deep brain stimulation via upconversion nanoparticle-mediated optogenetics. *Science* 2018, **359**(6376): 679.

62. Kravitz AV, Freeze BS, Parker PRL, Kay K, Thwin MT, Deisseroth K, *et al.*
Regulation of parkinsonian motor behaviours by optogenetic control of basal ganglia
circuitry. *Nature* 2010, **466**(7306): 622-626.
63. Keppeler D, Schwaerzle M, Harczos T, Jablonski L, Dieter A, Wolf B, *et al.*
Multichannel optogenetic stimulation of the auditory pathway using microfabricated
LED cochlear implants in rodents. *Sci Transl Med* 2020, **12**(553): eabb8086.
64. Michoud F, Seehus C, Schönle P, Brun N, Taub D, Zhang Z, *et al.* Epineural
optogenetic activation of nociceptors initiates and amplifies inflammation. *Nat*
Biotechnol 2021, **39**(2): 179-185.

Acknowledgements

This work is supported by National Natural Science Foundation of China (NSFC) (61874064), Beijing Municipal Natural Science Foundation (4202032), Beijing Innovation Center for Future Chips at Tsinghua University, Center for Flexible Electronics Technology at Tsinghua University, Beijing National Research Center for Information Science and Technology (BNR2019ZS01005). We thank Prof. Y. Li (Peking University) for support on GRAB_{DA2m}, Y. Zhou and Y. Ju (Chinese Institute for Brain Research, Beijing) for support on stGtACR2.

Author contributions

X. S. and M. L. developed the concepts. L. Li, G. T., Y. Z., X. C., Z. S., H. D., C. L., Y. X., H. W., L. Y. and X. S. performed material, device and circuit design, fabrication and characterization. L. Li, X. C., D. C, and X. S. performed the simulations. L. Li, L. Lu, Y. R., G. T., X. F., L. Y., M. L., and X. S. designed and performed biological experiments. L. Li, L. Lu, M. L. and X. S. wrote the paper in consultation with other authors.

Competing interests

The authors declare no competing interests.

Data and materials availability

All data needed to evaluate the conclusions in the paper are present in the paper and/or the Supplementary information.

Figure 1

Figure 1. Wirelessly operated, dual-color microscale LED (micro-LED) probes for bidirectional, colocalized optogenetic control of neural activities. **a**, Schematic view of the probe inserted into the target brain area to perform dual-color neural activation and inhibition. **b**, Schematic, cellular scale depiction of the modulation function: red light regulates the cation channel ChrimsonR for depolarization, while blue light regulates the anion channel stGtACR2 for hyperpolarization. **c**, Exploded view of the dual-color LED probe made from vertically stacked blue and red thin-film micro-LEDs assembled on a flexible polyimide (PI) substrate, with a thin-film filter interface for spectrally selective reflection and transmission. **d**, Cross-sectional scanning electron microscopic (SEM) image of the probe structure, with false colors highlighting (from bottom to top) the indium gallium arsenide (InGaP) red micro-LED, the multilayer silicon dioxide / titanium dioxide (SiO₂/TiO₂) filter, and the indium gallium nitride (InGaN) blue micro-LED. Layers of SU-8 photoresists serve as the bonding materials. **e**, Optical images of a micro-LED probe (with a thickness of ~120 μm and a width of ~320 μm), showing independent control of blue and red emissions under current injections. **f**, Emission spectra of the blue and red micro-LEDs, in comparison with the action spectra of ion channels stGtACR2 and ChrimsonR (data extracted from references^{7, 8}). **g**, Experimental photographs and simulated results illustrating blue and red light propagations for a micro-LED probe embedded into a brain phantom. Results for a probe with laterally assemble blue and red LEDs are also presented, showing misplaced blue and red emissions. Iso-intensity lines show 10% of the maximum power. **h**, Images of a micro-LED probe integrated with a wireless circuit for independent control of dual-color emissions. **i**, Photos of a behaving mouse after intracranial implantation of a probe, with (bottom) and without (top) the wireless circuit module.

Figure 2

Figure 2. Bidirectional optogenetic activation and inhibition of neurons with red and blue light stimulation. **a**, Schematic strategy for co-expressing ChrimsonR and stGtACR2 in the ventral tegmental area (VTA) of *DAT-cre* mice. **b**, Immunostained fluorescence images of the VTA region expressing DAPI (blue), stGtACR2 (green), and ChrimsonR (red) and a merged image, 3 weeks after viral injection. **c**, Zoomed view showing cells co-expressing stGtACR2 (green) and ChrimsonR (red). **d**, Percentages of stGtACR2-, ChrimsonR- and co-expressing neurons among DAPI cells in the VTA (mean $\pm$ s.e.m., 3 slices per animal, $n = 3$ mice). **e**, Illustration of the experiment setup with combined blue (480 nm) and red (633 nm) laser beams (spot diameter ~ 1 mm) for optogenetic stimulating brain slices during patch-clamp recordings *in vitro*. **f**, Example traces showing neural activation by pulsed red light at different irradiances (20 Hz, 10-ms pulse, 20-pulse train). **g**, Activated spike probability as a function of red irradiances (mean $\pm$ s.e.m., $n = 4$ cells). **h**, Example photocurrents of neurons at membrane potentials varied from -105 mV to -5 mV (bottom to top), under continuous blue light illumination (duration 5 s, irradiance 8 mW/mm²). **i**, Example traces of neural inhibition by continuous blue light at different irradiances (duration 0.5 s). **j**, Inhibited spike probability as a function of blue irradiances (mean $\pm$ s.e.m., $n = 5$ cells). **k**, Traces presenting bidirectional spike activation and inhibition with alternating red and blue illuminations (Patching current $+90$ pA; Red: duration 3 s, 20 Hz, 10-ms pulse, 2.2 mW/mm²; Blue: duration 3 s, 8.5 mW/mm²). **l**, Example traces of neural activities with simultaneous red and blue illuminations at different irradiances (duration 6 s). **m**, Summarized plot of measured spike probability (normalized mean values) as a function of red and blue irradiances ($n = 5$ cells).

Figure 3

Figure 3. Bidirectional, *in vivo* optogenetic regulation of animals' preference or aversion behaviors with wirelessly operated, dual-color microscale LED (micro-LED) probes. **a**, Diagram illustrating the combined viral injection (AAV2/9-CAG-DIO-ChrimsonR-mCherry and AAV2/9-EF1a-DIO-stGtACR2-EGFP) followed by implanting the micro-LED probe into the VTA of *DAT-cre* mice. **b**, Micrographs of the VTA showing co-expression of stGtACR2 and ChrimsonR, with a dual-color micro-LED probe embedded into the tissue. **c**, Illustration of the real-time place preference test with a two-compartment arena, where a mouse receives signals for dual-color control and its locomotion is recorded by a camera. **d**, Patterns used for optogenetic modulations, including a 10-min pretest, a 30-min red LED stimulation (20 Hz, 10-ms pulse, current 7 mA), a 30-min rest, and a 30-min blue LED stimulation (continuous, current 5 mA). **e**, Representative heat maps showing real-time preference and aversion behavior following red or blue stimulation for mice co-expressing stGtACR2 + ChrimsonR (experiment, or Exp) or EGFP + mCherry (control, or Ctrl). **f**, Summary of preference indices (the ratio of the time that mice spend in the left chamber to the whole recorded time) under red and blue stimulations for Exp and Ctrl groups ($n = 4$ mice each). Student's t test, ** $P < 0.01$, *** $P < 0.001$, n.s. $P > 0.1$. **g**, Preference indices measured at different times under red and blue stimulations for Exp and Ctrl groups ($n = 4$ mice each). **h**, Preference indices measured at different times under red and blue stimulations for groups expressing only ChrimsonR or stGtACR2 ($n = 3$ mice each).

Figure 4

Figure 4. Bidirectional, *in vivo* optogenetic control of dopamine (DA) release simultaneously recorded in the nucleus accumbens (NAc) during dual-color stimulation in the VTA. **a**, Diagram illustrating the combined viral injection (AAV2/9-CAG-DIO-ChrimsonR-mCherry and AAV2/9-EF1a-DIO-stGtACR2-EGFP) followed by implanting the micro-LED probe into the VTA, and the viral injection (AAV2-hSyn-GRAB_{DA2m}) followed by optical fiber implantation into the NAc of *DAT-cre* mice. **b**, Micrographs of (top) the NAc showing expression of DA2m and the fiber track, (bottom) the VTA showing co-expression of stGtACR2 and ChrimsonR (merged color) and the micro-LED probe track, 2 weeks after viral injection and probe implantation. **c**, Representative traces of DA signal increase or decrease in response to stimulations by the red LED (7 mA, 20 Hz, 20-ms pulse, 6 s) or the blue LED (2.4 mA, continuous, 12 s). The solid lines and shaded areas indicate the mean and s.e.m., respectively ($n = 4$ mice). **d**, Accumulative DA signals in response to red and blue stimulations under different LED currents. AUC: area under the curve during stimulation. **e**, Summary of measured DA signals (AUC) under red or blue stimulations (Red LED: 20 Hz, 20-ms pulse, 10 mA, 6 s; Blue LED, continuous, 3 mA, 12 s) for different animal groups expressing: ChrimsonR + stGtACR2 ($n = 4$ mice), ChrimsonR + EGFP ($n = 3$ mice), stGtACR2 + mCherry ($n = 3$ mice), EGFP + mCherry ($n = 3$ mice). Student's *t* test. **f**, Example traces of DA signals under combined red and blue stimulations with varied LED currents. **g**, Summarized plot of measured DA signals (AUC) (mean value) as a function of red and blue LED currents for 6 s ($n = 3$ mice). **h**, Plots of optogenetically evoked DA transients for mice co-expressing stGtACR2 + ChrimsonR, in response to different stimulation patterns (Red: red LED on, 20 Hz, 20-ms pulse, 10 mA, 0–12 s; Blue: blue LED on, continuous, 3 mA, 6–12 s; Green: red LED on 0–12 s and blue LED on 6–12 s). The solid lines and shaded areas indicate the mean and s.e.m., respectively. ($n = 4$ mice). **i–k**, Summary of measured DA signals (AUC) for mice with and without stGtACR2 and ChrimsonR expressions ($n = 3$ or 4 mice each), in response to different stimulation patterns. **i**, red LED on, 20 Hz, 20-ms pulse, 10 mA, 0–12 s. **j**, blue LED on, continuous, 3 mA, 6–12 s. **k**, red LED on 0–12 s and blue LED on 6–12 s.

Figure S1

Figure S1. Schematic illustration of the process flow for fabricating the dual-color micro-LED probes, with vertically stacked red LED, filter and blue LED printed on PI substrates.

Figure S2

Figure S2. Measured optoelectronic properties of the red and the blue micro-LEDs in a probe structure. (a) Current–voltage curves. (b) External quantum efficiencies (EQEs) as a function of current. (c) Irradiance (power density on the LED surface) versus current. (d) Angular dependent emission profiles (in arbitrary unit).

Figure S3

Figure S3. Thermal behaviors of a dual-color micro-LED probe. (a, b) Images and infrared thermographs showing a probe embedded into a brain phantom (~0.5 mm below the phantom surface). (a) Red LED on (injection current 7 mA, pulse frequency 20 Hz, pulse duration 10 ms). (b) Blue LED on (injection current 2 mA, continuous operation). (c–j) Measured and simulated temperature rise in phantom or brain when the red LED or the blue LED is on. (c) Temperature rise in phantom, measured by an infrared thermal camera when the red LED is on (current 0–10 mA, frequency 10–50 Hz, pulse duration 10 ms). (d) Temperature rise in the brain of living mice, measured with a thin wire thermocouple attached to the probe. (e) Simulated steady-state temperature distribution in the brain tissue surrounding the probe, when the red LED is on (current 7 mA, frequency 20 Hz, pulse duration 10 ms). (f) Simulated maximum temperature rise of the brain tissue when the red LED is in different operation conditions. (g–j) Corresponding results when the blue LED is on (continuous mode). current = 2 mA in (i).

Figure S4

Figure S4. Schematic diagram showing the operational principle for the wireless circuit system controlling the dual-color micro-LED probe.

Figure S6

Figure S6. Layout of the designed printed circuit board (top: front view; bottom: back view).

Figure S7

	components
L1, L2	High frequency chip inductor +/- 5%, 6.8 nH
L3	High frequency chip inductor +/- 5%, 5.6 nH
C1, C2	NP0 +/- 2%, 15 pF
C3, C4	X7R +/- 10%, 2.2 nF
C5	NP0 +/- 0.1pF, 1.5 pF
C6	NP0 +/- 0.1pF, 1.0 pF
C7,C9,C11	X7R +/- 10%, 100 nF
C8,C10	X7R +/- 10%, 33 nF
C12,C14	X7R +/- 10%, 1 μ F
Cout1,Cout2	2.2 μ F
C13	NP0 50V, 10 pF
D0, D1	Indicating red LED, 610 nm, SMD 0402, 1 mm * 0.5 mm * 0.4 mm
R1	22 k Ω , 1%
nRF24LE1	nRF24LE1-F16Q24, QFN24 4 mm * 4 mm package, MCU
Component_1 Component_2	ZLED7012 ZI1R, LED current driver
X1	Crystal oscillator, SMD-3225, 16 MHz, CL=9 pF, +/- 60 ppm 3.2 mm * 2.5 mm package
E1	Antenna, H2U34WGTQW0100, AA055, 3.2 mm * 1.6 mm * 0.5 mm package
P1, P2	Pin Header, 6 pin, 1.25 mm, for programming
P3	FPC connector, 4 pin, 0.5 mm, for LED probe
P4	PCB header Connector, male, DIP Straight, 2 pin, 2.54 mm, for battery
Battery	Rechargeable lithium ion battery, 3.7 V, 45 mAh

Figure S7. List of components used in the wireless circuit system.

Figure S8

Figure S8. Recorded traces of a representative neuron with combined red and blue illumination at different irradiances. Red LED: 20 Hz, 10-ms pulse; Blue LED: continuous operation.

Figure S9

Figure S9. Immunoreactivity results showing the biocompatibility of the micro-LED probe, in comparison with those obtained with a silica based optical fiber. (a) Representative confocal fluorescence images of horizontal brain slices, showing immunohistochemical staining of astrocytes (GFAP) and activated microglia (Iba1) for both the LED probe and the fiber, after 1 day and 21 days implantation. Green: GFAP; Red: Iba1; Blue: DAPI. (b) Percentages of GFAP and Iba1 cell populations among DAPI cells collected at a distance of 200 μm from the edge of implantation. The LED probe and the fiber show similar results of inflammatory glial responses occurring after implantation. ($n = 3$ mice for each group).

Figure S10

Figure S10. Number of functional micro-LEDs measured after chronic implantation in behaving mice.

Figure S11

Figure S11. (a) Schematic diagram and (b) top-view photograph of the real-time place preference test with a two-compartment arena.

Figure S12

Figure S12. Representative heat maps showing real-time preference and aversion behavior following red or blue stimulation for mice expressing stGtACR2 + ChromsonR (experiment, or Exp), EGFP + mCherry (control, or Ctrl), ChromsonR + EGFP and stGtACR2 + mCherry. The data for ChromsonR + stGtACR2 and EGFP + mCherry groups are identical to those presented in Fig. 3e.

Figure S13

Figure S13. Summary of preference indices (the ratio of the time that mice spend in the left chamber to the whole recorded time) under red and blue stimulations for different animal groups expressing ChromsonR + stGtACR2 ($n = 4$ mice), EGFP + mCherry ($n = 4$ mice), ChromsonR + EGFP ($n = 3$ mice), stGtACR2 + mCherry ($n = 3$ mice). Student's t test, ** $P < 0.01$, *** $P < 0.001$, n.s. $P > 0.1$. The data for ChromsonR + stGtACR2 and EGFP + mCherry groups are identical to those presented in Fig. 3f.

Figure S14

Figure S14. Schematic illustration (horizontal view) for the implantation of the fiber for DA recording in the NAc, and the dual-color LED probe for optogenetic stimulations in the VTA. The specific insertion angle is chosen for the LED probe, to minimize the optical crosstalk that disturbs the fiber photometric recording in the NAc.

Figure S15

Figure S15. DA signals recorded as the fluorescence of GRAB_{DA2m} in the NAc, in response to optogenetic stimulations applied in the VTA by red or blue micro-LEDs. (a, b) Representative traces of DA signals in response to stimulations by (a) the red LED (20 Hz, 20 ms, 6 s) and (b) the blue LED (continuous, 12 s) at various currents. The solid lines and shaded areas indicate the mean and s.e.m., respectively. ($n = 4$ mice). AUC: area under the curve during stimulation. The data for red LED at 7 mA and blue LED at 2.4 mA are identical to those presented in Figure 4c.

Figure S16

Figure S16. Detailed results of measured DA signals (AUC) with the combination of red and blue irradiance at different LED currents simultaneously for 6 s ($n = 3$ mice).

Figure S17

Figure S17. Example traces of optogenetically evoked DA transients for mice co-expressing stGtACR2 + ChrimsonR, in response to different stimulation patterns. (a) Red LED on, 20 Hz, 20-ms pulse, 10 mA, 0–12 s. (b) Blue LED on, continuous, 3 mA, 0–12 s. (c) Blue LED on, continuous, 3 mA, 6–12 s. (d) Red LED on 0–12 s and Blue LED on 6–12 s. (e) Red LED on 0–6 s and Blue LED on 0–12 s. The solid lines and shaded areas indicate the mean and s.e.m., respectively.

Figure S18

Figure S18. Plots of optogenetically evoked DA transients in response to different stimulation patterns (Red: red LED on, 20 Hz, 20-ms pulse, 10 mA, 0–12 s; Blue: blue LED on, continuous, 3 mA, 6–12 s; Green: red LED on 0–12 s and blue LED on 6–12 s), for different animal groups expressing (a) ChrimsonR + stGtACR2 ($n = 4$ mice), (b) ChrimsonR + EGFP ($n = 3$ mice), (c) stGtACR2 + mCherry ($n = 3$ mice), (d) EGFP + mCherry ($n = 3$ mice). The solid lines and shaded areas indicate the mean and s.e.m., respectively. The data in Figure S18a are identical to those presented in Figure 4h.

Movie S1

Movie S1. Video for dual-color micro-LED probe, displaying alternating blue and red emissions.

Movie S2

Movie S2. Video for multiple dual-color LED probes assembled with wirelessly operated circuit modules, showing capabilities for independent light emission control.

Movie S3

Movie S3. Video for multi-channel dual-color probes with 6 pairs of stacked red-blue micro-LEDs, displaying alternating blue and red emissions.

Movie S4

Movie S4. Video for four freely moving mice implanted with micro-LED probes and head-mounted circuits.

Movie S5

Movie S5. Video for a tri-color (red-green-blue) micro-LED needle structure, displaying different colors under current injection.

Supplementary Information

Device Fabrication

Fabrication of red and blue micro-LEDs

Details about the thin-film micro-LED structures and fabrication processes can be found in our previous work.^{1,2} Via metal-organic chemical vapor deposition (MOCVD), red and blue LEDs are originally grown on gallium arsenide (GaAs) and sapphire substrates, respectively. The red LED active device contains an indium gallium phosphide (InGaP) based multiple quantum well structure (thickness $\sim 5.6 \mu\text{m}$) grown on GaAs with an $\text{Al}_{0.95}\text{Ga}_{0.05}\text{As}$ sacrificial interlayer. The blue LED active device contains an indium gallium nitride (InGaN) based multiple quantum well structure (thickness $\sim 7 \mu\text{m}$) grown on sapphire. Photolithographic process, wet / dry etching and metallic electrode deposition define the LED geometry (size $\sim 180 \mu\text{m} \times 125 \mu\text{m}$). Freestanding, thin-film red LEDs are formed by selectively removing the $\text{Al}_{0.95}\text{Ga}_{0.05}\text{As}$ sacrificial layer in a hydrofluoric acid (HF) based solution. Freestanding, thin-film blue LEDs are formed by laser lift off (with a KrF excimer laser at 248 nm). Poly(dimethylsiloxane) (PDMS) based stamps are employed to pick up released thin-film micro-LEDs and transferred them onto other foreign substrates.

Fabrication of thin-film filters

Details about the thin-film filter fabrication can be found in our previous work.³ The wavelength selective thin-film filters is based on a multilayered titanium dioxide

(TiO₂) and silicon dioxide (SiO₂) structure (thickness ~ 6.6 μm) comprising (from top to bottom): 19 nm TiO₂ / 15 periods (89 nm SiO₂ + 52 nm TiO₂) / 63 nm SiO₂ / 66 nm TiO₂ / 22 periods of (73 nm SiO₂ + 42.5 nm TiO₂) / 154 nm SiO₂. The structure is deposited on GaAs by ion beam-assisted sputter deposition. The filter shape (size ~ 180 μm × 125 μm) is defined by laser milling (Nd:YVO₄ laser, 1064 nm). After removing the GaAs substrate in solutions (NH₄OH:H₂O₂:H₂O = 1:1:2), freestanding thin-film filters can be picked up and transfer printed onto other foreign substrates by PDMS stamps.

Fabrication of dual-color micro-LED probes

The process flow to fabricate the dual-color micro-LED probe is described in Figure S1. The micro-LED probe is based on a flexible substrate made of composite polyimide (PI) and copper (Cu) films (from top to bottom: Cu/PI/Cu = 18/25/18 μm, from DuPont). From bottom to top, a red LED, an optical filter and a blue LED are sequentially transfer printed on the film by PDMS, in a vertical stack. Between each device layer, SU-8 based epoxy (thickness ~ 2–5 μm) serves an optically transparent and electrically insulating bonding layer. Each individual LED is metallized with sputtered metal layers (Cr/Au/Cu/Au = 10/100/500/100 nm). Laser milling (ultraviolet laser 365 nm) defines the needle shape (width ~ 320 μm, length ~ 5 mm). A bilayer of PDMS (~20 μm) and parylene (~15 μm) is used for waterproof encapsulation. For comparison in Fig. 1g, laterally arranged red-blue micro-LED probes are also fabricated with similar procedures.

Device Characterization and Modeling

Structural and optoelectronic characterizations

Scanning electron microscopic (SEM) images are taken with a ZEISS Auriga SEM / FIB Crossbeam System. Optical images are captured by a microscope MC-D800U(C). The LED emission spectra are measured with a spectrometer (HR2000+, Ocean Optics). Current–voltage characteristics are recorded by a Keithley 2400 source meter. LED irradiances and external quantum efficiencies are measured using a spectroradiometer with an integrating sphere (LabSphere Inc.). The angular dependent emission profiles are measured by a standard Si photodetector (DET36A, Thorlabs) with the device mounted on a goniometer. For *in vitro* test, micro-LED probes are implanted into the brain phantom. The brain phantom is made of agarose (0.5% w/v), hemoglobin (0.2% w/v), and intralipid (1% w/v) mixed in phosphate buffer solutions (98.3% w/v). After stirring, the mixed liquid is heated to boiling and then naturally cooled to room temperature, forming the gels.

Optical modeling

A ray tracing method based on Monte Carlo simulations is used to simulate the light propagation in the brain tissue. In the model, the tissue has absorption coefficients of 0.2 /mm and 0.08 /mm, and scattering coefficients of 47 /mm and 35 /mm, for blue (475 nm) and red (630 nm), respectively. The tissue has an anisotropy factor 0.85 and a refractive index of 1.36.⁴

Thermal characterizations

For *in vitro* experiments, the micro-LED probes are implanted into the brain phantom (~0.5 mm underneath the surface, at a depth of ~5 mm), and various currents are injected into the LEDs. The temperature distributions are mapped with an infrared camera (FOTRIC 228). For *in vivo* experiments, the micro-LED probe is bonded with an ultra fine flexible micro thermocouple (Physitemp, W-TW-44P/50) and inserted into the anesthetized mouse brain. Steady-state temperature rises are recorded from the thermocouple with a thermometer (Physitemp, BAT-12).

Thermal modeling

3D steady-state heat transfer models are established by finite element analysis (Comsol Multiphysics). The parameters of the brain use in the model include: density 1.1 g/cm³, thermal conductivity 0.5 W/m/K, heat capacity 3650 J/kg/K, and the natural heat convection to air. The micro-LED probe serves as the heat source, with the input thermal power estimated by $P = V \times I \times (1 - EQE)$, where V , I and EQE are the measured voltage, current, and corresponding external quantum efficiencies for LEDs.

Wireless Circuit Design

A customized circuit module is designed to wirelessly operate the implantable micro-LED probe by independently controlling the red and the blue micro-LEDs for bi-directional modulation. The schematic diagram is depicted in Figure S4, with detailed circuit diagram, layouts and components presented in Figures S5, S6 and S7,

respectively. The core components include: a micro-controller (nRF24LE1, Nordic Semiconductor) operated at 2.4 GHz for data communication, and two LED drivers (ZLED7012, Renesas Electronics Corp.) with programmable current levels, frequencies, pulse widths for providing a constant current to micro-LEDs.⁵ Two red LEDs are used on the circuit board outside the brain for signal indication, without affecting the mouse behaviors.⁶ Control signals are wirelessly transmitted from the transmitting module connected to a laptop computer via an antenna operated at a radio frequency of 2.4 GHz, with a communication distance up to 50 m. The wireless circuit has a footprint of $\sim 2.2 \text{ cm} \times 1.3 \text{ cm}$ and a weight of 1.9 g (including a 0.9-g rechargeable lithium-ion battery with a capacity of 35 mAh). Used in the experimental conditions in this paper, the battery lifetime can reach up to 2 hours. The circuit can be connected with the implanted micro-LED probes via a flexible printed circuit (FPC) connector before animal experiment.

Biological Studies

Animals

Animal care is in accordance with the institutional guidelines of National Institute of Biological Sciences in Beijing (NIBS), with protocols proved by Institutional Animal Care and Use Committee (IACUC). All animals are socially housed in a 12 h/12 h (lights on at 8 am) light/dark cycle, with food and water *ad libitum*. Wildtype C57BL6/N mice are purchased from VitalRiver (Beijing, China). *DAT-Cre* mice

(strain name B6.SJLSlc6a3tm1.1(cre) Bkmn; JAX Strain 006660) are at least six weeks old at the time of surgery.

Virus production

The AAV2/9-CAG-DIO-ChrimsonR-tdTomato (Plasmid #130909) vector⁷ is from Addgene. The AAV2/9-EF1a-DIO-stGtACR2-EGFP-Kv2.1 plasmid is synthesized and constructed according to the original reports.^{8, 9} The AAV-hSyn-GRAB_{DA2m} plasmid is provided by Prof. Yulong Li (Peking University).¹⁰ AAV vectors are packaged into serotype2/9 vectors, which consist of AAV2 ITR genomes pseudotyped with AAV9 serotype capsid proteins. AAV vectors are replicated in HEK293 cells with the triple plasmid transfection system and purified by chloroform,¹¹ resulting in AAV vector titers of about 5×10^{12} particles/mL. Virus suspension is aliquoted and stored at -80°C . AAV titers are measured using real-time qPCR. Adeno-associated viral particles are produced at the Vector Core Facility at Chinese Institute for Brain Research, Beijing (CIBR).

Stereotaxic Surgery

Adult *DAT-Cre* mice are anesthetized with an intraperitoneal injection of Avertin (250 mg/kg) before surgery, and then placed in a mouse stereotaxic instrument. After disinfection with 0.3% hydrogen peroxide, a small incision of the scalp is created to expose the skull. Then, 0.3% hydrogen peroxide is applied to clean the skull. A small craniotomy is made and a calibrated pulled-glass pipette (Sutter Instrument) is lowered to the VTA (coordinates 3.64 mm posterior to the bregma, 0.5 mm lateral to midline, and 4.6 mm ventral to the skull surface). Virus is delivered through a small

durotomy by a glass micropipette using a microsyringe pump (Nanoliter 2000 injector with the Micro4 controller, WPI).

Four different combinations of virus are used for *in vivo* experiments:

1) ChrimsonR + stGtACR2 (AAV2/9-CAG-DIO-ChrimsonR-tdTomato and AAV2/9-EF1a-DIO-stGtACR2-EGFP-Kv2.1, 1:1);

2) mCherry + EGFP (AAV2/9-EF1a-DIO-mCherry and AAV2/9-EF1a-DIO-EGFP, 1:1);

3) ChrimsonR + EGFP (AAV2/9-CAG-DIO-ChrimsonR-tdTomato and AAV2/9-EF1a-DIO-EGFP, 1:1)

4) stGtACR2 + mCherry (AAV2/9-EF1a-DIO-stGtACR2-EGFP-Kv2.1 and AAV2/9-EF1a-DIO-mCherry, 1:1)

A bolus of 0.4 μ L of mixed virus is injected into the VTA at 46 nL/min. The pipette is held in place for 10 min after the injection and then slowly withdrawn. For real-time place preference (RTPP) behavioral test, the dual-color LED probe implantation is carried out after virus injection and diffusion. The probe was slowly implanted at the side of the target brain areas with an angle of 45° to midline (coordinates 3.64 mm posterior to the bregma, 0.8 mm lateral to midline, and 4.85 mm ventral to the skull surface). The probe is fixed to the skull with Cyanoacrylate adhesive (TONSAN 1454) and dental cement. For dopamine detection with fiber photometric recording, the AAV-hSyn-GRAB_{DA2m} (0.2 μ L) is injected into the NAc (coordinates: 1.18 mm front to the bregma, 0.75 mm lateral to midline, and 3.75 mm ventral to the skull

surface). Optical fiber implantation is carried out after virus injection and diffusion. A piece of optical fiber (FT200UMT, 200 μ m O.D., 0.39 NA, Thorlabs) is fit into an LC-sized ceramic fiber ferrule (Shanghai Fiblaser, China) and implanted over the target brain areas with the tip 0.1 mm above the virus injection sites in the NAc. The ceramic ferrule is fixed to the skull with Cyanoacrylate adhesive (TONSAN 1454) and dental cement. After the virus injection and device implantation, mice are housed individually to prevent potential damage to the probe. All subsequent experiments are performed at least 2 weeks after injection to allow sufficient time for transgene expression and animal recovery.

Immunohistochemistry

The micro-LED probes are implanted into the VTA region and fixed to the skull with Cyanoacrylate adhesive (TONSAN 1454) and dental cement. After 1 day or 3 weeks following the surgery, animals are transcardially perfused with 4% PFA in 0.1 M PBS. Brains are postfixed in 4% PFA and 40- μ m-thick vibratome sections are prepared. After 5-min PBS rinses for 3 times, the sections are blocked in PBST (PBS + 0.3% Triton X-100) with 3% bovine serum albumin for 1 hour, and incubated in rabbit anti-Iba1 (1:1000, Wako, Cat.#019-19741) and rabbit anti-GFAP antibody (1:2000, Sigma, Cat#AB5541) dissolved in PBST at 4 °C for 24 hours. After 5-min washes in PBS for 5 times, the sections are incubated with Alexa Cy3-conjugated goat anti-rabbit antibody (1:500, Jackson ImmunoResearch) and Alexa 488 goat anti-chicken antibody (1:500, Thermo Fisher Scientific) dissolved in PBST for 2 hours at room temperature followed by 5-min washes in PBS for 3 times. At last, the samples are coverslipped in 50% glyceroland slides and photographed with an

automated slide scanner (VS120 Virtual Slide, Olympus) or a confocal scanning microscope (DigitalEclipse A1, Nikon). The numbers of DAPI, stGtACR2 and ChrimsonR stained cells or the numbers of Iba1+ microglia and GFAP+ astrocytes are analyzed with Fiji software (<https://fiji.sc/>) and their percentages in cell population indicated by DAPI are calculated.

In vitro optogenetics evaluation

Electrophysiological properties of the neurons in brain slices are measured with whole-cell patch-clamp recording technique following a previously described procedure.¹² After two weeks of virus expression (ChrimsonR + stGtACR2), mice are deeply anesthetized with pentobarbital (100 mg/kg i.p.) and intracardially perfused with 5 mL ice-cold oxygenated modified artificial cerebrospinal fluid (ACSF) at a rate of 2 mL/min. After perfusion with modified ACSF containing (in mM): 225 sucrose, 119 NaCl, 2.5 KCl, 0.1 CaCl₂, 4.9 MgCl₂, 1.0 NaH₂PO₄, 26.2 NaHCO₃, 1.25 glucose, 3 kynurenic acid, and 1 Na L-ascorbate (Sigma-Aldrich), brains are quickly removed and placed in ice-cold oxygenated ACSF containing (in mM): 110 choline chloride, 2.5 KCl, 0.5 CaCl₂, 7 MgCl₂, 1.3 NaH₂PO₄, 25 NaHCO₃, 20 glucose, 1.3 Na ascorbate, and 0.6 Na pyruvate. Sagittal sections (200 µm thick) are prepared using a Leica VT1200S vibratome. Brain slices are incubated for 1 hour at 34 °C with ACSF saturated with 95% O₂/5% CO₂ and containing (in mM) 125 NaCl, 2.5 KCl, 2 CaCl₂, 1.3 MgCl₂, 1.3 NaH₂PO₄, 25 NaHCO₃, 10 glucose, 1.3 Na ascorbate, and 0.6 Na pyruvate. The internal solution within whole-cell recording pipettes (3–6 MΩ) contains (in mM): 130 K-gluconate, 10 HEPES, 0.6 EGTA, 5 KCl, 3 Na₂ATP, 0.3

Na₃GTP, 4 MgCl₂, and 10 Na₂-phosphocreatine (pH 7.2–7.4). Voltage-clamp recordings are performed using a MultiClamp 700B amplifier (Molecular Devices). Traces are low-pass filtered at 2.6 kHz and digitized at 10 kHz (DigiData 1440, Molecular Devices). The data are acquired and analyzed using Clampfit 10.0 software (Molecular Devices). Red and blue illuminations are provided by a fiber coupled to two laser sources (wavelengths 633 nm and 473 nm).

Real-time place preference test

After 2 weeks of recovery and AAV expression in *DAT-cre* mice, real-time place preference / avoidance (RTPP) tests are performed in a custom-made two-compartment apparatus. The RTPPT apparatus comprises two white plastic chambers (size: 25 × 25 × 30, L × W × H in cm), separated by a half-open opaque plastic walls in the middle (8 × 30, L × H cm). A RTPP test consists of 3 phases: 10-min baseline (pretest), 30-min red LED stimulation, 30-min rest (no stimulation), and 30-min blue LED stimulation. During the pretest phase, individual mice are allowed to freely explore the entire apparatus for 10 min. Animals that exhibit a strong initial preference or avoidance for one of the side chambers (preference index > 60% or < 40%) are excluded from further experiments. During the red LED stimulation phase, animals that entered the left chamber receive red light pulses (20 Hz, 20-ms pulse, 7 mA). During the blue light stimulation phase, animals that entered the left chamber receive constant blue light (continuous, 5 mA). The locomotion of animals is assessed from the video recording data using a custom-developed MATLAB program.

Fiber photometry for dopamine (DA) detection

Details for the fiber photometric system are described previously.¹³ The excitation source is a 488 nm semiconductor laser (Coherent, Inc. OBIS 488 LS, tunable power up to 60 mW). A dichroic mirror with a 452–490 nm reflection band and a 505–800 nm transmission band (Thorlabs, Inc. MD498) is employed for wavelength selection. A multimode fiber (Thorlabs, Inc., 200 μm in diameter and 0.39 in numerical aperture) coupled to an objective lens (Olympus Inc, 10 $\times$, NA 0.3) is used for optical transmission. The fluorescence signals are collected with a photomultiplier tube (PMT) (Hamamatsu, Inc. R3896) after filtering by a GFP bandpass filter (Thorlabs, Inc. MF525-39). An amplifier (C7319, Hamamatsu) is used to convert the current output from the PMT to voltage signals, which are further filtered through a lowpass filter (40 Hz cut-off; Brownlee 440, USA). The fluorescence signals are digitalized at 100 Hz and recorded by a Power 1401 digitizer and a Spike2 software (CED, UK). Photometric data are exported and analyzed with MATLAB. The fluorescence signals ($\Delta F/F_0$) are processed by averaging the baseline signal F_0 over a 4.5-s long control time window, and presented as heatmaps or per-event plots. AUC (area under the curve) of DA signal is used for statistical analysis. AUC is the integral of a curve that describes the variation of DA signals during light stimulation.

References

1. Ding H, Lu L, Shi Z, Wang D, Li L, Li X, *et al.* Microscale optoelectronic infrared-to-visible upconversion devices and their use as injectable light sources. *Proc Natl Acad Sci U S A* 2018, **115**(26): 6632-6637.
2. Li L, Liu C, Su Y, Bai J, Wu J, Han Y, *et al.* Heterogeneous Integration of Microscale GaN Light-Emitting Diodes and Their Electrical, Optical, and Thermal Characteristics on Flexible Substrates. *Adv Mater Technol* 2018, **3**(1): 1700239.
3. Liu C, Zhang Q, Wang D, Zhao G, Cai X, Li L, *et al.* High Performance, Biocompatible Dielectric Thin-Film Optical Filters Integrated with Flexible Substrates and Microscale Optoelectronic Devices. *Adv Opt Mater* 2018, **6**(15): 1800146.
4. Gebhart SC, Lin WC, Mahadevan-Jansen A. In vitro determination of normal and neoplastic human brain tissue optical properties using inverse adding-doubling. *Phys Med Biol* 2006, **51**(8): 2011-2027.
5. Zhao Y, Liu C, Liu Z, Luo W, Li L, Cai X, *et al.* Wirelessly Operated, Implantable Optoelectronic Probes for Optogenetics in Freely Moving Animals. *IEEE Trans Electron Devices* 2019, **66**(1): 785-792.
6. Peirson SN, Brown LA, Potheary CA, Benson LA, Fisk AS. Light and the laboratory mouse. *J Neurosci Methods* 2018, **300**: 26-36.
7. Klapoetke NC, Murata Y, Kim SS, Pulver SR, Birdsey-Benson A, Cho YK, *et al.* Independent optical excitation of distinct neural populations. *Nat Methods* 2014, **11**(3): 338-346.
8. Govorunova EG, Sineshchekov OA, Janz R, Liu X, Spudich JL. Natural light-gated anion channels: A family of microbial rhodopsins for advanced optogenetics. *Science* 2015, **349**(6248): 647.
9. Mahn M, Gibor L, Patil P, Cohen-Kashi Malina K, Oring S, Printz Y, *et al.* High-efficiency optogenetic silencing with soma-targeted anion-conducting channelrhodopsins. *Nat Commun* 2018, **9**(1): 4125.
10. Sun F, Zhou J, Dai B, Qian T, Zeng J, Li X, *et al.* Next-generation GRAB sensors for monitoring dopaminergic activity in vivo. *Nat Methods* 2020, **17**(11): 1156-1166.
11. Wu X, Dong X, Wu Z, Cao H, Niu D, Qu J, *et al.* A novel method for purification of recombinant adenoassociated virus vectors on a large scale. *Chinese Sci Bull* 2001, **46**(6): 485-488.
12. Lu L, Ren Y, Yu T, Liu Z, Wang S, Tan L, *et al.* Control of locomotor speed, arousal, and hippocampal theta rhythms by the nucleus incertus. *Nat Commun* 2020, **11**(1): 1-16.
13. Lu L, Wang R, Luo M. An optical brain-to-brain interface supports rapid information transmission for precise locomotion control. *Sci China Life Sci* 2020, **63**(6): 875-885.