Design of an islet-on-a-chip device reveals glucose-stimulated respiration is substrate limited by glycolytic flux through PKM2.

Romario Regeenes^{1,2}, Yufeng Wang^{1,2}, Anthony Piro^{1,3}, Aaron Au², Christopher Yip², Michael B. Wheeler^{1,3} and Jonathan V. Rocheleau^{1,2,3}

¹Advanced Diagnostics, Toronto General Hospital Research Institute, Toronto, ON

²Institute of Biomedical Engineering, University of Toronto, Toronto, ON

³Departments of Medicine and Physiology, University of Toronto, ON

1. Abstract

Pancreatic islets respond metabolically to glucose by closing K_{ATP} channels resulting in Ca^{2+} -influx and insulin secretion. Previous work showed that the importance of glycolytic flux in triggering insulin secretion. However, it is unclear whether the triggered ('first phase') secretion is further amplified by Ca^{2+} -stimulation of mitochondrial NADH production and/or oxidative phosphorylation (OxPhos). To tease apart the metabolism of first phase glucose-stimulated insulin secretion, we designed an islet-on-a-chip microfluidic device to simultaneously measure O_2 -consumption rate (OCR) and Ca^{2+} -activity of individual islets with high temporal resolution. We used finite element analysis to optimize the placement of sensor in optically clear microwells on a thin glass coverslip. The microfluidic channels were subsequently fabricated using O_2 -impermeable plastic to limit outside-in diffusion and push islets in the device against the microsensor. We validated our device using living mouse islets and well-established modulators of respiration. By inhibiting glycolysis and mitochondrial pyruvate transport, we showed islet OxPhos is limited by substrate-NADH rather than ADP in low and high glucose. We subsequently imaged glucose-stimulated OCR and Ca^{2+} -influx simultaneously to reveal a biphasic respiratory response that is determined by glycolytic flux through pyruvate kinase (PKM2) and independent of Ca^{2+} . Ultimately, we validated our device and successfully used it to show islet OxPhos is determined by glycolytic flux and not amplified by Ca^{2+} -activity.

Significance

Loss of the first phase glucose-stimulated insulin secretion is one of the earliest signs of Type 2 diabetes (T2D). Revealing the mechanisms that trigger and amplify first phase insulin secretion from pancreatic islets could lead to strategies to mitigate the onset of T2D. In this study, we designed an islet-on-a-chip microfluidic device to measure O_2 -consumption rate and Ca^{2+} activity of individual islets dynamically through first phase secretion (< 5 min). Our data show first phase islet respiration is limited by glycolytic production of pyruvate independent of Ca^{2+} -influx.

Author contributions

R.R., Y.W. and J.V.R. designed research; R.R. and Y.W. performed research; A.P. and M.B.W. provided tissue; A.A. and C.Y. provided technical assistance, R.R., Y.W., and J.V.R. wrote the paper.

Keywords

Islet-on-a-chip, microfluidics, pancreatic islets, widefield imaging, oxygen sensor, oxygen consumption rate, seahorse assay, mouse islets, Stern-Volmer.

2. Introduction

Glucose-stimulated insulin secretion (GSIS) occurs from pancreatic islets in a large first phase burst followed by smaller second phase oscillations [1] [2]. Critically, first phase insulin secretion is lost at the earliest signs of type 2 diabetes (T2D) [3] suggesting the metabolism behind the stepped response to glucose is mechanistically linked to the disease.

GSIS is triggered from islet beta-cells by a rise in ATP/ADP ratio that closes K_{ATP} channels resulting in membrane depolarization and Ca^{2+} -influx (termed ‘ K_{ATP} -dependent secretion’) [4] [5]. Early work revealed a central role for glycolysis in triggering insulin secretion [6] [7]. More recent work showed that inhibition of pyruvate kinase-M2 (PKM2) (i.e., the last step of glycolysis) abolishes insulin secretion suggesting glycolysis-derived pyruvate is also a prerequisite to trigger insulin secretion [8]. Insulin secretion is further amplified by glucose-metabolism independent of K_{ATP} channel activity (termed ‘ K_{ATP} -independent amplification’) effectively doubling insulin secretion [9] [10]. The amplification of second phase insulin secretion has been classically linked to mitochondrial anaplerosis and associated efflux of metabolic intermediates [11] as well as to Ca^{2+} -induced mitochondrial metabolism and respiration [12] [13]. Whether these same mechanisms also amplify first phase insulin secretion is less defined. For example, activation of PKM2 was recently shown to enhance second phase insulin secretion from islets without impacting first phase insulin secretion [14].

It is well recognized that mitochondrial metabolism and OxPhos are key determining factors of GSIS that are significantly impaired in T2D [15] [16]. O_2 -consumption rate (OCR) has been suggested to be the most informative parameter of mitochondrial function and cellular metabolism as it reveals electron transport chain (ETC) function and substrate supply as well as others [17]. Thus, measuring OCR could help reveal the metabolism of first phase insulin secretion. The lack of studies on K_{ATP} -independent amplification of first phase insulin secretion could be partly due to technological limitations of simultaneous measuring OCR (related to OxPhos) and Ca^{2+} (insulin secretion triggering and mitochondrial activation). Specifically, such measurements need high temporal resolution, since glucose-stimulated metabolism related to first phase insulin secretion occurs within 5 minutes. For example, Patterson et al. used two-photon NAD(P)H imaging to show a two-step rise in glucose-stimulated islet NAD(P)H response; however, due to the short time-lag between the steps (<20 s), they suggested that the second step was not triggered by Ca^{2+} -influx but by the regulated entry of pyruvate into mitochondria [18]. In addition, these simultaneous measurements should ideally be performed on single islets to avoid blurring of the temporal responses due to islet-to-islet variability [19].

Our goal was to simultaneously measure OCR and Ca^{2+} activity from individual islets with sufficient temporal resolution to distinguish their first phase glucose responses. For example,

sensing glucose-stimulated Ca^{2+} -influx prior to a rise in OCR would be consistent with Ca^{2+} -stimulated mitochondrial metabolism. Different methods are commercially available to measure islet respiration. However, these methods often do not provide optical access to simultaneously measure Ca^{2+} activity, and when they do, they lack the spatial and temporal resolution to measure individual islet responses [19] [20]. To address these limitations, we aimed to place an optical O_2 sensor inside microfluidic devices previously designed for quantitative live cell imaging of pancreatic islets [21] [22] [23]. These devices hold the islets directly against a thin glass coverslip, which here provides an ideal location for the placement of optical sensors. By placing the O_2 sensor in optically clear yet O_2 -impermeable microwells, we achieved the sensitivity to measure OCR from individual pancreatic islets and the ability to simultaneously image Ca^{2+} -influx. Finally, we used this device to show first phase respiration is NADH-substrate limited by glycolytic flux and independent of Ca^{2+} activity.

3. Experimental

Pancreatic islet isolation and culture

Animal procedures were approved by the Animal Care Committee of the University Health Network, Toronto, Ontario, Canada in accordance with the policies and guidelines of the Canadian Council on Animal Care (Animal Use Protocol #1531). Pancreatic islets were isolated from C57BL/6 male mice (12-14 weeks of age). A solution of collagenase type-V (0.9 mg/mL) in RPMI-1640 media supplemented with BSA (0.5mg/mL) was prepared for islet isolation. Mice were anesthetized with isoflurane and sacrificed via cervical dislocation, and perfusion of the pancreas was performed via injection of 3 mL collagenase type-V solution in the common bile duct with a 30G needle, inserted in the direction of liver to pancreas (common bile duct sutured at small intestine). Following successful perfusion, the pancreas was removed and placed in a 50 mL conical tube containing an additional 3 mL of the collagenase type-V solution. Pancreases in conical tubes were incubated in a water bath at 37°C for 11-12 minutes and mechanically disrupted via light shaking until exocrine pancreas was fully digested and the solution was homogenous with digested tissue. Finally, digestion was inhibited via addition of 45 mL islet culture media (RPMI medium 1640 (Sigma-Aldrich) supplemented with 11 mM glucose, 10% FBS, 5 U/ml penicillin-streptomycin, and 20 mM HEPES). Under a dissection microscope, islets were separated from digested exocrine pancreas material and incubated in islet culture media in a humidified incubator at 37°C and under 5% CO_2 overnight.

Microfluidic device fabrication

The custom microfluidic device comprised a ‘top piece’ and glass coverslip with O_2 -sensor microwells. The top piece with 80 μm tall channels and 60 μm tall dam-wall was fabricated using 3 mm thick PMMA and a Nomad 883 Pro to mill the channel. The channel dimensions were modelled in AutoCAD 2021, and then imported into Carbide Create V5 to create the GCode necessary to run the CNC mill. A 1/16” endmill was used to mill the device followed by a 1/8” μm endmill for the inlet and outlet ports. The top piece was washed in a 70% ethanol solution and left to dry before attachment to the microwell coverslip. The O_2 -sensor microwells were fabricated onto no. 1 glass coverslips using SU-8 photoresist (Micro-Chem) exposed through a transparent plastic film. The 10 μm tall microwells in the final design were 10 μm in

diameter and 20 μm in pitch size. The O_2 -sensor was prepared by mixing PDMS (10:1 ratio) and PtTFPP (0.1 mg/1 mL in toluene) at a volume ratio of 10:1 [24]. 100 μL of this mixture was left to dry for 30 min prior to being squeezed into the microwells. The slide was subsequently left to dry in the dark for 24 hours at room temperature. Prior to use, the microfluidic device and glass coverslip with O_2 -sensor microwells was carefully assembled using epoxy-based glue and clamped together for 1 hour.

Microfluidic fluid flow modelling/Oxygen Sensor

The device was constructed in two-dimensions using AutoCAD 2020 prior to being extruded to three-dimensions in COMSOL Multiphysics 5.4 (COMSOL). Pancreatic islets were modelled as 120 μm radii cylinders with the circumferential edge filleted and modeled to consume O_2 under a maximum rate of $0.034 \text{ mol}\cdot\text{s}^{-1}\cdot\text{m}^{-3}$ [25]. To determine O_2 -concentration underneath the islet a *cut line surface normal* was used with the same coordinates for all the simulations. Inlet pressure for all the simulations was set to 0 kPa while the exit flow rate was set to $200 \mu\text{L h}^{-1}$, unless otherwise stated. The simulated models had meshes with 50735-174199 domain elements, 11221-14864 boundary elements, and 818-11184 edge elements. All diffusion coefficients were used at 37°C for our numerical simulations.

Microscope Setup

Luminescence and fluorescence imaging was done using the $40\times/0.75$ NA air objective of an inverted widefield RAMM fluorescence microscope (ASI) equipped with 3 LEDs (460, 535, and 560 nm), stage top incubator (Okolab) and Iris 15 sCMOS camera (Photometrics). The microscope was controlled using MicroManager [26]. PtTFPP luminescence was excited using two to three 500 ms bursts of the 535 nm LED and collected through the 590-660 nm band-pass emission filter. Cal-520 fluorescence was excited by a 300 ms of the 465 nm LED attenuated to 32% using a neutral density filter and collected using a 495-525 nm band-pass emission filter. Images consisted of 1264×740 pixels sized at $0.425 \mu\text{m}/\text{pixel}$ after 4×4 pixel binning.

Sensor Calibration

The PMMA device with O_2 -sensing microwells was subjected to flowing solutions of 20% (air saturated ddH₂O) and 0% (10% sodium sulfite) O_2 . The PtTFPP was imaged to measure individual microwell mean intensity. The mean intensities at 20% and 0% O_2 were used to calculate the Stern-Volmer constant, K_{SV} . The mean luminescence intensity *versus* O_2 concentration was modelled as a linear regression using the Stern-Volmer equation (**Eq. 1**):

$$\frac{I_0}{I} = 1 + K_{SV} * [\text{O}_2]$$

where I_0 and I are the luminescence intensity in the absence and presence of molecular oxygen, $[\text{O}_2]$.

Glucose stimulated oxygen imaging

Freshly isolated islets were pre-incubated for 1 hr in BmHH imaging buffer (125 mM NaCl, 5.7 mM KCl, 0.42 mM CaCl_2 , 0.38 mM MgCl_2 , 10 mM HEPES, and 0.1% BSA; pH 7.4) with 2 mM

glucose at 37°C. Islets (2 to 5 at a time) were subsequently loaded into the stage-mounted microfluidic device at 37°C and maintained in 200 μ L/ml of flowing media using pull from a syringe pump. The islets were exposed to various treatments as indicated by exchange of the media in an on-chip well prior to imaging the PtTFPP luminescence.

Simultaneous OCR- and Ca^{2+} -imaging

To simultaneously image OCR and Ca^{2+} -activity, mouse islets were incubated for 1 hr at 37°C in BmHH with 4 μ M Cal-520 (Life Technologies), 2 mM glucose, and 1 mM N-acetylcysteine (NAC) [27]. The islets were subsequently loaded into the stage-mounted microfluidic device at 37°C and pre-incubated in 200 μ L/hr of flowing BMHH (2 mM glucose, 1 mM NAC) for 10 minutes prior to imaging. A custom macro was created to collect PtTFPP and Cal-520 images using 535 and 460 nm excitation, and at two different z-planes to account for the difference in imaging plane between the microwells and Cal-520 loaded islets. The islets were then variably treated as indicated by replacing media in the on-chip well while imaging.

Image analysis

PtTFPP luminescence and Cal-520 images were analyzed using ImageJ. Regions of interest (ROI) were manually drawn around each microwell underneath the islet as well as microwells loaded with only PDMS for background correction. After selection, the mean intensity for each microwell was recorded after background correction. The O_2 -concentration in each microwell was subsequently determined using the K_{SV} of the calibrated sensor and Eq. 1. Lastly, OCR was calculated as the change in O_2 -concentration multiplied by the volumetric flow rate across the islet, approximating islet volume to be a cylinder or sphere. The dissolved O_2 -concentration at 37 °C was calculated to be 207.3 μ M. Ca^{2+} images were analyzed by normalizing the fluorescence intensity of randomly selected ROIs by the mean intensity during 2 mM glucose treatment.

4. Results

Design and optimization of an islet-on-a-chip device with modular O_2 -sensors

We recognized that the coverslip of our islet-on-chip microfluidic devices would provide proximity to the tissue and optimal optics to place optical O_2 -sensors [21] [22] [23]. To explore placing the sensor in O_2 -permeable substrate underneath islets, we modelled O_2 -diffusion and OCR by islets in a dam-wall microfluidic device (**Fig. 1**). We chose an islet-on-a-chip device that traps islets in flowing media using a drop in channel height (e.g., 80 to 20 μ m) (**Fig. 1A**). The drawing (not to scale) shows a device fabricated out of O_2 -impermeable PMMA (grey) with PtTFPP sensor embedded inside O_2 -permeable PDMS (yellow) coating the glass coverslip. We normally fabricate microfluidic channels out of PDMS but switched to PMMA to avoid swamping/masking islet O_2 -consumption. We modelled both an empty device (left) and islet-filled-device (right) using finite element analysis to determine O_2 concentration in the sensor layer (**Fig. 1B**) including along the centerline plotted from left to right with the flow (**Fig. 1C**). These data show very little difference in the O_2 underneath islets (right image, red and blue circles), which is consistent with diffusion along the PDMS substrate dampening the sensor response to islet respiration. To maximize sensor response, we subsequently modelled placing

the sensor inside microwells (**Fig. 1D-H**). The drawing (not to scale) depicts microwells formed using O₂-impermeable SU-8 (black) to limit diffusion between the microwells (yellow) (**Fig. 1D**). To determine the impact of microwell size, we modelled the OCR of islets placed against microwells of varying widths (10, 50 and 100 μm) (**Fig. 1E-F**). These data show progressively larger responses to islet OCR with smaller microwell width. Of all well sizes examined, the 10 μm wide wells with 20 μm pitch were the most sensitive to the islet OCR. To determine the optimal dimensions, we also simulated 10 μm wide microwells with varying depths (10, 30 and 60 μm) (**Fig. 1G-H**). These data show no difference in the response based on the depth of the microwell. Thus, we chose to fabricate microwells that are 10 μm wide with 20 μm pitch due to sensor response and chose the shortest microwell height (10 μm) to minimally affect optical clarity.

Determining factors that affect sensitivity of O₂-sensors

To evaluate potential factors that could affect O₂ sensitivity, we modelled varying flow rates, islet proximity and islet O₂ consumption rates (**Fig. 2**). Flow rate is a key factor to consider with the PMMA fabrication since it is the sole source of replenishing O₂ inside the channel. Modelling various flow rates through the device showed only a moderate impact on O₂-sensitivity (**Fig. 2A**). These data suggest slower flow rates marginally enhance O₂-sensor responses, but the difference is small suggesting sensing would be stable even during fluctuations in flow. Our dam-wall design can optimally trap two different sized islets; ~20 μm diameter under the dam-wall, and >80 μm diameter in the channel. Although islets sink, we were concerned that this arrangement could result in some islets not being uniformly pressed against the microwells. To evaluate the effect of proximity, we simulated islets at various heights relative to the microwells (**Fig. 2B-C**). These data show a less dynamic sensor response with increasing distance between the islet and microwell suggesting close tissue proximity is essential for sensitivity. We also considered different OCR values since islets are normally studied in low and high glucose conditions. We modelled 2- and 20-mM glucose-stimulated OCR by changing the inward flux values of the islet walls. The goal was to determine if there was a visible difference between both treatments (**Fig. 2D**). It revealed a 12.8% difference which could be empirically measured. These data suggest relative OCR induced by 2- and 20-mM glucose will be detected by the sensor (**Fig. 2E**).

On-chip characterization of the O₂ sensor

O₂ sensors are generally calibrated prior to use so that luminescence intensities can be related back to O₂ concentrations based on the Stern-Volmer relationship [28]. To determine our sensor dynamic range, we flowed double distilled water (ddH₂O) and an O₂-scavenger (10% sodium sulfite) to obtain luminescence intensities for 100% air saturated and 0% O₂, respectively (**Fig. 3A**). These images show an ~25-fold change in intensity of the microwells in the presence of an O₂-scavenger. Based on linear regression, we calculate a K_{SV} of $113.3 \pm 0.9196 \text{ L} \cdot \text{mmol}^{-1}$, which is consistent with literature values for this type of sensor, albeit larger than existing sensors (**Fig. 3B**) [29]. To determine the temporal resolution of the O₂-sensor, we measured the response times of the microwells upon switching from ddH₂O to 10% sodium sulfite and vice versa (**Fig. 3C and D**). Response time, t_{95} , is defined as the time required to reach 95% of the

luminescence maximum or minimum. These data show a sensor t_{95} switching from 0% to 100% air and vice versa of 12.4 ± 0.7 s (**Fig. 3C**), and 4.2 ± 0.2 s (**Fig. 3D**). These times reflect a 0.21 mM change in dissolved O_2 suggesting we should expect much faster response times with the nanomolar changes induced by living islets.

Imaging OCR in individual pancreatic islets

We subsequently loaded living mouse pancreatic islets in our device to measure the dynamics of OCR (**Fig. 4**). Once loaded, islets sitting on the microsensor could be imaged by an inverted widefield epifluorescence microscope consistent with the optical transparency of the device (**Fig. 4A, top**). Islet bioenergetics are commonly measured in batch (~70 islets) using commercially available devices (e.g., Seahorse XF analyzer) and a series of mitochondrial respiration treatments including: (i) 2 mM glucose to measure basal OCR, (ii) 20 mM glucose to measure glucose-stimulated OCR, (iii) oligomycin to measure ATP-linked respiration, (iv) FCCP to measure maximal respiratory capacity, and (v) rotenone + antimycin to define non-mitochondrial respiration. To validate our device, we used these same treatments on individual islets (**Fig. 4A, bottom**). The microsensors underneath the islet showed clear changes in luminescence intensities reflecting their respective measurements of O_2 . These data show that the O_2 sensor can measure pharmacologically induced changes in OCR in individual pancreatic islets. However, unlike commercial assays, we are not limited to a narrow size of islets [19] (**Fig. 4B**). To determine the effect of variable islet size on the OCR, we plotted the basal OCR relative to islet volume estimated from the widefield image (**Fig. 4C**). These data show a linear relationship between islet volume and OCR ($R^2 = 0.79$) suggesting the responses should be normalized to islet volume. We therefore normalized the individual islet responses to the islet volumes prior to plotting the responses to the respiratory inhibitors (**Fig. 4D**). These data show a 2-fold increase from basal to glucose-stimulated OCR and maximal respiration consistent with responses from batch analysis using commercially available instruments [19] (**Fig. 4E**). These data also show proton leak is approximately 68% of basal OCR consistent with previous studies [30]. Overall, these data validate that our device can measure the OCR of individual pancreatic islets.

Temporal imaging of the glucose-stimulated OCR and Ca^{2+} responses

Glucose-stimulated OCR has previously been shown to rise prior to Ca^{2+} -influx suggesting that Ca^{2+} does not play an immediate role [31]. To determine the temporal contribution of Ca^{2+} to glucose-stimulated metabolism, we loaded mouse pancreatic islets with Cal-520 to simultaneously image OCR and Ca^{2+} activity in response to glucose (**Fig. 5A and B**). The Ca^{2+} traces upon glucose stimulation show a characteristic 'phase 0' drop upon glucose stimulation due to activation of Ca^{2+} -ATPase ER pump efflux and 'phase 1' rise ~3 min after stimulation due to voltage-gated Ca^{2+} channel influx (**Fig. 5B**). Notably, the glucose-stimulated OCR traces were biphasic showing a rise simultaneously with the drop in Phase 0 Ca^{2+} and a second rise near or before phase 1 Ca^{2+} -influx. To determine the magnitude of the OCR responses, we calculated the normalized OCR area under the curve (AUC) for phase 0 and phase 1 as defined by the Ca^{2+} traces (**Fig. 5C and D**). In other words, we used the changes in Ca^{2+} -activity to define the 2mM glucose, phase 0, and phase 1 regions for each OCR trace (**Fig. 5C**). Using this criterion, the OCR increased significantly and sequentially from basal to phase 0 and to phase 1 (**Fig. 5D**). To

further explore the biphasic OCR response, we focused on the relative changes in OCR within the first two minutes of phase 0 and phase 1 by setting “Time 0” for each curve based on the inflection points of the Ca^{2+} traces (**Fig. 5E and F**). Due to variability in the timing of the responses, this analysis could only be done due to our ability to measure traces from individual islets. The first rise in OCR occurs simultaneously with the phase 0 drop in Ca^{2+} plateauing within 1 min and prior to phase 1 Ca^{2+} -influx (**Fig. 5E**). These data are consistent with an increase in OxPhos due to the metabolic flux (supply) that is independent of mitochondrial stimulation by Ca^{2+} . The initial plateau was followed by a second rise in OCR that preceded Ca^{2+} -influx (**Fig. 5F**). Further analysis of the OCR curve using linear regression of the eight time points starting from time = 0 min fit to an x-intercept of -28 ± 20 s. In other words, the second OCR response preceded Ca^{2+} -influx by ~ 28 s. To further confirm the biphasic OCR response to glucose, we determined the slope of OCR responses for each trace before and after the -30 s inflection point (**Fig. 5G**). These data show significantly different slopes once again are consistent with a further inflection in OxPhos. Overall, these data illustrate that islet respiratory response to glucose is biphasic with the step in the response occurring before and independently of Ca^{2+} -influx.

PKM2 is responsible for the biphasic rise in OCR

The rate of OxPhos depends on the metabolic supply of NADH (‘substrate limited’) or when NADH is in excess by the availability of ADP (‘ADP limited’) (**Fig. 6A**). We postulated that the biphasic glucose-stimulated OCR response reflects islets in a substrate limited state with transitions in OCR reflecting greater production of NADH by glycolysis and the TCA cycle. To first confirm islet OCR is substrate limited at low glucose, we measured the OCR response of islets in low glucose to inhibition of glycolytic flux using 2-deoxyglucose (2DG) and in response to blocking entry of pyruvate into mitochondria using UK5099 (**Fig. 6B**). These data show 2DG did not induce a change in OCR consistent with low glycolytic flux in glucose below the K_m for glucokinase. However, UK5099 induced a small but significant decrease in OCR indicating that basal respiration is dependent on pyruvate entry. These data highlight the importance of pyruvate entry into the TCA cycle in stimulating OxPhos (3 NADH/pyruvate) and potentially reflect flux of amino acid oxidation (e.g., alanine, serine). To confirm that glucose-stimulated OCR is due to glycolytic flux and substrate limited, we measured the OCR response of glucose-stimulated islets to 2DG (**Fig. 6C and D**). These data show that glucose-stimulated OCR dropped to basal (low glucose) levels by inhibition of glycolytic flux again suggesting the respiratory transition of islets ultimately depends on substrate supply. Due to the varied 2DG and UK5099 responses in low glucose, we speculated that the biphasic nature of the glucose-stimulated OCR reflects downstream regulation of PKM2, which is allosterically regulated by the glycolytic intermediate fructose 1,6-bisphosphate (FBP) [32]. To investigate the role of PKM2 in setting basal OxPhos, we measured the OCR of islets treated with PKM2 activator TEPP-46 for 1 hr (**Fig. 6E**). These data show a small but significant increase in basal OCR induced by PKM2 activation again consistent with OCR being substrate limited (in this case by the supply of pyruvate from glycolysis). To further investigate the role of PKM2 activity in the glucose-stimulated response, we measured the glucose-stimulated OCR response of TEPP-46 treated islets (**Fig. 6F**). The higher basal OCR in the presence of TEPP-46 is followed by a sharp decrease induced by glucose. This rapid reduction in OCR is consistent with the rapid entry of pyruvate due to fully active PKM2 leading to lower NADH induced by lactate dehydrogenase activity, conversion to

oxaloacetate/malate through pyruvate carboxylase activity, and/or depolarization of mitochondrial membrane potential. Notably, the OCR steadily increases out to 15 min of stimulation with no sign of a biphasic response. In other words, the biphasic glucose-stimulated response observed in untreated islets is replaced by a single faster rise. To determine the role of Ca^{2+} -influx in this response, we simultaneously imaged the glucose-stimulated OCR and Ca^{2+} responses of TEPP-46 treated islets (**Fig. 6G**). These data again show a sharp drop in OCR that coincides with the beginning of phase 0 Ca^{2+} -efflux followed by a linear rise in OCR that noticeably precedes phase 1 Ca^{2+} -influx. Normalizing the curves from several islets relative to glucose arrival showed the linear increase in OCR precedes Ca^{2+} -influx by ~3.2 min (**Fig. 6H**). Collectively, these results confirm islet OCR is substrate-limited by glycolytic flux and suggest that PKM2 plays a critical role in regulating glucose-stimulated OCR.

5. Discussion

In this study, we redesigned a dam wall microfluidic device to hold pancreatic islets directly against O_2 -microsensors to simultaneously image OCR and Ca^{2+} activity. We digitized the sensor in microwells to gain spatial resolution sufficient to measure the responses of individual islets. Our measurements of basal OCR showed that islet responses should be normalized to islet size to reduce variability. Our measurements with respiratory chain inhibitors further validated the method by showing similar responses when comparing bulk assays to the pooled individual islet responses. The optical clarity and fabrication of the microwells using a thin glass coverslip allowed us to image glucose-stimulated OCR and Ca^{2+} -activity. These data revealed a biphasic respiratory response to glucose with an inflection point determined by PKM2 activity and independent of Ca^{2+} -influx. Overall, our data are consistent with first phase glucose-stimulated OxPhos being NADH-substrate limited by glycolytic flux.

Our microsensors are sensitive to nanomolar changes in O_2 , show fast response times (<15 s) and are easy to calibrate. To achieve these characteristics, we embedded optically responsive sensor (PtTFPP) within an O_2 permeable substrate (PDMS) based on previous work showing this combination had high photostability, large stokes shift and dynamic response [33] [24]. Conventional optical O_2 sensors for microfluidic devices are generally planar and open to the environment. Our initial modelling suggested that O_2 diffusion inside the planar sensor masks OCR responses of individual islets. Thus, we placed the PtTFPP/PDMS sensor into micron-sized wells surrounded by O_2 impermeable epoxy. We improved the sensitivity by reducing the microwell width to less than the footprint of islets (~50 μm in diameter), which highlights the importance of using microchannels to push the tissue against the sensor. We subsequently used PMMA to fabricate the channels (i.e., the top of the device) instead of more conventionally used and O_2 -permeable PDMS. Fabricating the device using air saturated PDMS abolished the sensor response during early trials consistent with O_2 diffusion from the air saturated polymer swamping the sensor response.

Static OCR responses using our device were consistent with previous work once we normalized our data against individual islet volume [19]. We showed that islet OCR is highly uncoupled from ATP production using oligomycin consistent with other bulk islet studies (**Fig. 4E**) [30] [19]. A unique feature of our microsensors is that they are fabricated using a thin glass coverslip

and are optically clear, both of which allow simultaneous imaging of responses. We used 10 μm thick SU-8 photoresist for rapid prototyping in the present study. This diminished the optical throughput and image quality somewhat due to SU-8 having a slightly larger refractive index than glass (~ 1.6 vs. 1.5) [34]. Future devices will be fabricated using a 3D glass printer to achieve optimal optics and thus improve dual color imaging [35]. We used this prototype to measure the OCR responses of individual islets while simultaneously imaging Ca^{2+} -activity of the individual islet. Our microsensor shows high temporal resolution due to the digitization of the sensor. To dissect mechanisms of glycolysis and OxPhos, many labs use commercial instruments that are limited to a bulk measurement every 2-5 minutes, thereby masking islet dynamics, and prohibiting multiparametric imaging. Consequently, failing to reveal early timepoints that may reflect respiratory control by NADH substrate supply prior to ATP demand.

Loss of first phase insulin secretion is a hallmark of the early stages of T2D suggesting metabolic flux of the first phase response is central to the disease [3]. Therefore, the temporal dynamics provided by our device could lead to unique insight into the disease. We showed OxPhos is substrate-limited throughout first phase glucose-stimulated insulin secretion (**Fig. 7**). In low glucose (i.e., below the K_m of Glucokinase), respiration was unchanged by 2DG and diminished by inhibiting mitochondrial pyruvate transport. These data suggest OxPhos in low glucose is NADH- rather than ADP-limited and set by a slow steady-state entry of pyruvate into mitochondria. Glucose stimulation resulted in a biphasic increase in respiration with the first response plateauing within 1 min of glucose-stimulation and a second steady rise in respiration initiated ~ 28 s prior to Ca^{2+} -influx (**Fig. 7A**). There was no obvious change in the OCR slope during or even shortly after Ca^{2+} -influx suggesting no change in NADH supply and/or OxPhos. We postulate that the initial plateau is due to glycolytic NADH to supply OxPhos (**Fig. 7A, lower left panel**). If islets were in an ADP-limited state, we would expect glycolysis (which produces 2 net ATP / glucose) to decrease OxPhos. Furthermore, we postulate that the plateau represents maximal flux of glycolysis through ‘inactive’ PKM2. The subsequent linear rise in OCR is due to allosteric activation of PKM2 (**Fig. 7B, lower right panel**). This conclusion is based on TEPP-46-induced rise in basal OxPhos and the early linear rise in glucose-stimulated OCR that was fully separated from Ca^{2+} influx. PKM2 is heavily expressed in pancreatic islets [36] and tightly regulated by numerous posttranslational mechanisms including phosphorylation and acetylation [37] and by allosteric activation by FBP [32]. Thus, PKM2 is expected to transition from being inactive (i.e., low pyruvate flux) in low glucose to being active (i.e., high pyruvate flux) in high glucose [32]. The immediate glucose-stimulated dip in OCR in the presence of TEPP-46 could be due to pyruvate flux into mitochondria resulting in depolarization of mitochondrial membrane potential [38] and/or flux of pyruvate into mitochondrial metabolism through PC [39]. Reaching maximal glycolytic NADH is needed to offset this pyruvate-stimulated dip in mitochondrial respiration. Alternatively, ATP produced by PKM2 could also decrease OxPhos as shown to regulate second phase secretion [14]; however, the speed of the dip is inconsistent with control by cytoplasmic ATP generation and our evidence suggests that NADH supply determines OxPhos just prior to stimulation. Consistent with a two-step glycolytic flux model, Patterson et al. previously showed a stepped rise in glucose-stimulated islet NAD(P)H response that due to the short time-lag between the steps (<20 s) suggested the stimulus was not triggered by Ca^{2+} -influx but rather stimulated by entry of pyruvate into the

TCA [18]. Overall, our data are consistent with a two-step response to glucose based on allosteric activation of PKM2 and a coincident increase in pyruvate flux into mitochondrial metabolism.

Two-step glycolytic models have been used to explain insulin secretion based on ADP-limited OxPhos [14] and Ca^{2+} stimulated mitochondrial NADH metabolism and OxPhos [40] [41]. In contrast, our model does not incorporate these mechanisms to explain first phase OxPhos. First, ADP-limited OxPhos is more relevant to second phase secretion. Second, we saw no evidence of Ca^{2+} -induced OCR. This is perhaps not surprising since entry through the tightly regulated mitochondrial Ca^{2+} uniporter (MCU) would be delayed relative to cytoplasmic Ca^{2+} -influx. This again suggests that mitochondrial Ca^{2+} activity would be more relevant in second phase insulin secretion [42]. Overall, our data show the importance of two-step glycolytic flux through PKM2 in first phase insulin secretion and highlights that not all mechanisms amplifying second phase insulin secretion are relevant to first phase secretion.

Our device opens the door to many more applications. First, we are interested in increasing the capability of multiparametric imaging with OCR. We used the optical clarity of the O_2 -sensor to image Ca^{2+} activity but could also easily work with other spectrally resolved sensors to simultaneously measure ATP, NADH, and glucose [43]. We are particularly interested in simultaneously imaging ATP using Perceval-HR to determine the extent to which ATP is coupled to OCR. Second, this device will allow us to explore respiratory heterogeneity of pancreatic islets. Beta-cells are genetically and functionally heterogeneous [44]; however, whether this impacts islet function is less clear. Preliminary evidence suggests some variability in glucose stimulated OxPhos in human and mice islets [19]. Our device offers the ability to explore this variability and further link it to Ca^{2+} activity. Third, our modular sensor design can be easily used with alternative islet-on-chip and with precision cut tissue slices [45]. We anticipate the O_2 -sensor will translate well to tissue slices if the tissue is pressed against the microwells. Finally, this device could be used clinically to assess tissue prior to transplantation. OCR has previously been shown to be a powerful predictor of insulin independence after transplantation [46]. This device could allow rapid screening of first phase OCR-response prior to islet transplantation for the treatment of type 1 diabetes [47]. Thus, we anticipate this study could inspire wide-ranging experimental and clinical applications.

In summary, our results show that pancreatic islets are in a substrate (NADH) limited state throughout first phase insulin secretion. We identified a role for PKM2 in controlling a biphasic glucose-stimulated OCR response. Activation of PKM2 further uncoupled the glucose-stimulated second rise from Ca^{2+} -influx suggesting the glucose-stimulated OCR response requires glycolytic NADH and pyruvate entry prior to Ca^{2+} -influx.

ACKNOWLEDGMENTS

R.R. was supported by an NSERC PGSD. This research was supported by grants from NSERC (RGPIN-2016-371705 and RTI-2018-00846) and CIHR (162330) to J.V.R. and stipend support to R.R. from NSERC (PGSD-535059-2019).

REFERENCES

- [1] T. Nilsson, V. Schultz, P.-O. Berggren, B. E. Corkey and K. Tornheim, "Temporal patterns of changes in ATP/ADP ratio, glucose 6-phosphate and cytoplasmic free Ca²⁺ in glucose-stimulated pancreatic β -cells," *Biochemical Journal*, vol. 314, pp. 91-94, 1996.
- [2] J.-C. Henquin, M. Nenquin, P. Stiernet and B. Ahren, "In Vivo and In Vitro Glucose-Induced Biphasic Insulin Secretion in the Mouse," *Diabetes*, vol. 55, no. 2, pp. 441-451, 2006.
- [3] G. C. Weir and S. Bonner-Weir, "Reduced glucose-induced first-phase insulin release is a danger signal that predicts diabetes," *The Journal of Clinical Investigation*, vol. 131, no. 12, pp. 1-3, 2021.
- [4] Z. Wang, T. Gurlo, A. V. Matveyenko, D. Elashoff, P. Wang, M. Rosenberger, J. A. Junge, R. C. Stevens, K. L. White, S. E. Fraser and P. C. Butler, "Live-cell imaging of glucose-induced metabolic coupling of β and α cell metabolism in health and type 2 diabetes," *Communications Biology*, vol. 4, no. 594, pp. 1-11, 2021.
- [5] R. Bertrama, M. G. Pedersen, D. S. Lucianic and A. Sherman, "A simplified model for mitochondrial ATP production," *Journal of Theoretical Biology*, vol. 242, pp. 575-586, 2006.
- [6] I. D. Dukes, M. S. McIntyre, R. J. Mertz, L. H. Philipson, M. W. Roe, B. Spencer and J. F. Worley, "Dependence on NADH Produced during Glycolysis for Beta-Cell Glucose Signaling," *The Journal of Biological Chemistry*, vol. 269, no. 15, pp. 10979-10982, 1994.
- [7] K. Eto, Y. Tsubamoto, Y. Terauchi, T. Sugiyama, T. Kishimoto, N. Takahashi, N. Yamauchi, N. Kubota, S. Murayama, T. Aizawa, Y. Akanuma, S. Aizawa, H. Kasai, Y. Yazaki and T. Kadowaki, "Role of NADH shuttle system in glucose-induced activation of mitochondrial metabolism and insulin secretion," *Science*, vol. 283, no. 5404, pp. 981-5, 1999.
- [8] D. Nakatsu, Y. Horiuchi, F. Kano, Y. Noguchi, T. Sugawara, I. Takamoto, N. Kubota, T. Kadowaki and M. Murata, "l-cysteine reversibly inhibits glucose-induced biphasic insulin secretion and ATP production by inactivating PKM2," *PNAS*, vol. Early Edition, pp. 1-10, 2015.

- [9] N. I. Mourad, M. Nenquin and J.-C. Henquin, "Metabolic amplifying pathway increases both phases of insulin secretion independently of Beta-cell actin microfilaments," *American Journal of Physiology-Cell Physiology*, vol. 299, no. 2, pp. C389-C398, 2010.
- [10] J. C. Henquin, "Regulation of insulin secretion: a matter of phase control and amplitude modulation," *Diabetologia*, vol. 52, pp. 739-751, 2009.
- [11] J. E. Campbell and C. B. Newgard, "Mechanisms controlling pancreatic islet cell function in insulin secretion," *Nature Reviews Molecular Cell Biology*, vol. 22, p. 142–158, 2021.
- [12] A. I. Tarasov, F. Semplici, M. A. Ravier, E. A. Bellomo, T. J. Pullen, P. Gilon, I. Sekler, R. Rizzuto and G. A. Rutter, "The Mitochondrial Ca²⁺ Uniporter MCU Is Essential for Glucose-Induced ATP Increases in Pancreatic β -Cells," *Plos One*, vol. 7, no. 7, pp. 1-12, 2012.
- [13] X. Quan, T. T. Nguyen, S.-K. Choi, S. Xu, R. Das, S.-K. Cha, N. Kim, J. Han, A. Wiederkehr, C. B. Wollheim and K.-S. Park, "Essential Role of Mitochondrial Ca²⁺ Uniporter in the Generation of Mitochondrial pH Gradient and Metabolism-Secretion Coupling in Insulin-releasing Cells," *The Journal of Biological Chemistry*, vol. 290, no. 7, p. 4086–4096, 2015.
- [14] S. L. Lewandowski, R. L. Cardone, H. R. Foster, T. Ho, E. Potapenko, C. P. H. R. VanDeusen, S. M. Sdao, T. C. Alves, X. Zhao, M. E. Capozzi, A. H. d. Souza, I. Jahan, C. J. Thomas and C. S. N, "Pyruvate kinase controls signal strength in the insulin secretory pathway," *Cell Metabolism*, vol. 32, no. 5, pp. 735-750, 2020.
- [15] H. I. C. B. C. B. Peter A. Antinozzi, "Mitochondrial Metabolism Sets the Maximal Limit of Fuel-stimulated Insulin Secretion in a Model Pancreatic Beta Cell: A SURVEY OF FOUR FUEL SECRETAGOGUES," *JBC*, 2002.
- [16] E. Haythorne, M. Rohm, M. v. d. Bunt, M. F. Brereton, A. I. Tarasov, T. S. Blacker, G. Sachse, R. T. Exposito, S. Davis, O. Baba, R. Fischer, M. R. Duchon, P. Rorsman, J. I. MacRae and F. M. A, "Diabetes causes marked inhibition of mitochondrial metabolism in pancreatic β -cells," *Nature Communications*, vol. 10, no. 2474, pp. 1-17, 2019.
- [17] Y. H. J. O. V. e. a. Will, "Analysis of mitochondrial function using phosphorescent oxygen-sensitive probes," *Nat Protoc*, 2006.
- [18] G. H. Patterson, S. M. Knobel, P. Arkhammar, O. Thastrup and D. W. Piston, "Separation of the glucose-stimulated cytoplasmic and mitochondrial NAD(P)H responses in pancreatic islet beta cells," *PNAS*, vol. 97, no. 10, p. 5203–5207, 2000.
- [19] E. P. Taddeo, L. Stiles, S. Sereda, E. Ritou, D. M. Wolf, M. Abdullah, Z. Swanson, J. Wilhelm, M. Bellin, P. McDonald, K. Caradonna, A. Neilson, M. Liesa and O. S. Shirihai, "Individual islet respirometry reveals functional diversity within the islet population of mice

- and human donors," *Molecular Metabolism*, vol. 16, pp. 150-159, 2018.
- [20] A. Nászai, E. Terhes, J. Kaszaki, M. Boros and L. Juhász, "Ca(2+)N It Be Measured? Detection of Extramitochondrial Calcium Movement With High-Resolution FluoRespirometry," *Scientific Reports*, vol. 9, no. 19229, pp. 1-13, 2019.
- [21] K. S. Sankar, B. J. Green, A. R. Crocker, J. E. Verity, S. M. Altamentova and J. V. Rocheleau, "Culturing Pancreatic Islets in Microfluidic Flow Enhances Morphology of the Associated Endothelial Cells," *PLoS One*, vol. 6, no. 9, pp. 1-9, 2011.
- [22] P. N. Silva, B. J. Green, S. M. Altamentovab and J. V. Rocheleau, "A microfluidic device designed to induce media flow throughout pancreatic islets while limiting shear-induced damage," *Lab on a Chip*, vol. 13, no. 4374, p. 4374–4384, 2013.
- [23] P. N. Silva, Z. Atto, R. Regeenes, U. Tufa, Y. Y. Chen, W. C. W. Chan, A. Volchuk, D. M. Kilkenny and J. V. Rocheleau, "Highly efficient adenoviral transduction of pancreatic islets using a microfluidic device," *Lab on a Chip*, vol. 16, no. 15, pp. 2921-2934, 2016.
- [24] P. C. Thomas, M. Halter, A. Tona, S. R. Raghavan, A. L. Plant and S. P. Forry, "A Noninvasive Thin Film Sensor for Monitoring Oxygen Tension during in Vitro Cell Culture," *Analytical Chemistry*, vol. 81, no. 22, p. 9239–9246, 2009.
- [25] P. Buchwald, "FEM-based oxygen consumption and cell viability models for avascular pancreatic islets," *Theoretical Biology and Medical Modelling*, vol. 6, no. 5, pp. 1-13, 2009.
- [26] A. D. Edelstein, M. A. Tsuchida, N. Amodaj, H. Pinkard, R. D. Vale and N. Stuurman, "Advanced methods of microscope control using µManager software," *Journal of Biological Methods*, vol. 1(2), pp. 1-10, 2014.
- [27] J. E. Paz-Miguel, R. Flores, P. Sánchez-Velasco, G. Ocejó-Vinyals, J. E. d. Diego, J. L. d. Rego and F. Leyva-Cobián, "Reactive Oxygen Intermediates During Programmed Cell Death Induced in the Thymus of the Ts(1716)65Dn Mouse, a Murine Model for Human Down's Syndrome," *The Journal of Immunology*, vol. 163, no. 10, pp. 5399-5410, 1999.
- [28] J. M. Vanderkooi, G. Maniara, T. J. Green and D. F. Wilson, "An Optical Method for Measurement of Dioxygen Concentration Based upon Quenching of Phosphorescence," *The Journal of Biological Chemistry*, vol. 262, no. 12, pp. 5476-5482, 1987.
- [29] Y. Mao, Y. Gao, S. Wu, S. Wua, J. Shi, B. Zhou and Y. Tian, "Highly enhanced sensitivity of optical oxygen sensors using microstructured PtTFPP/PDMS-pillar arrays sensing layer," *Sensors and Actuators B: Chemical*, vol. 251, pp. 495-502, 2017.
- [30] J. D. Wikstrom, S. B. Sereda, L. Stiles, A. Elorza, E. M. Allister, A. Neilson, D. A. Ferrick, M. B. Wheeler and O. S. Shirihai, "A Novel High-Throughput Assay for Islet Respiration Reveals Uncoupling of Rodent and Human Islets," *Plos One*, vol. 7, no. 5, pp. 1-7, 2012.

- [31] S.-K. Jung, L. M. Kauri, W.-J. Qian and R. T. Kennedy, "Correlated Oscillations in Glucose Consumption, Oxygen Consumption, and Intracellular Free Ca²⁺ in Single Islets of Langerhans," *Journal of Biological Chemistry*, vol. 275, no. 9, pp. 6642-3320, 2000.
- [32] K. Ashizawa, M. C. Willingham, C.-M. Liangs and S.-y. Cheng, "In Vivo Regulation of Monomer-Tetramer Conversion of Pyruvate Kinase Subtype M2 by Glucose Is Mediated via Fructose 1,6-Bisphosphate," *The Journal of Biological Chemistry*, vol. 266, no. 25, pp. 16842-16846, 1991.
- [33] S.-K. Lee and I. Okura, "Photostable Optical Oxygen Sensing Material: Platinum Tetrakis(pentafluorophenyl)porphyrin Immobilized in Polystyrene," *Analytical Communications*, vol. 34, pp. 185-188, 1994.
- [34] S. Ashraf, I. Niskanen, B. Kanyathare, E. Vartiainen, C. Mattsson, R. Keikkilä and G. Thungstrom, "Determination of complex refractive index of SU-8 by Kramers–Kronig dispersion relation method at the wavelength range 2.5–22.0 μm," *Journal of Quantitative Spectroscopy and Radiative Transfer*, vol. 224, pp. 309-311, 2019.
- [35] N. Burshtein, S. ToChan, K. Toda-Peters, A. Q. Shen and S. J. Haward, "3D-printed glass microfluidics for fluid dynamics and rheology," *Current Opinion in Colloid & Interface Science*, vol. 43, pp. 1-14, 2019.
- [36] G. A. Martens, L. Jiang, K. Verhaeghen, J. B. Connolly, S. G. Geromanos, G. Stangé, L. V. Oudenhove, B. Devreese, K. H. Hellemans, Z. Ling, C. V. Schravendijk, D. G. Pipeleers and J. P. C. Vissers, "Protein Markers for Insulin-Producing Beta Cells with Higher Glucose Sensitivity," *Plos One*, vol. 5, no. 12, pp. 1-13, 2010.
- [37] G. Prakasam, M. A. Iqbal, R. N. K. Bamezai and S. Mazurek, "Posttranslational Modifications of Pyruvate Kinase M2: Tweaks that Benefit Cancer," *Frontiers in Oncology*, vol. 8, pp. 1-12, 2018.
- [38] P. B. d. Andrade, M. Casimir and P. Maechler, "Mitochondrial activation and the pyruvate paradox in a human cell line," *FEBS Letters*, vol. 578, no. 3, pp. 224-228, 2004.
- [39] J. V. Rocheleau, W. S. Head, W. E. Nicholson, A. C. Powers and D. W. Piston, "Pancreatic Islet β-Cells Transiently Metabolize Pyruvate," *Journal of Biological Chemistry*, vol. 277, no. 34, p. 30914–30920, 2002.
- [40] A. Wiederkehr, G. Szanda, D. Akhmedov, C. Matak, C. W. Heizmann, K. Schoonjans, T. Pozzan, A. Spat and C. B. Wollheim, "Mitochondrial Matrix Calcium Is an Activating Signal for Hormone Secretion," *Cell Metabolism*, vol. 13, no. 5, pp. 601-11, 2011.
- [41] a. R. Sweet and M. Gilbert, "Contribution of Calcium Influx in Mediating Glucose-Stimulated Oxygen Consumption in Pancreatic Islets," *Diabetes*, vol. 55, no. 12, pp. 3509-3519, 2006.

- [42] S. Marchi and P. Pinton, "The mitochondrial calcium uniporter complex: molecular components, structure and physiopathological implications," *The Journal of Physiology*, vol. 592, no. 5, p. 829–839, 2013.
- [43] L. Sanford and A. Palmer, "Recent Advances in Development of Genetically Encoded Fluorescent Sensors," *Methods Enzymology*, vol. 589, pp. 1-49, 2017.
- [44] G. D. Gutierrez, J. Gromada and L. Sussel, "Heterogeneity of the Pancreatic Beta Cell," *Frontiers in Genetics*, vol. 8, pp. 1-9, 2017.
- [45] N. Rafiei, A. Au, M. G. Moghadam, R. Regeenes, Y. Wang, C. M. Yip, H. Gaisano and J. V. Rocheleau, "Design of a Versatile Microfluidic Device for Imaging Precision-Cut-Tissue Slices," *bioRxiv*, pp. 1-19, 2022.
- [46] K. K. Papas, M. D. Bellin, D. E. R. Sutherland, T. M. Suszynski, J. P. Kitzmann, E. S. Avgoustiniatos and A. C. Gruessner, "Islet Oxygen Consumption Rate (OCR) Dose Predicts Insulin Independence in Clinical Islet Autotransplantation," *Plos One*, vol. 10, no. 8, pp. 1-11, 2015.
- [47] A. M. J. Shapiro, C. Ricordi, B. J. Herin, H. Auchincloss, R. Lindblad, R. P. Robertson, A. Secchi, M. D. Brendel, T. Berney, D. C. Brennan, E. Cagliero, R. Alejandro, E. A. Ryan and B. DiMercurio, "International trial of the Edmonton protocol for islet transplantation," *The New England Journal of Medicine*, vol. 355, pp. 1318-1330, 2006.

Figure Captions

Figure 1

COMSOL simulations reveal greater sensitivity when patterning O₂ sensor in microwells.

A) A schematic of various entry points for external O₂ into the microfluidic device. The planar sensor made from PDMS allows O₂ from the environment to enter and leave lowering sensor sensitivity directly beneath the islet. Black arrows indicate regions where O₂ can enter and leave.

B) O₂ transport in the device in the absence (*left*) and presence (*right*) of an O₂ consuming islet with the PtTFPP/PDMS in a uniform plane underneath. White dotted line represents distance across islet (μm).

C) Simulated O₂ concentrations across the PtTFPP/PDMS in the presence and absence of an islet reveal very small response across the planar sensor.

D) A schematic with O₂ entry points sealed by patterning SU-8. This arrangement helps digitize the sensor readout and prevent O₂ from travelling between wells.

E) O₂ concentration distributions in 10 (*right*), 50 (*middle*) and 100 (*left*) μm wells with an islet above. 10 μm sized wells show greater sensitivity to O₂ compared to 50 and 100 μm sized well designs.

F) Simulated O₂ concentrations in 10, 50 and 100 μm wells underneath the islet reveal a dramatic increase in sensitivity with 10 μm wells.

G) O₂ concentration distributions in 10 (*bottom*), 30 (*middle*) and 60 (*top*) μm height wells reveal no changes in sensor response.

H) Simulated O₂ concentrations in 10, 30 and 60 μm height wells reveal no change in OCR sensitivity (*right*).

Figure 2

Determining the impact of flow rate and islet proximity on O₂ sensitivity.

A) Varying flow rates have minimal effect on O₂ levels.

B) Islet proximity is measured as distance between the bottom of the islet to the top of the microwell plane.

C) Simulated O₂ concentrations reveal proximity to enhance microwell sensitivity.

D) Glucose-stimulated O_2 concentrations in the wells when islets are subjected to 2- and 20-mM glucose reveal greater OCR in the center of the islet.

E) Simulated O_2 concentrations when islets are subjected to 2- and 20-mM glucose show a large response to an increase in OCR at 20 mM glucose.

Figure 3

Characterization of the O_2 sensor.

A) O_2 sensitive microwells in the presence of ddH₂O (*left*) and 10% sodium sulfite (*right*). Scale bar represents 10 μ m.

B) The relationship between O_2 concentration and normalized fluorescence modelled using the Stern-Volmer relationship. A representative two-point calibration curve to calculate changes in O_2 . Data in figure are means $\pm$ SD for 10 independent repeats.

C) A temporal trace of switching from a 0% to 100% air saturated solution revealing a t_{95} of 12.4 ± 0.69 s for 6 independent repeats. Red dotted lines represent t_{95} interval.

D) A temporal trace of switching from a 100% to 0% air saturated solution revealing a t_{95} of 4.24 ± 0.18 s for 6 independent repeats. Red dotted lines represent t_{95} interval.

Data in figure are means $\pm$ SEM unless otherwise stated.

Figure 4

An on-chip O_2 sensor to measure OCR in individual pancreatic islets.

A) Representative widefield fluorescence images of O_2 sensitive microwells underneath islet while flowing the indicated respiratory inhibitors. Scale bars represent 10 μ m.

B) Representative widefield white light images of different sized islets trapped in device. Scale bars represent 10 μ m.

C) Basal OCR of individual mouse islets relative to islet volume.

D) Averaged OCR traces of mouse islets exposed to 20 mM glucose, 5 μ M oligomycin, 2 μ M FCCP and 5 μ M rotenone/antimycin A.

E) Islet bioenergetic measurements extracted from the respiratory inhibitor treatments. Glucose stimulated OCR, proton leak and maximal respiration relative to basal OCR.

Data in figure are means $\pm$ SEM. **** indicates $p \leq 0.0001$ and *** indicates $p \leq 0.001$ by one-way ANOVA. N = 14 islets.

Figure 5

Dual-colour imaging of OCR and Ca^{2+} in individual islets reveals a two-step increases in OCR response.

A) Representative images of a Cal-520 stained mouse islet and O_2 sensitive microwells underneath the islet after stimulating with 2- (left) and 20-mM glucose (right) show a visible increase in fluorescence in both sensors. Scale bars represent 10 μm .

B) A representative OCR and Ca^{2+} traces of an individual mouse islet show OCR to precede Ca^{2+} influx. The black arrow indicates treatment change from 2- to 20-mM glucose. The Ca^{2+} response was normalized to the Cal-520 fluorescence intensity during low glucose treatment.

C-D) Averaged area under the curve (AUC) of Ca^{2+} (C) and OCR (D) traces during low glucose treatment, phase 0 and 1 Ca^{2+} responses. Analyzed areas are indicated by the blue, red and green lines in B).

E-F) Average normalized changes in OCR determined by setting “time = 0 min” based on the start of 20 mM glucose stimulation (E) and the beginning of phase 1 Ca^{2+} response (F). Changes in OCR were normalized to the six time points prior to $t = 0$ min. Cyan line in F) indicates the fitting for the start of the second OCR rise.

G) Averaged slopes of the two-step OCR responses to high glucose stimulation.

Data in figure are means $\pm$ SEM. **** indicates $p \leq 0.0001$ and *** indicates $p \leq 0.001$ by one-way ANOVA. $N = 7$ islets.

Figure 6

Glycolytic flux and pyruvate entry into mitochondria control OCR.

A) Simplified metabolic model depicting the impact of glycolysis derived NADH and PKM2 activity on OCR. NADH from glycolysis is shuttled into the mitochondria for charging the electron transport chain. PKM2 controls pyruvate flux into the mitochondria for tricarboxylic acid (TCA cycle) metabolism which also generates NADH to further fuel the electron transport chain. The substrate (NADH) supply establishes the proton gradient stimulating OCR and ATP production. In excess substrate supply, OCR will be inhibited by a rise in the ATP/ADP ratio.

B) Normalized basal OCR rates in the presence of 2DG, a glycolysis inhibitor, and UK5099, a mitochondrial pyruvate carrier inhibitor and 20 mM glucose. Significant changes relative to 2 mM glucose are seen with UK5099 and 20 mM glucose treatment.

C) Normalized changes in OCR of islets in 20 mM glucose in response to 20 mM + 2DG reveals substrate supply is needed to stimulate OCR.

D) Averaged OCR from 3-minute intervals during steady state of each treatment reveals 2DG to significantly decrease OCR back down to basal levels in the presence of 20 mM glucose.

E) Preincubating islets with PKM2 activator (4 μ M TEPP-46) shows a significant increase in basal OCR levels consistent with greater pyruvate flux.

F) Islets preincubated in 4 μ M TEPP-46 respond to 20 mM glucose with a rapid drop in OCR followed by a linear increase in OCR.

G) A representative OCR and Ca^{2+} trace of an individual islet TEPP-46 treated shows OCR precedes Ca^{2+} -influx.

H) Individual islet traces were normalized to the onset of 20 mM glucose stimulation to reveal a rapid drop followed by linear increase in OCR prior to Ca^{2+} -influx.

Data in figure are means $\pm$ SEM. *** indicates $p \leq 0.001$, ** indicates $p \leq 0.01$, and * indicates $p \leq 0.05$ by one-way ANOVA. N = 5-12 islets.

Figure 7

A model depicting the biphasic behavior of OCR during first phase insulin secretion.

A) Upon glucose stimulation, glycolytic NADH is shuttled into the mitochondria for OxPhos, stimulating the first rise in OCR. A plateau in the OCR response follows shortly due to inactivated PKM2 (in dimer form).

B) As glycolytic metabolism continues, the accumulation of upstream glycolytic metabolite, FBP, allosterically activates PKM2 (dimer to tetramer transition) causing a large pyruvate flux into the mitochondria. This pyruvate entry further stimulates TCA cycle metabolism to generate more NADH for OxPhos stimulating the second rise in OCR.

Fig 1

Fig 2

A

B

C

D

Fig 3

Fig 4

Fig 5

Fig 6

Fig 7